

January 05, 2011

Mr. Cesar Rodriguez
Sarasota County
1255 T. Mabry Carlton Parkway
Resource Management
Venice, FL 34293

RE: Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Dear Mr. Rodriguez:

Enclosed are the analytical results for sample(s) received by the laboratory between October 04, 2010 and October 29, 2010. The results relate only to the samples included in this report. Results reported herein conform to the most current NELAC standards, where applicable, unless otherwise narrated in the body of the report.

Analysis performed by Benchmark, E84167, identified on the COC and report attached.

Revised report with updated additional analytes requested.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Joe Vondrick

joe.vondrick@pacelabs.com
Project Manager

Enclosures

cc: Mr. Frank DeSteno, Sarasota County
Finance Dept., Sarasota County

REPORT OF LABORATORY ANALYSIS

Page 1 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



CERTIFICATIONS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Ormond Beach Certification IDs

8 East Tower Circle, Ormond Beach, FL 32174
Alabama Certification #: 41320
Arizona Certification #: AZ0735
Colorado Certification: FL NELAC Reciprocity
Connecticut Certification #: PH 0216
Florida Certification #: E83079
Georgia Certification #: 955
Guam Certification: FL NELAC Reciprocity
Hawaii Certification: FL NELAC Reciprocity
Kansas Certification #: E-10383
Kentucky Certification #: 90050
Louisiana Certification #: LA090012
Louisiana Environmental Certificate #: 05007
Maine Certification #: FL1264
Massachusetts Certification #: M-FL1264

Michigan Certification #: 9911
Mississippi Certification: FL NELAC Reciprocity
Montana Certification #: Cert 0074
Nevada Certification: FL NELAC Reciprocity
New Hampshire Certification #: 2958
New Jersey Certification #: FL765
New York Certification #: 11608
North Carolina Environmental Certificate #: 667
North Carolina Certification #: 12710
Pennsylvania Certification #: 68-547
Puerto Rico Certification #: FL01264
Tennessee Certification #: TN02974
Texas Certification: FL NELAC Reciprocity
Virginia Certification #: 00432
Wyoming Certification: FL NELAC Reciprocity

Green Bay Certification IDs

1241 Bellevue Street, Green Bay, WI 54302
California Certification #: 09268CA
Florida/NELAP Certification #: E87948
Illinois Certification #: 200050
Kentucky Certification #: 82
Louisiana Certification #: 04168
Minnesota Certification #: 055-999-334
New York Certification #: 11888

New York Certification #: 11888
North Carolina Certification #: 503
North Dakota Certification #: R-150
South Carolina Certification #: 83006001
US Dept of Agriculture #: S-76505
Wisconsin Certification #: 405132750
Wisconsin DATCP Certification #: 105-444

REPORT OF LABORATORY ANALYSIS

Page 2 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



SAMPLE SUMMARY

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Matrix	Date Collected	Date Received
3519325001	B-4R	Water	09/30/10 13:00	10/04/10 07:00
3519325002	B-4R DUP	Water	09/30/10 13:00	10/04/10 07:00
3519325003	B-4R Hg Blank	Water	09/30/10 13:00	10/04/10 07:00
3519325004	Trip Blank (9/30/10)	Water	09/30/10 08:00	10/04/10 07:00
3519325005	B-4R re	Water	10/05/10 14:45	10/06/10 10:45
3519325006	CW-19	Water	10/13/10 13:02	10/16/10 10:20
3519325007	CW-20	Water	10/13/10 09:46	10/16/10 10:20
3519325008	CW-19 DUP	Water	10/13/10 13:02	10/16/10 10:20
3519325009	Equip blank (10/13/10)	Water	10/13/10 10:40	10/16/10 10:20
3519325010	Trip blank (10/13/10)	Water	10/13/10 08:00	10/16/10 10:20
3519325011	CW-16	Water	10/13/10 16:17	10/19/10 07:00
3519325012	MW-15	Water	10/14/10 11:00	10/19/10 07:00
3519325013	MW-16	Water	10/15/10 11:55	10/19/10 07:00
3519325014	CW-15	Water	10/15/10 14:45	10/19/10 07:00
3519325015	MW-17	Water	10/15/10 13:36	10/19/10 07:00
3519325016	MW-9	Water	10/14/10 14:43	10/19/10 07:00
3519325017	CW-8A	Water	10/14/10 12:41	10/19/10 07:00
3519325018	Equip Blank (10/14/10)	Water	10/14/10 09:25	10/19/10 07:00
3519325019	MW-9 DUP	Water	10/14/10 14:43	10/19/10 07:00
3519325020	MW-8A	Water	10/14/10 10:13	10/19/10 07:00
3519325021	Trip blank appdx 2 (10/14/10)	Water	10/14/10 08:00	10/19/10 07:00
3519325022	Trip blank appdx 1 (10/14/10)	Water	10/14/10 08:00	10/19/10 07:00
3519325023	CW-9	Water	10/18/10 12:16	10/21/10 07:00
3519325024	CW-10R	Water	10/18/10 14:00	10/21/10 07:00
3519325025	MW-18	Water	10/18/10 09:23	10/21/10 07:00
3519325026	MW-19	Water	10/18/10 11:40	10/21/10 07:00
3519325027	MW-20	Water	10/18/10 13:45	10/21/10 07:00
3519325028	MW-1R	Water	10/18/10 10:35	10/21/10 07:00
3519325029	MW-10R	Water	10/18/10 14:58	10/21/10 07:00
3519325030	Trip Blank appdx 1 10-18	Water	10/18/10 14:58	10/21/10 07:00
3519325031	Trip Blank appdx 2 10-18	Water	10/18/10 14:58	10/21/10 07:00
3519325032	C-1	Water	10/27/10 09:20	10/29/10 07:00
3519325033	C-2	Water	10/27/10 09:50	10/29/10 07:00
3519325034	C-3	Water	10/27/10 10:25	10/29/10 07:00
3519325035	C-4	Water	10/27/10 12:30	10/29/10 07:00
3519325036	C-5	Water	10/27/10 13:00	10/29/10 07:00
3519325037	P2-1	Water	10/27/10 13:30	10/29/10 07:00

REPORT OF LABORATORY ANALYSIS

Page 3 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



SAMPLE SUMMARY

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Matrix	Date Collected	Date Received
3519325038	Trip Blank APPH (10/27/10)	Water	10/27/10 08:00	10/29/10 07:00
3519325039	Gas Condensate	Water	10/27/10 10:45	10/29/10 07:00

REPORT OF LABORATORY ANALYSIS

Page 4 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325001	B-4R	EPA 1631E	GMW	1	PASI-G
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	13	PASI-O
		EPA 6020	DRS	8	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 2540D	TP1	1	PASI-O
		SM10200	KHC	1	PASI-O
		TKN+NOx Calculation	AMD	1	PASI-O
		EPA 300.0	TLK	1	PASI-O
		EPA 350.1	AMD	2	PASI-O
		EPA 351.2	AMD	1	PASI-O
		EPA 353.2	HEM	1	PASI-O
		EPA 365.4	AMD	1	PASI-O
		EPA 410.4	MMD	1	PASI-O
		SM 5310B	HEM	1	PASI-O
3519325002	B-4R DUP	EPA 1631E	GMW	1	PASI-G
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	13	PASI-O
		EPA 6020	DRS	8	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 2540D	TP1	1	PASI-O
		SM10200	KHC	1	PASI-O
		TKN+NOx Calculation	AMD	1	PASI-O
		EPA 300.0	TLK	1	PASI-O
		EPA 350.1	AMD	2	PASI-O
		EPA 351.2	AMD	1	PASI-O
		EPA 353.2	HEM	1	PASI-O
		EPA 365.4	AMD	1	PASI-O
3519325003	B-4R Hg Blank	EPA 1631E	GMW	1	PASI-G
		EPA 8260	JBH	49	PASI-O
			JJV	5	PASI-O
3519325004	Trip Blank (9/30/10)				
3519325006	CW-19				

REPORT OF LABORATORY ANALYSIS

Page 5 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325007	CW-20	EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	WFH	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	17	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	TP1	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
3519325008	CW-19 DUP	EPA 8141	WFH	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	TP1	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 6 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325009	Equip blank (10/13/10)	EPA 8141	WFH	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	17	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	TP1	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	WFH	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	TP1	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325010	Trip blank (10/13/10)	EPA 8260	JBH	62	PASI-O
3519325011	CW-16		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	AE1	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	16	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 7 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325012	MW-15	EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	AE1	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	20	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
3519325013	MW-16	EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	AE1	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	19	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 8 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325014	CW-15	EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	AE1	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325015	MW-17		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325016	MW-9		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 9 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325017	CW-8A	EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 6010	TAP	2	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325018	Equip Blank (10/14/10)	EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325019	MW-9 DUP	EPA 350.1	AMD	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
3519325020	MW-8A	EPA 350.1	AMD	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 10 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
		EPA 300.0	HEM	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325021	Trip blank appdx 2 (10/14/10)	EPA 8260	JBH	62	PASI-O
3519325022	Trip blank appdx 1 (10/14/10)	EPA 8260	JBH	49	PASI-O
3519325023	CW-9	EPA 6010	TAP	2	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325024	CW-10R	EPA 6010	TAP	2	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325025	MW-18		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	TLK	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325026	MW-19		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	20	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	WFH	108	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	TLK	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 11 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325027	MW-20		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	19	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	WFH	108	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	TLK	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325028	MW-1R		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	JLY	6	PASI-O
		EPA 6010	TAP	19	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	DRS	1	PASI-O
		EPA 8270	WFH	108	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	TLK	2	PASI-O
		EPA 335.4	TLK	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325029	MW-10R		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 6010	TAP	18	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 12 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK	1	PASI-O
		EPA 8260	JBH	49	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	TLK	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325030	Trip Blank appdx 1 10-18	EPA 8260	JBH	49	PASI-O
3519325031	Trip Blank appdx 2 10-18	EPA 8260	JBH	62	PASI-O
3519325032	C-1		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325033	C-2		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 13 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325034	C-3	EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
3519325035	C-4	EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 14 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325036	C-5	EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8141	EAO	7	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
3519325037	P2-1		JJV	5	PASI-O
		EPA 8011	JLR	2	PASI-O
		EPA 8081	AE1	23	PASI-O
		EPA 8082	AE1	9	PASI-O
		EPA 8151	LJM	6	PASI-O
		EPA 6010	TAP	16	PASI-O
		EPA 6020	DRS	2	PASI-O
		EPA 7470	SK1	1	PASI-O
		EPA 8270	EAO	111	PASI-O
		EPA 8270 by SCAN	WFH	20	PASI-O
		EPA 8260	JBH	62	PASI-O
		SM 2540C	KDM	1	PASI-O
		SM 4500-S2E	AMD	1	PASI-O
		EPA 300.0	HEM	1	PASI-O
		EPA 335.4	HEM	1	PASI-O
		EPA 350.1	AMD	1	PASI-O
			JBH	62	PASI-O
3519325038	Trip Blank APPII (10/27/10)	EPA 8260	JBH	62	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 15 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

SAMPLE ANALYTE COUNT

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
3519325039	Gas Condensate	EPA 6010	TAP	20	PASI-O
		EPA 7470	SK1	1	PASI-O
		SM 2320B	AMD	3	PASI-O
		SM 2540C	KDM	1	PASI-O
		EPA 300.0	HEM	2	PASI-O
		EPA 350.1	AMD	1	PASI-O
		EPA 410.4	MMD	1	PASI-O

REPORT OF LABORATORY ANALYSIS

Page 16 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 1631E
Description: 1631E Mercury, Low Level
Client: Sarasota County
Date: January 05, 2011

General Information:

3 samples were analyzed for EPA 1631E. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Analyte Comments:

QC Batch: CVFS/2230

5p: Sample was received with headspace.

- B-4R DUP (Lab ID: 3519325002)
- Mercury

REPORT OF LABORATORY ANALYSIS

Page 17 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method:

Description: Field Data

Client: Sarasota County

Date: January 05, 2011

General Information:

22 samples were analyzed for . All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 18 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8011
Description: 8011 GCS EDB and DBCP
Client: Sarasota County
Date: January 05, 2011

General Information:

26 samples were analyzed for EPA 8011. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 8011 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: OEXT/3262

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325002

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 126472)
- 1,2-Dibromoethane (EDB)

QC Batch: OEXT/3459

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 10142001007

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MSD (Lab ID: 137449)
- 1,2-Dibromo-3-chloropropane

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 19 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8011

Description: 8011 GCS EDB and DBCP

Client: Sarasota County

Date: January 05, 2011

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 20 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8081
Description: 8081 GCS Pesticides
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 8081. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3510 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

QC Batch: OEXT/3326

J(S0): Estimated Value. Surrogate recovery outside laboratory control limits.

- LCS (Lab ID: 130318)
- Decachlorobiphenyl (S)

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- CW-15 (Lab ID: 3519325014)
- Decachlorobiphenyl (S)
- MW-15 (Lab ID: 3519325012)
- Decachlorobiphenyl (S)
- MW-16 (Lab ID: 3519325013)
- Tetrachloro-m-xylene (S)

S3: Surrogate recovery exceeded laboratory control limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- BLANK (Lab ID: 130317)
- Decachlorobiphenyl (S)

QC Batch: OEXT/3431

J(S2): Estimated Value. Surrogate recovery outside laboratory control limits due to matrix interferences (confirmed by similar results from sample re-analysis).

- C-1 (Lab ID: 3519325032)
- Decachlorobiphenyl (S)
- MS (Lab ID: 136032)
- Decachlorobiphenyl (S)

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-2 (Lab ID: 3519325033)
- Decachlorobiphenyl (S)

REPORT OF LABORATORY ANALYSIS

Page 21 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8081
Description: 8081 GCS Pesticides
Client: Sarasota County
Date: January 05, 2011

QC Batch: OEXT/3431

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-3 (Lab ID: 3519325034)
 - Decachlorobiphenyl (S)
- C-4 (Lab ID: 3519325035)
 - Decachlorobiphenyl (S)
 - Tetrachloro-m-xylene (S)
- C-5 (Lab ID: 3519325036)
 - Decachlorobiphenyl (S)
 - Tetrachloro-m-xylene (S)

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

QC Batch: OEXT/3369

L3: Analyte recovery in the laboratory control sample (LCS) exceeded QC limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- LCS (Lab ID: 133077)
 - 4,4'-DDT
 - Methoxychlor

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: OEXT/3431

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325032

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 136032)
 - 4,4'-DDD
 - 4,4'-DDE
 - Dieldrin
 - Endosulfan I
 - Endosulfan sulfate
 - Endrin
 - Endrin aldehyde
- MSD (Lab ID: 136033)
 - Endrin aldehyde

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8081

Description: 8081 GCS Pesticides

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: OEXT/3326

1p: A heavy emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.

- CW-15 (Lab ID: 3519325014)
 - Decachlorobiphenyl (S)
- MW-15 (Lab ID: 3519325012)
 - Decachlorobiphenyl (S)
- MW-16 (Lab ID: 3519325013)
 - Tetrachloro-m-xylene (S)

QC Batch: OEXT/3431

2p: An emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.

- C-1 (Lab ID: 3519325032)
 - Decachlorobiphenyl (S)
- C-2 (Lab ID: 3519325033)
 - Decachlorobiphenyl (S)
- C-3 (Lab ID: 3519325034)
 - Decachlorobiphenyl (S)
- C-4 (Lab ID: 3519325035)
 - Decachlorobiphenyl (S)
 - Tetrachloro-m-xylene (S)
- C-5 (Lab ID: 3519325036)
 - Decachlorobiphenyl (S)
 - Tetrachloro-m-xylene (S)
- MS (Lab ID: 136032)
 - Decachlorobiphenyl (S)

REPORT OF LABORATORY ANALYSIS

Page 23 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8082
Description: 8082 GCS PCB
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 8082. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3510 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

QC Batch: OEXT/3327

J(S0): Estimated Value. Surrogate recovery outside laboratory control limits.

- Equip blank (10/13/10) (Lab ID: 3519325009)
- Decachlorobiphenyl (S)

J(S2): Estimated Value. Surrogate recovery outside laboratory control limits due to matrix interferences (confirmed by similar results from sample re-analysis).

- MS (Lab ID: 130323)
 - Decachlorobiphenyl (S)
- MSD (Lab ID: 130324)
 - Decachlorobiphenyl (S)

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- CW-15 (Lab ID: 3519325014)
 - Decachlorobiphenyl (S)
- MW-15 (Lab ID: 3519325012)
 - Decachlorobiphenyl (S)
- MW-16 (Lab ID: 3519325013)
 - Decachlorobiphenyl (S)

QC Batch: OEXT/3370

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- MW-20 (Lab ID: 3519325027)
 - Decachlorobiphenyl (S)

QC Batch: OEXT/3432

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-1 (Lab ID: 3519325032)
 - Decachlorobiphenyl (S)

REPORT OF LABORATORY ANALYSIS

Page 24 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8082

Description: 8082 GCS PCB

Client: Sarasota County

Date: January 05, 2011

QC Batch: OEXT/3432

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-2 (Lab ID: 3519325033)
 - Decachlorobiphenyl (S)
- C-3 (Lab ID: 3519325034)
 - Decachlorobiphenyl (S)
- C-4 (Lab ID: 3519325035)
 - Decachlorobiphenyl (S)
- C-5 (Lab ID: 3519325036)
 - Decachlorobiphenyl (S)

S3: Surrogate recovery exceeded laboratory control limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- BLANK (Lab ID: 136034)
 - Decachlorobiphenyl (S)
 - Tetrachloro-m-xylene (S)
- P2-1 (Lab ID: 3519325037)
 - Tetrachloro-m-xylene (S)

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

QC Batch: OEXT/3370

L3: Analyte recovery in the laboratory control sample (LCS) exceeded QC limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- LCS (Lab ID: 133097)
 - PCB-1260 (Aroclor 1260)

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: OEXT/3327

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520442004

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130323)
 - PCB-1260 (Aroclor 1260)
- MSD (Lab ID: 130324)
 - PCB-1260 (Aroclor 1260)

REPORT OF LABORATORY ANALYSIS

Page 25 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8082

Description: 8082 GCS PCB

Client: Sarasota County

Date: January 05, 2011

QC Batch: OEXT/3432

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3518484037

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 136036)
 - PCB-1260 (Aroclor 1260)
- MSD (Lab ID: 136037)
 - PCB-1016 (Aroclor 1016)
 - PCB-1260 (Aroclor 1260)

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Analyte Comments:

QC Batch: OEXT/3327

1p: A heavy emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.

- CW-15 (Lab ID: 3519325014)
 - Decachlorobiphenyl (S)
- MW-15 (Lab ID: 3519325012)
 - Decachlorobiphenyl (S)
- MW-16 (Lab ID: 3519325013)
 - Decachlorobiphenyl (S)

QC Batch: OEXT/3370

7p: The continuing calibration for this compound is outside of method control limits for this compound (high/low bias). However an acceptable Reporting Limit standard was analyzed at the end of the sequence demonstrating appropriate instrument sensitivity.

- MW-20 (Lab ID: 3519325027)
 - Decachlorobiphenyl (S)

QC Batch: OEXT/3432

1p: A heavy emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.

- C-1 (Lab ID: 3519325032)
 - Decachlorobiphenyl (S)
- C-2 (Lab ID: 3519325033)
 - Decachlorobiphenyl (S)
- C-3 (Lab ID: 3519325034)
 - Decachlorobiphenyl (S)
- C-4 (Lab ID: 3519325035)
 - Decachlorobiphenyl (S)
- C-5 (Lab ID: 3519325036)
 - Decachlorobiphenyl (S)

REPORT OF LABORATORY ANALYSIS

Page 26 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8141
Description: 8141 GCS O/P Pesticides
Client: Sarasota County
Date: January 05, 2011

General Information:

16 samples were analyzed for EPA 8141. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3510 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

QC Batch: OEXT/3321

S3: Surrogate recovery exceeded laboratory control limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- CW-20 (Lab ID: 3519325007)
- 4-Chloro3nitrobenzotrifluoride

QC Batch: OEXT/3445

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-5 (Lab ID: 3519325036)
- 4-Chloro3nitrobenzotrifluoride

S3: Surrogate recovery exceeded laboratory control limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- C-1 (Lab ID: 3519325032)
- 4-Chloro3nitrobenzotrifluoride

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 27 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8141

Description: 8141 GCS O/P Pesticides

Client: Sarasota County

Date: January 05, 2011

QC Batch: OEXT/3338

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325014

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130927)
 - Phorate
- MSD (Lab ID: 130928)
 - Disulfoton

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Analyte Comments:

QC Batch: OEXT/3372

9p: The internal standard response associated with this result exceeds the upper control limit. However, the data is accepted based on surrogate compound recovery meeting control limits.

- LCS (Lab ID: 133137)
 - 4-Chloro3nitrobenzotrifluoride

QC Batch: OEXT/3445

9p: The internal standard response associated with this result exceeds the upper control limit. However, the data is accepted based on surrogate compound recovery meeting control limits.

- C-1 (Lab ID: 3519325032)
 - 4-Chloro3nitrobenzotrifluoride
- C-2 (Lab ID: 3519325033)
 - 4-Chloro3nitrobenzotrifluoride
- C-3 (Lab ID: 3519325034)
 - 4-Chloro3nitrobenzotrifluoride
- C-4 (Lab ID: 3519325035)
 - 4-Chloro3nitrobenzotrifluoride
- C-5 (Lab ID: 3519325036)
 - 4-Chloro3nitrobenzotrifluoride

REPORT OF LABORATORY ANALYSIS

Page 28 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8151
Description: 8151 Chlorinated Herbicides
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 8151. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 8151 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

QC Batch: OEXT/3366

J(S0): Estimated Value. Surrogate recovery outside laboratory control limits.

- MW-20 (Lab ID: 3519325027)
- 2,4-DCPA (S)

QC Batch: OEXT/3411

J(S2): Estimated Value. Surrogate recovery outside laboratory control limits due to matrix interferences (confirmed by similar results from sample re-analysis).

- C-3 (Lab ID: 3519325034)
- 2,4-DCPA (S)

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-4 (Lab ID: 3519325035)
- 2,4-DCPA (S)

QC Batch: OEXT/3440

J(S5): Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).

- C-2 (Lab ID: 3519325033)
- 2,4-DCPA (S)

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 29 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8151

Description: 8151 Chlorinated Herbicides

Client: Sarasota County

Date: January 05, 2011

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: OEXT/3411

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 9280763001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 135060)
 - Pentachlorophenol
- MSD (Lab ID: 135061)
 - Pentachlorophenol

QC Batch: OEXT/3440

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 9280813001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MSD (Lab ID: 136267)
 - Pentachlorophenol

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 30 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 6010
Description: 6010 MET ICP
Client: Sarasota County
Date: January 05, 2011

General Information:

30 samples were analyzed for EPA 6010. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3010 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

QC Batch: MPRP/3418

CH: The continuing calibration for this compound is outside of Pace Analytical acceptance limits. The results may be biased high.

- MW-15 (Lab ID: 3519325012)
- Manganese

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: MPRP/3418

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325007

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 131523)
- Calcium

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 31 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 6010

Description: 6010 MET ICP

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: MPRP/3418

D3: Sample was diluted due to the presence of high levels of non-target analytes or other matrix interference.

- MW-15 (Lab ID: 3519325012)

- Silver
- Barium
- Beryllium
- Cadmium
- Cobalt
- Chromium
- Copper
- Iron
- Manganese
- Sodium
- Nickel
- Lead
- Selenium
- Tin
- Vanadium
- Zinc

- MW-9 (Lab ID: 3519325016)

- Cobalt

QC Batch: MPRP/3530

D4: Sample was diluted due to the presence of high levels of target analytes.

- C-2 (Lab ID: 3519325033)

- Sodium

- C-3 (Lab ID: 3519325034)

- Sodium

- C-4 (Lab ID: 3519325035)

- Sodium

- C-5 (Lab ID: 3519325036)

- Sodium

REPORT OF LABORATORY ANALYSIS

Page 32 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 6020

Description: 6020 MET ICPMS

Client: Sarasota County

Date: January 05, 2011

General Information:

26 samples were analyzed for EPA 6020. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3010 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 33 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 7470
Description: 7470 Mercury
Client: Sarasota County
Date: January 05, 2011

General Information:

25 samples were analyzed for EPA 7470. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 7470 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: MERP/1583

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325016,3521204004

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 136740)
 - Mercury
- MSD (Lab ID: 136741)
 - Mercury

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 34 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8270
Description: 8270 MSSV SemiVOA App. II
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 8270. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3510 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

QC Batch: OEXT/3333

J(SS): Estimated Value. This analyte did not meet the secondary source verification criteria for the initial calibration. The reported result should be considered an estimated value.

- BLANK (Lab ID: 130562)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- CW-15 (Lab ID: 3519325014)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- CW-16 (Lab ID: 3519325011)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- CW-19 (Lab ID: 3519325006)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- CW-19 DUP (Lab ID: 3519325008)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- CW-20 (Lab ID: 3519325007)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- Equip blank (10/13/10) (Lab ID: 3519325009)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- LCS (Lab ID: 130563)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- MS (Lab ID: 130669)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- MSD (Lab ID: 130670)
 - Methapyrilene
 - P-Dimethylaminoazobenzene

REPORT OF LABORATORY ANALYSIS

Page 35 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8270

Description: 8270 MSSV SemiVOA App. II

Client: Sarasota County

Date: January 05, 2011

QC Batch: OEXT/3333

J(SS): Estimated Value. This analyte did not meet the secondary source verification criteria for the initial calibration. The reported result should be considered an estimated value.

- MW-15 (Lab ID: 3519325012)
 - Methapyrilene
 - P-Dimethylaminoazobenzene
- MW-16 (Lab ID: 3519325013)
 - Methapyrilene
 - P-Dimethylaminoazobenzene

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

QC Batch: OEXT/3333

L3: Analyte recovery in the laboratory control sample (LCS) exceeded QC limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- LCS (Lab ID: 130563)
 - 1,3,5-Trinitrobenzene
 - 1-Naphthylamine
 - 4-Chlorophenylphenyl ether
 - Diallate

QC Batch: OEXT/3361

L3: Analyte recovery in the laboratory control sample (LCS) exceeded QC limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.

- LCS (Lab ID: 132185)
 - 1,2-Dinitrobenzene
 - 1-Naphthylamine
 - 4-Chlorophenylphenyl ether
 - Naphthalene

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 36 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8270

Description: 8270 MSSV SemiVOA App. II

Client: Sarasota County

Date: January 05, 2011

QC Batch: OEXT/3333

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520108018

J(M0): Estimated Value. Matrix spike recovery was outside laboratory control limits.

- MS (Lab ID: 130669)
 - 1-Naphthylamine
 - 4-Chlorophenylphenyl ether
 - Diallate
- MSD (Lab ID: 130670)
 - 1-Naphthylamine

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130669)
 - 4-Nitroaniline
 - Diethylphthalate

QC Batch: OEXT/3361

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520572014

J(M0): Estimated Value. Matrix spike recovery was outside laboratory control limits.

- MS (Lab ID: 132323)
 - 1,2-Dinitrobenzene
 - 1-Naphthylamine
 - 4-Chlorophenylphenyl ether
- MSD (Lab ID: 132324)
 - 1,2-Dinitrobenzene
 - 1-Naphthylamine
 - 4-Chlorophenylphenyl ether

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 132323)
 - Benzo(a)pyrene

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Analyte Comments:

QC Batch: OEXT/3333

8p: The internal standard response associated with this result exceeds the lower control limit. However, the data is accepted based on surrogate compound recovery meeting control limits.

- CW-20 (Lab ID: 3519325007)
 - Phenol
- MS (Lab ID: 130669)
 - Phenol

REPORT OF LABORATORY ANALYSIS

Page 37 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8270

Description: 8270 MSSV SemiVOA App. II

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: OEXT/3429

D3: Sample was diluted due to the presence of high levels of non-target analytes or other matrix interference.

- C-1 (Lab ID: 3519325032)
 - Phenol
- C-3 (Lab ID: 3519325034)
 - Phenol
- C-4 (Lab ID: 3519325035)
 - Phenol
- C-5 (Lab ID: 3519325036)
 - Phenol
- MS (Lab ID: 135956)
 - Phenol
- MSD (Lab ID: 135957)
 - Phenol

REPORT OF LABORATORY ANALYSIS

Page 38 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8270 by SCAN
Description: 8270 MSSV PAH by SCAN
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 8270 by SCAN. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 3510 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: OEXT/3373

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325026

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 133219)
- Benzo(a)pyrene

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 39 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8270 by SCAN

Description: 8270 MSSV PAH by SCAN

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: OEXT/3415

3p: Reported result is estimated due to significant matrix interference to the related internal standard.

- C-1 (Lab ID: 3519325032)
- Naphthalene

REPORT OF LABORATORY ANALYSIS

Page 40 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 8260
Description: 8260 MSV
Client: Sarasota County
Date: January 05, 2011

General Information:

33 samples were analyzed for EPA 8260. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

QC Batch: MSV/2254

J(S0): Estimated Value. Surrogate recovery outside laboratory control limits.

- MS (Lab ID: 130888)
 - 1,2-Dichloroethane-d4 (S)
- MSD (Lab ID: 130889)
 - 1,2-Dichloroethane-d4 (S)

QC Batch: MSV/2258

J(S0): Estimated Value. Surrogate recovery outside laboratory control limits.

- MS (Lab ID: 130834)
 - 1,2-Dichloroethane-d4 (S)
- MSD (Lab ID: 130835)
 - 1,2-Dichloroethane-d4 (S)

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 41 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8260

Description: 8260 MSV

Client: Sarasota County

Date: January 05, 2011

QC Batch: MSV/2205

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519705002

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 124805)
 - Acetone
- MSD (Lab ID: 124806)
 - Acetone

QC Batch: MSV/2254

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520507007

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130888)
 - 1,1-Dichloroethene
 - 2-Butanone (MEK)
 - 2-Hexanone
 - Acetone
 - Acrolein
 - Bromomethane
 - Chloroethane
 - Chloromethane
 - Iodomethane
 - Vinyl chloride
- MSD (Lab ID: 130889)
 - 1,1-Dichloroethene
 - 2-Butanone (MEK)
 - 2-Hexanone
 - Acetone
 - Acrolein
 - Bromomethane
 - Chloroethane
 - Chloromethane
 - Vinyl chloride

QC Batch: MSV/2258

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519452046

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130834)
 - 1,1-Dichloroethene
 - 2-Butanone (MEK)
 - 2-Hexanone
 - Acetone
 - Acrolein
 - Bromomethane
 - Chloroethane

REPORT OF LABORATORY ANALYSIS

Page 42 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8260

Description: 8260 MSV

Client: Sarasota County

Date: January 05, 2011

QC Batch: MSV/2258

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519452046

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- Chloromethane
- Iodomethane
- Isobutyl Alcohol
- Vinyl acetate
- Vinyl chloride
- MSD (Lab ID: 130835)
 - 1,1-Dichloroethene
 - 2-Butanone (MEK)
 - 2-Hexanone
 - Acetone
 - Acrolein
 - Bromomethane
 - Chloroethane
 - Chloromethane
 - Vinyl chloride

QC Batch: MSV/2283

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520506003

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 134515)
 - Acetone
 - Iodomethane
- MSD (Lab ID: 134516)
 - 2-Butanone (MEK)
 - Acetone
 - Carbon disulfide
 - Iodomethane

QC Batch: MSV/2284

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520846001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MSD (Lab ID: 134520)
 - Bromomethane
 - Carbon disulfide
 - Iodomethane

QC Batch: MSV/2287

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520572001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 134721)

REPORT OF LABORATORY ANALYSIS

Page 43 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8260

Description: 8260 MSV

Client: Sarasota County

Date: January 05, 2011

QC Batch: MSV/2287

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520572001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- Carbon disulfide
- Chloromethane
- Vinyl acetate
- MSD (Lab ID: 134722)
 - Chloromethane
 - Vinyl acetate

QC Batch: MSV/2291

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519452053

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 135014)
 - Bromomethane
 - Carbon disulfide
- MSD (Lab ID: 135015)
 - Acetone
 - Bromomethane
 - Carbon disulfide

QC Batch: MSV/2294

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3520667001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 135451)
 - Bromomethane
 - Chloroethane
 - Chloromethane
 - Iodomethane
 - Vinyl chloride
- MSD (Lab ID: 135452)
 - 1,1-Dichloroethene
 - Bromomethane
 - Chloroethane
 - Chloromethane
 - Vinyl chloride

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 44 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 8260

Description: 8260 MSV

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: MSV/2284

6p: The continuing calibration for this compound is outside (HIGH) of method control limits. The result is estimated.

- LCS (Lab ID: 134518)
- Bromomethane

QC Batch: MSV/2294

D3: Sample was diluted due to the presence of high levels of non-target analytes or other matrix interference.

- C-1 (Lab ID: 3519325032)
 - 4-Bromofluorobenzene (S)
- C-2 (Lab ID: 3519325033)
 - 4-Bromofluorobenzene (S)
- C-3 (Lab ID: 3519325034)
 - Vinyl acetate
- C-4 (Lab ID: 3519325035)
 - 4-Bromofluorobenzene (S)
- C-5 (Lab ID: 3519325036)
 - 4-Bromofluorobenzene (S)

REPORT OF LABORATORY ANALYSIS

Page 45 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: SM 2320B
Description: 2320B Alkalinity
Client: Sarasota County
Date: January 05, 2011

General Information:

10 samples were analyzed for SM 2320B. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: WET/5629

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519243004,3519451001

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 125360)
- Alkalinity, Total as CaCO₃

QC Batch: WET/5776

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325015,3520168011

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 130983)
- Alkalinity, Total as CaCO₃

QC Batch: WET/5804

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325025,3520549005

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 132180)
- Alkalinity, Total as CaCO₃

QC Batch: WET/5978

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3521348005,3521433007

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 137477)
- Alkalinity, Total as CaCO₃

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

REPORT OF LABORATORY ANALYSIS

Page 46 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: SM 2320B

Description: 2320B Alkalinity

Client: Sarasota County

Date: January 05, 2011

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 47 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: SM 2540C
Description: 2540C Total Dissolved Solids
Client: Sarasota County
Date: January 05, 2011

General Information:

30 samples were analyzed for SM 2540C. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Q: Sample held beyond the accepted holding time.

- CW-16 (Lab ID: 3519325011)
- CW-19 (Lab ID: 3519325006)
- CW-19 DUP (Lab ID: 3519325008)
- CW-20 (Lab ID: 3519325007)
- Equip blank (10/13/10) (Lab ID: 3519325009)

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

QC Batch: WET/5756

J(D6): Estimated Value. The relative percent difference (RPD) between the sample and sample duplicate exceeded laboratory control limits.

- DUP (Lab ID: 135232)
- Total Dissolved Solids

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 48 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: SM 2540D
Description: 2540D Total Suspended Solids
Client: Sarasota County
Date: January 05, 2011

General Information:

2 samples were analyzed for SM 2540D. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 49 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: SM 4500-S2E

Description: 4500S2E Sulfide, Iodometric

Client: Sarasota County

Date: January 05, 2011

General Information:

17 samples were analyzed for SM 4500-S2E. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 50 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: SM10200

Description: Chlorophyll & Pheophytin

Client: Sarasota County

Date: January 05, 2011

General Information:

2 samples were analyzed for SM10200. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Q: Sample held beyond the accepted holding time.

- B-4R (Lab ID: 3519325001)
- B-4R DUP (Lab ID: 3519325002)

Sample Preparation:

The samples were prepared in accordance with SM10200 with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 51 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: TKN+NO_x Calculation

Description: Total Nitrogen Calculation

Client: Sarasota County

Date: January 05, 2011

General Information:

2 samples were analyzed for TKN+NO_x Calculation. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 52 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 300.0
Description: 300.0 IC Anions 28 Days
Client: Sarasota County
Date: January 05, 2011

General Information:

27 samples were analyzed for EPA 300.0. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: WETA/6986

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325025,3520634003

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 134639)
 - Sulfate
- MSD (Lab ID: 134640)
 - Sulfate

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 53 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 335.4
Description: 335.4 Cyanide, Total
Client: Sarasota County
Date: January 05, 2011

General Information:

17 samples were analyzed for EPA 335.4. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 335.4 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: WETA/7056

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325032, 3521176002

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 136073)
 - Cyanide
- MSD (Lab ID: 136074)
 - Cyanide

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Analyte Comments:

QC Batch: WETA/7056

4p: Sample required a dilution due to matrix interference, which resulted in elevated reporting limits for the target compound(s).

- C-1 (Lab ID: 3519325032)
 - Cyanide

REPORT OF LABORATORY ANALYSIS

Page 54 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 335.4

Description: 335.4 Cyanide, Total

Client: Sarasota County

Date: January 05, 2011

Analyte Comments:

QC Batch: WETA/7056

4p: Sample required a dilution due to matrix interference, which resulted in elevated reporting limits for the target compound(s).

- MS (Lab ID: 136073)
 - Cyanide
- MSD (Lab ID: 136074)
 - Cyanide

REPORT OF LABORATORY ANALYSIS

Page 55 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..

PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 350.1
Description: 350.1 Ammonia
Client: Sarasota County
Date: January 05, 2011

General Information:

30 samples were analyzed for EPA 350.1. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: WETA/7300

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3521739062

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 141536)
- Nitrogen, Ammonia

QC Batch: WETA/6895

A matrix spike and matrix spike duplicate (MS/MSD) were performed on the following sample(s): 3519325020

J(M1): Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- MS (Lab ID: 133083)
- Nitrogen, Ammonia

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 56 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 351.2

Description: 351.2 Total Kjeldahl Nitrogen

Client: Sarasota County

Date: January 05, 2011

General Information:

2 samples were analyzed for EPA 351.2. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 351.2 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 57 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Method: EPA 353.2

Description: 353.2 Nitrogen, NO₂/NO₃ pres.

Client: Sarasota County

Date: January 05, 2011

General Information:

2 samples were analyzed for EPA 353.2. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 58 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 365.4
Description: 365.4 Phosphorus, Total
Client: Sarasota County
Date: January 05, 2011

General Information:

2 samples were analyzed for EPA 365.4. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 365.4 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 59 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: EPA 410.4
Description: 410.4 COD
Client: Sarasota County
Date: January 05, 2011

General Information:

3 samples were analyzed for EPA 410.4. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

REPORT OF LABORATORY ANALYSIS

Page 60 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



PROJECT NARRATIVE

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Method: SM 5310B
Description: 5310B TOC
Client: Sarasota County
Date: January 05, 2011

General Information:

2 samples were analyzed for SM 5310B. All samples were received in acceptable condition with any exceptions noted below.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:

All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

This data package has been reviewed for quality and completeness and is approved for release.

REPORT OF LABORATORY ANALYSIS

Page 61 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R		Lab ID: 3519325001		Collected: 09/30/10 13:00		Received: 10/04/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
1631E Mercury, Low Level		Analytical Method: EPA 1631E							
Mercury	0.00352	ug/L	0.00050	0.00013	1	10/07/10 15:45	10/18/10 07:37	7439-97-6	
		Analytical Method:							
Field pH	6.68	Std. Units			1		09/30/10 13:00		
Field Temperature	25.75	deg C			1		09/30/10 13:00		
Field Specific Conductance	393	umhos/cm			1		09/30/10 13:00		
Oxygen, Dissolved	2.41	mg/L			1		09/30/10 13:00	7782-44-7	
Turbidity	6.71	NTU			1		09/30/10 13:00		
8011 GCS EDB and DBCP		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/11/10 15:15	10/12/10 04:50	96-12-8	
1,2-Dibromoethane (EDB)	0.0061U	ug/L	0.0099	0.0061	1	10/11/10 15:15	10/12/10 04:50	106-93-4	
6010 MET ICP		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Antimony	7.5U	ug/L	15.0	7.5	1	10/05/10 11:00	10/07/10 22:51	7440-36-0	
Arsenic	5.0U	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:51	7440-38-2	
Barium	20.1	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:51	7440-39-3	
Calcium	43.2	mg/L	0.50	0.25	1	10/05/10 11:00	10/07/10 22:51	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/07/10 22:51	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:51	7440-48-4	
Iron	511	ug/L	40.0	20.0	1	10/05/10 11:00	10/07/10 22:51	7439-89-6	
Magnesium	7.8	mg/L	0.50	0.25	1	10/05/10 11:00	10/07/10 22:51	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/07/10 22:51	7440-02-0	
Potassium	24.1	mg/L	1.0	0.50	1	10/05/10 11:00	10/07/10 22:51	7440-09-7	
Sodium	14.3	mg/L	1.0	0.50	1	10/05/10 11:00	10/07/10 22:51	7440-23-5	
Tot Hardness asCaCO3 (SM 2340B	140	mg/L	3.2	1.6	1	10/05/10 11:00	10/07/10 22:51		
Vanadium	5.3 I	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:51	7440-62-2	
6020 MET ICPMS		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Beryllium	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:14	7440-41-7	
Cadmium	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:14	7440-43-9	
Copper	0.93U	ug/L	1.0	0.93	1	10/05/10 11:00	10/22/10 04:14	7440-50-8	
Lead	0.50U	ug/L	1.0	0.50	1	10/05/10 11:00	10/21/10 05:27	7439-92-1	
Selenium	0.50U	ug/L	1.0	0.50	1	10/05/10 11:00	10/22/10 04:14	7782-49-2	
Silver	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:14	7440-22-4	
Thallium	0.61 I	ug/L	1.0	0.50	1	10/05/10 11:00	10/22/10 04:14	7440-28-0	
Zinc	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/22/10 04:14	7440-66-6	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/07/10 15:47	75-27-4	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 62 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R		Lab ID: 3519325001	Collected: 09/30/10 13:00	Received: 10/04/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Bromoform	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/07/10 15:47	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/07/10 15:47	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 15:47	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 15:47	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/07/10 15:47	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/07/10 15:47	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/07/10 15:47	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/07/10 15:47	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/07/10 15:47	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/07/10 15:47	1330-20-7	
4-Bromofluorobenzene (S)	100 %		70-114		1		10/07/10 15:47	460-00-4	
Dibromofluoromethane (S)	95 %		88-117		1		10/07/10 15:47	1868-53-7	
1,2-Dichloroethane-d4 (S)	112 %		86-125		1		10/07/10 15:47	17060-07-0	
Toluene-d8 (S)	99 %		87-113		1		10/07/10 15:47	2037-26-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 63 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R		Lab ID: 3519325001	Collected: 09/30/10 13:00	Received: 10/04/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity	Analytical Method: SM 2320B								
Alkalinity,Bicarbonate (CaCO3)	103	mg/L	5.0	5.0	1		10/07/10 15:58		
Alkalinity, Carbonate (CaCO3)	5.0U	mg/L	5.0	5.0	1		10/07/10 15:58		
Alkalinity, Total as CaCO3	103	mg/L	5.0	5.0	1		10/07/10 15:58		
2540C Total Dissolved Solids	Analytical Method: SM 2540C								
Total Dissolved Solids	265	mg/L	5.0	5.0	1		10/06/10 08:39		
2540D Total Suspended Solids	Analytical Method: SM 2540D								
Total Suspended Solids	10	mg/L	5.0	5.0	1		10/04/10 14:08		
Chlorophyll & Pheophytin	Analytical Method: SM10200 Preparation Method: SM10200								
Chlorophyll a	18.5	ug/L	1.0	1.0	1	09/30/10	10/15/10 10:51		
Total Nitrogen Calculation	Analytical Method: TKN+NOx Calculation								
Total Nitrogen	1.5	mg/L	0.50	0.25	1		10/06/10 15:18		
300.0 IC Anions 28 Days	Analytical Method: EPA 300.0								
Sulfate	32.3	mg/L	10.0	5.0	2		10/17/10 03:10	14808-79-8	
350.1 Ammonia	Analytical Method: EPA 350.1								
Nitrogen, Ammonia	0.082	mg/L	0.050	0.020	1		10/07/10 12:46	7664-41-7	
Nitrogen, Ammonia (Unionized)	0.020U	mg/L	0.050	0.020	1		10/07/10 12:46		
351.2 Total Kjeldahl Nitrogen	Analytical Method: EPA 351.2 Preparation Method: EPA 351.2								
Nitrogen, Kjeldahl, Total	1.5	mg/L	0.50	0.25	1	10/05/10 09:30	10/06/10 12:37	7727-37-9	
353.2 Nitrogen, NO2/NO3 pres.	Analytical Method: EPA 353.2								
Nitrogen, NO2 plus NO3	0.044 I	mg/L	0.050	0.025	1		10/05/10 11:34		
365.4 Phosphorus, Total	Analytical Method: EPA 365.4 Preparation Method: EPA 365.4								
Phosphorus, Total (as P)	0.61	mg/L	0.10	0.050	1	10/05/10 09:30	10/06/10 12:37	7723-14-0	
410.4 COD	Analytical Method: EPA 410.4								
Chemical Oxygen Demand	93.5	mg/L	25.0	12.5	1		10/06/10 17:55		
5310B TOC	Analytical Method: SM 5310B								
Total Organic Carbon	21.0	mg/L	1.0	0.50	1		10/06/10 13:25	7440-44-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R DUP		Lab ID: 3519325002		Collected: 09/30/10 13:00		Received: 10/04/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
1631E Mercury, Low Level		Analytical Method: EPA 1631E							
Mercury	0.00303	ug/L	0.00050	0.00013	1	10/07/10 15:45	10/18/10 07:43	7439-97-6	5p
		Analytical Method:							
Field pH	6.68	Std. Units			1		09/30/10 13:00		
Field Temperature	25.75	deg C			1		09/30/10 13:00		
Field Specific Conductance	393	umhos/cm			1		09/30/10 13:00		
Oxygen, Dissolved	2.41	mg/L			1		09/30/10 13:00	7782-44-7	
Turbidity	6.71	NTU			1		09/30/10 13:00		
8011 GCS EDB and DBCP		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/11/10 15:50	10/12/10 05:35	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.0099	0.0062	1	10/11/10 15:50	10/12/10 05:35	106-93-4	J(M1)
6010 MET ICP		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Antimony	7.5U	ug/L	15.0	7.5	1	10/05/10 11:00	10/07/10 22:54	7440-36-0	
Arsenic	5.0U	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:54	7440-38-2	
Barium	20.0	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:54	7440-39-3	
Calcium	43.3	mg/L	0.50	0.25	1	10/05/10 11:00	10/07/10 22:54	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/07/10 22:54	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:54	7440-48-4	
Iron	510	ug/L	40.0	20.0	1	10/05/10 11:00	10/07/10 22:54	7439-89-6	
Magnesium	7.8	mg/L	0.50	0.25	1	10/05/10 11:00	10/07/10 22:54	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/07/10 22:54	7440-02-0	
Potassium	24.2	mg/L	1.0	0.50	1	10/05/10 11:00	10/07/10 22:54	7440-09-7	
Sodium	14.1	mg/L	1.0	0.50	1	10/05/10 11:00	10/07/10 22:54	7440-23-5	
Tot Hardness asCaCO3 (SM 2340B	140	mg/L	3.2	1.6	1	10/05/10 11:00	10/07/10 22:54		
Vanadium	5.2 I	ug/L	10.0	5.0	1	10/05/10 11:00	10/07/10 22:54	7440-62-2	
6020 MET ICPMS		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Beryllium	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:19	7440-41-7	
Cadmium	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:19	7440-43-9	
Copper	1.2	ug/L	1.0	0.93	1	10/05/10 11:00	10/22/10 04:19	7440-50-8	
Lead	0.50U	ug/L	1.0	0.50	1	10/05/10 11:00	10/21/10 05:42	7439-92-1	
Selenium	0.50U	ug/L	1.0	0.50	1	10/05/10 11:00	10/22/10 04:19	7782-49-2	
Silver	0.050U	ug/L	0.10	0.050	1	10/05/10 11:00	10/22/10 04:19	7440-22-4	
Thallium	0.50U	ug/L	1.0	0.50	1	10/05/10 11:00	10/22/10 04:19	7440-28-0	
Zinc	2.5U	ug/L	5.0	2.5	1	10/05/10 11:00	10/22/10 04:19	7440-66-6	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/07/10 16:10	75-27-4	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 65 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R DUP **Lab ID: 3519325002** Collected: 09/30/10 13:00 Received: 10/04/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Bromoform	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/07/10 16:10	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/07/10 16:10	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 16:10	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 16:10	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/07/10 16:10	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/07/10 16:10	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/07/10 16:10	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/07/10 16:10	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/07/10 16:10	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/07/10 16:10	1330-20-7	
4-Bromofluorobenzene (S)	96 %		70-114		1		10/07/10 16:10	460-00-4	
Dibromofluoromethane (S)	102 %		88-117		1		10/07/10 16:10	1868-53-7	
1,2-Dichloroethane-d4 (S)	109 %		86-125		1		10/07/10 16:10	17060-07-0	
Toluene-d8 (S)	96 %		87-113		1		10/07/10 16:10	2037-26-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 66 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R DUP		Lab ID: 3519325002	Collected: 09/30/10 13:00	Received: 10/04/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity	Analytical Method: SM 2320B								
Alkalinity,Bicarbonate (CaCO3)	104	mg/L	5.0	5.0	1		10/07/10 16:02		
Alkalinity, Carbonate (CaCO3)	5.0U	mg/L	5.0	5.0	1		10/07/10 16:02		
Alkalinity, Total as CaCO3	104	mg/L	5.0	5.0	1		10/07/10 16:02		
2540C Total Dissolved Solids	Analytical Method: SM 2540C								
Total Dissolved Solids	265	mg/L	5.0	5.0	1		10/06/10 08:39		
2540D Total Suspended Solids	Analytical Method: SM 2540D								
Total Suspended Solids	5.5	mg/L	5.0	5.0	1		10/04/10 14:08		
Chlorophyll & Pheophytin	Analytical Method: SM10200 Preparation Method: SM10200								
Chlorophyll a	16.9	ug/L	1.0	1.0	1	09/30/10	10/15/10 10:51		
Total Nitrogen Calculation	Analytical Method: TKN+NOx Calculation								
Total Nitrogen	1.6	mg/L	0.50	0.25	1		10/06/10 15:18		
300.0 IC Anions 28 Days	Analytical Method: EPA 300.0								
Sulfate	34.3	mg/L	5.0	2.5	1		10/27/10 17:03	14808-79-8	
350.1 Ammonia	Analytical Method: EPA 350.1								
Nitrogen, Ammonia	0.099	mg/L	0.050	0.020	1		10/07/10 12:50	7664-41-7	
Nitrogen, Ammonia (Unionized)	0.020U	mg/L	0.050	0.020	1		10/07/10 12:50		
351.2 Total Kjeldahl Nitrogen	Analytical Method: EPA 351.2 Preparation Method: EPA 351.2								
Nitrogen, Kjeldahl, Total	1.5	mg/L	0.50	0.25	1	10/05/10 09:30	10/06/10 12:39	7727-37-9	
353.2 Nitrogen, NO2/NO3 pres.	Analytical Method: EPA 353.2								
Nitrogen, NO2 plus NO3	0.062	mg/L	0.050	0.025	1		10/05/10 11:36		
365.4 Phosphorus, Total	Analytical Method: EPA 365.4 Preparation Method: EPA 365.4								
Phosphorus, Total (as P)	0.61	mg/L	0.10	0.050	1	10/05/10 09:30	10/06/10 12:39	7723-14-0	
410.4 COD	Analytical Method: EPA 410.4								
Chemical Oxygen Demand	96.8	mg/L	25.0	12.5	1		10/06/10 17:55		
5310B TOC	Analytical Method: SM 5310B								
Total Organic Carbon	21.1	mg/L	1.0	0.50	1		10/06/10 13:41	7440-44-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: B-4R Hg Blank		Lab ID: 3519325003	Collected: 09/30/10 13:00	Received: 10/04/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
1631E Mercury, Low Level	Analytical Method: EPA 1631E								
Mercury	0.00103	ug/L	0.00050	0.00013	1	10/07/10 15:45	10/18/10 07:32	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank (9/30/10) **Lab ID:** 3519325004 **Collected:** 09/30/10 08:00 **Received:** 10/04/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/07/10 08:01	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/07/10 08:01	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/07/10 08:01	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 08:01	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/07/10 08:01	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/07/10 08:01	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/07/10 08:01	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/07/10 08:01	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	127-18-4	
Toluene	3.4	ug/L	1.0	0.50	1		10/07/10 08:01	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/07/10 08:01	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/07/10 08:01	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/07/10 08:01	1330-20-7	
4-Bromofluorobenzene (S)	98 %		70-114		1		10/07/10 08:01	460-00-4	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 69 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank (9/30/10) **Lab ID: 3519325004** Collected: 09/30/10 08:00 Received: 10/04/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Dibromofluoromethane (S)	98 %		88-117		1		10/07/10 08:01	1868-53-7	
1,2-Dichloroethane-d4 (S)	110 %		86-125		1		10/07/10 08:01	17060-07-0	
Toluene-d8 (S)	97 %		87-113		1		10/07/10 08:01	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19		Lab ID: 3519325006		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.56	Std. Units			1		10/13/10 13:02		
Field Temperature	29.0	deg C			1		10/13/10 13:02		
Field Specific Conductance	653	umhos/cm			1		10/13/10 13:02		
Oxygen, Dissolved	0.33	mg/L			1		10/13/10 13:02	7782-44-7	
Turbidity	5.0	NTU			1		10/13/10 13:02		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/21/10 16:20	10/24/10 09:49	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/21/10 16:20	10/24/10 09:49	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	10/19/10 15:17	11/12/10 21:28	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/19/10 15:17	11/12/10 21:28	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	10/19/10 15:17	11/12/10 21:28	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/19/10 15:17	11/12/10 21:28	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	10/19/10 15:17	11/12/10 21:28	58-89-9	
Chlordane (Technical)	0.082U	ug/L	0.51	0.082	1	10/19/10 15:17	11/12/10 21:28	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/19/10 15:17	11/12/10 21:28	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/19/10 15:17	11/12/10 21:28	72-54-8	
4,4'-DDE	0.00092U	ug/L	0.010	0.00092	1	10/19/10 15:17	11/12/10 21:28	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/19/10 15:17	11/12/10 21:28	50-29-3	
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	10/19/10 15:17	11/12/10 21:28	60-57-1	
Endosulfan I	0.00072U	ug/L	0.010	0.00072	1	10/19/10 15:17	11/12/10 21:28	959-98-8	
Endosulfan II	0.00072U	ug/L	0.010	0.00072	1	10/19/10 15:17	11/12/10 21:28	33213-65-9	
Endosulfan sulfate	0.00061U	ug/L	0.010	0.00061	1	10/19/10 15:17	11/12/10 21:28	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/19/10 15:17	11/12/10 21:28	72-20-8	
Endrin aldehyde	0.0073U	ug/L	0.010	0.0073	1	10/19/10 15:17	11/12/10 21:28	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/19/10 15:17	11/12/10 21:28	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/19/10 15:17	11/12/10 21:28	1024-57-3	
Methoxychlor	0.0072U	ug/L	0.010	0.0072	1	10/19/10 15:17	11/12/10 21:28	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/19/10 15:17	11/12/10 21:28	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	10/19/10 15:17	11/12/10 21:28	8001-35-2	
Tetrachloro-m-xylene (S)	93 %		66.5-120.3		1	10/19/10 15:17	11/12/10 21:28	877-09-8	
Decachlorobiphenyl (S)	98 %		41.7-109.1		1	10/19/10 15:17	11/12/10 21:28	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.082U	ug/L	0.51	0.082	1	10/19/10 15:18	11/12/10 21:28	12674-11-2	
PCB-1221 (Aroclor 1221)	0.083U	ug/L	0.51	0.083	1	10/19/10 15:18	11/12/10 21:28	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	10/19/10 15:18	11/12/10 21:28	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	10/19/10 15:18	11/12/10 21:28	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	10/19/10 15:18	11/12/10 21:28	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	10/19/10 15:18	11/12/10 21:28	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	10/19/10 15:18	11/12/10 21:28	11096-82-5	
Tetrachloro-m-xylene (S)	93 %		48-111		1	10/19/10 15:18	11/12/10 21:28	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 71 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19		Lab ID: 3519325006		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	97 %		63-121		1	10/19/10 15:18	11/12/10 21:28	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.51	0.19	1	10/18/10 17:07	10/21/10 17:50	60-51-5	
Disulfoton	0.16U	ug/L	0.51	0.16	1	10/18/10 17:07	10/21/10 17:50	298-04-4	
Famphur	0.15U	ug/L	0.51	0.15	1	10/18/10 17:07	10/21/10 17:50	52-85-7	
Methyl parathion	0.20U	ug/L	0.51	0.20	1	10/18/10 17:07	10/21/10 17:50	298-00-0	
Parathion (Ethyl parathion)	0.36U	ug/L	1.0	0.36	1	10/18/10 17:07	10/21/10 17:50	56-38-2	
Phorate	0.37U	ug/L	1.0	0.37	1	10/18/10 17:07	10/21/10 17:50	298-02-2	
4-Chloro3nitrobenzotrifluoride	80 %		34.2-122		1	10/18/10 17:07	10/21/10 17:50		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.96	0.23	1	10/19/10 17:00	10/22/10 01:26	94-75-7	
Dinoseb	0.058U	ug/L	0.19	0.058	1	10/19/10 17:00	10/22/10 01:26	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.029	0.017	1	10/19/10 17:00	10/22/10 01:26	87-86-5	
2,4,5-T	0.043U	ug/L	0.19	0.043	1	10/19/10 17:00	10/22/10 01:26	93-76-5	
2,4,5-TP (Silvex)	0.050U	ug/L	0.19	0.050	1	10/19/10 17:00	10/22/10 01:26	93-72-1	
2,4-DCPA (S)	90 %		65.5-125.7		1	10/19/10 17:00	10/22/10 01:26	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Aluminum	50.0U	ug/L	100	50.0	1	10/21/10 06:45	10/21/10 23:36	7429-90-5	
Arsenic	23.1	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:36	7440-38-2	
Barium	39.6	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:36	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:36	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:36	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:36	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:36	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:36	7440-50-8	
Iron	11800	ug/L	40.0	20.0	1	10/21/10 06:45	10/21/10 23:36	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:36	7439-92-1	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:36	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/21/10 23:36	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:36	7440-22-4	
Sodium	10.5	mg/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:36	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/21/10 23:36	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:36	7440-62-2	
Zinc	11.9 I	ug/L	20.0	10.0	1	10/21/10 06:45	10/21/10 23:36	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:41	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:41	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 08:51	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19 **Lab ID: 3519325006** Collected: 10/13/10 13:02 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	0.88U	ug/L	5.1	0.88	1	10/19/10 20:24	10/22/10 03:31	83-32-9	
Acenaphthylene	0.98U	ug/L	5.1	0.98	1	10/19/10 20:24	10/22/10 03:31	208-96-8	
Acetophenone	1.5U	ug/L	5.1	1.5	1	10/19/10 20:24	10/22/10 03:31	98-86-2	
2-Acetylaminofluorene	0.67U	ug/L	5.1	0.67	1	10/19/10 20:24	10/22/10 03:31	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.1	2.9	1	10/19/10 20:24	10/22/10 03:31	92-67-1	
Anthracene	0.62U	ug/L	5.1	0.62	1	10/19/10 20:24	10/22/10 03:31	120-12-7	
Benzo(a)anthracene	0.65U	ug/L	5.1	0.65	1	10/19/10 20:24	10/22/10 03:31	56-55-3	
Benzo(a)pyrene	0.60U	ug/L	1.0	0.60	1	10/19/10 20:24	10/22/10 03:31	50-32-8	
Benzo(b)fluoranthene	0.64U	ug/L	2.1	0.64	1	10/19/10 20:24	10/22/10 03:31	205-99-2	
Benzo(g,h,i)perylene	0.70U	ug/L	5.1	0.70	1	10/19/10 20:24	10/22/10 03:31	191-24-2	
Benzo(k)fluoranthene	0.52U	ug/L	4.1	0.52	1	10/19/10 20:24	10/22/10 03:31	207-08-9	
Benzyl alcohol	1.0U	ug/L	5.1	1.0	1	10/19/10 20:24	10/22/10 03:31	100-51-6	
4-Bromophenylphenyl ether	0.69U	ug/L	5.1	0.69	1	10/19/10 20:24	10/22/10 03:31	101-55-3	
Butylbenzylphthalate	0.74U	ug/L	5.1	0.74	1	10/19/10 20:24	10/22/10 03:31	85-68-7	
4-Chloro-3-methylphenol	0.64U	ug/L	20.5	0.64	1	10/19/10 20:24	10/22/10 03:31	59-50-7	
4-Chloroaniline	1.2U	ug/L	5.1	1.2	1	10/19/10 20:24	10/22/10 03:31	106-47-8	
bis(2-Chloroethoxy)methane	3.0U	ug/L	5.1	3.0	1	10/19/10 20:24	10/22/10 03:31	111-91-1	
bis(2-Chloroethyl) ether	0.77U	ug/L	4.1	0.77	1	10/19/10 20:24	10/22/10 03:31	111-44-4	
bis(2-Chloroisopropyl) ether	0.75U	ug/L	5.1	0.75	1	10/19/10 20:24	10/22/10 03:31	108-60-1	
2-Chloronaphthalene	0.82U	ug/L	5.1	0.82	1	10/19/10 20:24	10/22/10 03:31	91-58-7	
2-Chlorophenol	0.70U	ug/L	5.1	0.70	1	10/19/10 20:24	10/22/10 03:31	95-57-8	
4-Chlorophenylphenyl ether	0.65U	ug/L	5.1	0.65	1	10/19/10 20:24	10/22/10 03:31	7005-72-3	L3
Chrysene	0.38U	ug/L	5.1	0.38	1	10/19/10 20:24	10/22/10 03:31	218-01-9	
Diallylate	0.75U	ug/L	5.1	0.75	1	10/19/10 20:24	10/22/10 03:31	2303-16-4	L3
Dibenz(a,h)anthracene	0.67U	ug/L	2.1	0.67	1	10/19/10 20:24	10/22/10 03:31	53-70-3	
Dibenzofuran	0.69U	ug/L	5.1	0.69	1	10/19/10 20:24	10/22/10 03:31	132-64-9	
1,2-Dichlorobenzene	0.70U	ug/L	5.1	0.70	1	10/19/10 20:24	10/22/10 03:31	95-50-1	
1,3-Dichlorobenzene	0.78U	ug/L	5.1	0.78	1	10/19/10 20:24	10/22/10 03:31	541-73-1	
1,4-Dichlorobenzene	0.79U	ug/L	5.1	0.79	1	10/19/10 20:24	10/22/10 03:31	106-46-7	
3,3'-Dichlorobenzidine	0.71U	ug/L	10.3	0.71	1	10/19/10 20:24	10/22/10 03:31	91-94-1	
2,4-Dichlorophenol	0.57U	ug/L	2.1	0.57	1	10/19/10 20:24	10/22/10 03:31	120-83-2	
2,6-Dichlorophenol	0.64U	ug/L	4.1	0.64	1	10/19/10 20:24	10/22/10 03:31	87-65-0	
Diethylphthalate	0.52U	ug/L	5.1	0.52	1	10/19/10 20:24	10/22/10 03:31	84-66-2	
P-Dimethylaminoazobenzene	0.69U	ug/L	5.1	0.69	1	10/19/10 20:24	10/22/10 03:31	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.1	2.0	1	10/19/10 20:24	10/22/10 03:31	57-97-6	
3,3'-Dimethylbenzidine	3.2U	ug/L	10.3	3.2	1	10/19/10 20:24	10/22/10 03:31	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.1	1.6	1	10/19/10 20:24	10/22/10 03:31	105-67-9	
a,a-Dimethylphenylethylamine	10.3U	ug/L	20.5	10.3	1	10/19/10 20:24	10/22/10 03:31	122-09-8	
Dimethylphthalate	0.66U	ug/L	5.1	0.66	1	10/19/10 20:24	10/22/10 03:31	131-11-3	
Di-n-butylphthalate	0.42U	ug/L	5.1	0.42	1	10/19/10 20:24	10/22/10 03:31	84-74-2	
4,6-Dinitro-2-methylphenol	1.4U	ug/L	20.5	1.4	1	10/19/10 20:24	10/22/10 03:31	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/19/10 20:24	10/22/10 03:31	528-29-0	
1,3-Dinitrobenzene	0.70U	ug/L	8.2	0.70	1	10/19/10 20:24	10/22/10 03:31	99-65-0	
2,4-Dinitrophenol	1.6U	ug/L	20.5	1.6	1	10/19/10 20:24	10/22/10 03:31	51-28-5	
2,4-Dinitrotoluene	0.54U	ug/L	2.1	0.54	1	10/19/10 20:24	10/22/10 03:31	121-14-2	
2,6-Dinitrotoluene	1.3U	ug/L	2.1	1.3	1	10/19/10 20:24	10/22/10 03:31	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 73 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19		Lab ID: 3519325006		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Di-n-octylphthalate	0.92U	ug/L	5.1	0.92	1	10/19/10 20:24	10/22/10 03:31	117-84-0	
bis(2-Ethylhexyl)phthalate	0.82U	ug/L	5.1	0.82	1	10/19/10 20:24	10/22/10 03:31	117-81-7	
Ethyl methanesulfonate	0.92U	ug/L	5.1	0.92	1	10/19/10 20:24	10/22/10 03:31	62-50-0	
Fluoranthene	0.55U	ug/L	5.1	0.55	1	10/19/10 20:24	10/22/10 03:31	206-44-0	
Fluorene	0.57U	ug/L	5.1	0.57	1	10/19/10 20:24	10/22/10 03:31	86-73-7	
Hexachlorobenzene	0.82U	ug/L	1.0	0.82	1	10/19/10 20:24	10/22/10 03:31	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.1	1.3	1	10/19/10 20:24	10/22/10 03:31	77-47-4	
Hexachloroethane	0.73U	ug/L	5.1	0.73	1	10/19/10 20:24	10/22/10 03:31	67-72-1	
Hexachloropropene	1.4U	ug/L	5.1	1.4	1	10/19/10 20:24	10/22/10 03:31	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.75U	ug/L	2.1	0.75	1	10/19/10 20:24	10/22/10 03:31	193-39-5	
Isodrin	0.55U	ug/L	5.1	0.55	1	10/19/10 20:24	10/22/10 03:31	465-73-6	
Isophorone	0.75U	ug/L	5.1	0.75	1	10/19/10 20:24	10/22/10 03:31	78-59-1	
Isosafrole	0.62U	ug/L	5.1	0.62	1	10/19/10 20:24	10/22/10 03:31	120-58-1	
Kepone	10.3U	ug/L	20.5	10.3	1	10/19/10 20:24	10/22/10 03:31	143-50-0	
Methapyrilene	1.7U	ug/L	5.1	1.7	1	10/19/10 20:24	10/22/10 03:31	91-80-5	J(SS)
3-Methylcholanthrene	1.1U	ug/L	5.1	1.1	1	10/19/10 20:24	10/22/10 03:31	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.1	1.0	1	10/19/10 20:24	10/22/10 03:31	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/19/10 20:24	10/22/10 03:31	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/19/10 20:24	10/22/10 03:31	91-57-6	
2-Methylphenol(o-Cresol)	0.75U	ug/L	5.1	0.75	1	10/19/10 20:24	10/22/10 03:31	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.68U	ug/L	10.3	0.68	1	10/19/10 20:24	10/22/10 03:31		
2-Naphthylamine	2.3U	ug/L	5.1	2.3	1	10/19/10 20:24	10/22/10 03:31	91-59-8	
Naphthalene	0.80U	ug/L	5.1	0.80	1	10/19/10 20:24	10/22/10 03:31	91-20-3	
1-Naphthylamine	1.1U	ug/L	5.1	1.1	1	10/19/10 20:24	10/22/10 03:31	134-32-7	L3
1,4-Naphthoquinone	1.2U	ug/L	5.1	1.2	1	10/19/10 20:24	10/22/10 03:31	130-15-4	
2-Nitroaniline	0.62U	ug/L	5.1	0.62	1	10/19/10 20:24	10/22/10 03:31	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.1	1.0	1	10/19/10 20:24	10/22/10 03:31	99-09-2	
4-Nitroaniline	0.71U	ug/L	4.1	0.71	1	10/19/10 20:24	10/22/10 03:31	100-01-6	
Nitrobenzene	1.1U	ug/L	4.1	1.1	1	10/19/10 20:24	10/22/10 03:31	98-95-3	
2-Nitrophenol	0.83U	ug/L	5.1	0.83	1	10/19/10 20:24	10/22/10 03:31	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.5	1.1	1	10/19/10 20:24	10/22/10 03:31	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.1	1.3	1	10/19/10 20:24	10/22/10 03:31	99-55-8	
N-Nitrosodiethylamine	0.75U	ug/L	4.1	0.75	1	10/19/10 20:24	10/22/10 03:31	55-18-5	
N-Nitrosodimethylamine	1.0U	ug/L	2.1	1.0	1	10/19/10 20:24	10/22/10 03:31	62-75-9	
N-Nitroso-di-n-butylamine	0.56U	ug/L	4.1	0.56	1	10/19/10 20:24	10/22/10 03:31	924-16-3	
N-Nitroso-di-n-propylamine	0.96U	ug/L	4.1	0.96	1	10/19/10 20:24	10/22/10 03:31	621-64-7	
N-Nitrosodiphenylamine	0.51U	ug/L	5.1	0.51	1	10/19/10 20:24	10/22/10 03:31	86-30-6	
N-Nitrosomethylethylamine	0.76U	ug/L	5.1	0.76	1	10/19/10 20:24	10/22/10 03:31	10595-95-6	
N-Nitrosopiperidine	0.66U	ug/L	5.1	0.66	1	10/19/10 20:24	10/22/10 03:31	100-75-4	
N-Nitrosopyrrolidine	0.90U	ug/L	5.1	0.90	1	10/19/10 20:24	10/22/10 03:31	930-55-2	
O,O,O-Triethylphosphorothioate	0.71U	ug/L	5.1	0.71	1	10/19/10 20:24	10/22/10 03:31	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.1	1.2	1	10/19/10 20:24	10/22/10 03:31	56-38-2	
Pentachlorobenzene	0.80U	ug/L	5.1	0.80	1	10/19/10 20:24	10/22/10 03:31	608-93-5	
Pentachlorophenol	0.68U	ug/L	20.5	0.68	1	10/19/10 20:24	10/22/10 03:31	87-86-5	
Phenacetin	0.54U	ug/L	5.1	0.54	1	10/19/10 20:24	10/22/10 03:31	62-44-2	
Phenanthrene	0.53U	ug/L	5.1	0.53	1	10/19/10 20:24	10/22/10 03:31	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 74 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19		Lab ID: 3519325006		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenol	0.55U	ug/L	5.1	0.55	1	10/19/10 20:24	10/22/10 03:31	108-95-2	
p-Phenylenediamine	10.3U	ug/L	20.5	10.3	1	10/19/10 20:24	10/22/10 03:31	106-50-3	
Pronamide	1.2U	ug/L	5.1	1.2	1	10/19/10 20:24	10/22/10 03:31	23950-58-5	
Pyrene	0.70U	ug/L	5.1	0.70	1	10/19/10 20:24	10/22/10 03:31	129-00-0	
Safrole	0.87U	ug/L	5.1	0.87	1	10/19/10 20:24	10/22/10 03:31	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.72U	ug/L	5.1	0.72	1	10/19/10 20:24	10/22/10 03:31	95-94-3	
2,3,4,6-Tetrachlorophenol	4.0U	ug/L	5.1	4.0	1	10/19/10 20:24	10/22/10 03:31	58-90-2	
Thionazin	0.63U	ug/L	5.1	0.63	1	10/19/10 20:24	10/22/10 03:31	297-97-2	
O-Toluidine	1.1U	ug/L	5.1	1.1	1	10/19/10 20:24	10/22/10 03:31	95-53-4	
1,2,4-Trichlorobenzene	0.85U	ug/L	5.1	0.85	1	10/19/10 20:24	10/22/10 03:31	120-82-1	
2,4,5-Trichlorophenol	0.53U	ug/L	4.1	0.53	1	10/19/10 20:24	10/22/10 03:31	95-95-4	
2,4,6-Trichlorophenol	0.71U	ug/L	2.1	0.71	1	10/19/10 20:24	10/22/10 03:31	88-06-2	
1,3,5-Trinitrobenzene	1.3U	ug/L	5.1	1.3	1	10/19/10 20:24	10/22/10 03:31	99-35-4	L3
Nitrobenzene-d5 (S)	64	%	10-110		1	10/19/10 20:24	10/22/10 03:31	4165-60-0	
2-Fluorobiphenyl (S)	72	%	18-110		1	10/19/10 20:24	10/22/10 03:31	321-60-8	
Terphenyl-d14 (S)	83	%	10-123		1	10/19/10 20:24	10/22/10 03:31	1718-51-0	
Phenol-d6 (S)	27	%	10-110		1	10/19/10 20:24	10/22/10 03:31	13127-88-3	
2-Fluorophenol (S)	38	%	18-110		1	10/19/10 20:24	10/22/10 03:31	367-12-4	
2,4,6-Tribromophenol (S)	76	%	10-110		1	10/19/10 20:24	10/22/10 03:31	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/18/10 19:20	10/19/10 20:11	83-32-9	
Acenaphthylene	0.051U	ug/L	2.1	0.051	1	10/18/10 19:20	10/19/10 20:11	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/18/10 19:20	10/19/10 20:11	120-12-7	
Benzo(a)anthracene	0.062U	ug/L	0.21	0.062	1	10/18/10 19:20	10/19/10 20:11	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.21	0.051	1	10/18/10 19:20	10/19/10 20:11	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/18/10 19:20	10/19/10 20:11	205-99-2	
Benzo(g,h,i)perylene	0.062U	ug/L	1.0	0.062	1	10/18/10 19:20	10/19/10 20:11	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/18/10 19:20	10/19/10 20:11	207-08-9	
Chrysene	0.062U	ug/L	1.0	0.062	1	10/18/10 19:20	10/19/10 20:11	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.21	0.051	1	10/18/10 19:20	10/19/10 20:11	53-70-3	
Fluoranthene	0.062U	ug/L	1.0	0.062	1	10/18/10 19:20	10/19/10 20:11	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/18/10 19:20	10/19/10 20:11	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/18/10 19:20	10/19/10 20:11	193-39-5	
1-Methylnaphthalene	0.092U	ug/L	1.5	0.092	1	10/18/10 19:20	10/19/10 20:11	90-12-0	
2-Methylnaphthalene	0.062U	ug/L	1.5	0.062	1	10/18/10 19:20	10/19/10 20:11	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/18/10 19:20	10/19/10 20:11	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/18/10 19:20	10/19/10 20:11	85-01-8	
Pyrene	0.062U	ug/L	1.0	0.062	1	10/18/10 19:20	10/19/10 20:11	129-00-0	
2-Fluorobiphenyl (S)	77	%	43.9-113		1	10/18/10 19:20	10/19/10 20:11	321-60-8	
Terphenyl-d14 (S)	89	%	24.8-144		1	10/18/10 19:20	10/19/10 20:11	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/18/10 21:00	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 75 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 **Lab ID: 3519325006** Collected: 10/13/10 13:02 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/18/10 21:00	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-00-3	
Chloroform	0.78 U	ug/L	1.0	0.50	1		10/18/10 21:00	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/18/10 21:00	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/18/10 21:00	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/18/10 21:00	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/18/10 21:00	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/18/10 21:00	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/18/10 21:00	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:00	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/18/10 21:00	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 76 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19		Lab ID: 3519325006		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/18/10 21:00	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/18/10 21:00	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/18/10 21:00	1330-20-7	
4-Bromofluorobenzene (S)	92 %		70-114		1		10/18/10 21:00	460-00-4	
Dibromofluoromethane (S)	110 %		88-117		1		10/18/10 21:00	1868-53-7	
1,2-Dichloroethane-d4 (S)	106 %		86-125		1		10/18/10 21:00	17060-07-0	
Toluene-d8 (S)	106 %		87-113		1		10/18/10 21:00	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	371	mg/L	5.0	5.0	1		10/21/10 04:00		Q
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/19/10 08:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	12.1	mg/L	5.0	2.5	1		10/18/10 17:27	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/19/10 09:30	10/19/10 16:19	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	2.9	mg/L	0.050	0.020	1		10/22/10 13:25	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-20		Lab ID: 3519325007		Collected: 10/13/10 09:46		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.68	Std. Units			1		10/13/10 09:46		
Field Temperature	27.5	deg C			1		10/13/10 09:46		
Field Specific Conductance	854	umhos/cm			1		10/13/10 09:46		
Oxygen, Dissolved	0.37	mg/L			1		10/13/10 09:46	7782-44-7	
Turbidity	14.6	NTU			1		10/13/10 09:46		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/21/10 16:20	10/24/10 10:03	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/21/10 16:20	10/24/10 10:03	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00050U	ug/L	0.010	0.00050	1	10/19/10 15:17	11/12/10 21:45	309-00-2	
alpha-BHC	0.00030U	ug/L	0.010	0.00030	1	10/19/10 15:17	11/12/10 21:45	319-84-6	
beta-BHC	0.00050U	ug/L	0.010	0.00050	1	10/19/10 15:17	11/12/10 21:45	319-85-7	
delta-BHC	0.00040U	ug/L	0.010	0.00040	1	10/19/10 15:17	11/12/10 21:45	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	10/19/10 15:17	11/12/10 21:45	58-89-9	
Chlordane (Technical)	0.081U	ug/L	0.50	0.081	1	10/19/10 15:17	11/12/10 21:45	57-74-9	
Chlorobenzilate	0.021U	ug/L	0.10	0.021	1	10/19/10 15:17	11/12/10 21:45	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/19/10 15:17	11/12/10 21:45	72-54-8	
4,4'-DDE	0.00091U	ug/L	0.010	0.00091	1	10/19/10 15:17	11/12/10 21:45	72-55-9	
4,4'-DDT	0.0036U	ug/L	0.010	0.0036	1	10/19/10 15:17	11/12/10 21:45	50-29-3	
Dieldrin	0.00050U	ug/L	0.010	0.00050	1	10/19/10 15:17	11/12/10 21:45	60-57-1	
Endosulfan I	0.00070U	ug/L	0.010	0.00070	1	10/19/10 15:17	11/12/10 21:45	959-98-8	
Endosulfan II	0.00070U	ug/L	0.010	0.00070	1	10/19/10 15:17	11/12/10 21:45	33213-65-9	
Endosulfan sulfate	0.00060U	ug/L	0.010	0.00060	1	10/19/10 15:17	11/12/10 21:45	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/19/10 15:17	11/12/10 21:45	72-20-8	
Endrin aldehyde	0.0071U	ug/L	0.010	0.0071	1	10/19/10 15:17	11/12/10 21:45	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/19/10 15:17	11/12/10 21:45	76-44-8	
Heptachlor epoxide	0.00040U	ug/L	0.010	0.00040	1	10/19/10 15:17	11/12/10 21:45	1024-57-3	
Methoxychlor	0.0070U	ug/L	0.010	0.0070	1	10/19/10 15:17	11/12/10 21:45	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/19/10 15:17	11/12/10 21:45	82-68-8	
Toxaphene	0.29U	ug/L	0.50	0.29	1	10/19/10 15:17	11/12/10 21:45	8001-35-2	
Tetrachloro-m-xylene (S)	91 %		66.5-120.3		1	10/19/10 15:17	11/12/10 21:45	877-09-8	
Decachlorobiphenyl (S)	79 %		41.7-109.1		1	10/19/10 15:17	11/12/10 21:45	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.080U	ug/L	0.50	0.080	1	10/19/10 15:18	11/12/10 21:45	12674-11-2	
PCB-1221 (Aroclor 1221)	0.081U	ug/L	0.50	0.081	1	10/19/10 15:18	11/12/10 21:45	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.50	0.12	1	10/19/10 15:18	11/12/10 21:45	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.50	0.13	1	10/19/10 15:18	11/12/10 21:45	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.50	0.28	1	10/19/10 15:18	11/12/10 21:45	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.50	0.15	1	10/19/10 15:18	11/12/10 21:45	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.50	0.11	1	10/19/10 15:18	11/12/10 21:45	11096-82-5	
Tetrachloro-m-xylene (S)	91 %		48-111		1	10/19/10 15:18	11/12/10 21:45	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 78 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-20		Lab ID: 3519325007		Collected: 10/13/10 09:46		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	79 %		63-121		1	10/19/10 15:18	11/12/10 21:45	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.51	0.19	1	10/18/10 17:07	10/21/10 18:32	60-51-5	
Disulfoton	0.16U	ug/L	0.51	0.16	1	10/18/10 17:07	10/21/10 18:32	298-04-4	
Famphur	0.15U	ug/L	0.51	0.15	1	10/18/10 17:07	10/21/10 18:32	52-85-7	
Methyl parathion	0.20U	ug/L	0.51	0.20	1	10/18/10 17:07	10/21/10 18:32	298-00-0	
Parathion (Ethyl parathion)	0.36U	ug/L	1.0	0.36	1	10/18/10 17:07	10/21/10 18:32	56-38-2	
Phorate	0.38U	ug/L	1.0	0.38	1	10/18/10 17:07	10/21/10 18:32	298-02-2	
4-Chloro3nitrobenzotrifluoride	209 %		34.2-122		1	10/18/10 17:07	10/21/10 18:32		S3
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.97	0.23	1	10/19/10 17:00	10/22/10 01:53	94-75-7	
Dinoseb	0.059U	ug/L	0.20	0.059	1	10/19/10 17:00	10/22/10 01:53	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.029	0.018	1	10/19/10 17:00	10/22/10 01:53	87-86-5	
2,4,5-T	0.043U	ug/L	0.20	0.043	1	10/19/10 17:00	10/22/10 01:53	93-76-5	
2,4,5-TP (Silvex)	0.051U	ug/L	0.20	0.051	1	10/19/10 17:00	10/22/10 01:53	93-72-1	
2,4-DCPA (S)	91 %		65.5-125.7		1	10/19/10 17:00	10/22/10 01:53	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	26.6	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:40	7440-38-2	
Barium	45.7	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:40	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:40	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:40	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:40	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:40	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:40	7440-50-8	
Iron	7100	ug/L	40.0	20.0	1	10/21/10 06:45	10/21/10 23:40	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:40	7439-92-1	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:40	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/21/10 23:40	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:40	7440-22-4	
Sodium	24.1	mg/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:40	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/21/10 23:40	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:40	7440-62-2	
Zinc	14.3 I	ug/L	20.0	10.0	1	10/21/10 06:45	10/21/10 23:40	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:55	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:55	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:08	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-20 **Lab ID:** 3519325007 Collected: 10/13/10 09:46 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	0.89U	ug/L	5.2	0.89	1	10/19/10 20:24	10/22/10 04:02	83-32-9	
Acenaphthylene	0.99U	ug/L	5.2	0.99	1	10/19/10 20:24	10/22/10 04:02	208-96-8	
Acetophenone	1.5U	ug/L	5.2	1.5	1	10/19/10 20:24	10/22/10 04:02	98-86-2	
2-Acetylaminofluorene	0.68U	ug/L	5.2	0.68	1	10/19/10 20:24	10/22/10 04:02	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.2	2.9	1	10/19/10 20:24	10/22/10 04:02	92-67-1	
Anthracene	0.62U	ug/L	5.2	0.62	1	10/19/10 20:24	10/22/10 04:02	120-12-7	
Benzo(a)anthracene	0.66U	ug/L	5.2	0.66	1	10/19/10 20:24	10/22/10 04:02	56-55-3	
Benzo(a)pyrene	0.60U	ug/L	1.0	0.60	1	10/19/10 20:24	10/22/10 04:02	50-32-8	
Benzo(b)fluoranthene	0.64U	ug/L	2.1	0.64	1	10/19/10 20:24	10/22/10 04:02	205-99-2	
Benzo(g,h,i)perylene	0.71U	ug/L	5.2	0.71	1	10/19/10 20:24	10/22/10 04:02	191-24-2	
Benzo(k)fluoranthene	0.53U	ug/L	4.2	0.53	1	10/19/10 20:24	10/22/10 04:02	207-08-9	
Benzyl alcohol	1.1U	ug/L	5.2	1.1	1	10/19/10 20:24	10/22/10 04:02	100-51-6	
4-Bromophenylphenyl ether	0.70U	ug/L	5.2	0.70	1	10/19/10 20:24	10/22/10 04:02	101-55-3	
Butylbenzylphthalate	0.75U	ug/L	5.2	0.75	1	10/19/10 20:24	10/22/10 04:02	85-68-7	
4-Chloro-3-methylphenol	0.64U	ug/L	20.8	0.64	1	10/19/10 20:24	10/22/10 04:02	59-50-7	
4-Chloroaniline	1.3U	ug/L	5.2	1.3	1	10/19/10 20:24	10/22/10 04:02	106-47-8	
bis(2-Chloroethoxy)methane	3.1U	ug/L	5.2	3.1	1	10/19/10 20:24	10/22/10 04:02	111-91-1	
bis(2-Chloroethyl) ether	0.78U	ug/L	4.2	0.78	1	10/19/10 20:24	10/22/10 04:02	111-44-4	
bis(2-Chloroisopropyl) ether	0.76U	ug/L	5.2	0.76	1	10/19/10 20:24	10/22/10 04:02	108-60-1	
2-Chloronaphthalene	0.83U	ug/L	5.2	0.83	1	10/19/10 20:24	10/22/10 04:02	91-58-7	
2-Chlorophenol	0.71U	ug/L	5.2	0.71	1	10/19/10 20:24	10/22/10 04:02	95-57-8	
4-Chlorophenylphenyl ether	0.66U	ug/L	5.2	0.66	1	10/19/10 20:24	10/22/10 04:02	7005-72-3	L3
Chrysene	0.38U	ug/L	5.2	0.38	1	10/19/10 20:24	10/22/10 04:02	218-01-9	
Diallylate	0.76U	ug/L	5.2	0.76	1	10/19/10 20:24	10/22/10 04:02	2303-16-4	L3
Dibenz(a,h)anthracene	0.68U	ug/L	2.1	0.68	1	10/19/10 20:24	10/22/10 04:02	53-70-3	
Dibenzofuran	0.70U	ug/L	5.2	0.70	1	10/19/10 20:24	10/22/10 04:02	132-64-9	
1,2-Dichlorobenzene	0.71U	ug/L	5.2	0.71	1	10/19/10 20:24	10/22/10 04:02	95-50-1	
1,3-Dichlorobenzene	0.79U	ug/L	5.2	0.79	1	10/19/10 20:24	10/22/10 04:02	541-73-1	
1,4-Dichlorobenzene	0.80U	ug/L	5.2	0.80	1	10/19/10 20:24	10/22/10 04:02	106-46-7	
3,3'-Dichlorobenzidine	0.72U	ug/L	10.4	0.72	1	10/19/10 20:24	10/22/10 04:02	91-94-1	
2,4-Dichlorophenol	0.58U	ug/L	2.1	0.58	1	10/19/10 20:24	10/22/10 04:02	120-83-2	
2,6-Dichlorophenol	0.64U	ug/L	4.2	0.64	1	10/19/10 20:24	10/22/10 04:02	87-65-0	
Diethylphthalate	0.53U	ug/L	5.2	0.53	1	10/19/10 20:24	10/22/10 04:02	84-66-2	
P-Dimethylaminoazobenzene	0.70U	ug/L	5.2	0.70	1	10/19/10 20:24	10/22/10 04:02	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.2	2.0	1	10/19/10 20:24	10/22/10 04:02	57-97-6	
3,3'-Dimethylbenzidine	3.3U	ug/L	10.4	3.3	1	10/19/10 20:24	10/22/10 04:02	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.2	1.6	1	10/19/10 20:24	10/22/10 04:02	105-67-9	
a,a-Dimethylphenylethylamine	10.4U	ug/L	20.8	10.4	1	10/19/10 20:24	10/22/10 04:02	122-09-8	
Dimethylphthalate	0.67U	ug/L	5.2	0.67	1	10/19/10 20:24	10/22/10 04:02	131-11-3	
Di-n-butylphthalate	0.43U	ug/L	5.2	0.43	1	10/19/10 20:24	10/22/10 04:02	84-74-2	
4,6-Dinitro-2-methylphenol	1.4U	ug/L	20.8	1.4	1	10/19/10 20:24	10/22/10 04:02	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.2	1.2	1	10/19/10 20:24	10/22/10 04:02	528-29-0	
1,3-Dinitrobenzene	0.71U	ug/L	8.3	0.71	1	10/19/10 20:24	10/22/10 04:02	99-65-0	
2,4-Dinitrophenol	1.6U	ug/L	20.8	1.6	1	10/19/10 20:24	10/22/10 04:02	51-28-5	
2,4-Dinitrotoluene	0.55U	ug/L	2.1	0.55	1	10/19/10 20:24	10/22/10 04:02	121-14-2	
2,6-Dinitrotoluene	1.3U	ug/L	2.1	1.3	1	10/19/10 20:24	10/22/10 04:02	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 80 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-20 **Lab ID: 3519325007** Collected: 10/13/10 09:46 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	0.94U	ug/L	5.2	0.94	1	10/19/10 20:24	10/22/10 04:02	117-84-0	
bis(2-Ethylhexyl)phthalate	0.83U	ug/L	5.2	0.83	1	10/19/10 20:24	10/22/10 04:02	117-81-7	
Ethyl methanesulfonate	0.94U	ug/L	5.2	0.94	1	10/19/10 20:24	10/22/10 04:02	62-50-0	
Fluoranthene	0.56U	ug/L	5.2	0.56	1	10/19/10 20:24	10/22/10 04:02	206-44-0	
Fluorene	0.58U	ug/L	5.2	0.58	1	10/19/10 20:24	10/22/10 04:02	86-73-7	
Hexachlorobenzene	0.83U	ug/L	1.0	0.83	1	10/19/10 20:24	10/22/10 04:02	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.2	1.3	1	10/19/10 20:24	10/22/10 04:02	77-47-4	
Hexachloroethane	0.74U	ug/L	5.2	0.74	1	10/19/10 20:24	10/22/10 04:02	67-72-1	
Hexachloropropene	1.5U	ug/L	5.2	1.5	1	10/19/10 20:24	10/22/10 04:02	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.76U	ug/L	2.1	0.76	1	10/19/10 20:24	10/22/10 04:02	193-39-5	
Isodrin	0.56U	ug/L	5.2	0.56	1	10/19/10 20:24	10/22/10 04:02	465-73-6	
Isophorone	0.76U	ug/L	5.2	0.76	1	10/19/10 20:24	10/22/10 04:02	78-59-1	
Isosafrole	0.62U	ug/L	5.2	0.62	1	10/19/10 20:24	10/22/10 04:02	120-58-1	
Kepone	10.4U	ug/L	20.8	10.4	1	10/19/10 20:24	10/22/10 04:02	143-50-0	
Methapyrilene	1.7U	ug/L	5.2	1.7	1	10/19/10 20:24	10/22/10 04:02	91-80-5	J(SS)
3-Methylcholanthrene	1.1U	ug/L	5.2	1.1	1	10/19/10 20:24	10/22/10 04:02	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.2	1.0	1	10/19/10 20:24	10/22/10 04:02	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.2	1.0	1	10/19/10 20:24	10/22/10 04:02	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.2	1.0	1	10/19/10 20:24	10/22/10 04:02	91-57-6	
2-Methylphenol(o-Cresol)	0.76U	ug/L	5.2	0.76	1	10/19/10 20:24	10/22/10 04:02	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.69U	ug/L	10.4	0.69	1	10/19/10 20:24	10/22/10 04:02		
2-Naphthylamine	2.4U	ug/L	5.2	2.4	1	10/19/10 20:24	10/22/10 04:02	91-59-8	
Naphthalene	0.81U	ug/L	5.2	0.81	1	10/19/10 20:24	10/22/10 04:02	91-20-3	
1-Naphthylamine	1.1U	ug/L	5.2	1.1	1	10/19/10 20:24	10/22/10 04:02	134-32-7	L3
1,4-Naphthoquinone	1.2U	ug/L	5.2	1.2	1	10/19/10 20:24	10/22/10 04:02	130-15-4	
2-Nitroaniline	0.62U	ug/L	5.2	0.62	1	10/19/10 20:24	10/22/10 04:02	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.2	1.0	1	10/19/10 20:24	10/22/10 04:02	99-09-2	
4-Nitroaniline	0.72U	ug/L	4.2	0.72	1	10/19/10 20:24	10/22/10 04:02	100-01-6	
Nitrobenzene	1.1U	ug/L	4.2	1.1	1	10/19/10 20:24	10/22/10 04:02	98-95-3	
2-Nitrophenol	0.84U	ug/L	5.2	0.84	1	10/19/10 20:24	10/22/10 04:02	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.8	1.1	1	10/19/10 20:24	10/22/10 04:02	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.2	1.3	1	10/19/10 20:24	10/22/10 04:02	99-55-8	
N-Nitrosodiethylamine	0.76U	ug/L	4.2	0.76	1	10/19/10 20:24	10/22/10 04:02	55-18-5	
N-Nitrosodimethylamine	1.0U	ug/L	2.1	1.0	1	10/19/10 20:24	10/22/10 04:02	62-75-9	
N-Nitroso-di-n-butylamine	0.57U	ug/L	4.2	0.57	1	10/19/10 20:24	10/22/10 04:02	924-16-3	
N-Nitroso-di-n-propylamine	0.98U	ug/L	4.2	0.98	1	10/19/10 20:24	10/22/10 04:02	621-64-7	
N-Nitrosodiphenylamine	0.52U	ug/L	5.2	0.52	1	10/19/10 20:24	10/22/10 04:02	86-30-6	
N-Nitrosomethylethylamine	0.77U	ug/L	5.2	0.77	1	10/19/10 20:24	10/22/10 04:02	10595-95-6	
N-Nitrosopiperidine	0.67U	ug/L	5.2	0.67	1	10/19/10 20:24	10/22/10 04:02	100-75-4	
N-Nitrosopyrrolidine	0.92U	ug/L	5.2	0.92	1	10/19/10 20:24	10/22/10 04:02	930-55-2	
O,O,O-Triethylphosphorothioate	0.72U	ug/L	5.2	0.72	1	10/19/10 20:24	10/22/10 04:02	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.2	1.2	1	10/19/10 20:24	10/22/10 04:02	56-38-2	
Pentachlorobenzene	0.81U	ug/L	5.2	0.81	1	10/19/10 20:24	10/22/10 04:02	608-93-5	
Pentachlorophenol	0.69U	ug/L	20.8	0.69	1	10/19/10 20:24	10/22/10 04:02	87-86-5	
Phenacetin	0.55U	ug/L	5.2	0.55	1	10/19/10 20:24	10/22/10 04:02	62-44-2	
Phenanthrene	0.54U	ug/L	5.2	0.54	1	10/19/10 20:24	10/22/10 04:02	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 81 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-20 **Lab ID:** 3519325007 **Collected:** 10/13/10 09:46 **Received:** 10/16/10 10:20 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Phenol	0.56U	ug/L	5.2	0.56	1	10/19/10 20:24	10/22/10 04:02	108-95-2	8p
p-Phenylenediamine	10.4U	ug/L	20.8	10.4	1	10/19/10 20:24	10/22/10 04:02	106-50-3	
Pronamide	1.2U	ug/L	5.2	1.2	1	10/19/10 20:24	10/22/10 04:02	23950-58-5	
Pyrene	0.71U	ug/L	5.2	0.71	1	10/19/10 20:24	10/22/10 04:02	129-00-0	
Safrole	0.88U	ug/L	5.2	0.88	1	10/19/10 20:24	10/22/10 04:02	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.73U	ug/L	5.2	0.73	1	10/19/10 20:24	10/22/10 04:02	95-94-3	
2,3,4,6-Tetrachlorophenol	4.0U	ug/L	5.2	4.0	1	10/19/10 20:24	10/22/10 04:02	58-90-2	
Thionazin	0.63U	ug/L	5.2	0.63	1	10/19/10 20:24	10/22/10 04:02	297-97-2	
O-Toluidine	1.1U	ug/L	5.2	1.1	1	10/19/10 20:24	10/22/10 04:02	95-53-4	
1,2,4-Trichlorobenzene	0.86U	ug/L	5.2	0.86	1	10/19/10 20:24	10/22/10 04:02	120-82-1	
2,4,5-Trichlorophenol	0.54U	ug/L	4.2	0.54	1	10/19/10 20:24	10/22/10 04:02	95-95-4	
2,4,6-Trichlorophenol	0.72U	ug/L	2.1	0.72	1	10/19/10 20:24	10/22/10 04:02	88-06-2	
1,3,5-Trinitrobenzene	1.3U	ug/L	5.2	1.3	1	10/19/10 20:24	10/22/10 04:02	99-35-4	L3
Nitrobenzene-d5 (S)	76	%	10-110		1	10/19/10 20:24	10/22/10 04:02	4165-60-0	
2-Fluorobiphenyl (S)	86	%	18-110		1	10/19/10 20:24	10/22/10 04:02	321-60-8	
Terphenyl-d14 (S)	101	%	10-123		1	10/19/10 20:24	10/22/10 04:02	1718-51-0	
Phenol-d6 (S)	33	%	10-110		1	10/19/10 20:24	10/22/10 04:02	13127-88-3	
2-Fluorophenol (S)	47	%	18-110		1	10/19/10 20:24	10/22/10 04:02	367-12-4	
2,4,6-Tribromophenol (S)	96	%	10-110		1	10/19/10 20:24	10/22/10 04:02	118-79-6	
8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/18/10 19:20	10/19/10 20:32	83-32-9	
Acenaphthylene	0.051U	ug/L	2.0	0.051	1	10/18/10 19:20	10/19/10 20:32	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/18/10 19:20	10/19/10 20:32	120-12-7	
Benzo(a)anthracene	0.061U	ug/L	0.20	0.061	1	10/18/10 19:20	10/19/10 20:32	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.20	0.051	1	10/18/10 19:20	10/19/10 20:32	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/18/10 19:20	10/19/10 20:32	205-99-2	
Benzo(g,h,i)perylene	0.061U	ug/L	1.0	0.061	1	10/18/10 19:20	10/19/10 20:32	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.25	0.041	1	10/18/10 19:20	10/19/10 20:32	207-08-9	
Chrysene	0.061U	ug/L	1.0	0.061	1	10/18/10 19:20	10/19/10 20:32	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.20	0.051	1	10/18/10 19:20	10/19/10 20:32	53-70-3	
Fluoranthene	0.061U	ug/L	1.0	0.061	1	10/18/10 19:20	10/19/10 20:32	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/18/10 19:20	10/19/10 20:32	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/18/10 19:20	10/19/10 20:32	193-39-5	
1-Methylnaphthalene	0.092U	ug/L	1.5	0.092	1	10/18/10 19:20	10/19/10 20:32	90-12-0	
2-Methylnaphthalene	0.061U	ug/L	1.5	0.061	1	10/18/10 19:20	10/19/10 20:32	91-57-6	
Naphthalene	0.081U	ug/L	1.0	0.081	1	10/18/10 19:20	10/19/10 20:32	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/18/10 19:20	10/19/10 20:32	85-01-8	
Pyrene	0.061U	ug/L	1.0	0.061	1	10/18/10 19:20	10/19/10 20:32	129-00-0	
2-Fluorobiphenyl (S)	77	%	43.9-113		1	10/18/10 19:20	10/19/10 20:32	321-60-8	
Terphenyl-d14 (S)	86	%	24.8-144		1	10/18/10 19:20	10/19/10 20:32	1718-51-0	
8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/18/10 21:25	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 82 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-20 **Lab ID:** 3519325007 **Collected:** 10/13/10 09:46 **Received:** 10/16/10 10:20 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/18/10 21:25	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	78-93-3	
Carbon disulfide	0.58 U	ug/L	1.0	0.50	1		10/18/10 21:25	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-00-3	
Chloroform	2.2	ug/L	1.0	0.50	1		10/18/10 21:25	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/18/10 21:25	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/18/10 21:25	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/18/10 21:25	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/18/10 21:25	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/18/10 21:25	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/18/10 21:25	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/18/10 21:25	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/18/10 21:25	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 83 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-20		Lab ID: 3519325007		Collected: 10/13/10 09:46		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/18/10 21:25	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/18/10 21:25	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/18/10 21:25	1330-20-7	
4-Bromofluorobenzene (S)	92 %		70-114		1		10/18/10 21:25	460-00-4	
Dibromofluoromethane (S)	109 %		88-117		1		10/18/10 21:25	1868-53-7	
1,2-Dichloroethane-d4 (S)	110 %		86-125		1		10/18/10 21:25	17060-07-0	
Toluene-d8 (S)	106 %		87-113		1		10/18/10 21:25	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	604	mg/L	5.0	5.0	1		10/21/10 04:00		Q
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/19/10 08:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	18.2	mg/L	10.0	5.0	2		10/18/10 17:39	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/19/10 09:30	10/19/10 16:20	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	0.89	mg/L	0.050	0.020	1		10/22/10 13:26	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19 DUP		Lab ID: 3519325008		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.56	Std. Units			1		10/13/10 13:46		
Field Temperature	29.0	deg C			1		10/13/10 13:46		
Field Specific Conductance	653	umhos/cm			1		10/13/10 13:46		
Oxygen, Dissolved	0.33	mg/L			1		10/13/10 13:46	7782-44-7	
Turbidity	5.0	NTU			1		10/13/10 13:46		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/21/10 16:20	10/24/10 10:17	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/21/10 16:20	10/24/10 10:17	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:02	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/19/10 15:17	11/12/10 22:02	319-84-6	
beta-BHC	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:02	319-85-7	
delta-BHC	0.00042U	ug/L	0.010	0.00042	1	10/19/10 15:17	11/12/10 22:02	319-86-8	
gamma-BHC (Lindane)	0.00021U	ug/L	0.010	0.00021	1	10/19/10 15:17	11/12/10 22:02	58-89-9	
Chlordane (Technical)	0.083U	ug/L	0.52	0.083	1	10/19/10 15:17	11/12/10 22:02	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/19/10 15:17	11/12/10 22:02	510-15-6	
4,4'-DDD	0.0020U	ug/L	0.010	0.0020	1	10/19/10 15:17	11/12/10 22:02	72-54-8	
4,4'-DDE	0.00094U	ug/L	0.010	0.00094	1	10/19/10 15:17	11/12/10 22:02	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/19/10 15:17	11/12/10 22:02	50-29-3	
Dieldrin	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:02	60-57-1	
Endosulfan I	0.00073U	ug/L	0.010	0.00073	1	10/19/10 15:17	11/12/10 22:02	959-98-8	
Endosulfan II	0.00073U	ug/L	0.010	0.00073	1	10/19/10 15:17	11/12/10 22:02	33213-65-9	
Endosulfan sulfate	0.00062U	ug/L	0.010	0.00062	1	10/19/10 15:17	11/12/10 22:02	1031-07-8	
Endrin	0.0018U	ug/L	0.010	0.0018	1	10/19/10 15:17	11/12/10 22:02	72-20-8	
Endrin aldehyde	0.0074U	ug/L	0.010	0.0074	1	10/19/10 15:17	11/12/10 22:02	7421-93-4	
Heptachlor	0.0016U	ug/L	0.010	0.0016	1	10/19/10 15:17	11/12/10 22:02	76-44-8	
Heptachlor epoxide	0.00042U	ug/L	0.010	0.00042	1	10/19/10 15:17	11/12/10 22:02	1024-57-3	
Methoxychlor	0.0073U	ug/L	0.010	0.0073	1	10/19/10 15:17	11/12/10 22:02	72-43-5	
Pentachloronitrobenzene	0.016U	ug/L	0.10	0.016	1	10/19/10 15:17	11/12/10 22:02	82-68-8	
Toxaphene	0.30U	ug/L	0.52	0.30	1	10/19/10 15:17	11/12/10 22:02	8001-35-2	
Tetrachloro-m-xylene (S)	93 %		66.5-120.3		1	10/19/10 15:17	11/12/10 22:02	877-09-8	
Decachlorobiphenyl (S)	79 %		41.7-109.1		1	10/19/10 15:17	11/12/10 22:02	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.083U	ug/L	0.52	0.083	1	10/19/10 15:18	11/12/10 22:02	12674-11-2	
PCB-1221 (Aroclor 1221)	0.084U	ug/L	0.52	0.084	1	10/19/10 15:18	11/12/10 22:02	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.52	0.12	1	10/19/10 15:18	11/12/10 22:02	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.52	0.13	1	10/19/10 15:18	11/12/10 22:02	53469-21-9	
PCB-1248 (Aroclor 1248)	0.29U	ug/L	0.52	0.29	1	10/19/10 15:18	11/12/10 22:02	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.52	0.15	1	10/19/10 15:18	11/12/10 22:02	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.52	0.11	1	10/19/10 15:18	11/12/10 22:02	11096-82-5	
Tetrachloro-m-xylene (S)	93 %		48-111		1	10/19/10 15:18	11/12/10 22:02	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 85 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-19 DUP Lab ID: 3519325008 Collected: 10/13/10 13:02 Received: 10/16/10 10:20 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	79 %		63-121		1	10/19/10 15:18	11/12/10 22:02	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.20U	ug/L	0.52	0.20	1	10/18/10 17:07	10/21/10 19:13	60-51-5	
Disulfoton	0.17U	ug/L	0.52	0.17	1	10/18/10 17:07	10/21/10 19:13	298-04-4	
Famphur	0.15U	ug/L	0.52	0.15	1	10/18/10 17:07	10/21/10 19:13	52-85-7	
Methyl parathion	0.20U	ug/L	0.52	0.20	1	10/18/10 17:07	10/21/10 19:13	298-00-0	
Parathion (Ethyl parathion)	0.37U	ug/L	1.0	0.37	1	10/18/10 17:07	10/21/10 19:13	56-38-2	
Phorate	0.39U	ug/L	1.0	0.39	1	10/18/10 17:07	10/21/10 19:13	298-02-2	
4-Chloro3nitrobenzotrifluoride	55 %		34.2-122		1	10/18/10 17:07	10/21/10 19:13		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.97	0.23	1	10/19/10 17:00	10/22/10 02:19	94-75-7	
Dinoseb	0.059U	ug/L	0.19	0.059	1	10/19/10 17:00	10/22/10 02:19	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.029	0.017	1	10/19/10 17:00	10/22/10 02:19	87-86-5	
2,4,5-T	0.043U	ug/L	0.19	0.043	1	10/19/10 17:00	10/22/10 02:19	93-76-5	
2,4,5-TP (Silvex)	0.050U	ug/L	0.20	0.050	1	10/19/10 17:00	10/22/10 02:19	93-72-1	
2,4-DCPA (S)	89 %		65.5-125.7		1	10/19/10 17:00	10/22/10 02:19	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Aluminum	50.0U	ug/L	100	50.0	1	10/21/10 06:45	10/21/10 23:51	7429-90-5	
Arsenic	21.9	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:51	7440-38-2	
Barium	37.3	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:51	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:51	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:51	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:51	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:51	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:51	7440-50-8	
Iron	12600	ug/L	40.0	20.0	1	10/21/10 06:45	10/21/10 23:51	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:51	7439-92-1	
Nickel	2.6 I	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:51	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/21/10 23:51	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:51	7440-22-4	
Sodium	10.3	mg/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:51	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/21/10 23:51	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:51	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/21/10 23:51	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:59	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 03:59	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:11	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 DUP **Lab ID: 3519325008** Collected: 10/13/10 13:02 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	1.7U	ug/L	10.0	1.7	1	10/19/10 20:24	10/22/10 19:22	83-32-9	
Acenaphthylene	1.9U	ug/L	10.0	1.9	1	10/19/10 20:24	10/22/10 19:22	208-96-8	
Acetophenone	2.9U	ug/L	10.0	2.9	1	10/19/10 20:24	10/22/10 19:22	98-86-2	
2-Acetylaminofluorene	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	53-96-3	
4-Aminobiphenyl	5.7U	ug/L	10.0	5.7	1	10/19/10 20:24	10/22/10 19:22	92-67-1	
Anthracene	1.2U	ug/L	10.0	1.2	1	10/19/10 20:24	10/22/10 19:22	120-12-7	
Benzo(a)anthracene	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	56-55-3	
Benzo(a)pyrene	1.2U	ug/L	2.0	1.2	1	10/19/10 20:24	10/22/10 19:22	50-32-8	
Benzo(b)fluoranthene	1.2U	ug/L	4.0	1.2	1	10/19/10 20:24	10/22/10 19:22	205-99-2	
Benzo(g,h,i)perylene	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	191-24-2	
Benzo(k)fluoranthene	1.0U	ug/L	8.0	1.0	1	10/19/10 20:24	10/22/10 19:22	207-08-9	
Benzyl alcohol	2.0U	ug/L	10.0	2.0	1	10/19/10 20:24	10/22/10 19:22	100-51-6	
4-Bromophenylphenyl ether	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	101-55-3	
Butylbenzylphthalate	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	85-68-7	
4-Chloro-3-methylphenol	1.2U	ug/L	40.0	1.2	1	10/19/10 20:24	10/22/10 19:22	59-50-7	
4-Chloroaniline	2.4U	ug/L	10.0	2.4	1	10/19/10 20:24	10/22/10 19:22	106-47-8	
bis(2-Chloroethoxy)methane	5.9U	ug/L	10.0	5.9	1	10/19/10 20:24	10/22/10 19:22	111-91-1	
bis(2-Chloroethyl) ether	1.5U	ug/L	8.0	1.5	1	10/19/10 20:24	10/22/10 19:22	111-44-4	
bis(2-Chloroisopropyl) ether	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	108-60-1	
2-Chloronaphthalene	1.6U	ug/L	10.0	1.6	1	10/19/10 20:24	10/22/10 19:22	91-58-7	
2-Chlorophenol	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	95-57-8	
4-Chlorophenylphenyl ether	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	7005-72-3	
Chrysene	0.74U	ug/L	10.0	0.74	1	10/19/10 20:24	10/22/10 19:22	218-01-9	
Diallylate	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	2303-16-4	
Dibenz(a,h)anthracene	1.3U	ug/L	4.0	1.3	1	10/19/10 20:24	10/22/10 19:22	53-70-3	
Dibenzofuran	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	132-64-9	
1,2-Dichlorobenzene	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	95-50-1	
1,3-Dichlorobenzene	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	541-73-1	
1,4-Dichlorobenzene	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	106-46-7	
3,3'-Dichlorobenzidine	1.4U	ug/L	20.0	1.4	1	10/19/10 20:24	10/22/10 19:22	91-94-1	
2,4-Dichlorophenol	1.1U	ug/L	4.0	1.1	1	10/19/10 20:24	10/22/10 19:22	120-83-2	
2,6-Dichlorophenol	1.2U	ug/L	8.0	1.2	1	10/19/10 20:24	10/22/10 19:22	87-65-0	
Diethylphthalate	1.0U	ug/L	10.0	1.0	1	10/19/10 20:24	10/22/10 19:22	84-66-2	
P-Dimethylaminoazobenzene	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	3.9U	ug/L	10.0	3.9	1	10/19/10 20:24	10/22/10 19:22	57-97-6	
3,3'-Dimethylbenzidine	6.3U	ug/L	20.0	6.3	1	10/19/10 20:24	10/22/10 19:22	119-93-7	
2,4-Dimethylphenol	3.2U	ug/L	10.0	3.2	1	10/19/10 20:24	10/22/10 19:22	105-67-9	
a,a-Dimethylphenylethylamine	20.0U	ug/L	40.0	20.0	1	10/19/10 20:24	10/22/10 19:22	122-09-8	
Dimethylphthalate	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	131-11-3	
Di-n-butylphthalate	0.82U	ug/L	10.0	0.82	1	10/19/10 20:24	10/22/10 19:22	84-74-2	
4,6-Dinitro-2-methylphenol	2.6U	ug/L	40.0	2.6	1	10/19/10 20:24	10/22/10 19:22	534-52-1	
1,2-Dinitrobenzene	2.3U	ug/L	10.0	2.3	1	10/19/10 20:24	10/22/10 19:22	528-29-0	
1,3-Dinitrobenzene	1.4U	ug/L	16.0	1.4	1	10/19/10 20:24	10/22/10 19:22	99-65-0	
2,4-Dinitrophenol	3.1U	ug/L	40.0	3.1	1	10/19/10 20:24	10/22/10 19:22	51-28-5	
2,4-Dinitrotoluene	1.1U	ug/L	4.0	1.1	1	10/19/10 20:24	10/22/10 19:22	121-14-2	
2,6-Dinitrotoluene	2.4U	ug/L	4.0	2.4	1	10/19/10 20:24	10/22/10 19:22	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 87 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 DUP		Lab ID: 3519325008	Collected: 10/13/10 13:02	Received: 10/16/10 10:20	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Di-n-octylphthalate	1.8U	ug/L	10.0	1.8	1	10/19/10 20:24	10/22/10 19:22	117-84-0	
bis(2-Ethylhexyl)phthalate	1.6U	ug/L	10.0	1.6	1	10/19/10 20:24	10/22/10 19:22	117-81-7	
Ethyl methanesulfonate	1.8U	ug/L	10.0	1.8	1	10/19/10 20:24	10/22/10 19:22	62-50-0	
Fluoranthene	1.1U	ug/L	10.0	1.1	1	10/19/10 20:24	10/22/10 19:22	206-44-0	
Fluorene	1.1U	ug/L	10.0	1.1	1	10/19/10 20:24	10/22/10 19:22	86-73-7	
Hexachlorobenzene	1.6U	ug/L	2.0	1.6	1	10/19/10 20:24	10/22/10 19:22	118-74-1	
Hexachlorocyclopentadiene	2.6U	ug/L	10.0	2.6	1	10/19/10 20:24	10/22/10 19:22	77-47-4	
Hexachloroethane	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	67-72-1	
Hexachloropropene	2.8U	ug/L	10.0	2.8	1	10/19/10 20:24	10/22/10 19:22	1888-71-7	
Indeno(1,2,3-cd)pyrene	1.5U	ug/L	4.0	1.5	1	10/19/10 20:24	10/22/10 19:22	193-39-5	
Isodrin	1.1U	ug/L	10.0	1.1	1	10/19/10 20:24	10/22/10 19:22	465-73-6	
Isophorone	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	78-59-1	
Isosafrole	1.2U	ug/L	10.0	1.2	1	10/19/10 20:24	10/22/10 19:22	120-58-1	
Kepone	20.0U	ug/L	40.0	20.0	1	10/19/10 20:24	10/22/10 19:22	143-50-0	
Methapyrilene	3.3U	ug/L	10.0	3.3	1	10/19/10 20:24	10/22/10 19:22	91-80-5	J(SS)
3-Methylcholanthrene	2.1U	ug/L	10.0	2.1	1	10/19/10 20:24	10/22/10 19:22	56-49-5	
Methyl methanesulfonate	2.0U	ug/L	10.0	2.0	1	10/19/10 20:24	10/22/10 19:22	66-27-3	
1-Methylnaphthalene	2.0U	ug/L	10.0	2.0	1	10/19/10 20:24	10/22/10 19:22	90-12-0	
2-Methylnaphthalene	2.0U	ug/L	10.0	2.0	1	10/19/10 20:24	10/22/10 19:22	91-57-6	
2-Methylphenol(o-Cresol)	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	95-48-7	
3&4-Methylphenol(m&p Cresol)	1.3U	ug/L	20.0	1.3	1	10/19/10 20:24	10/22/10 19:22		
2-Naphthylamine	4.5U	ug/L	10.0	4.5	1	10/19/10 20:24	10/22/10 19:22	91-59-8	
Naphthalene	1.6U	ug/L	10.0	1.6	1	10/19/10 20:24	10/22/10 19:22	91-20-3	
1-Naphthylamine	2.1U	ug/L	10.0	2.1	1	10/19/10 20:24	10/22/10 19:22	134-32-7	
1,4-Naphthoquinone	2.4U	ug/L	10.0	2.4	1	10/19/10 20:24	10/22/10 19:22	130-15-4	
2-Nitroaniline	1.2U	ug/L	10.0	1.2	1	10/19/10 20:24	10/22/10 19:22	88-74-4	
3-Nitroaniline	2.0U	ug/L	10.0	2.0	1	10/19/10 20:24	10/22/10 19:22	99-09-2	
4-Nitroaniline	1.4U	ug/L	8.0	1.4	1	10/19/10 20:24	10/22/10 19:22	100-01-6	
Nitrobenzene	2.2U	ug/L	8.0	2.2	1	10/19/10 20:24	10/22/10 19:22	98-95-3	
2-Nitrophenol	1.6U	ug/L	10.0	1.6	1	10/19/10 20:24	10/22/10 19:22	88-75-5	
4-Nitrophenol	2.2U	ug/L	40.0	2.2	1	10/19/10 20:24	10/22/10 19:22	100-02-7	
5-Nitro-o-toluidine	2.6U	ug/L	10.0	2.6	1	10/19/10 20:24	10/22/10 19:22	99-55-8	
N-Nitrosodiethylamine	1.5U	ug/L	8.0	1.5	1	10/19/10 20:24	10/22/10 19:22	55-18-5	
N-Nitrosodimethylamine	1.9U	ug/L	4.0	1.9	1	10/19/10 20:24	10/22/10 19:22	62-75-9	
N-Nitroso-di-n-butylamine	1.1U	ug/L	8.0	1.1	1	10/19/10 20:24	10/22/10 19:22	924-16-3	
N-Nitroso-di-n-propylamine	1.9U	ug/L	8.0	1.9	1	10/19/10 20:24	10/22/10 19:22	621-64-7	
N-Nitrosodiphenylamine	1.0U	ug/L	10.0	1.0	1	10/19/10 20:24	10/22/10 19:22	86-30-6	
N-Nitrosomethylethylamine	1.5U	ug/L	10.0	1.5	1	10/19/10 20:24	10/22/10 19:22	10595-95-6	
N-Nitrosopiperidine	1.3U	ug/L	10.0	1.3	1	10/19/10 20:24	10/22/10 19:22	100-75-4	
N-Nitrosopyrrolidine	1.8U	ug/L	10.0	1.8	1	10/19/10 20:24	10/22/10 19:22	930-55-2	
O,O,O-Triethylphosphorothioate	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	126-68-1	
Parathion (Ethyl parathion)	2.3U	ug/L	10.0	2.3	1	10/19/10 20:24	10/22/10 19:22	56-38-2	
Pentachlorobenzene	1.6U	ug/L	10.0	1.6	1	10/19/10 20:24	10/22/10 19:22	608-93-5	
Pentachlorophenol	1.3U	ug/L	40.0	1.3	1	10/19/10 20:24	10/22/10 19:22	87-86-5	
Phenacetin	1.1U	ug/L	10.0	1.1	1	10/19/10 20:24	10/22/10 19:22	62-44-2	
Phenanthrene	1.0U	ug/L	10.0	1.0	1	10/19/10 20:24	10/22/10 19:22	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 88 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 DUP		Lab ID: 3519325008	Collected: 10/13/10 13:02	Received: 10/16/10 10:20	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenol	1.1U	ug/L	10.0	1.1	1	10/19/10 20:24	10/22/10 19:22	108-95-2	
p-Phenylenediamine	20.0U	ug/L	40.0	20.0	1	10/19/10 20:24	10/22/10 19:22	106-50-3	
Pronamide	2.3U	ug/L	10.0	2.3	1	10/19/10 20:24	10/22/10 19:22	23950-58-5	
Pyrene	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	129-00-0	
Safrole	1.7U	ug/L	10.0	1.7	1	10/19/10 20:24	10/22/10 19:22	94-59-7	
1,2,4,5-Tetrachlorobenzene	1.4U	ug/L	10.0	1.4	1	10/19/10 20:24	10/22/10 19:22	95-94-3	
2,3,4,6-Tetrachlorophenol	7.7U	ug/L	10.0	7.7	1	10/19/10 20:24	10/22/10 19:22	58-90-2	
Thionazin	1.2U	ug/L	10.0	1.2	1	10/19/10 20:24	10/22/10 19:22	297-97-2	
O-Toluidine	2.1U	ug/L	10.0	2.1	1	10/19/10 20:24	10/22/10 19:22	95-53-4	
1,2,4-Trichlorobenzene	1.7U	ug/L	10.0	1.7	1	10/19/10 20:24	10/22/10 19:22	120-82-1	
2,4,5-Trichlorophenol	1.0U	ug/L	8.0	1.0	1	10/19/10 20:24	10/22/10 19:22	95-95-4	
2,4,6-Trichlorophenol	1.4U	ug/L	4.0	1.4	1	10/19/10 20:24	10/22/10 19:22	88-06-2	
1,3,5-Trinitrobenzene	2.4U	ug/L	10.0	2.4	1	10/19/10 20:24	10/22/10 19:22	99-35-4	
Nitrobenzene-d5 (S)	73	%	10-110		1	10/19/10 20:24	10/22/10 19:22	4165-60-0	
2-Fluorobiphenyl (S)	74	%	18-110		1	10/19/10 20:24	10/22/10 19:22	321-60-8	
Terphenyl-d14 (S)	95	%	10-123		1	10/19/10 20:24	10/22/10 19:22	1718-51-0	
Phenol-d6 (S)	48	%	10-110		1	10/19/10 20:24	10/22/10 19:22	13127-88-3	
2-Fluorophenol (S)	53	%	18-110		1	10/19/10 20:24	10/22/10 19:22	367-12-4	
2,4,6-Tribromophenol (S)	71	%	10-110		1	10/19/10 20:24	10/22/10 19:22	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.060U	ug/L	2.0	0.060	1	10/18/10 19:20	10/19/10 20:54	83-32-9	
Acenaphthylene	0.10U	ug/L	4.0	0.10	1	10/18/10 19:20	10/19/10 20:54	208-96-8	
Anthracene	0.10U	ug/L	2.0	0.10	1	10/18/10 19:20	10/19/10 20:54	120-12-7	
Benzo(a)anthracene	0.12U	ug/L	0.40	0.12	1	10/18/10 19:20	10/19/10 20:54	56-55-3	
Benzo(a)pyrene	0.10U	ug/L	0.40	0.10	1	10/18/10 19:20	10/19/10 20:54	50-32-8	
Benzo(b)fluoranthene	0.10U	ug/L	0.20	0.10	1	10/18/10 19:20	10/19/10 20:54	205-99-2	
Benzo(g,h,i)perylene	0.12U	ug/L	2.0	0.12	1	10/18/10 19:20	10/19/10 20:54	191-24-2	
Benzo(k)fluoranthene	0.080U	ug/L	0.50	0.080	1	10/18/10 19:20	10/19/10 20:54	207-08-9	
Chrysene	0.12U	ug/L	2.0	0.12	1	10/18/10 19:20	10/19/10 20:54	218-01-9	
Dibenz(a,h)anthracene	0.10U	ug/L	0.40	0.10	1	10/18/10 19:20	10/19/10 20:54	53-70-3	
Fluoranthene	0.12U	ug/L	2.0	0.12	1	10/18/10 19:20	10/19/10 20:54	206-44-0	
Fluorene	0.060U	ug/L	2.0	0.060	1	10/18/10 19:20	10/19/10 20:54	86-73-7	
Indeno(1,2,3-cd)pyrene	0.080U	ug/L	0.30	0.080	1	10/18/10 19:20	10/19/10 20:54	193-39-5	
1-Methylnaphthalene	0.18U	ug/L	3.0	0.18	1	10/18/10 19:20	10/19/10 20:54	90-12-0	
2-Methylnaphthalene	0.12U	ug/L	3.0	0.12	1	10/18/10 19:20	10/19/10 20:54	91-57-6	
Naphthalene	0.16U	ug/L	2.0	0.16	1	10/18/10 19:20	10/19/10 20:54	91-20-3	
Phenanthrene	0.10U	ug/L	2.0	0.10	1	10/18/10 19:20	10/19/10 20:54	85-01-8	
Pyrene	0.12U	ug/L	2.0	0.12	1	10/18/10 19:20	10/19/10 20:54	129-00-0	
2-Fluorobiphenyl (S)	71	%	43.9-113		1	10/18/10 19:20	10/19/10 20:54	321-60-8	
Terphenyl-d14 (S)	89	%	24.8-144		1	10/18/10 19:20	10/19/10 20:54	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/27/10 20:54	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 89 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 DUP **Lab ID: 3519325008** Collected: 10/13/10 13:02 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/27/10 20:54	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/27/10 20:54	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/27/10 20:54	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 20:54	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 20:54	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/27/10 20:54	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/27/10 20:54	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 20:54	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/27/10 20:54	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 90 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-19 DUP		Lab ID: 3519325008		Collected: 10/13/10 13:02		Received: 10/16/10 10:20		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/27/10 20:54	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/27/10 20:54	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/27/10 20:54	1330-20-7	
4-Bromofluorobenzene (S)	97 %		70-114		1		10/27/10 20:54	460-00-4	
Dibromofluoromethane (S)	107 %		88-117		1		10/27/10 20:54	1868-53-7	
1,2-Dichloroethane-d4 (S)	109 %		86-125		1		10/27/10 20:54	17060-07-0	
Toluene-d8 (S)	102 %		87-113		1		10/27/10 20:54	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	362	mg/L	5.0	5.0	1		10/21/10 04:00		Q
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/19/10 08:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	11.9	mg/L	5.0	2.5	1		10/18/10 17:51	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/19/10 09:30	10/19/10 16:21	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	2.8	mg/L	0.050	0.020	1		10/22/10 13:28	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) Lab ID: 3519325009 Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8011 GCS EDB and DBCP Analytical Method: EPA 8011 Preparation Method: EPA 8011									
1,2-Dibromo-3-chloropropane	0.0048U	ug/L	0.020	0.0048	1	10/21/10 16:20	10/24/10 10:32	96-12-8	
1,2-Dibromoethane (EDB)	0.0061U	ug/L	0.0098	0.0061	1	10/21/10 16:20	10/24/10 10:32	106-93-4	
8081 GCS Pesticides Analytical Method: EPA 8081 Preparation Method: EPA 3510									
Aldrin	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:19	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/19/10 15:17	11/12/10 22:19	319-84-6	
beta-BHC	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:19	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/19/10 15:17	11/12/10 22:19	319-86-8	
gamma-BHC (Lindane)	0.00021U	ug/L	0.010	0.00021	1	10/19/10 15:17	11/12/10 22:19	58-89-9	
Chlordane (Technical)	0.083U	ug/L	0.52	0.083	1	10/19/10 15:17	11/12/10 22:19	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/19/10 15:17	11/12/10 22:19	510-15-6	
4,4'-DDD	0.0020U	ug/L	0.010	0.0020	1	10/19/10 15:17	11/12/10 22:19	72-54-8	
4,4'-DDE	0.00093U	ug/L	0.010	0.00093	1	10/19/10 15:17	11/12/10 22:19	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/19/10 15:17	11/12/10 22:19	50-29-3	
Dieldrin	0.00052U	ug/L	0.010	0.00052	1	10/19/10 15:17	11/12/10 22:19	60-57-1	
Endosulfan I	0.00073U	ug/L	0.010	0.00073	1	10/19/10 15:17	11/12/10 22:19	959-98-8	
Endosulfan II	0.00073U	ug/L	0.010	0.00073	1	10/19/10 15:17	11/12/10 22:19	33213-65-9	
Endosulfan sulfate	0.00062U	ug/L	0.010	0.00062	1	10/19/10 15:17	11/12/10 22:19	1031-07-8	
Endrin	0.0018U	ug/L	0.010	0.0018	1	10/19/10 15:17	11/12/10 22:19	72-20-8	
Endrin aldehyde	0.0074U	ug/L	0.010	0.0074	1	10/19/10 15:17	11/12/10 22:19	7421-93-4	
Heptachlor	0.0016U	ug/L	0.010	0.0016	1	10/19/10 15:17	11/12/10 22:19	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/19/10 15:17	11/12/10 22:19	1024-57-3	
Methoxychlor	0.0073U	ug/L	0.010	0.0073	1	10/19/10 15:17	11/12/10 22:19	72-43-5	
Pentachloronitrobenzene	0.016U	ug/L	0.10	0.016	1	10/19/10 15:17	11/12/10 22:19	82-68-8	
Toxaphene	0.30U	ug/L	0.52	0.30	1	10/19/10 15:17	11/12/10 22:19	8001-35-2	
Tetrachloro-m-xylene (S)	100 %		66.5-120.3		1	10/19/10 15:17	11/12/10 22:19	877-09-8	
Decachlorobiphenyl (S)	58 %		41.7-109.1		1	10/19/10 15:17	11/12/10 22:19	2051-24-3	
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
PCB-1016 (Aroclor 1016)	0.083U	ug/L	0.52	0.083	1	10/19/10 15:18	11/12/10 22:19	12674-11-2	
PCB-1221 (Aroclor 1221)	0.084U	ug/L	0.52	0.084	1	10/19/10 15:18	11/12/10 22:19	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.52	0.12	1	10/19/10 15:18	11/12/10 22:19	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.52	0.13	1	10/19/10 15:18	11/12/10 22:19	53469-21-9	
PCB-1248 (Aroclor 1248)	0.29U	ug/L	0.52	0.29	1	10/19/10 15:18	11/12/10 22:19	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.52	0.15	1	10/19/10 15:18	11/12/10 22:19	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.52	0.11	1	10/19/10 15:18	11/12/10 22:19	11096-82-5	
Tetrachloro-m-xylene (S)	100 %		48-111		1	10/19/10 15:18	11/12/10 22:19	877-09-8	
Decachlorobiphenyl (S)	58 %		63-121		1	10/19/10 15:18	11/12/10 22:19	2051-24-3	J(S0)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.20U	ug/L	0.53	0.20	1	10/18/10 17:07	10/21/10 19:54	60-51-5	
Disulfoton	0.17U	ug/L	0.53	0.17	1	10/18/10 17:07	10/21/10 19:54	298-04-4	
Famphur	0.15U	ug/L	0.53	0.15	1	10/18/10 17:07	10/21/10 19:54	52-85-7	
Methyl parathion	0.20U	ug/L	0.53	0.20	1	10/18/10 17:07	10/21/10 19:54	298-00-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 92 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) Lab ID: 3519325009 Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Parathion (Ethyl parathion)	0.37U	ug/L	1.1	0.37	1	10/18/10 17:07	10/21/10 19:54	56-38-2	
Phorate	0.39U	ug/L	1.1	0.39	1	10/18/10 17:07	10/21/10 19:54	298-02-2	
4-Chloro3nitrobenzotrifluoride	90 %		34.2-122		1	10/18/10 17:07	10/21/10 19:54		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.24U	ug/L	1.0	0.24	1	10/19/10 17:00	10/22/10 02:46	94-75-7	
Dinoseb	0.062U	ug/L	0.20	0.062	1	10/19/10 17:00	10/22/10 02:46	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.031	0.018	1	10/19/10 17:00	10/22/10 02:46	87-86-5	
2,4,5-T	0.046U	ug/L	0.21	0.046	1	10/19/10 17:00	10/22/10 02:46	93-76-5	
2,4,5-TP (Silvex)	0.053U	ug/L	0.21	0.053	1	10/19/10 17:00	10/22/10 02:46	93-72-1	
2,4-DCPA (S)	104 %		65.5-125.7		1	10/19/10 17:00	10/22/10 02:46	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:55	7440-38-2	
Barium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:55	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:55	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:55	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:55	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:55	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:55	7440-50-8	
Iron	20.0U	ug/L	40.0	20.0	1	10/21/10 06:45	10/21/10 23:55	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:55	7439-92-1	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:55	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/21/10 23:55	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:55	7440-22-4	
Sodium	0.50U	mg/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:55	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/21/10 23:55	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:55	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/21/10 23:55	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:04	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:04	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:13	7439-97-6	
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	0.91U	ug/L	5.3	0.91	1	10/19/10 20:24	10/22/10 19:53	83-32-9	
Acenaphthylene	1.0U	ug/L	5.3	1.0	1	10/19/10 20:24	10/22/10 19:53	208-96-8	
Acetophenone	1.5U	ug/L	5.3	1.5	1	10/19/10 20:24	10/22/10 19:53	98-86-2	
2-Acetylaminofluorene	0.69U	ug/L	5.3	0.69	1	10/19/10 20:24	10/22/10 19:53	53-96-3	
4-Aminobiphenyl	3.0U	ug/L	5.3	3.0	1	10/19/10 20:24	10/22/10 19:53	92-67-1	
Anthracene	0.63U	ug/L	5.3	0.63	1	10/19/10 20:24	10/22/10 19:53	120-12-7	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 93 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) **Lab ID: 3519325009** Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Benzo(a)anthracene	0.67U	ug/L	5.3	0.67	1	10/19/10 20:24	10/22/10 19:53	56-55-3	
Benzo(a)pyrene	0.61U	ug/L	1.1	0.61	1	10/19/10 20:24	10/22/10 19:53	50-32-8	
Benzo(b)fluoranthene	0.66U	ug/L	2.1	0.66	1	10/19/10 20:24	10/22/10 19:53	205-99-2	
Benzo(g,h,i)perylene	0.72U	ug/L	5.3	0.72	1	10/19/10 20:24	10/22/10 19:53	191-24-2	
Benzo(k)fluoranthene	0.54U	ug/L	4.2	0.54	1	10/19/10 20:24	10/22/10 19:53	207-08-9	
Benzyl alcohol	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	100-51-6	
4-Bromophenylphenyl ether	0.71U	ug/L	5.3	0.71	1	10/19/10 20:24	10/22/10 19:53	101-55-3	
Butylbenzylphthalate	0.76U	ug/L	5.3	0.76	1	10/19/10 20:24	10/22/10 19:53	85-68-7	
4-Chloro-3-methylphenol	0.66U	ug/L	21.2	0.66	1	10/19/10 20:24	10/22/10 19:53	59-50-7	
4-Chloroaniline	1.3U	ug/L	5.3	1.3	1	10/19/10 20:24	10/22/10 19:53	106-47-8	
bis(2-Chloroethoxy)methane	3.1U	ug/L	5.3	3.1	1	10/19/10 20:24	10/22/10 19:53	111-91-1	
bis(2-Chloroethyl) ether	0.79U	ug/L	4.2	0.79	1	10/19/10 20:24	10/22/10 19:53	111-44-4	
bis(2-Chloroisopropyl) ether	0.77U	ug/L	5.3	0.77	1	10/19/10 20:24	10/22/10 19:53	108-60-1	
2-Chloronaphthalene	0.85U	ug/L	5.3	0.85	1	10/19/10 20:24	10/22/10 19:53	91-58-7	
2-Chlorophenol	0.72U	ug/L	5.3	0.72	1	10/19/10 20:24	10/22/10 19:53	95-57-8	
4-Chlorophenylphenyl ether	0.67U	ug/L	5.3	0.67	1	10/19/10 20:24	10/22/10 19:53	7005-72-3	
Chrysene	0.39U	ug/L	5.3	0.39	1	10/19/10 20:24	10/22/10 19:53	218-01-9	
Diallate	0.77U	ug/L	5.3	0.77	1	10/19/10 20:24	10/22/10 19:53	2303-16-4	
Dibenz(a,h)anthracene	0.69U	ug/L	2.1	0.69	1	10/19/10 20:24	10/22/10 19:53	53-70-3	
Dibenzofuran	0.71U	ug/L	5.3	0.71	1	10/19/10 20:24	10/22/10 19:53	132-64-9	
1,2-Dichlorobenzene	0.72U	ug/L	5.3	0.72	1	10/19/10 20:24	10/22/10 19:53	95-50-1	
1,3-Dichlorobenzene	0.80U	ug/L	5.3	0.80	1	10/19/10 20:24	10/22/10 19:53	541-73-1	
1,4-Dichlorobenzene	0.81U	ug/L	5.3	0.81	1	10/19/10 20:24	10/22/10 19:53	106-46-7	
3,3'-Dichlorobenzidine	0.73U	ug/L	10.6	0.73	1	10/19/10 20:24	10/22/10 19:53	91-94-1	
2,4-Dichlorophenol	0.59U	ug/L	2.1	0.59	1	10/19/10 20:24	10/22/10 19:53	120-83-2	
2,6-Dichlorophenol	0.66U	ug/L	4.2	0.66	1	10/19/10 20:24	10/22/10 19:53	87-65-0	
Diethylphthalate	0.54U	ug/L	5.3	0.54	1	10/19/10 20:24	10/22/10 19:53	84-66-2	
P-Dimethylaminoazobenzene	0.71U	ug/L	5.3	0.71	1	10/19/10 20:24	10/22/10 19:53	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.1U	ug/L	5.3	2.1	1	10/19/10 20:24	10/22/10 19:53	57-97-6	
3,3'-Dimethylbenzidine	3.3U	ug/L	10.6	3.3	1	10/19/10 20:24	10/22/10 19:53	119-93-7	
2,4-Dimethylphenol	1.7U	ug/L	5.3	1.7	1	10/19/10 20:24	10/22/10 19:53	105-67-9	
a,a-Dimethylphenylethylamine	10.6U	ug/L	21.2	10.6	1	10/19/10 20:24	10/22/10 19:53	122-09-8	
Dimethylphthalate	0.68U	ug/L	5.3	0.68	1	10/19/10 20:24	10/22/10 19:53	131-11-3	
Di-n-butylphthalate	0.43U	ug/L	5.3	0.43	1	10/19/10 20:24	10/22/10 19:53	84-74-2	
4,6-Dinitro-2-methylphenol	1.4U	ug/L	21.2	1.4	1	10/19/10 20:24	10/22/10 19:53	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.3	1.2	1	10/19/10 20:24	10/22/10 19:53	528-29-0	
1,3-Dinitrobenzene	0.72U	ug/L	8.5	0.72	1	10/19/10 20:24	10/22/10 19:53	99-65-0	
2,4-Dinitrophenol	1.7U	ug/L	21.2	1.7	1	10/19/10 20:24	10/22/10 19:53	51-28-5	
2,4-Dinitrotoluene	0.56U	ug/L	2.1	0.56	1	10/19/10 20:24	10/22/10 19:53	121-14-2	
2,6-Dinitrotoluene	1.3U	ug/L	2.1	1.3	1	10/19/10 20:24	10/22/10 19:53	606-20-2	
Di-n-octylphthalate	0.95U	ug/L	5.3	0.95	1	10/19/10 20:24	10/22/10 19:53	117-84-0	
bis(2-Ethylhexyl)phthalate	0.85U	ug/L	5.3	0.85	1	10/19/10 20:24	10/22/10 19:53	117-81-7	
Ethyl methanesulfonate	0.95U	ug/L	5.3	0.95	1	10/19/10 20:24	10/22/10 19:53	62-50-0	
Fluoranthene	0.57U	ug/L	5.3	0.57	1	10/19/10 20:24	10/22/10 19:53	206-44-0	
Fluorene	0.59U	ug/L	5.3	0.59	1	10/19/10 20:24	10/22/10 19:53	86-73-7	
Hexachlorobenzene	0.85U	ug/L	1.1	0.85	1	10/19/10 20:24	10/22/10 19:53	118-74-1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 94 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) **Lab ID: 3519325009** Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Hexachlorocyclopentadiene	1.4U	ug/L	5.3	1.4	1	10/19/10 20:24	10/22/10 19:53	77-47-4	
Hexachloroethane	0.75U	ug/L	5.3	0.75	1	10/19/10 20:24	10/22/10 19:53	67-72-1	
Hexachloropropene	1.5U	ug/L	5.3	1.5	1	10/19/10 20:24	10/22/10 19:53	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.77U	ug/L	2.1	0.77	1	10/19/10 20:24	10/22/10 19:53	193-39-5	
Isodrin	0.57U	ug/L	5.3	0.57	1	10/19/10 20:24	10/22/10 19:53	465-73-6	
Isophorone	0.77U	ug/L	5.3	0.77	1	10/19/10 20:24	10/22/10 19:53	78-59-1	
Isosafrole	0.63U	ug/L	5.3	0.63	1	10/19/10 20:24	10/22/10 19:53	120-58-1	
Kepone	10.6U	ug/L	21.2	10.6	1	10/19/10 20:24	10/22/10 19:53	143-50-0	
Methapyrilene	1.7U	ug/L	5.3	1.7	1	10/19/10 20:24	10/22/10 19:53	91-80-5	J(SS)
3-Methylcholanthrene	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	56-49-5	
Methyl methanesulfonate	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	66-27-3	
1-Methylnaphthalene	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.3	1.0	1	10/19/10 20:24	10/22/10 19:53	91-57-6	
2-Methylphenol(o-Cresol)	0.77U	ug/L	5.3	0.77	1	10/19/10 20:24	10/22/10 19:53	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.70U	ug/L	10.6	0.70	1	10/19/10 20:24	10/22/10 19:53		
2-Naphthylamine	2.4U	ug/L	5.3	2.4	1	10/19/10 20:24	10/22/10 19:53	91-59-8	
Naphthalene	0.82U	ug/L	5.3	0.82	1	10/19/10 20:24	10/22/10 19:53	91-20-3	
1-Naphthylamine	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	134-32-7	
1,4-Naphthoquinone	1.2U	ug/L	5.3	1.2	1	10/19/10 20:24	10/22/10 19:53	130-15-4	
2-Nitroaniline	0.63U	ug/L	5.3	0.63	1	10/19/10 20:24	10/22/10 19:53	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.3	1.0	1	10/19/10 20:24	10/22/10 19:53	99-09-2	
4-Nitroaniline	0.73U	ug/L	4.2	0.73	1	10/19/10 20:24	10/22/10 19:53	100-01-6	
Nitrobenzene	1.2U	ug/L	4.2	1.2	1	10/19/10 20:24	10/22/10 19:53	98-95-3	
2-Nitrophenol	0.86U	ug/L	5.3	0.86	1	10/19/10 20:24	10/22/10 19:53	88-75-5	
4-Nitrophenol	1.1U	ug/L	21.2	1.1	1	10/19/10 20:24	10/22/10 19:53	100-02-7	
5-Nitro-o-toluidine	1.4U	ug/L	5.3	1.4	1	10/19/10 20:24	10/22/10 19:53	99-55-8	
N-Nitrosodiethylamine	0.77U	ug/L	4.2	0.77	1	10/19/10 20:24	10/22/10 19:53	55-18-5	
N-Nitrosodimethylamine	1.0U	ug/L	2.1	1.0	1	10/19/10 20:24	10/22/10 19:53	62-75-9	
N-Nitroso-di-n-butylamine	0.58U	ug/L	4.2	0.58	1	10/19/10 20:24	10/22/10 19:53	924-16-3	
N-Nitroso-di-n-propylamine	0.99U	ug/L	4.2	0.99	1	10/19/10 20:24	10/22/10 19:53	621-64-7	
N-Nitrosodiphenylamine	0.53U	ug/L	5.3	0.53	1	10/19/10 20:24	10/22/10 19:53	86-30-6	
N-Nitrosomethylethylamine	0.78U	ug/L	5.3	0.78	1	10/19/10 20:24	10/22/10 19:53	10595-95-6	
N-Nitrosopiperidine	0.68U	ug/L	5.3	0.68	1	10/19/10 20:24	10/22/10 19:53	100-75-4	
N-Nitrosopyrrolidine	0.93U	ug/L	5.3	0.93	1	10/19/10 20:24	10/22/10 19:53	930-55-2	
O,O,O-Triethylphosphorothioate	0.73U	ug/L	5.3	0.73	1	10/19/10 20:24	10/22/10 19:53	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.3	1.2	1	10/19/10 20:24	10/22/10 19:53	56-38-2	
Pentachlorobenzene	0.82U	ug/L	5.3	0.82	1	10/19/10 20:24	10/22/10 19:53	608-93-5	
Pentachlorophenol	0.70U	ug/L	21.2	0.70	1	10/19/10 20:24	10/22/10 19:53	87-86-5	
Phenacetin	0.56U	ug/L	5.3	0.56	1	10/19/10 20:24	10/22/10 19:53	62-44-2	
Phenanthrene	0.55U	ug/L	5.3	0.55	1	10/19/10 20:24	10/22/10 19:53	85-01-8	
Phenol	0.57U	ug/L	5.3	0.57	1	10/19/10 20:24	10/22/10 19:53	108-95-2	
p-Phenylenediamine	10.6U	ug/L	21.2	10.6	1	10/19/10 20:24	10/22/10 19:53	106-50-3	
Pronamide	1.2U	ug/L	5.3	1.2	1	10/19/10 20:24	10/22/10 19:53	23950-58-5	
Pyrene	0.72U	ug/L	5.3	0.72	1	10/19/10 20:24	10/22/10 19:53	129-00-0	
Safrole	0.90U	ug/L	5.3	0.90	1	10/19/10 20:24	10/22/10 19:53	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.74U	ug/L	5.3	0.74	1	10/19/10 20:24	10/22/10 19:53	95-94-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 95 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) **Lab ID: 3519325009** Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,3,4,6-Tetrachlorophenol	4.1U	ug/L	5.3	4.1	1	10/19/10 20:24	10/22/10 19:53	58-90-2	
Thionazin	0.65U	ug/L	5.3	0.65	1	10/19/10 20:24	10/22/10 19:53	297-97-2	
O-Toluidine	1.1U	ug/L	5.3	1.1	1	10/19/10 20:24	10/22/10 19:53	95-53-4	
1,2,4-Trichlorobenzene	0.88U	ug/L	5.3	0.88	1	10/19/10 20:24	10/22/10 19:53	120-82-1	
2,4,5-Trichlorophenol	0.55U	ug/L	4.2	0.55	1	10/19/10 20:24	10/22/10 19:53	95-95-4	
2,4,6-Trichlorophenol	0.73U	ug/L	2.1	0.73	1	10/19/10 20:24	10/22/10 19:53	88-06-2	
1,3,5-Trinitrobenzene	1.3U	ug/L	5.3	1.3	1	10/19/10 20:24	10/22/10 19:53	99-35-4	
Nitrobenzene-d5 (S)	81	%	10-110		1	10/19/10 20:24	10/22/10 19:53	4165-60-0	
2-Fluorobiphenyl (S)	82	%	18-110		1	10/19/10 20:24	10/22/10 19:53	321-60-8	
Terphenyl-d14 (S)	97	%	10-123		1	10/19/10 20:24	10/22/10 19:53	1718-51-0	
Phenol-d6 (S)	30	%	10-110		1	10/19/10 20:24	10/22/10 19:53	13127-88-3	
2-Fluorophenol (S)	46	%	18-110		1	10/19/10 20:24	10/22/10 19:53	367-12-4	
2,4,6-Tribromophenol (S)	76	%	10-110		1	10/19/10 20:24	10/22/10 19:53	118-79-6	

8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.032U	ug/L	1.1	0.032	1	10/18/10 19:20	10/19/10 21:15	83-32-9	
Acenaphthylene	0.053U	ug/L	2.1	0.053	1	10/18/10 19:20	10/19/10 21:15	208-96-8	
Anthracene	0.053U	ug/L	1.1	0.053	1	10/18/10 19:20	10/19/10 21:15	120-12-7	
Benzo(a)anthracene	0.064U	ug/L	0.21	0.064	1	10/18/10 19:20	10/19/10 21:15	56-55-3	
Benzo(a)pyrene	0.053U	ug/L	0.21	0.053	1	10/18/10 19:20	10/19/10 21:15	50-32-8	
Benzo(b)fluoranthene	0.053U	ug/L	0.11	0.053	1	10/18/10 19:20	10/19/10 21:15	205-99-2	
Benzo(g,h,i)perylene	0.064U	ug/L	1.1	0.064	1	10/18/10 19:20	10/19/10 21:15	191-24-2	
Benzo(k)fluoranthene	0.042U	ug/L	0.26	0.042	1	10/18/10 19:20	10/19/10 21:15	207-08-9	
Chrysene	0.064U	ug/L	1.1	0.064	1	10/18/10 19:20	10/19/10 21:15	218-01-9	
Dibenz(a,h)anthracene	0.053U	ug/L	0.21	0.053	1	10/18/10 19:20	10/19/10 21:15	53-70-3	
Fluoranthene	0.064U	ug/L	1.1	0.064	1	10/18/10 19:20	10/19/10 21:15	206-44-0	
Fluorene	0.032U	ug/L	1.1	0.032	1	10/18/10 19:20	10/19/10 21:15	86-73-7	
Indeno(1,2,3-cd)pyrene	0.042U	ug/L	0.16	0.042	1	10/18/10 19:20	10/19/10 21:15	193-39-5	
1-Methylnaphthalene	0.095U	ug/L	1.6	0.095	1	10/18/10 19:20	10/19/10 21:15	90-12-0	
2-Methylnaphthalene	0.064U	ug/L	1.6	0.064	1	10/18/10 19:20	10/19/10 21:15	91-57-6	
Naphthalene	0.085U	ug/L	1.1	0.085	1	10/18/10 19:20	10/19/10 21:15	91-20-3	
Phenanthrene	0.053U	ug/L	1.1	0.053	1	10/18/10 19:20	10/19/10 21:15	85-01-8	
Pyrene	0.064U	ug/L	1.1	0.064	1	10/18/10 19:20	10/19/10 21:15	129-00-0	
2-Fluorobiphenyl (S)	76	%	43.9-113		1	10/18/10 19:20	10/19/10 21:15	321-60-8	
Terphenyl-d14 (S)	92	%	24.8-144		1	10/18/10 19:20	10/19/10 21:15	1718-51-0	

8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/19/10 16:51	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/19/10 16:51	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-25-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 96 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) **Lab ID: 3519325009** Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/19/10 16:51	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/19/10 16:51	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/19/10 16:51	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/19/10 16:51	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/19/10 16:51	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/19/10 16:51	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 16:51	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/19/10 16:51	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/19/10 16:51	96-18-4	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 97 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip blank (10/13/10) Lab ID: 3519325009 Collected: 10/13/10 10:40 Received: 10/16/10 10:20 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/19/10 16:51	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/19/10 16:51	1330-20-7	
4-Bromofluorobenzene (S)	94	%	70-114		1		10/19/10 16:51	460-00-4	
Dibromofluoromethane (S)	103	%	88-117		1		10/19/10 16:51	1868-53-7	
1,2-Dichloroethane-d4 (S)	103	%	86-125		1		10/19/10 16:51	17060-07-0	
Toluene-d8 (S)	102	%	87-113		1		10/19/10 16:51	2037-26-5	
2540C Total Dissolved Solids Analytical Method: SM 2540C									
Total Dissolved Solids	5.0U	mg/L	5.0	5.0	1		10/21/10 04:00		Q
4500S2E Sulfide, Iodometric Analytical Method: SM 4500-S2E									
Sulfide	1.0U	mg/L	1.0	1.0	1		10/19/10 08:30	18496-25-8	
300.0 IC Anions 28 Days Analytical Method: EPA 300.0									
Chloride	2.5U	mg/L	5.0	2.5	1		10/18/10 18:03	16887-00-6	
335.4 Cyanide, Total Analytical Method: EPA 335.4 Preparation Method: EPA 335.4									
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/19/10 09:30	10/19/10 16:22	57-12-5	
350.1 Ammonia Analytical Method: EPA 350.1									
Nitrogen, Ammonia	0.020U	mg/L	0.050	0.020	1		10/22/10 13:32	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank (10/13/10) **Lab ID:** 3519325010 **Collected:** 10/13/10 08:00 **Received:** 10/16/10 10:20 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/19/10 17:16	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/19/10 17:16	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/19/10 17:16	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/19/10 17:16	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/19/10 17:16	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/19/10 17:16	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/19/10 17:16	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/19/10 17:16	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/19/10 17:16	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 99 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank (10/13/10) **Lab ID:** 3519325010 Collected: 10/13/10 08:00 Received: 10/16/10 10:20 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/19/10 17:16	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	127-18-4	
Toluene	0.79 I	ug/L	1.0	0.50	1		10/19/10 17:16	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/19/10 17:16	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/19/10 17:16	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/19/10 17:16	1330-20-7	
4-Bromofluorobenzene (S)	95 %		70-114		1		10/19/10 17:16	460-00-4	
Dibromofluoromethane (S)	104 %		88-117		1		10/19/10 17:16	1868-53-7	
1,2-Dichloroethane-d4 (S)	105 %		86-125		1		10/19/10 17:16	17060-07-0	
Toluene-d8 (S)	104 %		87-113		1		10/19/10 17:16	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-16		Lab ID: 3519325011		Collected: 10/13/10 16:17		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	5.97	Std. Units			1		10/13/10 16:17		
Field Temperature	27.7	deg C			1		10/13/10 16:17		
Field Specific Conductance	1536	umhos/cm			1		10/13/10 16:17		
Oxygen, Dissolved	0.259	mg/L			1		10/13/10 16:17	7782-44-7	
Turbidity	16.4	NTU			1		10/13/10 16:17		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/21/10 16:20	10/24/10 10:46	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/21/10 16:20	10/24/10 10:46	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 22:36	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/20/10 12:00	11/12/10 22:36	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 22:36	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/20/10 12:00	11/12/10 22:36	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	10/20/10 12:00	11/12/10 22:36	58-89-9	
Chlordane (Technical)	0.082U	ug/L	0.51	0.082	1	10/20/10 12:00	11/12/10 22:36	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/20/10 12:00	11/12/10 22:36	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/20/10 12:00	11/12/10 22:36	72-54-8	
4,4'-DDE	0.00092U	ug/L	0.010	0.00092	1	10/20/10 12:00	11/12/10 22:36	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/20/10 12:00	11/12/10 22:36	50-29-3	
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 22:36	60-57-1	
Endosulfan I	0.00071U	ug/L	0.010	0.00071	1	10/20/10 12:00	11/12/10 22:36	959-98-8	
Endosulfan II	0.00071U	ug/L	0.010	0.00071	1	10/20/10 12:00	11/12/10 22:36	33213-65-9	
Endosulfan sulfate	0.00061U	ug/L	0.010	0.00061	1	10/20/10 12:00	11/12/10 22:36	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/20/10 12:00	11/12/10 22:36	72-20-8	
Endrin aldehyde	0.0072U	ug/L	0.010	0.0072	1	10/20/10 12:00	11/12/10 22:36	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/20/10 12:00	11/12/10 22:36	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/20/10 12:00	11/12/10 22:36	1024-57-3	
Methoxychlor	0.0071U	ug/L	0.010	0.0071	1	10/20/10 12:00	11/12/10 22:36	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/20/10 12:00	11/12/10 22:36	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	10/20/10 12:00	11/12/10 22:36	8001-35-2	
Tetrachloro-m-xylene (S)	82 %		66.5-120.3		1	10/20/10 12:00	11/12/10 22:36	877-09-8	
Decachlorobiphenyl (S)	73 %		41.7-109.1		1	10/20/10 12:00	11/12/10 22:36	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.082U	ug/L	0.51	0.082	1	10/20/10 12:01	11/12/10 22:36	12674-11-2	
PCB-1221 (Aroclor 1221)	0.083U	ug/L	0.51	0.083	1	10/20/10 12:01	11/12/10 22:36	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	10/20/10 12:01	11/12/10 22:36	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	10/20/10 12:01	11/12/10 22:36	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	10/20/10 12:01	11/12/10 22:36	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	10/20/10 12:01	11/12/10 22:36	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	10/20/10 12:01	11/12/10 22:36	11096-82-5	
Tetrachloro-m-xylene (S)	81 %		48-111		1	10/20/10 12:01	11/12/10 22:36	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 101 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-16		Lab ID: 3519325011		Collected: 10/13/10 16:17		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	73 %		63-121		1	10/20/10 12:01	11/12/10 22:36	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.20U	ug/L	0.53	0.20	1	10/20/10 16:08	11/02/10 01:30	60-51-5	
Disulfoton	0.17U	ug/L	0.53	0.17	1	10/20/10 16:08	11/02/10 01:30	298-04-4	
Famphur	0.15U	ug/L	0.53	0.15	1	10/20/10 16:08	11/02/10 01:30	52-85-7	
Methyl parathion	0.20U	ug/L	0.53	0.20	1	10/20/10 16:08	11/02/10 01:30	298-00-0	
Parathion (Ethyl parathion)	0.37U	ug/L	1.1	0.37	1	10/20/10 16:08	11/02/10 01:30	56-38-2	
Phorate	0.39U	ug/L	1.1	0.39	1	10/20/10 16:08	11/02/10 01:30	298-02-2	
4-Chloro3nitrobenzotrifluoride	66 %		34.2-122		1	10/20/10 16:08	11/02/10 01:30		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.96	0.23	1	10/20/10 17:00	10/27/10 04:28	94-75-7	
Dinoseb	0.058U	ug/L	0.19	0.058	1	10/20/10 17:00	10/27/10 04:28	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.029	0.017	1	10/20/10 17:00	10/27/10 04:28	87-86-5	
2,4,5-T	0.043U	ug/L	0.19	0.043	1	10/20/10 17:00	10/27/10 04:28	93-76-5	
2,4,5-TP (Silvex)	0.050U	ug/L	0.19	0.050	1	10/20/10 17:00	10/27/10 04:28	93-72-1	
2,4-DCPA (S)	92 %		65.5-125.7		1	10/20/10 17:00	10/27/10 04:28	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	24.9	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:59	7440-38-2	
Barium	122	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:59	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:59	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:59	7440-43-9	
Chromium	3.4 I	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:59	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:59	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:59	7440-50-8	
Iron	71900	ug/L	40.0	20.0	1	10/21/10 06:45	10/21/10 23:59	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:59	7439-92-1	
Nickel	2.7 I	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:59	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/21/10 23:59	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/21/10 23:59	7440-22-4	
Sodium	173	mg/L	1.0	0.50	1	10/21/10 06:45	10/21/10 23:59	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/21/10 23:59	7440-31-5	
Vanadium	10.9	ug/L	10.0	5.0	1	10/21/10 06:45	10/21/10 23:59	7440-62-2	
Zinc	15.1 I	ug/L	20.0	10.0	1	10/21/10 06:45	10/21/10 23:59	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:09	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:09	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:16	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-16 **Lab ID: 3519325011** Collected: 10/13/10 16:17 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	0.88U	ug/L	5.1	0.88	1	10/20/10 18:20	10/22/10 20:24	83-32-9	
Acenaphthylene	0.97U	ug/L	5.1	0.97	1	10/20/10 18:20	10/22/10 20:24	208-96-8	
Acetophenone	1.5U	ug/L	5.1	1.5	1	10/20/10 18:20	10/22/10 20:24	98-86-2	
2-Acetylaminofluorene	0.66U	ug/L	5.1	0.66	1	10/20/10 18:20	10/22/10 20:24	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.1	2.9	1	10/20/10 18:20	10/22/10 20:24	92-67-1	
Anthracene	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 20:24	120-12-7	
Benzo(a)anthracene	0.64U	ug/L	5.1	0.64	1	10/20/10 18:20	10/22/10 20:24	56-55-3	
Benzo(a)pyrene	0.59U	ug/L	1.0	0.59	1	10/20/10 18:20	10/22/10 20:24	50-32-8	
Benzo(b)fluoranthene	0.63U	ug/L	2.0	0.63	1	10/20/10 18:20	10/22/10 20:24	205-99-2	
Benzo(g,h,i)perylene	0.69U	ug/L	5.1	0.69	1	10/20/10 18:20	10/22/10 20:24	191-24-2	
Benzo(k)fluoranthene	0.52U	ug/L	4.1	0.52	1	10/20/10 18:20	10/22/10 20:24	207-08-9	
Benzyl alcohol	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	100-51-6	
4-Bromophenylphenyl ether	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 20:24	101-55-3	
Butylbenzylphthalate	0.73U	ug/L	5.1	0.73	1	10/20/10 18:20	10/22/10 20:24	85-68-7	
4-Chloro-3-methylphenol	0.63U	ug/L	20.4	0.63	1	10/20/10 18:20	10/22/10 20:24	59-50-7	
4-Chloroaniline	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	106-47-8	
bis(2-Chloroethoxy)methane	3.0U	ug/L	5.1	3.0	1	10/20/10 18:20	10/22/10 20:24	111-91-1	
bis(2-Chloroethyl) ether	0.76U	ug/L	4.1	0.76	1	10/20/10 18:20	10/22/10 20:24	111-44-4	
bis(2-Chloroisopropyl) ether	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 20:24	108-60-1	
2-Chloronaphthalene	0.82U	ug/L	5.1	0.82	1	10/20/10 18:20	10/22/10 20:24	91-58-7	
2-Chlorophenol	0.69U	ug/L	5.1	0.69	1	10/20/10 18:20	10/22/10 20:24	95-57-8	
4-Chlorophenylphenyl ether	0.64U	ug/L	5.1	0.64	1	10/20/10 18:20	10/22/10 20:24	7005-72-3	
Chrysene	0.38U	ug/L	5.1	0.38	1	10/20/10 18:20	10/22/10 20:24	218-01-9	
Diallylate	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 20:24	2303-16-4	
Dibenz(a,h)anthracene	0.66U	ug/L	2.0	0.66	1	10/20/10 18:20	10/22/10 20:24	53-70-3	
Dibenzofuran	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 20:24	132-64-9	
1,2-Dichlorobenzene	0.69U	ug/L	5.1	0.69	1	10/20/10 18:20	10/22/10 20:24	95-50-1	
1,3-Dichlorobenzene	0.77U	ug/L	5.1	0.77	1	10/20/10 18:20	10/22/10 20:24	541-73-1	
1,4-Dichlorobenzene	0.78U	ug/L	5.1	0.78	1	10/20/10 18:20	10/22/10 20:24	106-46-7	
3,3'-Dichlorobenzidine	0.70U	ug/L	10.2	0.70	1	10/20/10 18:20	10/22/10 20:24	91-94-1	
2,4-Dichlorophenol	0.57U	ug/L	2.0	0.57	1	10/20/10 18:20	10/22/10 20:24	120-83-2	
2,6-Dichlorophenol	0.63U	ug/L	4.1	0.63	1	10/20/10 18:20	10/22/10 20:24	87-65-0	
Diethylphthalate	0.52U	ug/L	5.1	0.52	1	10/20/10 18:20	10/22/10 20:24	84-66-2	
P-Dimethylaminoazobenzene	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 20:24	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.1	2.0	1	10/20/10 18:20	10/22/10 20:24	57-97-6	
3,3'-Dimethylbenzidine	3.2U	ug/L	10.2	3.2	1	10/20/10 18:20	10/22/10 20:24	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.1	1.6	1	10/20/10 18:20	10/22/10 20:24	105-67-9	
a,a-Dimethylphenylethylamine	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 20:24	122-09-8	
Dimethylphthalate	0.65U	ug/L	5.1	0.65	1	10/20/10 18:20	10/22/10 20:24	131-11-3	
Di-n-butylphthalate	0.42U	ug/L	5.1	0.42	1	10/20/10 18:20	10/22/10 20:24	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	20.4	1.3	1	10/20/10 18:20	10/22/10 20:24	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	528-29-0	
1,3-Dinitrobenzene	0.69U	ug/L	8.2	0.69	1	10/20/10 18:20	10/22/10 20:24	99-65-0	
2,4-Dinitrophenol	1.6U	ug/L	20.4	1.6	1	10/20/10 18:20	10/22/10 20:24	51-28-5	
2,4-Dinitrotoluene	0.54U	ug/L	2.0	0.54	1	10/20/10 18:20	10/22/10 20:24	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	2.0	1.2	1	10/20/10 18:20	10/22/10 20:24	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 103 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-16 **Lab ID: 3519325011** Collected: 10/13/10 16:17 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	0.92U	ug/L	5.1	0.92	1	10/20/10 18:20	10/22/10 20:24	117-84-0	
bis(2-Ethylhexyl)phthalate	0.82U	ug/L	5.1	0.82	1	10/20/10 18:20	10/22/10 20:24	117-81-7	
Ethyl methanesulfonate	0.92U	ug/L	5.1	0.92	1	10/20/10 18:20	10/22/10 20:24	62-50-0	
Fluoranthene	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 20:24	206-44-0	
Fluorene	0.57U	ug/L	5.1	0.57	1	10/20/10 18:20	10/22/10 20:24	86-73-7	
Hexachlorobenzene	0.82U	ug/L	1.0	0.82	1	10/20/10 18:20	10/22/10 20:24	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.1	1.3	1	10/20/10 18:20	10/22/10 20:24	77-47-4	
Hexachloroethane	0.72U	ug/L	5.1	0.72	1	10/20/10 18:20	10/22/10 20:24	67-72-1	
Hexachloropropene	1.4U	ug/L	5.1	1.4	1	10/20/10 18:20	10/22/10 20:24	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.74U	ug/L	2.0	0.74	1	10/20/10 18:20	10/22/10 20:24	193-39-5	
Isodrin	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 20:24	465-73-6	
Isophorone	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 20:24	78-59-1	
Isosafrole	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 20:24	120-58-1	
Kepone	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 20:24	143-50-0	
Methapyrilene	1.7U	ug/L	5.1	1.7	1	10/20/10 18:20	10/22/10 20:24	91-80-5	J(SS)
3-Methylcholanthrene	1.1U	ug/L	5.1	1.1	1	10/20/10 18:20	10/22/10 20:24	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	91-57-6	
2-Methylphenol(o-Cresol)	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 20:24	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.67U	ug/L	10.2	0.67	1	10/20/10 18:20	10/22/10 20:24		
2-Naphthylamine	2.3U	ug/L	5.1	2.3	1	10/20/10 18:20	10/22/10 20:24	91-59-8	
Naphthalene	0.79U	ug/L	5.1	0.79	1	10/20/10 18:20	10/22/10 20:24	91-20-3	
1-Naphthylamine	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	134-32-7	
1,4-Naphthoquinone	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	130-15-4	
2-Nitroaniline	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 20:24	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 20:24	99-09-2	
4-Nitroaniline	0.70U	ug/L	4.1	0.70	1	10/20/10 18:20	10/22/10 20:24	100-01-6	
Nitrobenzene	1.1U	ug/L	4.1	1.1	1	10/20/10 18:20	10/22/10 20:24	98-95-3	
2-Nitrophenol	0.83U	ug/L	5.1	0.83	1	10/20/10 18:20	10/22/10 20:24	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.4	1.1	1	10/20/10 18:20	10/22/10 20:24	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.1	1.3	1	10/20/10 18:20	10/22/10 20:24	99-55-8	
N-Nitrosodiethylamine	0.74U	ug/L	4.1	0.74	1	10/20/10 18:20	10/22/10 20:24	55-18-5	
N-Nitrosodimethylamine	0.99U	ug/L	2.0	0.99	1	10/20/10 18:20	10/22/10 20:24	62-75-9	
N-Nitroso-di-n-butylamine	0.56U	ug/L	4.1	0.56	1	10/20/10 18:20	10/22/10 20:24	924-16-3	
N-Nitroso-di-n-propylamine	0.96U	ug/L	4.1	0.96	1	10/20/10 18:20	10/22/10 20:24	621-64-7	
N-Nitrosodiphenylamine	0.51U	ug/L	5.1	0.51	1	10/20/10 18:20	10/22/10 20:24	86-30-6	
N-Nitrosomethylethylamine	0.75U	ug/L	5.1	0.75	1	10/20/10 18:20	10/22/10 20:24	10595-95-6	
N-Nitrosopiperidine	0.65U	ug/L	5.1	0.65	1	10/20/10 18:20	10/22/10 20:24	100-75-4	
N-Nitrosopyrrolidine	0.90U	ug/L	5.1	0.90	1	10/20/10 18:20	10/22/10 20:24	930-55-2	
O,O,O-Triethylphosphorothioate	0.70U	ug/L	5.1	0.70	1	10/20/10 18:20	10/22/10 20:24	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	56-38-2	
Pentachlorobenzene	0.79U	ug/L	5.1	0.79	1	10/20/10 18:20	10/22/10 20:24	608-93-5	
Pentachlorophenol	0.67U	ug/L	20.4	0.67	1	10/20/10 18:20	10/22/10 20:24	87-86-5	
Phenacetin	0.54U	ug/L	5.1	0.54	1	10/20/10 18:20	10/22/10 20:24	62-44-2	
Phenanthrene	0.53U	ug/L	5.1	0.53	1	10/20/10 18:20	10/22/10 20:24	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 104 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-16 **Lab ID: 3519325011** Collected: 10/13/10 16:17 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Phenol	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 20:24	108-95-2	
p-Phenylenediamine	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 20:24	106-50-3	
Pronamide	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	23950-58-5	
Pyrene	0.69U	ug/L	5.1	0.69	1	10/20/10 18:20	10/22/10 20:24	129-00-0	
Safrole	0.87U	ug/L	5.1	0.87	1	10/20/10 18:20	10/22/10 20:24	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.71U	ug/L	5.1	0.71	1	10/20/10 18:20	10/22/10 20:24	95-94-3	
2,3,4,6-Tetrachlorophenol	3.9U	ug/L	5.1	3.9	1	10/20/10 18:20	10/22/10 20:24	58-90-2	
Thionazin	0.62U	ug/L	5.1	0.62	1	10/20/10 18:20	10/22/10 20:24	297-97-2	
O-Toluidine	1.1U	ug/L	5.1	1.1	1	10/20/10 18:20	10/22/10 20:24	95-53-4	
1,2,4-Trichlorobenzene	0.85U	ug/L	5.1	0.85	1	10/20/10 18:20	10/22/10 20:24	120-82-1	
2,4,5-Trichlorophenol	0.53U	ug/L	4.1	0.53	1	10/20/10 18:20	10/22/10 20:24	95-95-4	
2,4,6-Trichlorophenol	0.70U	ug/L	2.0	0.70	1	10/20/10 18:20	10/22/10 20:24	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 20:24	99-35-4	
Nitrobenzene-d5 (S)	70	%	10-110		1	10/20/10 18:20	10/22/10 20:24	4165-60-0	
2-Fluorobiphenyl (S)	75	%	18-110		1	10/20/10 18:20	10/22/10 20:24	321-60-8	
Terphenyl-d14 (S)	90	%	10-123		1	10/20/10 18:20	10/22/10 20:24	1718-51-0	
Phenol-d6 (S)	30	%	10-110		1	10/20/10 18:20	10/22/10 20:24	13127-88-3	
2-Fluorophenol (S)	42	%	18-110		1	10/20/10 18:20	10/22/10 20:24	367-12-4	
2,4,6-Tribromophenol (S)	84	%	10-110		1	10/20/10 18:20	10/22/10 20:24	118-79-6	
8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 12:32	83-32-9	
Acenaphthylene	0.051U	ug/L	2.0	0.051	1	10/20/10 13:02	10/21/10 12:32	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 12:32	120-12-7	
Benzo(a)anthracene	0.061U	ug/L	0.20	0.061	1	10/20/10 13:02	10/21/10 12:32	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.20	0.051	1	10/20/10 13:02	10/21/10 12:32	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/20/10 13:02	10/21/10 12:32	205-99-2	
Benzo(g,h,i)perylene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 12:32	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/20/10 13:02	10/21/10 12:32	207-08-9	
Chrysene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 12:32	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.20	0.051	1	10/20/10 13:02	10/21/10 12:32	53-70-3	
Fluoranthene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 12:32	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 12:32	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/20/10 13:02	10/21/10 12:32	193-39-5	
1-Methylnaphthalene	0.092U	ug/L	1.5	0.092	1	10/20/10 13:02	10/21/10 12:32	90-12-0	
2-Methylnaphthalene	0.061U	ug/L	1.5	0.061	1	10/20/10 13:02	10/21/10 12:32	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/20/10 13:02	10/21/10 12:32	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 12:32	85-01-8	
Pyrene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 12:32	129-00-0	
2-Fluorobiphenyl (S)	78	%	43.9-113		1	10/20/10 13:02	10/21/10 12:32	321-60-8	
Terphenyl-d14 (S)	80	%	24.8-144		1	10/20/10 13:02	10/21/10 12:32	1718-51-0	
8260 MSV Analytical Method: EPA 8260									
Acetone	5.9 I	ug/L	10.0	5.0	1		10/27/10 21:17	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/27/10 21:17	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 105 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-16 **Lab ID: 3519325011** Collected: 10/13/10 16:17 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	71-43-2	
Bromochloromethane	1.2	ug/L	1.0	0.50	1		10/27/10 21:17	74-97-5	
Bromodichloromethane	3.8	ug/L	0.60	0.27	1		10/27/10 21:17	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-00-3	
Chloroform	22.5	ug/L	1.0	0.50	1		10/27/10 21:17	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/27/10 21:17	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	126-99-8	
Dibromochloromethane	2.6	ug/L	0.50	0.26	1		10/27/10 21:17	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:17	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:17	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/27/10 21:17	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	126-98-7	
Methylene Chloride	2.6 I	ug/L	5.0	2.5	1		10/27/10 21:17	75-09-2	Z3
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:17	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/27/10 21:17	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 106 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-16		Lab ID: 3519325011		Collected: 10/13/10 16:17		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/27/10 21:17	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/27/10 21:17	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/27/10 21:17	1330-20-7	
4-Bromofluorobenzene (S)	96 %		70-114		1		10/27/10 21:17	460-00-4	
Dibromofluoromethane (S)	98 %		88-117		1		10/27/10 21:17	1868-53-7	
1,2-Dichloroethane-d4 (S)	111 %		86-125		1		10/27/10 21:17	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/27/10 21:17	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	930	mg/L	10.0	10.0	1		10/21/10 16:00		Q
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/20/10 10:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	158	mg/L	25.0	12.5	5		10/29/10 13:45	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 15:56	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	14.4	mg/L	0.050	0.020	1		10/22/10 13:33	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-15		Lab ID: 3519325012		Collected: 10/14/10 11:00		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.29	Std. Units			1		10/14/10 11:00		
Field Temperature	27.3	deg C			1		10/14/10 11:00		
Field Specific Conductance	3750	umhos/cm			1		10/14/10 11:00		
Oxygen, Dissolved	0.205	mg/L			1		10/14/10 11:00	7782-44-7	
Turbidity	19.9	NTU			1		10/14/10 11:00		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0048U	ug/L	0.020	0.0048	1	10/21/10 16:20	10/24/10 11:01	96-12-8	
1,2-Dibromoethane (EDB)	0.0061U	ug/L	0.0098	0.0061	1	10/21/10 16:20	10/24/10 11:01	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00050U	ug/L	0.0099	0.00050	1	10/20/10 12:00	11/12/10 22:53	309-00-2	
alpha-BHC	0.00030U	ug/L	0.0099	0.00030	1	10/20/10 12:00	11/12/10 22:53	319-84-6	
beta-BHC	0.00050U	ug/L	0.0099	0.00050	1	10/20/10 12:00	11/12/10 22:53	319-85-7	
delta-BHC	0.00040U	ug/L	0.0099	0.00040	1	10/20/10 12:00	11/12/10 22:53	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.0099	0.00020	1	10/20/10 12:00	11/12/10 22:53	58-89-9	
Chlordane (Technical)	0.079U	ug/L	0.50	0.079	1	10/20/10 12:00	11/12/10 22:53	57-74-9	
Chlorobenzilate	0.021U	ug/L	0.099	0.021	1	10/20/10 12:00	11/12/10 22:53	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.0099	0.0019	1	10/20/10 12:00	11/12/10 22:53	72-54-8	
4,4'-DDE	0.00089U	ug/L	0.0099	0.00089	1	10/20/10 12:00	11/12/10 22:53	72-55-9	
4,4'-DDT	0.0036U	ug/L	0.0099	0.0036	1	10/20/10 12:00	11/12/10 22:53	50-29-3	
Dieldrin	0.00050U	ug/L	0.0099	0.00050	1	10/20/10 12:00	11/12/10 22:53	60-57-1	
Endosulfan I	0.00069U	ug/L	0.0099	0.00069	1	10/20/10 12:00	11/12/10 22:53	959-98-8	
Endosulfan II	0.00069U	ug/L	0.0099	0.00069	1	10/20/10 12:00	11/12/10 22:53	33213-65-9	
Endosulfan sulfate	0.00060U	ug/L	0.0099	0.00060	1	10/20/10 12:00	11/12/10 22:53	1031-07-8	
Endrin	0.0017U	ug/L	0.0099	0.0017	1	10/20/10 12:00	11/12/10 22:53	72-20-8	
Endrin aldehyde	0.0070U	ug/L	0.0099	0.0070	1	10/20/10 12:00	11/12/10 22:53	7421-93-4	
Heptachlor	0.0015U	ug/L	0.0099	0.0015	1	10/20/10 12:00	11/12/10 22:53	76-44-8	
Heptachlor epoxide	0.00040U	ug/L	0.0099	0.00040	1	10/20/10 12:00	11/12/10 22:53	1024-57-3	
Methoxychlor	0.0069U	ug/L	0.0099	0.0069	1	10/20/10 12:00	11/12/10 22:53	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.099	0.015	1	10/20/10 12:00	11/12/10 22:53	82-68-8	
Toxaphene	0.28U	ug/L	0.50	0.28	1	10/20/10 12:00	11/12/10 22:53	8001-35-2	
Tetrachloro-m-xylene (S)	78 %		66.5-120.3		1	10/20/10 12:00	11/12/10 22:53	877-09-8	
Decachlorobiphenyl (S)	22 %		41.7-109.1		1	10/20/10 12:00	11/12/10 22:53	2051-24-3	1p, J(S5)
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.079U	ug/L	0.50	0.079	1	10/20/10 12:01	11/12/10 22:53	12674-11-2	
PCB-1221 (Aroclor 1221)	0.080U	ug/L	0.50	0.080	1	10/20/10 12:01	11/12/10 22:53	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.50	0.12	1	10/20/10 12:01	11/12/10 22:53	11141-16-5	
PCB-1242 (Aroclor 1242)	0.12U	ug/L	0.50	0.12	1	10/20/10 12:01	11/12/10 22:53	53469-21-9	
PCB-1248 (Aroclor 1248)	0.27U	ug/L	0.50	0.27	1	10/20/10 12:01	11/12/10 22:53	12672-29-6	
PCB-1254 (Aroclor 1254)	0.14U	ug/L	0.50	0.14	1	10/20/10 12:01	11/12/10 22:53	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.50	0.11	1	10/20/10 12:01	11/12/10 22:53	11096-82-5	
Tetrachloro-m-xylene (S)	78 %		48-111		1	10/20/10 12:01	11/12/10 22:53	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 108 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-15		Lab ID: 3519325012		Collected: 10/14/10 11:00		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	22 %		63-121		1	10/20/10 12:01	11/12/10 22:53	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.51	0.19	1	10/20/10 16:08	11/02/10 01:44	60-51-5	
Disulfoton	0.16U	ug/L	0.51	0.16	1	10/20/10 16:08	11/02/10 01:44	298-04-4	
Famphur	0.15U	ug/L	0.51	0.15	1	10/20/10 16:08	11/02/10 01:44	52-85-7	
Methyl parathion	0.20U	ug/L	0.51	0.20	1	10/20/10 16:08	11/02/10 01:44	298-00-0	
Parathion (Ethyl parathion)	0.36U	ug/L	1.0	0.36	1	10/20/10 16:08	11/02/10 01:44	56-38-2	
Phorate	0.37U	ug/L	1.0	0.37	1	10/20/10 16:08	11/02/10 01:44	298-02-2	
4-Chloro3nitrobenzotrifluoride	75 %		34.2-122		1	10/20/10 16:08	11/02/10 01:44		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.95	0.23	1	10/20/10 17:00	10/27/10 04:54	94-75-7	
Dinoseb	0.058U	ug/L	0.19	0.058	1	10/20/10 17:00	10/27/10 04:54	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.029	0.017	1	10/20/10 17:00	10/27/10 04:54	87-86-5	
2,4,5-T	0.042U	ug/L	0.19	0.042	1	10/20/10 17:00	10/27/10 04:54	93-76-5	
2,4,5-TP (Silvex)	0.049U	ug/L	0.19	0.049	1	10/20/10 17:00	10/27/10 04:54	93-72-1	
2,4-DCPA (S)	84 %		65.5-125.7		1	10/20/10 17:00	10/27/10 04:54	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	48.6	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:02	7440-38-2	
Barium	256	ug/L	50.0	25.0	5	10/21/10 06:45	10/25/10 02:23	7440-39-3	D3
Beryllium	2.5U	ug/L	5.0	2.5	5	10/21/10 06:45	10/25/10 02:23	7440-41-7	D3
Cadmium	2.5U	ug/L	5.0	2.5	5	10/21/10 06:45	10/25/10 02:23	7440-43-9	D3
Calcium	725	mg/L	2.5	1.2	5	10/21/10 06:45	10/25/10 02:23	7440-70-2	
Chromium	12.5U	ug/L	25.0	12.5	5	10/21/10 06:45	10/25/10 02:23	7440-47-3	D3
Cobalt	38.9	ug/L	50.0	25.0	5	10/21/10 06:45	10/25/10 02:23	7440-48-4	D3
Copper	12.5U	ug/L	25.0	12.5	5	10/21/10 06:45	10/25/10 02:23	7440-50-8	D3
Iron	49600	ug/L	200	100	5	10/21/10 06:45	10/25/10 02:23	7439-89-6	D3
Lead	25.0U	ug/L	50.0	25.0	5	10/21/10 06:45	10/25/10 02:23	7439-92-1	D3
Magnesium	164	mg/L	2.5	1.2	5	10/21/10 06:45	10/25/10 02:23	7439-95-4	
Manganese	995	ug/L	25.0	12.5	5	10/21/10 06:45	10/25/10 02:23	7439-96-5	CH,D3
Nickel	17.2	ug/L	25.0	12.5	5	10/21/10 06:45	10/25/10 02:23	7440-02-0	D3
Potassium	22.4	mg/L	5.0	2.5	5	10/21/10 06:45	10/25/10 02:23	7440-09-7	
Selenium	37.5U	ug/L	75.0	37.5	5	10/21/10 06:45	10/25/10 02:23	7782-49-2	D3
Silver	12.5U	ug/L	25.0	12.5	5	10/21/10 06:45	10/25/10 02:23	7440-22-4	D3
Sodium	77.5	mg/L	5.0	2.5	5	10/21/10 06:45	10/25/10 02:23	7440-23-5	D3
Tin	125U	ug/L	250	125	5	10/21/10 06:45	10/25/10 02:23	7440-31-5	D3
Vanadium	25.0U	ug/L	50.0	25.0	5	10/21/10 06:45	10/25/10 02:23	7440-62-2	D3
Zinc	50.0U	ug/L	100	50.0	5	10/21/10 06:45	10/25/10 02:23	7440-66-6	D3
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:37	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:37	7440-28-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 109 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-15		Lab ID: 3519325012	Collected: 10/14/10 11:00	Received: 10/19/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:19	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.86U	ug/L	5.0	0.86	1	10/20/10 18:20	10/22/10 20:55	83-32-9	
Acenaphthylene	0.96U	ug/L	5.0	0.96	1	10/20/10 18:20	10/22/10 20:55	208-96-8	
Acetophenone	1.5U	ug/L	5.0	1.5	1	10/20/10 18:20	10/22/10 20:55	98-86-2	
2-Acetylaminofluorene	0.65U	ug/L	5.0	0.65	1	10/20/10 18:20	10/22/10 20:55	53-96-3	
4-Aminobiphenyl	2.8U	ug/L	5.0	2.8	1	10/20/10 18:20	10/22/10 20:55	92-67-1	
Anthracene	0.60U	ug/L	5.0	0.60	1	10/20/10 18:20	10/22/10 20:55	120-12-7	
Benzo(a)anthracene	0.63U	ug/L	5.0	0.63	1	10/20/10 18:20	10/22/10 20:55	56-55-3	
Benzo(a)pyrene	0.58U	ug/L	1.0	0.58	1	10/20/10 18:20	10/22/10 20:55	50-32-8	
Benzo(b)fluoranthene	0.62U	ug/L	2.0	0.62	1	10/20/10 18:20	10/22/10 20:55	205-99-2	
Benzo(g,h,i)perylene	0.68U	ug/L	5.0	0.68	1	10/20/10 18:20	10/22/10 20:55	191-24-2	
Benzo(k)fluoranthene	0.51U	ug/L	4.0	0.51	1	10/20/10 18:20	10/22/10 20:55	207-08-9	
Benzyl alcohol	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	100-51-6	
4-Bromophenylphenyl ether	0.67U	ug/L	5.0	0.67	1	10/20/10 18:20	10/22/10 20:55	101-55-3	
Butylbenzylphthalate	0.72U	ug/L	5.0	0.72	1	10/20/10 18:20	10/22/10 20:55	85-68-7	
4-Chloro-3-methylphenol	0.62U	ug/L	20.1	0.62	1	10/20/10 18:20	10/22/10 20:55	59-50-7	
4-Chloroaniline	1.2U	ug/L	5.0	1.2	1	10/20/10 18:20	10/22/10 20:55	106-47-8	
bis(2-Chloroethoxy)methane	3.0U	ug/L	5.0	3.0	1	10/20/10 18:20	10/22/10 20:55	111-91-1	
bis(2-Chloroethyl) ether	0.75U	ug/L	4.0	0.75	1	10/20/10 18:20	10/22/10 20:55	111-44-4	
bis(2-Chloroisopropyl) ether	0.73U	ug/L	5.0	0.73	1	10/20/10 18:20	10/22/10 20:55	108-60-1	
2-Chloronaphthalene	0.80U	ug/L	5.0	0.80	1	10/20/10 18:20	10/22/10 20:55	91-58-7	
2-Chlorophenol	0.68U	ug/L	5.0	0.68	1	10/20/10 18:20	10/22/10 20:55	95-57-8	
4-Chlorophenylphenyl ether	0.63U	ug/L	5.0	0.63	1	10/20/10 18:20	10/22/10 20:55	7005-72-3	
Chrysene	0.37U	ug/L	5.0	0.37	1	10/20/10 18:20	10/22/10 20:55	218-01-9	
Diallate	0.73U	ug/L	5.0	0.73	1	10/20/10 18:20	10/22/10 20:55	2303-16-4	
Dibenz(a,h)anthracene	0.65U	ug/L	2.0	0.65	1	10/20/10 18:20	10/22/10 20:55	53-70-3	
Dibenzofuran	0.67U	ug/L	5.0	0.67	1	10/20/10 18:20	10/22/10 20:55	132-64-9	
1,2-Dichlorobenzene	0.68U	ug/L	5.0	0.68	1	10/20/10 18:20	10/22/10 20:55	95-50-1	
1,3-Dichlorobenzene	0.76U	ug/L	5.0	0.76	1	10/20/10 18:20	10/22/10 20:55	541-73-1	
1,4-Dichlorobenzene	0.77U	ug/L	5.0	0.77	1	10/20/10 18:20	10/22/10 20:55	106-46-7	
3,3'-Dichlorobenzidine	0.69U	ug/L	10.1	0.69	1	10/20/10 18:20	10/22/10 20:55	91-94-1	
2,4-Dichlorophenol	0.56U	ug/L	2.0	0.56	1	10/20/10 18:20	10/22/10 20:55	120-83-2	
2,6-Dichlorophenol	0.62U	ug/L	4.0	0.62	1	10/20/10 18:20	10/22/10 20:55	87-65-0	
Diethylphthalate	0.51U	ug/L	5.0	0.51	1	10/20/10 18:20	10/22/10 20:55	84-66-2	
P-Dimethylaminoazobenzene	0.67U	ug/L	5.0	0.67	1	10/20/10 18:20	10/22/10 20:55	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.0	2.0	1	10/20/10 18:20	10/22/10 20:55	57-97-6	
3,3'-Dimethylbenzidine	3.1U	ug/L	10.1	3.1	1	10/20/10 18:20	10/22/10 20:55	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.0	1.6	1	10/20/10 18:20	10/22/10 20:55	105-67-9	
a,a-Dimethylphenylethylamine	10.1U	ug/L	20.1	10.1	1	10/20/10 18:20	10/22/10 20:55	122-09-8	
Dimethylphthalate	0.64U	ug/L	5.0	0.64	1	10/20/10 18:20	10/22/10 20:55	131-11-3	
Di-n-butylphthalate	0.41U	ug/L	5.0	0.41	1	10/20/10 18:20	10/22/10 20:55	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	20.1	1.3	1	10/20/10 18:20	10/22/10 20:55	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.0	1.2	1	10/20/10 18:20	10/22/10 20:55	528-29-0	
1,3-Dinitrobenzene	0.68U	ug/L	8.0	0.68	1	10/20/10 18:20	10/22/10 20:55	99-65-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 110 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-15 **Lab ID: 3519325012** Collected: 10/14/10 11:00 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrophenol	1.6U	ug/L	20.1	1.6	1	10/20/10 18:20	10/22/10 20:55	51-28-5	
2,4-Dinitrotoluene	0.53U	ug/L	2.0	0.53	1	10/20/10 18:20	10/22/10 20:55	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	2.0	1.2	1	10/20/10 18:20	10/22/10 20:55	606-20-2	
Di-n-octylphthalate	0.90U	ug/L	5.0	0.90	1	10/20/10 18:20	10/22/10 20:55	117-84-0	
bis(2-Ethylhexyl)phthalate	0.80U	ug/L	5.0	0.80	1	10/20/10 18:20	10/22/10 20:55	117-81-7	
Ethyl methanesulfonate	0.90U	ug/L	5.0	0.90	1	10/20/10 18:20	10/22/10 20:55	62-50-0	
Fluoranthene	0.54U	ug/L	5.0	0.54	1	10/20/10 18:20	10/22/10 20:55	206-44-0	
Fluorene	0.56U	ug/L	5.0	0.56	1	10/20/10 18:20	10/22/10 20:55	86-73-7	
Hexachlorobenzene	0.80U	ug/L	1.0	0.80	1	10/20/10 18:20	10/22/10 20:55	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.0	1.3	1	10/20/10 18:20	10/22/10 20:55	77-47-4	
Hexachloroethane	0.71U	ug/L	5.0	0.71	1	10/20/10 18:20	10/22/10 20:55	67-72-1	
Hexachloropropene	1.4U	ug/L	5.0	1.4	1	10/20/10 18:20	10/22/10 20:55	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.73U	ug/L	2.0	0.73	1	10/20/10 18:20	10/22/10 20:55	193-39-5	
Isodrin	0.54U	ug/L	5.0	0.54	1	10/20/10 18:20	10/22/10 20:55	465-73-6	
Isophorone	0.73U	ug/L	5.0	0.73	1	10/20/10 18:20	10/22/10 20:55	78-59-1	
Isosafrole	0.60U	ug/L	5.0	0.60	1	10/20/10 18:20	10/22/10 20:55	120-58-1	
Kepone	10.1U	ug/L	20.1	10.1	1	10/20/10 18:20	10/22/10 20:55	143-50-0	
Methapyrilene	1.7U	ug/L	5.0	1.7	1	10/20/10 18:20	10/22/10 20:55	91-80-5	J(SS)
3-Methylcholanthrene	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	91-57-6	
2-Methylphenol(o-Cresol)	0.73U	ug/L	5.0	0.73	1	10/20/10 18:20	10/22/10 20:55	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.66U	ug/L	10.1	0.66	1	10/20/10 18:20	10/22/10 20:55		
2-Naphthylamine	2.3U	ug/L	5.0	2.3	1	10/20/10 18:20	10/22/10 20:55	91-59-8	
Naphthalene	0.78U	ug/L	5.0	0.78	1	10/20/10 18:20	10/22/10 20:55	91-20-3	
1-Naphthylamine	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	134-32-7	
1,4-Naphthoquinone	1.2U	ug/L	5.0	1.2	1	10/20/10 18:20	10/22/10 20:55	130-15-4	
2-Nitroaniline	0.60U	ug/L	5.0	0.60	1	10/20/10 18:20	10/22/10 20:55	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.0	1.0	1	10/20/10 18:20	10/22/10 20:55	99-09-2	
4-Nitroaniline	0.69U	ug/L	4.0	0.69	1	10/20/10 18:20	10/22/10 20:55	100-01-6	
Nitrobenzene	1.1U	ug/L	4.0	1.1	1	10/20/10 18:20	10/22/10 20:55	98-95-3	
2-Nitrophenol	0.81U	ug/L	5.0	0.81	1	10/20/10 18:20	10/22/10 20:55	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.1	1.1	1	10/20/10 18:20	10/22/10 20:55	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.0	1.3	1	10/20/10 18:20	10/22/10 20:55	99-55-8	
N-Nitrosodiethylamine	0.73U	ug/L	4.0	0.73	1	10/20/10 18:20	10/22/10 20:55	55-18-5	
N-Nitrosodimethylamine	0.98U	ug/L	2.0	0.98	1	10/20/10 18:20	10/22/10 20:55	62-75-9	
N-Nitroso-di-n-butylamine	0.55U	ug/L	4.0	0.55	1	10/20/10 18:20	10/22/10 20:55	924-16-3	
N-Nitroso-di-n-propylamine	0.95U	ug/L	4.0	0.95	1	10/20/10 18:20	10/22/10 20:55	621-64-7	
N-Nitrosodiphenylamine	0.50U	ug/L	5.0	0.50	1	10/20/10 18:20	10/22/10 20:55	86-30-6	
N-Nitrosomethylethylamine	0.74U	ug/L	5.0	0.74	1	10/20/10 18:20	10/22/10 20:55	10595-95-6	
N-Nitrosopiperidine	0.64U	ug/L	5.0	0.64	1	10/20/10 18:20	10/22/10 20:55	100-75-4	
N-Nitrosopyrrolidine	0.88U	ug/L	5.0	0.88	1	10/20/10 18:20	10/22/10 20:55	930-55-2	
O,O,O-Triethylphosphorothioate	0.69U	ug/L	5.0	0.69	1	10/20/10 18:20	10/22/10 20:55	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.0	1.2	1	10/20/10 18:20	10/22/10 20:55	56-38-2	
Pentachlorobenzene	0.78U	ug/L	5.0	0.78	1	10/20/10 18:20	10/22/10 20:55	608-93-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 111 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-15 **Lab ID: 3519325012** Collected: 10/14/10 11:00 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Pentachlorophenol	0.66U	ug/L	20.1	0.66	1	10/20/10 18:20	10/22/10 20:55	87-86-5	
Phenacetin	0.53U	ug/L	5.0	0.53	1	10/20/10 18:20	10/22/10 20:55	62-44-2	
Phenanthrene	0.52U	ug/L	5.0	0.52	1	10/20/10 18:20	10/22/10 20:55	85-01-8	
Phenol	0.54U	ug/L	5.0	0.54	1	10/20/10 18:20	10/22/10 20:55	108-95-2	
p-Phenylenediamine	10.1U	ug/L	20.1	10.1	1	10/20/10 18:20	10/22/10 20:55	106-50-3	
Pronamide	1.1U	ug/L	5.0	1.1	1	10/20/10 18:20	10/22/10 20:55	23950-58-5	
Pyrene	0.68U	ug/L	5.0	0.68	1	10/20/10 18:20	10/22/10 20:55	129-00-0	
Safrole	0.85U	ug/L	5.0	0.85	1	10/20/10 18:20	10/22/10 20:55	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.70U	ug/L	5.0	0.70	1	10/20/10 18:20	10/22/10 20:55	95-94-3	
2,3,4,6-Tetrachlorophenol	3.9U	ug/L	5.0	3.9	1	10/20/10 18:20	10/22/10 20:55	58-90-2	
Thionazin	0.61U	ug/L	5.0	0.61	1	10/20/10 18:20	10/22/10 20:55	297-97-2	
O-Toluidine	1.1U	ug/L	5.0	1.1	1	10/20/10 18:20	10/22/10 20:55	95-53-4	
1,2,4-Trichlorobenzene	0.83U	ug/L	5.0	0.83	1	10/20/10 18:20	10/22/10 20:55	120-82-1	
2,4,5-Trichlorophenol	0.52U	ug/L	4.0	0.52	1	10/20/10 18:20	10/22/10 20:55	95-95-4	
2,4,6-Trichlorophenol	0.69U	ug/L	2.0	0.69	1	10/20/10 18:20	10/22/10 20:55	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	5.0	1.2	1	10/20/10 18:20	10/22/10 20:55	99-35-4	
Nitrobenzene-d5 (S)	59 %		10-110		1	10/20/10 18:20	10/22/10 20:55	4165-60-0	
2-Fluorobiphenyl (S)	66 %		18-110		1	10/20/10 18:20	10/22/10 20:55	321-60-8	
Terphenyl-d14 (S)	86 %		10-123		1	10/20/10 18:20	10/22/10 20:55	1718-51-0	
Phenol-d6 (S)	26 %		10-110		1	10/20/10 18:20	10/22/10 20:55	13127-88-3	
2-Fluorophenol (S)	37 %		18-110		1	10/20/10 18:20	10/22/10 20:55	367-12-4	
2,4,6-Tribromophenol (S)	77 %		10-110		1	10/20/10 18:20	10/22/10 20:55	118-79-6	

8270 MSSV PAH by SCAN

Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510

Acenaphthene	0.030U	ug/L	1.0	0.030	1	10/20/10 13:02	10/21/10 12:54	83-32-9	
Acenaphthylene	0.050U	ug/L	2.0	0.050	1	10/20/10 13:02	10/21/10 12:54	208-96-8	
Anthracene	0.050U	ug/L	1.0	0.050	1	10/20/10 13:02	10/21/10 12:54	120-12-7	
Benzo(a)anthracene	0.060U	ug/L	0.20	0.060	1	10/20/10 13:02	10/21/10 12:54	56-55-3	
Benzo(a)pyrene	0.050U	ug/L	0.20	0.050	1	10/20/10 13:02	10/21/10 12:54	50-32-8	
Benzo(b)fluoranthene	0.050U	ug/L	0.10	0.050	1	10/20/10 13:02	10/21/10 12:54	205-99-2	
Benzo(g,h,i)perylene	0.060U	ug/L	1.0	0.060	1	10/20/10 13:02	10/21/10 12:54	191-24-2	
Benzo(k)fluoranthene	0.040U	ug/L	0.25	0.040	1	10/20/10 13:02	10/21/10 12:54	207-08-9	
Chrysene	0.060U	ug/L	1.0	0.060	1	10/20/10 13:02	10/21/10 12:54	218-01-9	
Dibenz(a,h)anthracene	0.050U	ug/L	0.20	0.050	1	10/20/10 13:02	10/21/10 12:54	53-70-3	
Fluoranthene	0.060U	ug/L	1.0	0.060	1	10/20/10 13:02	10/21/10 12:54	206-44-0	
Fluorene	0.030U	ug/L	1.0	0.030	1	10/20/10 13:02	10/21/10 12:54	86-73-7	
Indeno(1,2,3-cd)pyrene	0.040U	ug/L	0.15	0.040	1	10/20/10 13:02	10/21/10 12:54	193-39-5	
1-Methylnaphthalene	0.090U	ug/L	1.5	0.090	1	10/20/10 13:02	10/21/10 12:54	90-12-0	
2-Methylnaphthalene	0.060U	ug/L	1.5	0.060	1	10/20/10 13:02	10/21/10 12:54	91-57-6	
Naphthalene	0.080U	ug/L	1.0	0.080	1	10/20/10 13:02	10/21/10 12:54	91-20-3	
Phenanthrene	0.050U	ug/L	1.0	0.050	1	10/20/10 13:02	10/21/10 12:54	85-01-8	
Pyrene	0.060U	ug/L	1.0	0.060	1	10/20/10 13:02	10/21/10 12:54	129-00-0	
2-Fluorobiphenyl (S)	70 %		43.9-113		1	10/20/10 13:02	10/21/10 12:54	321-60-8	
Terphenyl-d14 (S)	42 %		24.8-144		1	10/20/10 13:02	10/21/10 12:54	1718-51-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-15 **Lab ID: 3519325012** Collected: 10/14/10 11:00 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.21	ug/L	10.0	5.0	1		10/28/10 04:46	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/28/10 04:46	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/28/10 04:46	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 04:46	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/28/10 04:46	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 04:46	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 04:46	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/28/10 04:46	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 04:46	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 04:46	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 113 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-15		Lab ID: 3519325012		Collected: 10/14/10 11:00		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 04:46	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 04:46	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 04:46	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 04:46	1330-20-7	
4-Bromofluorobenzene (S)	96 %		70-114		1		10/28/10 04:46	460-00-4	p2
Dibromofluoromethane (S)	101 %		88-117		1		10/28/10 04:46	1868-53-7	
1,2-Dichloroethane-d4 (S)	111 %		86-125		1		10/28/10 04:46	17060-07-0	
Toluene-d8 (S)	100 %		87-113		1		10/28/10 04:46	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	2810	mg/L	20.0	20.0	1		10/21/10 16:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.8	mg/L	1.0	1.0	1		10/20/10 10:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	131	mg/L	100	50.0	20		11/01/10 20:03	16887-00-6	
Sulfate	392	mg/L	100	50.0	20		11/01/10 20:03	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 15:59	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	5.0	mg/L	0.050	0.020	1		10/22/10 13:35	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-16		Lab ID: 3519325013		Collected: 10/15/10 11:55		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.32	Std. Units			1		10/15/10 11:55		
Field Temperature	26.3	deg C			1		10/15/10 11:55		
Field Specific Conductance	2713	umhos/cm			1		10/15/10 11:55		
Oxygen, Dissolved	0.731	mg/L			1		10/15/10 11:55	7782-44-7	
Turbidity	16.4	NTU			1		10/15/10 11:55		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/27/10 16:00	10/28/10 13:09	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.0099	0.0062	1	10/27/10 16:00	10/28/10 13:09	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 23:10	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/20/10 12:00	11/12/10 23:10	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 23:10	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/20/10 12:00	11/12/10 23:10	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	10/20/10 12:00	11/12/10 23:10	58-89-9	
Chlordane (Technical)	0.082U	ug/L	0.51	0.082	1	10/20/10 12:00	11/12/10 23:10	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/20/10 12:00	11/12/10 23:10	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/20/10 12:00	11/12/10 23:10	72-54-8	
4,4'-DDE	0.00092U	ug/L	0.010	0.00092	1	10/20/10 12:00	11/12/10 23:10	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/20/10 12:00	11/12/10 23:10	50-29-3	
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	10/20/10 12:00	11/12/10 23:10	60-57-1	
Endosulfan I	0.00072U	ug/L	0.010	0.00072	1	10/20/10 12:00	11/12/10 23:10	959-98-8	
Endosulfan II	0.00072U	ug/L	0.010	0.00072	1	10/20/10 12:00	11/12/10 23:10	33213-65-9	
Endosulfan sulfate	0.00061U	ug/L	0.010	0.00061	1	10/20/10 12:00	11/12/10 23:10	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/20/10 12:00	11/12/10 23:10	72-20-8	
Endrin aldehyde	0.011	ug/L	0.010	0.0073	1	10/20/10 12:00	11/12/10 23:10	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/20/10 12:00	11/12/10 23:10	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/20/10 12:00	11/12/10 23:10	1024-57-3	
Methoxychlor	0.0072U	ug/L	0.010	0.0072	1	10/20/10 12:00	11/12/10 23:10	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/20/10 12:00	11/12/10 23:10	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	10/20/10 12:00	11/12/10 23:10	8001-35-2	
Tetrachloro-m-xylene (S)	66 %		66.5-120.3		1	10/20/10 12:00	11/12/10 23:10	877-09-8	1p, J(S5)
Decachlorobiphenyl (S)	59 %		41.7-109.1		1	10/20/10 12:00	11/12/10 23:10	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.082U	ug/L	0.51	0.082	1	10/20/10 12:01	11/12/10 23:10	12674-11-2	
PCB-1221 (Aroclor 1221)	0.083U	ug/L	0.51	0.083	1	10/20/10 12:01	11/12/10 23:10	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	10/20/10 12:01	11/12/10 23:10	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	10/20/10 12:01	11/12/10 23:10	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	10/20/10 12:01	11/12/10 23:10	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	10/20/10 12:01	11/12/10 23:10	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	10/20/10 12:01	11/12/10 23:10	11096-82-5	
Tetrachloro-m-xylene (S)	66 %		48-111		1	10/20/10 12:01	11/12/10 23:10	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 115 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-16 Lab ID: 3519325013 Collected: 10/15/10 11:55 Received: 10/19/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	59 %		63-121		1	10/20/10 12:01	11/12/10 23:10	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.51	0.19	1	10/20/10 16:08	11/02/10 02:26	60-51-5	
Disulfoton	0.16U	ug/L	0.51	0.16	1	10/20/10 16:08	11/02/10 02:26	298-04-4	
Famphur	0.15U	ug/L	0.51	0.15	1	10/20/10 16:08	11/02/10 02:26	52-85-7	
Methyl parathion	0.20U	ug/L	0.51	0.20	1	10/20/10 16:08	11/02/10 02:26	298-00-0	
Parathion (Ethyl parathion)	0.36U	ug/L	1.0	0.36	1	10/20/10 16:08	11/02/10 02:26	56-38-2	
Phorate	0.38U	ug/L	1.0	0.38	1	10/20/10 16:08	11/02/10 02:26	298-02-2	
4-Chloro3nitrobenzotrifluoride	68 %		34.2-122		1	10/20/10 16:08	11/02/10 02:26		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.97	0.23	1	10/20/10 17:00	10/27/10 05:21	94-75-7	
Dinoseb	0.059U	ug/L	0.19	0.059	1	10/20/10 17:00	10/27/10 05:21	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.029	0.017	1	10/20/10 17:00	10/27/10 05:21	87-86-5	
2,4,5-T	0.043U	ug/L	0.19	0.043	1	10/20/10 17:00	10/27/10 05:21	93-76-5	
2,4,5-TP (Silvex)	0.050U	ug/L	0.20	0.050	1	10/20/10 17:00	10/27/10 05:21	93-72-1	
2,4-DCPA (S)	77 %		65.5-125.7		1	10/20/10 17:00	10/27/10 05:21	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	48.5	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:15	7440-38-2	
Barium	119	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:15	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:15	7440-41-7	
Cadmium	0.68	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:15	7440-43-9	
Calcium	234	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:15	7440-70-2	
Chromium	3.8	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:15	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:15	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:15	7440-50-8	
Iron	62000	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:15	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:15	7439-92-1	
Magnesium	84.0	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:15	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:15	7440-02-0	
Potassium	6.7	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:15	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:15	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:15	7440-22-4	
Sodium	275	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:15	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/22/10 00:15	7440-31-5	
Vanadium	10.4	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:15	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:15	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:42	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:42	7440-28-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 116 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-16		Lab ID: 3519325013	Collected: 10/15/10 11:55	Received: 10/19/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:22	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.88U	ug/L	5.1	0.88	1	10/20/10 18:20	10/22/10 21:26	83-32-9	
Acenaphthylene	0.97U	ug/L	5.1	0.97	1	10/20/10 18:20	10/22/10 21:26	208-96-8	
Acetophenone	1.5U	ug/L	5.1	1.5	1	10/20/10 18:20	10/22/10 21:26	98-86-2	
2-Acetylaminofluorene	0.66U	ug/L	5.1	0.66	1	10/20/10 18:20	10/22/10 21:26	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.1	2.9	1	10/20/10 18:20	10/22/10 21:26	92-67-1	
Anthracene	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 21:26	120-12-7	
Benzo(a)anthracene	0.64U	ug/L	5.1	0.64	1	10/20/10 18:20	10/22/10 21:26	56-55-3	
Benzo(a)pyrene	0.59U	ug/L	1.0	0.59	1	10/20/10 18:20	10/22/10 21:26	50-32-8	
Benzo(b)fluoranthene	0.63U	ug/L	2.0	0.63	1	10/20/10 18:20	10/22/10 21:26	205-99-2	
Benzo(g,h,i)perylene	0.70U	ug/L	5.1	0.70	1	10/20/10 18:20	10/22/10 21:26	191-24-2	
Benzo(k)fluoranthene	0.52U	ug/L	4.1	0.52	1	10/20/10 18:20	10/22/10 21:26	207-08-9	
Benzyl alcohol	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 21:26	100-51-6	
4-Bromophenylphenyl ether	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 21:26	101-55-3	
Butylbenzylphthalate	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 21:26	85-68-7	
4-Chloro-3-methylphenol	0.63U	ug/L	20.4	0.63	1	10/20/10 18:20	10/22/10 21:26	59-50-7	
4-Chloroaniline	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	106-47-8	
bis(2-Chloroethoxy)methane	3.0U	ug/L	5.1	3.0	1	10/20/10 18:20	10/22/10 21:26	111-91-1	
bis(2-Chloroethyl) ether	0.77U	ug/L	4.1	0.77	1	10/20/10 18:20	10/22/10 21:26	111-44-4	
bis(2-Chloroisopropyl) ether	0.75U	ug/L	5.1	0.75	1	10/20/10 18:20	10/22/10 21:26	108-60-1	
2-Chloronaphthalene	0.82U	ug/L	5.1	0.82	1	10/20/10 18:20	10/22/10 21:26	91-58-7	
2-Chlorophenol	0.70U	ug/L	5.1	0.70	1	10/20/10 18:20	10/22/10 21:26	95-57-8	
4-Chlorophenylphenyl ether	0.64U	ug/L	5.1	0.64	1	10/20/10 18:20	10/22/10 21:26	7005-72-3	
Chrysene	0.38U	ug/L	5.1	0.38	1	10/20/10 18:20	10/22/10 21:26	218-01-9	
Diallate	0.74U	ug/L	5.1	0.74	1	10/20/10 18:20	10/22/10 21:26	2303-16-4	
Dibenz(a,h)anthracene	0.66U	ug/L	2.0	0.66	1	10/20/10 18:20	10/22/10 21:26	53-70-3	
Dibenzofuran	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 21:26	132-64-9	
1,2-Dichlorobenzene	0.70U	ug/L	5.1	0.70	1	10/20/10 18:20	10/22/10 21:26	95-50-1	
1,3-Dichlorobenzene	0.78U	ug/L	5.1	0.78	1	10/20/10 18:20	10/22/10 21:26	541-73-1	
1,4-Dichlorobenzene	0.79U	ug/L	5.1	0.79	1	10/20/10 18:20	10/22/10 21:26	106-46-7	
3,3'-Dichlorobenzidine	0.71U	ug/L	10.2	0.71	1	10/20/10 18:20	10/22/10 21:26	91-94-1	
2,4-Dichlorophenol	0.57U	ug/L	2.0	0.57	1	10/20/10 18:20	10/22/10 21:26	120-83-2	
2,6-Dichlorophenol	0.63U	ug/L	4.1	0.63	1	10/20/10 18:20	10/22/10 21:26	87-65-0	
Diethylphthalate	0.52U	ug/L	5.1	0.52	1	10/20/10 18:20	10/22/10 21:26	84-66-2	
P-Dimethylaminoazobenzene	0.68U	ug/L	5.1	0.68	1	10/20/10 18:20	10/22/10 21:26	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.1	2.0	1	10/20/10 18:20	10/22/10 21:26	57-97-6	
3,3'-Dimethylbenzidine	3.2U	ug/L	10.2	3.2	1	10/20/10 18:20	10/22/10 21:26	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.1	1.6	1	10/20/10 18:20	10/22/10 21:26	105-67-9	
a,a-Dimethylphenylethylamine	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 21:26	122-09-8	
Dimethylphthalate	0.65U	ug/L	5.1	0.65	1	10/20/10 18:20	10/22/10 21:26	131-11-3	
Di-n-butylphthalate	0.42U	ug/L	5.1	0.42	1	10/20/10 18:20	10/22/10 21:26	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	20.4	1.3	1	10/20/10 18:20	10/22/10 21:26	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	528-29-0	
1,3-Dinitrobenzene	0.70U	ug/L	8.2	0.70	1	10/20/10 18:20	10/22/10 21:26	99-65-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 117 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-16 **Lab ID: 3519325013** Collected: 10/15/10 11:55 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrophenol	1.6U	ug/L	20.4	1.6	1	10/20/10 18:20	10/22/10 21:26	51-28-5	
2,4-Dinitrotoluene	0.54U	ug/L	2.0	0.54	1	10/20/10 18:20	10/22/10 21:26	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	2.0	1.2	1	10/20/10 18:20	10/22/10 21:26	606-20-2	
Di-n-octylphthalate	0.92U	ug/L	5.1	0.92	1	10/20/10 18:20	10/22/10 21:26	117-84-0	
bis(2-Ethylhexyl)phthalate	0.82U	ug/L	5.1	0.82	1	10/20/10 18:20	10/22/10 21:26	117-81-7	
Ethyl methanesulfonate	0.92U	ug/L	5.1	0.92	1	10/20/10 18:20	10/22/10 21:26	62-50-0	
Fluoranthene	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 21:26	206-44-0	
Fluorene	0.57U	ug/L	5.1	0.57	1	10/20/10 18:20	10/22/10 21:26	86-73-7	
Hexachlorobenzene	0.82U	ug/L	1.0	0.82	1	10/20/10 18:20	10/22/10 21:26	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.1	1.3	1	10/20/10 18:20	10/22/10 21:26	77-47-4	
Hexachloroethane	0.73U	ug/L	5.1	0.73	1	10/20/10 18:20	10/22/10 21:26	67-72-1	
Hexachloropropene	1.4U	ug/L	5.1	1.4	1	10/20/10 18:20	10/22/10 21:26	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.75U	ug/L	2.0	0.75	1	10/20/10 18:20	10/22/10 21:26	193-39-5	
Isodrin	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 21:26	465-73-6	
Isophorone	0.75U	ug/L	5.1	0.75	1	10/20/10 18:20	10/22/10 21:26	78-59-1	
Isosafrole	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 21:26	120-58-1	
Kepone	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 21:26	143-50-0	
Methapyrilene	1.7U	ug/L	5.1	1.7	1	10/20/10 18:20	10/22/10 21:26	91-80-5	J(SS)
3-Methylcholanthrene	1.1U	ug/L	5.1	1.1	1	10/20/10 18:20	10/22/10 21:26	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 21:26	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 21:26	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 21:26	91-57-6	
2-Methylphenol(o-Cresol)	0.75U	ug/L	5.1	0.75	1	10/20/10 18:20	10/22/10 21:26	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.67U	ug/L	10.2	0.67	1	10/20/10 18:20	10/22/10 21:26		
2-Naphthylamine	2.3U	ug/L	5.1	2.3	1	10/20/10 18:20	10/22/10 21:26	91-59-8	
Naphthalene	0.80U	ug/L	5.1	0.80	1	10/20/10 18:20	10/22/10 21:26	91-20-3	
1-Naphthylamine	1.1U	ug/L	5.1	1.1	1	10/20/10 18:20	10/22/10 21:26	134-32-7	
1,4-Naphthoquinone	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	130-15-4	
2-Nitroaniline	0.61U	ug/L	5.1	0.61	1	10/20/10 18:20	10/22/10 21:26	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.1	1.0	1	10/20/10 18:20	10/22/10 21:26	99-09-2	
4-Nitroaniline	0.71U	ug/L	4.1	0.71	1	10/20/10 18:20	10/22/10 21:26	100-01-6	
Nitrobenzene	1.1U	ug/L	4.1	1.1	1	10/20/10 18:20	10/22/10 21:26	98-95-3	
2-Nitrophenol	0.83U	ug/L	5.1	0.83	1	10/20/10 18:20	10/22/10 21:26	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.4	1.1	1	10/20/10 18:20	10/22/10 21:26	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.1	1.3	1	10/20/10 18:20	10/22/10 21:26	99-55-8	
N-Nitrosodiethylamine	0.75U	ug/L	4.1	0.75	1	10/20/10 18:20	10/22/10 21:26	55-18-5	
N-Nitrosodimethylamine	0.99U	ug/L	2.0	0.99	1	10/20/10 18:20	10/22/10 21:26	62-75-9	
N-Nitroso-di-n-butylamine	0.56U	ug/L	4.1	0.56	1	10/20/10 18:20	10/22/10 21:26	924-16-3	
N-Nitroso-di-n-propylamine	0.96U	ug/L	4.1	0.96	1	10/20/10 18:20	10/22/10 21:26	621-64-7	
N-Nitrosodiphenylamine	0.51U	ug/L	5.1	0.51	1	10/20/10 18:20	10/22/10 21:26	86-30-6	
N-Nitrosomethylethylamine	0.76U	ug/L	5.1	0.76	1	10/20/10 18:20	10/22/10 21:26	10595-95-6	
N-Nitrosopiperidine	0.65U	ug/L	5.1	0.65	1	10/20/10 18:20	10/22/10 21:26	100-75-4	
N-Nitrosopyrrolidine	0.90U	ug/L	5.1	0.90	1	10/20/10 18:20	10/22/10 21:26	930-55-2	
O,O,O-Triethylphosphorothioate	0.71U	ug/L	5.1	0.71	1	10/20/10 18:20	10/22/10 21:26	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	56-38-2	
Pentachlorobenzene	0.80U	ug/L	5.1	0.80	1	10/20/10 18:20	10/22/10 21:26	608-93-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 118 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-16 **Lab ID: 3519325013** Collected: 10/15/10 11:55 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Pentachlorophenol	0.67U	ug/L	20.4	0.67	1	10/20/10 18:20	10/22/10 21:26	87-86-5	
Phenacetin	0.54U	ug/L	5.1	0.54	1	10/20/10 18:20	10/22/10 21:26	62-44-2	
Phenanthrene	0.53U	ug/L	5.1	0.53	1	10/20/10 18:20	10/22/10 21:26	85-01-8	
Phenol	0.55U	ug/L	5.1	0.55	1	10/20/10 18:20	10/22/10 21:26	108-95-2	
p-Phenylenediamine	10.2U	ug/L	20.4	10.2	1	10/20/10 18:20	10/22/10 21:26	106-50-3	
Pronamide	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	23950-58-5	
Pyrene	0.70U	ug/L	5.1	0.70	1	10/20/10 18:20	10/22/10 21:26	129-00-0	
Safrole	0.87U	ug/L	5.1	0.87	1	10/20/10 18:20	10/22/10 21:26	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.72U	ug/L	5.1	0.72	1	10/20/10 18:20	10/22/10 21:26	95-94-3	
2,3,4,6-Tetrachlorophenol	3.9U	ug/L	5.1	3.9	1	10/20/10 18:20	10/22/10 21:26	58-90-2	
Thionazin	0.62U	ug/L	5.1	0.62	1	10/20/10 18:20	10/22/10 21:26	297-97-2	
O-Toluidine	1.1U	ug/L	5.1	1.1	1	10/20/10 18:20	10/22/10 21:26	95-53-4	
1,2,4-Trichlorobenzene	0.85U	ug/L	5.1	0.85	1	10/20/10 18:20	10/22/10 21:26	120-82-1	
2,4,5-Trichlorophenol	0.53U	ug/L	4.1	0.53	1	10/20/10 18:20	10/22/10 21:26	95-95-4	
2,4,6-Trichlorophenol	0.71U	ug/L	2.0	0.71	1	10/20/10 18:20	10/22/10 21:26	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/20/10 18:20	10/22/10 21:26	99-35-4	
Nitrobenzene-d5 (S)	66 %		10-110		1	10/20/10 18:20	10/22/10 21:26	4165-60-0	
2-Fluorobiphenyl (S)	68 %		18-110		1	10/20/10 18:20	10/22/10 21:26	321-60-8	
Terphenyl-d14 (S)	87 %		10-123		1	10/20/10 18:20	10/22/10 21:26	1718-51-0	
Phenol-d6 (S)	27 %		10-110		1	10/20/10 18:20	10/22/10 21:26	13127-88-3	
2-Fluorophenol (S)	41 %		18-110		1	10/20/10 18:20	10/22/10 21:26	367-12-4	
2,4,6-Tribromophenol (S)	89 %		10-110		1	10/20/10 18:20	10/22/10 21:26	118-79-6	
8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 13:15	83-32-9	
Acenaphthylene	0.051U	ug/L	2.0	0.051	1	10/20/10 13:02	10/21/10 13:15	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 13:15	120-12-7	
Benzo(a)anthracene	0.061U	ug/L	0.20	0.061	1	10/20/10 13:02	10/21/10 13:15	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.20	0.051	1	10/20/10 13:02	10/21/10 13:15	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/20/10 13:02	10/21/10 13:15	205-99-2	
Benzo(g,h,i)perylene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 13:15	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.25	0.041	1	10/20/10 13:02	10/21/10 13:15	207-08-9	
Chrysene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 13:15	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.20	0.051	1	10/20/10 13:02	10/21/10 13:15	53-70-3	
Fluoranthene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 13:15	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 13:15	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/20/10 13:02	10/21/10 13:15	193-39-5	
1-Methylnaphthalene	0.092U	ug/L	1.5	0.092	1	10/20/10 13:02	10/21/10 13:15	90-12-0	
2-Methylnaphthalene	0.061U	ug/L	1.5	0.061	1	10/20/10 13:02	10/21/10 13:15	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/20/10 13:02	10/21/10 13:15	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 13:15	85-01-8	
Pyrene	0.061U	ug/L	1.0	0.061	1	10/20/10 13:02	10/21/10 13:15	129-00-0	
2-Fluorobiphenyl (S)	63 %		43.9-113		1	10/20/10 13:02	10/21/10 13:15	321-60-8	
Terphenyl-d14 (S)	55 %		24.8-144		1	10/20/10 13:02	10/21/10 13:15	1718-51-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-16 **Lab ID: 3519325013** Collected: 10/15/10 11:55 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.6 I	ug/L	10.0	5.0	1		10/29/10 09:18	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 09:18	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 09:18	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 09:18	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 09:18	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 09:18	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 09:18	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 09:18	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 09:18	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 09:18	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 120 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-16		Lab ID: 3519325013		Collected: 10/15/10 11:55		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 09:18	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 09:18	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 09:18	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 09:18	1330-20-7	
4-Bromofluorobenzene (S)	97 %		70-114		1		10/29/10 09:18	460-00-4	p2
Dibromofluoromethane (S)	100 %		88-117		1		10/29/10 09:18	1868-53-7	
1,2-Dichloroethane-d4 (S)	101 %		86-125		1		10/29/10 09:18	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 09:18	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1630	mg/L	20.0	20.0	1		10/21/10 16:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	2.1	mg/L	1.0	1.0	1		10/20/10 10:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	317	mg/L	25.0	12.5	5		11/01/10 21:52	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		11/01/10 21:52	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 16:00	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	20.4	mg/L	0.10	0.040	2		10/25/10 08:38	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-15		Lab ID: 3519325014		Collected: 10/15/10 14:45		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.39	Std. Units			1		10/15/10 14:45		
Field Temperature	26.7	deg C			1		10/15/10 14:45		
Field Specific Conductance	2591	umhos/cm			1		10/15/10 14:45		
Oxygen, Dissolved	1.006	mg/L			1		10/15/10 14:45	7782-44-7	
Turbidity	16.3	NTU			1		10/15/10 14:45		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/27/10 16:30	10/28/10 14:01	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/27/10 16:30	10/28/10 14:01	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00048U	ug/L	0.0095	0.00048	1	10/20/10 12:00	11/12/10 23:27	309-00-2	
alpha-BHC	0.00029U	ug/L	0.0095	0.00029	1	10/20/10 12:00	11/12/10 23:27	319-84-6	
beta-BHC	0.00048U	ug/L	0.0095	0.00048	1	10/20/10 12:00	11/12/10 23:27	319-85-7	
delta-BHC	0.00038U	ug/L	0.0095	0.00038	1	10/20/10 12:00	11/12/10 23:27	319-86-8	
gamma-BHC (Lindane)	0.00019U	ug/L	0.0095	0.00019	1	10/20/10 12:00	11/12/10 23:27	58-89-9	
Chlordane (Technical)	0.076U	ug/L	0.48	0.076	1	10/20/10 12:00	11/12/10 23:27	57-74-9	
Chlorobenzilate	0.020U	ug/L	0.095	0.020	1	10/20/10 12:00	11/12/10 23:27	510-15-6	
4,4'-DDD	0.0018U	ug/L	0.0095	0.0018	1	10/20/10 12:00	11/12/10 23:27	72-54-8	
4,4'-DDE	0.00086U	ug/L	0.0095	0.00086	1	10/20/10 12:00	11/12/10 23:27	72-55-9	
4,4'-DDT	0.0034U	ug/L	0.0095	0.0034	1	10/20/10 12:00	11/12/10 23:27	50-29-3	
Dieldrin	0.00048U	ug/L	0.0095	0.00048	1	10/20/10 12:00	11/12/10 23:27	60-57-1	
Endosulfan I	0.00067U	ug/L	0.0095	0.00067	1	10/20/10 12:00	11/12/10 23:27	959-98-8	
Endosulfan II	0.00067U	ug/L	0.0095	0.00067	1	10/20/10 12:00	11/12/10 23:27	33213-65-9	
Endosulfan sulfate	0.00057U	ug/L	0.0095	0.00057	1	10/20/10 12:00	11/12/10 23:27	1031-07-8	
Endrin	0.0016U	ug/L	0.0095	0.0016	1	10/20/10 12:00	11/12/10 23:27	72-20-8	
Endrin aldehyde	0.0068U	ug/L	0.0095	0.0068	1	10/20/10 12:00	11/12/10 23:27	7421-93-4	
Heptachlor	0.0014U	ug/L	0.0095	0.0014	1	10/20/10 12:00	11/12/10 23:27	76-44-8	
Heptachlor epoxide	0.00038U	ug/L	0.0095	0.00038	1	10/20/10 12:00	11/12/10 23:27	1024-57-3	
Methoxychlor	0.0067U	ug/L	0.0095	0.0067	1	10/20/10 12:00	11/12/10 23:27	72-43-5	
Pentachloronitrobenzene	0.014U	ug/L	0.095	0.014	1	10/20/10 12:00	11/12/10 23:27	82-68-8	
Toxaphene	0.27U	ug/L	0.48	0.27	1	10/20/10 12:00	11/12/10 23:27	8001-35-2	
Tetrachloro-m-xylene (S)	72 %		66.5-120.3		1	10/20/10 12:00	11/12/10 23:27	877-09-8	
Decachlorobiphenyl (S)	28 %		41.7-109.1		1	10/20/10 12:00	11/12/10 23:27	2051-24-3	1p, J(S5)
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.076U	ug/L	0.48	0.076	1	10/20/10 12:01	11/12/10 23:27	12674-11-2	
PCB-1221 (Aroclor 1221)	0.077U	ug/L	0.48	0.077	1	10/20/10 12:01	11/12/10 23:27	11104-28-2	
PCB-1232 (Aroclor 1232)	0.11U	ug/L	0.48	0.11	1	10/20/10 12:01	11/12/10 23:27	11141-16-5	
PCB-1242 (Aroclor 1242)	0.12U	ug/L	0.48	0.12	1	10/20/10 12:01	11/12/10 23:27	53469-21-9	
PCB-1248 (Aroclor 1248)	0.26U	ug/L	0.48	0.26	1	10/20/10 12:01	11/12/10 23:27	12672-29-6	
PCB-1254 (Aroclor 1254)	0.14U	ug/L	0.48	0.14	1	10/20/10 12:01	11/12/10 23:27	11097-69-1	
PCB-1260 (Aroclor 1260)	0.10U	ug/L	0.48	0.10	1	10/20/10 12:01	11/12/10 23:27	11096-82-5	
Tetrachloro-m-xylene (S)	72 %		48-111		1	10/20/10 12:01	11/12/10 23:27	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 122 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-15 Lab ID: 3519325014 Collected: 10/15/10 14:45 Received: 10/19/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	28 %		63-121		1	10/20/10 12:01	11/12/10 23:27	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.18U	ug/L	0.47	0.18	1	10/20/10 16:08	11/02/10 03:07	60-51-5	
Disulfoton	0.15U	ug/L	0.47	0.15	1	10/20/10 16:08	11/02/10 03:07	298-04-4	J(M1)
Famphur	0.14U	ug/L	0.47	0.14	1	10/20/10 16:08	11/02/10 03:07	52-85-7	
Methyl parathion	0.18U	ug/L	0.47	0.18	1	10/20/10 16:08	11/02/10 03:07	298-00-0	
Parathion (Ethyl parathion)	0.34U	ug/L	0.95	0.34	1	10/20/10 16:08	11/02/10 03:07	56-38-2	
Phorate	0.35U	ug/L	0.95	0.35	1	10/20/10 16:08	11/02/10 03:07	298-02-2	J(M1)
4-Chloro3nitrobenzotrifluoride	69 %		34.2-122		1	10/20/10 16:08	11/02/10 03:07		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.21U	ug/L	0.89	0.21	1	10/20/10 17:00	10/27/10 05:47	94-75-7	
Dinoseb	0.054U	ug/L	0.18	0.054	1	10/20/10 17:00	10/27/10 05:47	88-85-7	
Pentachlorophenol	0.016U	ug/L	0.027	0.016	1	10/20/10 17:00	10/27/10 05:47	87-86-5	
2,4,5-T	0.040U	ug/L	0.18	0.040	1	10/20/10 17:00	10/27/10 05:47	93-76-5	
2,4,5-TP (Silvex)	0.046U	ug/L	0.18	0.046	1	10/20/10 17:00	10/27/10 05:47	93-72-1	
2,4-DCPA (S)	74 %		65.5-125.7		1	10/20/10 17:00	10/27/10 05:47	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Aluminum	75.2 I	ug/L	100	50.0	1	10/21/10 06:45	10/22/10 00:18	7429-90-5	
Arsenic	16.8	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:18	7440-38-2	
Barium	119	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:18	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:18	7440-41-7	
Cadmium	0.52 I	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:18	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:18	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:18	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:18	7440-50-8	
Iron	59900	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:18	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:18	7439-92-1	
Manganese	64.1	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:18	7439-96-5	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:18	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:18	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:18	7440-22-4	
Sodium	193	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:18	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/21/10 06:45	10/22/10 00:18	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:18	7440-62-2	
Zinc	33.7	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:18	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:46	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:46	7440-28-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-15		Lab ID: 3519325014		Collected: 10/15/10 14:45		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:24	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.82U	ug/L	4.7	0.82	1	10/20/10 18:20	10/22/10 21:57	83-32-9	
Acenaphthylene	0.90U	ug/L	4.7	0.90	1	10/20/10 18:20	10/22/10 21:57	208-96-8	
Acetophenone	1.4U	ug/L	4.7	1.4	1	10/20/10 18:20	10/22/10 21:57	98-86-2	
2-Acetylaminofluorene	0.62U	ug/L	4.7	0.62	1	10/20/10 18:20	10/22/10 21:57	53-96-3	
4-Aminobiphenyl	2.7U	ug/L	4.7	2.7	1	10/20/10 18:20	10/22/10 21:57	92-67-1	
Anthracene	0.57U	ug/L	4.7	0.57	1	10/20/10 18:20	10/22/10 21:57	120-12-7	
Benzo(a)anthracene	0.60U	ug/L	4.7	0.60	1	10/20/10 18:20	10/22/10 21:57	56-55-3	
Benzo(a)pyrene	0.55U	ug/L	0.95	0.55	1	10/20/10 18:20	10/22/10 21:57	50-32-8	
Benzo(b)fluoranthene	0.59U	ug/L	1.9	0.59	1	10/20/10 18:20	10/22/10 21:57	205-99-2	
Benzo(g,h,i)perylene	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	191-24-2	
Benzo(k)fluoranthene	0.48U	ug/L	3.8	0.48	1	10/20/10 18:20	10/22/10 21:57	207-08-9	
Benzyl alcohol	0.97U	ug/L	4.7	0.97	1	10/20/10 18:20	10/22/10 21:57	100-51-6	
4-Bromophenylphenyl ether	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	101-55-3	
Butylbenzylphthalate	0.68U	ug/L	4.7	0.68	1	10/20/10 18:20	10/22/10 21:57	85-68-7	
4-Chloro-3-methylphenol	0.59U	ug/L	19.0	0.59	1	10/20/10 18:20	10/22/10 21:57	59-50-7	
4-Chloroaniline	1.1U	ug/L	4.7	1.1	1	10/20/10 18:20	10/22/10 21:57	106-47-8	
bis(2-Chloroethoxy)methane	2.8U	ug/L	4.7	2.8	1	10/20/10 18:20	10/22/10 21:57	111-91-1	
bis(2-Chloroethyl) ether	0.71U	ug/L	3.8	0.71	1	10/20/10 18:20	10/22/10 21:57	111-44-4	
bis(2-Chloroisopropyl) ether	0.69U	ug/L	4.7	0.69	1	10/20/10 18:20	10/22/10 21:57	108-60-1	
2-Chloronaphthalene	0.76U	ug/L	4.7	0.76	1	10/20/10 18:20	10/22/10 21:57	91-58-7	
2-Chlorophenol	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	95-57-8	
4-Chlorophenylphenyl ether	0.60U	ug/L	4.7	0.60	1	10/20/10 18:20	10/22/10 21:57	7005-72-3	
Chrysene	0.35U	ug/L	4.7	0.35	1	10/20/10 18:20	10/22/10 21:57	218-01-9	
Diallate	0.69U	ug/L	4.7	0.69	1	10/20/10 18:20	10/22/10 21:57	2303-16-4	
Dibenz(a,h)anthracene	0.62U	ug/L	1.9	0.62	1	10/20/10 18:20	10/22/10 21:57	53-70-3	
Dibenzofuran	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	132-64-9	
1,2-Dichlorobenzene	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	95-50-1	
1,3-Dichlorobenzene	0.72U	ug/L	4.7	0.72	1	10/20/10 18:20	10/22/10 21:57	541-73-1	
1,4-Dichlorobenzene	0.73U	ug/L	4.7	0.73	1	10/20/10 18:20	10/22/10 21:57	106-46-7	
3,3'-Dichlorobenzidine	0.65U	ug/L	9.5	0.65	1	10/20/10 18:20	10/22/10 21:57	91-94-1	
2,4-Dichlorophenol	0.53U	ug/L	1.9	0.53	1	10/20/10 18:20	10/22/10 21:57	120-83-2	
2,6-Dichlorophenol	0.59U	ug/L	3.8	0.59	1	10/20/10 18:20	10/22/10 21:57	87-65-0	
Diethylphthalate	0.48U	ug/L	4.7	0.48	1	10/20/10 18:20	10/22/10 21:57	84-66-2	
P-Dimethylaminoazobenzene	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	60-11-7	J(SS)
7,12-Dimethylbenz(a)anthracene	1.8U	ug/L	4.7	1.8	1	10/20/10 18:20	10/22/10 21:57	57-97-6	
3,3'-Dimethylbenzidine	3.0U	ug/L	9.5	3.0	1	10/20/10 18:20	10/22/10 21:57	119-93-7	
2,4-Dimethylphenol	1.5U	ug/L	4.7	1.5	1	10/20/10 18:20	10/22/10 21:57	105-67-9	
a,a-Dimethylphenylethylamine	9.5U	ug/L	19.0	9.5	1	10/20/10 18:20	10/22/10 21:57	122-09-8	
Dimethylphthalate	0.61U	ug/L	4.7	0.61	1	10/20/10 18:20	10/22/10 21:57	131-11-3	
Di-n-butylphthalate	0.39U	ug/L	4.7	0.39	1	10/20/10 18:20	10/22/10 21:57	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	19.0	1.3	1	10/20/10 18:20	10/22/10 21:57	534-52-1	
1,2-Dinitrobenzene	1.1U	ug/L	4.7	1.1	1	10/20/10 18:20	10/22/10 21:57	528-29-0	
1,3-Dinitrobenzene	0.64U	ug/L	7.6	0.64	1	10/20/10 18:20	10/22/10 21:57	99-65-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 124 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-15 **Lab ID: 3519325014** Collected: 10/15/10 14:45 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrophenol	1.5U	ug/L	19.0	1.5	1	10/20/10 18:20	10/22/10 21:57	51-28-5	
2,4-Dinitrotoluene	0.50U	ug/L	1.9	0.50	1	10/20/10 18:20	10/22/10 21:57	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	1.9	1.2	1	10/20/10 18:20	10/22/10 21:57	606-20-2	
Di-n-octylphthalate	0.85U	ug/L	4.7	0.85	1	10/20/10 18:20	10/22/10 21:57	117-84-0	
bis(2-Ethylhexyl)phthalate	0.76U	ug/L	4.7	0.76	1	10/20/10 18:20	10/22/10 21:57	117-81-7	
Ethyl methanesulfonate	0.85U	ug/L	4.7	0.85	1	10/20/10 18:20	10/22/10 21:57	62-50-0	
Fluoranthene	0.51U	ug/L	4.7	0.51	1	10/20/10 18:20	10/22/10 21:57	206-44-0	
Fluorene	0.53U	ug/L	4.7	0.53	1	10/20/10 18:20	10/22/10 21:57	86-73-7	
Hexachlorobenzene	0.76U	ug/L	0.95	0.76	1	10/20/10 18:20	10/22/10 21:57	118-74-1	
Hexachlorocyclopentadiene	1.2U	ug/L	4.7	1.2	1	10/20/10 18:20	10/22/10 21:57	77-47-4	
Hexachloroethane	0.67U	ug/L	4.7	0.67	1	10/20/10 18:20	10/22/10 21:57	67-72-1	
Hexachloropropene	1.3U	ug/L	4.7	1.3	1	10/20/10 18:20	10/22/10 21:57	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.69U	ug/L	1.9	0.69	1	10/20/10 18:20	10/22/10 21:57	193-39-5	
Isodrin	0.51U	ug/L	4.7	0.51	1	10/20/10 18:20	10/22/10 21:57	465-73-6	
Isophorone	0.69U	ug/L	4.7	0.69	1	10/20/10 18:20	10/22/10 21:57	78-59-1	
Isosafrole	0.57U	ug/L	4.7	0.57	1	10/20/10 18:20	10/22/10 21:57	120-58-1	
Kepone	9.5U	ug/L	19.0	9.5	1	10/20/10 18:20	10/22/10 21:57	143-50-0	
Methapyrilene	1.6U	ug/L	4.7	1.6	1	10/20/10 18:20	10/22/10 21:57	91-80-5	J(SS)
3-Methylcholanthrene	0.99U	ug/L	4.7	0.99	1	10/20/10 18:20	10/22/10 21:57	56-49-5	
Methyl methanesulfonate	0.95U	ug/L	4.7	0.95	1	10/20/10 18:20	10/22/10 21:57	66-27-3	
1-Methylnaphthalene	0.95U	ug/L	4.7	0.95	1	10/20/10 18:20	10/22/10 21:57	90-12-0	
2-Methylnaphthalene	0.94U	ug/L	4.7	0.94	1	10/20/10 18:20	10/22/10 21:57	91-57-6	
2-Methylphenol(o-Cresol)	0.69U	ug/L	4.7	0.69	1	10/20/10 18:20	10/22/10 21:57	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.63U	ug/L	9.5	0.63	1	10/20/10 18:20	10/22/10 21:57		
2-Naphthylamine	2.2U	ug/L	4.7	2.2	1	10/20/10 18:20	10/22/10 21:57	91-59-8	
Naphthalene	0.74U	ug/L	4.7	0.74	1	10/20/10 18:20	10/22/10 21:57	91-20-3	
1-Naphthylamine	0.98U	ug/L	4.7	0.98	1	10/20/10 18:20	10/22/10 21:57	134-32-7	
1,4-Naphthoquinone	1.1U	ug/L	4.7	1.1	1	10/20/10 18:20	10/22/10 21:57	130-15-4	
2-Nitroaniline	0.57U	ug/L	4.7	0.57	1	10/20/10 18:20	10/22/10 21:57	88-74-4	
3-Nitroaniline	0.94U	ug/L	4.7	0.94	1	10/20/10 18:20	10/22/10 21:57	99-09-2	
4-Nitroaniline	0.65U	ug/L	3.8	0.65	1	10/20/10 18:20	10/22/10 21:57	100-01-6	
Nitrobenzene	1.0U	ug/L	3.8	1.0	1	10/20/10 18:20	10/22/10 21:57	98-95-3	
2-Nitrophenol	0.77U	ug/L	4.7	0.77	1	10/20/10 18:20	10/22/10 21:57	88-75-5	
4-Nitrophenol	1.0U	ug/L	19.0	1.0	1	10/20/10 18:20	10/22/10 21:57	100-02-7	
5-Nitro-o-toluidine	1.2U	ug/L	4.7	1.2	1	10/20/10 18:20	10/22/10 21:57	99-55-8	
N-Nitrosodiethylamine	0.69U	ug/L	3.8	0.69	1	10/20/10 18:20	10/22/10 21:57	55-18-5	
N-Nitrosodimethylamine	0.92U	ug/L	1.9	0.92	1	10/20/10 18:20	10/22/10 21:57	62-75-9	
N-Nitroso-di-n-butylamine	0.52U	ug/L	3.8	0.52	1	10/20/10 18:20	10/22/10 21:57	924-16-3	
N-Nitroso-di-n-propylamine	0.89U	ug/L	3.8	0.89	1	10/20/10 18:20	10/22/10 21:57	621-64-7	
N-Nitrosodiphenylamine	0.47U	ug/L	4.7	0.47	1	10/20/10 18:20	10/22/10 21:57	86-30-6	
N-Nitrosomethylethylamine	0.70U	ug/L	4.7	0.70	1	10/20/10 18:20	10/22/10 21:57	10595-95-6	
N-Nitrosopiperidine	0.61U	ug/L	4.7	0.61	1	10/20/10 18:20	10/22/10 21:57	100-75-4	
N-Nitrosopyrrolidine	0.83U	ug/L	4.7	0.83	1	10/20/10 18:20	10/22/10 21:57	930-55-2	
O,O,O-Triethylphosphorothioate	0.65U	ug/L	4.7	0.65	1	10/20/10 18:20	10/22/10 21:57	126-68-1	
Parathion (Ethyl parathion)	1.1U	ug/L	4.7	1.1	1	10/20/10 18:20	10/22/10 21:57	56-38-2	
Pentachlorobenzene	0.74U	ug/L	4.7	0.74	1	10/20/10 18:20	10/22/10 21:57	608-93-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 125 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-15 **Lab ID: 3519325014** Collected: 10/15/10 14:45 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Pentachlorophenol	0.63U	ug/L	19.0	0.63	1	10/20/10 18:20	10/22/10 21:57	87-86-5	
Phenacetin	0.50U	ug/L	4.7	0.50	1	10/20/10 18:20	10/22/10 21:57	62-44-2	
Phenanthrene	0.49U	ug/L	4.7	0.49	1	10/20/10 18:20	10/22/10 21:57	85-01-8	
Phenol	0.51U	ug/L	4.7	0.51	1	10/20/10 18:20	10/22/10 21:57	108-95-2	
p-Phenylenediamine	9.5U	ug/L	19.0	9.5	1	10/20/10 18:20	10/22/10 21:57	106-50-3	
Pronamide	1.1U	ug/L	4.7	1.1	1	10/20/10 18:20	10/22/10 21:57	23950-58-5	
Pyrene	0.64U	ug/L	4.7	0.64	1	10/20/10 18:20	10/22/10 21:57	129-00-0	
Safrole	0.81U	ug/L	4.7	0.81	1	10/20/10 18:20	10/22/10 21:57	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.66U	ug/L	4.7	0.66	1	10/20/10 18:20	10/22/10 21:57	95-94-3	
2,3,4,6-Tetrachlorophenol	3.7U	ug/L	4.7	3.7	1	10/20/10 18:20	10/22/10 21:57	58-90-2	
Thionazin	0.58U	ug/L	4.7	0.58	1	10/20/10 18:20	10/22/10 21:57	297-97-2	
O-Toluidine	1.0U	ug/L	4.7	1.0	1	10/20/10 18:20	10/22/10 21:57	95-53-4	
1,2,4-Trichlorobenzene	0.79U	ug/L	4.7	0.79	1	10/20/10 18:20	10/22/10 21:57	120-82-1	
2,4,5-Trichlorophenol	0.49U	ug/L	3.8	0.49	1	10/20/10 18:20	10/22/10 21:57	95-95-4	
2,4,6-Trichlorophenol	0.65U	ug/L	1.9	0.65	1	10/20/10 18:20	10/22/10 21:57	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	4.7	1.2	1	10/20/10 18:20	10/22/10 21:57	99-35-4	
Nitrobenzene-d5 (S)	69 %		10-110		1	10/20/10 18:20	10/22/10 21:57	4165-60-0	
2-Fluorobiphenyl (S)	80 %		18-110		1	10/20/10 18:20	10/22/10 21:57	321-60-8	
Terphenyl-d14 (S)	87 %		10-123		1	10/20/10 18:20	10/22/10 21:57	1718-51-0	
Phenol-d6 (S)	26 %		10-110		1	10/20/10 18:20	10/22/10 21:57	13127-88-3	
2-Fluorophenol (S)	40 %		18-110		1	10/20/10 18:20	10/22/10 21:57	367-12-4	
2,4,6-Tribromophenol (S)	88 %		10-110		1	10/20/10 18:20	10/22/10 21:57	118-79-6	

8270 MSSV PAH by SCAN

Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510

Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 13:37	83-32-9	
Acenaphthylene	0.051U	ug/L	2.1	0.051	1	10/20/10 13:02	10/21/10 13:37	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 13:37	120-12-7	
Benzo(a)anthracene	0.062U	ug/L	0.21	0.062	1	10/20/10 13:02	10/21/10 13:37	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.21	0.051	1	10/20/10 13:02	10/21/10 13:37	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/20/10 13:02	10/21/10 13:37	205-99-2	
Benzo(g,h,i)perylene	0.062U	ug/L	1.0	0.062	1	10/20/10 13:02	10/21/10 13:37	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/20/10 13:02	10/21/10 13:37	207-08-9	
Chrysene	0.062U	ug/L	1.0	0.062	1	10/20/10 13:02	10/21/10 13:37	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.21	0.051	1	10/20/10 13:02	10/21/10 13:37	53-70-3	
Fluoranthene	0.062U	ug/L	1.0	0.062	1	10/20/10 13:02	10/21/10 13:37	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/20/10 13:02	10/21/10 13:37	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/20/10 13:02	10/21/10 13:37	193-39-5	
1-Methylnaphthalene	0.093U	ug/L	1.5	0.093	1	10/20/10 13:02	10/21/10 13:37	90-12-0	
2-Methylnaphthalene	0.062U	ug/L	1.5	0.062	1	10/20/10 13:02	10/21/10 13:37	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/20/10 13:02	10/21/10 13:37	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/20/10 13:02	10/21/10 13:37	85-01-8	
Pyrene	0.062U	ug/L	1.0	0.062	1	10/20/10 13:02	10/21/10 13:37	129-00-0	
2-Fluorobiphenyl (S)	80 %		43.9-113		1	10/20/10 13:02	10/21/10 13:37	321-60-8	
Terphenyl-d14 (S)	68 %		24.8-144		1	10/20/10 13:02	10/21/10 13:37	1718-51-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-15 **Lab ID: 3519325014** Collected: 10/15/10 14:45 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	7.21	ug/L	10.0	5.0	1		10/28/10 19:00	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/28/10 19:00	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	74-97-5	
Bromodichloromethane	2.0	ug/L	0.60	0.27	1		10/28/10 19:00	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-00-3	
Chloroform	17.7	ug/L	1.0	0.50	1		10/28/10 19:00	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 19:00	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	126-99-8	
Dibromochloromethane	1.6	ug/L	0.50	0.26	1		10/28/10 19:00	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 19:00	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 19:00	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/28/10 19:00	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 19:00	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 19:00	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 127 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: CW-15		Lab ID: 3519325014		Collected: 10/15/10 14:45		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 19:00	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 19:00	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 19:00	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 19:00	1330-20-7	
4-Bromofluorobenzene (S)	96 %		70-114		1		10/28/10 19:00	460-00-4	p2
Dibromofluoromethane (S)	104 %		88-117		1		10/28/10 19:00	1868-53-7	
1,2-Dichloroethane-d4 (S)	108 %		86-125		1		10/28/10 19:00	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/28/10 19:00	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1700	mg/L	20.0	20.0	1		10/21/10 16:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.4	mg/L	1.0	1.0	1		10/20/10 10:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	283	mg/L	25.0	12.5	5		11/01/10 22:29	16887-00-6	
Sulfate	47.8	mg/L	25.0	12.5	5		11/01/10 22:29	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0054 I	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 16:01	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	25.1	mg/L	0.10	0.040	2		10/25/10 08:40	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-17		Lab ID: 3519325015		Collected: 10/15/10 13:36		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.20	Std. Units			1		10/15/10 13:36		
Field Temperature	26.0	deg C			1		10/15/10 13:36		
Field Specific Conductance	1628	umhos/cm			1		10/15/10 13:36		
Oxygen, Dissolved	0.77	mg/L			1		10/15/10 13:36	7782-44-7	
Turbidity	12.1	NTU			1		10/15/10 13:36		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.021	0.0050	1	10/27/10 16:30	10/28/10 14:54	96-12-8	
1,2-Dibromoethane (EDB)	0.0064U	ug/L	0.010	0.0064	1	10/27/10 16:30	10/28/10 14:54	106-93-4	
6010 MET ICP									
		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Arsenic	65.5	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:22	7440-38-2	
Barium	110	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:22	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:22	7440-41-7	
Cadmium	0.63 I	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:22	7440-43-9	
Calcium	213	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:22	7440-70-2	
Chromium	3.2 I	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:22	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:22	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:22	7440-50-8	
Iron	116000	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:22	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:22	7439-92-1	
Magnesium	18.2	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:22	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:22	7440-02-0	
Potassium	6.5	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:22	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:22	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:22	7440-22-4	
Sodium	64.6	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:22	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:22	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:22	7440-66-6	
6020 MET ICPMS									
		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:51	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:51	7440-28-0	
7470 Mercury									
		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:33	7439-97-6	
8260 MSV									
		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 07:21	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	74-83-9	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-17 **Lab ID: 3519325015** Collected: 10/15/10 13:36 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 07:21	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 07:21	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 07:21	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 07:21	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 07:21	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 07:21	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 07:21	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	127-18-4	
Toluene	0.52 U	ug/L	1.0	0.50	1		10/29/10 07:21	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 07:21	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 07:21	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 07:21	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 07:21	460-00-4	p2
Dibromofluoromethane (S)	103 %		88-117		1		10/29/10 07:21	1868-53-7	
1,2-Dichloroethane-d4 (S)	102 %		86-125		1		10/29/10 07:21	17060-07-0	
Toluene-d8 (S)	102 %		87-113		1		10/29/10 07:21	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity,Bicarbonate (CaCO ₃)	785	mg/L	5.0	5.0	1	10/20/10 10:29
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/20/10 10:29

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 130 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-17		Lab ID: 3519325015		Collected: 10/15/10 13:36		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO ₃	785	mg/L	5.0	5.0	1		10/20/10 10:29		J(M1)
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	910	mg/L	10.0	10.0	1		10/21/10 16:00		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	68.5	mg/L	25.0	12.5	5		10/29/10 14:58	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		10/29/10 14:58	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	25.2	mg/L	0.10	0.040	2		10/25/10 08:44	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-9		Lab ID: 3519325016		Collected: 10/14/10 14:43		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.45	Std. Units			1		10/14/10 14:43		
Field Temperature	29.07	deg C			1		10/14/10 14:43		
Field Specific Conductance	2099	umhos/cm			1		10/14/10 14:43		
Oxygen, Dissolved	0.11	mg/L			1		10/14/10 14:43	7782-44-7	
Turbidity	1.00	NTU			1		10/14/10 14:43		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/21/10 16:20	10/24/10 11:16	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/21/10 16:20	10/24/10 11:16	106-93-4	
6010 MET ICP									
		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Arsenic	43.2	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:27	7440-38-2	
Barium	117	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:27	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:27	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:27	7440-43-9	
Calcium	342	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:27	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:27	7440-47-3	
Cobalt	20.4	ug/L	20.0	10.0	2	10/21/10 06:45	10/26/10 16:11	7440-48-4	D3
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:27	7440-50-8	
Iron	39200	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:27	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:27	7439-92-1	
Magnesium	57.4	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:27	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:27	7440-02-0	
Potassium	6.0	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:27	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:27	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:27	7440-22-4	
Sodium	43.9	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:27	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:27	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:27	7440-66-6	
6020 MET ICPMS									
		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:56	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 04:56	7440-28-0	
7470 Mercury									
		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:36	7439-97-6	
8260 MSV									
		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/28/10 04:22	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	74-83-9	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-9 **Lab ID: 3519325016** Collected: 10/14/10 14:43 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 04:22	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/28/10 04:22	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 04:22	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 04:22	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 04:22	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 04:22	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 04:22	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	127-18-4	
Toluene	0.98 U	ug/L	1.0	0.50	1		10/28/10 04:22	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 04:22	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 04:22	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 04:22	1330-20-7	
4-Bromofluorobenzene (S)	95 %		70-114		1		10/28/10 04:22	460-00-4	
Dibromofluoromethane (S)	101 %		88-117		1		10/28/10 04:22	1868-53-7	
1,2-Dichloroethane-d4 (S)	111 %		86-125		1		10/28/10 04:22	17060-07-0	
Toluene-d8 (S)	99 %		87-113		1		10/28/10 04:22	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	1030	mg/L	5.0	5.0	1	10/20/10 11:18
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/20/10 11:18

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 133 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-9		Lab ID: 3519325016		Collected: 10/14/10 14:43		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO ₃	1030	mg/L	5.0	5.0	1		10/20/10 11:18		
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1160	mg/L	10.0	10.0	1		10/21/10 16:00		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	38.7	mg/L	25.0	12.5	5		10/29/10 15:10	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		10/29/10 15:10	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	17.3	mg/L	0.10	0.040	2		10/25/10 08:45	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-8A		Lab ID: 3519325017		Collected: 10/14/10 12:41		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Analytical Method:									
Field pH	5.98	Std. Units			1		10/14/10 12:41		
Field Temperature	27.41	deg C			1		10/14/10 12:41		
Field Specific Conductance	1193	umhos/cm			1		10/14/10 12:41		
Oxygen, Dissolved	0.17	mg/L			1		10/14/10 12:41	7782-44-7	
Turbidity	3.91	NTU			1		10/14/10 12:41		
6010 MET ICP									
Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	53.3	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:31	7440-38-2	
Iron	45100	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:31	7439-89-6	
2540C Total Dissolved Solids									
Analytical Method: SM 2540C									
Total Dissolved Solids	670	mg/L	10.0	10.0	1		10/21/10 16:00		
350.1 Ammonia									
Analytical Method: EPA 350.1									
Nitrogen, Ammonia	7.9	mg/L	0.050	0.020	1		10/22/10 13:42	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip Blank (10/14/10) **Lab ID: 3519325018** Collected: 10/14/10 09:25 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8011 GCS EDB and DBCP Analytical Method: EPA 8011 Preparation Method: EPA 8011									
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/21/10 16:20	10/24/10 11:45	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/21/10 16:20	10/24/10 11:45	106-93-4	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:35	7440-38-2	
Barium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:35	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:35	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:35	7440-43-9	
Calcium	0.25U	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:35	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:35	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:35	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:35	7440-50-8	
Iron	20.0U	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:35	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:35	7439-92-1	
Magnesium	0.25U	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:35	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:35	7440-02-0	
Potassium	0.50U	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:35	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:35	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:35	7440-22-4	
Sodium	0.50U	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:35	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:35	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:35	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:00	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:00	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:44	7439-97-6	
8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/28/10 03:11	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 03:11	74-87-3	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip Blank (10/14/10) **Lab ID: 3519325018** Collected: 10/14/10 09:25 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/28/10 03:11	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:11	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:11	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 03:11	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:11	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 03:11	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	127-18-4	
Toluene	1.1	ug/L	1.0	0.50	1		10/28/10 03:11	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 03:11	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 03:11	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 03:11	1330-20-7	
4-Bromofluorobenzene (S)	95 %		70-114		1		10/28/10 03:11	460-00-4	
Dibromofluoromethane (S)	103 %		88-117		1		10/28/10 03:11	1868-53-7	
1,2-Dichloroethane-d4 (S)	113 %		86-125		1		10/28/10 03:11	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/28/10 03:11	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1		10/20/10 11:22		
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1		10/20/10 11:22		
Alkalinity, Total as CaCO ₃	5.0U	mg/L	5.0	5.0	1		10/20/10 11:22		

2540C Total Dissolved Solids

Analytical Method: SM 2540C

Total Dissolved Solids	5.0U	mg/L	5.0	5.0	1		10/21/10 16:00		
------------------------	------	------	-----	-----	---	--	----------------	--	--

300.0 IC Anions 28 Days

Analytical Method: EPA 300.0

Chloride	2.5U	mg/L	5.0	2.5	1		10/29/10 15:22	16887-00-6	
----------	------	------	-----	-----	---	--	----------------	------------	--

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 137 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Equip Blank (10/14/10) Lab ID: 3519325018 Collected: 10/14/10 09:25 Received: 10/19/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
300.0 IC Anions 28 Days									
Analytical Method: EPA 300.0									
Sulfate	2.5U	mg/L	5.0	2.5	1		10/29/10 15:22	14808-79-8	
350.1 Ammonia									
Analytical Method: EPA 350.1									
Nitrogen, Ammonia	0.020U	mg/L	0.050	0.020	1		10/25/10 08:47	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-9 DUP Lab ID: 3519325019 Collected: 10/14/10 14:43 Received: 10/19/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data Analytical Method:									
Field pH	6.45	Std. Units			1		10/14/10 08:00		
Field Temperature	29.07	deg C			1		10/14/10 08:00		
Field Specific Conductance	2099	umhos/cm			1		10/14/10 08:00		
Oxygen, Dissolved	0.11	mg/L			1		10/14/10 08:00	7782-44-7	
Turbidity	1.00	NTU			1		10/14/10 08:00		
8011 GCS EDB and DBCP Analytical Method: EPA 8011 Preparation Method: EPA 8011									
1,2-Dibromo-3-chloropropane	0.0051U	ug/L	0.021	0.0051	1	10/21/10 16:20	10/24/10 11:59	96-12-8	
1,2-Dibromoethane (EDB)	0.0064U	ug/L	0.010	0.0064	1	10/21/10 16:20	10/24/10 11:59	106-93-4	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	40.0	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:39	7440-38-2	
Barium	66.8	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:39	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:39	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:39	7440-43-9	
Calcium	253	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:39	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:39	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:39	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:39	7440-50-8	
Iron	40900	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:39	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:39	7439-92-1	
Magnesium	79.3	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:39	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:39	7440-02-0	
Potassium	4.9	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:39	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:39	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:39	7440-22-4	
Sodium	64.3	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:39	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:39	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:39	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:05	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:05	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:47	7439-97-6	
8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/28/10 03:35	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	74-83-9	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 139 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-9 DUP **Lab ID: 3519325019** Collected: 10/14/10 14:43 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 03:35	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/28/10 03:35	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:35	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:35	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 03:35	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:35	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 03:35	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	127-18-4	
Toluene	0.56 U	ug/L	1.0	0.50	1		10/28/10 03:35	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 03:35	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 03:35	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 03:35	1330-20-7	
4-Bromofluorobenzene (S)	93 %		70-114		1		10/28/10 03:35	460-00-4	
Dibromofluoromethane (S)	102 %		88-117		1		10/28/10 03:35	1868-53-7	
1,2-Dichloroethane-d4 (S)	109 %		86-125		1		10/28/10 03:35	17060-07-0	
Toluene-d8 (S)	99 %		87-113		1		10/28/10 03:35	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	996	mg/L	5.0	5.0	1	10/20/10 11:41
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/20/10 11:41

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 140 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-9 DUP		Lab ID: 3519325019	Collected: 10/14/10 14:43			Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO3	996	mg/L	5.0	5.0	1		10/20/10 11:41		
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1080	mg/L	10.0	10.0	1		10/21/10 16:00		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	30.2	mg/L	25.0	12.5	5		10/29/10 15:34	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		10/29/10 15:34	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	23.0	mg/L	0.10	0.040	2		10/25/10 13:12	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-8A		Lab ID: 3519325020		Collected: 10/14/10 10:13		Received: 10/19/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.30	Std. Units			1		10/14/10 10:13		
Field Temperature	26.86	deg C			1		10/14/10 10:13		
Field Specific Conductance	2005	umhos/cm			1		10/14/10 10:13		
Oxygen, Dissolved	0.22	mg/L			1		10/14/10 10:13	7782-44-7	
Turbidity	2.35	NTU			1		10/14/10 10:13		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/21/10 16:20	10/24/10 12:14	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/21/10 16:20	10/24/10 12:14	106-93-4	
6010 MET ICP									
		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Arsenic	38.6	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:42	7440-38-2	
Barium	65.8	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:42	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:42	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:42	7440-43-9	
Calcium	253	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:42	7440-70-2	
Chromium	2.7 I	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:42	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:42	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:42	7440-50-8	
Iron	37700	ug/L	40.0	20.0	1	10/21/10 06:45	10/22/10 00:42	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:42	7439-92-1	
Magnesium	77.7	mg/L	0.50	0.25	1	10/21/10 06:45	10/22/10 00:42	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:42	7440-02-0	
Potassium	5.1	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:42	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/21/10 06:45	10/22/10 00:42	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/21/10 06:45	10/22/10 00:42	7440-22-4	
Sodium	62.9	mg/L	1.0	0.50	1	10/21/10 06:45	10/22/10 00:42	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/21/10 06:45	10/22/10 00:42	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/21/10 06:45	10/22/10 00:42	7440-66-6	
6020 MET ICPMS									
		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Antimony	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:10	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/21/10 06:45	10/28/10 05:10	7440-28-0	
7470 Mercury									
		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	11/03/10 08:30	11/04/10 09:50	7439-97-6	
8260 MSV									
		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/28/10 03:58	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	74-83-9	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 142 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-8A **Lab ID: 3519325020** Collected: 10/14/10 10:13 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/28/10 03:58	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/28/10 03:58	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:58	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/28/10 03:58	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/28/10 03:58	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/28/10 03:58	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/28/10 03:58	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	127-18-4	
Toluene	1.3	ug/L	1.0	0.50	1		10/28/10 03:58	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/28/10 03:58	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/28/10 03:58	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/28/10 03:58	1330-20-7	
4-Bromofluorobenzene (S)	97 %		70-114		1		10/28/10 03:58	460-00-4	
Dibromofluoromethane (S)	102 %		88-117		1		10/28/10 03:58	1868-53-7	
1,2-Dichloroethane-d4 (S)	107 %		86-125		1		10/28/10 03:58	17060-07-0	
Toluene-d8 (S)	100 %		87-113		1		10/28/10 03:58	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	1020	mg/L	5.0	5.0	1	10/20/10 11:59
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/20/10 11:59

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 143 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-8A		Lab ID: 3519325020	Collected: 10/14/10 10:13	Received: 10/19/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO ₃	1020	mg/L	5.0	5.0	1		10/20/10 11:59		
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1090	mg/L	10.0	10.0	1		10/21/10 16:00		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	29.2	mg/L	25.0	12.5	5		10/29/10 15:46	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		10/29/10 15:46	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	25.5	mg/L	0.25	0.10	5		10/25/10 13:13	7664-41-7	J(M1)

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank appdx 2 (10/14/10) **Lab ID:** 3519325021 **Collected:** 10/14/10 08:00 **Received:** 10/19/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/27/10 21:09	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/27/10 21:09	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/27/10 21:09	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/27/10 21:09	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:09	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:09	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/27/10 21:09	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/27/10 21:09	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:09	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	100-42-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank appdx 2
(10/14/10) **Lab ID:** 3519325021 Collected: 10/14/10 08:00 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/27/10 21:09	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	127-18-4	
Toluene	19.4	ug/L	1.0	0.50	1		10/27/10 21:09	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/27/10 21:09	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/27/10 21:09	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/27/10 21:09	1330-20-7	
4-Bromofluorobenzene (S)	98	%	70-114		1		10/27/10 21:09	460-00-4	
Dibromofluoromethane (S)	101	%	88-117		1		10/27/10 21:09	1868-53-7	
1,2-Dichloroethane-d4 (S)	101	%	86-125		1		10/27/10 21:09	17060-07-0	
Toluene-d8 (S)	100	%	87-113		1		10/27/10 21:09	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank appdx 1 (10/14/10) **Lab ID:** 3519325022 **Collected:** 10/14/10 08:00 **Received:** 10/19/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Acetone	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/27/10 21:32	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/27/10 21:32	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/27/10 21:32	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:32	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/27/10 21:32	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/27/10 21:32	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/27/10 21:32	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/27/10 21:32	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	127-18-4	
Toluene	3.2	ug/L	1.0	0.50	1		10/27/10 21:32	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/27/10 21:32	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/27/10 21:32	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/27/10 21:32	1330-20-7	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 147 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip blank appdx 1 (10/14/10) **Lab ID:** 3519325022 Collected: 10/14/10 08:00 Received: 10/19/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV									
Analytical Method: EPA 8260									
4-Bromofluorobenzene (S)	98 %		70-114		1		10/27/10 21:32	460-00-4	
Dibromofluoromethane (S)	101 %		88-117		1		10/27/10 21:32	1868-53-7	
1,2-Dichloroethane-d4 (S)	102 %		86-125		1		10/27/10 21:32	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/27/10 21:32	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-9		Lab ID: 3519325023		Collected: 10/18/10 12:16		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Analytical Method:									
Field pH	6.42	Std. Units			1		10/18/10 12:16		
Field Temperature	27.06	deg C			1		10/18/10 12:16		
Field Specific Conductance	1326	umhos/cm			1		10/18/10 12:16		
Oxygen, Dissolved	0.13	mg/L			1		10/18/10 12:16	7782-44-7	
Turbidity	8.31	NTU			1		10/18/10 12:16		
6010 MET ICP									
Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	53.5	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:34	7440-38-2	
Iron	25600	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:34	7439-89-6	
2540C Total Dissolved Solids									
Analytical Method: SM 2540C									
Total Dissolved Solids	770	mg/L	10.0	10.0	1		10/23/10 10:20		
350.1 Ammonia									
Analytical Method: EPA 350.1									
Nitrogen, Ammonia	10.5	mg/L	0.050	0.020	1		10/25/10 08:57	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: CW-10R		Lab ID: 3519325024		Collected: 10/18/10 14:00		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Analytical Method:									
Field pH	5.89	Std. Units			1		10/18/10 14:00		
Field Temperature	27.40	deg C			1		10/18/10 14:00		
Field Specific Conductance	2107	umhos/cm			1		10/18/10 14:00		
Oxygen, Dissolved	0.15	mg/L			1		10/18/10 14:00	7782-44-7	
Turbidity	2.79	NTU			1		10/18/10 14:00		
6010 MET ICP									
Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	7.21	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:38	7440-38-2	
Iron	6890	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:38	7439-89-6	
2540C Total Dissolved Solids									
Analytical Method: SM 2540C									
Total Dissolved Solids	1600	mg/L	10.0	10.0	1		10/23/10 10:20		
350.1 Ammonia									
Analytical Method: EPA 350.1									
Nitrogen, Ammonia	3.4	mg/L	0.050	0.020	1		10/25/10 09:01	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-18		Lab ID: 3519325025		Collected: 10/18/10 09:23		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.29	Std. Units			1		10/18/10 09:23		
Field Temperature	25.9	deg C			1		10/18/10 09:23		
Field Specific Conductance	1268	umhos/cm			1		10/18/10 09:23		
Oxygen, Dissolved	0.48	mg/L			1		10/18/10 09:23	7782-44-7	
Turbidity	14	NTU			1		10/18/10 09:23		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/27/10 16:30	10/29/10 12:11	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/27/10 16:30	10/29/10 12:11	106-93-4	
6010 MET ICP									
		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Arsenic	10.3	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:42	7440-38-2	
Barium	75.1	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:42	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:42	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:42	7440-43-9	
Calcium	246	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:42	7440-70-2	
Chromium	18.2	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:42	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:42	7440-48-4	
Copper	4.0 I	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:42	7440-50-8	
Iron	37100	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:42	7439-89-6	
Lead	7.2 I	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:42	7439-92-1	
Magnesium	27.6	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:42	7439-95-4	
Nickel	5.1	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:42	7440-02-0	
Potassium	1.5	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:42	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/22/10 10:00	10/23/10 14:42	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:42	7440-22-4	
Sodium	9.2	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:42	7440-23-5	
Vanadium	21.7	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:42	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/22/10 10:00	10/23/10 14:42	7440-66-6	
6020 MET ICPMS									
		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Antimony	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:18	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:18	7440-28-0	
7470 Mercury									
		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	10/22/10 09:30	10/25/10 14:11	7439-97-6	
8260 MSV									
		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/31/10 00:01	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	74-83-9	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-18 **Lab ID: 3519325025** Collected: 10/18/10 09:23 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/31/10 00:01	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/31/10 00:01	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/31/10 00:01	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/31/10 00:01	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/31/10 00:01	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/31/10 00:01	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/31/10 00:01	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/31/10 00:01	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/31/10 00:01	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/31/10 00:01	1330-20-7	
4-Bromofluorobenzene (S)	98 %		70-114		1		10/31/10 00:01	460-00-4	
Dibromofluoromethane (S)	105 %		88-117		1		10/31/10 00:01	1868-53-7	
1,2-Dichloroethane-d4 (S)	108 %		86-125		1		10/31/10 00:01	17060-07-0	
Toluene-d8 (S)	102 %		87-113		1		10/31/10 00:01	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	710	mg/L	5.0	5.0	1	10/22/10 08:47
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/22/10 08:47

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 152 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-18		Lab ID: 3519325025		Collected: 10/18/10 09:23		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO ₃	710	mg/L	5.0	5.0	1		10/22/10 08:47		
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	746	mg/L	10.0	10.0	1		10/23/10 10:20		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	16.4	mg/L	10.0	5.0	2		10/28/10 13:11	16887-00-6	
Sulfate	5.0U	mg/L	10.0	5.0	2		10/28/10 13:11	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	2.3	mg/L	0.050	0.020	1		10/25/10 09:03	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-19		Lab ID: 3519325026		Collected: 10/18/10 11:40		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
Analytical Method:									
Field pH	6.08	Std. Units			1		10/18/10 11:40		
Field Temperature	27.3	deg C			1		10/18/10 11:40		
Field Specific Conductance	823	umhos/cm			1		10/18/10 11:40		
Oxygen, Dissolved	0.92	mg/L			1		10/18/10 11:40	7782-44-7	
Turbidity	7	NTU			1		10/18/10 11:40		
8011 GCS EDB and DBCP									
Analytical Method: EPA 8011		Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/27/10 16:30	10/29/10 12:43	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/27/10 16:30	10/29/10 12:43	106-93-4	
8081 GCS Pesticides									
Analytical Method: EPA 8081		Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 00:50	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/25/10 15:20	10/29/10 00:50	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 00:50	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/25/10 15:20	10/29/10 00:50	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	10/25/10 15:20	10/29/10 00:50	58-89-9	
Chlordane (Technical)	0.082U	ug/L	0.51	0.082	1	10/25/10 15:20	10/29/10 00:50	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/25/10 15:20	10/29/10 00:50	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/25/10 15:20	10/29/10 00:50	72-54-8	
4,4'-DDE	0.00092U	ug/L	0.010	0.00092	1	10/25/10 15:20	10/29/10 00:50	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/25/10 15:20	10/29/10 00:50	50-29-3	L3
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 00:50	60-57-1	
Endosulfan I	0.00072U	ug/L	0.010	0.00072	1	10/25/10 15:20	10/29/10 00:50	959-98-8	
Endosulfan II	0.00072U	ug/L	0.010	0.00072	1	10/25/10 15:20	10/29/10 00:50	33213-65-9	
Endosulfan sulfate	0.00061U	ug/L	0.010	0.00061	1	10/25/10 15:20	10/29/10 00:50	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/25/10 15:20	10/29/10 00:50	72-20-8	
Endrin aldehyde	0.0073U	ug/L	0.010	0.0073	1	10/25/10 15:20	10/29/10 00:50	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/25/10 15:20	10/29/10 00:50	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/25/10 15:20	10/29/10 00:50	1024-57-3	
Methoxychlor	0.0072U	ug/L	0.010	0.0072	1	10/25/10 15:20	10/29/10 00:50	72-43-5	L3
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/25/10 15:20	10/29/10 00:50	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	10/25/10 15:20	10/29/10 00:50	8001-35-2	
Tetrachloro-m-xylene (S)	81 %		66.5-120.3		1	10/25/10 15:20	10/29/10 00:50	877-09-8	
Decachlorobiphenyl (S)	79 %		41.7-109.1		1	10/25/10 15:20	10/29/10 00:50	2051-24-3	
8082 GCS PCB									
Analytical Method: EPA 8082		Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.082U	ug/L	0.51	0.082	1	10/25/10 15:20	10/29/10 00:50	12674-11-2	
PCB-1221 (Aroclor 1221)	0.083U	ug/L	0.51	0.083	1	10/25/10 15:20	10/29/10 00:50	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	10/25/10 15:20	10/29/10 00:50	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	10/25/10 15:20	10/29/10 00:50	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	10/25/10 15:20	10/29/10 00:50	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	10/25/10 15:20	10/29/10 00:50	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	10/25/10 15:20	10/29/10 00:50	11096-82-5	L3
Tetrachloro-m-xylene (S)	76 %		48-111		1	10/25/10 15:20	10/29/10 00:50	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 154 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-19		Lab ID: 3519325026		Collected: 10/18/10 11:40		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	71 %		63-121		1	10/25/10 15:20	10/29/10 00:50	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.52	0.19	1	10/25/10 17:10	11/09/10 19:18	60-51-5	
Disulfoton	0.16U	ug/L	0.52	0.16	1	10/25/10 17:10	11/09/10 19:18	298-04-4	
Famphur	0.15U	ug/L	0.52	0.15	1	10/25/10 17:10	11/09/10 19:18	52-85-7	
Methyl parathion	0.20U	ug/L	0.52	0.20	1	10/25/10 17:10	11/09/10 19:18	298-00-0	
Parathion (Ethyl parathion)	0.37U	ug/L	1.0	0.37	1	10/25/10 17:10	11/09/10 19:18	56-38-2	
Phorate	0.38U	ug/L	1.0	0.38	1	10/25/10 17:10	11/09/10 19:18	298-02-2	
4-Chloro3nitrobenzotrifluoride	76 %		34.2-122		1	10/25/10 17:10	11/09/10 19:18		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.97	0.23	1	10/25/10 00:00	10/27/10 07:59	94-75-7	
Dinoseb	0.059U	ug/L	0.20	0.059	1	10/25/10 00:00	10/27/10 07:59	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.029	0.018	1	10/25/10 00:00	10/27/10 07:59	87-86-5	
2,4,5-T	0.044U	ug/L	0.20	0.044	1	10/25/10 00:00	10/27/10 07:59	93-76-5	
2,4,5-TP (Silvex)	0.051U	ug/L	0.20	0.051	1	10/25/10 00:00	10/27/10 07:59	93-72-1	
2,4-DCPA (S)	92 %		65.5-125.7		1	10/25/10 00:00	10/27/10 07:59	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Aluminum	471	ug/L	100	50.0	1	10/22/10 10:00	10/23/10 14:49	7429-90-5	
Arsenic	38.0	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:49	7440-38-2	
Barium	36.9	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:49	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:49	7440-41-7	
Cadmium	0.54 I	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:49	7440-43-9	
Calcium	64.6	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:49	7440-70-2	
Chromium	2.6 I	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:49	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:49	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:49	7440-50-8	
Iron	76000	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:49	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:49	7439-92-1	
Magnesium	22.0	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:49	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:49	7440-02-0	
Potassium	2.5	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:49	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/22/10 10:00	10/23/10 14:49	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:49	7440-22-4	
Sodium	18.5	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:49	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/22/10 10:00	10/23/10 14:49	7440-31-5	
Vanadium	8.9 I	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:49	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/22/10 10:00	10/23/10 14:49	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:25	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:25	7440-28-0	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 155 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-19		Lab ID: 3519325026	Collected: 10/18/10 11:40	Received: 10/21/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	10/27/10 09:00	10/28/10 14:26	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.89U	ug/L	5.2	0.89	1	10/22/10 15:15	10/30/10 08:12	83-32-9	
Acenaphthylene	0.99U	ug/L	5.2	0.99	1	10/22/10 15:15	10/30/10 08:12	208-96-8	
Acetophenone	1.5U	ug/L	5.2	1.5	1	10/22/10 15:15	10/30/10 08:12	98-86-2	
2-Acetylaminofluorene	0.68U	ug/L	5.2	0.68	1	10/22/10 15:15	10/30/10 08:12	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.2	2.9	1	10/22/10 15:15	10/30/10 08:12	92-67-1	
Anthracene	0.62U	ug/L	5.2	0.62	1	10/22/10 15:15	10/30/10 08:12	120-12-7	
Benzo(a)anthracene	0.65U	ug/L	5.2	0.65	1	10/22/10 15:15	10/30/10 08:12	56-55-3	
Benzo(a)pyrene	0.60U	ug/L	1.0	0.60	1	10/22/10 15:15	10/30/10 08:12	50-32-8	
Benzo(b)fluoranthene	0.64U	ug/L	2.1	0.64	1	10/22/10 15:15	10/30/10 08:12	205-99-2	
Benzo(g,h,i)perylene	0.71U	ug/L	5.2	0.71	1	10/22/10 15:15	10/30/10 08:12	191-24-2	
Benzo(k)fluoranthene	0.53U	ug/L	4.2	0.53	1	10/22/10 15:15	10/30/10 08:12	207-08-9	
Benzyl alcohol	1.1U	ug/L	5.2	1.1	1	10/22/10 15:15	10/30/10 08:12	100-51-6	
4-Bromophenylphenyl ether	0.70U	ug/L	5.2	0.70	1	10/22/10 15:15	10/30/10 08:12	101-55-3	
Butylbenzylphthalate	0.75U	ug/L	5.2	0.75	1	10/22/10 15:15	10/30/10 08:12	85-68-7	
4-Chloro-3-methylphenol	0.64U	ug/L	20.8	0.64	1	10/22/10 15:15	10/30/10 08:12	59-50-7	
4-Chloroaniline	1.3U	ug/L	5.2	1.3	1	10/22/10 15:15	10/30/10 08:12	106-47-8	
bis(2-Chloroethoxy)methane	3.1U	ug/L	5.2	3.1	1	10/22/10 15:15	10/30/10 08:12	111-91-1	
bis(2-Chloroethyl) ether	0.78U	ug/L	4.2	0.78	1	10/22/10 15:15	10/30/10 08:12	111-44-4	
bis(2-Chloroisopropyl) ether	0.76U	ug/L	5.2	0.76	1	10/22/10 15:15	10/30/10 08:12	108-60-1	
2-Chloronaphthalene	0.83U	ug/L	5.2	0.83	1	10/22/10 15:15	10/30/10 08:12	91-58-7	
2-Chlorophenol	0.71U	ug/L	5.2	0.71	1	10/22/10 15:15	10/30/10 08:12	95-57-8	
4-Chlorophenylphenyl ether	0.65U	ug/L	5.2	0.65	1	10/22/10 15:15	10/30/10 08:12	7005-72-3	L3
Chrysene	0.38U	ug/L	5.2	0.38	1	10/22/10 15:15	10/30/10 08:12	218-01-9	
Diallate	0.76U	ug/L	5.2	0.76	1	10/22/10 15:15	10/30/10 08:12	2303-16-4	
Dibenz(a,h)anthracene	0.68U	ug/L	2.1	0.68	1	10/22/10 15:15	10/30/10 08:12	53-70-3	
Dibenzofuran	0.70U	ug/L	5.2	0.70	1	10/22/10 15:15	10/30/10 08:12	132-64-9	
1,2-Dichlorobenzene	0.71U	ug/L	5.2	0.71	1	10/22/10 15:15	10/30/10 08:12	95-50-1	
1,3-Dichlorobenzene	0.79U	ug/L	5.2	0.79	1	10/22/10 15:15	10/30/10 08:12	541-73-1	
1,4-Dichlorobenzene	0.80U	ug/L	5.2	0.80	1	10/22/10 15:15	10/30/10 08:12	106-46-7	
3,3'-Dichlorobenzidine	0.72U	ug/L	10.4	0.72	1	10/22/10 15:15	10/30/10 08:12	91-94-1	
2,4-Dichlorophenol	0.58U	ug/L	2.1	0.58	1	10/22/10 15:15	10/30/10 08:12	120-83-2	
2,6-Dichlorophenol	0.64U	ug/L	4.2	0.64	1	10/22/10 15:15	10/30/10 08:12	87-65-0	
Diethylphthalate	0.53U	ug/L	5.2	0.53	1	10/22/10 15:15	10/30/10 08:12	84-66-2	
P-Dimethylaminoazobenzene	0.70U	ug/L	5.2	0.70	1	10/22/10 15:15	10/30/10 08:12	60-11-7	
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.2	2.0	1	10/22/10 15:15	10/30/10 08:12	57-97-6	
3,3'-Dimethylbenzidine	3.3U	ug/L	10.4	3.3	1	10/22/10 15:15	10/30/10 08:12	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.2	1.6	1	10/22/10 15:15	10/30/10 08:12	105-67-9	
Dimethylphthalate	0.67U	ug/L	5.2	0.67	1	10/22/10 15:15	10/30/10 08:12	131-11-3	
Di-n-butylphthalate	0.43U	ug/L	5.2	0.43	1	10/22/10 15:15	10/30/10 08:12	84-74-2	
4,6-Dinitro-2-methylphenol	1.4U	ug/L	20.8	1.4	1	10/22/10 15:15	10/30/10 08:12	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.2	1.2	1	10/22/10 15:15	10/30/10 08:12	528-29-0	L3
1,3-Dinitrobenzene	0.71U	ug/L	8.3	0.71	1	10/22/10 15:15	10/30/10 08:12	99-65-0	
2,4-Dinitrophenol	1.6U	ug/L	20.8	1.6	1	10/22/10 15:15	10/30/10 08:12	51-28-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 156 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-19 **Lab ID: 3519325026** Collected: 10/18/10 11:40 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrotoluene	0.55U	ug/L	2.1	0.55	1	10/22/10 15:15	10/30/10 08:12	121-14-2	
2,6-Dinitrotoluene	1.3U	ug/L	2.1	1.3	1	10/22/10 15:15	10/30/10 08:12	606-20-2	
Di-n-octylphthalate	0.94U	ug/L	5.2	0.94	1	10/22/10 15:15	10/30/10 08:12	117-84-0	
bis(2-Ethylhexyl)phthalate	0.83U	ug/L	5.2	0.83	1	10/22/10 15:15	10/30/10 08:12	117-81-7	
Ethyl methanesulfonate	0.94U	ug/L	5.2	0.94	1	10/22/10 15:15	10/30/10 08:12	62-50-0	
Fluoranthene	0.56U	ug/L	5.2	0.56	1	10/22/10 15:15	10/30/10 08:12	206-44-0	
Fluorene	0.58U	ug/L	5.2	0.58	1	10/22/10 15:15	10/30/10 08:12	86-73-7	
Hexachlorobenzene	0.83U	ug/L	1.0	0.83	1	10/22/10 15:15	10/30/10 08:12	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.2	1.3	1	10/22/10 15:15	10/30/10 08:12	77-47-4	
Hexachloroethane	0.74U	ug/L	5.2	0.74	1	10/22/10 15:15	10/30/10 08:12	67-72-1	
Hexachloropropene	1.5U	ug/L	5.2	1.5	1	10/22/10 15:15	10/30/10 08:12	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.76U	ug/L	2.1	0.76	1	10/22/10 15:15	10/30/10 08:12	193-39-5	
Isodrin	0.56U	ug/L	5.2	0.56	1	10/22/10 15:15	10/30/10 08:12	465-73-6	
Isophorone	0.76U	ug/L	5.2	0.76	1	10/22/10 15:15	10/30/10 08:12	78-59-1	
Isosafrole	0.62U	ug/L	5.2	0.62	1	10/22/10 15:15	10/30/10 08:12	120-58-1	
Methapyrilene	1.7U	ug/L	5.2	1.7	1	10/22/10 15:15	10/30/10 08:12	91-80-5	
3-Methylcholanthrene	1.1U	ug/L	5.2	1.1	1	10/22/10 15:15	10/30/10 08:12	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.2	1.0	1	10/22/10 15:15	10/30/10 08:12	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.2	1.0	1	10/22/10 15:15	10/30/10 08:12	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.2	1.0	1	10/22/10 15:15	10/30/10 08:12	91-57-6	
2-Methylphenol(o-Cresol)	0.76U	ug/L	5.2	0.76	1	10/22/10 15:15	10/30/10 08:12	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.69U	ug/L	10.4	0.69	1	10/22/10 15:15	10/30/10 08:12		
2-Naphthylamine	2.4U	ug/L	5.2	2.4	1	10/22/10 15:15	10/30/10 08:12	91-59-8	
Naphthalene	0.81U	ug/L	5.2	0.81	1	10/22/10 15:15	10/30/10 08:12	91-20-3	L3
1-Naphthylamine	1.1U	ug/L	5.2	1.1	1	10/22/10 15:15	10/30/10 08:12	134-32-7	L3
1,4-Naphthoquinone	1.2U	ug/L	5.2	1.2	1	10/22/10 15:15	10/30/10 08:12	130-15-4	
2-Nitroaniline	0.62U	ug/L	5.2	0.62	1	10/22/10 15:15	10/30/10 08:12	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.2	1.0	1	10/22/10 15:15	10/30/10 08:12	99-09-2	
4-Nitroaniline	0.72U	ug/L	4.2	0.72	1	10/22/10 15:15	10/30/10 08:12	100-01-6	
Nitrobenzene	1.1U	ug/L	4.2	1.1	1	10/22/10 15:15	10/30/10 08:12	98-95-3	
2-Nitrophenol	0.84U	ug/L	5.2	0.84	1	10/22/10 15:15	10/30/10 08:12	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.8	1.1	1	10/22/10 15:15	10/30/10 08:12	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.2	1.3	1	10/22/10 15:15	10/30/10 08:12	99-55-8	
N-Nitrosodiethylamine	0.76U	ug/L	4.2	0.76	1	10/22/10 15:15	10/30/10 08:12	55-18-5	
N-Nitrosodimethylamine	1.0U	ug/L	2.1	1.0	1	10/22/10 15:15	10/30/10 08:12	62-75-9	
N-Nitroso-di-n-butylamine	0.57U	ug/L	4.2	0.57	1	10/22/10 15:15	10/30/10 08:12	924-16-3	
N-Nitroso-di-n-propylamine	0.98U	ug/L	4.2	0.98	1	10/22/10 15:15	10/30/10 08:12	621-64-7	
N-Nitrosodiphenylamine	0.52U	ug/L	5.2	0.52	1	10/22/10 15:15	10/30/10 08:12	86-30-6	
N-Nitrosomethylethylamine	0.77U	ug/L	5.2	0.77	1	10/22/10 15:15	10/30/10 08:12	10595-95-6	
N-Nitrosopiperidine	0.67U	ug/L	5.2	0.67	1	10/22/10 15:15	10/30/10 08:12	100-75-4	
N-Nitrosopyrrolidine	0.91U	ug/L	5.2	0.91	1	10/22/10 15:15	10/30/10 08:12	930-55-2	
O,O,O-Triethylphosphorothioate	0.72U	ug/L	5.2	0.72	1	10/22/10 15:15	10/30/10 08:12	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.2	1.2	1	10/22/10 15:15	10/30/10 08:12	56-38-2	
Pentachlorobenzene	0.81U	ug/L	5.2	0.81	1	10/22/10 15:15	10/30/10 08:12	608-93-5	
Pentachlorophenol	0.69U	ug/L	20.8	0.69	1	10/22/10 15:15	10/30/10 08:12	87-86-5	
Phenacetin	0.55U	ug/L	5.2	0.55	1	10/22/10 15:15	10/30/10 08:12	62-44-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 157 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-19		Lab ID: 3519325026		Collected: 10/18/10 11:40		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenanthrene	0.54U	ug/L	5.2	0.54	1	10/22/10 15:15	10/30/10 08:12	85-01-8	
Phenol	0.56U	ug/L	5.2	0.56	1	10/22/10 15:15	10/30/10 08:12	108-95-2	
Pronamide	1.2U	ug/L	5.2	1.2	1	10/22/10 15:15	10/30/10 08:12	23950-58-5	
Pyrene	0.71U	ug/L	5.2	0.71	1	10/22/10 15:15	10/30/10 08:12	129-00-0	
Safrole	0.88U	ug/L	5.2	0.88	1	10/22/10 15:15	10/30/10 08:12	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.73U	ug/L	5.2	0.73	1	10/22/10 15:15	10/30/10 08:12	95-94-3	
2,3,4,6-Tetrachlorophenol	4.0U	ug/L	5.2	4.0	1	10/22/10 15:15	10/30/10 08:12	58-90-2	
Thionazin	0.63U	ug/L	5.2	0.63	1	10/22/10 15:15	10/30/10 08:12	297-97-2	
O-Toluidine	1.1U	ug/L	5.2	1.1	1	10/22/10 15:15	10/30/10 08:12	95-53-4	
1,2,4-Trichlorobenzene	0.86U	ug/L	5.2	0.86	1	10/22/10 15:15	10/30/10 08:12	120-82-1	
2,4,5-Trichlorophenol	0.54U	ug/L	4.2	0.54	1	10/22/10 15:15	10/30/10 08:12	95-95-4	
2,4,6-Trichlorophenol	0.72U	ug/L	2.1	0.72	1	10/22/10 15:15	10/30/10 08:12	88-06-2	
1,3,5-Trinitrobenzene	1.3U	ug/L	5.2	1.3	1	10/22/10 15:15	10/30/10 08:12	99-35-4	
Nitrobenzene-d5 (S)	75	%	10-110		1	10/22/10 15:15	10/30/10 08:12	4165-60-0	
2-Fluorobiphenyl (S)	80	%	18-110		1	10/22/10 15:15	10/30/10 08:12	321-60-8	
Terphenyl-d14 (S)	96	%	10-123		1	10/22/10 15:15	10/30/10 08:12	1718-51-0	
Phenol-d6 (S)	24	%	10-110		1	10/22/10 15:15	10/30/10 08:12	13127-88-3	
2-Fluorophenol (S)	42	%	18-110		1	10/22/10 15:15	10/30/10 08:12	367-12-4	
2,4,6-Tribromophenol (S)	92	%	10-110		1	10/22/10 15:15	10/30/10 08:12	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/25/10 20:05	10/26/10 13:24	83-32-9	
Acenaphthylene	0.051U	ug/L	2.0	0.051	1	10/25/10 20:05	10/26/10 13:24	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/25/10 20:05	10/26/10 13:24	120-12-7	
Benzo(a)anthracene	0.061U	ug/L	0.20	0.061	1	10/25/10 20:05	10/26/10 13:24	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.20	0.051	1	10/25/10 20:05	10/26/10 13:24	50-32-8	J(M1)
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/25/10 20:05	10/26/10 13:24	205-99-2	
Benzo(g,h,i)perylene	0.061U	ug/L	1.0	0.061	1	10/25/10 20:05	10/26/10 13:24	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/25/10 20:05	10/26/10 13:24	207-08-9	
Chrysene	0.061U	ug/L	1.0	0.061	1	10/25/10 20:05	10/26/10 13:24	218-01-9	
Dibenz(a,h)anthracene	0.051U	ug/L	0.20	0.051	1	10/25/10 20:05	10/26/10 13:24	53-70-3	
Fluoranthene	0.061U	ug/L	1.0	0.061	1	10/25/10 20:05	10/26/10 13:24	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/25/10 20:05	10/26/10 13:24	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/25/10 20:05	10/26/10 13:24	193-39-5	
1-Methylnaphthalene	0.092U	ug/L	1.5	0.092	1	10/25/10 20:05	10/26/10 13:24	90-12-0	
2-Methylnaphthalene	0.061U	ug/L	1.5	0.061	1	10/25/10 20:05	10/26/10 13:24	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/25/10 20:05	10/26/10 13:24	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/25/10 20:05	10/26/10 13:24	85-01-8	
Pyrene	0.061U	ug/L	1.0	0.061	1	10/25/10 20:05	10/26/10 13:24	129-00-0	
2-Fluorobiphenyl (S)	60	%	43.9-113		1	10/25/10 20:05	10/26/10 13:24	321-60-8	
Terphenyl-d14 (S)	67	%	24.8-144		1	10/25/10 20:05	10/26/10 13:24	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.7 I	ug/L	10.0	5.0	1		10/29/10 08:55	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 08:55	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 158 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-19 **Lab ID: 3519325026** Collected: 10/18/10 11:40 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 08:55	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 08:55	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 08:55	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:55	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:55	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 08:55	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 08:55	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:55	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 08:55	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 159 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-19		Lab ID: 3519325026		Collected: 10/18/10 11:40		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 08:55	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 08:55	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 08:55	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 08:55	460-00-4	p2
Dibromofluoromethane (S)	99 %		88-117		1		10/29/10 08:55	1868-53-7	
1,2-Dichloroethane-d4 (S)	102 %		86-125		1		10/29/10 08:55	17060-07-0	
Toluene-d8 (S)	100 %		87-113		1		10/29/10 08:55	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	497	mg/L	5.0	5.0	1		10/23/10 10:20		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/25/10 13:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	12.7	mg/L	5.0	2.5	1		10/28/10 13:47	16887-00-6	
Sulfate	4.9 I	mg/L	5.0	2.5	1		10/28/10 13:47	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 16:15	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	12.5	mg/L	0.050	0.020	1		10/25/10 09:04	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-20		Lab ID: 3519325027		Collected: 10/18/10 13:45		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.58	Std. Units			1		10/18/10 13:45		
Field Temperature	30.4	deg C			1		10/18/10 13:45		
Field Specific Conductance	2372	umhos/cm			1		10/18/10 13:45		
Oxygen, Dissolved	0.65	mg/L			1		10/18/10 13:45	7782-44-7	
Turbidity	8	NTU			1		10/18/10 13:45		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	10/27/10 16:30	10/29/10 12:59	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	10/27/10 16:30	10/29/10 12:59	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 01:07	309-00-2	
alpha-BHC	0.00031U	ug/L	0.010	0.00031	1	10/25/10 15:20	10/29/10 01:07	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 01:07	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	10/25/10 15:20	10/29/10 01:07	319-86-8	
gamma-BHC (Lindane)	0.00021U	ug/L	0.010	0.00021	1	10/25/10 15:20	10/29/10 01:07	58-89-9	
Chlordane (Technical)	0.082U	ug/L	0.51	0.082	1	10/25/10 15:20	10/29/10 01:07	57-74-9	
Chlorobenzilate	0.022U	ug/L	0.10	0.022	1	10/25/10 15:20	10/29/10 01:07	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	10/25/10 15:20	10/29/10 01:07	72-54-8	
4,4'-DDE	0.00092U	ug/L	0.010	0.00092	1	10/25/10 15:20	10/29/10 01:07	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	10/25/10 15:20	10/29/10 01:07	50-29-3	L3
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	10/25/10 15:20	10/29/10 01:07	60-57-1	
Endosulfan I	0.00072U	ug/L	0.010	0.00072	1	10/25/10 15:20	10/29/10 01:07	959-98-8	
Endosulfan II	0.00072U	ug/L	0.010	0.00072	1	10/25/10 15:20	10/29/10 01:07	33213-65-9	
Endosulfan sulfate	0.00062U	ug/L	0.010	0.00062	1	10/25/10 15:20	10/29/10 01:07	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	10/25/10 15:20	10/29/10 01:07	72-20-8	
Endrin aldehyde	0.0073U	ug/L	0.010	0.0073	1	10/25/10 15:20	10/29/10 01:07	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	10/25/10 15:20	10/29/10 01:07	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	10/25/10 15:20	10/29/10 01:07	1024-57-3	
Methoxychlor	0.0072U	ug/L	0.010	0.0072	1	10/25/10 15:20	10/29/10 01:07	72-43-5	L3
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	10/25/10 15:20	10/29/10 01:07	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	10/25/10 15:20	10/29/10 01:07	8001-35-2	
Tetrachloro-m-xylene (S)	93 %		66.5-120.3		1	10/25/10 15:20	10/29/10 01:07	877-09-8	
Decachlorobiphenyl (S)	58 %		41.7-109.1		1	10/25/10 15:20	10/29/10 01:07	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.082U	ug/L	0.51	0.082	1	10/25/10 15:20	10/29/10 01:07	12674-11-2	
PCB-1221 (Aroclor 1221)	0.083U	ug/L	0.51	0.083	1	10/25/10 15:20	10/29/10 01:07	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	10/25/10 15:20	10/29/10 01:07	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	10/25/10 15:20	10/29/10 01:07	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	10/25/10 15:20	10/29/10 01:07	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	10/25/10 15:20	10/29/10 01:07	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	10/25/10 15:20	10/29/10 01:07	11096-82-5	L3
Tetrachloro-m-xylene (S)	83 %		48-111		1	10/25/10 15:20	10/29/10 01:07	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 161 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-20		Lab ID: 3519325027		Collected: 10/18/10 13:45		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	53 %		63-121		1	10/25/10 15:20	10/29/10 01:07	2051-24-3	7p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.51	0.19	1	10/25/10 17:10	11/09/10 19:59	60-51-5	
Disulfoton	0.16U	ug/L	0.51	0.16	1	10/25/10 17:10	11/09/10 19:59	298-04-4	
Famphur	0.15U	ug/L	0.51	0.15	1	10/25/10 17:10	11/09/10 19:59	52-85-7	
Methyl parathion	0.20U	ug/L	0.51	0.20	1	10/25/10 17:10	11/09/10 19:59	298-00-0	
Parathion (Ethyl parathion)	0.36U	ug/L	1.0	0.36	1	10/25/10 17:10	11/09/10 19:59	56-38-2	
Phorate	0.38U	ug/L	1.0	0.38	1	10/25/10 17:10	11/09/10 19:59	298-02-2	
4-Chloro3nitrobenzotrifluoride	81 %		34.2-122		1	10/25/10 17:10	11/09/10 19:59		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.23U	ug/L	0.98	0.23	1	10/25/10 00:00	10/27/10 08:26	94-75-7	
Dinoseb	0.059U	ug/L	0.20	0.059	1	10/25/10 00:00	10/27/10 08:26	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.030	0.018	1	10/25/10 00:00	10/27/10 08:26	87-86-5	
2,4,5-T	0.044U	ug/L	0.20	0.044	1	10/25/10 00:00	10/27/10 08:26	93-76-5	
2,4,5-TP (Silvex)	0.051U	ug/L	0.20	0.051	1	10/25/10 00:00	10/27/10 08:26	93-72-1	
2,4-DCPA (S)	26 %		65.5-125.7		1	10/25/10 00:00	10/27/10 08:26	19719-28-9	J(S0)
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	55.6	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:53	7440-38-2	
Barium	105	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:53	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:53	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:53	7440-43-9	
Calcium	255	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:53	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:53	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:53	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:53	7440-50-8	
Iron	38900	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:53	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:53	7439-92-1	
Magnesium	86.6	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:53	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:53	7440-02-0	
Potassium	1.1	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:53	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/22/10 10:00	10/23/10 14:53	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:53	7440-22-4	
Sodium	110	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:53	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/22/10 10:00	10/23/10 14:53	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:53	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/22/10 10:00	10/23/10 14:53	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:29	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:29	7440-28-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-20		Lab ID: 3519325027	Collected: 10/18/10 13:45	Received: 10/21/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	10/27/10 09:00	10/28/10 14:29	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.88U	ug/L	5.1	0.88	1	10/22/10 15:15	10/30/10 08:42	83-32-9	
Acenaphthylene	0.97U	ug/L	5.1	0.97	1	10/22/10 15:15	10/30/10 08:42	208-96-8	
Acetophenone	1.5U	ug/L	5.1	1.5	1	10/22/10 15:15	10/30/10 08:42	98-86-2	
2-Acetylaminofluorene	0.66U	ug/L	5.1	0.66	1	10/22/10 15:15	10/30/10 08:42	53-96-3	
4-Aminobiphenyl	2.9U	ug/L	5.1	2.9	1	10/22/10 15:15	10/30/10 08:42	92-67-1	
Anthracene	0.61U	ug/L	5.1	0.61	1	10/22/10 15:15	10/30/10 08:42	120-12-7	
Benzo(a)anthracene	0.64U	ug/L	5.1	0.64	1	10/22/10 15:15	10/30/10 08:42	56-55-3	
Benzo(a)pyrene	0.59U	ug/L	1.0	0.59	1	10/22/10 15:15	10/30/10 08:42	50-32-8	
Benzo(b)fluoranthene	0.63U	ug/L	2.0	0.63	1	10/22/10 15:15	10/30/10 08:42	205-99-2	
Benzo(g,h,i)perylene	0.69U	ug/L	5.1	0.69	1	10/22/10 15:15	10/30/10 08:42	191-24-2	
Benzo(k)fluoranthene	0.52U	ug/L	4.1	0.52	1	10/22/10 15:15	10/30/10 08:42	207-08-9	
Benzyl alcohol	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	100-51-6	
4-Bromophenylphenyl ether	0.68U	ug/L	5.1	0.68	1	10/22/10 15:15	10/30/10 08:42	101-55-3	
Butylbenzylphthalate	0.73U	ug/L	5.1	0.73	1	10/22/10 15:15	10/30/10 08:42	85-68-7	
4-Chloro-3-methylphenol	0.63U	ug/L	20.4	0.63	1	10/22/10 15:15	10/30/10 08:42	59-50-7	
4-Chloroaniline	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	106-47-8	
bis(2-Chloroethoxy)methane	3.0U	ug/L	5.1	3.0	1	10/22/10 15:15	10/30/10 08:42	111-91-1	
bis(2-Chloroethyl) ether	0.76U	ug/L	4.1	0.76	1	10/22/10 15:15	10/30/10 08:42	111-44-4	
bis(2-Chloroisopropyl) ether	0.74U	ug/L	5.1	0.74	1	10/22/10 15:15	10/30/10 08:42	108-60-1	
2-Chloronaphthalene	0.82U	ug/L	5.1	0.82	1	10/22/10 15:15	10/30/10 08:42	91-58-7	
2-Chlorophenol	0.69U	ug/L	5.1	0.69	1	10/22/10 15:15	10/30/10 08:42	95-57-8	
4-Chlorophenylphenyl ether	0.64U	ug/L	5.1	0.64	1	10/22/10 15:15	10/30/10 08:42	7005-72-3	L3
Chrysene	0.38U	ug/L	5.1	0.38	1	10/22/10 15:15	10/30/10 08:42	218-01-9	
Diallate	0.74U	ug/L	5.1	0.74	1	10/22/10 15:15	10/30/10 08:42	2303-16-4	
Dibenz(a,h)anthracene	0.66U	ug/L	2.0	0.66	1	10/22/10 15:15	10/30/10 08:42	53-70-3	
Dibenzofuran	0.68U	ug/L	5.1	0.68	1	10/22/10 15:15	10/30/10 08:42	132-64-9	
1,2-Dichlorobenzene	0.69U	ug/L	5.1	0.69	1	10/22/10 15:15	10/30/10 08:42	95-50-1	
1,3-Dichlorobenzene	0.77U	ug/L	5.1	0.77	1	10/22/10 15:15	10/30/10 08:42	541-73-1	
1,4-Dichlorobenzene	0.78U	ug/L	5.1	0.78	1	10/22/10 15:15	10/30/10 08:42	106-46-7	
3,3'-Dichlorobenzidine	0.70U	ug/L	10.2	0.70	1	10/22/10 15:15	10/30/10 08:42	91-94-1	
2,4-Dichlorophenol	0.57U	ug/L	2.0	0.57	1	10/22/10 15:15	10/30/10 08:42	120-83-2	
2,6-Dichlorophenol	0.63U	ug/L	4.1	0.63	1	10/22/10 15:15	10/30/10 08:42	87-65-0	
Diethylphthalate	0.52U	ug/L	5.1	0.52	1	10/22/10 15:15	10/30/10 08:42	84-66-2	
P-Dimethylaminoazobenzene	0.68U	ug/L	5.1	0.68	1	10/22/10 15:15	10/30/10 08:42	60-11-7	
7,12-Dimethylbenz(a)anthracene	2.0U	ug/L	5.1	2.0	1	10/22/10 15:15	10/30/10 08:42	57-97-6	
3,3'-Dimethylbenzidine	3.2U	ug/L	10.2	3.2	1	10/22/10 15:15	10/30/10 08:42	119-93-7	
2,4-Dimethylphenol	1.6U	ug/L	5.1	1.6	1	10/22/10 15:15	10/30/10 08:42	105-67-9	
Dimethylphthalate	0.65U	ug/L	5.1	0.65	1	10/22/10 15:15	10/30/10 08:42	131-11-3	
Di-n-butylphthalate	0.42U	ug/L	5.1	0.42	1	10/22/10 15:15	10/30/10 08:42	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	20.4	1.3	1	10/22/10 15:15	10/30/10 08:42	534-52-1	
1,2-Dinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	528-29-0	L3
1,3-Dinitrobenzene	0.69U	ug/L	8.2	0.69	1	10/22/10 15:15	10/30/10 08:42	99-65-0	
2,4-Dinitrophenol	1.6U	ug/L	20.4	1.6	1	10/22/10 15:15	10/30/10 08:42	51-28-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 163 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-20 **Lab ID: 3519325027** Collected: 10/18/10 13:45 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrotoluene	0.54U	ug/L	2.0	0.54	1	10/22/10 15:15	10/30/10 08:42	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	2.0	1.2	1	10/22/10 15:15	10/30/10 08:42	606-20-2	
Di-n-octylphthalate	0.92U	ug/L	5.1	0.92	1	10/22/10 15:15	10/30/10 08:42	117-84-0	
bis(2-Ethylhexyl)phthalate	0.82U	ug/L	5.1	0.82	1	10/22/10 15:15	10/30/10 08:42	117-81-7	
Ethyl methanesulfonate	0.92U	ug/L	5.1	0.92	1	10/22/10 15:15	10/30/10 08:42	62-50-0	
Fluoranthene	0.55U	ug/L	5.1	0.55	1	10/22/10 15:15	10/30/10 08:42	206-44-0	
Fluorene	0.57U	ug/L	5.1	0.57	1	10/22/10 15:15	10/30/10 08:42	86-73-7	
Hexachlorobenzene	0.82U	ug/L	1.0	0.82	1	10/22/10 15:15	10/30/10 08:42	118-74-1	
Hexachlorocyclopentadiene	1.3U	ug/L	5.1	1.3	1	10/22/10 15:15	10/30/10 08:42	77-47-4	
Hexachloroethane	0.72U	ug/L	5.1	0.72	1	10/22/10 15:15	10/30/10 08:42	67-72-1	
Hexachloropropene	1.4U	ug/L	5.1	1.4	1	10/22/10 15:15	10/30/10 08:42	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.74U	ug/L	2.0	0.74	1	10/22/10 15:15	10/30/10 08:42	193-39-5	
Isodrin	0.55U	ug/L	5.1	0.55	1	10/22/10 15:15	10/30/10 08:42	465-73-6	
Isophorone	0.74U	ug/L	5.1	0.74	1	10/22/10 15:15	10/30/10 08:42	78-59-1	
Isosafrole	0.61U	ug/L	5.1	0.61	1	10/22/10 15:15	10/30/10 08:42	120-58-1	
Methapyrilene	1.7U	ug/L	5.1	1.7	1	10/22/10 15:15	10/30/10 08:42	91-80-5	
3-Methylcholanthrene	1.1U	ug/L	5.1	1.1	1	10/22/10 15:15	10/30/10 08:42	56-49-5	
Methyl methanesulfonate	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	66-27-3	
1-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	90-12-0	
2-Methylnaphthalene	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	91-57-6	
2-Methylphenol(o-Cresol)	0.74U	ug/L	5.1	0.74	1	10/22/10 15:15	10/30/10 08:42	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.67U	ug/L	10.2	0.67	1	10/22/10 15:15	10/30/10 08:42		
2-Naphthylamine	2.3U	ug/L	5.1	2.3	1	10/22/10 15:15	10/30/10 08:42	91-59-8	
Naphthalene	0.79U	ug/L	5.1	0.79	1	10/22/10 15:15	10/30/10 08:42	91-20-3	L3
1-Naphthylamine	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	134-32-7	L3
1,4-Naphthoquinone	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	130-15-4	
2-Nitroaniline	0.61U	ug/L	5.1	0.61	1	10/22/10 15:15	10/30/10 08:42	88-74-4	
3-Nitroaniline	1.0U	ug/L	5.1	1.0	1	10/22/10 15:15	10/30/10 08:42	99-09-2	
4-Nitroaniline	0.70U	ug/L	4.1	0.70	1	10/22/10 15:15	10/30/10 08:42	100-01-6	
Nitrobenzene	1.1U	ug/L	4.1	1.1	1	10/22/10 15:15	10/30/10 08:42	98-95-3	
2-Nitrophenol	0.83U	ug/L	5.1	0.83	1	10/22/10 15:15	10/30/10 08:42	88-75-5	
4-Nitrophenol	1.1U	ug/L	20.4	1.1	1	10/22/10 15:15	10/30/10 08:42	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	5.1	1.3	1	10/22/10 15:15	10/30/10 08:42	99-55-8	
N-Nitrosodiethylamine	0.74U	ug/L	4.1	0.74	1	10/22/10 15:15	10/30/10 08:42	55-18-5	
N-Nitrosodimethylamine	0.99U	ug/L	2.0	0.99	1	10/22/10 15:15	10/30/10 08:42	62-75-9	
N-Nitroso-di-n-butylamine	0.56U	ug/L	4.1	0.56	1	10/22/10 15:15	10/30/10 08:42	924-16-3	
N-Nitroso-di-n-propylamine	0.96U	ug/L	4.1	0.96	1	10/22/10 15:15	10/30/10 08:42	621-64-7	
N-Nitrosodiphenylamine	0.51U	ug/L	5.1	0.51	1	10/22/10 15:15	10/30/10 08:42	86-30-6	
N-Nitrosomethylethylamine	0.75U	ug/L	5.1	0.75	1	10/22/10 15:15	10/30/10 08:42	10595-95-6	
N-Nitrosopiperidine	0.65U	ug/L	5.1	0.65	1	10/22/10 15:15	10/30/10 08:42	100-75-4	
N-Nitrosopyrrolidine	0.90U	ug/L	5.1	0.90	1	10/22/10 15:15	10/30/10 08:42	930-55-2	
O,O,O-Triethylphosphorothioate	0.70U	ug/L	5.1	0.70	1	10/22/10 15:15	10/30/10 08:42	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	56-38-2	
Pentachlorobenzene	0.79U	ug/L	5.1	0.79	1	10/22/10 15:15	10/30/10 08:42	608-93-5	
Pentachlorophenol	0.67U	ug/L	20.4	0.67	1	10/22/10 15:15	10/30/10 08:42	87-86-5	
Phenacetin	0.54U	ug/L	5.1	0.54	1	10/22/10 15:15	10/30/10 08:42	62-44-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 164 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-20 **Lab ID: 3519325027** Collected: 10/18/10 13:45 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Phenanthrene	0.53U	ug/L	5.1	0.53	1	10/22/10 15:15	10/30/10 08:42	85-01-8	
Phenol	0.55U	ug/L	5.1	0.55	1	10/22/10 15:15	10/30/10 08:42	108-95-2	
Pronamide	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	23950-58-5	
Pyrene	0.69U	ug/L	5.1	0.69	1	10/22/10 15:15	10/30/10 08:42	129-00-0	
Safrole	0.87U	ug/L	5.1	0.87	1	10/22/10 15:15	10/30/10 08:42	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.71U	ug/L	5.1	0.71	1	10/22/10 15:15	10/30/10 08:42	95-94-3	
2,3,4,6-Tetrachlorophenol	3.9U	ug/L	5.1	3.9	1	10/22/10 15:15	10/30/10 08:42	58-90-2	
Thionazin	0.62U	ug/L	5.1	0.62	1	10/22/10 15:15	10/30/10 08:42	297-97-2	
O-Toluidine	1.1U	ug/L	5.1	1.1	1	10/22/10 15:15	10/30/10 08:42	95-53-4	
1,2,4-Trichlorobenzene	0.85U	ug/L	5.1	0.85	1	10/22/10 15:15	10/30/10 08:42	120-82-1	
2,4,5-Trichlorophenol	0.53U	ug/L	4.1	0.53	1	10/22/10 15:15	10/30/10 08:42	95-95-4	
2,4,6-Trichlorophenol	0.70U	ug/L	2.0	0.70	1	10/22/10 15:15	10/30/10 08:42	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	5.1	1.2	1	10/22/10 15:15	10/30/10 08:42	99-35-4	
Nitrobenzene-d5 (S)	72	%	10-110		1	10/22/10 15:15	10/30/10 08:42	4165-60-0	
2-Fluorobiphenyl (S)	75	%	18-110		1	10/22/10 15:15	10/30/10 08:42	321-60-8	
Terphenyl-d14 (S)	87	%	10-123		1	10/22/10 15:15	10/30/10 08:42	1718-51-0	
Phenol-d6 (S)	17	%	10-110		1	10/22/10 15:15	10/30/10 08:42	13127-88-3	
2-Fluorophenol (S)	34	%	18-110		1	10/22/10 15:15	10/30/10 08:42	367-12-4	
2,4,6-Tribromophenol (S)	85	%	10-110		1	10/22/10 15:15	10/30/10 08:42	118-79-6	
8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.031U	ug/L	1.0	0.031	1	10/25/10 20:05	10/26/10 13:46	83-32-9	
Acenaphthylene	0.052U	ug/L	2.1	0.052	1	10/25/10 20:05	10/26/10 13:46	208-96-8	
Anthracene	0.052U	ug/L	1.0	0.052	1	10/25/10 20:05	10/26/10 13:46	120-12-7	
Benzo(a)anthracene	0.062U	ug/L	0.21	0.062	1	10/25/10 20:05	10/26/10 13:46	56-55-3	
Benzo(a)pyrene	0.052U	ug/L	0.21	0.052	1	10/25/10 20:05	10/26/10 13:46	50-32-8	
Benzo(b)fluoranthene	0.052U	ug/L	0.10	0.052	1	10/25/10 20:05	10/26/10 13:46	205-99-2	
Benzo(g,h,i)perylene	0.062U	ug/L	1.0	0.062	1	10/25/10 20:05	10/26/10 13:46	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/25/10 20:05	10/26/10 13:46	207-08-9	
Chrysene	0.062U	ug/L	1.0	0.062	1	10/25/10 20:05	10/26/10 13:46	218-01-9	
Dibenz(a,h)anthracene	0.052U	ug/L	0.21	0.052	1	10/25/10 20:05	10/26/10 13:46	53-70-3	
Fluoranthene	0.062U	ug/L	1.0	0.062	1	10/25/10 20:05	10/26/10 13:46	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/25/10 20:05	10/26/10 13:46	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/25/10 20:05	10/26/10 13:46	193-39-5	
1-Methylnaphthalene	0.093U	ug/L	1.5	0.093	1	10/25/10 20:05	10/26/10 13:46	90-12-0	
2-Methylnaphthalene	0.062U	ug/L	1.5	0.062	1	10/25/10 20:05	10/26/10 13:46	91-57-6	
Naphthalene	0.083U	ug/L	1.0	0.083	1	10/25/10 20:05	10/26/10 13:46	91-20-3	
Phenanthrene	0.052U	ug/L	1.0	0.052	1	10/25/10 20:05	10/26/10 13:46	85-01-8	
Pyrene	0.062U	ug/L	1.0	0.062	1	10/25/10 20:05	10/26/10 13:46	129-00-0	
2-Fluorobiphenyl (S)	85	%	43.9-113		1	10/25/10 20:05	10/26/10 13:46	321-60-8	
Terphenyl-d14 (S)	102	%	24.8-144		1	10/25/10 20:05	10/26/10 13:46	1718-51-0	
8260 MSV Analytical Method: EPA 8260									
Acetone	6.1 I	ug/L	10.0	5.0	1		10/29/10 07:45	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 07:45	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 165 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-20 **Lab ID: 3519325027** Collected: 10/18/10 13:45 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 07:45	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 07:45	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 07:45	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 07:45	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 07:45	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 07:45	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 07:45	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 07:45	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 07:45	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 166 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-20		Lab ID: 3519325027		Collected: 10/18/10 13:45		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 07:45	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 07:45	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 07:45	75-01-4	
Xylene (Total)	0.58 I	ug/L	1.0	0.50	1		10/29/10 07:45	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 07:45	460-00-4	
Dibromofluoromethane (S)	101 %		88-117		1		10/29/10 07:45	1868-53-7	
1,2-Dichloroethane-d4 (S)	100 %		86-125		1		10/29/10 07:45	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 07:45	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	1350	mg/L	10.0	10.0	1		10/23/10 10:20		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.6	mg/L	1.0	1.0	1		10/25/10 13:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	86.0	mg/L	25.0	12.5	5		10/28/10 13:59	16887-00-6	
Sulfate	27.4	mg/L	25.0	12.5	5		10/28/10 13:59	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 16:19	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	2.6	mg/L	0.050	0.020	1		10/25/10 09:06	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-1R		Lab ID: 3519325028		Collected: 10/18/10 10:35		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.61	Std. Units			1		10/18/10 10:35		
Field Temperature	25.84	deg C			1		10/18/10 10:35		
Field Specific Conductance	562	umhos/cm			1		10/18/10 10:35		
Oxygen, Dissolved	0.13	mg/L			1		10/18/10 10:35	7782-44-7	
Turbidity	1.77	NTU			1		10/18/10 10:35		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/27/10 16:30	10/29/10 13:15	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/27/10 16:30	10/29/10 13:15	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00054U	ug/L	0.011	0.00054	1	10/25/10 15:20	10/29/10 01:58	309-00-2	
alpha-BHC	0.00032U	ug/L	0.011	0.00032	1	10/25/10 15:20	10/29/10 01:58	319-84-6	
beta-BHC	0.00054U	ug/L	0.011	0.00054	1	10/25/10 15:20	10/29/10 01:58	319-85-7	
delta-BHC	0.00043U	ug/L	0.011	0.00043	1	10/25/10 15:20	10/29/10 01:58	319-86-8	
gamma-BHC (Lindane)	0.00021U	ug/L	0.011	0.00021	1	10/25/10 15:20	10/29/10 01:58	58-89-9	
Chlordane (Technical)	0.086U	ug/L	0.54	0.086	1	10/25/10 15:20	10/29/10 01:58	57-74-9	
Chlorobenzilate	0.023U	ug/L	0.11	0.023	1	10/25/10 15:20	10/29/10 01:58	510-15-6	
4,4'-DDD	0.0020U	ug/L	0.011	0.0020	1	10/25/10 15:20	10/29/10 01:58	72-54-8	
4,4'-DDE	0.00097U	ug/L	0.011	0.00097	1	10/25/10 15:20	10/29/10 01:58	72-55-9	
4,4'-DDT	0.0039U	ug/L	0.011	0.0039	1	10/25/10 15:20	10/29/10 01:58	50-29-3	L3
Dieldrin	0.00054U	ug/L	0.011	0.00054	1	10/25/10 15:20	10/29/10 01:58	60-57-1	
Endosulfan I	0.00075U	ug/L	0.011	0.00075	1	10/25/10 15:20	10/29/10 01:58	959-98-8	
Endosulfan II	0.00075U	ug/L	0.011	0.00075	1	10/25/10 15:20	10/29/10 01:58	33213-65-9	
Endosulfan sulfate	0.00064U	ug/L	0.011	0.00064	1	10/25/10 15:20	10/29/10 01:58	1031-07-8	
Endrin	0.0018U	ug/L	0.011	0.0018	1	10/25/10 15:20	10/29/10 01:58	72-20-8	
Endrin aldehyde	0.0076U	ug/L	0.011	0.0076	1	10/25/10 15:20	10/29/10 01:58	7421-93-4	
Heptachlor	0.0016U	ug/L	0.011	0.0016	1	10/25/10 15:20	10/29/10 01:58	76-44-8	
Heptachlor epoxide	0.00043U	ug/L	0.011	0.00043	1	10/25/10 15:20	10/29/10 01:58	1024-57-3	
Methoxychlor	0.0075U	ug/L	0.011	0.0075	1	10/25/10 15:20	10/29/10 01:58	72-43-5	L3
Pentachloronitrobenzene	0.016U	ug/L	0.11	0.016	1	10/25/10 15:20	10/29/10 01:58	82-68-8	
Toxaphene	0.31U	ug/L	0.54	0.31	1	10/25/10 15:20	10/29/10 01:58	8001-35-2	
Tetrachloro-m-xylene (S)	83 %		66.5-120.3		1	10/25/10 15:20	10/29/10 01:58	877-09-8	
Decachlorobiphenyl (S)	77 %		41.7-109.1		1	10/25/10 15:20	10/29/10 01:58	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.086U	ug/L	0.54	0.086	1	10/25/10 15:20	10/29/10 01:58	12674-11-2	
PCB-1221 (Aroclor 1221)	0.087U	ug/L	0.54	0.087	1	10/25/10 15:20	10/29/10 01:58	11104-28-2	
PCB-1232 (Aroclor 1232)	0.13U	ug/L	0.54	0.13	1	10/25/10 15:20	10/29/10 01:58	11141-16-5	
PCB-1242 (Aroclor 1242)	0.14U	ug/L	0.54	0.14	1	10/25/10 15:20	10/29/10 01:58	53469-21-9	
PCB-1248 (Aroclor 1248)	0.30U	ug/L	0.54	0.30	1	10/25/10 15:20	10/29/10 01:58	12672-29-6	
PCB-1254 (Aroclor 1254)	0.16U	ug/L	0.54	0.16	1	10/25/10 15:20	10/29/10 01:58	11097-69-1	
PCB-1260 (Aroclor 1260)	0.12U	ug/L	0.54	0.12	1	10/25/10 15:20	10/29/10 01:58	11096-82-5	L3
Tetrachloro-m-xylene (S)	77 %		48-111		1	10/25/10 15:20	10/29/10 01:58	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 168 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-1R		Lab ID: 3519325028		Collected: 10/18/10 10:35		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	69 %		63-121		1	10/25/10 15:20	10/29/10 01:58	2051-24-3	
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.20U	ug/L	0.54	0.20	1	10/25/10 17:10	11/09/10 20:41	60-51-5	
Disulfoton	0.17U	ug/L	0.54	0.17	1	10/25/10 17:10	11/09/10 20:41	298-04-4	
Famphur	0.16U	ug/L	0.54	0.16	1	10/25/10 17:10	11/09/10 20:41	52-85-7	
Methyl parathion	0.21U	ug/L	0.54	0.21	1	10/25/10 17:10	11/09/10 20:41	298-00-0	
Parathion (Ethyl parathion)	0.38U	ug/L	1.1	0.38	1	10/25/10 17:10	11/09/10 20:41	56-38-2	
Phorate	0.40U	ug/L	1.1	0.40	1	10/25/10 17:10	11/09/10 20:41	298-02-2	
4-Chloro3nitrobenzotrifluoride	65 %		34.2-122		1	10/25/10 17:10	11/09/10 20:41		
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.24U	ug/L	1.0	0.24	1	10/25/10 00:00	10/27/10 08:52	94-75-7	
Dinoseb	0.061U	ug/L	0.20	0.061	1	10/25/10 00:00	10/27/10 08:52	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.030	0.018	1	10/25/10 00:00	10/27/10 08:52	87-86-5	
2,4,5-T	0.045U	ug/L	0.20	0.045	1	10/25/10 00:00	10/27/10 08:52	93-76-5	
2,4,5-TP (Silvex)	0.052U	ug/L	0.20	0.052	1	10/25/10 00:00	10/27/10 08:52	93-72-1	
2,4-DCPA (S)	89 %		65.5-125.7		1	10/25/10 00:00	10/27/10 08:52	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 15:04	7440-38-2	
Barium	34.8	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 15:04	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 15:04	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 15:04	7440-43-9	
Calcium	82.0	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 15:04	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 15:04	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 15:04	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 15:04	7440-50-8	
Iron	6550	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 15:04	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 15:04	7439-92-1	
Magnesium	12.8	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 15:04	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 15:04	7440-02-0	
Potassium	1.0	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 15:04	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/22/10 10:00	10/23/10 15:04	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 15:04	7440-22-4	
Sodium	18.2	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 15:04	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	10/22/10 10:00	10/23/10 15:04	7440-31-5	
Vanadium	5.2	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 15:04	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/22/10 10:00	10/23/10 15:04	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:47	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:47	7440-28-0	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-1R		Lab ID: 3519325028	Collected: 10/18/10 10:35	Received: 10/21/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
7470 Mercury		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	10/27/10 09:00	10/28/10 14:32	7439-97-6	
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	0.93U	ug/L	5.4	0.93	1	10/22/10 15:15	10/30/10 09:13	83-32-9	
Acenaphthylene	1.0U	ug/L	5.4	1.0	1	10/22/10 15:15	10/30/10 09:13	208-96-8	
Acetophenone	1.6U	ug/L	5.4	1.6	1	10/22/10 15:15	10/30/10 09:13	98-86-2	
2-Acetylaminofluorene	0.70U	ug/L	5.4	0.70	1	10/22/10 15:15	10/30/10 09:13	53-96-3	
4-Aminobiphenyl	3.1U	ug/L	5.4	3.1	1	10/22/10 15:15	10/30/10 09:13	92-67-1	
Anthracene	0.65U	ug/L	5.4	0.65	1	10/22/10 15:15	10/30/10 09:13	120-12-7	
Benzo(a)anthracene	0.68U	ug/L	5.4	0.68	1	10/22/10 15:15	10/30/10 09:13	56-55-3	
Benzo(a)pyrene	0.63U	ug/L	1.1	0.63	1	10/22/10 15:15	10/30/10 09:13	50-32-8	
Benzo(b)fluoranthene	0.67U	ug/L	2.2	0.67	1	10/22/10 15:15	10/30/10 09:13	205-99-2	
Benzo(g,h,i)perylene	0.73U	ug/L	5.4	0.73	1	10/22/10 15:15	10/30/10 09:13	191-24-2	
Benzo(k)fluoranthene	0.55U	ug/L	4.3	0.55	1	10/22/10 15:15	10/30/10 09:13	207-08-9	
Benzyl alcohol	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	100-51-6	
4-Bromophenylphenyl ether	0.72U	ug/L	5.4	0.72	1	10/22/10 15:15	10/30/10 09:13	101-55-3	
Butylbenzylphthalate	0.78U	ug/L	5.4	0.78	1	10/22/10 15:15	10/30/10 09:13	85-68-7	
4-Chloro-3-methylphenol	0.67U	ug/L	21.6	0.67	1	10/22/10 15:15	10/30/10 09:13	59-50-7	
4-Chloroaniline	1.3U	ug/L	5.4	1.3	1	10/22/10 15:15	10/30/10 09:13	106-47-8	
bis(2-Chloroethoxy)methane	3.2U	ug/L	5.4	3.2	1	10/22/10 15:15	10/30/10 09:13	111-91-1	
bis(2-Chloroethyl) ether	0.81U	ug/L	4.3	0.81	1	10/22/10 15:15	10/30/10 09:13	111-44-4	
bis(2-Chloroisopropyl) ether	0.79U	ug/L	5.4	0.79	1	10/22/10 15:15	10/30/10 09:13	108-60-1	
2-Chloronaphthalene	0.86U	ug/L	5.4	0.86	1	10/22/10 15:15	10/30/10 09:13	91-58-7	
2-Chlorophenol	0.73U	ug/L	5.4	0.73	1	10/22/10 15:15	10/30/10 09:13	95-57-8	
4-Chlorophenylphenyl ether	0.68U	ug/L	5.4	0.68	1	10/22/10 15:15	10/30/10 09:13	7005-72-3	L3
Chrysene	0.40U	ug/L	5.4	0.40	1	10/22/10 15:15	10/30/10 09:13	218-01-9	
Diallate	0.78U	ug/L	5.4	0.78	1	10/22/10 15:15	10/30/10 09:13	2303-16-4	
Dibenz(a,h)anthracene	0.70U	ug/L	2.2	0.70	1	10/22/10 15:15	10/30/10 09:13	53-70-3	
Dibenzofuran	0.72U	ug/L	5.4	0.72	1	10/22/10 15:15	10/30/10 09:13	132-64-9	
1,2-Dichlorobenzene	0.73U	ug/L	5.4	0.73	1	10/22/10 15:15	10/30/10 09:13	95-50-1	
1,3-Dichlorobenzene	0.82U	ug/L	5.4	0.82	1	10/22/10 15:15	10/30/10 09:13	541-73-1	
1,4-Dichlorobenzene	0.83U	ug/L	5.4	0.83	1	10/22/10 15:15	10/30/10 09:13	106-46-7	
3,3'-Dichlorobenzidine	0.74U	ug/L	10.8	0.74	1	10/22/10 15:15	10/30/10 09:13	91-94-1	
2,4-Dichlorophenol	0.60U	ug/L	2.2	0.60	1	10/22/10 15:15	10/30/10 09:13	120-83-2	
2,6-Dichlorophenol	0.67U	ug/L	4.3	0.67	1	10/22/10 15:15	10/30/10 09:13	87-65-0	
Diethylphthalate	0.55U	ug/L	5.4	0.55	1	10/22/10 15:15	10/30/10 09:13	84-66-2	
P-Dimethylaminoazobenzene	0.72U	ug/L	5.4	0.72	1	10/22/10 15:15	10/30/10 09:13	60-11-7	
7,12-Dimethylbenz(a)anthracene	2.1U	ug/L	5.4	2.1	1	10/22/10 15:15	10/30/10 09:13	57-97-6	
3,3'-Dimethylbenzidine	3.4U	ug/L	10.8	3.4	1	10/22/10 15:15	10/30/10 09:13	119-93-7	
2,4-Dimethylphenol	1.7U	ug/L	5.4	1.7	1	10/22/10 15:15	10/30/10 09:13	105-67-9	
Dimethylphthalate	0.69U	ug/L	5.4	0.69	1	10/22/10 15:15	10/30/10 09:13	131-11-3	
Di-n-butylphthalate	0.44U	ug/L	5.4	0.44	1	10/22/10 15:15	10/30/10 09:13	84-74-2	
4,6-Dinitro-2-methylphenol	1.4U	ug/L	21.6	1.4	1	10/22/10 15:15	10/30/10 09:13	534-52-1	
1,2-Dinitrobenzene	1.3U	ug/L	5.4	1.3	1	10/22/10 15:15	10/30/10 09:13	528-29-0	L3
1,3-Dinitrobenzene	0.73U	ug/L	8.6	0.73	1	10/22/10 15:15	10/30/10 09:13	99-65-0	
2,4-Dinitrophenol	1.7U	ug/L	21.6	1.7	1	10/22/10 15:15	10/30/10 09:13	51-28-5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 170 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-1R **Lab ID: 3519325028** Collected: 10/18/10 10:35 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
2,4-Dinitrotoluene	0.57U	ug/L	2.2	0.57	1	10/22/10 15:15	10/30/10 09:13	121-14-2	
2,6-Dinitrotoluene	1.3U	ug/L	2.2	1.3	1	10/22/10 15:15	10/30/10 09:13	606-20-2	
Di-n-octylphthalate	0.97U	ug/L	5.4	0.97	1	10/22/10 15:15	10/30/10 09:13	117-84-0	
bis(2-Ethylhexyl)phthalate	0.86U	ug/L	5.4	0.86	1	10/22/10 15:15	10/30/10 09:13	117-81-7	
Ethyl methanesulfonate	0.97U	ug/L	5.4	0.97	1	10/22/10 15:15	10/30/10 09:13	62-50-0	
Fluoranthene	0.58U	ug/L	5.4	0.58	1	10/22/10 15:15	10/30/10 09:13	206-44-0	
Fluorene	0.60U	ug/L	5.4	0.60	1	10/22/10 15:15	10/30/10 09:13	86-73-7	
Hexachlorobenzene	0.86U	ug/L	1.1	0.86	1	10/22/10 15:15	10/30/10 09:13	118-74-1	
Hexachlorocyclopentadiene	1.4U	ug/L	5.4	1.4	1	10/22/10 15:15	10/30/10 09:13	77-47-4	
Hexachloroethane	0.77U	ug/L	5.4	0.77	1	10/22/10 15:15	10/30/10 09:13	67-72-1	
Hexachloropropene	1.5U	ug/L	5.4	1.5	1	10/22/10 15:15	10/30/10 09:13	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.79U	ug/L	2.2	0.79	1	10/22/10 15:15	10/30/10 09:13	193-39-5	
Isodrin	0.58U	ug/L	5.4	0.58	1	10/22/10 15:15	10/30/10 09:13	465-73-6	
Isophorone	0.79U	ug/L	5.4	0.79	1	10/22/10 15:15	10/30/10 09:13	78-59-1	
Isosafrole	0.65U	ug/L	5.4	0.65	1	10/22/10 15:15	10/30/10 09:13	120-58-1	
Methapyrilene	1.8U	ug/L	5.4	1.8	1	10/22/10 15:15	10/30/10 09:13	91-80-5	
3-Methylcholanthrene	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	56-49-5	
Methyl methanesulfonate	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	66-27-3	
1-Methylnaphthalene	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	90-12-0	
2-Methylnaphthalene	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	91-57-6	
2-Methylphenol(o-Cresol)	0.79U	ug/L	5.4	0.79	1	10/22/10 15:15	10/30/10 09:13	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.71U	ug/L	10.8	0.71	1	10/22/10 15:15	10/30/10 09:13		
2-Naphthylamine	2.4U	ug/L	5.4	2.4	1	10/22/10 15:15	10/30/10 09:13	91-59-8	
Naphthalene	0.84U	ug/L	5.4	0.84	1	10/22/10 15:15	10/30/10 09:13	91-20-3	L3
1-Naphthylamine	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	134-32-7	L3
1,4-Naphthoquinone	1.3U	ug/L	5.4	1.3	1	10/22/10 15:15	10/30/10 09:13	130-15-4	
2-Nitroaniline	0.65U	ug/L	5.4	0.65	1	10/22/10 15:15	10/30/10 09:13	88-74-4	
3-Nitroaniline	1.1U	ug/L	5.4	1.1	1	10/22/10 15:15	10/30/10 09:13	99-09-2	
4-Nitroaniline	0.74U	ug/L	4.3	0.74	1	10/22/10 15:15	10/30/10 09:13	100-01-6	
Nitrobenzene	1.2U	ug/L	4.3	1.2	1	10/22/10 15:15	10/30/10 09:13	98-95-3	
2-Nitrophenol	0.87U	ug/L	5.4	0.87	1	10/22/10 15:15	10/30/10 09:13	88-75-5	
4-Nitrophenol	1.2U	ug/L	21.6	1.2	1	10/22/10 15:15	10/30/10 09:13	100-02-7	
5-Nitro-o-toluidine	1.4U	ug/L	5.4	1.4	1	10/22/10 15:15	10/30/10 09:13	99-55-8	
N-Nitrosodiethylamine	0.79U	ug/L	4.3	0.79	1	10/22/10 15:15	10/30/10 09:13	55-18-5	
N-Nitrosodimethylamine	1.0U	ug/L	2.2	1.0	1	10/22/10 15:15	10/30/10 09:13	62-75-9	
N-Nitroso-di-n-butylamine	0.59U	ug/L	4.3	0.59	1	10/22/10 15:15	10/30/10 09:13	924-16-3	
N-Nitroso-di-n-propylamine	1.0U	ug/L	4.3	1.0	1	10/22/10 15:15	10/30/10 09:13	621-64-7	
N-Nitrosodiphenylamine	0.54U	ug/L	5.4	0.54	1	10/22/10 15:15	10/30/10 09:13	86-30-6	
N-Nitrosomethylethylamine	0.80U	ug/L	5.4	0.80	1	10/22/10 15:15	10/30/10 09:13	10595-95-6	
N-Nitrosopiperidine	0.69U	ug/L	5.4	0.69	1	10/22/10 15:15	10/30/10 09:13	100-75-4	
N-Nitrosopyrrolidine	0.95U	ug/L	5.4	0.95	1	10/22/10 15:15	10/30/10 09:13	930-55-2	
O,O,O-Triethylphosphorothioate	0.74U	ug/L	5.4	0.74	1	10/22/10 15:15	10/30/10 09:13	126-68-1	
Parathion (Ethyl parathion)	1.2U	ug/L	5.4	1.2	1	10/22/10 15:15	10/30/10 09:13	56-38-2	
Pentachlorobenzene	0.84U	ug/L	5.4	0.84	1	10/22/10 15:15	10/30/10 09:13	608-93-5	
Pentachlorophenol	0.71U	ug/L	21.6	0.71	1	10/22/10 15:15	10/30/10 09:13	87-86-5	
Phenacetin	0.57U	ug/L	5.4	0.57	1	10/22/10 15:15	10/30/10 09:13	62-44-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 171 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-1R		Lab ID: 3519325028		Collected: 10/18/10 10:35		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenanthrene	0.56U	ug/L	5.4	0.56	1	10/22/10 15:15	10/30/10 09:13	85-01-8	
Phenol	0.58U	ug/L	5.4	0.58	1	10/22/10 15:15	10/30/10 09:13	108-95-2	
Pronamide	1.2U	ug/L	5.4	1.2	1	10/22/10 15:15	10/30/10 09:13	23950-58-5	
Pyrene	0.73U	ug/L	5.4	0.73	1	10/22/10 15:15	10/30/10 09:13	129-00-0	
Safrole	0.92U	ug/L	5.4	0.92	1	10/22/10 15:15	10/30/10 09:13	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.75U	ug/L	5.4	0.75	1	10/22/10 15:15	10/30/10 09:13	95-94-3	
2,3,4,6-Tetrachlorophenol	4.2U	ug/L	5.4	4.2	1	10/22/10 15:15	10/30/10 09:13	58-90-2	
Thionazin	0.66U	ug/L	5.4	0.66	1	10/22/10 15:15	10/30/10 09:13	297-97-2	
O-Toluidine	1.2U	ug/L	5.4	1.2	1	10/22/10 15:15	10/30/10 09:13	95-53-4	
1,2,4-Trichlorobenzene	0.89U	ug/L	5.4	0.89	1	10/22/10 15:15	10/30/10 09:13	120-82-1	
2,4,5-Trichlorophenol	0.56U	ug/L	4.3	0.56	1	10/22/10 15:15	10/30/10 09:13	95-95-4	
2,4,6-Trichlorophenol	0.74U	ug/L	2.2	0.74	1	10/22/10 15:15	10/30/10 09:13	88-06-2	
1,3,5-Trinitrobenzene	1.3U	ug/L	5.4	1.3	1	10/22/10 15:15	10/30/10 09:13	99-35-4	
Nitrobenzene-d5 (S)	62 %		10-110		1	10/22/10 15:15	10/30/10 09:13	4165-60-0	
2-Fluorobiphenyl (S)	67 %		18-110		1	10/22/10 15:15	10/30/10 09:13	321-60-8	
Terphenyl-d14 (S)	88 %		10-123		1	10/22/10 15:15	10/30/10 09:13	1718-51-0	
Phenol-d6 (S)	30 %		10-110		1	10/22/10 15:15	10/30/10 09:13	13127-88-3	
2-Fluorophenol (S)	43 %		18-110		1	10/22/10 15:15	10/30/10 09:13	367-12-4	
2,4,6-Tribromophenol (S)	84 %		10-110		1	10/22/10 15:15	10/30/10 09:13	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.033U	ug/L	1.1	0.033	1	10/25/10 20:05	10/26/10 14:08	83-32-9	
Acenaphthylene	0.055U	ug/L	2.2	0.055	1	10/25/10 20:05	10/26/10 14:08	208-96-8	
Anthracene	0.055U	ug/L	1.1	0.055	1	10/25/10 20:05	10/26/10 14:08	120-12-7	
Benzo(a)anthracene	0.066U	ug/L	0.22	0.066	1	10/25/10 20:05	10/26/10 14:08	56-55-3	
Benzo(a)pyrene	0.055U	ug/L	0.22	0.055	1	10/25/10 20:05	10/26/10 14:08	50-32-8	
Benzo(b)fluoranthene	0.055U	ug/L	0.11	0.055	1	10/25/10 20:05	10/26/10 14:08	205-99-2	
Benzo(g,h,i)perylene	0.066U	ug/L	1.1	0.066	1	10/25/10 20:05	10/26/10 14:08	191-24-2	
Benzo(k)fluoranthene	0.044U	ug/L	0.27	0.044	1	10/25/10 20:05	10/26/10 14:08	207-08-9	
Chrysene	0.066U	ug/L	1.1	0.066	1	10/25/10 20:05	10/26/10 14:08	218-01-9	
Dibenz(a,h)anthracene	0.055U	ug/L	0.22	0.055	1	10/25/10 20:05	10/26/10 14:08	53-70-3	
Fluoranthene	0.066U	ug/L	1.1	0.066	1	10/25/10 20:05	10/26/10 14:08	206-44-0	
Fluorene	0.033U	ug/L	1.1	0.033	1	10/25/10 20:05	10/26/10 14:08	86-73-7	
Indeno(1,2,3-cd)pyrene	0.044U	ug/L	0.16	0.044	1	10/25/10 20:05	10/26/10 14:08	193-39-5	
1-Methylnaphthalene	0.098U	ug/L	1.6	0.098	1	10/25/10 20:05	10/26/10 14:08	90-12-0	
2-Methylnaphthalene	0.066U	ug/L	1.6	0.066	1	10/25/10 20:05	10/26/10 14:08	91-57-6	
Naphthalene	0.087U	ug/L	1.1	0.087	1	10/25/10 20:05	10/26/10 14:08	91-20-3	
Phenanthrene	0.055U	ug/L	1.1	0.055	1	10/25/10 20:05	10/26/10 14:08	85-01-8	
Pyrene	0.066U	ug/L	1.1	0.066	1	10/25/10 20:05	10/26/10 14:08	129-00-0	
2-Fluorobiphenyl (S)	70 %		43.9-113		1	10/25/10 20:05	10/26/10 14:08	321-60-8	
Terphenyl-d14 (S)	90 %		24.8-144		1	10/25/10 20:05	10/26/10 14:08	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 08:08	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 172 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-1R **Lab ID: 3519325028** Collected: 10/18/10 10:35 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 08:08	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 08:08	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 08:08	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:08	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:08	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 08:08	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 08:08	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:08	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 08:08	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 173 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: MW-1R		Lab ID: 3519325028		Collected: 10/18/10 10:35		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 08:08	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 08:08	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 08:08	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 08:08	460-00-4	
Dibromofluoromethane (S)	101 %		88-117		1		10/29/10 08:08	1868-53-7	
1,2-Dichloroethane-d4 (S)	103 %		86-125		1		10/29/10 08:08	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 08:08	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	397	mg/L	5.0	5.0	1		10/23/10 10:20		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		10/25/10 13:30	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	35.1	mg/L	5.0	2.5	1		10/28/10 14:11	16887-00-6	
Sulfate	7.2	mg/L	5.0	2.5	1		10/28/10 14:11	14808-79-8	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	10/22/10 10:00	10/22/10 16:20	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	0.10	mg/L	0.050	0.020	1		10/25/10 09:07	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-10R		Lab ID: 3519325029		Collected: 10/18/10 14:58		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.17	Std. Units			1		10/18/10 14:58		
Field Temperature	28.04	deg C			1		10/18/10 14:58		
Field Specific Conductance	1656	umhos/cm			1		10/18/10 14:58		
Oxygen, Dissolved	0.14	mg/L			1		10/18/10 14:58	7782-44-7	
Turbidity	1.13	NTU			1		10/18/10 14:58		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0049U	ug/L	0.020	0.0049	1	10/27/10 16:30	10/29/10 13:31	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	10/27/10 16:30	10/29/10 13:31	106-93-4	
6010 MET ICP									
		Analytical Method: EPA 6010 Preparation Method: EPA 3010							
Arsenic	11.3	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:46	7440-38-2	
Barium	73.3	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:46	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:46	7440-41-7	
Cadmium	0.64	ug/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:46	7440-43-9	
Calcium	178	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:46	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:46	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:46	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:46	7440-50-8	
Iron	52900	ug/L	40.0	20.0	1	10/22/10 10:00	10/23/10 14:46	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:46	7439-92-1	
Magnesium	48.6	mg/L	0.50	0.25	1	10/22/10 10:00	10/23/10 14:46	7439-95-4	
Nickel	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:46	7440-02-0	
Potassium	1.7	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:46	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	10/22/10 10:00	10/23/10 14:46	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	10/22/10 10:00	10/23/10 14:46	7440-22-4	
Sodium	87.4	mg/L	1.0	0.50	1	10/22/10 10:00	10/23/10 14:46	7440-23-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	10/22/10 10:00	10/23/10 14:46	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	10/22/10 10:00	10/23/10 14:46	7440-66-6	
6020 MET ICPMS									
		Analytical Method: EPA 6020 Preparation Method: EPA 3010							
Antimony	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:22	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	10/22/10 10:00	11/07/10 01:22	7440-28-0	
7470 Mercury									
		Analytical Method: EPA 7470 Preparation Method: EPA 7470							
Mercury	0.10U	ug/L	0.20	0.10	1	10/22/10 09:30	10/25/10 14:20	7439-97-6	
8260 MSV									
		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 08:31	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	74-83-9	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-10R **Lab ID: 3519325029** Collected: 10/18/10 14:58 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 08:31	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 08:31	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:31	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 08:31	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 08:31	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 08:31	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 08:31	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 08:31	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 08:31	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 08:31	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 08:31	460-00-4	
Dibromofluoromethane (S)	101 %		88-117		1		10/29/10 08:31	1868-53-7	
1,2-Dichloroethane-d4 (S)	102 %		86-125		1		10/29/10 08:31	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 08:31	2037-26-5	

2320B Alkalinity

Analytical Method: SM 2320B

Alkalinity, Bicarbonate (CaCO ₃)	700	mg/L	5.0	5.0	1	10/22/10 09:32
Alkalinity, Carbonate (CaCO ₃)	5.0U	mg/L	5.0	5.0	1	10/22/10 09:32

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 176 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: MW-10R		Lab ID: 3519325029		Collected: 10/18/10 14:58		Received: 10/21/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
2320B Alkalinity		Analytical Method: SM 2320B							
Alkalinity, Total as CaCO ₃	700	mg/L	5.0	5.0	1		10/22/10 09:32		
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	904	mg/L	10.0	10.0	1		10/23/10 10:20		
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	70.5	mg/L	25.0	12.5	5		10/28/10 14:23	16887-00-6	
Sulfate	12.5U	mg/L	25.0	12.5	5		10/28/10 14:23	14808-79-8	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	8.4	mg/L	0.050	0.020	1		10/25/10 09:09	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank appdx 1 10-18 **Lab ID:** 3519325030 **Collected:** 10/18/10 14:58 **Received:** 10/21/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	67-64-1	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	107-13-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 04:14	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 04:14	74-87-3	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 04:14	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	74-95-3	
1,2-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	95-50-1	
1,4-Dichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	106-46-7	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	110-57-6	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	78-87-5	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 04:14	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 04:14	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	100-41-4	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	74-88-4	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 04:14	75-09-2	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 04:14	108-10-1	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 04:14	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	127-18-4	
Toluene	7.2	ug/L	1.0	0.50	1		10/29/10 04:14	108-88-3	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 04:14	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 04:14	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 04:14	1330-20-7	
4-Bromofluorobenzene (S)	97 %		70-114		1		10/29/10 04:14	460-00-4	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 178 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank appdx 1 10-18 **Lab ID:** 3519325030 Collected: 10/18/10 14:58 Received: 10/21/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV Analytical Method: EPA 8260									
Dibromofluoromethane (S)	101 %		88-117		1		10/29/10 04:14	1868-53-7	
1,2-Dichloroethane-d4 (S)	101 %		86-125		1		10/29/10 04:14	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 04:14	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank appdx 2 10-18 **Lab ID:** 3519325031 **Collected:** 10/18/10 14:58 **Received:** 10/21/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 04:38	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 04:38	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 04:38	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 04:38	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 04:38	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 04:38	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 04:38	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 04:38	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 04:38	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 180 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank appdx 2 10-18 **Lab ID:** 3519325031 **Collected:** 10/18/10 14:58 **Received:** 10/21/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 04:38	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 04:38	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 04:38	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 04:38	1330-20-7	
4-Bromofluorobenzene (S)	98 %		70-114		1		10/29/10 04:38	460-00-4	
Dibromofluoromethane (S)	102 %		88-117		1		10/29/10 04:38	1868-53-7	
1,2-Dichloroethane-d4 (S)	102 %		86-125		1		10/29/10 04:38	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 04:38	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-1		Lab ID: 3519325032		Collected: 10/27/10 09:20		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	7.25	Std. Units			1		10/27/10 09:20		
Field Temperature	31.08	deg C			1		10/27/10 09:20		
Field Specific Conductance	9922	umhos/cm			1		10/27/10 09:20		
Oxygen, Dissolved	3.65	mg/L			1		10/27/10 09:20	7782-44-7	
Turbidity	13.7	NTU			1		10/27/10 09:20		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.14	ug/L	0.020	0.0049	1	11/04/10 15:15	11/04/10 20:53	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	11/04/10 15:15	11/04/10 20:53	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 02:08	309-00-2	
alpha-BHC	0.00033U	ug/L	0.011	0.00033	1	11/01/10 17:10	11/19/10 02:08	319-84-6	
beta-BHC	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 02:08	319-85-7	
delta-BHC	0.00044U	ug/L	0.011	0.00044	1	11/01/10 17:10	11/19/10 02:08	319-86-8	
gamma-BHC (Lindane)	0.00022U	ug/L	0.011	0.00022	1	11/01/10 17:10	11/19/10 02:08	58-89-9	
Chlordane (Technical)	0.087U	ug/L	0.55	0.087	1	11/01/10 17:10	11/19/10 02:08	57-74-9	
Chlorobenzilate	0.023U	ug/L	0.11	0.023	1	11/01/10 17:10	11/19/10 02:08	510-15-6	
4,4'-DDD	0.0021U	ug/L	0.011	0.0021	1	11/01/10 17:10	11/19/10 02:08	72-54-8	J(M1)
4,4'-DDE	0.00098U	ug/L	0.011	0.00098	1	11/01/10 17:10	11/19/10 02:08	72-55-9	J(M1)
4,4'-DDT	0.0039U	ug/L	0.011	0.0039	1	11/01/10 17:10	11/19/10 02:08	50-29-3	
Dieldrin	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 02:08	60-57-1	J(M1)
Endosulfan I	0.00076U	ug/L	0.011	0.00076	1	11/01/10 17:10	11/19/10 02:08	959-98-8	J(M1)
Endosulfan II	0.00076U	ug/L	0.011	0.00076	1	11/01/10 17:10	11/19/10 02:08	33213-65-9	
Endosulfan sulfate	0.00066U	ug/L	0.011	0.00066	1	11/01/10 17:10	11/19/10 02:08	1031-07-8	J(M1)
Endrin	0.0019U	ug/L	0.011	0.0019	1	11/01/10 17:10	11/19/10 02:08	72-20-8	J(M1)
Endrin aldehyde	0.0078U	ug/L	0.011	0.0078	1	11/01/10 17:10	11/19/10 02:08	7421-93-4	J(M1)
Heptachlor	0.0016U	ug/L	0.011	0.0016	1	11/01/10 17:10	11/19/10 02:08	76-44-8	
Heptachlor epoxide	0.00044U	ug/L	0.011	0.00044	1	11/01/10 17:10	11/19/10 02:08	1024-57-3	
Methoxychlor	0.0076U	ug/L	0.011	0.0076	1	11/01/10 17:10	11/19/10 02:08	72-43-5	
Pentachloronitrobenzene	0.016U	ug/L	0.11	0.016	1	11/01/10 17:10	11/19/10 02:08	82-68-8	
Toxaphene	0.31U	ug/L	0.55	0.31	1	11/01/10 17:10	11/19/10 02:08	8001-35-2	
Tetrachloro-m-xylene (S)	70 %		66.5-120.3		1	11/01/10 17:10	11/19/10 02:08	877-09-8	
Decachlorobiphenyl (S)	17 %		41.7-109.1		1	11/01/10 17:10	11/19/10 02:08	2051-24-3	2p, J(S2)
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.087U	ug/L	0.55	0.087	1	11/01/10 17:10	11/19/10 02:08	12674-11-2	
PCB-1221 (Aroclor 1221)	0.088U	ug/L	0.55	0.088	1	11/01/10 17:10	11/19/10 02:08	11104-28-2	
PCB-1232 (Aroclor 1232)	0.13U	ug/L	0.55	0.13	1	11/01/10 17:10	11/19/10 02:08	11141-16-5	
PCB-1242 (Aroclor 1242)	0.14U	ug/L	0.55	0.14	1	11/01/10 17:10	11/19/10 02:08	53469-21-9	
PCB-1248 (Aroclor 1248)	0.30U	ug/L	0.55	0.30	1	11/01/10 17:10	11/19/10 02:08	12672-29-6	
PCB-1254 (Aroclor 1254)	0.16U	ug/L	0.55	0.16	1	11/01/10 17:10	11/19/10 02:08	11097-69-1	
PCB-1260 (Aroclor 1260)	0.12U	ug/L	0.55	0.12	1	11/01/10 17:10	11/19/10 02:08	11096-82-5	
Tetrachloro-m-xylene (S)	86 %		48-111		1	11/01/10 17:10	11/19/10 02:08	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 182 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-1		Lab ID: 3519325032		Collected: 10/27/10 09:20		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	23 %		63-121		1	11/01/10 17:10	11/19/10 02:08	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.21U	ug/L	0.56	0.21	1	11/03/10 13:00	11/29/10 19:16	60-51-5	
Disulfoton	0.18U	ug/L	0.56	0.18	1	11/03/10 13:00	11/29/10 19:16	298-04-4	
Famphur	0.16U	ug/L	0.56	0.16	1	11/03/10 13:00	11/29/10 19:16	52-85-7	
Methyl parathion	0.22U	ug/L	0.56	0.22	1	11/03/10 13:00	11/29/10 19:16	298-00-0	
Parathion (Ethyl parathion)	0.40U	ug/L	1.1	0.40	1	11/03/10 13:00	11/29/10 19:16	56-38-2	
Phorate	0.41U	ug/L	1.1	0.41	1	11/03/10 13:00	11/29/10 19:16	298-02-2	
4-Chloro3nitrobenzotrifluoride	125 %		34.2-122		1	11/03/10 13:00	11/29/10 19:16		9p,S3
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.25U	ug/L	1.0	0.25	1	10/29/10 18:00	11/02/10 19:44	94-75-7	
Dinoseb	0.063U	ug/L	0.21	0.063	1	10/29/10 18:00	11/02/10 19:44	88-85-7	
Pentachlorophenol	0.019U	ug/L	0.031	0.019	1	10/29/10 18:00	11/02/10 19:44	87-86-5	
2,4,5-T	0.046U	ug/L	0.21	0.046	1	10/29/10 18:00	11/02/10 19:44	93-76-5	
2,4,5-TP (Silvex)	0.054U	ug/L	0.21	0.054	1	10/29/10 18:00	11/02/10 19:44	93-72-1	
2,4-DCPA (S)	100 %		65.5-125.7		1	10/29/10 18:00	11/02/10 19:44	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 10:56	7440-38-2	
Barium	7.4 I	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 10:56	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 10:56	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 10:56	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 10:56	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 10:56	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 10:56	7440-50-8	
Iron	66.9	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 10:56	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 10:56	7439-92-1	
Nickel	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 10:56	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 10:56	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 10:56	7440-22-4	
Sodium	38.8	mg/L	1.0	0.50	1	11/05/10 11:40	11/10/10 10:56	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 10:56	7440-31-5	
Vanadium	13.1	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 10:56	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 10:56	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	42.3	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 09:54	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 09:54	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 11:42	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-1 Lab ID: 3519325032 Collected: 10/27/10 09:20 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Acenaphthene	10.4U	ug/L	60.2	10.4	10	11/01/10 18:00	11/11/10 07:32	83-32-9	M6
Acenaphthylene	11.4U	ug/L	60.2	11.4	10	11/01/10 18:00	11/11/10 07:32	208-96-8	
Acetophenone	17.5U	ug/L	60.2	17.5	10	11/01/10 18:00	11/11/10 07:32	98-86-2	
2-Acetylaminofluorene	7.8U	ug/L	60.2	7.8	10	11/01/10 18:00	11/11/10 07:32	53-96-3	M6
4-Aminobiphenyl	34.1U	ug/L	60.2	34.1	10	11/01/10 18:00	11/11/10 07:32	92-67-1	M6
Anthracene	7.2U	ug/L	60.2	7.2	10	11/01/10 18:00	11/11/10 07:32	120-12-7	
Benzo(a)anthracene	7.6U	ug/L	60.2	7.6	10	11/01/10 18:00	11/11/10 07:32	56-55-3	
Benzo(a)pyrene	7.0U	ug/L	12.0	7.0	10	11/01/10 18:00	11/11/10 07:32	50-32-8	
Benzo(b)fluoranthene	7.5U	ug/L	24.1	7.5	10	11/01/10 18:00	11/11/10 07:32	205-99-2	
Benzo(g,h,i)perylene	8.2U	ug/L	60.2	8.2	10	11/01/10 18:00	11/11/10 07:32	191-24-2	
Benzo(k)fluoranthene	6.1U	ug/L	48.2	6.1	10	11/01/10 18:00	11/11/10 07:32	207-08-9	
Benzyl alcohol	12.3U	ug/L	60.2	12.3	10	11/01/10 18:00	11/11/10 07:32	100-51-6	
4-Bromophenylphenyl ether	8.1U	ug/L	60.2	8.1	10	11/01/10 18:00	11/11/10 07:32	101-55-3	
Butylbenzylphthalate	8.7U	ug/L	60.2	8.7	10	11/01/10 18:00	11/11/10 07:32	85-68-7	
4-Chloro-3-methylphenol	7.5U	ug/L	241	7.5	10	11/01/10 18:00	11/11/10 07:32	59-50-7	
4-Chloroaniline	14.6U	ug/L	60.2	14.6	10	11/01/10 18:00	11/11/10 07:32	106-47-8	
bis(2-Chloroethoxy)methane	35.5U	ug/L	60.2	35.5	10	11/01/10 18:00	11/11/10 07:32	111-91-1	
bis(2-Chloroethyl) ether	9.0U	ug/L	48.2	9.0	10	11/01/10 18:00	11/11/10 07:32	111-44-4	
bis(2-Chloroisopropyl) ether	8.8U	ug/L	60.2	8.8	10	11/01/10 18:00	11/11/10 07:32	108-60-1	
2-Chloronaphthalene	9.6U	ug/L	60.2	9.6	10	11/01/10 18:00	11/11/10 07:32	91-58-7	
2-Chlorophenol	8.2U	ug/L	60.2	8.2	10	11/01/10 18:00	11/11/10 07:32	95-57-8	
4-Chlorophenylphenyl ether	7.6U	ug/L	60.2	7.6	10	11/01/10 18:00	11/11/10 07:32	7005-72-3	M6
Chrysene	4.5U	ug/L	60.2	4.5	10	11/01/10 18:00	11/11/10 07:32	218-01-9	
Diallylate	8.8U	ug/L	60.2	8.8	10	11/01/10 18:00	11/11/10 07:32	2303-16-4	
Dibenz(a,h)anthracene	7.8U	ug/L	24.1	7.8	10	11/01/10 18:00	11/11/10 07:32	53-70-3	
Dibenzofuran	8.1U	ug/L	60.2	8.1	10	11/01/10 18:00	11/11/10 07:32	132-64-9	
1,2-Dichlorobenzene	8.2U	ug/L	60.2	8.2	10	11/01/10 18:00	11/11/10 07:32	95-50-1	
1,3-Dichlorobenzene	9.1U	ug/L	60.2	9.1	10	11/01/10 18:00	11/11/10 07:32	541-73-1	
1,4-Dichlorobenzene	19.7 U	ug/L	60.2	9.3	10	11/01/10 18:00	11/11/10 07:32	106-46-7	
3,3'-Dichlorobenzidine	8.3U	ug/L	120	8.3	10	11/01/10 18:00	11/11/10 07:32	91-94-1	M6
2,4-Dichlorophenol	6.7U	ug/L	24.1	6.7	10	11/01/10 18:00	11/11/10 07:32	120-83-2	
2,6-Dichlorophenol	7.5U	ug/L	48.2	7.5	10	11/01/10 18:00	11/11/10 07:32	87-65-0	
Diethylphthalate	8.5 U	ug/L	60.2	6.1	10	11/01/10 18:00	11/11/10 07:32	84-66-2	M6
P-Dimethylaminoazobenzene	8.1U	ug/L	60.2	8.1	10	11/01/10 18:00	11/11/10 07:32	60-11-7	
7,12-Dimethylbenz(a)anthracene	23.5U	ug/L	60.2	23.5	10	11/01/10 18:00	11/11/10 07:32	57-97-6	
3,3'-Dimethylbenzidine	37.7U	ug/L	120	37.7	10	11/01/10 18:00	11/11/10 07:32	119-93-7	M6
2,4-Dimethylphenol	19.0U	ug/L	60.2	19.0	10	11/01/10 18:00	11/11/10 07:32	105-67-9	
a,a-Dimethylphenylethylamine	120U	ug/L	241	120	10	11/01/10 18:00	11/11/10 07:32	122-09-8	
Dimethylphthalate	7.7U	ug/L	60.2	7.7	10	11/01/10 18:00	11/11/10 07:32	131-11-3	
Di-n-butylphthalate	4.9U	ug/L	60.2	4.9	10	11/01/10 18:00	11/11/10 07:32	84-74-2	
4,6-Dinitro-2-methylphenol	15.9U	ug/L	241	15.9	10	11/01/10 18:00	11/11/10 07:32	534-52-1	M6
1,2-Dinitrobenzene	14.1U	ug/L	60.2	14.1	10	11/01/10 18:00	11/11/10 07:32	528-29-0	
1,3-Dinitrobenzene	8.2U	ug/L	96.3	8.2	10	11/01/10 18:00	11/11/10 07:32	99-65-0	
2,4-Dinitrophenol	18.9U	ug/L	241	18.9	10	11/01/10 18:00	11/11/10 07:32	51-28-5	M6
2,4-Dinitrotoluene	6.4U	ug/L	24.1	6.4	10	11/01/10 18:00	11/11/10 07:32	121-14-2	M6
2,6-Dinitrotoluene	14.7U	ug/L	24.1	14.7	10	11/01/10 18:00	11/11/10 07:32	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 184 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-1		Lab ID: 3519325032	Collected: 10/27/10 09:20	Received: 10/29/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Di-n-octylphthalate	10.8U	ug/L	60.2	10.8	10	11/01/10 18:00	11/11/10 07:32	117-84-0	
bis(2-Ethylhexyl)phthalate	9.6U	ug/L	60.2	9.6	10	11/01/10 18:00	11/11/10 07:32	117-81-7	
Ethyl methanesulfonate	10.8U	ug/L	60.2	10.8	10	11/01/10 18:00	11/11/10 07:32	62-50-0	
Fluoranthene	6.5U	ug/L	60.2	6.5	10	11/01/10 18:00	11/11/10 07:32	206-44-0	
Fluorene	6.7U	ug/L	60.2	6.7	10	11/01/10 18:00	11/11/10 07:32	86-73-7	
Hexachlorobenzene	9.6U	ug/L	12.0	9.6	10	11/01/10 18:00	11/11/10 07:32	118-74-1	
Hexachlorocyclopentadiene	15.4U	ug/L	60.2	15.4	10	11/01/10 18:00	11/11/10 07:32	77-47-4	M6
Hexachloroethane	8.5U	ug/L	60.2	8.5	10	11/01/10 18:00	11/11/10 07:32	67-72-1	
Hexachloropropene	17.0U	ug/L	60.2	17.0	10	11/01/10 18:00	11/11/10 07:32	1888-71-7	
Indeno(1,2,3-cd)pyrene	8.8U	ug/L	24.1	8.8	10	11/01/10 18:00	11/11/10 07:32	193-39-5	
Isodrin	6.5U	ug/L	60.2	6.5	10	11/01/10 18:00	11/11/10 07:32	465-73-6	
Isophorone	8.8U	ug/L	60.2	8.8	10	11/01/10 18:00	11/11/10 07:32	78-59-1	
Isosafrole	7.2U	ug/L	60.2	7.2	10	11/01/10 18:00	11/11/10 07:32	120-58-1	
Kepone	120U	ug/L	241	120	10	11/01/10 18:00	11/11/10 07:32	143-50-0	
Methapyrilene	19.9U	ug/L	60.2	19.9	10	11/01/10 18:00	11/11/10 07:32	91-80-5	
3-Methylcholanthrene	12.5U	ug/L	60.2	12.5	10	11/01/10 18:00	11/11/10 07:32	56-49-5	
Methyl methanesulfonate	12.0U	ug/L	60.2	12.0	10	11/01/10 18:00	11/11/10 07:32	66-27-3	
1-Methylnaphthalene	12.0U	ug/L	60.2	12.0	10	11/01/10 18:00	11/11/10 07:32	90-12-0	
2-Methylnaphthalene	11.9U	ug/L	60.2	11.9	10	11/01/10 18:00	11/11/10 07:32	91-57-6	
2-Methylphenol(o-Cresol)	8.8U	ug/L	60.2	8.8	10	11/01/10 18:00	11/11/10 07:32	95-48-7	M6
3&4-Methylphenol(m&p Cresol)	178	ug/L	120	7.9	10	11/01/10 18:00	11/11/10 07:32		M6
2-Naphthylamine	27.3U	ug/L	60.2	27.3	10	11/01/10 18:00	11/11/10 07:32	91-59-8	M6
Naphthalene	15.7 I	ug/L	60.2	9.4	10	11/01/10 18:00	11/11/10 07:32	91-20-3	
1-Naphthylamine	12.4U	ug/L	60.2	12.4	10	11/01/10 18:00	11/11/10 07:32	134-32-7	M6
1,4-Naphthoquinone	14.2U	ug/L	60.2	14.2	10	11/01/10 18:00	11/11/10 07:32	130-15-4	M6
2-Nitroaniline	7.2U	ug/L	60.2	7.2	10	11/01/10 18:00	11/11/10 07:32	88-74-4	
3-Nitroaniline	11.9U	ug/L	60.2	11.9	10	11/01/10 18:00	11/11/10 07:32	99-09-2	
4-Nitroaniline	8.3U	ug/L	48.2	8.3	10	11/01/10 18:00	11/11/10 07:32	100-01-6	
Nitrobenzene	13.1U	ug/L	48.2	13.1	10	11/01/10 18:00	11/11/10 07:32	98-95-3	
2-Nitrophenol	9.8U	ug/L	60.2	9.8	10	11/01/10 18:00	11/11/10 07:32	88-75-5	
4-Nitrophenol	13.0U	ug/L	241	13.0	10	11/01/10 18:00	11/11/10 07:32	100-02-7	
5-Nitro-o-toluidine	15.5U	ug/L	60.2	15.5	10	11/01/10 18:00	11/11/10 07:32	99-55-8	
N-Nitrosodiethylamine	8.8U	ug/L	48.2	8.8	10	11/01/10 18:00	11/11/10 07:32	55-18-5	
N-Nitrosodimethylamine	11.7U	ug/L	24.1	11.7	10	11/01/10 18:00	11/11/10 07:32	62-75-9	
N-Nitroso-di-n-butylamine	6.6U	ug/L	48.2	6.6	10	11/01/10 18:00	11/11/10 07:32	924-16-3	
N-Nitroso-di-n-propylamine	11.3U	ug/L	48.2	11.3	10	11/01/10 18:00	11/11/10 07:32	621-64-7	
N-Nitrosodiphenylamine	6.0U	ug/L	60.2	6.0	10	11/01/10 18:00	11/11/10 07:32	86-30-6	M6
N-Nitrosomethylethylamine	8.9U	ug/L	60.2	8.9	10	11/01/10 18:00	11/11/10 07:32	10595-95-6	
N-Nitrosopiperidine	7.7U	ug/L	60.2	7.7	10	11/01/10 18:00	11/11/10 07:32	100-75-4	
N-Nitrosopyrrolidine	10.6U	ug/L	60.2	10.6	10	11/01/10 18:00	11/11/10 07:32	930-55-2	
O,O,O-Triethylphosphorothioate	8.3U	ug/L	60.2	8.3	10	11/01/10 18:00	11/11/10 07:32	126-68-1	
Parathion (Ethyl parathion)	13.8U	ug/L	60.2	13.8	10	11/01/10 18:00	11/11/10 07:32	56-38-2	
Pentachlorobenzene	9.4U	ug/L	60.2	9.4	10	11/01/10 18:00	11/11/10 07:32	608-93-5	
Pentachlorophenol	7.9U	ug/L	241	7.9	10	11/01/10 18:00	11/11/10 07:32	87-86-5	M6
Phenacetin	6.4U	ug/L	60.2	6.4	10	11/01/10 18:00	11/11/10 07:32	62-44-2	
Phenanthrene	6.3U	ug/L	60.2	6.3	10	11/01/10 18:00	11/11/10 07:32	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 185 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-1 Lab ID: 3519325032 Collected: 10/27/10 09:20 Received: 10/29/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Phenol	116	ug/L	60.2	6.5	10	11/01/10 18:00	11/11/10 07:32	108-95-2	D3
p-Phenylenediamine	120U	ug/L	241	120	10	11/01/10 18:00	11/11/10 07:32	106-50-3	
Pronamide	13.6U	ug/L	60.2	13.6	10	11/01/10 18:00	11/11/10 07:32	23950-58-5	
Pyrene	8.2U	ug/L	60.2	8.2	10	11/01/10 18:00	11/11/10 07:32	129-00-0	
Safrole	10.2U	ug/L	60.2	10.2	10	11/01/10 18:00	11/11/10 07:32	94-59-7	
1,2,4,5-Tetrachlorobenzene	8.4U	ug/L	60.2	8.4	10	11/01/10 18:00	11/11/10 07:32	95-94-3	
2,3,4,6-Tetrachlorophenol	46.3U	ug/L	60.2	46.3	10	11/01/10 18:00	11/11/10 07:32	58-90-2	
Thionazin	7.3U	ug/L	60.2	7.3	10	11/01/10 18:00	11/11/10 07:32	297-97-2	
O-Toluidine	12.9U	ug/L	60.2	12.9	10	11/01/10 18:00	11/11/10 07:32	95-53-4	
1,2,4-Trichlorobenzene	10U	ug/L	60.2	10	10	11/01/10 18:00	11/11/10 07:32	120-82-1	
2,4,5-Trichlorophenol	6.3U	ug/L	48.2	6.3	10	11/01/10 18:00	11/11/10 07:32	95-95-4	
2,4,6-Trichlorophenol	8.3U	ug/L	24.1	8.3	10	11/01/10 18:00	11/11/10 07:32	88-06-2	M6
1,3,5-Trinitrobenzene	14.7U	ug/L	60.2	14.7	10	11/01/10 18:00	11/11/10 07:32	99-35-4	M6
Nitrobenzene-d5 (S)	77	%	10-110		10	11/01/10 18:00	11/11/10 07:32	4165-60-0	
2-Fluorobiphenyl (S)	96	%	18-110		10	11/01/10 18:00	11/11/10 07:32	321-60-8	
Terphenyl-d14 (S)	97	%	10-123		10	11/01/10 18:00	11/11/10 07:32	1718-51-0	
Phenol-d6 (S)	57	%	10-110		10	11/01/10 18:00	11/11/10 07:32	13127-88-3	
2-Fluorophenol (S)	60	%	18-110		10	11/01/10 18:00	11/11/10 07:32	367-12-4	
2,4,6-Tribromophenol (S)	99	%	10-110		10	11/01/10 18:00	11/11/10 07:32	118-79-6	
8270 MSSV PAH by SCAN Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510									
Acenaphthene	0.037U	ug/L	1.2	0.037	1	10/29/10 18:14	11/05/10 15:17	83-32-9	
Acenaphthylene	0.062U	ug/L	2.5	0.062	1	10/29/10 18:14	11/05/10 15:17	208-96-8	
Anthracene	0.062U	ug/L	1.2	0.062	1	10/29/10 18:14	11/05/10 15:17	120-12-7	
Benzo(a)anthracene	0.074U	ug/L	0.25	0.074	1	10/29/10 18:14	11/05/10 15:17	56-55-3	
Benzo(a)pyrene	0.062U	ug/L	0.25	0.062	1	10/29/10 18:14	11/05/10 15:17	50-32-8	
Benzo(b)fluoranthene	0.062U	ug/L	0.12	0.062	1	10/29/10 18:14	11/05/10 15:17	205-99-2	
Benzo(g,h,i)perylene	0.074U	ug/L	1.2	0.074	1	10/29/10 18:14	11/05/10 15:17	191-24-2	
Benzo(k)fluoranthene	0.049U	ug/L	0.31	0.049	1	10/29/10 18:14	11/05/10 15:17	207-08-9	
Chrysene	0.074U	ug/L	1.2	0.074	1	10/29/10 18:14	11/05/10 15:17	218-01-9	
Dibenz(a,h)anthracene	0.062U	ug/L	0.25	0.062	1	10/29/10 18:14	11/05/10 15:17	53-70-3	
Fluoranthene	0.074U	ug/L	1.2	0.074	1	10/29/10 18:14	11/05/10 15:17	206-44-0	
Fluorene	0.037U	ug/L	1.2	0.037	1	10/29/10 18:14	11/05/10 15:17	86-73-7	
Indeno(1,2,3-cd)pyrene	0.049U	ug/L	0.18	0.049	1	10/29/10 18:14	11/05/10 15:17	193-39-5	
1-Methylnaphthalene	1.2 I	ug/L	1.8	0.11	1	10/29/10 18:14	11/05/10 15:17	90-12-0	
2-Methylnaphthalene	1.4 I	ug/L	1.8	0.074	1	10/29/10 18:14	11/05/10 15:17	91-57-6	
Naphthalene	15.6	ug/L	6.2	0.49	5	10/29/10 18:14	11/05/10 14:56	91-20-3	3p
Phenanthrene	0.062U	ug/L	1.2	0.062	1	10/29/10 18:14	11/05/10 15:17	85-01-8	
Pyrene	0.074U	ug/L	1.2	0.074	1	10/29/10 18:14	11/05/10 15:17	129-00-0	
2-Fluorobiphenyl (S)	64	%	43.9-113		1	10/29/10 18:14	11/05/10 15:17	321-60-8	
Terphenyl-d14 (S)	65	%	24.8-144		1	10/29/10 18:14	11/05/10 15:17	1718-51-0	
8260 MSV Analytical Method: EPA 8260									
Acetone	61.5	ug/L	50.0	25.0	5		10/29/10 21:27	67-64-1	
Acetonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	75-05-8	
Acrolein	50.0U	ug/L	100	50.0	5		10/29/10 21:27	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 186 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-1 **Lab ID: 3519325032** Collected: 10/27/10 09:20 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	107-13-1	
Allyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	107-05-1	
Benzene	4.1 I	ug/L	5.0	2.5	5		10/29/10 21:27	71-43-2	
Bromochloromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	74-97-5	
Bromodichloromethane	1.4U	ug/L	3.0	1.4	5		10/29/10 21:27	75-27-4	
Bromoform	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-25-2	
Bromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	74-83-9	
2-Butanone (MEK)	59.7	ug/L	50.0	25.0	5		10/29/10 21:27	78-93-3	
Carbon disulfide	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-15-0	
Carbon tetrachloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	56-23-5	
Chlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	108-90-7	
Chloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-00-3	
Chloroform	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	67-66-3	
Chloromethane	3.1U	ug/L	5.0	3.1	5		10/29/10 21:27	74-87-3	
Chloroprene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	126-99-8	
Dibromochloromethane	1.3U	ug/L	2.5	1.3	5		10/29/10 21:27	124-48-1	
Dibromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	74-95-3	
trans-1,4-Dichloro-2-butene	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	110-57-6	
Dichlorodifluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-71-8	
1,1-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-34-3	
1,2-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	107-06-2	
1,1-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-35-4	
cis-1,2-Dichloroethene	2.6 I	ug/L	5.0	2.5	5		10/29/10 21:27	156-59-2	
trans-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	156-60-5	
1,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	78-87-5	
1,3-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	142-28-9	
2,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	594-20-7	
1,1-Dichloropropene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	563-58-6	
cis-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 21:27	10061-01-5	
trans-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 21:27	10061-02-6	
Ethylbenzene	18.8	ug/L	5.0	2.5	5		10/29/10 21:27	100-41-4	
Ethyl methacrylate	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	97-63-2	
Hexachloro-1,3-butadiene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	87-68-3	
2-Hexanone	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	591-78-6	
Iodomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	74-88-4	
Isobutyl Alcohol	50.0U	ug/L	100	50.0	5		10/29/10 21:27	78-83-1	
Methacrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	126-98-7	
Methylene Chloride	12.5U	ug/L	25.0	12.5	5		10/29/10 21:27	75-09-2	
Methyl methacrylate	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	80-62-6	
4-Methyl-2-pentanone (MIBK)	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	108-10-1	
Propionitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:27	107-12-0	
Styrene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	100-42-5	
1,1,1,2-Tetrachloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	630-20-6	
1,1,2,2-Tetrachloroethane	0.90U	ug/L	2.5	0.90	5		10/29/10 21:27	79-34-5	
Tetrachloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	127-18-4	
Toluene	3.1 I	ug/L	5.0	2.5	5		10/29/10 21:27	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 187 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-1		Lab ID: 3519325032		Collected: 10/27/10 09:20		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	120-82-1	
1,1,1-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	71-55-6	
1,1,2-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	79-00-5	
Trichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	79-01-6	
Trichlorofluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-69-4	
1,2,3-Trichloropropane	1.8U	ug/L	2.5	1.8	5		10/29/10 21:27	96-18-4	
Vinyl acetate	5.0U	ug/L	10.0	5.0	5		10/29/10 21:27	108-05-4	
Vinyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:27	75-01-4	
Xylene (Total)	26.0	ug/L	5.0	2.5	5		10/29/10 21:27	1330-20-7	
4-Bromofluorobenzene (S)	97	%	70-114		5		10/29/10 21:27	460-00-4	D3
Dibromofluoromethane (S)	101	%	88-117		5		10/29/10 21:27	1868-53-7	
1,2-Dichloroethane-d4 (S)	109	%	86-125		5		10/29/10 21:27	17060-07-0	
Toluene-d8 (S)	102	%	87-113		5		10/29/10 21:27	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	3220	mg/L	100	100	1		11/03/10 11:58		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	6.8	mg/L	1.0	1.0	1		11/01/10 11:50	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	792	mg/L	500	250	100		11/02/10 00:18	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.025U	mg/L	0.050	0.025	5	11/01/10 11:30	11/08/10 15:01	57-12-5	4p, J(M1)
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	599	mg/L	5.0	2.0	100		11/01/10 16:01	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-2		Lab ID: 3519325033		Collected: 10/27/10 09:50		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
Analytical Method:									
Field pH	7.63	Std. Units			1		10/27/10 09:50		
Field Temperature	37.76	deg C			1		10/27/10 09:50		
Field Specific Conductance	19413	umhos/cm			1		10/27/10 09:50		
Oxygen, Dissolved	2.23	mg/L			1		10/27/10 09:50	7782-44-7	
Turbidity	26.5	NTU			1		10/27/10 09:50		
8011 GCS EDB and DBCP									
Analytical Method: EPA 8011		Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.22	ug/L	0.020	0.0049	1	11/04/10 15:15	11/04/10 21:08	96-12-8	
1,2-Dibromoethane (EDB)	0.0062U	ug/L	0.010	0.0062	1	11/04/10 15:15	11/04/10 21:08	106-93-4	
8081 GCS Pesticides									
Analytical Method: EPA 8081		Preparation Method: EPA 3510							
Aldrin	0.00057U	ug/L	0.011	0.00057	1	11/01/10 17:10	11/19/10 02:59	309-00-2	
alpha-BHC	0.00034U	ug/L	0.011	0.00034	1	11/01/10 17:10	11/19/10 02:59	319-84-6	
beta-BHC	0.00057U	ug/L	0.011	0.00057	1	11/01/10 17:10	11/19/10 02:59	319-85-7	
delta-BHC	0.00046U	ug/L	0.011	0.00046	1	11/01/10 17:10	11/19/10 02:59	319-86-8	
gamma-BHC (Lindane)	0.00023U	ug/L	0.011	0.00023	1	11/01/10 17:10	11/19/10 02:59	58-89-9	
Chlordane (Technical)	0.091U	ug/L	0.57	0.091	1	11/01/10 17:10	11/19/10 02:59	57-74-9	
Chlorobenzilate	0.024U	ug/L	0.11	0.024	1	11/01/10 17:10	11/19/10 02:59	510-15-6	
4,4'-DDD	0.0022U	ug/L	0.011	0.0022	1	11/01/10 17:10	11/19/10 02:59	72-54-8	
4,4'-DDE	0.0010U	ug/L	0.011	0.0010	1	11/01/10 17:10	11/19/10 02:59	72-55-9	
4,4'-DDT	0.0041U	ug/L	0.011	0.0041	1	11/01/10 17:10	11/19/10 02:59	50-29-3	
Dieldrin	0.00057U	ug/L	0.011	0.00057	1	11/01/10 17:10	11/19/10 02:59	60-57-1	
Endosulfan I	0.00080U	ug/L	0.011	0.00080	1	11/01/10 17:10	11/19/10 02:59	959-98-8	
Endosulfan II	0.00080U	ug/L	0.011	0.00080	1	11/01/10 17:10	11/19/10 02:59	33213-65-9	
Endosulfan sulfate	0.00069U	ug/L	0.011	0.00069	1	11/01/10 17:10	11/19/10 02:59	1031-07-8	
Endrin	0.0019U	ug/L	0.011	0.0019	1	11/01/10 17:10	11/19/10 02:59	72-20-8	
Endrin aldehyde	0.0081U	ug/L	0.011	0.0081	1	11/01/10 17:10	11/19/10 02:59	7421-93-4	
Heptachlor	0.0017U	ug/L	0.011	0.0017	1	11/01/10 17:10	11/19/10 02:59	76-44-8	
Heptachlor epoxide	0.00046U	ug/L	0.011	0.00046	1	11/01/10 17:10	11/19/10 02:59	1024-57-3	
Methoxychlor	0.0080U	ug/L	0.011	0.0080	1	11/01/10 17:10	11/19/10 02:59	72-43-5	
Pentachloronitrobenzene	0.017U	ug/L	0.11	0.017	1	11/01/10 17:10	11/19/10 02:59	82-68-8	
Toxaphene	0.33U	ug/L	0.57	0.33	1	11/01/10 17:10	11/19/10 02:59	8001-35-2	
Tetrachloro-m-xylene (S)	75 %		66.5-120.3		1	11/01/10 17:10	11/19/10 02:59	877-09-8	
Decachlorobiphenyl (S)	4 %		41.7-109.1		1	11/01/10 17:10	11/19/10 02:59	2051-24-3	2p, J(S5)
8082 GCS PCB									
Analytical Method: EPA 8082		Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.091U	ug/L	0.57	0.091	1	11/01/10 17:10	11/19/10 02:59	12674-11-2	
PCB-1221 (Aroclor 1221)	0.093U	ug/L	0.57	0.093	1	11/01/10 17:10	11/19/10 02:59	11104-28-2	
PCB-1232 (Aroclor 1232)	0.13U	ug/L	0.57	0.13	1	11/01/10 17:10	11/19/10 02:59	11141-16-5	
PCB-1242 (Aroclor 1242)	0.14U	ug/L	0.57	0.14	1	11/01/10 17:10	11/19/10 02:59	53469-21-9	
PCB-1248 (Aroclor 1248)	0.31U	ug/L	0.57	0.31	1	11/01/10 17:10	11/19/10 02:59	12672-29-6	
PCB-1254 (Aroclor 1254)	0.17U	ug/L	0.57	0.17	1	11/01/10 17:10	11/19/10 02:59	11097-69-1	
PCB-1260 (Aroclor 1260)	0.13U	ug/L	0.57	0.13	1	11/01/10 17:10	11/19/10 02:59	11096-82-5	
Tetrachloro-m-xylene (S)	60 %		48-111		1	11/01/10 17:10	11/19/10 02:59	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 189 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-2		Lab ID: 3519325033		Collected: 10/27/10 09:50		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB									
Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	5 %		63-121		1	11/01/10 17:10	11/19/10 02:59	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides									
Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 19:57	60-51-5	
Disulfoton	1.6U	ug/L	5.0	1.6	1	11/03/10 13:00	11/29/10 19:57	298-04-4	
Famphur	1.4U	ug/L	5.0	1.4	1	11/03/10 13:00	11/29/10 19:57	52-85-7	
Methyl parathion	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 19:57	298-00-0	
Parathion (Ethyl parathion)	3.5U	ug/L	10.0	3.5	1	11/03/10 13:00	11/29/10 19:57	56-38-2	
Phorate	3.7U	ug/L	10.0	3.7	1	11/03/10 13:00	11/29/10 19:57	298-02-2	
4-Chloro3nitrobenzotrifluoride	91 %		34.2-122		1	11/03/10 13:00	11/29/10 19:57		9p
8151 Chlorinated Herbicides									
Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.25U	ug/L	1.0	0.25	1	11/02/10 18:00	11/05/10 06:09	94-75-7	
Dinoseb	0.063U	ug/L	0.21	0.063	1	11/02/10 18:00	11/05/10 06:09	88-85-7	
Pentachlorophenol	0.046	ug/L	0.032	0.019	1	11/02/10 18:00	11/05/10 06:09	87-86-5	
2,4,5-T	0.047U	ug/L	0.21	0.047	1	11/02/10 18:00	11/05/10 06:09	93-76-5	
2,4,5-TP (Silvex)	0.054U	ug/L	0.21	0.054	1	11/02/10 18:00	11/05/10 06:09	93-72-1	
2,4-DCPA (S)	33 %		65.5-125.7		1	11/02/10 18:00	11/05/10 06:09	19719-28-9	J(S5)
6010 MET ICP									
Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	253	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:00	7440-38-2	
Barium	122	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:00	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:00	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:00	7440-43-9	
Chromium	41.0	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:00	7440-47-3	
Cobalt	51.8	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:00	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:00	7440-50-8	
Iron	3140	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:00	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:00	7439-92-1	
Nickel	51.6	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:00	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:00	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:00	7440-22-4	
Sodium	1790	mg/L	5.0	2.5	5	11/05/10 11:40	11/11/10 19:03	7440-23-5	D4
Tin	25.0U	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:00	7440-31-5	
Vanadium	32.5	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:00	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:00	7440-66-6	
6020 MET ICPMS									
Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	28.8	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:36	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:36	7440-28-0	
7470 Mercury									
Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 11:50	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-2 **Lab ID: 3519325033** Collected: 10/27/10 09:50 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	8.6U	ug/L	50.0	8.6	1	11/01/10 18:00	11/11/10 05:28	83-32-9	
Acenaphthylene	9.5U	ug/L	50.0	9.5	1	11/01/10 18:00	11/11/10 05:28	208-96-8	
Acetophenone	14.5U	ug/L	50.0	14.5	1	11/01/10 18:00	11/11/10 05:28	98-86-2	
2-Acetylaminofluorene	6.5U	ug/L	50.0	6.5	1	11/01/10 18:00	11/11/10 05:28	53-96-3	
4-Aminobiphenyl	28.3U	ug/L	50.0	28.3	1	11/01/10 18:00	11/11/10 05:28	92-67-1	
Anthracene	6.0U	ug/L	50.0	6.0	1	11/01/10 18:00	11/11/10 05:28	120-12-7	
Benzo(a)anthracene	6.3U	ug/L	50.0	6.3	1	11/01/10 18:00	11/11/10 05:28	56-55-3	
Benzo(a)pyrene	5.8U	ug/L	10.0	5.8	1	11/01/10 18:00	11/11/10 05:28	50-32-8	
Benzo(b)fluoranthene	6.2U	ug/L	20.0	6.2	1	11/01/10 18:00	11/11/10 05:28	205-99-2	
Benzo(g,h,i)perylene	6.8U	ug/L	50.0	6.8	1	11/01/10 18:00	11/11/10 05:28	191-24-2	
Benzo(k)fluoranthene	5.1U	ug/L	40.0	5.1	1	11/01/10 18:00	11/11/10 05:28	207-08-9	
Benzyl alcohol	10.2U	ug/L	50.0	10.2	1	11/01/10 18:00	11/11/10 05:28	100-51-6	
4-Bromophenylphenyl ether	6.7U	ug/L	50.0	6.7	1	11/01/10 18:00	11/11/10 05:28	101-55-3	
Butylbenzylphthalate	7.2U	ug/L	50.0	7.2	1	11/01/10 18:00	11/11/10 05:28	85-68-7	
4-Chloro-3-methylphenol	6.2U	ug/L	200	6.2	1	11/01/10 18:00	11/11/10 05:28	59-50-7	
4-Chloroaniline	12.1U	ug/L	50.0	12.1	1	11/01/10 18:00	11/11/10 05:28	106-47-8	
bis(2-Chloroethoxy)methane	29.5U	ug/L	50.0	29.5	1	11/01/10 18:00	11/11/10 05:28	111-91-1	
bis(2-Chloroethyl) ether	7.5U	ug/L	40.0	7.5	1	11/01/10 18:00	11/11/10 05:28	111-44-4	
bis(2-Chloroisopropyl) ether	7.3U	ug/L	50.0	7.3	1	11/01/10 18:00	11/11/10 05:28	108-60-1	
2-Chloronaphthalene	8.0U	ug/L	50.0	8.0	1	11/01/10 18:00	11/11/10 05:28	91-58-7	
2-Chlorophenol	6.8U	ug/L	50.0	6.8	1	11/01/10 18:00	11/11/10 05:28	95-57-8	
4-Chlorophenylphenyl ether	6.3U	ug/L	50.0	6.3	1	11/01/10 18:00	11/11/10 05:28	7005-72-3	
Chrysene	3.7U	ug/L	50.0	3.7	1	11/01/10 18:00	11/11/10 05:28	218-01-9	
Diallylate	7.3U	ug/L	50.0	7.3	1	11/01/10 18:00	11/11/10 05:28	2303-16-4	
Dibenz(a,h)anthracene	6.5U	ug/L	20.0	6.5	1	11/01/10 18:00	11/11/10 05:28	53-70-3	
Dibenzofuran	6.7U	ug/L	50.0	6.7	1	11/01/10 18:00	11/11/10 05:28	132-64-9	
1,2-Dichlorobenzene	6.8U	ug/L	50.0	6.8	1	11/01/10 18:00	11/11/10 05:28	95-50-1	
1,3-Dichlorobenzene	7.6U	ug/L	50.0	7.6	1	11/01/10 18:00	11/11/10 05:28	541-73-1	
1,4-Dichlorobenzene	12.5 U	ug/L	50.0	7.7	1	11/01/10 18:00	11/11/10 05:28	106-46-7	
3,3'-Dichlorobenzidine	6.9U	ug/L	100	6.9	1	11/01/10 18:00	11/11/10 05:28	91-94-1	
2,4-Dichlorophenol	5.6U	ug/L	20.0	5.6	1	11/01/10 18:00	11/11/10 05:28	120-83-2	
2,6-Dichlorophenol	6.2U	ug/L	40.0	6.2	1	11/01/10 18:00	11/11/10 05:28	87-65-0	
Diethylphthalate	5.1U	ug/L	50.0	5.1	1	11/01/10 18:00	11/11/10 05:28	84-66-2	
P-Dimethylaminoazobenzene	6.7U	ug/L	50.0	6.7	1	11/01/10 18:00	11/11/10 05:28	60-11-7	
7,12-Dimethylbenz(a)anthracene	19.5U	ug/L	50.0	19.5	1	11/01/10 18:00	11/11/10 05:28	57-97-6	
3,3'-Dimethylbenzidine	31.3U	ug/L	100	31.3	1	11/01/10 18:00	11/11/10 05:28	119-93-7	
2,4-Dimethylphenol	15.8U	ug/L	50.0	15.8	1	11/01/10 18:00	11/11/10 05:28	105-67-9	
a,a-Dimethylphenylethylamine	100U	ug/L	200	100	1	11/01/10 18:00	11/11/10 05:28	122-09-8	
Dimethylphthalate	6.4U	ug/L	50.0	6.4	1	11/01/10 18:00	11/11/10 05:28	131-11-3	
Di-n-butylphthalate	4.1U	ug/L	50.0	4.1	1	11/01/10 18:00	11/11/10 05:28	84-74-2	
4,6-Dinitro-2-methylphenol	13.2U	ug/L	200	13.2	1	11/01/10 18:00	11/11/10 05:28	534-52-1	
1,2-Dinitrobenzene	11.7U	ug/L	50.0	11.7	1	11/01/10 18:00	11/11/10 05:28	528-29-0	
1,3-Dinitrobenzene	6.8U	ug/L	80.0	6.8	1	11/01/10 18:00	11/11/10 05:28	99-65-0	
2,4-Dinitrophenol	15.7U	ug/L	200	15.7	1	11/01/10 18:00	11/11/10 05:28	51-28-5	
2,4-Dinitrotoluene	5.3U	ug/L	20.0	5.3	1	11/01/10 18:00	11/11/10 05:28	121-14-2	
2,6-Dinitrotoluene	12.2U	ug/L	20.0	12.2	1	11/01/10 18:00	11/11/10 05:28	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 191 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-2 **Lab ID: 3519325033** Collected: 10/27/10 09:50 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	9.0U	ug/L	50.0	9.0	1	11/01/10 18:00	11/11/10 05:28	117-84-0	
bis(2-Ethylhexyl)phthalate	13.9 I	ug/L	50.0	8.0	1	11/01/10 18:00	11/11/10 05:28	117-81-7	
Ethyl methanesulfonate	9.0U	ug/L	50.0	9.0	1	11/01/10 18:00	11/11/10 05:28	62-50-0	
Fluoranthene	5.4U	ug/L	50.0	5.4	1	11/01/10 18:00	11/11/10 05:28	206-44-0	
Fluorene	5.6U	ug/L	50.0	5.6	1	11/01/10 18:00	11/11/10 05:28	86-73-7	
Hexachlorobenzene	8.0U	ug/L	10.0	8.0	1	11/01/10 18:00	11/11/10 05:28	118-74-1	
Hexachlorocyclopentadiene	12.8U	ug/L	50.0	12.8	1	11/01/10 18:00	11/11/10 05:28	77-47-4	
Hexachloroethane	7.1U	ug/L	50.0	7.1	1	11/01/10 18:00	11/11/10 05:28	67-72-1	
Hexachloropropene	14.1U	ug/L	50.0	14.1	1	11/01/10 18:00	11/11/10 05:28	1888-71-7	
Indeno(1,2,3-cd)pyrene	7.3U	ug/L	20.0	7.3	1	11/01/10 18:00	11/11/10 05:28	193-39-5	
Isodrin	5.4U	ug/L	50.0	5.4	1	11/01/10 18:00	11/11/10 05:28	465-73-6	
Isophorone	7.3U	ug/L	50.0	7.3	1	11/01/10 18:00	11/11/10 05:28	78-59-1	
Isosafrole	23.1 I	ug/L	50.0	6.0	1	11/01/10 18:00	11/11/10 05:28	120-58-1	
Kepone	100U	ug/L	200	100	1	11/01/10 18:00	11/11/10 05:28	143-50-0	
Methapyrilene	16.5U	ug/L	50.0	16.5	1	11/01/10 18:00	11/11/10 05:28	91-80-5	
3-Methylcholanthrene	10.4U	ug/L	50.0	10.4	1	11/01/10 18:00	11/11/10 05:28	56-49-5	
Methyl methanesulfonate	10.0U	ug/L	50.0	10.0	1	11/01/10 18:00	11/11/10 05:28	66-27-3	
1-Methylnaphthalene	10.0U	ug/L	50.0	10.0	1	11/01/10 18:00	11/11/10 05:28	90-12-0	
2-Methylnaphthalene	9.9U	ug/L	50.0	9.9	1	11/01/10 18:00	11/11/10 05:28	91-57-6	
2-Methylphenol(o-Cresol)	20.0 I	ug/L	50.0	7.3	1	11/01/10 18:00	11/11/10 05:28	95-48-7	
3&4-Methylphenol(m&p Cresol)	35.0 I	ug/L	100	6.6	1	11/01/10 18:00	11/11/10 05:28		
2-Naphthylamine	22.7U	ug/L	50.0	22.7	1	11/01/10 18:00	11/11/10 05:28	91-59-8	
Naphthalene	18.5 I	ug/L	50.0	7.8	1	11/01/10 18:00	11/11/10 05:28	91-20-3	
1-Naphthylamine	10.3U	ug/L	50.0	10.3	1	11/01/10 18:00	11/11/10 05:28	134-32-7	
1,4-Naphthoquinone	11.8U	ug/L	50.0	11.8	1	11/01/10 18:00	11/11/10 05:28	130-15-4	
2-Nitroaniline	6.0U	ug/L	50.0	6.0	1	11/01/10 18:00	11/11/10 05:28	88-74-4	
3-Nitroaniline	9.9U	ug/L	50.0	9.9	1	11/01/10 18:00	11/11/10 05:28	99-09-2	
4-Nitroaniline	6.9U	ug/L	40.0	6.9	1	11/01/10 18:00	11/11/10 05:28	100-01-6	
Nitrobenzene	10.9U	ug/L	40.0	10.9	1	11/01/10 18:00	11/11/10 05:28	98-95-3	
2-Nitrophenol	8.1U	ug/L	50.0	8.1	1	11/01/10 18:00	11/11/10 05:28	88-75-5	
4-Nitrophenol	10.8U	ug/L	200	10.8	1	11/01/10 18:00	11/11/10 05:28	100-02-7	
5-Nitro-o-toluidine	12.9U	ug/L	50.0	12.9	1	11/01/10 18:00	11/11/10 05:28	99-55-8	
N-Nitrosodiethylamine	7.3U	ug/L	40.0	7.3	1	11/01/10 18:00	11/11/10 05:28	55-18-5	
N-Nitrosodimethylamine	9.7U	ug/L	20.0	9.7	1	11/01/10 18:00	11/11/10 05:28	62-75-9	
N-Nitroso-di-n-butylamine	5.5U	ug/L	40.0	5.5	1	11/01/10 18:00	11/11/10 05:28	924-16-3	
N-Nitroso-di-n-propylamine	9.4U	ug/L	40.0	9.4	1	11/01/10 18:00	11/11/10 05:28	621-64-7	
N-Nitrosodiphenylamine	5.0U	ug/L	50.0	5.0	1	11/01/10 18:00	11/11/10 05:28	86-30-6	
N-Nitrosomethylethylamine	7.4U	ug/L	50.0	7.4	1	11/01/10 18:00	11/11/10 05:28	10595-95-6	
N-Nitrosopiperidine	6.4U	ug/L	50.0	6.4	1	11/01/10 18:00	11/11/10 05:28	100-75-4	
N-Nitrosopyrrolidine	8.8U	ug/L	50.0	8.8	1	11/01/10 18:00	11/11/10 05:28	930-55-2	
O,O,O-Triethylphosphorothioate	6.9U	ug/L	50.0	6.9	1	11/01/10 18:00	11/11/10 05:28	126-68-1	
Parathion (Ethyl parathion)	11.5U	ug/L	50.0	11.5	1	11/01/10 18:00	11/11/10 05:28	56-38-2	
Pentachlorobenzene	7.8U	ug/L	50.0	7.8	1	11/01/10 18:00	11/11/10 05:28	608-93-5	
Pentachlorophenol	6.6U	ug/L	200	6.6	1	11/01/10 18:00	11/11/10 05:28	87-86-5	
Phenacetin	5.3U	ug/L	50.0	5.3	1	11/01/10 18:00	11/11/10 05:28	62-44-2	
Phenanthrene	5.2U	ug/L	50.0	5.2	1	11/01/10 18:00	11/11/10 05:28	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 192 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-2		Lab ID: 3519325033		Collected: 10/27/10 09:50		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenol	5.4U	ug/L	50.0	5.4	1	11/01/10 18:00	11/11/10 05:28	108-95-2	
p-Phenylenediamine	100U	ug/L	200	100	1	11/01/10 18:00	11/11/10 05:28	106-50-3	
Pronamide	11.3U	ug/L	50.0	11.3	1	11/01/10 18:00	11/11/10 05:28	23950-58-5	
Pyrene	6.8U	ug/L	50.0	6.8	1	11/01/10 18:00	11/11/10 05:28	129-00-0	
Safrole	8.5U	ug/L	50.0	8.5	1	11/01/10 18:00	11/11/10 05:28	94-59-7	
1,2,4,5-Tetrachlorobenzene	7.0U	ug/L	50.0	7.0	1	11/01/10 18:00	11/11/10 05:28	95-94-3	
2,3,4,6-Tetrachlorophenol	38.5U	ug/L	50.0	38.5	1	11/01/10 18:00	11/11/10 05:28	58-90-2	
Thionazin	6.1U	ug/L	50.0	6.1	1	11/01/10 18:00	11/11/10 05:28	297-97-2	
O-Toluidine	10.7U	ug/L	50.0	10.7	1	11/01/10 18:00	11/11/10 05:28	95-53-4	
1,2,4-Trichlorobenzene	8.3U	ug/L	50.0	8.3	1	11/01/10 18:00	11/11/10 05:28	120-82-1	
2,4,5-Trichlorophenol	5.2U	ug/L	40.0	5.2	1	11/01/10 18:00	11/11/10 05:28	95-95-4	
2,4,6-Trichlorophenol	6.9U	ug/L	20.0	6.9	1	11/01/10 18:00	11/11/10 05:28	88-06-2	
1,3,5-Trinitrobenzene	12.2U	ug/L	50.0	12.2	1	11/01/10 18:00	11/11/10 05:28	99-35-4	
Nitrobenzene-d5 (S)	53	%	10-110		1	11/01/10 18:00	11/11/10 05:28	4165-60-0	
2-Fluorobiphenyl (S)	67	%	18-110		1	11/01/10 18:00	11/11/10 05:28	321-60-8	
Terphenyl-d14 (S)	82	%	10-123		1	11/01/10 18:00	11/11/10 05:28	1718-51-0	
Phenol-d6 (S)	30	%	10-110		1	11/01/10 18:00	11/11/10 05:28	13127-88-3	
2-Fluorophenol (S)	38	%	18-110		1	11/01/10 18:00	11/11/10 05:28	367-12-4	
2,4,6-Tribromophenol (S)	81	%	10-110		1	11/01/10 18:00	11/11/10 05:28	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 13:57	83-32-9	
Acenaphthylene	0.50U	ug/L	20.0	0.50	1	10/29/10 18:14	11/01/10 13:57	208-96-8	
Anthracene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 13:57	120-12-7	
Benzo(a)anthracene	0.60U	ug/L	2.0	0.60	1	10/29/10 18:14	11/01/10 13:57	56-55-3	
Benzo(a)pyrene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 13:57	50-32-8	
Benzo(b)fluoranthene	0.50U	ug/L	1.0	0.50	1	10/29/10 18:14	11/01/10 13:57	205-99-2	
Benzo(g,h,i)perylene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 13:57	191-24-2	
Benzo(k)fluoranthene	0.40U	ug/L	2.5	0.40	1	10/29/10 18:14	11/01/10 13:57	207-08-9	
Chrysene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 13:57	218-01-9	
Dibenz(a,h)anthracene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 13:57	53-70-3	
Fluoranthene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 13:57	206-44-0	
Fluorene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 13:57	86-73-7	
Indeno(1,2,3-cd)pyrene	0.40U	ug/L	1.5	0.40	1	10/29/10 18:14	11/01/10 13:57	193-39-5	
1-Methylnaphthalene	0.90U	ug/L	15.0	0.90	1	10/29/10 18:14	11/01/10 13:57	90-12-0	
2-Methylnaphthalene	1.8 I	ug/L	15.0	0.60	1	10/29/10 18:14	11/01/10 13:57	91-57-6	
Naphthalene	26.6	ug/L	10.0	0.80	1	10/29/10 18:14	11/01/10 13:57	91-20-3	
Phenanthrene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 13:57	85-01-8	
Pyrene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 13:57	129-00-0	
2-Fluorobiphenyl (S)	72	%	43.9-113		1	10/29/10 18:14	11/01/10 13:57	321-60-8	
Terphenyl-d14 (S)	63	%	24.8-144		1	10/29/10 18:14	11/01/10 13:57	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	38.6 I	ug/L	50.0	25.0	5		10/29/10 22:37	67-64-1	
Acetonitrile	222	ug/L	50.0	25.0	5		10/29/10 22:37	75-05-8	
Acrolein	50.0U	ug/L	100	50.0	5		10/29/10 22:37	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 193 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-2 **Lab ID: 3519325033** Collected: 10/27/10 09:50 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	107-13-1	
Allyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	107-05-1	
Benzene	3.3 I	ug/L	5.0	2.5	5		10/29/10 22:37	71-43-2	
Bromochloromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	74-97-5	
Bromodichloromethane	1.4U	ug/L	3.0	1.4	5		10/29/10 22:37	75-27-4	
Bromoform	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-25-2	
Bromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	74-83-9	
2-Butanone (MEK)	53.5	ug/L	50.0	25.0	5		10/29/10 22:37	78-93-3	
Carbon disulfide	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-15-0	
Carbon tetrachloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	56-23-5	
Chlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	108-90-7	
Chloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-00-3	
Chloroform	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	67-66-3	
Chloromethane	3.1U	ug/L	5.0	3.1	5		10/29/10 22:37	74-87-3	
Chloroprene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	126-99-8	
Dibromochloromethane	1.3U	ug/L	2.5	1.3	5		10/29/10 22:37	124-48-1	
Dibromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	74-95-3	
trans-1,4-Dichloro-2-butene	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	110-57-6	
Dichlorodifluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-71-8	
1,1-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-34-3	
1,2-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	107-06-2	
1,1-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-35-4	
cis-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	156-59-2	
trans-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	156-60-5	
1,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	78-87-5	
1,3-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	142-28-9	
2,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	594-20-7	
1,1-Dichloropropene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	563-58-6	
cis-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 22:37	10061-01-5	
trans-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 22:37	10061-02-6	
Ethylbenzene	47.5	ug/L	5.0	2.5	5		10/29/10 22:37	100-41-4	
Ethyl methacrylate	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	97-63-2	
Hexachloro-1,3-butadiene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	87-68-3	
2-Hexanone	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	591-78-6	
Iodomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	74-88-4	
Isobutyl Alcohol	50.0U	ug/L	100	50.0	5		10/29/10 22:37	78-83-1	
Methacrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	126-98-7	
Methylene Chloride	12.5U	ug/L	25.0	12.5	5		10/29/10 22:37	75-09-2	
Methyl methacrylate	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	80-62-6	
4-Methyl-2-pentanone (MIBK)	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	108-10-1	
Propionitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:37	107-12-0	
Styrene	4.8 I	ug/L	5.0	2.5	5		10/29/10 22:37	100-42-5	
1,1,1,2-Tetrachloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	630-20-6	
1,1,2,2-Tetrachloroethane	0.90U	ug/L	2.5	0.90	5		10/29/10 22:37	79-34-5	
Tetrachloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	127-18-4	
Toluene	62.4	ug/L	5.0	2.5	5		10/29/10 22:37	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 194 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-2		Lab ID: 3519325033	Collected: 10/27/10 09:50	Received: 10/29/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	120-82-1	
1,1,1-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	71-55-6	
1,1,2-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	79-00-5	
Trichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	79-01-6	
Trichlorofluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-69-4	
1,2,3-Trichloropropane	1.8U	ug/L	2.5	1.8	5		10/29/10 22:37	96-18-4	
Vinyl acetate	5.0U	ug/L	10.0	5.0	5		10/29/10 22:37	108-05-4	
Vinyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:37	75-01-4	
Xylene (Total)	113	ug/L	5.0	2.5	5		10/29/10 22:37	1330-20-7	
4-Bromofluorobenzene (S)	96	%	70-114		5		10/29/10 22:37	460-00-4	D3
Dibromofluoromethane (S)	98	%	88-117		5		10/29/10 22:37	1868-53-7	
1,2-Dichloroethane-d4 (S)	109	%	86-125		5		10/29/10 22:37	17060-07-0	
Toluene-d8 (S)	101	%	87-113		5		10/29/10 22:37	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	7380	mg/L	100	100	1		11/03/10 11:59		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	12.4	mg/L	5.0	5.0	1		11/01/10 11:50	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	1850	mg/L	500	250	100		11/02/10 00:30	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.042 I	mg/L	0.050	0.025	1	11/01/10 11:30	11/08/10 15:04	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	1150	mg/L	5.0	2.0	100		11/12/10 15:38	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-3		Lab ID: 3519325034		Collected: 10/27/10 10:25		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
Analytical Method:									
Field pH	7.79	Std. Units			1		10/27/10 10:25		
Field Temperature	40.08	deg C			1		10/27/10 10:25		
Field Specific Conductance	25433	umhos/cm			1		10/27/10 10:25		
Oxygen, Dissolved	1.78	mg/L			1		10/27/10 10:25	7782-44-7	
Turbidity	17.6	NTU			1		10/27/10 10:25		
8011 GCS EDB and DBCP									
Analytical Method: EPA 8011		Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.16	ug/L	0.020	0.0048	1	11/04/10 15:15	11/04/10 21:37	96-12-8	
1,2-Dibromoethane (EDB)	0.0061U	ug/L	0.0098	0.0061	1	11/04/10 15:15	11/04/10 21:37	106-93-4	
8081 GCS Pesticides									
Analytical Method: EPA 8081		Preparation Method: EPA 3510							
Aldrin	0.00054U	ug/L	0.011	0.00054	1	11/01/10 17:10	11/19/10 03:16	309-00-2	
alpha-BHC	0.00032U	ug/L	0.011	0.00032	1	11/01/10 17:10	11/19/10 03:16	319-84-6	
beta-BHC	0.00054U	ug/L	0.011	0.00054	1	11/01/10 17:10	11/19/10 03:16	319-85-7	
delta-BHC	0.00043U	ug/L	0.011	0.00043	1	11/01/10 17:10	11/19/10 03:16	319-86-8	
gamma-BHC (Lindane)	0.00022U	ug/L	0.011	0.00022	1	11/01/10 17:10	11/19/10 03:16	58-89-9	
Chlordane (Technical)	0.086U	ug/L	0.54	0.086	1	11/01/10 17:10	11/19/10 03:16	57-74-9	
Chlorobenzilate	0.023U	ug/L	0.11	0.023	1	11/01/10 17:10	11/19/10 03:16	510-15-6	
4,4'-DDD	0.0020U	ug/L	0.011	0.0020	1	11/01/10 17:10	11/19/10 03:16	72-54-8	
4,4'-DDE	0.00097U	ug/L	0.011	0.00097	1	11/01/10 17:10	11/19/10 03:16	72-55-9	
4,4'-DDT	0.0039U	ug/L	0.011	0.0039	1	11/01/10 17:10	11/19/10 03:16	50-29-3	
Dieldrin	0.00054U	ug/L	0.011	0.00054	1	11/01/10 17:10	11/19/10 03:16	60-57-1	
Endosulfan I	0.00075U	ug/L	0.011	0.00075	1	11/01/10 17:10	11/19/10 03:16	959-98-8	
Endosulfan II	0.00075U	ug/L	0.011	0.00075	1	11/01/10 17:10	11/19/10 03:16	33213-65-9	
Endosulfan sulfate	0.00065U	ug/L	0.011	0.00065	1	11/01/10 17:10	11/19/10 03:16	1031-07-8	
Endrin	0.0018U	ug/L	0.011	0.0018	1	11/01/10 17:10	11/19/10 03:16	72-20-8	
Endrin aldehyde	0.0076U	ug/L	0.011	0.0076	1	11/01/10 17:10	11/19/10 03:16	7421-93-4	
Heptachlor	0.0016U	ug/L	0.011	0.0016	1	11/01/10 17:10	11/19/10 03:16	76-44-8	
Heptachlor epoxide	0.00043U	ug/L	0.011	0.00043	1	11/01/10 17:10	11/19/10 03:16	1024-57-3	
Methoxychlor	0.0075U	ug/L	0.011	0.0075	1	11/01/10 17:10	11/19/10 03:16	72-43-5	
Pentachloronitrobenzene	0.016U	ug/L	0.11	0.016	1	11/01/10 17:10	11/19/10 03:16	82-68-8	
Toxaphene	0.31U	ug/L	0.54	0.31	1	11/01/10 17:10	11/19/10 03:16	8001-35-2	
Tetrachloro-m-xylene (S)	75 %		66.5-120.3		1	11/01/10 17:10	11/19/10 03:16	877-09-8	
Decachlorobiphenyl (S)	4 %		41.7-109.1		1	11/01/10 17:10	11/19/10 03:16	2051-24-3	2p, J(S5)
8082 GCS PCB									
Analytical Method: EPA 8082		Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.086U	ug/L	0.54	0.086	1	11/01/10 17:10	11/19/10 03:16	12674-11-2	
PCB-1221 (Aroclor 1221)	0.087U	ug/L	0.54	0.087	1	11/01/10 17:10	11/19/10 03:16	11104-28-2	
PCB-1232 (Aroclor 1232)	0.13U	ug/L	0.54	0.13	1	11/01/10 17:10	11/19/10 03:16	11141-16-5	
PCB-1242 (Aroclor 1242)	0.14U	ug/L	0.54	0.14	1	11/01/10 17:10	11/19/10 03:16	53469-21-9	
PCB-1248 (Aroclor 1248)	0.30U	ug/L	0.54	0.30	1	11/01/10 17:10	11/19/10 03:16	12672-29-6	
PCB-1254 (Aroclor 1254)	0.16U	ug/L	0.54	0.16	1	11/01/10 17:10	11/19/10 03:16	11097-69-1	
PCB-1260 (Aroclor 1260)	0.12U	ug/L	0.54	0.12	1	11/01/10 17:10	11/19/10 03:16	11096-82-5	
Tetrachloro-m-xylene (S)	105 %		48-111		1	11/01/10 17:10	11/19/10 03:16	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 196 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3 Lab ID: 3519325034 Collected: 10/27/10 10:25 Received: 10/29/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	8 %		63-121		1	11/01/10 17:10	11/19/10 03:16	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 20:38	60-51-5	
Disulfoton	1.6U	ug/L	5.0	1.6	1	11/03/10 13:00	11/29/10 20:38	298-04-4	
Famphur	1.4U	ug/L	5.0	1.4	1	11/03/10 13:00	11/29/10 20:38	52-85-7	
Methyl parathion	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 20:38	298-00-0	
Parathion (Ethyl parathion)	3.5U	ug/L	10.0	3.5	1	11/03/10 13:00	11/29/10 20:38	56-38-2	
Phorate	3.7U	ug/L	10.0	3.7	1	11/03/10 13:00	11/29/10 20:38	298-02-2	
4-Chloro3nitrobenzotrifluoride	65 %		34.2-122		1	11/03/10 13:00	11/29/10 20:38		9p
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.25U	ug/L	1.0	0.25	1	10/29/10 18:00	11/02/10 20:10	94-75-7	
Dinoseb	0.063U	ug/L	0.21	0.063	1	10/29/10 18:00	11/02/10 20:10	88-85-7	
Pentachlorophenol	0.019U	ug/L	0.031	0.019	1	10/29/10 18:00	11/02/10 20:10	87-86-5	
2,4,5-T	0.046U	ug/L	0.21	0.046	1	10/29/10 18:00	11/02/10 20:10	93-76-5	
2,4,5-TP (Silvex)	9.4	ug/L	2.1	0.54	10	10/29/10 18:00	11/04/10 23:33	93-72-1	
2,4-DCPA (S)	199 %		65.5-125.7		10	10/29/10 18:00	11/04/10 23:33	19719-28-9	J(S2)
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	436	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:05	7440-38-2	
Barium	188	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:05	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:05	7440-41-7	
Cadmium	0.94	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:05	7440-43-9	
Chromium	592	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:05	7440-47-3	
Cobalt	68.7	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:05	7440-48-4	
Copper	9.9	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:05	7440-50-8	
Iron	3220	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:05	7439-89-6	
Lead	5.2	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:05	7439-92-1	
Nickel	128	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:05	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:05	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:05	7440-22-4	
Sodium	2100	mg/L	5.0	2.5	5	11/05/10 11:40	11/11/10 19:07	7440-23-5	D4
Tin	92.5	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:05	7440-31-5	
Vanadium	140	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:05	7440-62-2	
Zinc	175	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:05	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	85.7	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:40	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:40	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 11:53	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3 **Lab ID: 3519325034** Collected: 10/27/10 10:25 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	86.0U	ug/L	500	86.0	10	11/01/10 18:00	11/11/10 07:01	83-32-9	
Acenaphthylene	95.0U	ug/L	500	95.0	10	11/01/10 18:00	11/11/10 07:01	208-96-8	
Acetophenone	145U	ug/L	500	145	10	11/01/10 18:00	11/11/10 07:01	98-86-2	
2-Acetylaminofluorene	65.0U	ug/L	500	65.0	10	11/01/10 18:00	11/11/10 07:01	53-96-3	
4-Aminobiphenyl	283U	ug/L	500	283	10	11/01/10 18:00	11/11/10 07:01	92-67-1	
Anthracene	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 07:01	120-12-7	
Benzo(a)anthracene	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 07:01	56-55-3	
Benzo(a)pyrene	58.0U	ug/L	100	58.0	10	11/01/10 18:00	11/11/10 07:01	50-32-8	
Benzo(b)fluoranthene	62.0U	ug/L	200	62.0	10	11/01/10 18:00	11/11/10 07:01	205-99-2	
Benzo(g,h,i)perylene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 07:01	191-24-2	
Benzo(k)fluoranthene	51.0U	ug/L	400	51.0	10	11/01/10 18:00	11/11/10 07:01	207-08-9	
Benzyl alcohol	102U	ug/L	500	102	10	11/01/10 18:00	11/11/10 07:01	100-51-6	
4-Bromophenylphenyl ether	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 07:01	101-55-3	
Butylbenzylphthalate	72.0U	ug/L	500	72.0	10	11/01/10 18:00	11/11/10 07:01	85-68-7	
4-Chloro-3-methylphenol	62.0U	ug/L	2000	62.0	10	11/01/10 18:00	11/11/10 07:01	59-50-7	
4-Chloroaniline	121U	ug/L	500	121	10	11/01/10 18:00	11/11/10 07:01	106-47-8	
bis(2-Chloroethoxy)methane	295U	ug/L	500	295	10	11/01/10 18:00	11/11/10 07:01	111-91-1	
bis(2-Chloroethyl) ether	75.0U	ug/L	400	75.0	10	11/01/10 18:00	11/11/10 07:01	111-44-4	
bis(2-Chloroisopropyl) ether	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 07:01	108-60-1	
2-Chloronaphthalene	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 07:01	91-58-7	
2-Chlorophenol	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 07:01	95-57-8	
4-Chlorophenylphenyl ether	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 07:01	7005-72-3	
Chrysene	37.0U	ug/L	500	37.0	10	11/01/10 18:00	11/11/10 07:01	218-01-9	
Diallylate	72.7U	ug/L	500	72.7	10	11/01/10 18:00	11/11/10 07:01	2303-16-4	
Dibenz(a,h)anthracene	65.0U	ug/L	200	65.0	10	11/01/10 18:00	11/11/10 07:01	53-70-3	
Dibenzofuran	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 07:01	132-64-9	
1,2-Dichlorobenzene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 07:01	95-50-1	
1,3-Dichlorobenzene	76.0U	ug/L	500	76.0	10	11/01/10 18:00	11/11/10 07:01	541-73-1	
1,4-Dichlorobenzene	77.0U	ug/L	500	77.0	10	11/01/10 18:00	11/11/10 07:01	106-46-7	
3,3'-Dichlorobenzidine	69.0U	ug/L	1000	69.0	10	11/01/10 18:00	11/11/10 07:01	91-94-1	
2,4-Dichlorophenol	56.0U	ug/L	200	56.0	10	11/01/10 18:00	11/11/10 07:01	120-83-2	
2,6-Dichlorophenol	62.0U	ug/L	400	62.0	10	11/01/10 18:00	11/11/10 07:01	87-65-0	
Diethylphthalate	51.0U	ug/L	500	51.0	10	11/01/10 18:00	11/11/10 07:01	84-66-2	
P-Dimethylaminoazobenzene	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 07:01	60-11-7	
7,12-Dimethylbenz(a)anthracene	195U	ug/L	500	195	10	11/01/10 18:00	11/11/10 07:01	57-97-6	
3,3'-Dimethylbenzidine	313U	ug/L	1000	313	10	11/01/10 18:00	11/11/10 07:01	119-93-7	
2,4-Dimethylphenol	158U	ug/L	500	158	10	11/01/10 18:00	11/11/10 07:01	105-67-9	
a,a-Dimethylphenylethylamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 07:01	122-09-8	
Dimethylphthalate	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 07:01	131-11-3	
Di-n-butylphthalate	41.0U	ug/L	500	41.0	10	11/01/10 18:00	11/11/10 07:01	84-74-2	
4,6-Dinitro-2-methylphenol	132U	ug/L	2000	132	10	11/01/10 18:00	11/11/10 07:01	534-52-1	
1,2-Dinitrobenzene	117U	ug/L	500	117	10	11/01/10 18:00	11/11/10 07:01	528-29-0	
1,3-Dinitrobenzene	68.0U	ug/L	800	68.0	10	11/01/10 18:00	11/11/10 07:01	99-65-0	
2,4-Dinitrophenol	157U	ug/L	2000	157	10	11/01/10 18:00	11/11/10 07:01	51-28-5	
2,4-Dinitrotoluene	53.0U	ug/L	200	53.0	10	11/01/10 18:00	11/11/10 07:01	121-14-2	
2,6-Dinitrotoluene	122U	ug/L	200	122	10	11/01/10 18:00	11/11/10 07:01	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 198 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3 **Lab ID: 3519325034** Collected: 10/27/10 10:25 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 07:01	117-84-0	
bis(2-Ethylhexyl)phthalate	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 07:01	117-81-7	
Ethyl methanesulfonate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 07:01	62-50-0	
Fluoranthene	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 07:01	206-44-0	
Fluorene	56.0U	ug/L	500	56.0	10	11/01/10 18:00	11/11/10 07:01	86-73-7	
Hexachlorobenzene	80.0U	ug/L	100	80.0	10	11/01/10 18:00	11/11/10 07:01	118-74-1	
Hexachlorocyclopentadiene	128U	ug/L	500	128	10	11/01/10 18:00	11/11/10 07:01	77-47-4	
Hexachloroethane	71.0U	ug/L	500	71.0	10	11/01/10 18:00	11/11/10 07:01	67-72-1	
Hexachloropropene	141U	ug/L	500	141	10	11/01/10 18:00	11/11/10 07:01	1888-71-7	
Indeno(1,2,3-cd)pyrene	73.0U	ug/L	200	73.0	10	11/01/10 18:00	11/11/10 07:01	193-39-5	
Isodrin	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 07:01	465-73-6	
Isophorone	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 07:01	78-59-1	
Isosafrole	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 07:01	120-58-1	
Kepone	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 07:01	143-50-0	
Methapyrilene	165U	ug/L	500	165	10	11/01/10 18:00	11/11/10 07:01	91-80-5	
3-Methylcholanthrene	104U	ug/L	500	104	10	11/01/10 18:00	11/11/10 07:01	56-49-5	
Methyl methanesulfonate	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 07:01	66-27-3	
1-Methylnaphthalene	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 07:01	90-12-0	
2-Methylnaphthalene	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 07:01	91-57-6	
2-Methylphenol(o-Cresol)	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 07:01	95-48-7	
3&4-Methylphenol(m&p Cresol)	68.7 U	ug/L	1000	66.0	10	11/01/10 18:00	11/11/10 07:01		
2-Naphthylamine	227U	ug/L	500	227	10	11/01/10 18:00	11/11/10 07:01	91-59-8	
Naphthalene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 07:01	91-20-3	
1-Naphthylamine	103U	ug/L	500	103	10	11/01/10 18:00	11/11/10 07:01	134-32-7	
1,4-Naphthoquinone	118U	ug/L	500	118	10	11/01/10 18:00	11/11/10 07:01	130-15-4	
2-Nitroaniline	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 07:01	88-74-4	
3-Nitroaniline	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 07:01	99-09-2	
4-Nitroaniline	69.0U	ug/L	400	69.0	10	11/01/10 18:00	11/11/10 07:01	100-01-6	
Nitrobenzene	109U	ug/L	400	109	10	11/01/10 18:00	11/11/10 07:01	98-95-3	
2-Nitrophenol	81.0U	ug/L	500	81.0	10	11/01/10 18:00	11/11/10 07:01	88-75-5	
4-Nitrophenol	108U	ug/L	2000	108	10	11/01/10 18:00	11/11/10 07:01	100-02-7	
5-Nitro-o-toluidine	129U	ug/L	500	129	10	11/01/10 18:00	11/11/10 07:01	99-55-8	
N-Nitrosodiethylamine	73.0U	ug/L	400	73.0	10	11/01/10 18:00	11/11/10 07:01	55-18-5	
N-Nitrosodimethylamine	97.0U	ug/L	200	97.0	10	11/01/10 18:00	11/11/10 07:01	62-75-9	
N-Nitroso-di-n-butylamine	55.0U	ug/L	400	55.0	10	11/01/10 18:00	11/11/10 07:01	924-16-3	
N-Nitroso-di-n-propylamine	94.0U	ug/L	400	94.0	10	11/01/10 18:00	11/11/10 07:01	621-64-7	
N-Nitrosodiphenylamine	50.0U	ug/L	500	50.0	10	11/01/10 18:00	11/11/10 07:01	86-30-6	
N-Nitrosomethylethylamine	74.0U	ug/L	500	74.0	10	11/01/10 18:00	11/11/10 07:01	10595-95-6	
N-Nitrosopiperidine	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 07:01	100-75-4	
N-Nitrosopyrrolidine	88.0U	ug/L	500	88.0	10	11/01/10 18:00	11/11/10 07:01	930-55-2	
O,O,O-Triethylphosphorothioate	69.0U	ug/L	500	69.0	10	11/01/10 18:00	11/11/10 07:01	126-68-1	
Parathion (Ethyl parathion)	115U	ug/L	500	115	10	11/01/10 18:00	11/11/10 07:01	56-38-2	
Pentachlorobenzene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 07:01	608-93-5	
Pentachlorophenol	66.0U	ug/L	2000	66.0	10	11/01/10 18:00	11/11/10 07:01	87-86-5	
Phenacetin	53.0U	ug/L	500	53.0	10	11/01/10 18:00	11/11/10 07:01	62-44-2	
Phenanthrene	52.0U	ug/L	500	52.0	10	11/01/10 18:00	11/11/10 07:01	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 199 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3 **Lab ID: 3519325034** Collected: 10/27/10 10:25 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
------------	---------	-------	-----	-----	----	----------	----------	---------	------

8270 MSSV SemiVOA App. II

Analytical Method: EPA 8270 Preparation Method: EPA 3510

Phenol	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 07:01	108-95-2	D3
p-Phenylenediamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 07:01	106-50-3	
Pronamide	113U	ug/L	500	113	10	11/01/10 18:00	11/11/10 07:01	23950-58-5	
Pyrene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 07:01	129-00-0	
Safrole	85.0U	ug/L	500	85.0	10	11/01/10 18:00	11/11/10 07:01	94-59-7	
1,2,4,5-Tetrachlorobenzene	70.0U	ug/L	500	70.0	10	11/01/10 18:00	11/11/10 07:01	95-94-3	
2,3,4,6-Tetrachlorophenol	385U	ug/L	500	385	10	11/01/10 18:00	11/11/10 07:01	58-90-2	
Thionazin	61.0U	ug/L	500	61.0	10	11/01/10 18:00	11/11/10 07:01	297-97-2	
O-Toluidine	107U	ug/L	500	107	10	11/01/10 18:00	11/11/10 07:01	95-53-4	
1,2,4-Trichlorobenzene	83.0U	ug/L	500	83.0	10	11/01/10 18:00	11/11/10 07:01	120-82-1	
2,4,5-Trichlorophenol	52.0U	ug/L	400	52.0	10	11/01/10 18:00	11/11/10 07:01	95-95-4	
2,4,6-Trichlorophenol	69.0U	ug/L	200	69.0	10	11/01/10 18:00	11/11/10 07:01	88-06-2	
1,3,5-Trinitrobenzene	122U	ug/L	500	122	10	11/01/10 18:00	11/11/10 07:01	99-35-4	
Nitrobenzene-d5 (S)	53	%	10-110		10	11/01/10 18:00	11/11/10 07:01	4165-60-0	
2-Fluorobiphenyl (S)	62	%	18-110		10	11/01/10 18:00	11/11/10 07:01	321-60-8	
Terphenyl-d14 (S)	59	%	10-123		10	11/01/10 18:00	11/11/10 07:01	1718-51-0	
Phenol-d6 (S)	26	%	10-110		10	11/01/10 18:00	11/11/10 07:01	13127-88-3	
2-Fluorophenol (S)	33	%	18-110		10	11/01/10 18:00	11/11/10 07:01	367-12-4	
2,4,6-Tribromophenol (S)	59	%	10-110		10	11/01/10 18:00	11/11/10 07:01	118-79-6	

8270 MSSV PAH by SCAN

Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510

Acenaphthene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 14:19	83-32-9	
Acenaphthylene	0.50U	ug/L	20.0	0.50	1	10/29/10 18:14	11/01/10 14:19	208-96-8	
Anthracene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 14:19	120-12-7	
Benzo(a)anthracene	0.60U	ug/L	2.0	0.60	1	10/29/10 18:14	11/01/10 14:19	56-55-3	
Benzo(a)pyrene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 14:19	50-32-8	
Benzo(b)fluoranthene	0.50U	ug/L	1.0	0.50	1	10/29/10 18:14	11/01/10 14:19	205-99-2	
Benzo(g,h,i)perylene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:19	191-24-2	
Benzo(k)fluoranthene	0.40U	ug/L	2.5	0.40	1	10/29/10 18:14	11/01/10 14:19	207-08-9	
Chrysene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:19	218-01-9	
Dibenz(a,h)anthracene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 14:19	53-70-3	
Fluoranthene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:19	206-44-0	
Fluorene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 14:19	86-73-7	
Indeno(1,2,3-cd)pyrene	0.40U	ug/L	1.5	0.40	1	10/29/10 18:14	11/01/10 14:19	193-39-5	
1-Methylnaphthalene	1.2 I	ug/L	15.0	0.90	1	10/29/10 18:14	11/01/10 14:19	90-12-0	
2-Methylnaphthalene	1.5 I	ug/L	15.0	0.60	1	10/29/10 18:14	11/01/10 14:19	91-57-6	
Naphthalene	37.6	ug/L	10.0	0.80	1	10/29/10 18:14	11/01/10 14:19	91-20-3	
Phenanthrene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 14:19	85-01-8	
Pyrene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:19	129-00-0	
2-Fluorobiphenyl (S)	55	%	43.9-113		1	10/29/10 18:14	11/01/10 14:19	321-60-8	
Terphenyl-d14 (S)	55	%	24.8-144		1	10/29/10 18:14	11/01/10 14:19	1718-51-0	

8260 MSV

Analytical Method: EPA 8260

Acetone	161	ug/L	50.0	25.0	5	10/29/10 23:00	67-64-1
Acetonitrile	252	ug/L	50.0	25.0	5	10/29/10 23:00	75-05-8
Acrolein	50.0U	ug/L	100	50.0	5	10/29/10 23:00	107-02-8

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 200 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3 **Lab ID: 3519325034** Collected: 10/27/10 10:25 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	107-13-1	
Allyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	107-05-1	
Benzene	4.0 I	ug/L	5.0	2.5	5		10/29/10 23:00	71-43-2	
Bromochloromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	74-97-5	
Bromodichloromethane	1.4U	ug/L	3.0	1.4	5		10/29/10 23:00	75-27-4	
Bromoform	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-25-2	
Bromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	74-83-9	
2-Butanone (MEK)	130	ug/L	50.0	25.0	5		10/29/10 23:00	78-93-3	
Carbon disulfide	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-15-0	
Carbon tetrachloride	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	56-23-5	
Chlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	108-90-7	
Chloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-00-3	
Chloroform	13.9	ug/L	5.0	2.5	5		10/29/10 23:00	67-66-3	
Chloromethane	3.1U	ug/L	5.0	3.1	5		10/29/10 23:00	74-87-3	
Chloroprene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	126-99-8	
Dibromochloromethane	1.3U	ug/L	2.5	1.3	5		10/29/10 23:00	124-48-1	
Dibromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	74-95-3	
trans-1,4-Dichloro-2-butene	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	110-57-6	
Dichlorodifluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-71-8	
1,1-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-34-3	
1,2-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	107-06-2	
1,1-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-35-4	
cis-1,2-Dichloroethene	3.4 I	ug/L	5.0	2.5	5		10/29/10 23:00	156-59-2	
trans-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	156-60-5	
1,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	78-87-5	
1,3-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	142-28-9	
2,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	594-20-7	
1,1-Dichloropropene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	563-58-6	
cis-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 23:00	10061-01-5	
trans-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 23:00	10061-02-6	
Ethylbenzene	45.7	ug/L	5.0	2.5	5		10/29/10 23:00	100-41-4	
Ethyl methacrylate	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	97-63-2	
Hexachloro-1,3-butadiene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	87-68-3	
2-Hexanone	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	591-78-6	
Iodomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	74-88-4	
Isobutyl Alcohol	50.0U	ug/L	100	50.0	5		10/29/10 23:00	78-83-1	
Methacrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	126-98-7	
Methylene Chloride	12.5U	ug/L	25.0	12.5	5		10/29/10 23:00	75-09-2	
Methyl methacrylate	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	80-62-6	
4-Methyl-2-pentanone (MIBK)	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	108-10-1	
Propionitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 23:00	107-12-0	
Styrene	4.1 I	ug/L	5.0	2.5	5		10/29/10 23:00	100-42-5	
1,1,1,2-Tetrachloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	630-20-6	
1,1,2,2-Tetrachloroethane	0.90U	ug/L	2.5	0.90	5		10/29/10 23:00	79-34-5	
Tetrachloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	127-18-4	
Toluene	64.1	ug/L	5.0	2.5	5		10/29/10 23:00	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 201 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-3		Lab ID: 3519325034		Collected: 10/27/10 10:25		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	120-82-1	
1,1,1-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	71-55-6	
1,1,2-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	79-00-5	
Trichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	79-01-6	
Trichlorofluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-69-4	
1,2,3-Trichloropropane	1.8U	ug/L	2.5	1.8	5		10/29/10 23:00	96-18-4	
Vinyl acetate	5.0U	ug/L	10.0	5.0	5		10/29/10 23:00	108-05-4	D3
Vinyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 23:00	75-01-4	
Xylene (Total)	116	ug/L	5.0	2.5	5		10/29/10 23:00	1330-20-7	
4-Bromofluorobenzene (S)	96	%	70-114		5		10/29/10 23:00	460-00-4	
Dibromofluoromethane (S)	99	%	88-117		5		10/29/10 23:00	1868-53-7	
1,2-Dichloroethane-d4 (S)	109	%	86-125		5		10/29/10 23:00	17060-07-0	
Toluene-d8 (S)	101	%	87-113		5		10/29/10 23:00	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	12100	mg/L	100	100	1		11/03/10 11:59		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	32.3	mg/L	5.0	5.0	1		11/02/10 09:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	2560	mg/L	1000	500	200		11/02/10 00:42	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.053	mg/L	0.050	0.025	1	11/01/10 11:30	11/08/10 14:58	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	1890	mg/L	10.0	4.0	200		11/12/10 16:29	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-4		Lab ID: 3519325035		Collected: 10/27/10 12:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
Analytical Method:									
Field pH	7.66	Std. Units			1		10/27/10 12:30		
Field Temperature	39.33	deg C			1		10/27/10 12:30		
Field Specific Conductance	21991	umhos/cm			1		10/27/10 12:30		
Oxygen, Dissolved	3.40	mg/L			1		10/27/10 12:30	7782-44-7	
Turbidity	21.2	NTU			1		10/27/10 12:30		
8011 GCS EDB and DBCP									
Analytical Method: EPA 8011		Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.29	ug/L	0.020	0.0050	1	11/04/10 15:15	11/04/10 21:51	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	11/04/10 15:15	11/04/10 21:51	106-93-4	
8081 GCS Pesticides									
Analytical Method: EPA 8081		Preparation Method: EPA 3510							
Aldrin	0.00051U	ug/L	0.010	0.00051	1	11/01/10 17:10	11/19/10 03:33	309-00-2	
alpha-BHC	0.00030U	ug/L	0.010	0.00030	1	11/01/10 17:10	11/19/10 03:33	319-84-6	
beta-BHC	0.00051U	ug/L	0.010	0.00051	1	11/01/10 17:10	11/19/10 03:33	319-85-7	
delta-BHC	0.00041U	ug/L	0.010	0.00041	1	11/01/10 17:10	11/19/10 03:33	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.010	0.00020	1	11/01/10 17:10	11/19/10 03:33	58-89-9	
Chlordane (Technical)	0.081U	ug/L	0.51	0.081	1	11/01/10 17:10	11/19/10 03:33	57-74-9	
Chlorobenzilate	0.021U	ug/L	0.10	0.021	1	11/01/10 17:10	11/19/10 03:33	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.010	0.0019	1	11/01/10 17:10	11/19/10 03:33	72-54-8	
4,4'-DDE	0.00091U	ug/L	0.010	0.00091	1	11/01/10 17:10	11/19/10 03:33	72-55-9	
4,4'-DDT	0.0037U	ug/L	0.010	0.0037	1	11/01/10 17:10	11/19/10 03:33	50-29-3	
Dieldrin	0.00051U	ug/L	0.010	0.00051	1	11/01/10 17:10	11/19/10 03:33	60-57-1	
Endosulfan I	0.00071U	ug/L	0.010	0.00071	1	11/01/10 17:10	11/19/10 03:33	959-98-8	
Endosulfan II	0.00071U	ug/L	0.010	0.00071	1	11/01/10 17:10	11/19/10 03:33	33213-65-9	
Endosulfan sulfate	0.00061U	ug/L	0.010	0.00061	1	11/01/10 17:10	11/19/10 03:33	1031-07-8	
Endrin	0.0017U	ug/L	0.010	0.0017	1	11/01/10 17:10	11/19/10 03:33	72-20-8	
Endrin aldehyde	0.0072U	ug/L	0.010	0.0072	1	11/01/10 17:10	11/19/10 03:33	7421-93-4	
Heptachlor	0.0015U	ug/L	0.010	0.0015	1	11/01/10 17:10	11/19/10 03:33	76-44-8	
Heptachlor epoxide	0.00041U	ug/L	0.010	0.00041	1	11/01/10 17:10	11/19/10 03:33	1024-57-3	
Methoxychlor	0.0071U	ug/L	0.010	0.0071	1	11/01/10 17:10	11/19/10 03:33	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.10	0.015	1	11/01/10 17:10	11/19/10 03:33	82-68-8	
Toxaphene	0.29U	ug/L	0.51	0.29	1	11/01/10 17:10	11/19/10 03:33	8001-35-2	
Tetrachloro-m-xylene (S)	65 %		66.5-120.3		1	11/01/10 17:10	11/19/10 03:33	877-09-8	2p, J(S5)
Decachlorobiphenyl (S)	8 %		41.7-109.1		1	11/01/10 17:10	11/19/10 03:33	2051-24-3	2p, J(S5)
8082 GCS PCB									
Analytical Method: EPA 8082		Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.081U	ug/L	0.51	0.081	1	11/01/10 17:10	11/19/10 03:33	12674-11-2	
PCB-1221 (Aroclor 1221)	0.082U	ug/L	0.51	0.082	1	11/01/10 17:10	11/19/10 03:33	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.51	0.12	1	11/01/10 17:10	11/19/10 03:33	11141-16-5	
PCB-1242 (Aroclor 1242)	0.13U	ug/L	0.51	0.13	1	11/01/10 17:10	11/19/10 03:33	53469-21-9	
PCB-1248 (Aroclor 1248)	0.28U	ug/L	0.51	0.28	1	11/01/10 17:10	11/19/10 03:33	12672-29-6	
PCB-1254 (Aroclor 1254)	0.15U	ug/L	0.51	0.15	1	11/01/10 17:10	11/19/10 03:33	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.51	0.11	1	11/01/10 17:10	11/19/10 03:33	11096-82-5	
Tetrachloro-m-xylene (S)	84 %		48-111		1	11/01/10 17:10	11/19/10 03:33	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 203 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-4 Lab ID: 3519325035 Collected: 10/27/10 12:30 Received: 10/29/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	11 %		63-121		1	11/01/10 17:10	11/19/10 03:33	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 21:20	60-51-5	
Disulfoton	1.6U	ug/L	5.0	1.6	1	11/03/10 13:00	11/29/10 21:20	298-04-4	
Famphur	1.4U	ug/L	5.0	1.4	1	11/03/10 13:00	11/29/10 21:20	52-85-7	
Methyl parathion	1.9U	ug/L	5.0	1.9	1	11/03/10 13:00	11/29/10 21:20	298-00-0	
Parathion (Ethyl parathion)	3.5U	ug/L	10.0	3.5	1	11/03/10 13:00	11/29/10 21:20	56-38-2	
Phorate	3.7U	ug/L	10.0	3.7	1	11/03/10 13:00	11/29/10 21:20	298-02-2	
4-Chloro3nitrobenzotrifluoride	94 %		34.2-122		1	11/03/10 13:00	11/29/10 21:20		9p
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.22U	ug/L	0.92	0.22	1	10/29/10 18:00	11/02/10 20:37	94-75-7	
Dinoseb	0.056U	ug/L	0.19	0.056	1	10/29/10 18:00	11/02/10 20:37	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.028	0.017	1	10/29/10 18:00	11/02/10 20:37	87-86-5	
2,4,5-T	0.041U	ug/L	0.19	0.041	1	10/29/10 18:00	11/02/10 20:37	93-76-5	
2,4,5-TP (Silvex)	14.0	ug/L	1.9	0.48	10	10/29/10 18:00	11/04/10 23:59	93-72-1	
2,4-DCPA (S)	146 %		65.5-125.7		1	10/29/10 18:00	11/02/10 20:37	19719-28-9	J(S5)
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	156	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:10	7440-38-2	
Barium	104	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:10	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:10	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:10	7440-43-9	
Chromium	371	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:10	7440-47-3	
Cobalt	78.7	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:10	7440-48-4	
Copper	5.4	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:10	7440-50-8	
Iron	1890	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:10	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:10	7439-92-1	
Nickel	138	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:10	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:10	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:10	7440-22-4	
Sodium	1490	mg/L	5.0	2.5	5	11/05/10 11:40	11/11/10 19:18	7440-23-5	D4
Tin	63.7	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:10	7440-31-5	
Vanadium	90.6	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:10	7440-62-2	
Zinc	55.8	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:10	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	9.9	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:44	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/18/10 11:44	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 11:56	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-4 **Lab ID: 3519325035** Collected: 10/27/10 12:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	86.0U	ug/L	500	86.0	10	11/01/10 18:00	11/11/10 06:30	83-32-9	
Acenaphthylene	95.0U	ug/L	500	95.0	10	11/01/10 18:00	11/11/10 06:30	208-96-8	
Acetophenone	145U	ug/L	500	145	10	11/01/10 18:00	11/11/10 06:30	98-86-2	
2-Acetylaminofluorene	65.0U	ug/L	500	65.0	10	11/01/10 18:00	11/11/10 06:30	53-96-3	
4-Aminobiphenyl	283U	ug/L	500	283	10	11/01/10 18:00	11/11/10 06:30	92-67-1	
Anthracene	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 06:30	120-12-7	
Benzo(a)anthracene	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 06:30	56-55-3	
Benzo(a)pyrene	58.0U	ug/L	100	58.0	10	11/01/10 18:00	11/11/10 06:30	50-32-8	
Benzo(b)fluoranthene	62.0U	ug/L	200	62.0	10	11/01/10 18:00	11/11/10 06:30	205-99-2	
Benzo(g,h,i)perylene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 06:30	191-24-2	
Benzo(k)fluoranthene	51.0U	ug/L	400	51.0	10	11/01/10 18:00	11/11/10 06:30	207-08-9	
Benzyl alcohol	102U	ug/L	500	102	10	11/01/10 18:00	11/11/10 06:30	100-51-6	
4-Bromophenylphenyl ether	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 06:30	101-55-3	
Butylbenzylphthalate	72.0U	ug/L	500	72.0	10	11/01/10 18:00	11/11/10 06:30	85-68-7	
4-Chloro-3-methylphenol	62.0U	ug/L	2000	62.0	10	11/01/10 18:00	11/11/10 06:30	59-50-7	
4-Chloroaniline	121U	ug/L	500	121	10	11/01/10 18:00	11/11/10 06:30	106-47-8	
bis(2-Chloroethoxy)methane	295U	ug/L	500	295	10	11/01/10 18:00	11/11/10 06:30	111-91-1	
bis(2-Chloroethyl) ether	75.0U	ug/L	400	75.0	10	11/01/10 18:00	11/11/10 06:30	111-44-4	
bis(2-Chloroisopropyl) ether	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 06:30	108-60-1	
2-Chloronaphthalene	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 06:30	91-58-7	
2-Chlorophenol	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 06:30	95-57-8	
4-Chlorophenylphenyl ether	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 06:30	7005-72-3	
Chrysene	37.0U	ug/L	500	37.0	10	11/01/10 18:00	11/11/10 06:30	218-01-9	
Diallylate	72.7U	ug/L	500	72.7	10	11/01/10 18:00	11/11/10 06:30	2303-16-4	
Dibenz(a,h)anthracene	65.0U	ug/L	200	65.0	10	11/01/10 18:00	11/11/10 06:30	53-70-3	
Dibenzofuran	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 06:30	132-64-9	
1,2-Dichlorobenzene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 06:30	95-50-1	
1,3-Dichlorobenzene	76.0U	ug/L	500	76.0	10	11/01/10 18:00	11/11/10 06:30	541-73-1	
1,4-Dichlorobenzene	77.0U	ug/L	500	77.0	10	11/01/10 18:00	11/11/10 06:30	106-46-7	
3,3'-Dichlorobenzidine	69.0U	ug/L	1000	69.0	10	11/01/10 18:00	11/11/10 06:30	91-94-1	
2,4-Dichlorophenol	56.0U	ug/L	200	56.0	10	11/01/10 18:00	11/11/10 06:30	120-83-2	
2,6-Dichlorophenol	62.0U	ug/L	400	62.0	10	11/01/10 18:00	11/11/10 06:30	87-65-0	
Diethylphthalate	51.0U	ug/L	500	51.0	10	11/01/10 18:00	11/11/10 06:30	84-66-2	
P-Dimethylaminoazobenzene	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 06:30	60-11-7	
7,12-Dimethylbenz(a)anthracene	195U	ug/L	500	195	10	11/01/10 18:00	11/11/10 06:30	57-97-6	
3,3'-Dimethylbenzidine	313U	ug/L	1000	313	10	11/01/10 18:00	11/11/10 06:30	119-93-7	
2,4-Dimethylphenol	158U	ug/L	500	158	10	11/01/10 18:00	11/11/10 06:30	105-67-9	
a,a-Dimethylphenylethylamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 06:30	122-09-8	
Dimethylphthalate	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 06:30	131-11-3	
Di-n-butylphthalate	41.0U	ug/L	500	41.0	10	11/01/10 18:00	11/11/10 06:30	84-74-2	
4,6-Dinitro-2-methylphenol	132U	ug/L	2000	132	10	11/01/10 18:00	11/11/10 06:30	534-52-1	
1,2-Dinitrobenzene	117U	ug/L	500	117	10	11/01/10 18:00	11/11/10 06:30	528-29-0	
1,3-Dinitrobenzene	68.0U	ug/L	800	68.0	10	11/01/10 18:00	11/11/10 06:30	99-65-0	
2,4-Dinitrophenol	157U	ug/L	2000	157	10	11/01/10 18:00	11/11/10 06:30	51-28-5	
2,4-Dinitrotoluene	53.0U	ug/L	200	53.0	10	11/01/10 18:00	11/11/10 06:30	121-14-2	
2,6-Dinitrotoluene	122U	ug/L	200	122	10	11/01/10 18:00	11/11/10 06:30	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 205 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-4 **Lab ID: 3519325035** Collected: 10/27/10 12:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 06:30	117-84-0	
bis(2-Ethylhexyl)phthalate	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 06:30	117-81-7	
Ethyl methanesulfonate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 06:30	62-50-0	
Fluoranthene	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 06:30	206-44-0	
Fluorene	56.0U	ug/L	500	56.0	10	11/01/10 18:00	11/11/10 06:30	86-73-7	
Hexachlorobenzene	80.0U	ug/L	100	80.0	10	11/01/10 18:00	11/11/10 06:30	118-74-1	
Hexachlorocyclopentadiene	128U	ug/L	500	128	10	11/01/10 18:00	11/11/10 06:30	77-47-4	
Hexachloroethane	71.0U	ug/L	500	71.0	10	11/01/10 18:00	11/11/10 06:30	67-72-1	
Hexachloropropene	141U	ug/L	500	141	10	11/01/10 18:00	11/11/10 06:30	1888-71-7	
Indeno(1,2,3-cd)pyrene	73.0U	ug/L	200	73.0	10	11/01/10 18:00	11/11/10 06:30	193-39-5	
Isodrin	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 06:30	465-73-6	
Isophorone	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 06:30	78-59-1	
Isosafrole	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 06:30	120-58-1	
Kepone	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 06:30	143-50-0	
Methapyrilene	165U	ug/L	500	165	10	11/01/10 18:00	11/11/10 06:30	91-80-5	
3-Methylcholanthrene	104U	ug/L	500	104	10	11/01/10 18:00	11/11/10 06:30	56-49-5	
Methyl methanesulfonate	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 06:30	66-27-3	
1-Methylnaphthalene	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 06:30	90-12-0	
2-Methylnaphthalene	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 06:30	91-57-6	
2-Methylphenol(o-Cresol)	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 06:30	95-48-7	
3&4-Methylphenol(m&p Cresol)	66.0U	ug/L	1000	66.0	10	11/01/10 18:00	11/11/10 06:30		
2-Naphthylamine	227U	ug/L	500	227	10	11/01/10 18:00	11/11/10 06:30	91-59-8	
Naphthalene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 06:30	91-20-3	
1-Naphthylamine	103U	ug/L	500	103	10	11/01/10 18:00	11/11/10 06:30	134-32-7	
1,4-Naphthoquinone	118U	ug/L	500	118	10	11/01/10 18:00	11/11/10 06:30	130-15-4	
2-Nitroaniline	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 06:30	88-74-4	
3-Nitroaniline	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 06:30	99-09-2	
4-Nitroaniline	69.0U	ug/L	400	69.0	10	11/01/10 18:00	11/11/10 06:30	100-01-6	
Nitrobenzene	109U	ug/L	400	109	10	11/01/10 18:00	11/11/10 06:30	98-95-3	
2-Nitrophenol	81.0U	ug/L	500	81.0	10	11/01/10 18:00	11/11/10 06:30	88-75-5	
4-Nitrophenol	108U	ug/L	2000	108	10	11/01/10 18:00	11/11/10 06:30	100-02-7	
5-Nitro-o-toluidine	129U	ug/L	500	129	10	11/01/10 18:00	11/11/10 06:30	99-55-8	
N-Nitrosodiethylamine	73.0U	ug/L	400	73.0	10	11/01/10 18:00	11/11/10 06:30	55-18-5	
N-Nitrosodimethylamine	97.0U	ug/L	200	97.0	10	11/01/10 18:00	11/11/10 06:30	62-75-9	
N-Nitroso-di-n-butylamine	55.0U	ug/L	400	55.0	10	11/01/10 18:00	11/11/10 06:30	924-16-3	
N-Nitroso-di-n-propylamine	94.0U	ug/L	400	94.0	10	11/01/10 18:00	11/11/10 06:30	621-64-7	
N-Nitrosodiphenylamine	50.0U	ug/L	500	50.0	10	11/01/10 18:00	11/11/10 06:30	86-30-6	
N-Nitrosomethylethylamine	74.0U	ug/L	500	74.0	10	11/01/10 18:00	11/11/10 06:30	10595-95-6	
N-Nitrosopiperidine	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 06:30	100-75-4	
N-Nitrosopyrrolidine	88.0U	ug/L	500	88.0	10	11/01/10 18:00	11/11/10 06:30	930-55-2	
O,O,O-Triethylphosphorothioate	69.0U	ug/L	500	69.0	10	11/01/10 18:00	11/11/10 06:30	126-68-1	
Parathion (Ethyl parathion)	115U	ug/L	500	115	10	11/01/10 18:00	11/11/10 06:30	56-38-2	
Pentachlorobenzene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 06:30	608-93-5	
Pentachlorophenol	66.0U	ug/L	2000	66.0	10	11/01/10 18:00	11/11/10 06:30	87-86-5	
Phenacetin	53.0U	ug/L	500	53.0	10	11/01/10 18:00	11/11/10 06:30	62-44-2	
Phenanthrene	52.0U	ug/L	500	52.0	10	11/01/10 18:00	11/11/10 06:30	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 206 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-4		Lab ID: 3519325035		Collected: 10/27/10 12:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenol	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 06:30	108-95-2	D3
p-Phenylenediamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 06:30	106-50-3	
Pronamide	113U	ug/L	500	113	10	11/01/10 18:00	11/11/10 06:30	23950-58-5	
Pyrene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 06:30	129-00-0	
Safrole	85.0U	ug/L	500	85.0	10	11/01/10 18:00	11/11/10 06:30	94-59-7	
1,2,4,5-Tetrachlorobenzene	70.0U	ug/L	500	70.0	10	11/01/10 18:00	11/11/10 06:30	95-94-3	
2,3,4,6-Tetrachlorophenol	385U	ug/L	500	385	10	11/01/10 18:00	11/11/10 06:30	58-90-2	
Thionazin	61.0U	ug/L	500	61.0	10	11/01/10 18:00	11/11/10 06:30	297-97-2	
O-Toluidine	107U	ug/L	500	107	10	11/01/10 18:00	11/11/10 06:30	95-53-4	
1,2,4-Trichlorobenzene	83.0U	ug/L	500	83.0	10	11/01/10 18:00	11/11/10 06:30	120-82-1	
2,4,5-Trichlorophenol	52.0U	ug/L	400	52.0	10	11/01/10 18:00	11/11/10 06:30	95-95-4	
2,4,6-Trichlorophenol	69.0U	ug/L	200	69.0	10	11/01/10 18:00	11/11/10 06:30	88-06-2	
1,3,5-Trinitrobenzene	122U	ug/L	500	122	10	11/01/10 18:00	11/11/10 06:30	99-35-4	
Nitrobenzene-d5 (S)	74 %		10-110		10	11/01/10 18:00	11/11/10 06:30	4165-60-0	
2-Fluorobiphenyl (S)	84 %		18-110		10	11/01/10 18:00	11/11/10 06:30	321-60-8	
Terphenyl-d14 (S)	88 %		10-123		10	11/01/10 18:00	11/11/10 06:30	1718-51-0	
Phenol-d6 (S)	31 %		10-110		10	11/01/10 18:00	11/11/10 06:30	13127-88-3	
2-Fluorophenol (S)	44 %		18-110		10	11/01/10 18:00	11/11/10 06:30	367-12-4	
2,4,6-Tribromophenol (S)	83 %		10-110		10	11/01/10 18:00	11/11/10 06:30	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 14:40	83-32-9	
Acenaphthylene	0.50U	ug/L	20.0	0.50	1	10/29/10 18:14	11/01/10 14:40	208-96-8	
Anthracene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 14:40	120-12-7	
Benzo(a)anthracene	0.60U	ug/L	2.0	0.60	1	10/29/10 18:14	11/01/10 14:40	56-55-3	
Benzo(a)pyrene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 14:40	50-32-8	
Benzo(b)fluoranthene	0.50U	ug/L	1.0	0.50	1	10/29/10 18:14	11/01/10 14:40	205-99-2	
Benzo(g,h,i)perylene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:40	191-24-2	
Benzo(k)fluoranthene	0.40U	ug/L	2.5	0.40	1	10/29/10 18:14	11/01/10 14:40	207-08-9	
Chrysene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:40	218-01-9	
Dibenz(a,h)anthracene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 14:40	53-70-3	
Fluoranthene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:40	206-44-0	
Fluorene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 14:40	86-73-7	
Indeno(1,2,3-cd)pyrene	0.40U	ug/L	1.5	0.40	1	10/29/10 18:14	11/01/10 14:40	193-39-5	
1-Methylnaphthalene	0.90U	ug/L	15.0	0.90	1	10/29/10 18:14	11/01/10 14:40	90-12-0	
2-Methylnaphthalene	0.60U	ug/L	15.0	0.60	1	10/29/10 18:14	11/01/10 14:40	91-57-6	
Naphthalene	16.6	ug/L	10.0	0.80	1	10/29/10 18:14	11/01/10 14:40	91-20-3	
Phenanthrene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 14:40	85-01-8	
Pyrene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 14:40	129-00-0	
2-Fluorobiphenyl (S)	77 %		43.9-113		1	10/29/10 18:14	11/01/10 14:40	321-60-8	
Terphenyl-d14 (S)	65 %		24.8-144		1	10/29/10 18:14	11/01/10 14:40	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	67-64-1	
Acetonitrile	116	ug/L	50.0	25.0	5		10/29/10 21:50	75-05-8	
Acrolein	50.0U	ug/L	100	50.0	5		10/29/10 21:50	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 207 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-4 **Lab ID: 3519325035** Collected: 10/27/10 12:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	107-13-1	
Allyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	107-05-1	
Benzene	5.6	ug/L	5.0	2.5	5		10/29/10 21:50	71-43-2	
Bromochloromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	74-97-5	
Bromodichloromethane	1.4U	ug/L	3.0	1.4	5		10/29/10 21:50	75-27-4	
Bromoform	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-25-2	
Bromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	74-83-9	
2-Butanone (MEK)	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	78-93-3	
Carbon disulfide	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-15-0	
Carbon tetrachloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	56-23-5	
Chlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	108-90-7	
Chloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-00-3	
Chloroform	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	67-66-3	
Chloromethane	3.1U	ug/L	5.0	3.1	5		10/29/10 21:50	74-87-3	
Chloroprene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	126-99-8	
Dibromochloromethane	1.3U	ug/L	2.5	1.3	5		10/29/10 21:50	124-48-1	
Dibromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	74-95-3	
trans-1,4-Dichloro-2-butene	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	110-57-6	
Dichlorodifluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-71-8	
1,1-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-34-3	
1,2-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	107-06-2	
1,1-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-35-4	
cis-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	156-59-2	
trans-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	156-60-5	
1,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	78-87-5	
1,3-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	142-28-9	
2,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	594-20-7	
1,1-Dichloropropene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	563-58-6	
cis-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 21:50	10061-01-5	
trans-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 21:50	10061-02-6	
Ethylbenzene	47.7	ug/L	5.0	2.5	5		10/29/10 21:50	100-41-4	
Ethyl methacrylate	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	97-63-2	
Hexachloro-1,3-butadiene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	87-68-3	
2-Hexanone	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	591-78-6	
Iodomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	74-88-4	
Isobutyl Alcohol	50.0U	ug/L	100	50.0	5		10/29/10 21:50	78-83-1	
Methacrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	126-98-7	
Methylene Chloride	12.5U	ug/L	25.0	12.5	5		10/29/10 21:50	75-09-2	
Methyl methacrylate	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	80-62-6	
4-Methyl-2-pentanone (MIBK)	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	108-10-1	
Propionitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 21:50	107-12-0	
Styrene	2.6 I	ug/L	5.0	2.5	5		10/29/10 21:50	100-42-5	
1,1,1,2-Tetrachloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	630-20-6	
1,1,2,2-Tetrachloroethane	0.90U	ug/L	2.5	0.90	5		10/29/10 21:50	79-34-5	
Tetrachloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	127-18-4	
Toluene	26.8	ug/L	5.0	2.5	5		10/29/10 21:50	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 208 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-4		Lab ID: 3519325035		Collected: 10/27/10 12:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	120-82-1	
1,1,1-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	71-55-6	
1,1,2-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	79-00-5	
Trichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	79-01-6	
Trichlorofluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-69-4	
1,2,3-Trichloropropane	1.8U	ug/L	2.5	1.8	5		10/29/10 21:50	96-18-4	
Vinyl acetate	5.0U	ug/L	10.0	5.0	5		10/29/10 21:50	108-05-4	
Vinyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 21:50	75-01-4	
Xylene (Total)	82.5	ug/L	5.0	2.5	5		10/29/10 21:50	1330-20-7	
4-Bromofluorobenzene (S)	96 %		70-114		5		10/29/10 21:50	460-00-4	D3
Dibromofluoromethane (S)	98 %		88-117		5		10/29/10 21:50	1868-53-7	
1,2-Dichloroethane-d4 (S)	110 %		86-125		5		10/29/10 21:50	17060-07-0	
Toluene-d8 (S)	101 %		87-113		5		10/29/10 21:50	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	9140	mg/L	100	100	1		11/03/10 12:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	31.9	mg/L	5.0	5.0	1		11/02/10 09:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	2320	mg/L	1000	500	200		11/02/10 00:54	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.027 I	mg/L	0.050	0.025	1	11/01/10 11:30	11/08/10 17:51	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	1150	mg/L	5.0	2.0	100		11/12/10 16:23	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-5		Lab ID: 3519325036		Collected: 10/27/10 13:00		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
Analytical Method:									
Field pH	7.38	Std. Units			1		12/22/10 12:19		
Field Temperature	37.67	deg C			1		12/22/10 12:19		
Field Specific Conductance	11308	umhos/cm			1		12/22/10 12:19		
Oxygen, Dissolved	0.90	mg/L			1		12/22/10 12:19	7782-44-7	
Turbidity	23.0	NTU			1		12/22/10 12:19		
8011 GCS EDB and DBCP									
Analytical Method: EPA 8011		Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.11	ug/L	0.020	0.0050	1	11/04/10 15:15	11/04/10 22:05	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	11/04/10 15:15	11/04/10 22:05	106-93-4	
8081 GCS Pesticides									
Analytical Method: EPA 8081		Preparation Method: EPA 3510							
Aldrin	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 03:50	309-00-2	
alpha-BHC	0.00033U	ug/L	0.011	0.00033	1	11/01/10 17:10	11/19/10 03:50	319-84-6	
beta-BHC	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 03:50	319-85-7	
delta-BHC	0.00044U	ug/L	0.011	0.00044	1	11/01/10 17:10	11/19/10 03:50	319-86-8	
gamma-BHC (Lindane)	0.00022U	ug/L	0.011	0.00022	1	11/01/10 17:10	11/19/10 03:50	58-89-9	
Chlordane (Technical)	0.088U	ug/L	0.55	0.088	1	11/01/10 17:10	11/19/10 03:50	57-74-9	
Chlorobenzilate	0.023U	ug/L	0.11	0.023	1	11/01/10 17:10	11/19/10 03:50	510-15-6	
4,4'-DDD	0.0021U	ug/L	0.011	0.0021	1	11/01/10 17:10	11/19/10 03:50	72-54-8	
4,4'-DDE	0.00098U	ug/L	0.011	0.00098	1	11/01/10 17:10	11/19/10 03:50	72-55-9	
4,4'-DDT	0.0039U	ug/L	0.011	0.0039	1	11/01/10 17:10	11/19/10 03:50	50-29-3	
Dieldrin	0.00055U	ug/L	0.011	0.00055	1	11/01/10 17:10	11/19/10 03:50	60-57-1	
Endosulfan I	0.00077U	ug/L	0.011	0.00077	1	11/01/10 17:10	11/19/10 03:50	959-98-8	
Endosulfan II	0.00077U	ug/L	0.011	0.00077	1	11/01/10 17:10	11/19/10 03:50	33213-65-9	
Endosulfan sulfate	0.00066U	ug/L	0.011	0.00066	1	11/01/10 17:10	11/19/10 03:50	1031-07-8	
Endrin	0.0019U	ug/L	0.011	0.0019	1	11/01/10 17:10	11/19/10 03:50	72-20-8	
Endrin aldehyde	0.0078U	ug/L	0.011	0.0078	1	11/01/10 17:10	11/19/10 03:50	7421-93-4	
Heptachlor	0.0016U	ug/L	0.011	0.0016	1	11/01/10 17:10	11/19/10 03:50	76-44-8	
Heptachlor epoxide	0.00044U	ug/L	0.011	0.00044	1	11/01/10 17:10	11/19/10 03:50	1024-57-3	
Methoxychlor	0.0077U	ug/L	0.011	0.0077	1	11/01/10 17:10	11/19/10 03:50	72-43-5	
Pentachloronitrobenzene	0.016U	ug/L	0.11	0.016	1	11/01/10 17:10	11/19/10 03:50	82-68-8	
Toxaphene	0.31U	ug/L	0.55	0.31	1	11/01/10 17:10	11/19/10 03:50	8001-35-2	
Tetrachloro-m-xylene (S)	56 %		66.5-120.3		1	11/01/10 17:10	11/19/10 03:50	877-09-8	2p, J(S5)
Decachlorobiphenyl (S)	18 %		41.7-109.1		1	11/01/10 17:10	11/19/10 03:50	2051-24-3	2p, J(S5)
8082 GCS PCB									
Analytical Method: EPA 8082		Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.087U	ug/L	0.55	0.087	1	11/01/10 17:10	11/19/10 03:50	12674-11-2	
PCB-1221 (Aroclor 1221)	0.089U	ug/L	0.55	0.089	1	11/01/10 17:10	11/19/10 03:50	11104-28-2	
PCB-1232 (Aroclor 1232)	0.13U	ug/L	0.55	0.13	1	11/01/10 17:10	11/19/10 03:50	11141-16-5	
PCB-1242 (Aroclor 1242)	0.14U	ug/L	0.55	0.14	1	11/01/10 17:10	11/19/10 03:50	53469-21-9	
PCB-1248 (Aroclor 1248)	0.30U	ug/L	0.55	0.30	1	11/01/10 17:10	11/19/10 03:50	12672-29-6	
PCB-1254 (Aroclor 1254)	0.16U	ug/L	0.55	0.16	1	11/01/10 17:10	11/19/10 03:50	11097-69-1	
PCB-1260 (Aroclor 1260)	0.12U	ug/L	0.55	0.12	1	11/01/10 17:10	11/19/10 03:50	11096-82-5	
Tetrachloro-m-xylene (S)	76 %		48-111		1	11/01/10 17:10	11/19/10 03:50	877-09-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 210 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-5		Lab ID: 3519325036		Collected: 10/27/10 13:00		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	20 %		63-121		1	11/01/10 17:10	11/19/10 03:50	2051-24-3	1p, J(S5)
8141 GCS O/P Pesticides Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.21U	ug/L	0.57	0.21	1	11/03/10 13:00	11/29/10 22:01	60-51-5	
Disulfoton	0.18U	ug/L	0.57	0.18	1	11/03/10 13:00	11/29/10 22:01	298-04-4	
Famphur	0.16U	ug/L	0.57	0.16	1	11/03/10 13:00	11/29/10 22:01	52-85-7	
Methyl parathion	0.22U	ug/L	0.57	0.22	1	11/03/10 13:00	11/29/10 22:01	298-00-0	
Parathion (Ethyl parathion)	0.40U	ug/L	1.1	0.40	1	11/03/10 13:00	11/29/10 22:01	56-38-2	
Phorate	0.42U	ug/L	1.1	0.42	1	11/03/10 13:00	11/29/10 22:01	298-02-2	
4-Chloro3nitrobenzotrifluoride	162 %		34.2-122		1	11/03/10 13:00	11/29/10 22:01		9p, J(S5)
8151 Chlorinated Herbicides Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.24U	ug/L	1.0	0.24	1	11/02/10 18:00	11/05/10 07:02	94-75-7	
Dinoseb	0.061U	ug/L	0.20	0.061	1	11/02/10 18:00	11/05/10 07:02	88-85-7	
Pentachlorophenol	0.018U	ug/L	0.031	0.018	1	11/02/10 18:00	11/05/10 07:02	87-86-5	
2,4,5-T	0.045U	ug/L	0.20	0.045	1	11/02/10 18:00	11/05/10 07:02	93-76-5	
2,4,5-TP (Silvex)	1.8	ug/L	0.20	0.053	1	11/02/10 18:00	11/05/10 07:02	93-72-1	
2,4-DCPA (S)	93 %		65.5-125.7		1	11/02/10 18:00	11/05/10 07:02	19719-28-9	
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	102	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:15	7440-38-2	
Barium	125	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:15	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:15	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:15	7440-43-9	
Chromium	141	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:15	7440-47-3	
Cobalt	61.1	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:15	7440-48-4	
Copper	13.0	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:15	7440-50-8	
Iron	1290	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:15	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:15	7439-92-1	
Nickel	61.4	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:15	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:15	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:15	7440-22-4	
Sodium	855	mg/L	5.0	2.5	5	11/05/10 11:40	11/11/10 19:22	7440-23-5	D4
Tin	25.0U	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:15	7440-31-5	
Vanadium	52.6	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:15	7440-62-2	
Zinc	16.6 I	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:15	7440-66-6	
6020 MET ICPMS Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	11.5	ug/L	10.0	5.0	10	11/05/10 11:40	11/19/10 16:54	7440-36-0	
Thallium	5.0U	ug/L	10.0	5.0	10	11/05/10 11:40	11/19/10 16:54	7440-28-0	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 11:59	7439-97-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 211 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-5 **Lab ID: 3519325036** Collected: 10/27/10 13:00 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	86.0U	ug/L	500	86.0	10	11/01/10 18:00	11/11/10 05:59	83-32-9	
Acenaphthylene	95.0U	ug/L	500	95.0	10	11/01/10 18:00	11/11/10 05:59	208-96-8	
Acetophenone	145U	ug/L	500	145	10	11/01/10 18:00	11/11/10 05:59	98-86-2	
2-Acetylaminofluorene	65.0U	ug/L	500	65.0	10	11/01/10 18:00	11/11/10 05:59	53-96-3	
4-Aminobiphenyl	283U	ug/L	500	283	10	11/01/10 18:00	11/11/10 05:59	92-67-1	
Anthracene	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 05:59	120-12-7	
Benzo(a)anthracene	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 05:59	56-55-3	
Benzo(a)pyrene	58.0U	ug/L	100	58.0	10	11/01/10 18:00	11/11/10 05:59	50-32-8	
Benzo(b)fluoranthene	62.0U	ug/L	200	62.0	10	11/01/10 18:00	11/11/10 05:59	205-99-2	
Benzo(g,h,i)perylene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 05:59	191-24-2	
Benzo(k)fluoranthene	51.0U	ug/L	400	51.0	10	11/01/10 18:00	11/11/10 05:59	207-08-9	
Benzyl alcohol	102U	ug/L	500	102	10	11/01/10 18:00	11/11/10 05:59	100-51-6	
4-Bromophenylphenyl ether	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 05:59	101-55-3	
Butylbenzylphthalate	72.0U	ug/L	500	72.0	10	11/01/10 18:00	11/11/10 05:59	85-68-7	
4-Chloro-3-methylphenol	62.0U	ug/L	2000	62.0	10	11/01/10 18:00	11/11/10 05:59	59-50-7	
4-Chloroaniline	121U	ug/L	500	121	10	11/01/10 18:00	11/11/10 05:59	106-47-8	
bis(2-Chloroethoxy)methane	295U	ug/L	500	295	10	11/01/10 18:00	11/11/10 05:59	111-91-1	
bis(2-Chloroethyl) ether	75.0U	ug/L	400	75.0	10	11/01/10 18:00	11/11/10 05:59	111-44-4	
bis(2-Chloroisopropyl) ether	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 05:59	108-60-1	
2-Chloronaphthalene	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 05:59	91-58-7	
2-Chlorophenol	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 05:59	95-57-8	
4-Chlorophenylphenyl ether	63.0U	ug/L	500	63.0	10	11/01/10 18:00	11/11/10 05:59	7005-72-3	
Chrysene	37.0U	ug/L	500	37.0	10	11/01/10 18:00	11/11/10 05:59	218-01-9	
Diallylate	72.7U	ug/L	500	72.7	10	11/01/10 18:00	11/11/10 05:59	2303-16-4	
Dibenz(a,h)anthracene	65.0U	ug/L	200	65.0	10	11/01/10 18:00	11/11/10 05:59	53-70-3	
Dibenzofuran	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 05:59	132-64-9	
1,2-Dichlorobenzene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 05:59	95-50-1	
1,3-Dichlorobenzene	76.0U	ug/L	500	76.0	10	11/01/10 18:00	11/11/10 05:59	541-73-1	
1,4-Dichlorobenzene	77.0U	ug/L	500	77.0	10	11/01/10 18:00	11/11/10 05:59	106-46-7	
3,3'-Dichlorobenzidine	69.0U	ug/L	1000	69.0	10	11/01/10 18:00	11/11/10 05:59	91-94-1	
2,4-Dichlorophenol	56.0U	ug/L	200	56.0	10	11/01/10 18:00	11/11/10 05:59	120-83-2	
2,6-Dichlorophenol	62.0U	ug/L	400	62.0	10	11/01/10 18:00	11/11/10 05:59	87-65-0	
Diethylphthalate	51.0U	ug/L	500	51.0	10	11/01/10 18:00	11/11/10 05:59	84-66-2	
P-Dimethylaminoazobenzene	67.0U	ug/L	500	67.0	10	11/01/10 18:00	11/11/10 05:59	60-11-7	
7,12-Dimethylbenz(a)anthracene	195U	ug/L	500	195	10	11/01/10 18:00	11/11/10 05:59	57-97-6	
3,3'-Dimethylbenzidine	313U	ug/L	1000	313	10	11/01/10 18:00	11/11/10 05:59	119-93-7	
2,4-Dimethylphenol	158U	ug/L	500	158	10	11/01/10 18:00	11/11/10 05:59	105-67-9	
a,a-Dimethylphenylethylamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 05:59	122-09-8	
Dimethylphthalate	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 05:59	131-11-3	
Di-n-butylphthalate	41.0U	ug/L	500	41.0	10	11/01/10 18:00	11/11/10 05:59	84-74-2	
4,6-Dinitro-2-methylphenol	132U	ug/L	2000	132	10	11/01/10 18:00	11/11/10 05:59	534-52-1	
1,2-Dinitrobenzene	117U	ug/L	500	117	10	11/01/10 18:00	11/11/10 05:59	528-29-0	
1,3-Dinitrobenzene	68.0U	ug/L	800	68.0	10	11/01/10 18:00	11/11/10 05:59	99-65-0	
2,4-Dinitrophenol	157U	ug/L	2000	157	10	11/01/10 18:00	11/11/10 05:59	51-28-5	
2,4-Dinitrotoluene	53.0U	ug/L	200	53.0	10	11/01/10 18:00	11/11/10 05:59	121-14-2	
2,6-Dinitrotoluene	122U	ug/L	200	122	10	11/01/10 18:00	11/11/10 05:59	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 212 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-5 **Lab ID: 3519325036** Collected: 10/27/10 13:00 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 05:59	117-84-0	
bis(2-Ethylhexyl)phthalate	80.0U	ug/L	500	80.0	10	11/01/10 18:00	11/11/10 05:59	117-81-7	
Ethyl methanesulfonate	90.0U	ug/L	500	90.0	10	11/01/10 18:00	11/11/10 05:59	62-50-0	
Fluoranthene	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 05:59	206-44-0	
Fluorene	56.0U	ug/L	500	56.0	10	11/01/10 18:00	11/11/10 05:59	86-73-7	
Hexachlorobenzene	80.0U	ug/L	100	80.0	10	11/01/10 18:00	11/11/10 05:59	118-74-1	
Hexachlorocyclopentadiene	128U	ug/L	500	128	10	11/01/10 18:00	11/11/10 05:59	77-47-4	
Hexachloroethane	71.0U	ug/L	500	71.0	10	11/01/10 18:00	11/11/10 05:59	67-72-1	
Hexachloropropene	141U	ug/L	500	141	10	11/01/10 18:00	11/11/10 05:59	1888-71-7	
Indeno(1,2,3-cd)pyrene	73.0U	ug/L	200	73.0	10	11/01/10 18:00	11/11/10 05:59	193-39-5	
Isodrin	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 05:59	465-73-6	
Isophorone	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 05:59	78-59-1	
Isosafrole	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 05:59	120-58-1	
Kepone	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 05:59	143-50-0	
Methapyrilene	165U	ug/L	500	165	10	11/01/10 18:00	11/11/10 05:59	91-80-5	
3-Methylcholanthrene	104U	ug/L	500	104	10	11/01/10 18:00	11/11/10 05:59	56-49-5	
Methyl methanesulfonate	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 05:59	66-27-3	
1-Methylnaphthalene	100U	ug/L	500	100	10	11/01/10 18:00	11/11/10 05:59	90-12-0	
2-Methylnaphthalene	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 05:59	91-57-6	
2-Methylphenol(o-Cresol)	73.0U	ug/L	500	73.0	10	11/01/10 18:00	11/11/10 05:59	95-48-7	
3&4-Methylphenol(m&p Cresol)	66.0U	ug/L	1000	66.0	10	11/01/10 18:00	11/11/10 05:59		
2-Naphthylamine	227U	ug/L	500	227	10	11/01/10 18:00	11/11/10 05:59	91-59-8	
Naphthalene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 05:59	91-20-3	
1-Naphthylamine	103U	ug/L	500	103	10	11/01/10 18:00	11/11/10 05:59	134-32-7	
1,4-Naphthoquinone	118U	ug/L	500	118	10	11/01/10 18:00	11/11/10 05:59	130-15-4	
2-Nitroaniline	60.0U	ug/L	500	60.0	10	11/01/10 18:00	11/11/10 05:59	88-74-4	
3-Nitroaniline	99.0U	ug/L	500	99.0	10	11/01/10 18:00	11/11/10 05:59	99-09-2	
4-Nitroaniline	69.0U	ug/L	400	69.0	10	11/01/10 18:00	11/11/10 05:59	100-01-6	
Nitrobenzene	109U	ug/L	400	109	10	11/01/10 18:00	11/11/10 05:59	98-95-3	
2-Nitrophenol	81.0U	ug/L	500	81.0	10	11/01/10 18:00	11/11/10 05:59	88-75-5	
4-Nitrophenol	108U	ug/L	2000	108	10	11/01/10 18:00	11/11/10 05:59	100-02-7	
5-Nitro-o-toluidine	129U	ug/L	500	129	10	11/01/10 18:00	11/11/10 05:59	99-55-8	
N-Nitrosodiethylamine	73.0U	ug/L	400	73.0	10	11/01/10 18:00	11/11/10 05:59	55-18-5	
N-Nitrosodimethylamine	97.0U	ug/L	200	97.0	10	11/01/10 18:00	11/11/10 05:59	62-75-9	
N-Nitroso-di-n-butylamine	55.0U	ug/L	400	55.0	10	11/01/10 18:00	11/11/10 05:59	924-16-3	
N-Nitroso-di-n-propylamine	94.0U	ug/L	400	94.0	10	11/01/10 18:00	11/11/10 05:59	621-64-7	
N-Nitrosodiphenylamine	50.0U	ug/L	500	50.0	10	11/01/10 18:00	11/11/10 05:59	86-30-6	
N-Nitrosomethylethylamine	74.0U	ug/L	500	74.0	10	11/01/10 18:00	11/11/10 05:59	10595-95-6	
N-Nitrosopiperidine	64.0U	ug/L	500	64.0	10	11/01/10 18:00	11/11/10 05:59	100-75-4	
N-Nitrosopyrrolidine	88.0U	ug/L	500	88.0	10	11/01/10 18:00	11/11/10 05:59	930-55-2	
O,O,O-Triethylphosphorothioate	69.0U	ug/L	500	69.0	10	11/01/10 18:00	11/11/10 05:59	126-68-1	
Parathion (Ethyl parathion)	115U	ug/L	500	115	10	11/01/10 18:00	11/11/10 05:59	56-38-2	
Pentachlorobenzene	78.0U	ug/L	500	78.0	10	11/01/10 18:00	11/11/10 05:59	608-93-5	
Pentachlorophenol	66.0U	ug/L	2000	66.0	10	11/01/10 18:00	11/11/10 05:59	87-86-5	
Phenacetin	53.0U	ug/L	500	53.0	10	11/01/10 18:00	11/11/10 05:59	62-44-2	
Phenanthrene	52.0U	ug/L	500	52.0	10	11/01/10 18:00	11/11/10 05:59	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 213 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: C-5		Lab ID: 3519325036		Collected: 10/27/10 13:00		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II		Analytical Method: EPA 8270 Preparation Method: EPA 3510							
Phenol	54.0U	ug/L	500	54.0	10	11/01/10 18:00	11/11/10 05:59	108-95-2	D3
p-Phenylenediamine	1000U	ug/L	2000	1000	10	11/01/10 18:00	11/11/10 05:59	106-50-3	
Pronamide	113U	ug/L	500	113	10	11/01/10 18:00	11/11/10 05:59	23950-58-5	
Pyrene	68.0U	ug/L	500	68.0	10	11/01/10 18:00	11/11/10 05:59	129-00-0	
Safrole	85.0U	ug/L	500	85.0	10	11/01/10 18:00	11/11/10 05:59	94-59-7	
1,2,4,5-Tetrachlorobenzene	70.0U	ug/L	500	70.0	10	11/01/10 18:00	11/11/10 05:59	95-94-3	
2,3,4,6-Tetrachlorophenol	385U	ug/L	500	385	10	11/01/10 18:00	11/11/10 05:59	58-90-2	
Thionazin	61.0U	ug/L	500	61.0	10	11/01/10 18:00	11/11/10 05:59	297-97-2	
O-Toluidine	107U	ug/L	500	107	10	11/01/10 18:00	11/11/10 05:59	95-53-4	
1,2,4-Trichlorobenzene	83.0U	ug/L	500	83.0	10	11/01/10 18:00	11/11/10 05:59	120-82-1	
2,4,5-Trichlorophenol	52.0U	ug/L	400	52.0	10	11/01/10 18:00	11/11/10 05:59	95-95-4	
2,4,6-Trichlorophenol	69.0U	ug/L	200	69.0	10	11/01/10 18:00	11/11/10 05:59	88-06-2	
1,3,5-Trinitrobenzene	122U	ug/L	500	122	10	11/01/10 18:00	11/11/10 05:59	99-35-4	
Nitrobenzene-d5 (S)	64 %		10-110		10	11/01/10 18:00	11/11/10 05:59	4165-60-0	
2-Fluorobiphenyl (S)	89 %		18-110		10	11/01/10 18:00	11/11/10 05:59	321-60-8	
Terphenyl-d14 (S)	95 %		10-123		10	11/01/10 18:00	11/11/10 05:59	1718-51-0	
Phenol-d6 (S)	30 %		10-110		10	11/01/10 18:00	11/11/10 05:59	13127-88-3	
2-Fluorophenol (S)	45 %		18-110		10	11/01/10 18:00	11/11/10 05:59	367-12-4	
2,4,6-Tribromophenol (S)	84 %		10-110		10	11/01/10 18:00	11/11/10 05:59	118-79-6	
8270 MSSV PAH by SCAN		Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510							
Acenaphthene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 15:02	83-32-9	
Acenaphthylene	0.50U	ug/L	20.0	0.50	1	10/29/10 18:14	11/01/10 15:02	208-96-8	
Anthracene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 15:02	120-12-7	
Benzo(a)anthracene	0.60U	ug/L	2.0	0.60	1	10/29/10 18:14	11/01/10 15:02	56-55-3	
Benzo(a)pyrene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 15:02	50-32-8	
Benzo(b)fluoranthene	0.50U	ug/L	1.0	0.50	1	10/29/10 18:14	11/01/10 15:02	205-99-2	
Benzo(g,h,i)perylene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 15:02	191-24-2	
Benzo(k)fluoranthene	0.40U	ug/L	2.5	0.40	1	10/29/10 18:14	11/01/10 15:02	207-08-9	
Chrysene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 15:02	218-01-9	
Dibenz(a,h)anthracene	0.50U	ug/L	2.0	0.50	1	10/29/10 18:14	11/01/10 15:02	53-70-3	
Fluoranthene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 15:02	206-44-0	
Fluorene	0.30U	ug/L	10.0	0.30	1	10/29/10 18:14	11/01/10 15:02	86-73-7	
Indeno(1,2,3-cd)pyrene	0.40U	ug/L	1.5	0.40	1	10/29/10 18:14	11/01/10 15:02	193-39-5	
1-Methylnaphthalene	0.90U	ug/L	15.0	0.90	1	10/29/10 18:14	11/01/10 15:02	90-12-0	
2-Methylnaphthalene	0.60U	ug/L	15.0	0.60	1	10/29/10 18:14	11/01/10 15:02	91-57-6	
Naphthalene	7.8 I	ug/L	10.0	0.80	1	10/29/10 18:14	11/01/10 15:02	91-20-3	
Phenanthrene	0.50U	ug/L	10.0	0.50	1	10/29/10 18:14	11/01/10 15:02	85-01-8	
Pyrene	0.60U	ug/L	10.0	0.60	1	10/29/10 18:14	11/01/10 15:02	129-00-0	
2-Fluorobiphenyl (S)	75 %		43.9-113		1	10/29/10 18:14	11/01/10 15:02	321-60-8	
Terphenyl-d14 (S)	76 %		24.8-144		1	10/29/10 18:14	11/01/10 15:02	1718-51-0	
8260 MSV		Analytical Method: EPA 8260							
Acetone	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	67-64-1	
Acetonitrile	135	ug/L	50.0	25.0	5		10/29/10 22:13	75-05-8	
Acrolein	50.0U	ug/L	100	50.0	5		10/29/10 22:13	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 214 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-5 **Lab ID: 3519325036** Collected: 10/27/10 13:00 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	107-13-1	
Allyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	107-05-1	
Benzene	9.6	ug/L	5.0	2.5	5		10/29/10 22:13	71-43-2	
Bromochloromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	74-97-5	
Bromodichloromethane	1.4U	ug/L	3.0	1.4	5		10/29/10 22:13	75-27-4	
Bromoform	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-25-2	
Bromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	74-83-9	
2-Butanone (MEK)	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	78-93-3	
Carbon disulfide	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-15-0	
Carbon tetrachloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	56-23-5	
Chlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	108-90-7	
Chloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-00-3	
Chloroform	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	67-66-3	
Chloromethane	3.1U	ug/L	5.0	3.1	5		10/29/10 22:13	74-87-3	
Chloroprene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	126-99-8	
Dibromochloromethane	1.3U	ug/L	2.5	1.3	5		10/29/10 22:13	124-48-1	
Dibromomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	74-95-3	
trans-1,4-Dichloro-2-butene	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	110-57-6	
Dichlorodifluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-71-8	
1,1-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-34-3	
1,2-Dichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	107-06-2	
1,1-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-35-4	
cis-1,2-Dichloroethene	3.5 I	ug/L	5.0	2.5	5		10/29/10 22:13	156-59-2	
trans-1,2-Dichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	156-60-5	
1,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	78-87-5	
1,3-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	142-28-9	
2,2-Dichloropropane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	594-20-7	
1,1-Dichloropropene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	563-58-6	
cis-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 22:13	10061-01-5	
trans-1,3-Dichloropropene	1.2U	ug/L	2.5	1.2	5		10/29/10 22:13	10061-02-6	
Ethylbenzene	50.3	ug/L	5.0	2.5	5		10/29/10 22:13	100-41-4	
Ethyl methacrylate	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	97-63-2	
Hexachloro-1,3-butadiene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	87-68-3	
2-Hexanone	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	591-78-6	
Iodomethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	74-88-4	
Isobutyl Alcohol	50.0U	ug/L	100	50.0	5		10/29/10 22:13	78-83-1	
Methacrylonitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	126-98-7	
Methylene Chloride	12.5U	ug/L	25.0	12.5	5		10/29/10 22:13	75-09-2	
Methyl methacrylate	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	80-62-6	
4-Methyl-2-pentanone (MIBK)	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	108-10-1	
Propionitrile	25.0U	ug/L	50.0	25.0	5		10/29/10 22:13	107-12-0	
Styrene	3.3 I	ug/L	5.0	2.5	5		10/29/10 22:13	100-42-5	
1,1,1,2-Tetrachloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	630-20-6	
1,1,2,2-Tetrachloroethane	0.90U	ug/L	2.5	0.90	5		10/29/10 22:13	79-34-5	
Tetrachloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	127-18-4	
Toluene	35.6	ug/L	5.0	2.5	5		10/29/10 22:13	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 215 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: C-5		Lab ID: 3519325036		Collected: 10/27/10 13:00		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	120-82-1	
1,1,1-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	71-55-6	
1,1,2-Trichloroethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	79-00-5	
Trichloroethene	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	79-01-6	
Trichlorofluoromethane	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-69-4	
1,2,3-Trichloropropane	1.8U	ug/L	2.5	1.8	5		10/29/10 22:13	96-18-4	
Vinyl acetate	5.0U	ug/L	10.0	5.0	5		10/29/10 22:13	108-05-4	
Vinyl chloride	2.5U	ug/L	5.0	2.5	5		10/29/10 22:13	75-01-4	
Xylene (Total)	85.7	ug/L	5.0	2.5	5		10/29/10 22:13	1330-20-7	
4-Bromofluorobenzene (S)	97	%	70-114		5		10/29/10 22:13	460-00-4	D3
Dibromofluoromethane (S)	100	%	88-117		5		10/29/10 22:13	1868-53-7	
1,2-Dichloroethane-d4 (S)	106	%	86-125		5		10/29/10 22:13	17060-07-0	
Toluene-d8 (S)	100	%	87-113		5		10/29/10 22:13	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	4900	mg/L	100	100	1		11/03/10 12:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	6.5	mg/L	5.0	5.0	1		11/02/10 09:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	936	mg/L	500	250	100		11/02/10 01:06	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.010U	mg/L	0.020	0.010	1	11/01/10 11:30	11/08/10 15:05	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	581	mg/L	5.0	2.0	100		11/15/10 12:22	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: P2-1		Lab ID: 3519325037		Collected: 10/27/10 13:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Field Data									
		Analytical Method:							
Field pH	6.59	Std. Units			1		10/27/10 13:30		
Field Temperature	29.31	deg C			1		10/27/10 13:30		
Field Specific Conductance	780	umhos/cm			1		10/27/10 13:30		
Oxygen, Dissolved	2.34	mg/L			1		10/27/10 13:30	7782-44-7	
Turbidity	5.38	NTU			1		10/27/10 13:30		
8011 GCS EDB and DBCP									
		Analytical Method: EPA 8011 Preparation Method: EPA 8011							
1,2-Dibromo-3-chloropropane	0.0050U	ug/L	0.020	0.0050	1	11/04/10 15:15	11/04/10 22:20	96-12-8	
1,2-Dibromoethane (EDB)	0.0063U	ug/L	0.010	0.0063	1	11/04/10 15:15	11/04/10 22:20	106-93-4	
8081 GCS Pesticides									
		Analytical Method: EPA 8081 Preparation Method: EPA 3510							
Aldrin	0.00049U	ug/L	0.0098	0.00049	1	11/01/10 17:10	11/19/10 04:07	309-00-2	
alpha-BHC	0.00029U	ug/L	0.0098	0.00029	1	11/01/10 17:10	11/19/10 04:07	319-84-6	
beta-BHC	0.00049U	ug/L	0.0098	0.00049	1	11/01/10 17:10	11/19/10 04:07	319-85-7	
delta-BHC	0.00039U	ug/L	0.0098	0.00039	1	11/01/10 17:10	11/19/10 04:07	319-86-8	
gamma-BHC (Lindane)	0.00020U	ug/L	0.0098	0.00020	1	11/01/10 17:10	11/19/10 04:07	58-89-9	
Chlordane (Technical)	0.079U	ug/L	0.49	0.079	1	11/01/10 17:10	11/19/10 04:07	57-74-9	
Chlorobenzilate	0.021U	ug/L	0.098	0.021	1	11/01/10 17:10	11/19/10 04:07	510-15-6	
4,4'-DDD	0.0019U	ug/L	0.0098	0.0019	1	11/01/10 17:10	11/19/10 04:07	72-54-8	
4,4'-DDE	0.00088U	ug/L	0.0098	0.00088	1	11/01/10 17:10	11/19/10 04:07	72-55-9	
4,4'-DDT	0.0035U	ug/L	0.0098	0.0035	1	11/01/10 17:10	11/19/10 04:07	50-29-3	
Dieldrin	0.00049U	ug/L	0.0098	0.00049	1	11/01/10 17:10	11/19/10 04:07	60-57-1	
Endosulfan I	0.00069U	ug/L	0.0098	0.00069	1	11/01/10 17:10	11/19/10 04:07	959-98-8	
Endosulfan II	0.00069U	ug/L	0.0098	0.00069	1	11/01/10 17:10	11/19/10 04:07	33213-65-9	
Endosulfan sulfate	0.00059U	ug/L	0.0098	0.00059	1	11/01/10 17:10	11/19/10 04:07	1031-07-8	
Endrin	0.0017U	ug/L	0.0098	0.0017	1	11/01/10 17:10	11/19/10 04:07	72-20-8	
Endrin aldehyde	0.0070U	ug/L	0.0098	0.0070	1	11/01/10 17:10	11/19/10 04:07	7421-93-4	
Heptachlor	0.0015U	ug/L	0.0098	0.0015	1	11/01/10 17:10	11/19/10 04:07	76-44-8	
Heptachlor epoxide	0.00039U	ug/L	0.0098	0.00039	1	11/01/10 17:10	11/19/10 04:07	1024-57-3	
Methoxychlor	0.0069U	ug/L	0.0098	0.0069	1	11/01/10 17:10	11/19/10 04:07	72-43-5	
Pentachloronitrobenzene	0.015U	ug/L	0.098	0.015	1	11/01/10 17:10	11/19/10 04:07	82-68-8	
Toxaphene	0.28U	ug/L	0.49	0.28	1	11/01/10 17:10	11/19/10 04:07	8001-35-2	
Tetrachloro-m-xylene (S)	104	%	66.5-120.3		1	11/01/10 17:10	11/19/10 04:07	877-09-8	
Decachlorobiphenyl (S)	87	%	41.7-109.1		1	11/01/10 17:10	11/19/10 04:07	2051-24-3	
8082 GCS PCB									
		Analytical Method: EPA 8082 Preparation Method: EPA 3510							
PCB-1016 (Aroclor 1016)	0.078U	ug/L	0.49	0.078	1	11/01/10 17:10	11/19/10 04:07	12674-11-2	
PCB-1221 (Aroclor 1221)	0.079U	ug/L	0.49	0.079	1	11/01/10 17:10	11/19/10 04:07	11104-28-2	
PCB-1232 (Aroclor 1232)	0.12U	ug/L	0.49	0.12	1	11/01/10 17:10	11/19/10 04:07	11141-16-5	
PCB-1242 (Aroclor 1242)	0.12U	ug/L	0.49	0.12	1	11/01/10 17:10	11/19/10 04:07	53469-21-9	
PCB-1248 (Aroclor 1248)	0.27U	ug/L	0.49	0.27	1	11/01/10 17:10	11/19/10 04:07	12672-29-6	
PCB-1254 (Aroclor 1254)	0.14U	ug/L	0.49	0.14	1	11/01/10 17:10	11/19/10 04:07	11097-69-1	
PCB-1260 (Aroclor 1260)	0.11U	ug/L	0.49	0.11	1	11/01/10 17:10	11/19/10 04:07	11096-82-5	
Tetrachloro-m-xylene (S)	140	%	48-111		1	11/01/10 17:10	11/19/10 04:07	877-09-8	S3

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 217 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: P2-1		Lab ID: 3519325037		Collected: 10/27/10 13:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082 GCS PCB									
Analytical Method: EPA 8082 Preparation Method: EPA 3510									
Decachlorobiphenyl (S)	93 %		63-121		1	11/01/10 17:10	11/19/10 04:07	2051-24-3	
8151 Chlorinated Herbicides									
Analytical Method: EPA 8151 Preparation Method: EPA 8151									
2,4-D	0.22U	ug/L	0.91	0.22	1	10/29/10 18:00	11/02/10 21:03	94-75-7	
Dinoseb	0.055U	ug/L	0.18	0.055	1	10/29/10 18:00	11/02/10 21:03	88-85-7	
Pentachlorophenol	0.017U	ug/L	0.028	0.017	1	10/29/10 18:00	11/02/10 21:03	87-86-5	
2,4,5-T	0.041U	ug/L	0.18	0.041	1	10/29/10 18:00	11/02/10 21:03	93-76-5	
2,4,5-TP (Silvex)	0.048U	ug/L	0.18	0.048	1	10/29/10 18:00	11/02/10 21:03	93-72-1	
2,4-DCPA (S)	89 %		65.5-125.7		1	10/29/10 18:00	11/02/10 21:03	19719-28-9	
Analytical Method: EPA 8141 Preparation Method: EPA 3510									
Dimethoate	0.19U	ug/L	0.50	0.19	1	11/03/10 13:00	11/09/10 23:26	60-51-5	
Disulfoton	0.16U	ug/L	0.50	0.16	1	11/03/10 13:00	11/09/10 23:26	298-04-4	
Famphur	0.14U	ug/L	0.50	0.14	1	11/03/10 13:00	11/09/10 23:26	52-85-7	
Methyl parathion	0.19U	ug/L	0.50	0.19	1	11/03/10 13:00	11/09/10 23:26	298-00-0	
Parathion (Ethyl parathion)	0.35U	ug/L	1.0	0.35	1	11/03/10 13:00	11/09/10 23:26	56-38-2	
Phorate	0.37U	ug/L	1.0	0.37	1	11/03/10 13:00	11/09/10 23:26	298-02-2	
4-Chloro3nitrobenzotrifluoride	88 %		34.2-122		1	11/03/10 13:00	11/09/10 23:26		
6010 MET ICP									
Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	5.2 I	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:20	7440-38-2	
Barium	55.6	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:20	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:20	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:20	7440-43-9	
Chromium	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:20	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:20	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:20	7440-50-8	
Iron	3520	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:20	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:20	7439-92-1	
Nickel	2.7 I	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:20	7440-02-0	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:20	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:20	7440-22-4	
Sodium	13.2	mg/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:20	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:20	7440-31-5	
Vanadium	7.4 I	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:20	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:20	7440-66-6	
6020 MET ICPMS									
Analytical Method: EPA 6020 Preparation Method: EPA 3010									
Antimony	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/19/10 17:04	7440-36-0	
Thallium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/19/10 17:04	7440-28-0	
7470 Mercury									
Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.10U	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 12:02	7439-97-6	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: P2-1 **Lab ID: 3519325037** Collected: 10/27/10 13:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Acenaphthene	0.84U	ug/L	4.9	0.84	1	11/01/10 18:00	11/11/10 04:57	83-32-9	
Acenaphthylene	0.92U	ug/L	4.9	0.92	1	11/01/10 18:00	11/11/10 04:57	208-96-8	
Acetophenone	1.4U	ug/L	4.9	1.4	1	11/01/10 18:00	11/11/10 04:57	98-86-2	
2-Acetylaminofluorene	0.63U	ug/L	4.9	0.63	1	11/01/10 18:00	11/11/10 04:57	53-96-3	
4-Aminobiphenyl	2.8U	ug/L	4.9	2.8	1	11/01/10 18:00	11/11/10 04:57	92-67-1	
Anthracene	0.58U	ug/L	4.9	0.58	1	11/01/10 18:00	11/11/10 04:57	120-12-7	
Benzo(a)anthracene	0.61U	ug/L	4.9	0.61	1	11/01/10 18:00	11/11/10 04:57	56-55-3	
Benzo(a)pyrene	0.56U	ug/L	0.97	0.56	1	11/01/10 18:00	11/11/10 04:57	50-32-8	
Benzo(b)fluoranthene	0.60U	ug/L	1.9	0.60	1	11/01/10 18:00	11/11/10 04:57	205-99-2	
Benzo(g,h,i)perylene	0.66U	ug/L	4.9	0.66	1	11/01/10 18:00	11/11/10 04:57	191-24-2	
Benzo(k)fluoranthene	0.50U	ug/L	3.9	0.50	1	11/01/10 18:00	11/11/10 04:57	207-08-9	
Benzyl alcohol	0.99U	ug/L	4.9	0.99	1	11/01/10 18:00	11/11/10 04:57	100-51-6	
4-Bromophenylphenyl ether	0.65U	ug/L	4.9	0.65	1	11/01/10 18:00	11/11/10 04:57	101-55-3	
Butylbenzylphthalate	0.70U	ug/L	4.9	0.70	1	11/01/10 18:00	11/11/10 04:57	85-68-7	
4-Chloro-3-methylphenol	0.60U	ug/L	19.5	0.60	1	11/01/10 18:00	11/11/10 04:57	59-50-7	
4-Chloroaniline	1.2U	ug/L	4.9	1.2	1	11/01/10 18:00	11/11/10 04:57	106-47-8	
bis(2-Chloroethoxy)methane	2.9U	ug/L	4.9	2.9	1	11/01/10 18:00	11/11/10 04:57	111-91-1	
bis(2-Chloroethyl) ether	0.73U	ug/L	3.9	0.73	1	11/01/10 18:00	11/11/10 04:57	111-44-4	
bis(2-Chloroisopropyl) ether	0.71U	ug/L	4.9	0.71	1	11/01/10 18:00	11/11/10 04:57	108-60-1	
2-Chloronaphthalene	0.78U	ug/L	4.9	0.78	1	11/01/10 18:00	11/11/10 04:57	91-58-7	
2-Chlorophenol	0.66U	ug/L	4.9	0.66	1	11/01/10 18:00	11/11/10 04:57	95-57-8	
4-Chlorophenylphenyl ether	0.61U	ug/L	4.9	0.61	1	11/01/10 18:00	11/11/10 04:57	7005-72-3	
Chrysene	0.36U	ug/L	4.9	0.36	1	11/01/10 18:00	11/11/10 04:57	218-01-9	
Diallylate	0.71U	ug/L	4.9	0.71	1	11/01/10 18:00	11/11/10 04:57	2303-16-4	
Dibenz(a,h)anthracene	0.63U	ug/L	1.9	0.63	1	11/01/10 18:00	11/11/10 04:57	53-70-3	
Dibenzofuran	0.65U	ug/L	4.9	0.65	1	11/01/10 18:00	11/11/10 04:57	132-64-9	
1,2-Dichlorobenzene	0.66U	ug/L	4.9	0.66	1	11/01/10 18:00	11/11/10 04:57	95-50-1	
1,3-Dichlorobenzene	0.74U	ug/L	4.9	0.74	1	11/01/10 18:00	11/11/10 04:57	541-73-1	
1,4-Dichlorobenzene	1.0 I	ug/L	4.9	0.75	1	11/01/10 18:00	11/11/10 04:57	106-46-7	
3,3'-Dichlorobenzidine	0.67U	ug/L	9.7	0.67	1	11/01/10 18:00	11/11/10 04:57	91-94-1	
2,4-Dichlorophenol	0.54U	ug/L	1.9	0.54	1	11/01/10 18:00	11/11/10 04:57	120-83-2	
2,6-Dichlorophenol	0.60U	ug/L	3.9	0.60	1	11/01/10 18:00	11/11/10 04:57	87-65-0	
Diethylphthalate	0.50U	ug/L	4.9	0.50	1	11/01/10 18:00	11/11/10 04:57	84-66-2	
P-Dimethylaminoazobenzene	0.65U	ug/L	4.9	0.65	1	11/01/10 18:00	11/11/10 04:57	60-11-7	
7,12-Dimethylbenz(a)anthracene	1.9U	ug/L	4.9	1.9	1	11/01/10 18:00	11/11/10 04:57	57-97-6	
3,3'-Dimethylbenzidine	3.0U	ug/L	9.7	3.0	1	11/01/10 18:00	11/11/10 04:57	119-93-7	
2,4-Dimethylphenol	1.5U	ug/L	4.9	1.5	1	11/01/10 18:00	11/11/10 04:57	105-67-9	
a,a-Dimethylphenylethylamine	9.7U	ug/L	19.5	9.7	1	11/01/10 18:00	11/11/10 04:57	122-09-8	
Dimethylphthalate	0.62U	ug/L	4.9	0.62	1	11/01/10 18:00	11/11/10 04:57	131-11-3	
Di-n-butylphthalate	0.40U	ug/L	4.9	0.40	1	11/01/10 18:00	11/11/10 04:57	84-74-2	
4,6-Dinitro-2-methylphenol	1.3U	ug/L	19.5	1.3	1	11/01/10 18:00	11/11/10 04:57	534-52-1	
1,2-Dinitrobenzene	1.1U	ug/L	4.9	1.1	1	11/01/10 18:00	11/11/10 04:57	528-29-0	
1,3-Dinitrobenzene	0.66U	ug/L	7.8	0.66	1	11/01/10 18:00	11/11/10 04:57	99-65-0	
2,4-Dinitrophenol	1.5U	ug/L	19.5	1.5	1	11/01/10 18:00	11/11/10 04:57	51-28-5	
2,4-Dinitrotoluene	0.52U	ug/L	1.9	0.52	1	11/01/10 18:00	11/11/10 04:57	121-14-2	
2,6-Dinitrotoluene	1.2U	ug/L	1.9	1.2	1	11/01/10 18:00	11/11/10 04:57	606-20-2	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 219 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: P2-1 **Lab ID: 3519325037** Collected: 10/27/10 13:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8270 MSSV SemiVOA App. II Analytical Method: EPA 8270 Preparation Method: EPA 3510									
Di-n-octylphthalate	0.88U	ug/L	4.9	0.88	1	11/01/10 18:00	11/11/10 04:57	117-84-0	
bis(2-Ethylhexyl)phthalate	0.78U	ug/L	4.9	0.78	1	11/01/10 18:00	11/11/10 04:57	117-81-7	
Ethyl methanesulfonate	0.88U	ug/L	4.9	0.88	1	11/01/10 18:00	11/11/10 04:57	62-50-0	
Fluoranthene	0.53U	ug/L	4.9	0.53	1	11/01/10 18:00	11/11/10 04:57	206-44-0	
Fluorene	0.54U	ug/L	4.9	0.54	1	11/01/10 18:00	11/11/10 04:57	86-73-7	
Hexachlorobenzene	0.78U	ug/L	0.97	0.78	1	11/01/10 18:00	11/11/10 04:57	118-74-1	
Hexachlorocyclopentadiene	1.2U	ug/L	4.9	1.2	1	11/01/10 18:00	11/11/10 04:57	77-47-4	
Hexachloroethane	0.69U	ug/L	4.9	0.69	1	11/01/10 18:00	11/11/10 04:57	67-72-1	
Hexachloropropene	1.4U	ug/L	4.9	1.4	1	11/01/10 18:00	11/11/10 04:57	1888-71-7	
Indeno(1,2,3-cd)pyrene	0.71U	ug/L	1.9	0.71	1	11/01/10 18:00	11/11/10 04:57	193-39-5	
Isodrin	0.53U	ug/L	4.9	0.53	1	11/01/10 18:00	11/11/10 04:57	465-73-6	
Isophorone	0.71U	ug/L	4.9	0.71	1	11/01/10 18:00	11/11/10 04:57	78-59-1	
Isosafrole	0.58U	ug/L	4.9	0.58	1	11/01/10 18:00	11/11/10 04:57	120-58-1	
Kepone	9.7U	ug/L	19.5	9.7	1	11/01/10 18:00	11/11/10 04:57	143-50-0	
Methapyrilene	1.6U	ug/L	4.9	1.6	1	11/01/10 18:00	11/11/10 04:57	91-80-5	
3-Methylcholanthrene	1.0U	ug/L	4.9	1.0	1	11/01/10 18:00	11/11/10 04:57	56-49-5	
Methyl methanesulfonate	0.97U	ug/L	4.9	0.97	1	11/01/10 18:00	11/11/10 04:57	66-27-3	
1-Methylnaphthalene	0.97U	ug/L	4.9	0.97	1	11/01/10 18:00	11/11/10 04:57	90-12-0	
2-Methylnaphthalene	0.96U	ug/L	4.9	0.96	1	11/01/10 18:00	11/11/10 04:57	91-57-6	
2-Methylphenol(o-Cresol)	0.71U	ug/L	4.9	0.71	1	11/01/10 18:00	11/11/10 04:57	95-48-7	
3&4-Methylphenol(m&p Cresol)	0.64U	ug/L	9.7	0.64	1	11/01/10 18:00	11/11/10 04:57		
2-Naphthylamine	2.2U	ug/L	4.9	2.2	1	11/01/10 18:00	11/11/10 04:57	91-59-8	
Naphthalene	0.76U	ug/L	4.9	0.76	1	11/01/10 18:00	11/11/10 04:57	91-20-3	
1-Naphthylamine	1.0U	ug/L	4.9	1.0	1	11/01/10 18:00	11/11/10 04:57	134-32-7	
1,4-Naphthoquinone	1.1U	ug/L	4.9	1.1	1	11/01/10 18:00	11/11/10 04:57	130-15-4	
2-Nitroaniline	0.58U	ug/L	4.9	0.58	1	11/01/10 18:00	11/11/10 04:57	88-74-4	
3-Nitroaniline	0.96U	ug/L	4.9	0.96	1	11/01/10 18:00	11/11/10 04:57	99-09-2	
4-Nitroaniline	0.67U	ug/L	3.9	0.67	1	11/01/10 18:00	11/11/10 04:57	100-01-6	
Nitrobenzene	1.1U	ug/L	3.9	1.1	1	11/01/10 18:00	11/11/10 04:57	98-95-3	
2-Nitrophenol	0.79U	ug/L	4.9	0.79	1	11/01/10 18:00	11/11/10 04:57	88-75-5	
4-Nitrophenol	1.1U	ug/L	19.5	1.1	1	11/01/10 18:00	11/11/10 04:57	100-02-7	
5-Nitro-o-toluidine	1.3U	ug/L	4.9	1.3	1	11/01/10 18:00	11/11/10 04:57	99-55-8	
N-Nitrosodiethylamine	0.71U	ug/L	3.9	0.71	1	11/01/10 18:00	11/11/10 04:57	55-18-5	
N-Nitrosodimethylamine	0.94U	ug/L	1.9	0.94	1	11/01/10 18:00	11/11/10 04:57	62-75-9	
N-Nitroso-di-n-butylamine	0.54U	ug/L	3.9	0.54	1	11/01/10 18:00	11/11/10 04:57	924-16-3	
N-Nitroso-di-n-propylamine	0.91U	ug/L	3.9	0.91	1	11/01/10 18:00	11/11/10 04:57	621-64-7	
N-Nitrosodiphenylamine	0.49U	ug/L	4.9	0.49	1	11/01/10 18:00	11/11/10 04:57	86-30-6	
N-Nitrosomethylethylamine	0.72U	ug/L	4.9	0.72	1	11/01/10 18:00	11/11/10 04:57	10595-95-6	
N-Nitrosopiperidine	0.62U	ug/L	4.9	0.62	1	11/01/10 18:00	11/11/10 04:57	100-75-4	
N-Nitrosopyrrolidine	0.86U	ug/L	4.9	0.86	1	11/01/10 18:00	11/11/10 04:57	930-55-2	
O,O,O-Triethylphosphorothioate	0.67U	ug/L	4.9	0.67	1	11/01/10 18:00	11/11/10 04:57	126-68-1	
Parathion (Ethyl parathion)	1.1U	ug/L	4.9	1.1	1	11/01/10 18:00	11/11/10 04:57	56-38-2	
Pentachlorobenzene	0.76U	ug/L	4.9	0.76	1	11/01/10 18:00	11/11/10 04:57	608-93-5	
Pentachlorophenol	0.64U	ug/L	19.5	0.64	1	11/01/10 18:00	11/11/10 04:57	87-86-5	
Phenacetin	0.52U	ug/L	4.9	0.52	1	11/01/10 18:00	11/11/10 04:57	62-44-2	
Phenanthrene	0.51U	ug/L	4.9	0.51	1	11/01/10 18:00	11/11/10 04:57	85-01-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 220 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: P2-1 **Lab ID: 3519325037** Collected: 10/27/10 13:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
------------	---------	-------	-----	-----	----	----------	----------	---------	------

8270 MSSV SemiVOA App. II

Analytical Method: EPA 8270 Preparation Method: EPA 3510

Phenol	0.53U	ug/L	4.9	0.53	1	11/01/10 18:00	11/11/10 04:57	108-95-2	
p-Phenylenediamine	9.7U	ug/L	19.5	9.7	1	11/01/10 18:00	11/11/10 04:57	106-50-3	
Pronamide	1.1U	ug/L	4.9	1.1	1	11/01/10 18:00	11/11/10 04:57	23950-58-5	
Pyrene	0.66U	ug/L	4.9	0.66	1	11/01/10 18:00	11/11/10 04:57	129-00-0	
Safrole	0.83U	ug/L	4.9	0.83	1	11/01/10 18:00	11/11/10 04:57	94-59-7	
1,2,4,5-Tetrachlorobenzene	0.68U	ug/L	4.9	0.68	1	11/01/10 18:00	11/11/10 04:57	95-94-3	
2,3,4,6-Tetrachlorophenol	3.7U	ug/L	4.9	3.7	1	11/01/10 18:00	11/11/10 04:57	58-90-2	
Thionazin	0.59U	ug/L	4.9	0.59	1	11/01/10 18:00	11/11/10 04:57	297-97-2	
O-Toluidine	1.0U	ug/L	4.9	1.0	1	11/01/10 18:00	11/11/10 04:57	95-53-4	
1,2,4-Trichlorobenzene	0.81U	ug/L	4.9	0.81	1	11/01/10 18:00	11/11/10 04:57	120-82-1	
2,4,5-Trichlorophenol	0.51U	ug/L	3.9	0.51	1	11/01/10 18:00	11/11/10 04:57	95-95-4	
2,4,6-Trichlorophenol	0.67U	ug/L	1.9	0.67	1	11/01/10 18:00	11/11/10 04:57	88-06-2	
1,3,5-Trinitrobenzene	1.2U	ug/L	4.9	1.2	1	11/01/10 18:00	11/11/10 04:57	99-35-4	
Nitrobenzene-d5 (S)	73	%	10-110		1	11/01/10 18:00	11/11/10 04:57	4165-60-0	
2-Fluorobiphenyl (S)	78	%	18-110		1	11/01/10 18:00	11/11/10 04:57	321-60-8	
Terphenyl-d14 (S)	86	%	10-123		1	11/01/10 18:00	11/11/10 04:57	1718-51-0	
Phenol-d6 (S)	32	%	10-110		1	11/01/10 18:00	11/11/10 04:57	13127-88-3	
2-Fluorophenol (S)	44	%	18-110		1	11/01/10 18:00	11/11/10 04:57	367-12-4	
2,4,6-Tribromophenol (S)	87	%	10-110		1	11/01/10 18:00	11/11/10 04:57	118-79-6	

8270 MSSV PAH by SCAN

Analytical Method: EPA 8270 by SCAN Preparation Method: EPA 3510

Acenaphthene	0.53 I	ug/L	1.0	0.031	1	10/29/10 18:14	11/01/10 15:24	83-32-9	
Acenaphthylene	0.051U	ug/L	2.1	0.051	1	10/29/10 18:14	11/01/10 15:24	208-96-8	
Anthracene	0.051U	ug/L	1.0	0.051	1	10/29/10 18:14	11/01/10 15:24	120-12-7	
Benzo(a)anthracene	0.062U	ug/L	0.21	0.062	1	10/29/10 18:14	11/01/10 15:24	56-55-3	
Benzo(a)pyrene	0.051U	ug/L	0.21	0.051	1	10/29/10 18:14	11/01/10 15:24	50-32-8	
Benzo(b)fluoranthene	0.051U	ug/L	0.10	0.051	1	10/29/10 18:14	11/01/10 15:24	205-99-2	
Benzo(g,h,i)perylene	0.062U	ug/L	1.0	0.062	1	10/29/10 18:14	11/01/10 15:24	191-24-2	
Benzo(k)fluoranthene	0.041U	ug/L	0.26	0.041	1	10/29/10 18:14	11/01/10 15:24	207-08-9	
Chrysene	0.062U	ug/L	1.0	0.062	1	10/29/10 18:14	11/01/10 15:24	218-01-9	
Dibenz(a,h)anthracene	0.64	ug/L	0.21	0.051	1	10/29/10 18:14	11/01/10 15:24	53-70-3	
Fluoranthene	0.062U	ug/L	1.0	0.062	1	10/29/10 18:14	11/01/10 15:24	206-44-0	
Fluorene	0.031U	ug/L	1.0	0.031	1	10/29/10 18:14	11/01/10 15:24	86-73-7	
Indeno(1,2,3-cd)pyrene	0.041U	ug/L	0.15	0.041	1	10/29/10 18:14	11/01/10 15:24	193-39-5	
1-Methylnaphthalene	0.093U	ug/L	1.5	0.093	1	10/29/10 18:14	11/01/10 15:24	90-12-0	
2-Methylnaphthalene	0.062U	ug/L	1.5	0.062	1	10/29/10 18:14	11/01/10 15:24	91-57-6	
Naphthalene	0.082U	ug/L	1.0	0.082	1	10/29/10 18:14	11/01/10 15:24	91-20-3	
Phenanthrene	0.051U	ug/L	1.0	0.051	1	10/29/10 18:14	11/01/10 15:24	85-01-8	
Pyrene	0.062U	ug/L	1.0	0.062	1	10/29/10 18:14	11/01/10 15:24	129-00-0	
2-Fluorobiphenyl (S)	82	%	43.9-113		1	10/29/10 18:14	11/01/10 15:24	321-60-8	
Terphenyl-d14 (S)	69	%	24.8-144		1	10/29/10 18:14	11/01/10 15:24	1718-51-0	

8260 MSV

Analytical Method: EPA 8260

Acetone	5.0U	ug/L	10.0	5.0	1	10/29/10 19:06	67-64-1	
Acetonitrile	6.8 I	ug/L	10.0	5.0	1	10/29/10 19:06	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1	10/29/10 19:06	107-02-8	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 221 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: P2-1 **Lab ID: 3519325037** Collected: 10/27/10 13:30 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 19:06	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 19:06	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 19:06	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-34-3	
1,2-Dichloroethane	0.65 I	ug/L	1.0	0.50	1		10/29/10 19:06	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 19:06	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 19:06	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 19:06	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 19:06	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 19:06	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	630-20-6	
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 19:06	79-34-5	
Tetrachloroethene	0.66 I	ug/L	1.0	0.50	1		10/29/10 19:06	127-18-4	
Toluene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	108-88-3	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 222 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

Sample: P2-1		Lab ID: 3519325037		Collected: 10/27/10 13:30		Received: 10/29/10 07:00		Matrix: Water	
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 19:06	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 19:06	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 19:06	75-01-4	
Xylene (Total)	0.74 I	ug/L	1.0	0.50	1		10/29/10 19:06	1330-20-7	
4-Bromofluorobenzene (S)	97 %		70-114		1		10/29/10 19:06	460-00-4	
Dibromofluoromethane (S)	100 %		88-117		1		10/29/10 19:06	1868-53-7	
1,2-Dichloroethane-d4 (S)	108 %		86-125		1		10/29/10 19:06	17060-07-0	
Toluene-d8 (S)	101 %		87-113		1		10/29/10 19:06	2037-26-5	
2540C Total Dissolved Solids		Analytical Method: SM 2540C							
Total Dissolved Solids	422	mg/L	5.0	5.0	1		11/03/10 12:00		
4500S2E Sulfide, Iodometric		Analytical Method: SM 4500-S2E							
Sulfide	1.0U	mg/L	1.0	1.0	1		11/02/10 09:00	18496-25-8	
300.0 IC Anions 28 Days		Analytical Method: EPA 300.0							
Chloride	21.2	mg/L	5.0	2.5	1		11/02/10 01:43	16887-00-6	
335.4 Cyanide, Total		Analytical Method: EPA 335.4 Preparation Method: EPA 335.4							
Cyanide	0.0050U	mg/L	0.010	0.0050	1	11/01/10 11:30	11/08/10 15:06	57-12-5	
350.1 Ammonia		Analytical Method: EPA 350.1							
Nitrogen, Ammonia	2.6	mg/L	0.050	0.020	1		11/01/10 09:56	7664-41-7	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank APPII (10/27/10) **Lab ID:** 3519325038 **Collected:** 10/27/10 08:00 **Received:** 10/29/10 07:00 **Matrix:** Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
Acetone	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	67-64-1	
Acetonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	75-05-8	
Acrolein	10.0U	ug/L	20.0	10.0	1		10/29/10 15:41	107-02-8	
Acrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	107-13-1	
Allyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	107-05-1	
Benzene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	71-43-2	
Bromochloromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	74-97-5	
Bromodichloromethane	0.27U	ug/L	0.60	0.27	1		10/29/10 15:41	75-27-4	
Bromoform	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-25-2	
Bromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	74-83-9	
2-Butanone (MEK)	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	78-93-3	
Carbon disulfide	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-15-0	
Carbon tetrachloride	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	56-23-5	
Chlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	108-90-7	
Chloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-00-3	
Chloroform	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	67-66-3	
Chloromethane	0.62U	ug/L	1.0	0.62	1		10/29/10 15:41	74-87-3	
Chloroprene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	126-99-8	
Dibromochloromethane	0.26U	ug/L	0.50	0.26	1		10/29/10 15:41	124-48-1	
Dibromomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	74-95-3	
trans-1,4-Dichloro-2-butene	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	110-57-6	
Dichlorodifluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-71-8	
1,1-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-34-3	
1,2-Dichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	107-06-2	
1,1-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-35-4	
cis-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	156-59-2	
trans-1,2-Dichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	156-60-5	
1,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	78-87-5	
1,3-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	142-28-9	
2,2-Dichloropropane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	594-20-7	
1,1-Dichloropropene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	563-58-6	
cis-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 15:41	10061-01-5	
trans-1,3-Dichloropropene	0.25U	ug/L	0.50	0.25	1		10/29/10 15:41	10061-02-6	
Ethylbenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	100-41-4	
Ethyl methacrylate	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	97-63-2	
Hexachloro-1,3-butadiene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	87-68-3	
2-Hexanone	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	591-78-6	
Iodomethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	74-88-4	
Isobutyl Alcohol	10.0U	ug/L	20.0	10.0	1		10/29/10 15:41	78-83-1	
Methacrylonitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	126-98-7	
Methylene Chloride	2.5U	ug/L	5.0	2.5	1		10/29/10 15:41	75-09-2	
Methyl methacrylate	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	80-62-6	
4-Methyl-2-pentanone (MIBK)	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	108-10-1	
Propionitrile	5.0U	ug/L	10.0	5.0	1		10/29/10 15:41	107-12-0	
Styrene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	100-42-5	
1,1,1,2-Tetrachloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	630-20-6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 224 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Trip Blank APP11 (10/27/10) **Lab ID:** 3519325038 Collected: 10/27/10 08:00 Received: 10/29/10 07:00 Matrix: Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8260 MSV		Analytical Method: EPA 8260							
1,1,2,2-Tetrachloroethane	0.18U	ug/L	0.50	0.18	1		10/29/10 15:41	79-34-5	
Tetrachloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	127-18-4	
Toluene	1.2	ug/L	1.0	0.50	1		10/29/10 15:41	108-88-3	
1,2,4-Trichlorobenzene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	120-82-1	
1,1,1-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	71-55-6	
1,1,2-Trichloroethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	79-00-5	
Trichloroethene	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	79-01-6	
Trichlorofluoromethane	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-69-4	
1,2,3-Trichloropropane	0.36U	ug/L	0.50	0.36	1		10/29/10 15:41	96-18-4	
Vinyl acetate	1.0U	ug/L	2.0	1.0	1		10/29/10 15:41	108-05-4	
Vinyl chloride	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	75-01-4	
Xylene (Total)	0.50U	ug/L	1.0	0.50	1		10/29/10 15:41	1330-20-7	
4-Bromofluorobenzene (S)	99 %		70-114		1		10/29/10 15:41	460-00-4	
Dibromofluoromethane (S)	100 %		88-117		1		10/29/10 15:41	1868-53-7	
1,2-Dichloroethane-d4 (S)	106 %		86-125		1		10/29/10 15:41	17060-07-0	
Toluene-d8 (S)	99 %		87-113		1		10/29/10 15:41	2037-26-5	

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Gas Condensate Lab ID: 3519325039 Collected: 10/27/10 10:45 Received: 10/29/10 07:00 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
Analytical Method:									
Field pH	7.23	Std. Units			1		10/27/10 10:45		
Field Temperature	31.21	deg C			1		10/27/10 10:45		
Field Specific Conductance	7304	umhos/cm			1		10/27/10 10:45		
Oxygen, Dissolved	3.64	mg/L			1		10/27/10 10:45	7782-44-7	
Turbidity	2.39	NTU			1		10/27/10 10:45		
6010 MET ICP Analytical Method: EPA 6010 Preparation Method: EPA 3010									
Arsenic	426	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:23	7440-38-2	
Barium	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:23	7440-39-3	
Beryllium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:23	7440-41-7	
Cadmium	0.50U	ug/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:23	7440-43-9	
Calcium	0.63	mg/L	0.50	0.25	1	11/05/10 11:40	11/10/10 11:23	7440-70-2	
Chromium	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:23	7440-47-3	
Cobalt	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:23	7440-48-4	
Copper	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:23	7440-50-8	
Iron	81.8	ug/L	40.0	20.0	1	11/05/10 11:40	11/10/10 11:23	7439-89-6	
Lead	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:23	7439-92-1	
Magnesium	0.25U	mg/L	0.50	0.25	1	11/05/10 11:40	11/10/10 11:23	7439-95-4	
Manganese	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:23	7439-96-5	
Nickel	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:23	7440-02-0	
Potassium	0.50U	mg/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:23	7440-09-7	
Selenium	7.5U	ug/L	15.0	7.5	1	11/05/10 11:40	11/10/10 11:23	7782-49-2	
Silver	2.5U	ug/L	5.0	2.5	1	11/05/10 11:40	11/10/10 11:23	7440-22-4	
Sodium	0.50U	mg/L	1.0	0.50	1	11/05/10 11:40	11/10/10 11:23	7440-23-5	
Tin	25.0U	ug/L	50.0	25.0	1	11/05/10 11:40	11/10/10 11:23	7440-31-5	
Vanadium	5.0U	ug/L	10.0	5.0	1	11/05/10 11:40	11/10/10 11:23	7440-62-2	
Zinc	10.0U	ug/L	20.0	10.0	1	11/05/10 11:40	11/10/10 11:23	7440-66-6	
7470 Mercury Analytical Method: EPA 7470 Preparation Method: EPA 7470									
Mercury	0.14 I	ug/L	0.20	0.10	1	11/05/10 10:40	11/08/10 12:05	7439-97-6	
2320B Alkalinity Analytical Method: SM 2320B									
Alkalinity, Bicarbonate (CaCO ₃)	3410	mg/L	25.0	25.0	5		11/04/10 10:20		
Alkalinity, Carbonate (CaCO ₃)	25.0U	mg/L	25.0	25.0	5		11/04/10 10:20		
Alkalinity, Total as CaCO ₃	3410	mg/L	25.0	25.0	5		11/04/10 10:20		
2540C Total Dissolved Solids Analytical Method: SM 2540C									
Total Dissolved Solids	80.0	mg/L	50.0	50.0	1		11/03/10 12:01		
300.0 IC Anions 28 Days Analytical Method: EPA 300.0									
Chloride	125U	mg/L	250	125	50		11/02/10 01:55	16887-00-6	
Sulfate	125U	mg/L	250	125	50		11/02/10 01:55	14808-79-8	
350.1 Ammonia Analytical Method: EPA 350.1									
Nitrogen, Ammonia	1010	mg/L	5.0	2.0	100		11/15/10 08:52	7664-41-7	M6

ANALYTICAL RESULTS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

Sample: Gas Condensate		Lab ID: 3519325039	Collected: 10/27/10 10:45	Received: 10/29/10 07:00	Matrix: Water				
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
410.4 COD		Analytical Method: EPA 410.4							
Chemical Oxygen Demand	955	mg/L	25.0	12.5	1		11/02/10 14:24		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: CVFS/2230 Analysis Method: EPA 1631E
QC Batch Method: EPA 1631E Analysis Description: 1631E Mercury
Associated Lab Samples: 3519325001, 3519325002, 3519325003

METHOD BLANK: 371415 Matrix: Water

Associated Lab Samples: 3519325001, 3519325002, 3519325003

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	0.000363 I	0.00050	10/18/10 07:27	

METHOD BLANK: 371416 Matrix: Water

Associated Lab Samples: 3519325001, 3519325002, 3519325003

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	ND	0.00050	10/18/10 09:07	

METHOD BLANK: 371417 Matrix: Water

Associated Lab Samples: 3519325001, 3519325002, 3519325003

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	ND	0.00050	10/18/10 10:33	

LABORATORY CONTROL SAMPLE & LCSD: 371418 371419

Parameter	Units	Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec	% Rec Limits	RPD	Max RPD	Qualifiers
Mercury	ug/L	.005	0.00494	0.00553	99	111	79-121	11	24	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 371420 371421

Parameter	Units	3519325002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.00303	.01	.01	0.0133	0.0121	102	91	75-125	9	24	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 371422 371423

Parameter	Units	4037891002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	4.07 ng/L	.02	.02	0.0232	0.0221	96	90	75-125	5	24	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3368

Analysis Method: EPA 608

QC Batch Method: EPA 608 SF

Analysis Description: 608 GCS Pest PCB

Associated Lab Samples: 3519325027, 3519325028

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3443

Analysis Method: EPA 608

QC Batch Method: EPA 608 SF

Analysis Description: 608 GCS Pest PCB

Associated Lab Samples: 3519325032

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3261

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325001

METHOD BLANK: 126466

Matrix: Water

Associated Lab Samples: 3519325001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0050U	0.021	10/11/10 22:40	
1,2-Dibromoethane (EDB)	ug/L	0.0064U	0.010	10/11/10 22:40	

LABORATORY CONTROL SAMPLE: 126467

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.26	0.32	124	60-140	
1,2-Dibromoethane (EDB)	ug/L	.26	0.32	125	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 126468

126469

Parameter	Units	3518753055 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L	0.0047 U	.43	.42	0.48	0.47	112	112	60-140	2	40	
1,2-Dibromoethane (EDB)	ug/L	0.0059 U	.43	.42	0.54	0.52	126	124	60-140	3	40	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3262

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325002

METHOD BLANK: 126470

Matrix: Water

Associated Lab Samples: 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0050U	0.021	10/12/10 05:05	
1,2-Dibromoethane (EDB)	ug/L	0.0064U	0.010	10/12/10 05:05	

LABORATORY CONTROL SAMPLE: 126471

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.26	0.30	118	60-140	
1,2-Dibromoethane (EDB)	ug/L	.26	0.31	122	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 126472

126473

Parameter	Units	3519325002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L	0.0049 U	.44	.44	0.60	0.56	137	128	60-140	7	40	
1,2-Dibromoethane (EDB)	ug/L	0.0062 U	.44	.44	0.63	0.59	145	135	60-140	7	40	J(M1)

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3350

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325016, 3519325018, 3519325019, 3519325020

METHOD BLANK: 131594

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325016, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0049U	0.020	10/22/10 21:42	
1,2-Dibromoethane (EDB)	ug/L	0.0062U	0.010	10/22/10 21:42	

LABORATORY CONTROL SAMPLE: 131595

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.25	0.31	125	60-140	
1,2-Dibromoethane (EDB)	ug/L	.25	0.28	113	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 131596

131597

Parameter	Units	3520108019 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L	0.0049 U	.44	.44	0.59	0.57	135	130	60-140	3	40	
1,2-Dibromoethane (EDB)	ug/L	0.0062 U	.44	.44	0.56	0.54	128	123	60-140	4	40	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3389

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325013

METHOD BLANK: 133969

Matrix: Water

Associated Lab Samples: 3519325013

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0050U	0.021	10/28/10 05:27	
1,2-Dibromoethane (EDB)	ug/L	0.0064U	0.010	10/28/10 05:27	

LABORATORY CONTROL SAMPLE: 133970

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.26	0.26	100	60-140	
1,2-Dibromoethane (EDB)	ug/L	.26	0.27	104	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133971

133972

Parameter	Units	10140959001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L		.42	.44	0.43	0.42	102	96	60-140	1	40	
1,2-Dibromoethane (EDB)	ug/L	ND	.42	.44	0.53	0.51	127	116	60-140	4	40	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3390

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325014, 3519325015, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

METHOD BLANK: 133973

Matrix: Water

Associated Lab Samples: 3519325014, 3519325015, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0050U	0.021	10/28/10 13:27	
1,2-Dibromoethane (EDB)	ug/L	0.0064U	0.010	10/28/10 13:27	

LABORATORY CONTROL SAMPLE: 133974

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.26	0.28	109	60-140	
1,2-Dibromoethane (EDB)	ug/L	.26	0.29	111	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133975

133976

Parameter	Units	3519325014 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L	0.0049 U	.44	.44	0.47	0.50	108	115	60-140	7	40	
1,2-Dibromoethane (EDB)	ug/L	0.0062 U	.44	.44	0.49	0.53	112	120	60-140	7	40	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3459

Analysis Method: EPA 8011

QC Batch Method: EPA 8011

Analysis Description: 8011 EDB DBCP

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 137446

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	0.0049U	0.020	11/04/10 16:19	
1,2-Dibromoethane (EDB)	ug/L	0.0062U	0.010	11/04/10 16:19	

LABORATORY CONTROL SAMPLE: 137447

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2-Dibromo-3-chloropropane	ug/L	.25	0.27	109	60-140	
1,2-Dibromoethane (EDB)	ug/L	.25	0.27	109	60-140	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 137448

137449

Parameter	Units	10142001007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2-Dibromo-3-chloropropane	ug/L		.44	.44	0.61	0.62	140	142	60-140	2	40	J(M1)
1,2-Dibromoethane (EDB)	ug/L	ND	.44	.44	0.52	0.55	120	126	60-140	5	40	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	OEXT/3326	Analysis Method:	EPA 8081
QC Batch Method:	EPA 3510	Analysis Description:	8081 GCS Pesticides
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014		

METHOD BLANK: 130317 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
4,4'-DDD	ug/L	0.0019U	0.010	11/12/10 20:37	
4,4'-DDE	ug/L	0.00090U	0.010	11/12/10 20:37	
4,4'-DDT	ug/L	0.0036U	0.010	11/12/10 20:37	
Aldrin	ug/L	0.00050U	0.010	11/12/10 20:37	
alpha-BHC	ug/L	0.00030U	0.010	11/12/10 20:37	
beta-BHC	ug/L	0.00050U	0.010	11/12/10 20:37	
Chlordane (Technical)	ug/L	0.080U	0.50	11/12/10 20:37	
Chlorobenzilate	ug/L	0.021U	0.10	11/12/10 20:37	
delta-BHC	ug/L	0.00040U	0.010	11/12/10 20:37	
Dieldrin	ug/L	0.00050U	0.010	11/12/10 20:37	
Endosulfan I	ug/L	0.00070U	0.010	11/12/10 20:37	
Endosulfan II	ug/L	0.00070U	0.010	11/12/10 20:37	
Endosulfan sulfate	ug/L	0.00060U	0.010	11/12/10 20:37	
Endrin	ug/L	0.0017U	0.010	11/12/10 20:37	
Endrin aldehyde	ug/L	0.0071U	0.010	11/12/10 20:37	
gamma-BHC (Lindane)	ug/L	0.00020U	0.010	11/12/10 20:37	
Heptachlor	ug/L	0.0015U	0.010	11/12/10 20:37	
Heptachlor epoxide	ug/L	0.00040U	0.010	11/12/10 20:37	
Methoxychlor	ug/L	0.0070U	0.010	11/12/10 20:37	
Pentachloronitrobenzene	ug/L	0.015U	0.10	11/12/10 20:37	
Toxaphene	ug/L	0.28U	0.50	11/12/10 20:37	
Decachlorobiphenyl (S)	%	112	41.7-109.1	11/12/10 20:37	S3
Tetrachloro-m-xylene (S)	%	94	66.5-120.3	11/12/10 20:37	

LABORATORY CONTROL SAMPLE: 130318

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
4,4'-DDD	ug/L	.5	0.50	100	74-122	
4,4'-DDE	ug/L	.5	0.52	105	74-117	
4,4'-DDT	ug/L	.5	0.52	105	81-117	
Aldrin	ug/L	.5	0.47	95	56-112	
alpha-BHC	ug/L	.5	0.47	94	66-110	
beta-BHC	ug/L	.5	0.51	102	77-121	
delta-BHC	ug/L	.5	0.45	89	46-108	
Dieldrin	ug/L	.5	0.52	104	76-122	
Endosulfan I	ug/L	.5	0.51	102	75-122	
Endosulfan II	ug/L	.5	0.53	106	75-126	
Endosulfan sulfate	ug/L	.5	0.53	107	74-118	
Endrin	ug/L	.5	0.51	102	71-122	
Endrin aldehyde	ug/L	.5	0.50	99	76-122	
gamma-BHC (Lindane)	ug/L	.5	0.52	103	64-119	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 237 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130318

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Heptachlor	ug/L	.5	0.49	98	64-116	
Heptachlor epoxide	ug/L	.5	0.52	103	76-120	
Methoxychlor	ug/L	.5	0.58	116	76-129	
Decachlorobiphenyl (S)	%			111	41.7-109.1	J(S0)
Tetrachloro-m-xylene (S)	%			93	66.5-120.3	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130319 130320

Parameter	Units	3520442001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
4,4'-DDD	ug/L	0.0019 U	1		0.98	0.95	98		74-122			
4,4'-DDE	ug/L				0.92	0.84				9	40	
4,4'-DDT	ug/L	0.0035 U	1	1	0.99	0.93	99	93	81-117	6	40	
Aldrin	ug/L	0.00049 U	1	1	0.95	0.92	95	92	46-112	3	40	
alpha-BHC	ug/L	0.00030 U	1	1	0.87	0.86	87	86	66-110	.4	40	
beta-BHC	ug/L				1.0	1.0				.6	40	
delta-BHC	ug/L	0.00039 U	1	1	0.87	0.87	87	87	46-108	.3	40	
Dieldrin	ug/L				1.0	0.98				2	40	
Endosulfan I	ug/L				0.95	0.94				1	40	
Endosulfan II	ug/L				1.0	1.0				.7	40	
Endosulfan sulfate	ug/L				1.0	1.0				.5	40	
Endrin	ug/L	0.0017 U	1	1	0.97	0.96	97	96	71-122	2	40	
Endrin aldehyde	ug/L				0.99	0.97				1	40	
gamma-BHC (Lindane)	ug/L	0.00020 U	1	1	0.96	0.95	96	95	64-119	1	40	
Heptachlor	ug/L				0.97	0.95				2	40	
Heptachlor epoxide	ug/L				0.97	0.96				1	40	
Methoxychlor	ug/L	0.0069 U	1	1	1.1	1.1	113	112	76-129	.9	40	
Decachlorobiphenyl (S)	%						52	43	41.7-109			
Tetrachloro-m-xylene (S)	%						93	94	66.5-120			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3369 Analysis Method: EPA 8081
QC Batch Method: EPA 3510 Analysis Description: 8081 GCS Pesticides
Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133076 Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
4,4'-DDD	ug/L	0.0019U	0.010	10/28/10 23:58	
4,4'-DDE	ug/L	0.00090U	0.010	10/28/10 23:58	
4,4'-DDT	ug/L	0.0036U	0.010	10/28/10 23:58	
Aldrin	ug/L	0.00050U	0.010	10/28/10 23:58	
alpha-BHC	ug/L	0.00030U	0.010	10/28/10 23:58	
beta-BHC	ug/L	0.00050U	0.010	10/28/10 23:58	
Chlordane (Technical)	ug/L	0.080U	0.50	10/28/10 23:58	
Chlorobenzilate	ug/L	0.021U	0.10	10/28/10 23:58	
delta-BHC	ug/L	0.00040U	0.010	10/28/10 23:58	
Dieldrin	ug/L	0.00050U	0.010	10/28/10 23:58	
Endosulfan I	ug/L	0.00070U	0.010	10/28/10 23:58	
Endosulfan II	ug/L	0.00070U	0.010	10/28/10 23:58	
Endosulfan sulfate	ug/L	0.00060U	0.010	10/28/10 23:58	
Endrin	ug/L	0.0017U	0.010	10/28/10 23:58	
Endrin aldehyde	ug/L	0.0071U	0.010	10/28/10 23:58	
gamma-BHC (Lindane)	ug/L	0.00020U	0.010	10/28/10 23:58	
Heptachlor	ug/L	0.0015U	0.010	10/28/10 23:58	
Heptachlor epoxide	ug/L	0.00040U	0.010	10/28/10 23:58	
Methoxychlor	ug/L	0.0070U	0.010	10/28/10 23:58	
Pentachloronitrobenzene	ug/L	0.015U	0.10	10/28/10 23:58	
Toxaphene	ug/L	0.28U	0.50	10/28/10 23:58	
Decachlorobiphenyl (S)	%	104	41.7-109.1	10/28/10 23:58	
Tetrachloro-m-xylene (S)	%	92	66.5-120.3	10/28/10 23:58	

LABORATORY CONTROL SAMPLE: 133077

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
4,4'-DDD	ug/L	.5	0.49	97	74-122	
4,4'-DDE	ug/L	.5	0.48	96	74-117	
4,4'-DDT	ug/L	.5	0.60	121	81-117	L3
Aldrin	ug/L	.5	0.45	89	56-112	
alpha-BHC	ug/L	.5	0.41	83	66-110	
beta-BHC	ug/L	.5	0.49	97	77-121	
delta-BHC	ug/L	.5	0.37	75	46-108	
Dieldrin	ug/L	.5	0.49	99	76-122	
Endosulfan I	ug/L	.5	0.47	94	75-122	
Endosulfan II	ug/L	.5	0.49	98	75-126	
Endosulfan sulfate	ug/L	.5	0.47	94	74-118	
Endrin	ug/L	.5	0.52	104	71-122	
Endrin aldehyde	ug/L	.5	0.49	97	76-122	
gamma-BHC (Lindane)	ug/L	.5	0.47	93	64-119	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 239 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 133077

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Heptachlor	ug/L	.5	0.50	101	64-116	
Heptachlor epoxide	ug/L	.5	0.50	100	76-120	
Methoxychlor	ug/L	.5	0.66	131	76-129 L3	
Decachlorobiphenyl (S)	%			87	41.7-109.1	
Tetrachloro-m-xylene (S)	%			87	66.5-120.3	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133078 133079

Parameter	Units	3519325027 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
4,4'-DDD	ug/L	0.0019 U	1	1	0.97	0.98	97	98	74-122	1	40	
4,4'-DDE	ug/L				0.99	1.0				2	40	
4,4'-DDT	ug/L				1.1	1.1				4	40	
Aldrin	ug/L				0.93	0.94				.4	40	
alpha-BHC	ug/L	0.00031 U	1	1	0.80	0.82	80	82	66-110	1	40	
beta-BHC	ug/L				0.97	0.93				4	40	
delta-BHC	ug/L	0.00041 U	1	1	0.67	0.71	67	71	46-108	6	40	
Dieldrin	ug/L				0.99	1.0				1	40	
Endosulfan I	ug/L				0.95	0.98				3	40	
Endosulfan II	ug/L				0.98	1.0				4	40	
Endosulfan sulfate	ug/L	0.00062 U	1	1	0.75	0.77	75	77	74-118	3	40	
Endrin	ug/L				1.0	1.0				2	40	
Endrin aldehyde	ug/L				0.96	0.97				2	40	
gamma-BHC (Lindane)	ug/L	0.00021 U	1	1	0.89	0.90	89	90	64-119	.5	40	
Heptachlor	ug/L	0.0015 U	1	1	0.96	0.97	96	97	64-116	.9	40	
Heptachlor epoxide	ug/L				1.1	1.1				4	40	
Methoxychlor	ug/L				1.3	1.2				5	40	
Decachlorobiphenyl (S)	%						49	55	41.7-109			
Tetrachloro-m-xylene (S)	%						90	87	66.5-120			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3431 Analysis Method: EPA 8081
QC Batch Method: EPA 3510 Analysis Description: 8081 GCS Pesticides
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 136030 Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
4,4'-DDD	ug/L	0.0019U	0.010	11/19/10 01:17	
4,4'-DDE	ug/L	0.00090U	0.010	11/19/10 01:17	
4,4'-DDT	ug/L	0.0036U	0.010	11/19/10 01:17	
Aldrin	ug/L	0.00050U	0.010	11/19/10 01:17	
alpha-BHC	ug/L	0.00030U	0.010	11/19/10 01:17	
beta-BHC	ug/L	0.00050U	0.010	11/19/10 01:17	
Chlordane (Technical)	ug/L	0.080U	0.50	11/19/10 01:17	
Chlorobenzilate	ug/L	0.021U	0.10	11/19/10 01:17	
delta-BHC	ug/L	0.00040U	0.010	11/19/10 01:17	
Dieldrin	ug/L	0.00050U	0.010	11/19/10 01:17	
Endosulfan I	ug/L	0.00070U	0.010	11/19/10 01:17	
Endosulfan II	ug/L	0.00070U	0.010	11/19/10 01:17	
Endosulfan sulfate	ug/L	0.00060U	0.010	11/19/10 01:17	
Endrin	ug/L	0.0017U	0.010	11/19/10 01:17	
Endrin aldehyde	ug/L	0.0071U	0.010	11/19/10 01:17	
gamma-BHC (Lindane)	ug/L	0.00020U	0.010	11/19/10 01:17	
Heptachlor	ug/L	0.0015U	0.010	11/19/10 01:17	
Heptachlor epoxide	ug/L	0.00040U	0.010	11/19/10 01:17	
Methoxychlor	ug/L	0.0070U	0.010	11/19/10 01:17	
Pentachloronitrobenzene	ug/L	0.015U	0.10	11/19/10 01:17	
Toxaphene	ug/L	0.28U	0.50	11/19/10 01:17	
Decachlorobiphenyl (S)	%	106	41.7-109.1	11/19/10 01:17	
Tetrachloro-m-xylene (S)	%	94	66.5-120.3	11/19/10 01:17	

LABORATORY CONTROL SAMPLE: 136031

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
4,4'-DDD	ug/L	.5	0.46	92	74-122	
4,4'-DDE	ug/L	.5	0.47	93	74-117	
4,4'-DDT	ug/L	.5	0.44	88	81-117	
Aldrin	ug/L	.5	0.45	90	56-112	
alpha-BHC	ug/L	.5	0.42	83	66-110	
beta-BHC	ug/L	.5	0.47	95	77-121	
delta-BHC	ug/L	.5	0.31	63	46-108	
Dieldrin	ug/L	.5	0.47	93	76-122	
Endosulfan I	ug/L	.5	0.45	90	75-122	
Endosulfan II	ug/L	.5	0.46	92	75-126	
Endosulfan sulfate	ug/L	.5	0.41	82	74-118	
Endrin	ug/L	.5	0.48	95	71-122	
Endrin aldehyde	ug/L	.5	0.47	94	76-122	
gamma-BHC (Lindane)	ug/L	.5	0.45	89	64-119	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 241 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 136031

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Heptachlor	ug/L	.5	0.48	97	64-116	
Heptachlor epoxide	ug/L	.5	0.46	92	76-120	
Methoxychlor	ug/L	.5	0.49	97	76-129	
Decachlorobiphenyl (S)	%			107	41.7-109.1	
Tetrachloro-m-xylene (S)	%			101	66.5-120.3	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136032 136033

Parameter	Units	3519325032 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
4,4'-DDD	ug/L	0.0021 U	1	1.1	0.71	0.98	71	88	74-122	31	40	J(M1)
4,4'-DDE	ug/L	0.00098 U	1		0.54	0.86	54		74-117			J(M1)
4,4'-DDT	ug/L				0.84	1.2				35	40	
Aldrin	ug/L				0.49	0.74				41	40	J(D6)
alpha-BHC	ug/L				0.60	0.90				41	40	J(D6)
beta-BHC	ug/L				0.77	1.1				34	40	
delta-BHC	ug/L	0.00044 U	1	1.1	0.63	1.0	63	94	46-108	50	40	J(D6)
Dieldrin	ug/L	0.00055 U	1		0.65	1.0	65		76-122			J(M1)
Endosulfan I	ug/L	0.00076 U	1	1.1	0.68	1.0	68	90	75-122	38	40	J(M1)
Endosulfan II	ug/L				0.68	0.98				36	40	
Endosulfan sulfate	ug/L	0.00066 U	1	1.1	0.69	0.99	69	89	74-118	35	40	J(M1)
Endrin	ug/L	0.0019 U	1	1.1	0.69	0.99	69	89	71-122	35	40	J(M1)
Endrin aldehyde	ug/L	0.0078 U	1	1.1	0.19	0.12	19	11	76-122	46	40	J(D6), J(M1)
gamma-BHC (Lindane)	ug/L				0.65	0.98				40	40	
Heptachlor	ug/L	0.0016 U	1	1.1	0.67	1.1	67	96	64-116	46	40	J(D6)
Heptachlor epoxide	ug/L			1.1	0.61	0.91		82				
Methoxychlor	ug/L				0.86	1.2				36	40	
Decachlorobiphenyl (S)	%						29	45	41.7-109			2p, J(S2)
Tetrachloro-m-xylene (S)	%						76	93	66.5-120			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	OEXT/3327	Analysis Method:	EPA 8082
QC Batch Method:	EPA 3510	Analysis Description:	8082 GCS PCB
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014		

METHOD BLANK: 130321 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	0.080U	0.50	11/12/10 20:37	
PCB-1221 (Aroclor 1221)	ug/L	0.081U	0.50	11/12/10 20:37	
PCB-1232 (Aroclor 1232)	ug/L	0.12U	0.50	11/12/10 20:37	
PCB-1242 (Aroclor 1242)	ug/L	0.13U	0.50	11/12/10 20:37	
PCB-1248 (Aroclor 1248)	ug/L	0.28U	0.50	11/12/10 20:37	
PCB-1254 (Aroclor 1254)	ug/L	0.14U	0.50	11/12/10 20:37	
PCB-1260 (Aroclor 1260)	ug/L	0.11U	0.50	11/12/10 20:37	
Decachlorobiphenyl (S)	%	112	63-121	11/12/10 20:37	
Tetrachloro-m-xylene (S)	%	94	48-111	11/12/10 20:37	

LABORATORY CONTROL SAMPLE: 130322

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	2.5	2.0	81	70-130	
PCB-1260 (Aroclor 1260)	ug/L	2.5	2.2	90	70-130	
Decachlorobiphenyl (S)	%			108	63-121	
Tetrachloro-m-xylene (S)	%			93	48-111	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130323 130324

Parameter	Units	3520442004 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
PCB-1016 (Aroclor 1016)	ug/L	0.082U	5	5	4.6	4.6	92	93	70-130	1	40	
PCB-1260 (Aroclor 1260)	ug/L	0.11U	5	5	3.5	3.3	69	66	70-130	5	40	J(M1)
Decachlorobiphenyl (S)	%						52	44	63-121			J(S2)
Tetrachloro-m-xylene (S)	%						70	69	48-111			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Project No.: 3519325

QC Batch: OEXT/3370

Analysis Method: EPA 8082

QC Batch Method: EPA 3510

Analysis Description: 8082 GCS PCB

Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133096

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	0.080U	0.50	10/28/10 23:58	
PCB-1221 (Aroclor 1221)	ug/L	0.081U	0.50	10/28/10 23:58	
PCB-1232 (Aroclor 1232)	ug/L	0.12U	0.50	10/28/10 23:58	
PCB-1242 (Aroclor 1242)	ug/L	0.13U	0.50	10/28/10 23:58	
PCB-1248 (Aroclor 1248)	ug/L	0.28U	0.50	10/28/10 23:58	
PCB-1254 (Aroclor 1254)	ug/L	0.14U	0.50	10/28/10 23:58	
PCB-1260 (Aroclor 1260)	ug/L	0.11U	0.50	10/28/10 23:58	
Decachlorobiphenyl (S)	%	94	63-121	10/28/10 23:58	
Tetrachloro-m-xylene (S)	%	85	48-111	10/28/10 23:58	

LABORATORY CONTROL SAMPLE: 133097

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	2.5	2.0	82	70-130	
PCB-1260 (Aroclor 1260)	ug/L	2.5	3.4	135	70-130 L3	
Decachlorobiphenyl (S)	%			83	63-121	
Tetrachloro-m-xylene (S)	%			78	48-111	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133098

133099

Parameter	Units	3519325028 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
PCB-1016 (Aroclor 1016)	ug/L	0.086U	2.5	2.5	1.9	2.0	77	78	70-130	1	40	
PCB-1260 (Aroclor 1260)	ug/L	0.12U	2.5	2.5	1.9	2.0	78	79	70-130	1	40	
Decachlorobiphenyl (S)	%						79	82	63-121			
Tetrachloro-m-xylene (S)	%						74	76	48-111			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3432

Analysis Method: EPA 8082

QC Batch Method: EPA 3510

Analysis Description: 8082 GCS PCB

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 136034

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	0.080U	0.50	11/19/10 01:17	
PCB-1221 (Aroclor 1221)	ug/L	0.081U	0.50	11/19/10 01:17	
PCB-1232 (Aroclor 1232)	ug/L	0.12U	0.50	11/19/10 01:17	
PCB-1242 (Aroclor 1242)	ug/L	0.13U	0.50	11/19/10 01:17	
PCB-1248 (Aroclor 1248)	ug/L	0.28U	0.50	11/19/10 01:17	
PCB-1254 (Aroclor 1254)	ug/L	0.14U	0.50	11/19/10 01:17	
PCB-1260 (Aroclor 1260)	ug/L	0.11U	0.50	11/19/10 01:17	
Decachlorobiphenyl (S)	%	125	63-121	11/19/10 01:17	S3
Tetrachloro-m-xylene (S)	%	122	48-111	11/19/10 01:17	S3

LABORATORY CONTROL SAMPLE: 136035

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/L	2.5	2.6	105	70-130	
PCB-1260 (Aroclor 1260)	ug/L	2.5	2.9	118	70-130	
Decachlorobiphenyl (S)	%			121	63-121	
Tetrachloro-m-xylene (S)	%			111	48-111	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136036

136037

Parameter	Units	3518484037 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
PCB-1016 (Aroclor 1016)	ug/L	0.089U	5	5	5.6	3.3	112	66	70-130	52	40	J(D6), J(M1)
PCB-1260 (Aroclor 1260)	ug/L	0.12U	5	5	3.1	3.2	62	64	70-130	3	40	J(M1)
Decachlorobiphenyl (S)	%						88	88	63-121			
Tetrachloro-m-xylene (S)	%						104	105	48-111			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3321 Analysis Method: EPA 8141
QC Batch Method: EPA 3510 Analysis Description: 8141 GCS, O/P Pesticides
Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 129949 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Dimethoate	ug/L	0.19U	0.50	10/21/10 05:27	
Disulfoton	ug/L	0.16U	0.50	10/21/10 05:27	
Famphur	ug/L	0.14U	0.50	10/21/10 05:27	
Methyl parathion	ug/L	0.19U	0.50	10/21/10 05:27	
Parathion (Ethyl parathion)	ug/L	0.35U	1.0	10/21/10 05:27	
Phorate	ug/L	0.37U	1.0	10/21/10 05:27	
4-Chloro3nitrobenzotrifluoride	%	86	34.2-122	10/21/10 05:27	

LABORATORY CONTROL SAMPLE: 129950

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dimethoate	ug/L	2	1.8	90	59.8-116.4	
Disulfoton	ug/L	2	2.0	98	66.9-116.2	
Famphur	ug/L	2	1.8	88	56.3-133	
Methyl parathion	ug/L	2	1.8	88	67.3-119.6	
Parathion (Ethyl parathion)	ug/L	4	3.3	84	70-130	
Phorate	ug/L	4	4.0	100	58-119.3	
4-Chloro3nitrobenzotrifluoride	%			80	34.2-122	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 129951 129952

Parameter	Units	3520442001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Dimethoate	ug/L	0.19U	4	4	3.7	3.4	92	85	59.8-116	8	40	
Disulfoton	ug/L	0.16U	4	4	4.0	4.0	100	100	66.9-116	1	40	
Famphur	ug/L	0.14U	4	4	3.6	3.5	90	88	56.3-133	3	40	
Methyl parathion	ug/L	0.19U	4	4	3.6	3.6	90	89	67.3-119	1	40	
Parathion (Ethyl parathion)	ug/L	0.35U	8	8	6.8	6.7	85	84	70-130	.9	40	
Phorate	ug/L	0.36U	8	8	7.7	8.3	97	103	58-119.3	6	40	
4-Chloro3nitrobenzotrifluoride	%						94	90	34.2-122			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3338

Analysis Method: EPA 8141

QC Batch Method: EPA 3510

Analysis Description: 8141 GCS, O/P Pesticides

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

METHOD BLANK: 130925

Matrix: Water

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Dimethoate	ug/L	0.19U	0.50	11/01/10 23:40	
Disulfoton	ug/L	0.16U	0.50	11/01/10 23:40	
Famphur	ug/L	0.14U	0.50	11/01/10 23:40	
Methyl parathion	ug/L	0.19U	0.50	11/01/10 23:40	
Parathion (Ethyl parathion)	ug/L	0.35U	1.0	11/01/10 23:40	
Phorate	ug/L	0.37U	1.0	11/01/10 23:40	
4-Chloro3nitrobenzotrifluoride	%	75	34.2-122	11/01/10 23:40	

LABORATORY CONTROL SAMPLE: 130926

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dimethoate	ug/L	2	1.7	84	59.8-116.4	
Disulfoton	ug/L	2	1.6	82	66.9-116.2	J(F3)
Famphur	ug/L	2	1.6	80	56.3-133	
Methyl parathion	ug/L	2	1.5	76	67.3-119.6	
Parathion (Ethyl parathion)	ug/L	4	3.0	74	70-130	
Phorate	ug/L	4	3.2	81	58-119.3	
4-Chloro3nitrobenzotrifluoride	%			79	34.2-122	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130927

130928

Parameter	Units	3519325014 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Dimethoate	ug/L	0.18U	4	4	3.7	3.5	93	88	59.8-116	6	40	
Disulfoton	ug/L	0.15U	4	4	2.7	2.6	67	65	66.9-116	3	40	J(M1)
Famphur	ug/L	0.14U	4	4	3.7	3.1	92	78	56.3-133	17	40	
Methyl parathion	ug/L	0.18U	4	4	3.4	3.2	86	81	67.3-119	6	40	
Parathion (Ethyl parathion)	ug/L	0.34U	8	8	6.7	6.2	84	78	70-130	7	40	
Phorate	ug/L	0.35U	8	8	4.6	4.6	57	58	58-119.3	.9	40	J(M1)
4-Chloro3nitrobenzotrifluoride	%						68	71	34.2-122			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3372

Analysis Method: EPA 8141

QC Batch Method: EPA 3510

Analysis Description: 8141 GCS, O/P Pesticides

Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133136

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Dimethoate	ug/L	0.19U	0.50	11/09/10 17:14	
Disulfoton	ug/L	0.16U	0.50	11/09/10 17:14	
Famphur	ug/L	0.14U	0.50	11/09/10 17:14	
Methyl parathion	ug/L	0.19U	0.50	11/09/10 17:14	
Parathion (Ethyl parathion)	ug/L	0.35U	1.0	11/09/10 17:14	
Phorate	ug/L	0.37U	1.0	11/09/10 17:14	
4-Chloro3nitrobenzotrifluoride	%	87	34.2-122	11/09/10 17:14	

LABORATORY CONTROL SAMPLE & LCSD: 133137

133181

Parameter	Units	Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec	% Rec Limits	RPD	Max RPD	Qualifiers
Dimethoate	ug/L	2	1.4	1.5	68	76	59.8-116.	11	40	
Disulfoton	ug/L	2	1.4	1.6	71	78	66.9-116.	10	40	
Famphur	ug/L	2	1.4	1.6	69	78	56.3-133	13	40	
Methyl parathion	ug/L	2	1.4	1.5	69	77	67.3-119.	11	40	
Parathion (Ethyl parathion)	ug/L	4	2.9	3.4	72	84	70-130	15	40	
Phorate	ug/L	4	2.4	2.7	60	68	58-119.3	11	40	
4-Chloro3nitrobenzotrifluoride	%				80	84	34.2-122			9p

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3445 Analysis Method: EPA 8141
QC Batch Method: EPA 3510 Analysis Description: 8141 GCS, O/P Pesticides
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 136722 Matrix: Water
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Dimethoate	ug/L	0.19U	0.50	11/09/10 22:03	
Disulfoton	ug/L	0.16U	0.50	11/09/10 22:03	
Famphur	ug/L	0.14U	0.50	11/09/10 22:03	
Methyl parathion	ug/L	0.19U	0.50	11/09/10 22:03	
Parathion (Ethyl parathion)	ug/L	0.35U	1.0	11/09/10 22:03	
Phorate	ug/L	0.37U	1.0	11/09/10 22:03	
4-Chloro3nitrobenzotrifluoride	%	89	34.2-122	11/09/10 22:03	

LABORATORY CONTROL SAMPLE: 136723

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dimethoate	ug/L	2	1.5	74	59.8-116.4	
Disulfoton	ug/L	2	1.6	79	66.9-116.2	
Famphur	ug/L	2	1.6	78	56.3-133	
Methyl parathion	ug/L	2	1.6	78	67.3-119.6	
Parathion (Ethyl parathion)	ug/L	4	3.6	90	70-130	
Phorate	ug/L	4	2.7	68	58-119.3	
4-Chloro3nitrobenzotrifluoride	%			77	34.2-122	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136724 136725

Parameter	Units	3519325037 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Dimethoate	ug/L	0.19U	4	4	3.5	3.6	88	91	59.8-116	3	40	
Disulfoton	ug/L	0.16U	4	4	3.9	3.4	98	86	66.9-116	14	40	
Famphur	ug/L	0.14U	4	4	3.0	3.1	75	77	56.3-133	3	40	
Methyl parathion	ug/L	0.19U	4	4	3.3	3.5	84	87	67.3-119	4	40	
Parathion (Ethyl parathion)	ug/L	0.35U	8	8	6.0	6.1	74	76	70-130	2	40	
Phorate	ug/L	0.37U	8	8	6.0	6.3	76	78	58-119.3	4	40	
4-Chloro3nitrobenzotrifluoride	%						82	83	34.2-122			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3452

Analysis Method: EPA 8141

QC Batch Method: EPA 3510

Analysis Description: 8141 GCS, O/P Pesticides

Associated Lab Samples: 3519325037

METHOD BLANK: 137033

Matrix: Water

Associated Lab Samples: 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
4-Chloro3nitrobenzotrifluoride	%	89	34.2-122	11/09/10 22:03	

LABORATORY CONTROL SAMPLE: 137034

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
4-Chloro3nitrobenzotrifluoride	%			77	34.2-122	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 137035

137036

Parameter	Units	3519325037 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
4-Chloro3nitrobenzotrifluoride	%						82	83	34.2-122			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3329

Analysis Method: EPA 8151

QC Batch Method: EPA 8151

Analysis Description: 8151A GCS Herbicides

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 130329

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
2,4,5-T	ug/L	0.042U	0.19	10/21/10 20:36	
2,4,5-TP (Silvex)	ug/L	0.049U	0.19	10/21/10 20:36	
2,4-D	ug/L	0.22U	0.94	10/21/10 20:36	
Dinoseb	ug/L	0.057U	0.19	10/21/10 20:36	
Pentachlorophenol	ug/L	0.017U	0.028	10/21/10 20:36	
2,4-DCPA (S)	%	98	65.5-125.7	10/21/10 20:36	

LABORATORY CONTROL SAMPLE: 130330

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2,4,5-T	ug/L	1.2	1.1	90	66.7-116.7	
2,4,5-TP (Silvex)	ug/L	1.2	1.1	92	65.6-127.9	
2,4-D	ug/L	6	5.6	93	62-135.6	
Dinoseb	ug/L	1.2	1.0	84	44-111.8	
Pentachlorophenol	ug/L	.18	0.16	87	55.5-114	
2,4-DCPA (S)	%			86	65.5-125.7	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130331

130332

Parameter	Units	3520108015 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
2,4,5-T	ug/L	0.043U	2.4	2.4	2.0	1.9	85	77	66.7-116	10	40	
2,4,5-TP (Silvex)	ug/L	0.051U	2.4	2.4	2.0	1.7	82	72	65.6-127	12	40	
2,4-D	ug/L	0.23U	12	12	10.1	9.0	84	75	62-135.6	12	40	
Dinoseb	ug/L	0.059U	2.4	2.4	1.9	1.6	78	68	44-111.8	13	40	
Pentachlorophenol	ug/L	0.018U	.36	.36	0.26	0.22	73	61	55.5-114	17	40	
2,4-DCPA (S)	%						85	77	65.5-125			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3345

Analysis Method: EPA 8151

QC Batch Method: EPA 8151

Analysis Description: 8151A GCS Herbicides

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

METHOD BLANK: 131462

Matrix: Water

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
2,4,5-T	ug/L	0.042U	0.19	10/27/10 00:56	
2,4,5-TP (Silvex)	ug/L	0.049U	0.19	10/27/10 00:56	
2,4-D	ug/L	0.22U	0.94	10/27/10 00:56	
Dinoseb	ug/L	0.057U	0.19	10/27/10 00:56	
Pentachlorophenol	ug/L	0.017U	0.028	10/27/10 00:56	
2,4-DCPA (S)	%	82	65.5-125.7	10/27/10 00:56	

LABORATORY CONTROL SAMPLE & LCSD: 131463

131836

Parameter	Units	Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec	% Rec Limits	RPD	Max RPD	Qualifiers
2,4,5-T	ug/L	1.2	1.1	1.1	96	94	66.7-116.	2	40	
2,4,5-TP (Silvex)	ug/L	1.2	0.99	1.0	82	85	65.6-127.	3	40	
2,4-D	ug/L	6	5.3	4.9	88	82	62-135.6	7	40	
Dinoseb	ug/L	1.2	1.2	1.2	99	97	44-111.8	2	40	
Pentachlorophenol	ug/L	.18	0.15	0.15	83	86	55.5-114	4	40	
2,4-DCPA (S)	%				83	83	65.5-125.			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3366

Analysis Method: EPA 8151

QC Batch Method: EPA 8151

Analysis Description: 8151A GCS Herbicides

Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133057

Matrix: Water

Associated Lab Samples: 3519325026, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
2,4,5-T	ug/L	0.042U	0.19	10/27/10 02:15	
2,4,5-TP (Silvex)	ug/L	0.049U	0.19	10/27/10 02:15	
2,4-D	ug/L	0.22U	0.94	10/27/10 02:15	
Dinoseb	ug/L	0.057U	0.19	10/27/10 02:15	
Pentachlorophenol	ug/L	0.017U	0.028	10/27/10 02:15	
2,4-DCPA (S)	%	95	65.5-125.7	10/27/10 02:15	

LABORATORY CONTROL SAMPLE: 133058

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2,4,5-T	ug/L	1.2	1.2	99	66.7-116.7	
2,4,5-TP (Silvex)	ug/L	1.2	1.0	86	65.6-127.9	
2,4-D	ug/L	6	5.7	95	62-135.6	
Dinoseb	ug/L	1.2	1.0	86	44-111.8	
Pentachlorophenol	ug/L	.18	0.15	85	55.5-114	
2,4-DCPA (S)	%			85	65.5-125.7	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133059

133060

Parameter	Units	10141191001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
2,4,5-T	ug/L	<0.047	2.4	2.4	2.2	2.0	90	84	66.7-116	7	40	
2,4,5-TP (Silvex)	ug/L	<0.055	2.4	2.4	2.2	2.0	92	82	65.6-127	11	40	
2,4-D	ug/L	<0.25	12	12	10.8	10.1	90	84	62-135.6	6	40	
Dinoseb	ug/L	<0.064	2.4	2.4	2.0	1.9	82	79	44-111.8	4	40	
Pentachlorophenol	ug/L	<0.019	.36	.36	0.27	0.26	76	72	55.5-114	6	40	
2,4-DCPA (S)	%						83	78	65.5-125			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3411 Analysis Method: EPA 8151
QC Batch Method: EPA 8151 Analysis Description: 8151A GCS Herbicides
Associated Lab Samples: 3519325032, 3519325034, 3519325035, 3519325037

METHOD BLANK: 135058 Matrix: Water

Associated Lab Samples: 3519325032, 3519325034, 3519325035, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
2,4,5-T	ug/L	0.042U	0.19	11/02/10 07:00	
2,4,5-TP (Silvex)	ug/L	0.049U	0.19	11/02/10 07:00	
2,4-D	ug/L	0.22U	0.94	11/02/10 07:00	
Dinoseb	ug/L	0.057U	0.19	11/02/10 07:00	
Pentachlorophenol	ug/L	0.017U	0.028	11/02/10 07:00	
2,4-DCPA (S)	%	101	65.5-125.7	11/02/10 07:00	

LABORATORY CONTROL SAMPLE: 135059

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2,4,5-T	ug/L	1.2	0.95	79	66.7-116.7	
2,4,5-TP (Silvex)	ug/L	1.2	0.87	73	65.6-127.9	
2,4-D	ug/L	6	5.1	85	62-135.6	
Dinoseb	ug/L	1.2	0.97	81	44-111.8	
Pentachlorophenol	ug/L	.18	0.15	84	55.5-114	
2,4-DCPA (S)	%			88	65.5-125.7	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135060 135061

Parameter	Units	9280763001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
2,4,5-T	ug/L	ND	2.4	2.4	2.1	2.2	88	91	66.7-116	4	40	
2,4,5-TP (Silvex)	ug/L	ND	2.4	2.4	2.2	2.2	93	93	65.6-127	.1	40	
2,4-D	ug/L	ND	12	12	11.2	9.6	94	80	62-135.6	16	40	
Dinoseb	ug/L	ND	2.4	2.4	2.1	2.0	89	84	44-111.8	6	40	
Pentachlorophenol	ug/L	ND	.36	.36	0.11	0.058	30	16	55.5-114	61	40	J(D6), J(M1)
2,4-DCPA (S)	%						99	87	65.5-125			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

QC Batch: OEXT/3440 Analysis Method: EPA 8151
QC Batch Method: EPA 8151 Analysis Description: 8151A GCS Herbicides
Associated Lab Samples: 3519325033, 3519325036

METHOD BLANK: 136264 Matrix: Water
Associated Lab Samples: 3519325033, 3519325036

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
2,4,5-T	ug/L	0.042U	0.19	11/04/10 21:20	
2,4,5-TP (Silvex)	ug/L	0.049U	0.19	11/04/10 21:20	
2,4-D	ug/L	0.22U	0.94	11/04/10 21:20	
Dinoseb	ug/L	0.057U	0.19	11/04/10 21:20	
Pentachlorophenol	ug/L	0.017U	0.028	11/04/10 21:20	
2,4-DCPA (S)	%	101	65.5-125.7	11/04/10 21:20	

LABORATORY CONTROL SAMPLE: 136265

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2,4,5-T	ug/L	1.2	1.2	97	66.7-116.7	
2,4,5-TP (Silvex)	ug/L	1.2	1.1	90	65.6-127.9	
2,4-D	ug/L	6	5.9	99	62-135.6	
Dinoseb	ug/L	1.2	1.0	84	44-111.8	
Pentachlorophenol	ug/L	.18	0.15	86	55.5-114	
2,4-DCPA (S)	%			95	65.5-125.7	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136266 136267

Parameter	Units	9280813001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
2,4,5-T	ug/L	ND	2.4	2.4	2.1	1.9	87	79	66.7-116	10	40	
2,4,5-TP (Silvex)	ug/L	ND	2.4	2.4	2.4	2.0	99	84	65.6-127	16	40	
2,4-D	ug/L	ND	12	12	11.6	10.0	97	84	62-135.6	15	40	
Dinoseb	ug/L	ND	2.4	2.4	2.1	1.2	88	52	44-111.8	51	40	J(D6)
Pentachlorophenol	ug/L	ND	.36	.36	0.22	0.15	60	43	55.5-114	34	40	J(M1)
2,4-DCPA (S)	%						99	84	65.5-125			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3260

Analysis Method: EPA 6010

QC Batch Method: EPA 3010

Analysis Description: 6010 MET

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123780

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	7.5U	15.0	10/07/10 21:57	
Arsenic	ug/L	5.0U	10.0	10/07/10 21:57	
Barium	ug/L	5.0U	10.0	10/07/10 21:57	
Beryllium	ug/L	0.50U	1.0	10/07/10 21:57	
Cadmium	ug/L	0.50U	1.0	10/07/10 21:57	
Calcium	mg/L	0.25U	0.50	10/07/10 21:57	
Chromium	ug/L	2.5U	5.0	10/07/10 21:57	
Cobalt	ug/L	5.0U	10.0	10/07/10 21:57	
Copper	ug/L	2.5U	5.0	10/07/10 21:57	
Iron	ug/L	20.0U	40.0	10/07/10 21:57	
Lead	ug/L	5.0U	10.0	10/07/10 21:57	
Magnesium	mg/L	0.25U	0.50	10/07/10 21:57	
Nickel	ug/L	2.5U	5.0	10/07/10 21:57	
Potassium	mg/L	0.50U	1.0	10/07/10 21:57	
Selenium	ug/L	7.5U	15.0	10/07/10 21:57	
Silver	ug/L	2.5U	5.0	10/07/10 21:57	
Sodium	mg/L	0.50U	1.0	10/07/10 21:57	
Tin	ug/L	25.0U	50.0	10/07/10 21:57	
Tot Hardness asCaCO3 (SM 2340B	mg/L	1.6U	3.2	10/07/10 21:57	
Vanadium	ug/L	5.0U	10.0	10/07/10 21:57	
Zinc	ug/L	10.0U	20.0	10/07/10 21:57	

LABORATORY CONTROL SAMPLE: 123781

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	250	273	109	80-120	
Arsenic	ug/L	250	276	110	80-120	
Barium	ug/L	250	283	113	80-120	
Beryllium	ug/L	25	28.0	112	80-120	
Cadmium	ug/L	25	29.1	116	80-120	
Calcium	mg/L	12.5	13.8	110	80-120	
Chromium	ug/L	250	284	114	80-120	
Cobalt	ug/L	250	279	112	80-120	
Copper	ug/L	250	270	108	80-120	
Iron	ug/L	2500	2880	115	80-120	
Lead	ug/L	250	280	112	80-120	
Magnesium	mg/L	12.5	13.8	110	80-120	
Nickel	ug/L	250	286	114	80-120	
Potassium	mg/L	12.5	12.6	101	80-120	
Selenium	ug/L	250	278	111	80-120	
Silver	ug/L	25	27.8	111	80-120	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 256 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 123781

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sodium	mg/L	12.5	13.0	104	80-120	
Tin	ug/L	1250	1340	107	80-120	
Tot Hardness asCaCO3 (SM 2340B	mg/L		91.3			
Vanadium	ug/L	250	279	112	80-120	
Zinc	ug/L	1250	1400	112	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 123782 123783

Parameter	Units	3519325002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Antimony	ug/L	7.5U	250	250	269	270	107	107	75-125	.4	20	
Arsenic	ug/L	5.0U	250	250	277	281	110	111	75-125	1	20	
Barium	ug/L	20.0	250	250	297	300	111	112	75-125	1	20	
Beryllium	ug/L	0.50U	25	25	27.6	27.7	110	110	75-125	.4	20	
Cadmium	ug/L	0.50U	25	25	28.3	28.6	112	113	75-125	1	20	
Calcium	mg/L	43.3	12.5	12.5	57.3	56.9	112	109	75-125	.7	20	
Chromium	ug/L	2.5U	250	250	277	277	111	111	75-125	0	20	
Cobalt	ug/L	5.0U	250	250	279	280	111	112	75-125	.4	20	
Copper	ug/L	2.5U	250	250	271	273	108	109	75-125	.7	20	
Iron	ug/L	510	2500	2500	3350	3350	114	114	75-125	0	20	
Lead	ug/L	5.0U	250	250	277	277	110	110	75-125	0	20	
Magnesium	mg/L	7.8	12.5	12.5	21.5	21.4	110	109	75-125	.5	20	
Nickel	ug/L	2.5U	250	250	280	280	112	112	75-125	0	20	
Potassium	mg/L	24.2	12.5	12.5	38.1	38.2	111	112	75-125	.3	20	
Selenium	ug/L	7.5U	250	250	280	275	111	109	75-125	2	20	
Silver	ug/L	2.5U	25	25	27.2	26.7	107	105	75-125	2	20	
Sodium	mg/L	14.1	12.5	12.5	27.5	27.6	107	108	75-125	.4	20	
Tin	ug/L	25.0U	1250	1250	1340	1340	107	107	75-125	0	20	
Tot Hardness asCaCO3 (SM 2340B	mg/L	140			232	230				.6	20	
Vanadium	ug/L	5.2 I	250	250	280	281	110	110	75-125	.4	20	
Zinc	ug/L	10.0U	1250	1250	1380	1380	110	110	75-125	0	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	MPRP/3418	Analysis Method:	EPA 6010
QC Batch Method:	EPA 3010	Analysis Description:	6010 MET
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325017, 3519325018, 3519325019, 3519325020		

METHOD BLANK: 131521 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325017, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Aluminum	ug/L	50.0U	100	10/21/10 23:29	
Arsenic	ug/L	5.0U	10.0	10/21/10 23:29	
Barium	ug/L	5.0U	10.0	10/21/10 23:29	
Beryllium	ug/L	0.50U	1.0	10/21/10 23:29	
Cadmium	ug/L	0.50U	1.0	10/21/10 23:29	
Calcium	mg/L	0.25U	0.50	10/21/10 23:29	
Chromium	ug/L	2.5U	5.0	10/21/10 23:29	
Cobalt	ug/L	5.0U	10.0	10/21/10 23:29	
Copper	ug/L	2.5U	5.0	10/21/10 23:29	
Iron	ug/L	20.0U	40.0	10/21/10 23:29	
Lead	ug/L	5.0U	10.0	10/21/10 23:29	
Magnesium	mg/L	0.25U	0.50	10/21/10 23:29	
Manganese	ug/L	2.5U	5.0	10/21/10 23:29	
Nickel	ug/L	2.5U	5.0	10/21/10 23:29	
Potassium	mg/L	0.50U	1.0	10/21/10 23:29	
Selenium	ug/L	7.5U	15.0	10/21/10 23:29	
Silver	ug/L	2.5U	5.0	10/21/10 23:29	
Sodium	mg/L	0.50U	1.0	10/21/10 23:29	
Tin	ug/L	25.0U	50.0	10/21/10 23:29	
Vanadium	ug/L	5.0U	10.0	10/21/10 23:29	
Zinc	ug/L	10.0U	20.0	10/21/10 23:29	

LABORATORY CONTROL SAMPLE: 131522

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Aluminum	ug/L	2500	2540	102	80-120	
Arsenic	ug/L	250	274	110	80-120	
Barium	ug/L	250	278	111	80-120	
Beryllium	ug/L	25	27.3	109	80-120	
Cadmium	ug/L	25	28.1	112	80-120	
Calcium	mg/L	12.5	13.3	106	80-120	
Chromium	ug/L	250	276	110	80-120	
Cobalt	ug/L	250	276	110	80-120	
Copper	ug/L	250	267	107	80-120	
Iron	ug/L	2500	2740	110	80-120	
Lead	ug/L	250	276	110	80-120	
Magnesium	mg/L	12.5	13.3	106	80-120	
Manganese	ug/L	250	279	112	80-120	
Nickel	ug/L	250	278	111	80-120	
Potassium	mg/L	12.5	13.0	104	80-120	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 258 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 131522

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Selenium	ug/L	250	277	111	80-120	
Silver	ug/L	25	27.8	111	80-120	
Sodium	mg/L	12.5	13.1	105	80-120	
Tin	ug/L	1250	1360	109	80-120	
Vanadium	ug/L	250	273	109	80-120	
Zinc	ug/L	1250	1380	110	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 131523 131524

Parameter	Units	3519325007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Aluminum	ug/L	50.0U	2500	2500	2580	2630	103	105	75-125	2	20	
Arsenic	ug/L	26.6	250	250	288	295	105	107	75-125	2	20	
Barium	ug/L	45.7	250	250	311	317	106	109	75-125	2	20	
Beryllium	ug/L	0.50U	25	25	25.8	26.1	103	104	75-125	1	20	
Cadmium	ug/L	0.50U	25	25	25.7	26.1	103	104	75-125	2	20	
Calcium	mg/L	155	12.5	12.5	164	168	72	104	75-125	2	20	J(M1), J(P6)
Chromium	ug/L	2.5U	250	250	261	264	104	106	75-125	1	20	
Cobalt	ug/L	5.0U	250	250	255	260	102	104	75-125	2	20	
Copper	ug/L	2.5U	250	250	266	270	106	108	75-125	1	20	
Iron	ug/L	7100	2500	2500	9500	9700	96	104	75-125	2	20	
Lead	ug/L	5.0U	250	250	260	266	104	106	75-125	2	20	
Magnesium	mg/L	42.3	12.5	12.5	54.5	55.8	98	108	75-125	2	20	
Manganese	ug/L	25.3	250	250	288	292	105	107	75-125	1	20	
Nickel	ug/L	2.5U	250	250	262	265	104	105	75-125	1	20	
Potassium	mg/L	1.4	12.5	12.5	15.9	16.3	116	119	75-125	2	20	
Selenium	ug/L	7.5U	250	250	265	274	105	108	75-125	3	20	
Silver	ug/L	2.5U	25	25	27.3	27.4	107	107	75-125	.4	20	
Sodium	mg/L	24.1	12.5	12.5	37.2	38.2	105	113	75-125	3	20	
Tin	ug/L	25.0U	1250	1250	1320	1340	106	107	75-125	2	20	
Vanadium	ug/L	5.0U	250	250	262	267	104	106	75-125	2	20	
Zinc	ug/L	14.3 I	1250	1250	1320	1340	104	106	75-125	2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3439

Analysis Method: EPA 6010

QC Batch Method: EPA 3010

Analysis Description: 6010 MET

Associated Lab Samples: 3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

METHOD BLANK: 132237

Matrix: Water

Associated Lab Samples: 3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Aluminum	ug/L	50.0U	100	10/23/10 13:43	
Arsenic	ug/L	5.0U	10.0	10/23/10 13:43	
Barium	ug/L	5.0U	10.0	10/23/10 13:43	
Beryllium	ug/L	0.50U	1.0	10/23/10 13:43	
Cadmium	ug/L	0.50U	1.0	10/23/10 13:43	
Calcium	mg/L	0.25U	0.50	10/23/10 13:43	
Chromium	ug/L	2.5U	5.0	10/23/10 13:43	
Cobalt	ug/L	5.0U	10.0	10/23/10 13:43	
Copper	ug/L	2.5U	5.0	10/23/10 13:43	
Iron	ug/L	20.0U	40.0	10/23/10 13:43	
Lead	ug/L	5.0U	10.0	10/23/10 13:43	
Magnesium	mg/L	0.25U	0.50	10/23/10 13:43	
Nickel	ug/L	2.5U	5.0	10/23/10 13:43	
Potassium	mg/L	0.50U	1.0	10/23/10 13:43	
Selenium	ug/L	7.5U	15.0	10/23/10 13:43	
Silver	ug/L	2.5U	5.0	10/23/10 13:43	
Sodium	mg/L	0.50U	1.0	10/23/10 13:43	
Tin	ug/L	25.0U	50.0	10/23/10 13:43	
Vanadium	ug/L	5.0U	10.0	10/23/10 13:43	
Zinc	ug/L	10.0U	20.0	10/23/10 13:43	

LABORATORY CONTROL SAMPLE: 132238

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Aluminum	ug/L	2500	2540	102	80-120	
Arsenic	ug/L	250	270	108	80-120	
Barium	ug/L	250	272	109	80-120	
Beryllium	ug/L	25	27.6	110	80-120	
Cadmium	ug/L	25	28.0	112	80-120	
Calcium	mg/L	12.5	13.1	105	80-120	
Chromium	ug/L	250	278	111	80-120	
Cobalt	ug/L	250	278	111	80-120	
Copper	ug/L	250	268	107	80-120	
Iron	ug/L	2500	2770	111	80-120	
Lead	ug/L	250	275	110	80-120	
Magnesium	mg/L	12.5	13.1	105	80-120	
Nickel	ug/L	250	278	111	80-120	
Potassium	mg/L	12.5	12.1	97	80-120	
Selenium	ug/L	250	265	106	80-120	
Silver	ug/L	25	27.5	110	80-120	
Sodium	mg/L	12.5	12.8	102	80-120	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 260 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 132238

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Tin	ug/L	1250	1370	110	80-120	
Vanadium	ug/L	250	274	110	80-120	
Zinc	ug/L	1250	1370	110	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132239 132240

Parameter	Units	3519325028 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Aluminum	ug/L		2500	2500	2710	2760	103	105	75-125	2	20	
Arsenic	ug/L	5.0U	250	250	271	275	108	109	75-125	1	20	
Barium	ug/L	34.8	250	250	298	300	105	106	75-125	.7	20	
Beryllium	ug/L	0.50U	25	25	26.8	27.3	107	109	75-125	2	20	
Cadmium	ug/L	0.50U	25	25	27.1	27.6	108	110	75-125	2	20	
Calcium	mg/L	82.0	12.5	12.5	94.8	97.6	102	125	75-125	3	20	
Chromium	ug/L	2.5U	250	250	273	278	108	110	75-125	2	20	
Cobalt	ug/L	5.0U	250	250	271	276	108	110	75-125	2	20	
Copper	ug/L	2.5U	250	250	273	274	109	109	75-125	.4	20	
Iron	ug/L	6550	2500	2500	9170	9520	105	119	75-125	4	20	
Lead	ug/L	5.0U	250	250	268	277	107	111	75-125	3	20	
Magnesium	mg/L	12.8	12.5	12.5	25.8	26.5	104	110	75-125	3	20	
Nickel	ug/L	2.5U	250	250	271	274	108	109	75-125	1	20	
Potassium	mg/L	1.0	12.5	12.5	14.3	14.4	106	107	75-125	.7	20	
Selenium	ug/L	7.5U	250	250	263	263	105	105	75-125	0	20	
Silver	ug/L	2.5U	25	25	27.9	26.9	112	108	75-125	4	20	
Sodium	mg/L	18.2	12.5	12.5	31.4	32.0	106	110	75-125	2	20	
Tin	ug/L	25.0U	1250	1250	1370	1400	110	112	75-125	2	20	
Vanadium	ug/L	5.2 I	250	250	276	279	108	110	75-125	1	20	
Zinc	ug/L	10.0U	1250	1250	1330	1350	106	108	75-125	1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3530

Analysis Method: EPA 6010

QC Batch Method: EPA 3010

Analysis Description: 6010 MET

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039

METHOD BLANK: 138074

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Arsenic	ug/L	5.0U	10.0	11/11/10 17:47	
Barium	ug/L	5.0U	10.0	11/11/10 17:47	
Beryllium	ug/L	0.50U	1.0	11/11/10 17:47	
Cadmium	ug/L	0.50U	1.0	11/11/10 17:47	
Calcium	mg/L	0.25U	0.50	11/11/10 17:47	
Chromium	ug/L	2.5U	5.0	11/11/10 17:47	
Cobalt	ug/L	5.0U	10.0	11/11/10 17:47	
Copper	ug/L	2.5U	5.0	11/11/10 17:47	
Iron	ug/L	20.0U	40.0	11/11/10 17:47	
Lead	ug/L	5.0U	10.0	11/11/10 17:47	
Magnesium	mg/L	0.25U	0.50	11/11/10 17:47	
Manganese	ug/L	2.5U	5.0	11/11/10 17:47	
Nickel	ug/L	2.5U	5.0	11/11/10 17:47	
Potassium	mg/L	0.50U	1.0	11/11/10 17:47	
Selenium	ug/L	7.5U	15.0	11/11/10 17:47	
Silver	ug/L	2.5U	5.0	11/11/10 17:47	
Sodium	mg/L	0.50U	1.0	11/11/10 17:47	
Tin	ug/L	25.0U	50.0	11/11/10 17:47	
Vanadium	ug/L	5.0U	10.0	11/11/10 17:47	
Zinc	ug/L	10.0U	20.0	11/11/10 17:47	

LABORATORY CONTROL SAMPLE: 138075

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Arsenic	ug/L	250	272	109	80-120	
Barium	ug/L	250	270	108	80-120	
Beryllium	ug/L	25	26.9	108	80-120	
Cadmium	ug/L	25	27.8	111	80-120	
Calcium	mg/L	12.5	13.1	105	80-120	
Chromium	ug/L	250	280	112	80-120	
Cobalt	ug/L	250	275	110	80-120	
Copper	ug/L	250	271	108	80-120	
Iron	ug/L	2500	2730	109	80-120	
Lead	ug/L	250	270	108	80-120	
Magnesium	mg/L	12.5	13.2	106	80-120	
Manganese	ug/L	250	282	113	80-120	
Nickel	ug/L	250	275	110	80-120	
Potassium	mg/L	12.5	12.5	100	80-120	
Selenium	ug/L	250	259	104	80-120	
Silver	ug/L	25	28.6	114	80-120	
Sodium	mg/L	12.5	13.3	106	80-120	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 262 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 138075

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Tin	ug/L	1250	1360	109	80-120	
Vanadium	ug/L	250	273	109	80-120	
Zinc	ug/L	1250	1340	107	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 138212 138213

Parameter	Units	3521205022 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Arsenic	ug/L	16.3	250	250	289	283	109	107	75-125	2	20	
Barium	ug/L	12.4	250	250	287	286	110	109	75-125	.3	20	
Beryllium	ug/L	0.50U	25	25	27.3	26.7	109	107	75-125	2	20	
Cadmium	ug/L	0.50U	25	25	27.8	27.5	111	109	75-125	1	20	
Calcium	mg/L	5740	12.5	12.5	18.9	18.7	105	104	75-125	1	20	
Chromium	ug/L	2.5U	250	250	283	270	113	107	75-125	5	20	
Cobalt	ug/L	5.0U	250	250	281	272	111	108	75-125	3	20	
Copper	ug/L	8.6	250	250	286	272	111	105	75-125	5	20	
Iron	ug/L	3320	2500	2500	6090	5970	111	106	75-125	2	20	
Lead	ug/L	5.0U	250	250	271	264	108	106	75-125	3	20	
Magnesium	mg/L	1450	12.5	12.5	14.8	14.4	107	104	75-125	3	20	
Manganese	ug/L	41.3	250	250	324	312	113	108	75-125	4	20	
Nickel	ug/L	4.4 I	250	250	283	273	111	107	75-125	4	20	
Potassium	mg/L	500U	12.5	12.5	13.1	12.8	103	100	75-125	2	20	
Selenium	ug/L	7.5U	250	250	260	263	103	104	75-125	1	20	
Silver	ug/L	2.5U	25	25	28.4	27.6	113	109	75-125	3	20	
Sodium	mg/L	2.8	12.5	12.5	16.0	15.2	105	99	75-125	5	20	
Tin	ug/L	25.0U	1250	1250	1380	1290	110	103	75-125	7	20	
Vanadium	ug/L	5.0U	250	250	279	271	110	107	75-125	3	20	
Zinc	ug/L	32.4	1250	1250	1380	1380	108	108	75-125	0	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3259

Analysis Method: EPA 6020

QC Batch Method: EPA 3010

Analysis Description: 6020 MET

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123776

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	0.50U	1.0	10/21/10 03:46	
Beryllium	ug/L	0.050U	0.10	10/21/10 03:46	
Cadmium	ug/L	0.050U	0.10	10/21/10 03:46	
Copper	ug/L	0.93U	1.0	10/21/10 03:46	
Lead	ug/L	0.50U	1.0	10/21/10 03:46	
Selenium	ug/L	0.50U	1.0	10/21/10 03:46	
Silver	ug/L	0.050U	0.10	10/21/10 03:46	
Thallium	ug/L	0.50U	1.0	10/21/10 03:46	
Zinc	ug/L	2.5U	5.0	10/21/10 03:46	

LABORATORY CONTROL SAMPLE: 123777

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	50	50.1	100	90-110	
Beryllium	ug/L	5	5.5	110	90-110	
Cadmium	ug/L	5	5.4	108	90-110	
Copper	ug/L	50	52.1	104	90-110	
Lead	ug/L	50	49.6	99	90-110	
Selenium	ug/L	50	54.2	108	90-110	
Silver	ug/L	5	5.3	106	90-110	
Thallium	ug/L	50	51.0	102	90-110	
Zinc	ug/L	250	248	99	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 123778

123779

Parameter	Units	3519325001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Antimony	ug/L	0.52 I	50	50	51.6	50.3	102	100	70-130	2	20	
Beryllium	ug/L	0.050U	5	5	5.3	5.3	106	105	70-130	1	20	
Cadmium	ug/L	0.050U	5	5	5.0	5.0	100	99	70-130	.5	20	
Copper	ug/L	0.93U	50	50	52.0	51.7	103	102	70-130	.6	20	
Lead	ug/L	0.50U	50	50	50.2	49.5	100	98	70-130	1	20	
Selenium	ug/L	0.50U	50	50	50.5	47.9	101	95	70-130	5	20	
Silver	ug/L	0.050U	5	5	5.1	5.1	102	101	70-130	1	20	
Thallium	ug/L	0.61 I	50	50	52.2	52.6	103	104	70-130	.8	20	
Zinc	ug/L	2.5U	250	250	256	259	102	103	70-130	1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	MPRP/3417	Analysis Method:	EPA 6020
QC Batch Method:	EPA 3010	Analysis Description:	6020 MET
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020		

METHOD BLANK: 131517 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	0.50U	1.0	10/28/10 03:22	
Thallium	ug/L	0.50U	1.0	10/28/10 03:22	

LABORATORY CONTROL SAMPLE: 131518

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	50	50.4	101	90-110	
Thallium	ug/L	50	50.2	100	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 131519 131520

Parameter	Units	3519325006 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Antimony	ug/L	0.50U	50	50	50.5	50.7	101	101	70-130	.3	20	
Thallium	ug/L	0.50U	50	50	51.3	51.4	102	103	70-130	.3	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3438 Analysis Method: EPA 6020
QC Batch Method: EPA 3010 Analysis Description: 6020 MET
Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

METHOD BLANK: 132231 Matrix: Water
Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	0.50U	1.0	11/07/10 00:22	
Thallium	ug/L	0.50U	1.0	11/07/10 00:22	

LABORATORY CONTROL SAMPLE: 132232

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	50	50.7	101	90-110	
Thallium	ug/L	50	50.8	102	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132233 132234

Parameter	Units	3519325027 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Antimony	ug/L	0.50U	50	50	49.7	50.1	99	100	70-130	.9	20	
Thallium	ug/L	0.50U	50	50	50.9	51.2	102	102	70-130	.6	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MPRP/3529

Analysis Method: EPA 6020

QC Batch Method: EPA 3010

Analysis Description: 6020 MET

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 138068

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	0.50U	1.0	11/18/10 08:46	
Thallium	ug/L	0.50U	1.0	11/18/10 08:46	

LABORATORY CONTROL SAMPLE: 138069

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	50	51.4	103	90-110	
Thallium	ug/L	50	50.4	101	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 138070

138071

Parameter	Units	3521205021 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Antimony	ug/L	0.50U	50	50	51.4	50.9	103	102	70-130	.8	20	
Thallium	ug/L	0.50U	50	50	52.4	52.4	104	104	70-130	.04	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MERP/1565

Analysis Method: EPA 7470

QC Batch Method: EPA 7470

Analysis Description: 7470 Mercury

Associated Lab Samples: 3519325025, 3519325029

METHOD BLANK: 132150

Matrix: Water

Associated Lab Samples: 3519325025, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	0.10U	0.20	10/25/10 12:54	

LABORATORY CONTROL SAMPLE: 132151

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Mercury	ug/L	2	2.1	103	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132152

132153

Parameter	Units	3520572010 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	2.2	2.2	110	110	85-115	.7	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132154

132155

Parameter	Units	3519325029 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.6	1.6	81	80	85-115	2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MERP/1570

Analysis Method: EPA 7470

QC Batch Method: EPA 7470

Analysis Description: 7470 Mercury

Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133995

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	0.10U	0.20	10/28/10 13:52	

LABORATORY CONTROL SAMPLE: 133996

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Mercury	ug/L	2	1.9	93	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133997

133998

Parameter	Units	3520667021 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.9	1.9	94	93	85-115	.2	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133999

134000

Parameter	Units	3520667026 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.7	1.7	85	85	85-115	.4	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	MERP/1583	Analysis Method:	EPA 7470
QC Batch Method:	EPA 7470	Analysis Description:	7470 Mercury
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020		

METHOD BLANK:	136736	Matrix:	Water
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	0.10U	0.20	11/04/10 08:45	

LABORATORY CONTROL SAMPLE: 136737

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Mercury	ug/L	2	2.0	100	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE:		136738	136739									
Parameter	Units	3519325016 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.8	1.8	89	88	85-115	1	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE:		136740	136741									
Parameter	Units	3521204004 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.6	1.5	78	75	85-115	4	20	J(M1)

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	MERP/1592	Analysis Method:	EPA 7470
QC Batch Method:	EPA 7470	Analysis Description:	7470 Mercury
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

METHOD BLANK:	138000	Matrix:	Water
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Mercury	ug/L	0.10U	0.20	11/08/10 11:16	

LABORATORY CONTROL SAMPLE: 138001

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Mercury	ug/L	2	2.0	98	80-120	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 138002 138003

Parameter	Units	3519325039 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.14 I	2	2	2.3	2.1	107	97	85-115	9	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 138004 138005

Parameter	Units	3521520008 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Mercury	ug/L	0.10U	2	2	1.9	1.8	94	88	85-115	7	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3333

Analysis Method: EPA 8270

QC Batch Method: EPA 3510

Analysis Description: 8270 Water MSSV App II

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

METHOD BLANK: 130562

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	0.70U	5.0	10/21/10 17:47	
1,2,4-Trichlorobenzene	ug/L	0.83U	5.0	10/21/10 17:47	
1,2-Dichlorobenzene	ug/L	0.68U	5.0	10/21/10 17:47	
1,2-Dinitrobenzene	ug/L	1.2U	5.0	10/21/10 17:47	
1,3,5-Trinitrobenzene	ug/L	1.2U	5.0	10/21/10 17:47	
1,3-Dichlorobenzene	ug/L	0.76U	5.0	10/21/10 17:47	
1,3-Dinitrobenzene	ug/L	0.68U	8.0	10/21/10 17:47	
1,4-Dichlorobenzene	ug/L	0.77U	5.0	10/21/10 17:47	
1,4-Naphthoquinone	ug/L	1.2U	5.0	10/21/10 17:47	
1-Methylnaphthalene	ug/L	1.0U	5.0	10/21/10 17:47	
1-Naphthylamine	ug/L	1.0U	5.0	10/21/10 17:47	
2,3,4,6-Tetrachlorophenol	ug/L	3.8U	5.0	10/21/10 17:47	
2,4,5-Trichlorophenol	ug/L	0.52U	4.0	10/21/10 17:47	
2,4,6-Trichlorophenol	ug/L	0.69U	2.0	10/21/10 17:47	
2,4-Dichlorophenol	ug/L	0.56U	2.0	10/21/10 17:47	
2,4-Dimethylphenol	ug/L	1.6U	5.0	10/21/10 17:47	
2,4-Dinitrophenol	ug/L	1.6U	20.0	10/21/10 17:47	
2,4-Dinitrotoluene	ug/L	0.53U	2.0	10/21/10 17:47	
2,6-Dichlorophenol	ug/L	0.62U	4.0	10/21/10 17:47	
2,6-Dinitrotoluene	ug/L	1.2U	2.0	10/21/10 17:47	
2-Acetylaminofluorene	ug/L	0.65U	5.0	10/21/10 17:47	
2-Chloronaphthalene	ug/L	0.80U	5.0	10/21/10 17:47	
2-Chlorophenol	ug/L	0.68U	5.0	10/21/10 17:47	
2-Methylnaphthalene	ug/L	0.99U	5.0	10/21/10 17:47	
2-Methylphenol(o-Cresol)	ug/L	0.73U	5.0	10/21/10 17:47	
2-Naphthylamine	ug/L	2.3U	5.0	10/21/10 17:47	
2-Nitroaniline	ug/L	0.60U	5.0	10/21/10 17:47	
2-Nitrophenol	ug/L	0.81U	5.0	10/21/10 17:47	
3&4-Methylphenol(m&p Cresol)	ug/L	0.66U	10.0	10/21/10 17:47	
3,3'-Dichlorobenzidine	ug/L	0.69U	10.0	10/21/10 17:47	
3,3'-Dimethylbenzidine	ug/L	3.1U	10.0	10/21/10 17:47	
3-Methylcholanthrene	ug/L	1.0U	5.0	10/21/10 17:47	
3-Nitroaniline	ug/L	0.99U	5.0	10/21/10 17:47	
4,6-Dinitro-2-methylphenol	ug/L	1.3U	20.0	10/21/10 17:47	
4-Aminobiphenyl	ug/L	2.8U	5.0	10/21/10 17:47	
4-Bromophenylphenyl ether	ug/L	0.67U	5.0	10/21/10 17:47	
4-Chloro-3-methylphenol	ug/L	0.62U	20.0	10/21/10 17:47	
4-Chloroaniline	ug/L	1.2U	5.0	10/21/10 17:47	
4-Chlorophenylphenyl ether	ug/L	0.63U	5.0	10/21/10 17:47	
4-Nitroaniline	ug/L	0.69U	4.0	10/21/10 17:47	
4-Nitrophenol	ug/L	1.1U	20.0	10/21/10 17:47	
5-Nitro-o-toluidine	ug/L	1.3U	5.0	10/21/10 17:47	
7,12-Dimethylbenz(a)anthracene	ug/L	2.0U	5.0	10/21/10 17:47	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 272 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 130562

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
a,a-Dimethylphenylethylamine	ug/L	10.0U	20.0	10/21/10 17:47	
Acenaphthene	ug/L	0.86U	5.0	10/21/10 17:47	
Acenaphthylene	ug/L	0.95U	5.0	10/21/10 17:47	
Acetophenone	ug/L	1.4U	5.0	10/21/10 17:47	
Anthracene	ug/L	0.60U	5.0	10/21/10 17:47	
Benzo(a)anthracene	ug/L	0.63U	5.0	10/21/10 17:47	
Benzo(a)pyrene	ug/L	0.58U	1.0	10/21/10 17:47	
Benzo(b)fluoranthene	ug/L	0.62U	2.0	10/21/10 17:47	
Benzo(g,h,i)perylene	ug/L	0.68U	5.0	10/21/10 17:47	
Benzo(k)fluoranthene	ug/L	0.51U	4.0	10/21/10 17:47	
Benzyl alcohol	ug/L	1.0U	5.0	10/21/10 17:47	
bis(2-Chloroethoxy)methane	ug/L	3.0U	5.0	10/21/10 17:47	
bis(2-Chloroethyl) ether	ug/L	0.75U	4.0	10/21/10 17:47	
bis(2-Chloroisopropyl) ether	ug/L	0.73U	5.0	10/21/10 17:47	
bis(2-Ethylhexyl)phthalate	ug/L	0.80U	5.0	10/21/10 17:47	
Butylbenzylphthalate	ug/L	0.72U	5.0	10/21/10 17:47	
Chrysene	ug/L	0.37U	5.0	10/21/10 17:47	
Di-n-butylphthalate	ug/L	0.41U	5.0	10/21/10 17:47	
Di-n-octylphthalate	ug/L	0.90U	5.0	10/21/10 17:47	
Diallate	ug/L	0.73U	5.0	10/21/10 17:47	
Dibenz(a,h)anthracene	ug/L	0.65U	2.0	10/21/10 17:47	
Dibenzofuran	ug/L	0.67U	5.0	10/21/10 17:47	
Diethylphthalate	ug/L	0.51U	5.0	10/21/10 17:47	
Dimethylphthalate	ug/L	0.64U	5.0	10/21/10 17:47	
Ethyl methanesulfonate	ug/L	0.90U	5.0	10/21/10 17:47	
Fluoranthene	ug/L	0.54U	5.0	10/21/10 17:47	
Fluorene	ug/L	0.56U	5.0	10/21/10 17:47	
Hexachlorobenzene	ug/L	0.80U	1.0	10/21/10 17:47	
Hexachlorocyclopentadiene	ug/L	1.3U	5.0	10/21/10 17:47	
Hexachloroethane	ug/L	0.71U	5.0	10/21/10 17:47	
Hexachloropropene	ug/L	1.4U	5.0	10/21/10 17:47	
Indeno(1,2,3-cd)pyrene	ug/L	0.73U	2.0	10/21/10 17:47	
Isodrin	ug/L	0.54U	5.0	10/21/10 17:47	
Isophorone	ug/L	0.73U	5.0	10/21/10 17:47	
Isosafrole	ug/L	0.60U	5.0	10/21/10 17:47	
Kepone	ug/L	10.0U	20.0	10/21/10 17:47	
Methapyrilene	ug/L	1.6U	5.0	10/21/10 17:47	J(SS)
Methyl methanesulfonate	ug/L	1.0U	5.0	10/21/10 17:47	
N-Nitroso-di-n-butylamine	ug/L	0.55U	4.0	10/21/10 17:47	
N-Nitroso-di-n-propylamine	ug/L	0.94U	4.0	10/21/10 17:47	
N-Nitrosodiethylamine	ug/L	0.73U	4.0	10/21/10 17:47	
N-Nitrosodimethylamine	ug/L	0.97U	2.0	10/21/10 17:47	
N-Nitrosodiphenylamine	ug/L	0.50U	5.0	10/21/10 17:47	
N-Nitrosomethylethylamine	ug/L	0.74U	5.0	10/21/10 17:47	
N-Nitrosopiperidine	ug/L	0.64U	5.0	10/21/10 17:47	
N-Nitrosopyrrolidine	ug/L	0.88U	5.0	10/21/10 17:47	
Naphthalene	ug/L	0.78U	5.0	10/21/10 17:47	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 273 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 130562

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrobenzene	ug/L	1.1U	4.0	10/21/10 17:47	
O,O,O-Triethylphosphorothioate	ug/L	0.69U	5.0	10/21/10 17:47	
O-Toluidine	ug/L	1.1U	5.0	10/21/10 17:47	
P-Dimethylaminoazobenzene	ug/L	0.67U	5.0	10/21/10 17:47	J(SS)
p-Phenylenediamine	ug/L	10.0U	20.0	10/21/10 17:47	
Parathion (Ethyl parathion)	ug/L	1.2U	5.0	10/21/10 17:47	
Pentachlorobenzene	ug/L	0.78U	5.0	10/21/10 17:47	
Pentachlorophenol	ug/L	0.66U	20.0	10/21/10 17:47	
Phenacetin	ug/L	0.53U	5.0	10/21/10 17:47	
Phenanthrene	ug/L	0.52U	5.0	10/21/10 17:47	
Phenol	ug/L	0.54U	5.0	10/21/10 17:47	
Pronamide	ug/L	1.1U	5.0	10/21/10 17:47	
Pyrene	ug/L	0.68U	5.0	10/21/10 17:47	
Safrole	ug/L	0.85U	5.0	10/21/10 17:47	
Thionazin	ug/L	0.61U	5.0	10/21/10 17:47	
2,4,6-Tribromophenol (S)	%	72	10-110	10/21/10 17:47	
2-Fluorobiphenyl (S)	%	71	18-110	10/21/10 17:47	
2-Fluorophenol (S)	%	41	18-110	10/21/10 17:47	
Nitrobenzene-d5 (S)	%	65	10-110	10/21/10 17:47	
Phenol-d6 (S)	%	28	10-110	10/21/10 17:47	
Terphenyl-d14 (S)	%	84	10-123	10/21/10 17:47	

LABORATORY CONTROL SAMPLE: 130563

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	50	41.7	83	10-146.9	
1,2,4-Trichlorobenzene	ug/L	50	39.1	78	19.7-141.6	
1,2-Dichlorobenzene	ug/L	50	37.5	75	23.5-105.7	
1,2-Dinitrobenzene	ug/L	50	47.9	96	52.6-113.9	
1,3,5-Trinitrobenzene	ug/L	50	52.2	104	41.4-102.4	L3
1,3-Dichlorobenzene	ug/L	50	37.1	74	25.5-94.5	
1,3-Dinitrobenzene	ug/L	50	47.4	95	45.3-116.4	
1,4-Dichlorobenzene	ug/L	50	37.7	75	33.2-90.7	
1,4-Naphthoquinone	ug/L	50	50.0	100	39.3-113.3	
1-Methylnaphthalene	ug/L	50	41.2	82	27.8-110.7	
1-Naphthylamine	ug/L	50	48.6	97	37.1-90.4	L3
2,3,4,6-Tetrachlorophenol	ug/L	50	49.2	98	14.3-115.3	
2,4,5-Trichlorophenol	ug/L	50	46.3	93	10-121.3	
2,4,6-Trichlorophenol	ug/L	50	45.5	91	40.3-101.7	
2,4-Dichlorophenol	ug/L	50	42.9	86	35.8-108.5	
2,4-Dimethylphenol	ug/L	50	40.7	81	25-104.5	
2,4-Dinitrophenol	ug/L	50	51.3	103	10-147.1	
2,4-Dinitrotoluene	ug/L	50	49.5	99	47.9-113.6	
2,6-Dichlorophenol	ug/L	50	42.5	85	41.1-101.8	
2,6-Dinitrotoluene	ug/L	50	48.2	96	44.6-111.5	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130563

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2-Acetylaminofluorene	ug/L	50	48.1	96	58.3-112.7	
2-Chloronaphthalene	ug/L	50	42.5	85	41.2-101.2	
2-Chlorophenol	ug/L	50	37.9	76	32.1-96.5	
2-Methylnaphthalene	ug/L	50	41.0	82	40-93.6	
2-Methylphenol(o-Cresol)	ug/L	50	35.3	71	36.7-95.1	
2-Naphthylamine	ug/L	50	36.8	74	42.4-98.8	
2-Nitroaniline	ug/L	50	46.9	94	59-103.2	
2-Nitrophenol	ug/L	50	41.6	83	38.4-108.9	
3&4-Methylphenol(m&p Cresol)	ug/L	50	34.6	69	35.2-94.3	
3,3'-Dichlorobenzidine	ug/L	50	46.8	94	30.7-106	
3,3'-Dimethylbenzidine	ug/L	50	31.4	63	10-160.1	
3-Methylcholanthrene	ug/L	50	44.7	89	15.5-121.4	
3-Nitroaniline	ug/L	50	48.0	96	25.3-131.5	
4,6-Dinitro-2-methylphenol	ug/L	50	54.4	109	35.2-130.5	
4-Aminobiphenyl	ug/L	50	36.9	74	50.8-112.6	
4-Bromophenylphenyl ether	ug/L	50	44.6	89	51.9-110.4	
4-Chloro-3-methylphenol	ug/L	50	46.5	93	19.4-128.8	
4-Chloroaniline	ug/L	50	46.7	93	30.1-108.4	
4-Chlorophenylphenyl ether	ug/L	50	46.3	93	49.7-91.5	L3
4-Nitroaniline	ug/L	50	51.5	103	48.1-112.2	
4-Nitrophenol	ug/L	50	22.0	44	10-121.8	
5-Nitro-o-toluidine	ug/L	50	48.8	98	43-113	
7,12-Dimethylbenz(a)anthracene	ug/L	50	46.8	94	52.5-108.6	
Acenaphthene	ug/L	50	45.5	91	50.3-98.3	
Acenaphthylene	ug/L	50	44.6	89	49-98.1	
Acetophenone	ug/L	50	43.4	87	40.6-94.3	
Anthracene	ug/L	50	46.0	92	55-112.5	
Benzo(a)anthracene	ug/L	50	47.0	94	10-150.1	
Benzo(a)pyrene	ug/L	50	48.7	97	59.7-108.4	
Benzo(b)fluoranthene	ug/L	50	50.5	101	58.4-111.8	
Benzo(g,h,i)perylene	ug/L	50	44.6	89	57.6-115.1	
Benzo(k)fluoranthene	ug/L	50	50.1	100	57.6-112.4	
Benzyl alcohol	ug/L	50	37.3	75	26.1-118	
bis(2-Chloroethoxy)methane	ug/L	50	44.0	88	41.2-96.2	
bis(2-Chloroethyl) ether	ug/L	50	39.5	79	35.3-99.5	
bis(2-Chloroisopropyl) ether	ug/L	50	40.2	80	36.3-91	
bis(2-Ethylhexyl)phthalate	ug/L	50	50.4	101	43.1-118.3	
Butylbenzylphthalate	ug/L	50	50.1	100	57.5-118.2	
Chrysene	ug/L	50	47.5	95	42.4-113.9	
Di-n-butylphthalate	ug/L	50	52.4	105	22.2-139.3	
Di-n-octylphthalate	ug/L	50	47.7	95	57.4-116.9	
Diallate	ug/L	50	56.2	112	44.3-111.7	L3
Dibenz(a,h)anthracene	ug/L	50	46.1	92	59.1-111.8	
Dibenzofuran	ug/L	50	45.0	90	45.3-108.3	
Diethylphthalate	ug/L	50	51.0	102	51.1-107.5	
Dimethylphthalate	ug/L	50	48.0	96	47.4-110.6	
Ethyl methanesulfonate	ug/L	50	38.5	77	35.9-103.6	
Fluoranthene	ug/L	50	49.5	99	48.2-118.6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 275 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130563

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Fluorene	ug/L	50	46.2	92	44.7-106.4	
Hexachlorobenzene	ug/L	50	44.4	89	54-113.2	
Hexachlorocyclopentadiene	ug/L	50	36.8	74	16.5-105.1	
Hexachloroethane	ug/L	50	39.9	80	10-102	
Hexachloropropene	ug/L	50	39.4	79	29.1-84.2	
Indeno(1,2,3-cd)pyrene	ug/L	50	45.3	91	33.7-120.7	
Isodrin	ug/L	50	47.0	94	32.4-130.4	
Isophorone	ug/L	50	44.6	89	42.5-107.7	
Isosafrole	ug/L	50	41.3	83	45.8-99.3	
Methapyrilene	ug/L	50	37.9	76	17.8-119.5	J(SS)
Methyl methanesulfonate	ug/L	50	31.4	63	10-107	
N-Nitroso-di-n-butylamine	ug/L	50	41.1	82	15.2-107.9	
N-Nitroso-di-n-propylamine	ug/L	50	42.6	85	19.1-111.6	
N-Nitrosodiethylamine	ug/L	50	36.0	72	10-130.6	
N-Nitrosodimethylamine	ug/L	50	25.1	50	10-132	
N-Nitrosodiphenylamine	ug/L	50	44.1	88	37-104.4	
N-Nitrosomethylethylamine	ug/L	50	33.4	67	10-135	
N-Nitrosopiperidine	ug/L	50	40.1	80	43.3-96.3	
N-Nitrosopyrrolidine	ug/L	50	39.4	79	43.1-97.2	
Naphthalene	ug/L	50	40.2	80	40.1-85.7	
Nitrobenzene	ug/L	50	41.9	84	32.9-115.9	
O,O,O-Triethylphosphorothioate	ug/L	50	41.3	83	48.5-99.9	
O-Toluidine	ug/L	50	42.1	84	21.2-134.1	
P-Dimethylaminoazobenzene	ug/L	50	55.8	112	44.6-142.5	J(SS)
Parathion (Ethyl parathion)	ug/L	50	54.1	108	46.8-113.9	
Pentachlorobenzene	ug/L	50	45.5	91	37.5-128.1	
Pentachlorophenol	ug/L	50	43.2	86	44.6-115.6	
Phenacetin	ug/L	50	53.8	108	19.3-143.2	
Phenanthrene	ug/L	50	46.6	93	49.2-124.2	
Phenol	ug/L	50	18.0	36	10-158.5	
Pronamide	ug/L	50	51.5	103	10-128.9	
Pyrene	ug/L	50	45.2	90	10-150.1	
Safrole	ug/L	50	42.7	85	10-135.9	
Thionazin	ug/L	50	47.2	94	45-105.7	
2,4,6-Tribromophenol (S)	%			97	10-110	
2-Fluorobiphenyl (S)	%			87	18-110	
2-Fluorophenol (S)	%			47	18-110	
Nitrobenzene-d5 (S)	%			89	10-110	
Phenol-d6 (S)	%			34	10-110	
Terphenyl-d14 (S)	%			98	10-123	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130669 130670

Parameter	Units	3520108018 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2,4,5-Tetrachlorobenzene	ug/L	0.72U	100	100	82.1	70.6	82	71	10-146.9	15	40	
1,2,4-Trichlorobenzene	ug/L	0.85U	100	100	65.6	56.8	66	57	19.7-141	14	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 276 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130669 130670											
Parameter	Units	3520108018 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
1,2-Dichlorobenzene	ug/L	0.70U	100	100	64.9	56.1	65	56	23.5-105	15	40
1,2-Dinitrobenzene	ug/L	1.2U	100	100	106	95.4	106	95	52.6-113	11	40
1,3,5-Trinitrobenzene	ug/L	1.3U	100	100	94.7	93.4	95	93	41.4-102	1	40
1,3-Dichlorobenzene	ug/L	0.78U	100	100	60.6	53.4	61	53	25.5-94.	13	40
1,3-Dinitrobenzene	ug/L	0.70U	100	100	98.7	89.9	99	90	45.3-116	9	40
1,4-Dichlorobenzene	ug/L	2.0 U	100	100	64.7	56.5	63	54	33.2-90.	14	40
1,4-Naphthoquinone	ug/L	1.2U	100	100	91.0	83.7	91	84	39.3-113	8	40
1-Methylnaphthalene	ug/L	1.0U	100	100	82.1	73.2	82	73	27.8-110	12	40
1-Naphthylamine	ug/L	1.1U	100	100	124	120	124	119	37.1-90.	4	40 J(M0)
2,3,4,6-Tetrachlorophenol	ug/L	4.0U	100	100	102	90.4	102	90	14.3-115	12	40
2,4,5-Trichlorophenol	ug/L	0.53U	100	100	90.8	85.2	91	85	10-121.3	6	40
2,4,6-Trichlorophenol	ug/L	0.71U	100	100	88.6	81.9	89	82	40.3-101	8	40
2,4-Dichlorophenol	ug/L	0.58U	100	100	88.6	80.7	89	81	35.8-108	9	40
2,4-Dimethylphenol	ug/L	1.6U	100	100	92.5	85.4	93	85	25-104.5	8	40
2,4-Dinitrophenol	ug/L	1.6U	100	100	107	68.7	107	69	10-147.1	44	40 J(D6)
2,4-Dinitrotoluene	ug/L	0.55U	100	100	111	94.3	111	94	47.9-113	16	40
2,6-Dichlorophenol	ug/L	0.64U	100	100	87.8	79.1	88	79	41.1-101	10	40
2,6-Dinitrotoluene	ug/L	1.3U	100	100	99.1	88.9	99	89	44.6-111	11	40
2-Acetylaminofluorene	ug/L	0.67U	100	100	67.6	75.3	68	75	58.3-112	11	40
2-Chloronaphthalene	ug/L	0.82U	100	100	76.2	74.3	76	74	41.2-101	2	40
2-Chlorophenol	ug/L	0.70U	100	100	78.2	71.9	78	72	32.1-96.	8	40
2-Methylnaphthalene	ug/L	1.0U	100	100	81.6	70.3	81	70	40-93.6	15	40
2-Methylphenol(o-Cresol)	ug/L	0.75U	100	100	85.1	81.0	85	81	36.7-95.	5	40
2-Naphthylamine	ug/L	2.3U	100	100	77.1	82.5	77	82	42.4-98.	7	40
2-Nitroaniline	ug/L	0.62U	100	100	97.2	88.7	97	89	59-103.2	9	40
2-Nitrophenol	ug/L	0.83U	100	100	82.6	78.2	83	78	38.4-108	5	40
3&4-Methylphenol(m&p Cresol)	ug/L	0.68U	100	100	90.1	81.7	90	82	35.2-94.	10	40
3,3'-Dichlorobenzidine	ug/L	0.71U	100	100	67.5	78.6	68	79	30.7-106	15	40
3,3'-Dimethylbenzidine	ug/L	3.2U	100	100	24.0	71.1	24	71	10-160.1	99	40 J(D6)
3-Methylcholanthrene	ug/L	1.1U	100	100	89.2	87.8	89	88	15.5-121	2	40
3-Nitroaniline	ug/L	1.0U	100	100	108	92.6	108	93	25.3-131	16	40
4,6-Dinitro-2-methylphenol	ug/L	1.4U	100	100	104	103	104	103	35.2-130	.7	40
4-Aminobiphenyl	ug/L	2.9U	100	100	93.7	108	94	108	50.8-112	15	40
4-Bromophenylphenyl ether	ug/L	0.69U	100	100	88.5	86.0	88	86	51.9-110	3	40
4-Chloro-3-methylphenol	ug/L	0.64U	100	100	117	92.5	117	93	19.4-128	23	40
4-Chloroaniline	ug/L	1.2U	100	100	96.7	82.5	97	82	30.1-108	16	40
4-Chlorophenylphenyl ether	ug/L	0.65U	100	100	94.1	84.3	94	84	49.7-91.	11	40 J(M0)
4-Nitroaniline	ug/L	0.71U	100	100	123	99.5	123	100	48.1-112	21	40 J(M1)
4-Nitrophenol	ug/L	1.1U	100	100	91.2	64.5	91	65	10-121.8	34	40
5-Nitro-o-toluidine	ug/L	1.3U	100	100	109	91.8	109	92	43-113	17	40
7,12-Dimethylbenz(a)anthracene	ug/L	2.0U	100	100	98.8	90.1	99	90	52.5-108	9	40
Acenaphthene	ug/L	0.88U	100	100	87.2	78.9	87	79	50.3-98.	10	40
Acenaphthylene	ug/L	0.98U	100	100	85.7	81.2	86	81	49-98.1	5	40
Acetophenone	ug/L	1.5U	100	100	84.3	78.1	84	78	40.6-94.	8	40
Anthracene	ug/L	0.62U	100	100	91.1	88.3	91	88	55-112.5	3	40
Benzo(a)anthracene	ug/L	0.65U	100	100	91.8	86.5	92	86	10-150.1	6	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 277 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130669 130670											
Parameter	Units	3520108018 Result	MS		MSD		MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Conc.	Spike Conc.	Result				RPD	RPD
Benzo(a)pyrene	ug/L	0.60U	100	100	100	98.5	92.7	98	93 59.7-108	6	40
Benzo(b)fluoranthene	ug/L	0.64U	100	100	100	102	94.7	102	95 58.4-111	7	40
Benzo(g,h,i)perylene	ug/L	0.70U	100	100	100	87.0	86.9	87	87 57.6-115	.2	40
Benzo(k)fluoranthene	ug/L	0.52U	100	100	100	100	92.3	100	92 57.6-112	8	40
Benzyl alcohol	ug/L	1.0U	100	100	100	85.7	80.3	86	80 26.1-118	7	40
bis(2-Chloroethoxy)methane	ug/L	3.0U	100	100	100	83.2	77.8	83	78 41.2-96.	7	40
bis(2-Chloroethyl) ether	ug/L	0.77U	100	100	100	79.0	72.7	79	73 35.3-99.	8	40
bis(2-Chloroisopropyl) ether	ug/L	0.75U	100	100	100	80.3	73.2	80	73 36.3-91	9	40
bis(2-Ethylhexyl)phthalate	ug/L	0.82U	100	100	100	102	93.8	102	94 43.1-118	9	40
Butylbenzylphthalate	ug/L	0.74U	100	100	100	104	97.6	104	98 57.5-118	6	40
Chrysene	ug/L	0.38U	100	100	100	91.3	87.0	91	87 42.4-113	5	40
Di-n-butylphthalate	ug/L	0.42U	100	100	100	104	98.1	104	98 22.2-139	6	40
Di-n-octylphthalate	ug/L	0.93U	100	100	100	91.2	85.8	91	86 57.4-116	6	40
Diallylate	ug/L	0.75U	100	100	100	124	107	124	107 44.3-111	15	40 J(M0)
Dibenz(a,h)anthracene	ug/L	0.67U	100	100	100	91.4	91.2	91	91 59.1-111	.2	40
Dibenzofuran	ug/L	0.69U	100	100	100	91.2	83.2	91	83 45.3-108	9	40
Diethylphthalate	ug/L	0.52U	100	100	100	112	97.6	112	97 51.1-107	13	40 J(M1)
Dimethylphthalate	ug/L	0.66U	100	100	100	101	89.5	101	89 47.4-110	12	40
Ethyl methanesulfonate	ug/L	0.93U	100	100	100	81.2	75.2	81	75 35.9-103	8	40
Fluoranthene	ug/L	0.56U	100	100	100	96.7	89.7	97	90 48.2-118	8	40
Fluorene	ug/L	0.58U	100	100	100	95.9	86.6	96	87 44.7-106	10	40
Hexachlorobenzene	ug/L	0.82U	100	100	100	82.7	82.3	83	82 54-113.2	.4	40
Hexachlorocyclopentadiene	ug/L	1.3U	100	100	100	51.6	48.1	52	48 16.5-105	7	40
Hexachloroethane	ug/L	0.73U	100	100	100	61.6	52.0	62	52 10-102	17	40
Hexachloropropene	ug/L	1.5U	100	100	100	60.1	51.2	60	51 29.1-84.	16	40
Indeno(1,2,3-cd)pyrene	ug/L	0.75U	100	100	100	89.0	90.1	89	90 33.7-120	1	40
Isodrin	ug/L	0.56U	100	100	100	89.6	83.5	90	84 32.4-130	7	40
Isophorone	ug/L	0.75U	100	100	100	89.1	82.6	89	83 42.5-107	8	40
Isosafrole	ug/L	0.62U	100	100	100	83.5	71.4	83	71 45.8-99.	16	40
Methapyrilene	ug/L	1.7U	100	100	100	72.1	84.0	72	84 17.8-119	15	40 J(SS)
Methyl methanesulfonate	ug/L	1.0U	100	100	100	77.6	71.2	78	71 10-107	9	40
N-Nitroso-di-n-butylamine	ug/L	0.57U	100	100	100	79.5	75.0	79	75 15.2-107	6	40
N-Nitroso-di-n-propylamine	ug/L	0.97U	100	100	100	82.0	80.2	82	80 19.1-111	2	40
N-Nitrosodiethylamine	ug/L	0.75U	100	100	100	65.7	69.9	66	70 10-130.6	6	40
N-Nitrosodimethylamine	ug/L	1.0U	100	100	100	65.3	59.6	65	60 10-132	9	40
N-Nitrosodiphenylamine	ug/L	0.51U	100	100	100	90.8	87.2	91	87 37-104.4	4	40
N-Nitrosomethylethylamine	ug/L	0.76U	100	100	100	66.9	65.7	67	66 10-135	2	40
N-Nitrosopiperidine	ug/L	0.66U	100	100	100	75.7	76.0	76	76 43.3-96.	.5	40
N-Nitrosopyrrolidine	ug/L	0.91U	100	100	100	71.7	76.4	72	76 43.1-97.	6	40
Naphthalene	ug/L	0.80U	100	100	100	73.6	63.4	74	63 40.1-85.	15	40
Nitrobenzene	ug/L	1.1U	100	100	100	79.6	73.5	80	74 32.9-115	8	40
O,O,O-Triethylphosphorothioate	ug/L	0.71U	100	100	100	77.1	70.3	77	70 48.5-99.	9	40
O-Toluidine	ug/L	1.1U	100	100	100	79.9	77.5	80	77 21.2-134	3	40
P-Dimethylaminoazobenzene	ug/L	0.69U	100	100	100	102	101	102	101 44.6-142	1	40 J(SS)
Parathion (Ethyl parathion)	ug/L	1.2U	100	100	100	107	103	107	103 46.8-113	4	40
Pentachlorobenzene	ug/L	0.80U	100	100	100	86.2	80.8	86	81 37.5-128	6	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 278 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE:											
			130669			130670					
Parameter	Units	3520108018 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Pentachlorophenol	ug/L	0.68U	100	100	95.9	84.0	96	84	44.6-115	13	40
Phenacetin	ug/L	0.55U	100	100	108	102	108	102	19.3-143	5	40
Phenanthrene	ug/L	0.53U	100	100	90.5	87.0	90	87	49.2-124	4	40
Phenol	ug/L	0.56U	100	100	68.3	56.2	68	56	10-158.5	19	40 8p
Pronamide	ug/L	1.2U	100	100	103	96.8	103	97	10-128.9	6	40
Pyrene	ug/L	0.70U	100	100	98.9	93.5	99	94	10-150.1	6	40
Safrole	ug/L	0.87U	100	100	76.3	78.5	76	79	10-135.9	3	40
Thionazin	ug/L	0.63U	100	100	94.3	90.5	94	91	45-105.7	4	40
2,4,6-Tribromophenol (S)	%						105	89	10-110		
2-Fluorobiphenyl (S)	%						73	78	18-110		
2-Fluorophenol (S)	%						65	57	18-110		
Nitrobenzene-d5 (S)	%						83	77	10-110		
Phenol-d6 (S)	%						65	53	10-110		
Terphenyl-d14 (S)	%						105	101	10-123		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3361

Analysis Method: EPA 8270

QC Batch Method: EPA 3510

Analysis Description: 8270 Water MSSV App II

Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 132184

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	0.70U	5.0	10/30/10 06:08	
1,2,4-Trichlorobenzene	ug/L	0.83U	5.0	10/30/10 06:08	
1,2-Dichlorobenzene	ug/L	0.68U	5.0	10/30/10 06:08	
1,2-Dinitrobenzene	ug/L	1.2U	5.0	10/30/10 06:08	
1,3,5-Trinitrobenzene	ug/L	1.2U	5.0	10/30/10 06:08	
1,3-Dichlorobenzene	ug/L	0.76U	5.0	10/30/10 06:08	
1,3-Dinitrobenzene	ug/L	0.68U	8.0	10/30/10 06:08	
1,4-Dichlorobenzene	ug/L	0.77U	5.0	10/30/10 06:08	
1,4-Naphthoquinone	ug/L	1.2U	5.0	10/30/10 06:08	
1-Methylnaphthalene	ug/L	1.0U	5.0	10/30/10 06:08	
1-Naphthylamine	ug/L	1.0U	5.0	10/30/10 06:08	
2,3,4,6-Tetrachlorophenol	ug/L	3.8U	5.0	10/30/10 06:08	
2,4,5-Trichlorophenol	ug/L	0.52U	4.0	10/30/10 06:08	
2,4,6-Trichlorophenol	ug/L	0.69U	2.0	10/30/10 06:08	
2,4-Dichlorophenol	ug/L	0.56U	2.0	10/30/10 06:08	
2,4-Dimethylphenol	ug/L	1.6U	5.0	10/30/10 06:08	
2,4-Dinitrophenol	ug/L	1.6U	20.0	10/30/10 06:08	
2,4-Dinitrotoluene	ug/L	0.53U	2.0	10/30/10 06:08	
2,6-Dichlorophenol	ug/L	0.62U	4.0	10/30/10 06:08	
2,6-Dinitrotoluene	ug/L	1.2U	2.0	10/30/10 06:08	
2-Acetylaminofluorene	ug/L	0.65U	5.0	10/30/10 06:08	
2-Chloronaphthalene	ug/L	0.80U	5.0	10/30/10 06:08	
2-Chlorophenol	ug/L	0.68U	5.0	10/30/10 06:08	
2-Methylnaphthalene	ug/L	0.99U	5.0	10/30/10 06:08	
2-Methylphenol(o-Cresol)	ug/L	0.73U	5.0	10/30/10 06:08	
2-Naphthylamine	ug/L	2.3U	5.0	10/30/10 06:08	
2-Nitroaniline	ug/L	0.60U	5.0	10/30/10 06:08	
2-Nitrophenol	ug/L	0.81U	5.0	10/30/10 06:08	
3&4-Methylphenol(m&p Cresol)	ug/L	0.66U	10.0	10/30/10 06:08	
3,3'-Dichlorobenzidine	ug/L	0.69U	10.0	10/30/10 06:08	
3,3'-Dimethylbenzidine	ug/L	3.1U	10.0	10/30/10 06:08	
3-Methylcholanthrene	ug/L	1.0U	5.0	10/30/10 06:08	
3-Nitroaniline	ug/L	0.99U	5.0	10/30/10 06:08	
4,6-Dinitro-2-methylphenol	ug/L	1.3U	20.0	10/30/10 06:08	
4-Aminobiphenyl	ug/L	2.8U	5.0	10/30/10 06:08	
4-Bromophenylphenyl ether	ug/L	0.67U	5.0	10/30/10 06:08	
4-Chloro-3-methylphenol	ug/L	0.62U	20.0	10/30/10 06:08	
4-Chloroaniline	ug/L	1.2U	5.0	10/30/10 06:08	
4-Chlorophenylphenyl ether	ug/L	0.63U	5.0	10/30/10 06:08	
4-Nitroaniline	ug/L	0.69U	4.0	10/30/10 06:08	
4-Nitrophenol	ug/L	1.1U	20.0	10/30/10 06:08	
5-Nitro-o-toluidine	ug/L	1.3U	5.0	10/30/10 06:08	
7,12-Dimethylbenz(a)anthracene	ug/L	2.0U	5.0	10/30/10 06:08	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 280 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 132184

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Acenaphthene	ug/L	0.86U	5.0	10/30/10 06:08	
Acenaphthylene	ug/L	0.95U	5.0	10/30/10 06:08	
Acetophenone	ug/L	1.4U	5.0	10/30/10 06:08	
Anthracene	ug/L	0.60U	5.0	10/30/10 06:08	
Benzo(a)anthracene	ug/L	0.63U	5.0	10/30/10 06:08	
Benzo(a)pyrene	ug/L	0.58U	1.0	10/30/10 06:08	
Benzo(b)fluoranthene	ug/L	0.62U	2.0	10/30/10 06:08	
Benzo(g,h,i)perylene	ug/L	0.68U	5.0	10/30/10 06:08	
Benzo(k)fluoranthene	ug/L	0.51U	4.0	10/30/10 06:08	
Benzyl alcohol	ug/L	1.0U	5.0	10/30/10 06:08	
bis(2-Chloroethoxy)methane	ug/L	3.0U	5.0	10/30/10 06:08	
bis(2-Chloroethyl) ether	ug/L	0.75U	4.0	10/30/10 06:08	
bis(2-Chloroisopropyl) ether	ug/L	0.73U	5.0	10/30/10 06:08	
bis(2-Ethylhexyl)phthalate	ug/L	0.80U	5.0	10/30/10 06:08	
Butylbenzylphthalate	ug/L	0.72U	5.0	10/30/10 06:08	
Chrysene	ug/L	0.37U	5.0	10/30/10 06:08	
Di-n-butylphthalate	ug/L	0.41U	5.0	10/30/10 06:08	
Di-n-octylphthalate	ug/L	0.90U	5.0	10/30/10 06:08	
Diallylate	ug/L	0.73U	5.0	10/30/10 06:08	
Dibenz(a,h)anthracene	ug/L	0.65U	2.0	10/30/10 06:08	
Dibenzofuran	ug/L	0.67U	5.0	10/30/10 06:08	
Diethylphthalate	ug/L	0.51U	5.0	10/30/10 06:08	
Dimethylphthalate	ug/L	0.64U	5.0	10/30/10 06:08	
Ethyl methanesulfonate	ug/L	0.90U	5.0	10/30/10 06:08	
Fluoranthene	ug/L	0.54U	5.0	10/30/10 06:08	
Fluorene	ug/L	0.56U	5.0	10/30/10 06:08	
Hexachlorobenzene	ug/L	0.80U	1.0	10/30/10 06:08	
Hexachlorocyclopentadiene	ug/L	1.3U	5.0	10/30/10 06:08	
Hexachloroethane	ug/L	0.71U	5.0	10/30/10 06:08	
Hexachloropropene	ug/L	1.4U	5.0	10/30/10 06:08	
Indeno(1,2,3-cd)pyrene	ug/L	0.73U	2.0	10/30/10 06:08	
Isodrin	ug/L	0.54U	5.0	10/30/10 06:08	
Isophorone	ug/L	0.73U	5.0	10/30/10 06:08	
Isosafrole	ug/L	0.60U	5.0	10/30/10 06:08	
Methapyrilene	ug/L	1.6U	5.0	10/30/10 06:08	
Methyl methanesulfonate	ug/L	1.0U	5.0	10/30/10 06:08	
N-Nitroso-di-n-butylamine	ug/L	0.55U	4.0	10/30/10 06:08	
N-Nitroso-di-n-propylamine	ug/L	0.94U	4.0	10/30/10 06:08	
N-Nitrosodiethylamine	ug/L	0.73U	4.0	10/30/10 06:08	
N-Nitrosodimethylamine	ug/L	0.97U	2.0	10/30/10 06:08	
N-Nitrosodiphenylamine	ug/L	0.50U	5.0	10/30/10 06:08	
N-Nitrosomethylethylamine	ug/L	0.74U	5.0	10/30/10 06:08	
N-Nitrosopiperidine	ug/L	0.64U	5.0	10/30/10 06:08	
N-Nitrosopyrrolidine	ug/L	0.88U	5.0	10/30/10 06:08	
Naphthalene	ug/L	0.78U	5.0	10/30/10 06:08	
Nitrobenzene	ug/L	1.1U	4.0	10/30/10 06:08	
O,O,O-Triethylphosphorothioate	ug/L	0.69U	5.0	10/30/10 06:08	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 281 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 132184

Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
O-Toluidine	ug/L	1.1U	5.0	10/30/10 06:08	
P-Dimethylaminoazobenzene	ug/L	0.67U	5.0	10/30/10 06:08	
Parathion (Ethyl parathion)	ug/L	1.2U	5.0	10/30/10 06:08	
Pentachlorobenzene	ug/L	0.78U	5.0	10/30/10 06:08	
Pentachlorophenol	ug/L	0.66U	20.0	10/30/10 06:08	
Phenacetin	ug/L	0.53U	5.0	10/30/10 06:08	
Phenanthrene	ug/L	0.52U	5.0	10/30/10 06:08	
Phenol	ug/L	0.54U	5.0	10/30/10 06:08	
Pronamide	ug/L	1.1U	5.0	10/30/10 06:08	
Pyrene	ug/L	0.68U	5.0	10/30/10 06:08	
Safrole	ug/L	0.85U	5.0	10/30/10 06:08	
Thionazin	ug/L	0.61U	5.0	10/30/10 06:08	
2,4,6-Tribromophenol (S)	%	81	10-110	10/30/10 06:08	
2-Fluorobiphenyl (S)	%	77	18-110	10/30/10 06:08	
2-Fluorophenol (S)	%	43	18-110	10/30/10 06:08	
Nitrobenzene-d5 (S)	%	70	10-110	10/30/10 06:08	
Phenol-d6 (S)	%	29	10-110	10/30/10 06:08	
Terphenyl-d14 (S)	%	101	10-123	10/30/10 06:08	

LABORATORY CONTROL SAMPLE: 132185

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	50	43.6	87	10-146.9	
1,2,4-Trichlorobenzene	ug/L	50	41.9	84	19.7-141.6	
1,2-Dichlorobenzene	ug/L	50	38.9	78	23.5-105.7	
1,2-Dinitrobenzene	ug/L	50	80.4	161	52.6-113.9	L3
1,3,5-Trinitrobenzene	ug/L	50	47.4	95	41.4-102.4	
1,3-Dichlorobenzene	ug/L	50	38.5	77	25.5-94.5	
1,3-Dinitrobenzene	ug/L	50	49.6	99	45.3-116.4	
1,4-Dichlorobenzene	ug/L	50	38.4	77	33.2-90.7	
1,4-Naphthoquinone	ug/L	50	52.4	105	39.3-113.3	
1-Methylnaphthalene	ug/L	50	43.9	88	27.8-110.7	
1-Naphthylamine	ug/L	50	51.6	103	37.1-90.4	L3
2,3,4,6-Tetrachlorophenol	ug/L	50	53.6	107	14.3-115.3	
2,4,5-Trichlorophenol	ug/L	50	46.5	93	10-121.3	
2,4,6-Trichlorophenol	ug/L	50	48.5	97	40.3-101.7	
2,4-Dichlorophenol	ug/L	50	46.8	94	35.8-108.5	
2,4-Dimethylphenol	ug/L	50	45.1	90	25-104.5	
2,4-Dinitrophenol	ug/L	50	47.9	96	10-147.1	
2,4-Dinitrotoluene	ug/L	50	51.4	103	47.9-113.6	
2,6-Dichlorophenol	ug/L	50	45.7	91	41.1-101.8	
2,6-Dinitrotoluene	ug/L	50	49.6	99	44.6-111.5	
2-Acetylaminofluorene	ug/L	50	47.4	95	58.3-112.7	
2-Chloronaphthalene	ug/L	50	44.5	89	41.2-101.2	
2-Chlorophenol	ug/L	50	39.4	79	32.1-96.5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 282 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 132185

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2-Methylnaphthalene	ug/L	50	43.1	86	40-93.6	
2-Methylphenol(o-Cresol)	ug/L	50	39.6	79	36.7-95.1	
2-Naphthylamine	ug/L	50	40.9	82	42.4-98.8	
2-Nitroaniline	ug/L	50	48.9	98	59-103.2	
2-Nitrophenol	ug/L	50	44.2	88	38.4-108.9	
3&4-Methylphenol(m&p Cresol)	ug/L	50	37.5	75	35.2-94.3	
3,3'-Dichlorobenzidine	ug/L	50	49.4	99	30.7-106	
3,3'-Dimethylbenzidine	ug/L	50	47.4	95	10-160.1	
3-Methylcholanthrene	ug/L	50	49.6	99	15.5-121.4	
3-Nitroaniline	ug/L	50	48.6	97	25.3-131.5	
4,6-Dinitro-2-methylphenol	ug/L	50	47.9	96	35.2-130.5	
4-Aminobiphenyl	ug/L	50	48.4	97	50.8-112.6	
4-Bromophenylphenyl ether	ug/L	50	49.8	100	51.9-110.4	
4-Chloro-3-methylphenol	ug/L	50	46.5	93	19.4-128.8	
4-Chloroaniline	ug/L	50	46.8	94	30.1-108.4	
4-Chlorophenylphenyl ether	ug/L	50	47.6	95	49.7-91.5	L3
4-Nitroaniline	ug/L	50	48.2	96	48.1-112.2	
4-Nitrophenol	ug/L	50	20.5	41	10-121.8	
5-Nitro-o-toluidine	ug/L	50	48.3	97	43-113	
7,12-Dimethylbenz(a)anthracene	ug/L	50	45.0	90	52.5-108.6	
Acenaphthene	ug/L	50	47.4	95	50.3-98.3	
Acenaphthylene	ug/L	50	47.3	95	49-98.1	
Acetophenone	ug/L	50	44.3	89	40.6-94.3	
Anthracene	ug/L	50	50.3	101	55-112.5	
Benzo(a)anthracene	ug/L	50	48.8	98	10-150.1	
Benzo(a)pyrene	ug/L	50	51.6	103	59.7-108.4	
Benzo(b)fluoranthene	ug/L	50	51.4	103	58.4-111.8	
Benzo(g,h,i)perylene	ug/L	50	51.5	103	57.6-115.1	
Benzo(k)fluoranthene	ug/L	50	52.2	104	57.6-112.4	
Benzyl alcohol	ug/L	50	40.5	81	26.1-118	
bis(2-Chloroethoxy)methane	ug/L	50	45.6	91	41.2-96.2	
bis(2-Chloroethyl) ether	ug/L	50	41.2	82	35.3-99.5	
bis(2-Chloroisopropyl) ether	ug/L	50	41.4	83	36.3-91	
bis(2-Ethylhexyl)phthalate	ug/L	50	49.7	99	43.1-118.3	
Butylbenzylphthalate	ug/L	50	50.0	100	57.5-118.2	
Chrysene	ug/L	50	49.6	99	42.4-113.9	
Di-n-butylphthalate	ug/L	50	50.7	101	22.2-139.3	
Di-n-octylphthalate	ug/L	50	52.0	104	57.4-116.9	
Diallate	ug/L	50	48.6	97	44.3-111.7	
Dibenz(a,h)anthracene	ug/L	50	48.6	97	59.1-111.8	
Dibenzofuran	ug/L	50	48.1	96	45.3-108.3	
Diethylphthalate	ug/L	50	50.3	101	51.1-107.5	
Dimethylphthalate	ug/L	50	49.8	100	47.4-110.6	
Ethyl methanesulfonate	ug/L	50	41.6	83	35.9-103.6	
Fluoranthene	ug/L	50	50.0	100	48.2-118.6	
Fluorene	ug/L	50	47.9	96	44.7-106.4	
Hexachlorobenzene	ug/L	50	49.0	98	54-113.2	
Hexachlorocyclopentadiene	ug/L	50	38.0	76	16.5-105.1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 283 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 132185

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Hexachloroethane	ug/L	50	38.2	76	10-102	
Hexachloropropene	ug/L	50	38.9	78	29.1-84.2	
Indeno(1,2,3-cd)pyrene	ug/L	50	52.9	106	33.7-120.7	
Isodrin	ug/L	50	49.6	99	32.4-130.4	
Isophorone	ug/L	50	47.0	94	42.5-107.7	
Isosafrole	ug/L	50	44.2	88	45.8-99.3	
Methapyrilene	ug/L	50	33.0	66	17.8-119.5	
Methyl methanesulfonate	ug/L	50	33.0	66	10-107	
N-Nitroso-di-n-butylamine	ug/L	50	48.4	97	15.2-107.9	
N-Nitroso-di-n-propylamine	ug/L	50	45.9	92	19.1-111.6	
N-Nitrosodiethylamine	ug/L	50	42.8	86	10-130.6	
N-Nitrosodimethylamine	ug/L	50	26.8	54	10-132	
N-Nitrosodiphenylamine	ug/L	50	50.1	100	37-104.4	
N-Nitrosomethylethylamine	ug/L	50	39.0	78	10-135	
N-Nitrosopiperidine	ug/L	50	43.5	87	43.3-96.3	
N-Nitrosopyrrolidine	ug/L	50	44.2	88	43.1-97.2	
Naphthalene	ug/L	50	43.9	88	40.1-85.7	L3
Nitrobenzene	ug/L	50	41.0	82	32.9-115.9	
O,O,O-Triethylphosphorothioate	ug/L	50	43.7	87	48.5-99.9	
O-Toluidine	ug/L	50	41.9	84	21.2-134.1	
P-Dimethylaminoazobenzene	ug/L	50	50.6	101	44.6-142.5	
Parathion (Ethyl parathion)	ug/L	50	52.1	104	46.8-113.9	
Pentachlorobenzene	ug/L	50	46.5	93	37.5-128.1	
Pentachlorophenol	ug/L	50	51.7	103	44.6-115.6	
Phenacetin	ug/L	50	50.8	102	19.3-143.2	
Phenanthrene	ug/L	50	50.4	101	49.2-124.2	
Phenol	ug/L	50	18.9	38	10-158.5	
Pronamide	ug/L	50	50.2	100	10-128.9	
Pyrene	ug/L	50	48.0	96	10-150.1	
Safrole	ug/L	50	46.5	93	10-135.9	
Thionazin	ug/L	50	49.1	98	45-105.7	
2,4,6-Tribromophenol (S)	%			96	10-110	
2-Fluorobiphenyl (S)	%			82	18-110	
2-Fluorophenol (S)	%			46	18-110	
Nitrobenzene-d5 (S)	%			85	10-110	
Phenol-d6 (S)	%			35	10-110	
Terphenyl-d14 (S)	%			93	10-123	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132323 132324

Parameter	Units	3520572014 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2,4,5-Tetrachlorobenzene	ug/L	0.67U	100	100	78.7	75.7	79	76	10-146.9	4	40	
1,2,4-Trichlorobenzene	ug/L	0.80U	100	100	63.9	58.1	64	58	19.7-141	10	40	
1,2-Dichlorobenzene	ug/L	0.65U	100	100	53.8	48.5	54	48	23.5-105	10	40	
1,2-Dinitrobenzene	ug/L	1.1U	100	100	161	162	161	162	52.6-113	1	40	J(M0)
1,3,5-Trinitrobenzene	ug/L	1.2U	100	100	97.9	96.7	98	97	41.4-102	1	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 284 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132323 132324											
Parameter	Units	3520572014 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
1,3-Dichlorobenzene	ug/L	0.73U	100	100	50.4	44.0	50	44	25.5-94.	14	40
1,3-Dinitrobenzene	ug/L	0.65U	100	100	99.2	100	99	100	45.3-116	1	40
1,4-Dichlorobenzene	ug/L	0.74U	100	100	51.7	45.9	52	46	33.2-90.	12	40
1,4-Naphthoquinone	ug/L	1.1U	100	100	104	105	104	105	39.3-113	1	40
1-Methylnaphthalene	ug/L	0.96U	100	100	77.7	76.0	78	76	27.8-110	2	40
1-Naphthylamine	ug/L	0.99U	100	100	99.8	97.9	100	98	37.1-90.	2	40 J(M0)
2,3,4,6-Tetrachlorophenol	ug/L	3.7U	100	100	102	98.7	102	99	14.3-115	3	40
2,4,5-Trichlorophenol	ug/L	0.50U	100	100	99.9	99.6	100	100	10-121.3	.4	40
2,4,6-Trichlorophenol	ug/L	0.66U	100	100	96.4	98.8	96	99	40.3-101	3	40
2,4-Dichlorophenol	ug/L	0.54U	100	100	87.2	91.9	87	92	35.8-108	5	40
2,4-Dimethylphenol	ug/L	1.5U	100	100	81.8	86.6	82	87	25-104.5	6	40
2,4-Dinitrophenol	ug/L	1.5U	100	100	97.5	98.5	97	98	10-147.1	1	40
2,4-Dinitrotoluene	ug/L	0.51U	100	100	102	101	102	101	47.9-113	.4	40
2,6-Dichlorophenol	ug/L	0.60U	100	100	85.7	88.6	86	89	41.1-101	3	40
2,6-Dinitrotoluene	ug/L	1.2U	100	100	100	99.2	100	99	44.6-111	1	40
2-Acetylaminofluorene	ug/L	0.62U	100	100	95.1	93.3	95	93	58.3-112	2	40
2-Chloronaphthalene	ug/L	0.77U	100	100	83.1	81.4	83	81	41.2-101	2	40
2-Chlorophenol	ug/L	0.65U	100	100	76.3	76.8	76	77	32.1-96.	.6	40
2-Methylnaphthalene	ug/L	0.95U	100	100	74.8	73.5	75	74	40-93.6	2	40
2-Methylphenol(o-Cresol)	ug/L	0.70U	100	100	79.2	78.3	79	78	36.7-95.	1	40
2-Naphthylamine	ug/L	2.2U	100	100	81.7	79.1	82	79	42.4-98.	3	40
2-Nitroaniline	ug/L	0.58U	100	100	98.8	97.0	99	97	59-103.2	2	40
2-Nitrophenol	ug/L	0.78U	100	100	82.2	82.4	82	82	38.4-108	.2	40
3&4-Methylphenol(m&p Cresol)	ug/L	0.63U	100	100	81.5	82.9	82	83	35.2-94.	2	40
3,3'-Dichlorobenzidine	ug/L	0.66U	100	100	101	95.2	101	95	30.7-106	6	40
3,3'-Dimethylbenzidine	ug/L	3.0U	100	100	99.1	83.3	99	83	10-160.1	17	40
3-Methylcholanthrene	ug/L	1.0U	100	100	101	96.7	101	97	15.5-121	4	40
3-Nitroaniline	ug/L	0.95U	100	100	99.5	95.2	100	95	25.3-131	4	40
4,6-Dinitro-2-methylphenol	ug/L	1.3U	100	100	94.8	96.0	95	96	35.2-130	1	40
4-Aminobiphenyl	ug/L	2.7U	100	100	92.2	87.2	92	87	50.8-112	6	40
4-Bromophenylphenyl ether	ug/L	0.64U	100	100	98.4	99.7	98	100	51.9-110	1	40
4-Chloro-3-methylphenol	ug/L	0.60U	100	100	93.2	92.8	93	93	19.4-128	.4	40
4-Chloroaniline	ug/L	1.2U	100	100	85.3	84.9	85	85	30.1-108	.5	40
4-Chlorophenylphenyl ether	ug/L	0.60U	100	100	91.8	93.0	92	93	49.7-91.	1	40 J(M0)
4-Nitroaniline	ug/L	0.66U	100	100	98.4	93.2	98	93	48.1-112	5	40
4-Nitrophenol	ug/L	1.0U	100	100	65.2	63.6	65	64	10-121.8	2	40
5-Nitro-o-toluidine	ug/L	1.2U	100	100	98.2	95.2	98	95	43-113	3	40
7,12-Dimethylbenz(a)anthracene	ug/L	1.9U	100	100	94.2	90.5	94	91	52.5-108	4	40
Acenaphthene	ug/L	0.83U	100	100	93.0	89.0	93	89	50.3-98.	4	40
Acenaphthylene	ug/L	0.91U	100	100	91.7	90.0	92	90	49-98.1	2	40
Acetophenone	ug/L	1.4U	100	100	79.1	83.7	79	84	40.6-94.	6	40
Anthracene	ug/L	0.58U	100	100	100	101	100	101	55-112.5	1	40
Benzo(a)anthracene	ug/L	0.60U	100	100	97.0	95.2	97	95	10-150.1	2	40
Benzo(a)pyrene	ug/L	0.56U	100	100	1.2U	101	.03	101	59.7-108	40	J(M1)
Benzo(b)fluoranthene	ug/L	0.60U	100	100	105	100	105	100	58.4-111	4	40
Benzo(g,h,i)perylene	ug/L	0.65U	100	100	103	100	103	100	57.6-115	3	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 285 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132323 132324											
Parameter	Units	3520572014 Result	MS	MSD	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Spike Conc.						RPD	RPD
Benzo(k)fluoranthene	ug/L	0.49U	100	100	104	102	104	102	57.6-112	2	40
Benzyl alcohol	ug/L	0.98U	100	100	82.6	83.4	83	83	26.1-118	1	40
bis(2-Chloroethoxy)methane	ug/L	2.8U	100	100	83.2	81.0	83	81	41.2-96.	3	40
bis(2-Chloroethyl) ether	ug/L	0.72U	100	100	73.6	74.7	74	75	35.3-99.	1	40
bis(2-Chloroisopropyl) ether	ug/L	0.70U	100	100	75.4	74.4	75	74	36.3-91	1	40
bis(2-Ethylhexyl)phthalate	ug/L	0.77U	100	100	101	99.2	101	99	43.1-118	2	40
Butylbenzylphthalate	ug/L	0.69U	100	100	98.8	99.3	99	99	57.5-118	.5	40
Chrysene	ug/L	0.36U	100	100	98.7	96.8	99	97	42.4-113	2	40
Di-n-butylphthalate	ug/L	0.39U	100	100	102	101	102	101	22.2-139	1	40
Di-n-octylphthalate	ug/L	0.86U	100	100	105	103	104	103	57.4-116	2	40
Diallate	ug/L	0.70U	100	100	98.3	96.7	98	97	44.3-111	2	40
Dibenz(a,h)anthracene	ug/L	0.62U	100	100	99.3	95.6	99	96	59.1-111	4	40
Dibenzofuran	ug/L	0.64U	100	100	94.5	91.4	94	91	45.3-108	3	40
Diethylphthalate	ug/L	0.49U	100	100	99.8	98.4	100	98	51.1-107	1	40
Dimethylphthalate	ug/L	0.61U	100	100	95.5	97.2	95	97	47.4-110	2	40
Ethyl methanesulfonate	ug/L	0.86U	100	100	80.4	81.5	80	82	35.9-103	1	40
Fluoranthene	ug/L	0.52U	100	100	100	99.6	100	100	48.2-118	.5	40
Fluorene	ug/L	0.54U	100	100	94.8	92.8	95	93	44.7-106	2	40
Hexachlorobenzene	ug/L	0.77U	100	100	99.2	101	99	101	54-113.2	1	40
Hexachlorocyclopentadiene	ug/L	1.2U	100	100	64.6	62.7	65	63	16.5-105	3	40
Hexachloroethane	ug/L	0.68U	100	100	48.6	39.5	49	40	10-102	21	40
Hexachloropropene	ug/L	1.4U	100	100	56.9	51.1	57	51	29.1-84.	11	40
Indeno(1,2,3-cd)pyrene	ug/L	0.70U	100	100	61.4	106	61	106	33.7-120	53	40 J(D6)
Isodrin	ug/L	0.52U	100	100	100	97.6	100	98	32.4-130	2	40
Isophorone	ug/L	0.70U	100	100	86.2	88.1	86	88	42.5-107	2	40
Isosafrole	ug/L	0.58U	100	100	79.2	77.5	79	78	45.8-99.	2	40
Methapyrilene	ug/L	1.6U	100	100	71.2	66.8	71	67	17.8-119	6	40
Methyl methanesulfonate	ug/L	0.96U	100	100	72.8	73.2	73	73	10-107	.5	40
N-Nitroso-di-n-butylamine	ug/L	0.53U	100	100	87.0	88.5	87	89	15.2-107	2	40
N-Nitroso-di-n-propylamine	ug/L	0.90U	100	100	85.6	88.9	86	89	19.1-111	4	40
N-Nitrosodiethylamine	ug/L	0.70U	100	100	80.8	80.9	81	81	10-130.6	.06	40
N-Nitrosodimethylamine	ug/L	0.93U	100	100	67.5	67.1	68	67	10-132	.5	40
N-Nitrosodiphenylamine	ug/L	0.48U	100	100	98.2	98.3	98	98	37-104.4	.1	40
N-Nitrosomethylethylamine	ug/L	0.71U	100	100	76.8	76.2	77	76	10-135	.8	40
N-Nitrosopiperidine	ug/L	0.61U	100	100	82.5	85.3	82	85	43.3-96.	3	40
N-Nitrosopyrrolidine	ug/L	0.84U	100	100	88.2	93.4	88	93	43.1-97.	6	40
Naphthalene	ug/L	0.75U	100	100	70.0	66.7	70	67	40.1-85.	5	40
Nitrobenzene	ug/L	1.0U	100	100	74.6	74.4	75	74	32.9-115	.2	40
O,O,O-Triethylphosphorothioate	ug/L	0.66U	100	100	75.2	78.0	75	78	48.5-99.	4	40
O-Toluidine	ug/L	1.0U	100	100	78.9	81.0	79	81	21.2-134	3	40
P-Dimethylaminoazobenzene	ug/L	0.64U	100	100	100	100	100	100	44.6-142	.1	40
Parathion (Ethyl parathion)	ug/L	1.1U	100	100	104	102	104	102	46.8-113	1	40
Pentachlorobenzene	ug/L	0.75U	100	100	88.8	90.2	89	90	37.5-128	2	40
Pentachlorophenol	ug/L	0.63U	100	100	103	103	103	103	44.6-115	.2	40
Phenacetin	ug/L	0.51U	100	100	104	102	104	102	19.3-143	2	40
Phenanthrene	ug/L	0.50U	100	100	98.0	99.7	98	100	49.2-124	2	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 286 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132323 132324												
Parameter	Units	3520572014 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Phenol	ug/L	0.52U	100	100	55.5	57.4	55	57	10-158.5	3	40	
Pronamide	ug/L	1.1U	100	100	104	99.6	104	100	10-128.9	4	40	
Pyrene	ug/L	0.65U	100	100	97.0	93.5	97	93	10-150.1	4	40	
Safrole	ug/L	0.82U	100	100	86.9	86.1	87	86	10-135.9	1	40	
Thionazin	ug/L	0.59U	100	100	97.1	94.2	97	94	45-105.7	3	40	
2,4,6-Tribromophenol (S)	%						92	89	10-110			
2-Fluorobiphenyl (S)	%						84	81	18-110			
2-Fluorophenol (S)	%						58	56	18-110			
Nitrobenzene-d5 (S)	%						78	77	10-110			
Phenol-d6 (S)	%						53	52	10-110			
Terphenyl-d14 (S)	%						94	93	10-123			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3429

Analysis Method: EPA 8270

QC Batch Method: EPA 3510

Analysis Description: 8270 Water MSSV App II

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 135954

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	0.70U	5.0	11/11/10 02:21	
1,2,4-Trichlorobenzene	ug/L	0.83U	5.0	11/11/10 02:21	
1,2-Dichlorobenzene	ug/L	0.68U	5.0	11/11/10 02:21	
1,2-Dinitrobenzene	ug/L	1.2U	5.0	11/11/10 02:21	
1,3,5-Trinitrobenzene	ug/L	1.2U	5.0	11/11/10 02:21	
1,3-Dichlorobenzene	ug/L	0.76U	5.0	11/11/10 02:21	
1,3-Dinitrobenzene	ug/L	0.68U	8.0	11/11/10 02:21	
1,4-Dichlorobenzene	ug/L	0.77U	5.0	11/11/10 02:21	
1,4-Naphthoquinone	ug/L	1.2U	5.0	11/11/10 02:21	
1-Methylnaphthalene	ug/L	1.0U	5.0	11/11/10 02:21	
1-Naphthylamine	ug/L	1.0U	5.0	11/11/10 02:21	
2,3,4,6-Tetrachlorophenol	ug/L	3.8U	5.0	11/11/10 02:21	
2,4,5-Trichlorophenol	ug/L	0.52U	4.0	11/11/10 02:21	
2,4,6-Trichlorophenol	ug/L	0.69U	2.0	11/11/10 02:21	
2,4-Dichlorophenol	ug/L	0.56U	2.0	11/11/10 02:21	
2,4-Dimethylphenol	ug/L	1.6U	5.0	11/11/10 02:21	
2,4-Dinitrophenol	ug/L	1.6U	20.0	11/11/10 02:21	
2,4-Dinitrotoluene	ug/L	0.53U	2.0	11/11/10 02:21	
2,6-Dichlorophenol	ug/L	0.62U	4.0	11/11/10 02:21	
2,6-Dinitrotoluene	ug/L	1.2U	2.0	11/11/10 02:21	
2-Acetylaminofluorene	ug/L	0.65U	5.0	11/11/10 02:21	
2-Chloronaphthalene	ug/L	0.80U	5.0	11/11/10 02:21	
2-Chlorophenol	ug/L	0.68U	5.0	11/11/10 02:21	
2-Methylnaphthalene	ug/L	0.99U	5.0	11/11/10 02:21	
2-Methylphenol(o-Cresol)	ug/L	0.73U	5.0	11/11/10 02:21	
2-Naphthylamine	ug/L	2.3U	5.0	11/11/10 02:21	
2-Nitroaniline	ug/L	0.60U	5.0	11/11/10 02:21	
2-Nitrophenol	ug/L	0.81U	5.0	11/11/10 02:21	
3&4-Methylphenol(m&p Cresol)	ug/L	0.66U	10.0	11/11/10 02:21	
3,3'-Dichlorobenzidine	ug/L	0.69U	10.0	11/11/10 02:21	
3,3'-Dimethylbenzidine	ug/L	3.1U	10.0	11/11/10 02:21	
3-Methylcholanthrene	ug/L	1.0U	5.0	11/11/10 02:21	
3-Nitroaniline	ug/L	0.99U	5.0	11/11/10 02:21	
4,6-Dinitro-2-methylphenol	ug/L	1.3U	20.0	11/11/10 02:21	
4-Aminobiphenyl	ug/L	2.8U	5.0	11/11/10 02:21	
4-Bromophenylphenyl ether	ug/L	0.67U	5.0	11/11/10 02:21	
4-Chloro-3-methylphenol	ug/L	0.62U	20.0	11/11/10 02:21	
4-Chloroaniline	ug/L	1.2U	5.0	11/11/10 02:21	
4-Chlorophenylphenyl ether	ug/L	0.63U	5.0	11/11/10 02:21	
4-Nitroaniline	ug/L	0.69U	4.0	11/11/10 02:21	
4-Nitrophenol	ug/L	1.1U	20.0	11/11/10 02:21	
5-Nitro-o-toluidine	ug/L	1.3U	5.0	11/11/10 02:21	
7,12-Dimethylbenz(a)anthracene	ug/L	2.0U	5.0	11/11/10 02:21	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 288 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 135954

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
a,a-Dimethylphenylethylamine	ug/L	10.0U	20.0	11/11/10 02:21	
Acenaphthene	ug/L	0.86U	5.0	11/11/10 02:21	
Acenaphthylene	ug/L	0.95U	5.0	11/11/10 02:21	
Acetophenone	ug/L	1.4U	5.0	11/11/10 02:21	
Anthracene	ug/L	0.60U	5.0	11/11/10 02:21	
Benzo(a)anthracene	ug/L	0.63U	5.0	11/11/10 02:21	
Benzo(a)pyrene	ug/L	0.58U	1.0	11/11/10 02:21	
Benzo(b)fluoranthene	ug/L	0.62U	2.0	11/11/10 02:21	
Benzo(g,h,i)perylene	ug/L	0.68U	5.0	11/11/10 02:21	
Benzo(k)fluoranthene	ug/L	0.51U	4.0	11/11/10 02:21	
Benzyl alcohol	ug/L	1.0U	5.0	11/11/10 02:21	
bis(2-Chloroethoxy)methane	ug/L	3.0U	5.0	11/11/10 02:21	
bis(2-Chloroethyl) ether	ug/L	0.75U	4.0	11/11/10 02:21	
bis(2-Chloroisopropyl) ether	ug/L	0.73U	5.0	11/11/10 02:21	
bis(2-Ethylhexyl)phthalate	ug/L	0.80U	5.0	11/11/10 02:21	
Butylbenzylphthalate	ug/L	0.72U	5.0	11/11/10 02:21	
Chrysene	ug/L	0.37U	5.0	11/11/10 02:21	
Di-n-butylphthalate	ug/L	0.41U	5.0	11/11/10 02:21	
Di-n-octylphthalate	ug/L	0.90U	5.0	11/11/10 02:21	
Diallate	ug/L	0.73U	5.0	11/11/10 02:21	
Dibenz(a,h)anthracene	ug/L	0.65U	2.0	11/11/10 02:21	
Dibenzofuran	ug/L	0.67U	5.0	11/11/10 02:21	
Diethylphthalate	ug/L	0.51U	5.0	11/11/10 02:21	
Dimethylphthalate	ug/L	0.64U	5.0	11/11/10 02:21	
Ethyl methanesulfonate	ug/L	0.90U	5.0	11/11/10 02:21	
Fluoranthene	ug/L	0.54U	5.0	11/11/10 02:21	
Fluorene	ug/L	0.56U	5.0	11/11/10 02:21	
Hexachlorobenzene	ug/L	0.80U	1.0	11/11/10 02:21	
Hexachlorocyclopentadiene	ug/L	1.3U	5.0	11/11/10 02:21	
Hexachloroethane	ug/L	0.71U	5.0	11/11/10 02:21	
Hexachloropropene	ug/L	1.4U	5.0	11/11/10 02:21	
Indeno(1,2,3-cd)pyrene	ug/L	0.73U	2.0	11/11/10 02:21	
Isodrin	ug/L	0.54U	5.0	11/11/10 02:21	
Isophorone	ug/L	0.73U	5.0	11/11/10 02:21	
Isosafrole	ug/L	0.60U	5.0	11/11/10 02:21	
Kepone	ug/L	10.0U	20.0	11/11/10 02:21	
Methapyrilene	ug/L	1.6U	5.0	11/11/10 02:21	
Methyl methanesulfonate	ug/L	1.0U	5.0	11/11/10 02:21	
N-Nitroso-di-n-butylamine	ug/L	0.55U	4.0	11/11/10 02:21	
N-Nitroso-di-n-propylamine	ug/L	0.94U	4.0	11/11/10 02:21	
N-Nitrosodiethylamine	ug/L	0.73U	4.0	11/11/10 02:21	
N-Nitrosodimethylamine	ug/L	0.97U	2.0	11/11/10 02:21	
N-Nitrosodiphenylamine	ug/L	0.50U	5.0	11/11/10 02:21	
N-Nitrosomethylethylamine	ug/L	0.74U	5.0	11/11/10 02:21	
N-Nitrosopiperidine	ug/L	0.64U	5.0	11/11/10 02:21	
N-Nitrosopyrrolidine	ug/L	0.88U	5.0	11/11/10 02:21	
Naphthalene	ug/L	0.78U	5.0	11/11/10 02:21	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 289 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 135954

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrobenzene	ug/L	1.1U	4.0	11/11/10 02:21	
O,O,O-Triethylphosphorothioate	ug/L	0.69U	5.0	11/11/10 02:21	
O-Toluidine	ug/L	1.1U	5.0	11/11/10 02:21	
P-Dimethylaminoazobenzene	ug/L	0.67U	5.0	11/11/10 02:21	
p-Phenylenediamine	ug/L	10.0U	20.0	11/11/10 02:21	
Parathion (Ethyl parathion)	ug/L	1.2U	5.0	11/11/10 02:21	
Pentachlorobenzene	ug/L	0.78U	5.0	11/11/10 02:21	
Pentachlorophenol	ug/L	0.66U	20.0	11/11/10 02:21	
Phenacetin	ug/L	0.53U	5.0	11/11/10 02:21	
Phenanthrene	ug/L	0.52U	5.0	11/11/10 02:21	
Phenol	ug/L	0.54U	5.0	11/11/10 02:21	
Pronamide	ug/L	1.1U	5.0	11/11/10 02:21	
Pyrene	ug/L	0.68U	5.0	11/11/10 02:21	
Safrole	ug/L	0.85U	5.0	11/11/10 02:21	
Thionazin	ug/L	0.61U	5.0	11/11/10 02:21	
2,4,6-Tribromophenol (S)	%	89	10-110	11/11/10 02:21	
2-Fluorobiphenyl (S)	%	78	18-110	11/11/10 02:21	
2-Fluorophenol (S)	%	50	18-110	11/11/10 02:21	
Nitrobenzene-d5 (S)	%	74	10-110	11/11/10 02:21	
Phenol-d6 (S)	%	34	10-110	11/11/10 02:21	
Terphenyl-d14 (S)	%	90	10-123	11/11/10 02:21	

LABORATORY CONTROL SAMPLE: 135955

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,2,4,5-Tetrachlorobenzene	ug/L	50	41.1	82	10-146.9	
1,2,4-Trichlorobenzene	ug/L	50	38.8	78	19.7-141.6	
1,2-Dichlorobenzene	ug/L	50	36.1	72	23.5-105.7	
1,2-Dinitrobenzene	ug/L	50	46.5	93	52.6-113.9	
1,3,5-Trinitrobenzene	ug/L	50	45.9	92	41.4-102.4	
1,3-Dichlorobenzene	ug/L	50	36.5	73	25.5-94.5	
1,3-Dinitrobenzene	ug/L	50	45.9	92	45.3-116.4	
1,4-Dichlorobenzene	ug/L	50	34.8	70	33.2-90.7	
1,4-Naphthoquinone	ug/L	50	42.7	85	39.3-113.3	
1-Methylnaphthalene	ug/L	50	40.8	82	27.8-110.7	
1-Naphthylamine	ug/L	50	45.0	90	37.1-90.4	
2,3,4,6-Tetrachlorophenol	ug/L	50	47.4	95	14.3-115.3	
2,4,5-Trichlorophenol	ug/L	50	48.4	97	10-121.3	
2,4,6-Trichlorophenol	ug/L	50	45.3	91	40.3-101.7	
2,4-Dichlorophenol	ug/L	50	42.4	85	35.8-108.5	
2,4-Dimethylphenol	ug/L	50	40.4	81	25-104.5	
2,4-Dinitrophenol	ug/L	50	47.9	96	10-147.1	
2,4-Dinitrotoluene	ug/L	50	46.3	93	47.9-113.6	
2,6-Dichlorophenol	ug/L	50	42.3	85	41.1-101.8	
2,6-Dinitrotoluene	ug/L	50	47.7	95	44.6-111.5	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 135955

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
2-Acetylaminofluorene	ug/L	50	44.9	90	58.3-112.7	
2-Chloronaphthalene	ug/L	50	42.2	84	41.2-101.2	
2-Chlorophenol	ug/L	50	36.8	74	32.1-96.5	
2-Methylnaphthalene	ug/L	50	40.7	81	40-93.6	
2-Methylphenol(o-Cresol)	ug/L	50	35.3	71	36.7-95.1	
2-Naphthylamine	ug/L	50	34.9	70	42.4-98.8	
2-Nitroaniline	ug/L	50	47.5	95	59-103.2	
2-Nitrophenol	ug/L	50	40.9	82	38.4-108.9	
3&4-Methylphenol(m&p Cresol)	ug/L	50	33.7	67	35.2-94.3	
3,3'-Dichlorobenzidine	ug/L	50	46.2	92	30.7-106	
3,3'-Dimethylbenzidine	ug/L	50	37.8	76	10-160.1	
3-Methylcholanthrene	ug/L	50	44.9	90	15.5-121.4	
3-Nitroaniline	ug/L	50	43.1	86	25.3-131.5	
4,6-Dinitro-2-methylphenol	ug/L	50	45.1	90	35.2-130.5	
4-Aminobiphenyl	ug/L	50	39.5	79	50.8-112.6	
4-Bromophenylphenyl ether	ug/L	50	43.8	88	51.9-110.4	
4-Chloro-3-methylphenol	ug/L	50	43.5	87	19.4-128.8	
4-Chloroaniline	ug/L	50	40.1	80	30.1-108.4	
4-Chlorophenylphenyl ether	ug/L	50	45.2	90	49.7-91.5	
4-Nitroaniline	ug/L	50	46.5	93	48.1-112.2	
4-Nitrophenol	ug/L	50	20.9	42	10-121.8	
5-Nitro-o-toluidine	ug/L	50	44.0	88	43-113	
7,12-Dimethylbenz(a)anthracene	ug/L	50	38.1	76	52.5-108.6	
Acenaphthene	ug/L	50	43.4	87	50.3-98.3	
Acenaphthylene	ug/L	50	43.7	87	49-98.1	
Acetophenone	ug/L	50	39.5	79	40.6-94.3	
Anthracene	ug/L	50	44.4	89	55-112.5	
Benzo(a)anthracene	ug/L	50	44.1	88	10-150.1	
Benzo(a)pyrene	ug/L	50	46.1	92	59.7-108.4	
Benzo(b)fluoranthene	ug/L	50	45.0	90	58.4-111.8	
Benzo(g,h,i)perylene	ug/L	50	46.5	93	57.6-115.1	
Benzo(k)fluoranthene	ug/L	50	43.5	87	57.6-112.4	
Benzyl alcohol	ug/L	50	37.1	74	26.1-118	
bis(2-Chloroethoxy)methane	ug/L	50	41.0	82	41.2-96.2	
bis(2-Chloroethyl) ether	ug/L	50	35.5	71	35.3-99.5	
bis(2-Chloroisopropyl) ether	ug/L	50	37.9	76	36.3-91	
bis(2-Ethylhexyl)phthalate	ug/L	50	45.6	91	43.1-118.3	
Butylbenzylphthalate	ug/L	50	46.0	92	57.5-118.2	
Chrysene	ug/L	50	44.9	90	42.4-113.9	
Di-n-butylphthalate	ug/L	50	45.8	92	22.2-139.3	
Di-n-octylphthalate	ug/L	50	47.4	95	57.4-116.9	
Diallate	ug/L	50	38.3	77	44.3-111.7	
Dibenz(a,h)anthracene	ug/L	50	46.1	92	59.1-111.8	
Dibenzofuran	ug/L	50	42.0	84	45.3-108.3	
Diethylphthalate	ug/L	50	43.9	88	51.1-107.5	
Dimethylphthalate	ug/L	50	45.3	91	47.4-110.6	
Ethyl methanesulfonate	ug/L	50	37.5	75	35.9-103.6	
Fluoranthene	ug/L	50	45.4	91	48.2-118.6	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 291 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 135955

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Fluorene	ug/L	50	42.9	86	44.7-106.4	
Hexachlorobenzene	ug/L	50	41.2	82	54-113.2	
Hexachlorocyclopentadiene	ug/L	50	37.2	74	16.5-105.1	ES
Hexachloroethane	ug/L	50	35.8	72	10-102	
Hexachloropropene	ug/L	50	39.7	79	29.1-84.2	
Indeno(1,2,3-cd)pyrene	ug/L	50	46.1	92	33.7-120.7	
Isodrin	ug/L	50	45.1	90	32.4-130.4	
Isophorone	ug/L	50	41.7	83	42.5-107.7	
Isosafrole	ug/L	50	41.9	84	45.8-99.3	
Methapyrilene	ug/L	50	32.3	65	17.8-119.5	
Methyl methanesulfonate	ug/L	50	30.6	61	10-107	
N-Nitroso-di-n-butylamine	ug/L	50	40.8	82	15.2-107.9	
N-Nitroso-di-n-propylamine	ug/L	50	38.6	77	19.1-111.6	
N-Nitrosodiethylamine	ug/L	50	37.1	74	10-130.6	
N-Nitrosodimethylamine	ug/L	50	24.0	48	10-132	
N-Nitrosodiphenylamine	ug/L	50	42.7	85	37-104.4	
N-Nitrosomethylethylamine	ug/L	50	35.9	72	10-135	
N-Nitrosopiperidine	ug/L	50	41.4	83	43.3-96.3	
N-Nitrosopyrrolidine	ug/L	50	37.1	74	43.1-97.2	
Naphthalene	ug/L	50	39.8	80	40.1-85.7	
Nitrobenzene	ug/L	50	36.9	74	32.9-115.9	
O,O,O-Triethylphosphorothioate	ug/L	50	40.3	81	48.5-99.9	
O-Toluidine	ug/L	50	37.9	76	21.2-134.1	
P-Dimethylaminoazobenzene	ug/L	50	43.7	87	44.6-142.5	
Parathion (Ethyl parathion)	ug/L	50	46.0	92	46.8-113.9	
Pentachlorobenzene	ug/L	50	42.9	86	37.5-128.1	
Pentachlorophenol	ug/L	50	43.7	87	44.6-115.6	
Phenacetin	ug/L	50	41.7	83	19.3-143.2	
Phenanthrene	ug/L	50	44.2	88	49.2-124.2	
Phenol	ug/L	50	19.1	38	10-158.5	
Pronamide	ug/L	50	43.5	87	10-128.9	
Pyrene	ug/L	50	44.5	89	10-150.1	
Safrole	ug/L	50	44.5	89	10-135.9	
Thionazin	ug/L	50	42.8	86	45-105.7	
2,4,6-Tribromophenol (S)	%			89	10-110	
2-Fluorobiphenyl (S)	%			78	18-110	
2-Fluorophenol (S)	%			46	18-110	
Nitrobenzene-d5 (S)	%			78	10-110	
Phenol-d6 (S)	%			34	10-110	
Terphenyl-d14 (S)	%			85	10-123	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135956 135957

Parameter	Units	3519325032 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,2,4,5-Tetrachlorobenzene	ug/L	8.4U	100	100	91.1 I	84.7 I	91	85	10-146.9		40	
1,2,4-Trichlorobenzene	ug/L	10U	100	100	76.8 I	71.3 I	77	71	19.7-141		40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 292 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135956 135957											
Parameter	Units	3519325032		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						Qual
1,2-Dichlorobenzene	ug/L	8.2U	100	100	62.9 I	55.8 I	62	55	23.5-105		40
1,2-Dinitrobenzene	ug/L	14.1U	100	100	97.8 I	89.4 I	98	89	52.6-113		40
1,3,5-Trinitrobenzene	ug/L	14.7U	100	100	149	147	149	147	41.4-102	2	40 M6
1,3-Dichlorobenzene	ug/L	9.1U	100	100	59.8 I	55.1 I	60	55	25.5-94.		40
1,3-Dinitrobenzene	ug/L	8.2U	100	100	93.0 I	91.8 I	93	92	45.3-116		40
1,4-Dichlorobenzene	ug/L	19.7 I	100	100	79.6 I	72.4 I	60	53	33.2-90.		40
1,4-Naphthoquinone	ug/L	14.2U	100	100	40.4 I	34.7 I	40	35	39.3-113		40 M6
1-Methylnaphthalene	ug/L	12.0U	100	100	94.9 I	80.5 I	94	79	27.8-110		40
1-Naphthylamine	ug/L	12.4U	100	100	20.6U	20.6U	20	19	37.1-90.		40 M6
2,3,4,6-Tetrachlorophenol	ug/L	46.3U	100	100	109	101	109	101	14.3-115	7	40
2,4,5-Trichlorophenol	ug/L	6.3U	100	100	109	98.4	109	98	10-121.3	11	40
2,4,6-Trichlorophenol	ug/L	8.3U	100	100	111	102	111	102	40.3-101	9	40 M6
2,4-Dichlorophenol	ug/L	6.7U	100	100	94.8	87.5	95	87	35.8-108	8	40
2,4-Dimethylphenol	ug/L	19.0U	100	100	104	100	104	100	25-104.5	4	40
2,4-Dinitrophenol	ug/L	18.9U	100	100	219 I	215 I	219	215	10-147.1		40 M6
2,4-Dinitrotoluene	ug/L	6.4U	100	100	117	111	117	111	47.9-113	6	40 M6
2,6-Dichlorophenol	ug/L	7.5U	100	100	96.2	91.6	96	92	41.1-101	5	40
2,6-Dinitrotoluene	ug/L	14.7U	100	100	111	102	101	92	44.6-111	8	40
2-Acetylaminofluorene	ug/L	7.8U	100	100	118	123	118	123	58.3-112	4	40 M6
2-Chloronaphthalene	ug/L	9.6U	100	100	95.8 I	88.1 I	96	88	41.2-101		40
2-Chlorophenol	ug/L	8.2U	100	100	91.3 I	80.9 I	91	81	32.1-96.		40
2-Methylnaphthalene	ug/L	11.9U	100	100	88.4 I	80.1 I	88	80	40-93.6		40
2-Methylphenol(o-Cresol)	ug/L	8.8U	100	100	104	89.6 I	99	85	36.7-95.		40 M6
2-Naphthylamine	ug/L	27.3U	100	100	45.4U	45.4U	35	40	42.4-98.		40 M6
2-Nitroaniline	ug/L	7.2U	100	100	99.6 I	93.1 I	100	93	59-103.2		40
2-Nitrophenol	ug/L	9.8U	100	100	88.8 I	86.6 I	89	87	38.4-108		40
3&4-Methylphenol(m&p Cresol)	ug/L	178	100	100	288	251	110	73	35.2-94.	14	40 M6
3,3'-Dichlorobenzidine	ug/L	8.3U	100	100	13.8U	13.8U	8	5	30.7-106		40 M6
3,3'-Dimethylbenzidine	ug/L	37.7U	100	100	62.6U	62.6U	1	0	10-160.1		40 M6
3-Methylcholanthrene	ug/L	12.5U	100	100	89.9 I	87.1 I	90	87	15.5-121		40
3-Nitroaniline	ug/L	11.9U	100	100	81.6 I	78.3 I	82	78	25.3-131		40
4,6-Dinitro-2-methylphenol	ug/L	15.9U	100	100	170 I	161 I	170	161	35.2-130		40 M6
4-Aminobiphenyl	ug/L	34.1U	100	100	56.6U	56.6U	20	22	50.8-112		40 M6
4-Bromophenylphenyl ether	ug/L	8.1U	100	100	100	93.3 I	100	93	51.9-110		40
4-Chloro-3-methylphenol	ug/L	7.5U	100	100	124 I	109 I	124	109	19.4-128		40
4-Chloroaniline	ug/L	14.6U	100	100	66.7 I	65.5 I	67	66	30.1-108		40
4-Chlorophenylphenyl ether	ug/L	7.6U	100	100	100	96.9 I	100	97	49.7-91.		40 M6
4-Nitroaniline	ug/L	8.3U	100	100	69.9 I	73.4 I	70	73	48.1-112		40
4-Nitrophenol	ug/L	13.0U	100	100	99.4 I	102 I	99	102	10-121.8		40
5-Nitro-o-toluidine	ug/L	15.5U	100	100	90.3 I	85.7 I	90	86	43-113		40
7,12-Dimethylbenz(a)anthracene	ug/L	23.5U	100	100	97.3 I	95.0 I	97	95	52.5-108		40
Acenaphthene	ug/L	10.4U	100	100	98.5 I	87.3 I	99	87	50.3-98.		40 M6
Acenaphthylene	ug/L	11.4U	100	100	97.1 I	90.8 I	97	91	49-98.1		40
Acetophenone	ug/L	17.5U	100	100	99.7 I	93.0 I	90	83	40.6-94.		40
Anthracene	ug/L	7.2U	100	100	107	98.7 I	107	99	55-112.5		40
Benzo(a)anthracene	ug/L	7.6U	100	100	96.7 I	91.9 I	97	92	10-150.1		40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 293 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135956 135957											
Parameter	Units	3519325032		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						
						Result	Result	% Rec	% Rec	Limits	RPD
Benzo(a)pyrene	ug/L	7.0U	100	100	100	97.2	90.6	97	91	59.7-108	7
Benzo(b)fluoranthene	ug/L	7.5U	100	100	100	98.3	96.8	98	97	58.4-111	1
Benzo(g,h,i)perylene	ug/L	8.2U	100	100	100	88.3 I	88.0 I	88	88	57.6-115	40
Benzo(k)fluoranthene	ug/L	6.1U	100	100	100	102	95.1	102	95	57.6-112	7
Benzyl alcohol	ug/L	12.3U	100	100	100	101	88.0 I	101	88	26.1-118	40
bis(2-Chloroethoxy)methane	ug/L	35.5U	100	100	100	93.7 I	77.6 I	94	78	41.2-96.	40
bis(2-Chloroethyl) ether	ug/L	9.0U	100	100	100	87.0	74.4 I	87	74	35.3-99.	40
bis(2-Chloroisopropyl) ether	ug/L	8.8U	100	100	100	83.4 I	74.2 I	83	74	36.3-91	40
bis(2-Ethylhexyl)phthalate	ug/L	9.6U	100	100	100	94.0 I	90.6 I	94	91	43.1-118	40
Butylbenzylphthalate	ug/L	8.7U	100	100	100	101	93.0 I	101	93	57.5-118	40
Chrysene	ug/L	4.5U	100	100	100	94.7 I	90.7 I	95	91	42.4-113	40
Di-n-butylphthalate	ug/L	4.9U	100	100	100	106	101	105	101	22.2-139	5
Di-n-octylphthalate	ug/L	10.8U	100	100	100	84.7 I	85.9 I	85	86	57.4-116	40
Diallylate	ug/L	8.8U	100	100	100	111	108	111	108	44.3-111	3
Dibenz(a,h)anthracene	ug/L	7.8U	100	100	100	91.2	89.7	91	90	59.1-111	2
Dibenzofuran	ug/L	8.1U	100	100	100	101	97.0 I	101	97	45.3-108	40
Diethylphthalate	ug/L	8.5 I	100	100	100	116	114	108	106	51.1-107	2
Dimethylphthalate	ug/L	7.7U	100	100	100	103	98.4 I	102	97	47.4-110	40
Ethyl methanesulfonate	ug/L	10.8U	100	100	100	87.8 I	81.1 I	88	81	35.9-103	40
Fluoranthene	ug/L	6.5U	100	100	100	101	104	101	104	48.2-118	3
Fluorene	ug/L	6.7U	100	100	100	105	104	105	104	44.7-106	2
Hexachlorobenzene	ug/L	9.6U	100	100	100	99.5	91.7	100	92	54-113.2	8
Hexachlorocyclopentadiene	ug/L	15.4U	100	100	100	141	141	141	141	16.5-105	.6
Hexachloroethane	ug/L	8.5U	100	100	100	73.7 I	63.5 I	74	64	10-102	40
Hexachloropropene	ug/L	17.0U	100	100	100	64.2 I	61.1 I	64	61	29.1-84.	40
Indeno(1,2,3-cd)pyrene	ug/L	8.8U	100	100	100	89.7	88.9	90	89	33.7-120	.9
Isodrin	ug/L	6.5U	100	100	100	92.5 I	89.7 I	92	90	32.4-130	40
Isophorone	ug/L	8.8U	100	100	100	92.7 I	82.5 I	93	82	42.5-107	40
Isosafrole	ug/L	7.2U	100	100	100	93.6 I	86.1 I	94	86	45.8-99.	40
Methapyrilene	ug/L	19.9U	100	100	100	63.2 I	76.2 I	63	76	17.8-119	40
Methyl methanesulfonate	ug/L	12.0U	100	100	100	82.9 I	74.9 I	83	75	10-107	40
N-Nitroso-di-n-butylamine	ug/L	6.6U	100	100	100	96.6	98.4	97	98	15.2-107	2
N-Nitroso-di-n-propylamine	ug/L	11.3U	100	100	100	106	88.9	106	89	19.1-111	18
N-Nitrosodiethylamine	ug/L	8.8U	100	100	100	73.5 I	79.3 I	74	79	10-130.6	40
N-Nitrosodimethylamine	ug/L	11.7U	100	100	100	52.2	59.3	52	59	10-132	13
N-Nitrosodiphenylamine	ug/L	6.0U	100	100	100	105	95.1 I	105	95	37-104.4	40
N-Nitrosomethylethylamine	ug/L	8.9U	100	100	100	67.1 I	70.0 I	67	70	10-135	40
N-Nitrosopiperidine	ug/L	7.7U	100	100	100	88.7 I	86.1 I	89	86	43.3-96.	40
N-Nitrosopyrrolidine	ug/L	10.6U	100	100	100	72.9 I	87.2 I	73	87	43.1-97.	40
Naphthalene	ug/L	15.7 I	100	100	100	94.8 I	90.3 I	79	75	40.1-85.	40
Nitrobenzene	ug/L	13.1U	100	100	100	83.5	76.2 I	83	76	32.9-115	40
O,O,O-Triethylphosphorothioate	ug/L	8.3U	100	100	100	83.7 I	78.1 I	84	78	48.5-99.	40
O-Toluidine	ug/L	12.9U	100	100	100	87.6 I	80.2 I	81	74	21.2-134	40
P-Dimethylaminoazobenzene	ug/L	8.1U	100	100	100	80.9 I	78.2 I	81	78	44.6-142	40
Parathion (Ethyl parathion)	ug/L	13.8U	100	100	100	106	101	106	101	46.8-113	4
Pentachlorobenzene	ug/L	9.4U	100	100	100	99.1 I	89.3 I	99	89	37.5-128	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 294 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135956 135957											
Parameter	Units	3519325032 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Pentachlorophenol	ug/L	7.9U	100	100	289 I	283 I	289	283	44.6-115	40	M6
Phenacetin	ug/L	6.4U	100	100	93.6 I	89.9 I	94	90	19.3-143	40	
Phenanthrene	ug/L	6.3U	100	100	104	98.9 I	104	99	49.2-124	40	
Phenol	ug/L	116	100	100	211	194	95	79	10-158.5	8	40 D3
Pronamide	ug/L	13.6U	100	100	108	104	108	104	10-128.9	3	40
Pyrene	ug/L	8.2U	100	100	107	98.6 I	107	99	10-150.1		40
Safrole	ug/L	10.2U	100	100	115	103	115	103	10-135.9	11	40
Thionazin	ug/L	7.3U	100	100	103	95.5 I	103	96	45-105.7		40
2,4,6-Tribromophenol (S)	%						95	97	10-110		
2-Fluorobiphenyl (S)	%						92	84	18-110		
2-Fluorophenol (S)	%						68	63	18-110		
Nitrobenzene-d5 (S)	%						82	74	10-110		
Phenol-d6 (S)	%						71	63	10-110		
Terphenyl-d14 (S)	%						100	91	10-123		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3322 Analysis Method: EPA 8270 by SCAN
QC Batch Method: EPA 3510 Analysis Description: 8270 Water CPAH by SCAN MSSV
Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 130138 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1-Methylnaphthalene	ug/L	0.090U	1.5	10/19/10 18:44	
2-Methylnaphthalene	ug/L	0.060U	1.5	10/19/10 18:44	
Acenaphthene	ug/L	0.030U	1.0	10/19/10 18:44	
Acenaphthylene	ug/L	0.050U	2.0	10/19/10 18:44	
Anthracene	ug/L	0.050U	1.0	10/19/10 18:44	
Benzo(a)anthracene	ug/L	0.060U	0.20	10/19/10 18:44	
Benzo(a)pyrene	ug/L	0.050U	0.20	10/19/10 18:44	
Benzo(b)fluoranthene	ug/L	0.050U	0.10	10/19/10 18:44	
Benzo(g,h,i)perylene	ug/L	0.060U	1.0	10/19/10 18:44	
Benzo(k)fluoranthene	ug/L	0.040U	0.25	10/19/10 18:44	
Chrysene	ug/L	0.060U	1.0	10/19/10 18:44	
Dibenz(a,h)anthracene	ug/L	0.050U	0.20	10/19/10 18:44	
Fluoranthene	ug/L	0.060U	1.0	10/19/10 18:44	
Fluorene	ug/L	0.030U	1.0	10/19/10 18:44	
Indeno(1,2,3-cd)pyrene	ug/L	0.040U	0.15	10/19/10 18:44	
Naphthalene	ug/L	0.080U	1.0	10/19/10 18:44	
Phenanthrene	ug/L	0.050U	1.0	10/19/10 18:44	
Pyrene	ug/L	0.060U	1.0	10/19/10 18:44	
2-Fluorobiphenyl (S)	%	78	43.9-113	10/19/10 18:44	
Terphenyl-d14 (S)	%	93	24.8-144	10/19/10 18:44	

LABORATORY CONTROL SAMPLE: 130139

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1-Methylnaphthalene	ug/L	5	4.0	81	46.7-104	
2-Methylnaphthalene	ug/L	5	4.0	80	49.4-106	
Acenaphthene	ug/L	5	3.8	76	42.7-109	
Acenaphthylene	ug/L	5	4.2	84	53.2-107	
Anthracene	ug/L	5	4.3	86	52.2-112	
Benzo(a)anthracene	ug/L	5	4.3	86	57.5-115	
Benzo(a)pyrene	ug/L	5	3.8	76	61.8-104	
Benzo(b)fluoranthene	ug/L	5	4.4	88	61.6-120	
Benzo(g,h,i)perylene	ug/L	5	3.5	70	41.6-122	
Benzo(k)fluoranthene	ug/L	5	3.8	76	53.3-106	
Chrysene	ug/L	5	3.8	76	48-121	
Dibenz(a,h)anthracene	ug/L	5	3.6	71	38.3-110	
Fluoranthene	ug/L	5	4.5	89	46.8-122	
Fluorene	ug/L	5	4.0	80	50.5-107	
Indeno(1,2,3-cd)pyrene	ug/L	5	3.6	72	42.4-108	
Naphthalene	ug/L	5	3.7	74	43.9-99.6	
Phenanthrene	ug/L	5	4.0	80	54.3-107	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 296 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130139

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Pyrene	ug/L	5	4.3	86	48.5-120	
2-Fluorobiphenyl (S)	%			81	43.9-113	
Terphenyl-d14 (S)	%			87	24.8-144	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130140 130141

Parameter	Units	3520504001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1-Methylnaphthalene	ug/L	0.086U	10	10	7.3	7.3	73	73	46.7-104	.6	40	
2-Methylnaphthalene	ug/L	0.057U	10	10	6.9	7.1	69	71	49.4-106	4	40	
Acenaphthene	ug/L	0.029U	10	10	7.0	7.0	70	70	42.7-109	.7	40	
Acenaphthylene	ug/L	0.048U	10	10	7.9	8.2	79	82	53.2-107	4	40	
Anthracene	ug/L	0.048U	10	10	8.7	7.9	87	79	52.2-112	9	40	
Benzo(a)anthracene	ug/L	0.057U	10	10	8.7	7.8	87	78	57.5-115	11	40	
Benzo(a)pyrene	ug/L	0.048U	10	10	8.6	7.2	86	72	61.8-104	17	40	
Benzo(b)fluoranthene	ug/L	0.048U	10	10	9.8	8.1	98	81	61.6-120	19	40	
Benzo(g,h,i)perylene	ug/L	0.057U	10	10	8.6	7.3	86	73	41.6-122	16	40	
Benzo(k)fluoranthene	ug/L	0.038U	10	10	8.9	7.7	89	77	53.3-106	14	40	
Chrysene	ug/L	0.057U	10	10	7.8	6.7	78	67	48-121	16	40	
Dibenz(a,h)anthracene	ug/L	0.048U	10	10	9.2	7.9	92	79	38.3-110	15	40	
Fluoranthene	ug/L	0.057U	10	10	8.9	8.3	89	83	46.8-122	7	40	
Fluorene	ug/L	0.029U	10	10	7.2	7.4	72	74	50.5-107	3	40	
Indeno(1,2,3-cd)pyrene	ug/L	0.038U	10	10	8.7	7.3	87	73	42.4-108	18	40	
Naphthalene	ug/L	0.076U	10	10	6.9	6.8	69	68	43.9-99.	1	40	
Phenanthrene	ug/L	0.048U	10	10	8.0	7.2	80	72	54.3-107	11	40	
Pyrene	ug/L	0.057U	10	10	8.9	8.2	89	82	48.5-120	8	40	
2-Fluorobiphenyl (S)	%						74	80	43.9-113			
Terphenyl-d14 (S)	%						96	93	24.8-144			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3340 Analysis Method: EPA 8270 by SCAN
QC Batch Method: EPA 3510 Analysis Description: 8270 Water CPAH by SCAN MSSV
Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

METHOD BLANK: 130933 Matrix: Water

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1-Methylnaphthalene	ug/L	0.090U	1.5	10/21/10 11:06	
2-Methylnaphthalene	ug/L	0.060U	1.5	10/21/10 11:06	
Acenaphthene	ug/L	0.030U	1.0	10/21/10 11:06	
Acenaphthylene	ug/L	0.050U	2.0	10/21/10 11:06	
Anthracene	ug/L	0.050U	1.0	10/21/10 11:06	
Benzo(a)anthracene	ug/L	0.060U	0.20	10/21/10 11:06	
Benzo(a)pyrene	ug/L	0.050U	0.20	10/21/10 11:06	
Benzo(b)fluoranthene	ug/L	0.050U	0.10	10/21/10 11:06	
Benzo(g,h,i)perylene	ug/L	0.060U	1.0	10/21/10 11:06	
Benzo(k)fluoranthene	ug/L	0.040U	0.25	10/21/10 11:06	
Chrysene	ug/L	0.060U	1.0	10/21/10 11:06	
Dibenz(a,h)anthracene	ug/L	0.050U	0.20	10/21/10 11:06	
Fluoranthene	ug/L	0.060U	1.0	10/21/10 11:06	
Fluorene	ug/L	0.030U	1.0	10/21/10 11:06	
Indeno(1,2,3-cd)pyrene	ug/L	0.040U	0.15	10/21/10 11:06	
Naphthalene	ug/L	0.080U	1.0	10/21/10 11:06	
Phenanthrene	ug/L	0.050U	1.0	10/21/10 11:06	
Pyrene	ug/L	0.060U	1.0	10/21/10 11:06	
2-Fluorobiphenyl (S)	%	82	43.9-113	10/21/10 11:06	
Terphenyl-d14 (S)	%	91	24.8-144	10/21/10 11:06	

LABORATORY CONTROL SAMPLE & LCSD: 130934

131003

Parameter	Units	Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec	% Rec Limits	RPD	Max RPD	Qualifiers
1-Methylnaphthalene	ug/L	5	4.0	4.4	81	87	46.7-104	8	40	
2-Methylnaphthalene	ug/L	5	4.3	4.0	87	80	49.4-106	9	40	
Acenaphthene	ug/L	5	3.8	3.9	76	79	42.7-109	3	40	
Acenaphthylene	ug/L	5	3.9	4.4	79	88	53.2-107	11	40	
Anthracene	ug/L	5	4.2	4.1	84	82	52.2-112	3	40	
Benzo(a)anthracene	ug/L	5	4.5	4.4	89	89	57.5-115	.8	40	
Benzo(a)pyrene	ug/L	5	4.0	4.0	81	80	61.8-104	1	40	
Benzo(b)fluoranthene	ug/L	5	4.6	4.7	93	94	61.6-120	1	40	
Benzo(g,h,i)perylene	ug/L	5	3.5	3.6	69	73	41.6-122	6	40	
Benzo(k)fluoranthene	ug/L	5	4.5	4.1	90	83	53.3-106	8	40	
Chrysene	ug/L	5	4.0	3.9	81	78	48-121	4	40	
Dibenz(a,h)anthracene	ug/L	5	3.1	3.5	62	69	38.3-110	10	40	
Fluoranthene	ug/L	5	4.3	4.3	87	86	46.8-122	1	40	
Fluorene	ug/L	5	3.9	4.1	79	82	50.5-107	5	40	
Indeno(1,2,3-cd)pyrene	ug/L	5	3.2	3.5	64	70	42.4-108	10	40	
Naphthalene	ug/L	5	4.0	3.9	80	78	43.9-99.6	2	40	
Phenanthrene	ug/L	5	3.9	3.9	78	78	54.3-107	.003	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 298 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE & LCSD:		130934	131003							
Parameter	Units	Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec	% Rec Limits	RPD	Max RPD	Qualifiers
Pyrene	ug/L	5	4.3	4.3	86	86	48.5-120	.4	40	
2-Fluorobiphenyl (S)	%				86	85	43.9-113			
Terphenyl-d14 (S)	%				94	90	24.8-144			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3373 Analysis Method: EPA 8270 by SCAN
QC Batch Method: EPA 3510 Analysis Description: 8270 Water CPAH by SCAN MSSV
Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133217 Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1-Methylnaphthalene	ug/L	0.090U	1.5	10/26/10 11:36	
2-Methylnaphthalene	ug/L	0.060U	1.5	10/26/10 11:36	
Acenaphthene	ug/L	0.030U	1.0	10/26/10 11:36	
Acenaphthylene	ug/L	0.050U	2.0	10/26/10 11:36	
Anthracene	ug/L	0.050U	1.0	10/26/10 11:36	
Benzo(a)anthracene	ug/L	0.060U	0.20	10/26/10 11:36	
Benzo(a)pyrene	ug/L	0.050U	0.20	10/26/10 11:36	
Benzo(b)fluoranthene	ug/L	0.050U	0.10	10/26/10 11:36	
Benzo(g,h,i)perylene	ug/L	0.060U	1.0	10/26/10 11:36	
Benzo(k)fluoranthene	ug/L	0.040U	0.25	10/26/10 11:36	
Chrysene	ug/L	0.060U	1.0	10/26/10 11:36	
Dibenz(a,h)anthracene	ug/L	0.050U	0.20	10/26/10 11:36	
Fluoranthene	ug/L	0.060U	1.0	10/26/10 11:36	
Fluorene	ug/L	0.030U	1.0	10/26/10 11:36	
Indeno(1,2,3-cd)pyrene	ug/L	0.040U	0.15	10/26/10 11:36	
Naphthalene	ug/L	0.080U	1.0	10/26/10 11:36	
Phenanthrene	ug/L	0.050U	1.0	10/26/10 11:36	
Pyrene	ug/L	0.060U	1.0	10/26/10 11:36	
2-Fluorobiphenyl (S)	%	77	43.9-113	10/26/10 11:36	
Terphenyl-d14 (S)	%	93	24.8-144	10/26/10 11:36	

LABORATORY CONTROL SAMPLE: 133218

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1-Methylnaphthalene	ug/L	5	3.5	69	46.7-104	
2-Methylnaphthalene	ug/L	5	3.6	72	49.4-106	
Acenaphthene	ug/L	5	3.4	67	42.7-109	
Acenaphthylene	ug/L	5	3.7	75	53.2-107	
Anthracene	ug/L	5	3.9	77	52.2-112	
Benzo(a)anthracene	ug/L	5	4.5	89	57.5-115	
Benzo(a)pyrene	ug/L	5	3.6	72	61.8-104	
Benzo(b)fluoranthene	ug/L	5	3.8	77	61.6-120	
Benzo(g,h,i)perylene	ug/L	5	2.3	45	41.6-122	
Benzo(k)fluoranthene	ug/L	5	3.8	75	53.3-106	
Chrysene	ug/L	5	3.7	74	48-121	
Dibenz(a,h)anthracene	ug/L	5	1.9	39	38.3-110	
Fluoranthene	ug/L	5	4.0	80	46.8-122	
Fluorene	ug/L	5	3.5	70	50.5-107	
Indeno(1,2,3-cd)pyrene	ug/L	5	2.4	48	42.4-108	
Naphthalene	ug/L	5	3.3	66	43.9-99.6	
Phenanthrene	ug/L	5	3.6	71	54.3-107	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 300 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 133218

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Pyrene	ug/L	5	4.0	80	48.5-120	
2-Fluorobiphenyl (S)	%			78	43.9-113	
Terphenyl-d14 (S)	%			93	24.8-144	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 133219 133220

Parameter	Units	3519325026	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
		Result	Conc.	Conc.	Result	Result	% Rec	% Rec	Limits			
1-Methylnaphthalene	ug/L	0.092U	10	10	7.8	7.4	78	74	46.7-104	5	40	
2-Methylnaphthalene	ug/L	0.061U	10	10	6.9	6.9	69	69	49.4-106	.6	40	
Acenaphthene	ug/L	0.031U	10	10	7.6	6.7	76	67	42.7-109	12	40	
Acenaphthylene	ug/L	0.051U	10	10	7.8	7.6	78	76	53.2-107	2	40	
Anthracene	ug/L	0.051U	10	10	7.9	7.9	79	79	52.2-112	.1	40	
Benzo(a)anthracene	ug/L	0.061U	10	10	8.1	7.9	81	79	57.5-115	2	40	
Benzo(a)pyrene	ug/L	0.051U	10	10	6.0	6.5	60	65	61.8-104	8	40	J(M1)
Benzo(b)fluoranthene	ug/L	0.051U	10	10	7.1	7.9	71	79	61.6-120	11	40	
Benzo(g,h,i)perylene	ug/L	0.061U	10	10	5.9	6.2	59	62	41.6-122	5	40	
Benzo(k)fluoranthene	ug/L	0.041U	10	10	6.5	6.4	65	64	53.3-106	.7	40	
Chrysene	ug/L	0.061U	10	10	6.8	6.7	68	67	48-121	.5	40	
Dibenz(a,h)anthracene	ug/L	0.051U	10	10	5.3	5.4	53	54	38.3-110	2	40	
Fluoranthene	ug/L	0.061U	10	10	8.1	8.4	81	84	46.8-122	4	40	
Fluorene	ug/L	0.031U	10	10	7.3	7.5	73	75	50.5-107	2	40	
Indeno(1,2,3-cd)pyrene	ug/L	0.041U	10	10	5.0	5.8	50	58	42.4-108	16	40	
Naphthalene	ug/L	0.082U	10	10	7.4	6.6	74	66	43.9-99.	11	40	
Phenanthrene	ug/L	0.051U	10	10	7.6	7.2	76	72	54.3-107	4	40	
Pyrene	ug/L	0.061U	10	10	7.9	8.2	79	82	48.5-120	4	40	
2-Fluorobiphenyl (S)	%						79	81	43.9-113			
Terphenyl-d14 (S)	%						63	71	24.8-144			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: OEXT/3415 Analysis Method: EPA 8270 by SCAN
QC Batch Method: EPA 3510 Analysis Description: 8270 Water CPAH by SCAN MSSV
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 135074 Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1-Methylnaphthalene	ug/L	0.090U	1.5	11/01/10 09:16	
2-Methylnaphthalene	ug/L	0.060U	1.5	11/01/10 09:16	
Acenaphthene	ug/L	0.030U	1.0	11/01/10 09:16	
Acenaphthylene	ug/L	0.050U	2.0	11/01/10 09:16	
Anthracene	ug/L	0.050U	1.0	11/01/10 09:16	
Benzo(a)anthracene	ug/L	0.060U	0.20	11/01/10 09:16	
Benzo(a)pyrene	ug/L	0.050U	0.20	11/01/10 09:16	
Benzo(b)fluoranthene	ug/L	0.050U	0.10	11/01/10 09:16	
Benzo(g,h,i)perylene	ug/L	0.060U	1.0	11/01/10 09:16	
Benzo(k)fluoranthene	ug/L	0.040U	0.25	11/01/10 09:16	
Chrysene	ug/L	0.060U	1.0	11/01/10 09:16	
Dibenz(a,h)anthracene	ug/L	0.050U	0.20	11/01/10 09:16	
Fluoranthene	ug/L	0.060U	1.0	11/01/10 09:16	
Fluorene	ug/L	0.030U	1.0	11/01/10 09:16	
Indeno(1,2,3-cd)pyrene	ug/L	0.040U	0.15	11/01/10 09:16	
Naphthalene	ug/L	0.080U	1.0	11/01/10 09:16	
Phenanthrene	ug/L	0.050U	1.0	11/01/10 09:16	
Pyrene	ug/L	0.060U	1.0	11/01/10 09:16	
2-Fluorobiphenyl (S)	%	86	43.9-113	11/01/10 09:16	
Terphenyl-d14 (S)	%	86	24.8-144	11/01/10 09:16	

LABORATORY CONTROL SAMPLE: 135075

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1-Methylnaphthalene	ug/L	5	3.8	76	46.7-104	
2-Methylnaphthalene	ug/L	5	3.6	72	49.4-106	
Acenaphthene	ug/L	5	3.4	68	42.7-109	
Acenaphthylene	ug/L	5	3.5	70	53.2-107	
Anthracene	ug/L	5	3.6	72	52.2-112	
Benzo(a)anthracene	ug/L	5	4.5	91	57.5-115	
Benzo(a)pyrene	ug/L	5	4.0	81	61.8-104	
Benzo(b)fluoranthene	ug/L	5	4.1	81	61.6-120	
Benzo(g,h,i)perylene	ug/L	5	4.2	84	41.6-122	
Benzo(k)fluoranthene	ug/L	5	4.4	88	53.3-106	
Chrysene	ug/L	5	4.6	91	48-121	
Dibenz(a,h)anthracene	ug/L	5	3.5	70	38.3-110	
Fluoranthene	ug/L	5	3.9	78	46.8-122	
Fluorene	ug/L	5	3.6	73	50.5-107	
Indeno(1,2,3-cd)pyrene	ug/L	5	4.3	85	42.4-108	
Naphthalene	ug/L	5	3.6	72	43.9-99.6	
Phenanthrene	ug/L	5	3.6	73	54.3-107	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 302 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 135075

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Pyrene	ug/L	5	3.8	76	48.5-120	
2-Fluorobiphenyl (S)	%			82	43.9-113	
Terphenyl-d14 (S)	%			84	24.8-144	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135757 135758

Parameter	Units	3521149001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1-Methylnaphthalene	ug/L	0.086U	10	10	7.7	7.2	77	72	46.7-104	7	40	
2-Methylnaphthalene	ug/L	0.057U	10	10	7.9	7.2	79	72	49.4-106	9	40	
Acenaphthene	ug/L	0.029U	10	10	7.5	6.9	75	69	42.7-109	8	40	
Acenaphthylene	ug/L	0.048U	10	10	7.1	7.0	71	70	53.2-107	2	40	
Anthracene	ug/L	0.048U	10	10	7.7	7.6	77	76	52.2-112	.7	40	
Benzo(a)anthracene	ug/L	0.057U	10	10	8.6	9.0	86	90	57.5-115	5	40	
Benzo(a)pyrene	ug/L	0.048U	10	10	8.1	7.9	81	79	61.8-104	2	40	
Benzo(b)fluoranthene	ug/L	0.048U	10	10	8.2	7.7	82	77	61.6-120	7	40	
Benzo(g,h,i)perylene	ug/L	0.057U	10	10	9.1	8.7	91	87	41.6-122	4	40	
Benzo(k)fluoranthene	ug/L	0.038U	10	10	8.6	9.0	86	90	53.3-106	5	40	
Chrysene	ug/L	0.057U	10	10	8.8	9.1	88	91	48-121	4	40	
Dibenz(a,h)anthracene	ug/L	0.048U	10	10	8.3	7.7	83	77	38.3-110	8	40	
Fluoranthene	ug/L	0.057U	10	10	8.1	7.6	81	76	46.8-122	7	40	
Fluorene	ug/L	0.029U	10	10	7.4	7.2	74	72	50.5-107	3	40	
Indeno(1,2,3-cd)pyrene	ug/L	0.038U	10	10	9.2	8.8	92	88	42.4-108	4	40	
Naphthalene	ug/L	0.076U	10	10	7.6	7.7	76	77	43.9-99.	3	40	
Phenanthrene	ug/L	0.048U	10	10	7.6	7.4	76	74	54.3-107	3	40	
Pyrene	ug/L	0.057U	10	10	8.0	7.8	80	78	48.5-120	2	40	
2-Fluorobiphenyl (S)	%						84	86	43.9-113			
Terphenyl-d14 (S)	%						83	80	24.8-144			

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2205 Analysis Method: EPA 8260
QC Batch Method: EPA 8260 Analysis Description: 8260 MSV
Associated Lab Samples: 3519325001, 3519325002, 3519325004

METHOD BLANK: 124803 Matrix: Water

Associated Lab Samples: 3519325001, 3519325002, 3519325004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/07/10 07:14	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/07/10 07:14	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/07/10 07:14	
1,2-Dichlorobenzene	ug/L	0.50U	1.0	10/07/10 07:14	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/07/10 07:14	
1,4-Dichlorobenzene	ug/L	0.50U	1.0	10/07/10 07:14	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/07/10 07:14	
2-Hexanone	ug/L	5.0U	10.0	10/07/10 07:14	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/07/10 07:14	
Acetone	ug/L	5.0U	10.0	10/07/10 07:14	
Acrylonitrile	ug/L	5.0U	10.0	10/07/10 07:14	
Benzene	ug/L	0.50U	1.0	10/07/10 07:14	
Bromochloromethane	ug/L	0.50U	1.0	10/07/10 07:14	
Bromodichloromethane	ug/L	0.27U	0.60	10/07/10 07:14	
Bromoform	ug/L	0.50U	1.0	10/07/10 07:14	
Bromomethane	ug/L	0.50U	1.0	10/07/10 07:14	
Carbon disulfide	ug/L	0.50U	1.0	10/07/10 07:14	
Carbon tetrachloride	ug/L	0.50U	1.0	10/07/10 07:14	
Chlorobenzene	ug/L	0.50U	1.0	10/07/10 07:14	
Chloroethane	ug/L	0.50U	1.0	10/07/10 07:14	
Chloroform	ug/L	0.50U	1.0	10/07/10 07:14	
Chloromethane	ug/L	0.62U	1.0	10/07/10 07:14	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/07/10 07:14	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/07/10 07:14	
Dibromochloromethane	ug/L	0.26U	0.50	10/07/10 07:14	
Dibromomethane	ug/L	0.50U	1.0	10/07/10 07:14	
Ethylbenzene	ug/L	0.50U	1.0	10/07/10 07:14	
Iodomethane	ug/L	0.50U	1.0	10/07/10 07:14	
Methylene Chloride	ug/L	2.5U	5.0	10/07/10 07:14	
Styrene	ug/L	0.50U	1.0	10/07/10 07:14	
Tetrachloroethene	ug/L	0.50U	1.0	10/07/10 07:14	
Toluene	ug/L	0.50U	1.0	10/07/10 07:14	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/07/10 07:14	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/07/10 07:14	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/07/10 07:14	
Trichloroethene	ug/L	0.50U	1.0	10/07/10 07:14	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/07/10 07:14	
Vinyl acetate	ug/L	1.0U	2.0	10/07/10 07:14	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 304 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 124803

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002, 3519325004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Vinyl chloride	ug/L	0.50U	1.0	10/07/10 07:14	
Xylene (Total)	ug/L	0.50U	1.0	10/07/10 07:14	
1,2-Dichloroethane-d4 (S)	%	110	86-125	10/07/10 07:14	
4-Bromofluorobenzene (S)	%	96	70-114	10/07/10 07:14	
Dibromofluoromethane (S)	%	99	88-117	10/07/10 07:14	
Toluene-d8 (S)	%	98	87-113	10/07/10 07:14	

LABORATORY CONTROL SAMPLE: 124804

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	23.7	119	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	23.0	115	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	22.5	113	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	22.5	112	84.1-122.6	
1,1-Dichloroethane	ug/L	20	23.0	115	66.4-138.6	
1,1-Dichloroethene	ug/L	20	22.1	110	79.3-127.5	
1,2,3-Trichloropropane	ug/L	20	22.1	110	58.2-134.6	
1,2-Dichlorobenzene	ug/L	20	22.8	114	91.7-127	
1,2-Dichloroethane	ug/L	20	22.7	113	85.9-121.9	
1,2-Dichloropropane	ug/L	20	23.4	117	82.2-129.1	
1,4-Dichlorobenzene	ug/L	20	21.9	109	91.9-121.7	
2-Butanone (MEK)	ug/L	20	21.0	105	53.8-156.3	
2-Hexanone	ug/L	20	20.6	103	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	21.4	107	71.8-134.4	
Acetone	ug/L	20	21.9	110	47.2-184.1	
Acrylonitrile	ug/L	200	224	112	57.8-125.9	
Benzene	ug/L	20	22.9	114	77.3-132.8	
Bromochloromethane	ug/L	20	23.9	120	87.4-122.8	
Bromodichloromethane	ug/L	20	23.0	115	77.2-121.1	
Bromoform	ug/L	20	18.7	94	65.9-133.5	
Bromomethane	ug/L	20	28.2	141	48.2-223.9	
Carbon disulfide	ug/L	20	25.9	130	20.3-195.4	
Carbon tetrachloride	ug/L	20	22.4	112	69-155.5	
Chlorobenzene	ug/L	20	23.1	116	76.9-123.9	
Chloroethane	ug/L	20	24.5	122	46.7-157.8	
Chloroform	ug/L	20	22.4	112	69.7-132	
Chloromethane	ug/L	20	25.2	126	54.4-153.8	
cis-1,2-Dichloroethene	ug/L	20	22.8	114	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	23.1	116	73-121.6	
Dibromochloromethane	ug/L	20	23.2	116	65.4-126.2	
Dibromomethane	ug/L	20	22.7	113	85.3-121.7	
Ethylbenzene	ug/L	20	23.4	117	66.4-134.4	
Iodomethane	ug/L	20	26.9	134	1-243.3	
Methylene Chloride	ug/L	20	23.8	119	65.7-137.3	
Styrene	ug/L	20	23.2	116	76.5-118.5	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 305 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 124804

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Tetrachloroethene	ug/L	20	25.4	127	71-134	
Toluene	ug/L	20	22.9	115	75-129	
trans-1,2-Dichloroethene	ug/L	20	23.2	116	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	22.4	112	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	23.2	116	36.1-177.4	
Trichloroethene	ug/L	20	22.1	111	81.1-122.4	
Trichlorofluoromethane	ug/L	20	23.8	119	75.4-124.6	
Vinyl acetate	ug/L	20	22.9	114	72.2-139	
Vinyl chloride	ug/L	20	22.7	114	70.2-136.9	
Xylene (Total)	ug/L	60	69.0	115	82.3-126	
1,2-Dichloroethane-d4 (S)	%			99	86-125	
4-Bromofluorobenzene (S)	%			100	70-114	
Dibromofluoromethane (S)	%			104	88-117	
Toluene-d8 (S)	%			100	87-113	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 124805

124806

Parameter	Units	3519705002		MS Spike Conc.	MSD Spike Conc.	MS Result		MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
		Result	Conc.	Conc.	Conc.	Result	Result	Result	Result	Result	Result	RPD	RPD
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	21.1	20.8	106	104	70-130	2	40		
1,1,1-Trichloroethane	ug/L	0.50U	20	20	21.7	21.8	109	109	70-130	.6	40		
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	19.8	20.2	99	101	70-130	2	40		
1,1,2-Trichloroethane	ug/L	0.50U	20	20	19.7	20.0	98	100	70-130	2	40		
1,1-Dichloroethane	ug/L	3.4	20	20	24.3	24.9	104	107	70-130	2	40		
1,1-Dichloroethene	ug/L	0.50U	20	20	20.8	21.2	104	106	70-130	2	40		
1,2,3-Trichloropropane	ug/L	0.36U	20	20	18.3	19.4	92	97	70-130	5	40		
1,2-Dichlorobenzene	ug/L	3.2	20	20	22.7	23.3	98	101	70-130	2	40		
1,2-Dichloroethane	ug/L	0.50U	20	20	20.0	20.6	100	103	70-130	3	40		
1,2-Dichloropropane	ug/L	0.50U	20	20	19.8	21.6	99	108	70-130	9	40		
1,4-Dichlorobenzene	ug/L	0.50U	20	20	19.7	20.2	96	99	70-130	3	40		
2-Butanone (MEK)	ug/L	5.0U	20	20	16.9	20.5	85	103	70-130	19	40		
2-Hexanone	ug/L	5.0U	20	20	16.5	18.5	83	92	70-130	11	40		
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	18.4	21.1	92	106	70-130	14	40		
Acetone	ug/L	5.0U	20	20	14.2	17.5	51	68	70-130	21	40	J(M1)	
Acrylonitrile	ug/L	5.0U	200	200	207	210	103	105	70-130	2	40		
Benzene	ug/L	1.6	20	20	21.7	22.4	101	104	70-130	3	40		
Bromochloromethane	ug/L	0.50U	20	20	20.5	21.4	102	107	70-130	4	40		
Bromodichloromethane	ug/L	0.27U	20	20	22.3	21.4	111	107	70-130	4	40		
Bromoform	ug/L	0.50U	20	20	18.6	18.0	93	90	70-130	3	40		
Bromomethane	ug/L	0.50U	20	20	19.2	22.2	96	111	70-130	14	40		
Carbon disulfide	ug/L	0.50U	20	20	20.9	23.6	104	117	70-130	12	40		
Carbon tetrachloride	ug/L	0.50U	20	20	23.0	22.3	115	112	70-130	3	40		
Chlorobenzene	ug/L	1.0	20	20	20.6	21.3	98	101	70-130	3	40		
Chloroethane	ug/L	0.50U	20	20	19.3	19.8	96	99	70-130	2	40		
Chloroform	ug/L	3.0	20	20	22.8	22.9	99	100	70-130	.5	40		
Chloromethane	ug/L	0.62U	20	20	22.6	20.9	113	104	70-130	8	40		
cis-1,2-Dichloroethene	ug/L	1.7	20	20	20.8	23.0	96	106	70-130	10	40		

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 306 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 124805 124806											
Parameter	Units	3519705002		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						
				Conc.	Conc.	Result	Result	% Rec	% Rec	Limits	RPD
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	20	21.0	21.6	105	108	70-130	3
Dibromochloromethane	ug/L	0.26U	20	20	20	21.7	20.8	108	104	70-130	4
Dibromomethane	ug/L	0.50U	20	20	20	20.0	20.5	100	102	70-130	2
Ethylbenzene	ug/L	0.50U	20	20	20	20.8	21.4	103	106	70-130	3
Iodomethane	ug/L	0.50U	20	20	20	20.8	23.2	104	116	70-130	11
Methylene Chloride	ug/L	2.5U	20	20	20	19.5	20.9	97	104	70-130	7
Styrene	ug/L	0.50U	20	20	20	20.4	20.7	102	103	70-130	1
Tetrachloroethene	ug/L	0.50U	20	20	20	19.5	19.9	97	99	70-130	2
Toluene	ug/L	1.4	20	20	20	21.6	22.0	101	103	70-130	2
trans-1,2-Dichloroethene	ug/L	1.6	20	20	20	22.8	22.8	106	106	70-130	.07
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	20	21.5	20.8	107	104	70-130	3
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	20	19.8	19.8	99	99	70-130	.1
Trichloroethene	ug/L	0.50U	20	20	20	20.2	21.5	101	108	70-130	6
Trichlorofluoromethane	ug/L	0.50U	20	20	20	22.2	22.4	111	112	70-130	1
Vinyl acetate	ug/L	1.0U	20	20	20	18.9	21.6	94	108	70-130	14
Vinyl chloride	ug/L	12.6	20	20	20	33.3	33.4	104	104	70-130	.07
Xylene (Total)	ug/L	0.62 I	60	60	60	61.1	62.9	101	104	70-130	3
1,2-Dichloroethane-d4 (S)	%							100	101	86-125	
4-Bromofluorobenzene (S)	%							102	102	70-114	
Dibromofluoromethane (S)	%							103	105	88-117	
Toluene-d8 (S)	%							101	102	87-113	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2254

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325006, 3519325007

METHOD BLANK: 130162

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/18/10 12:18	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/18/10 12:18	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/18/10 12:18	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/18/10 12:18	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/18/10 12:18	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/18/10 12:18	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/18/10 12:18	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/18/10 12:18	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/18/10 12:18	
2-Hexanone	ug/L	5.0U	10.0	10/18/10 12:18	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/18/10 12:18	
Acetone	ug/L	5.0U	10.0	10/18/10 12:18	
Acetonitrile	ug/L	5.0U	10.0	10/18/10 12:18	
Acrolein	ug/L	10.0U	20.0	10/18/10 12:18	
Acrylonitrile	ug/L	5.0U	10.0	10/18/10 12:18	
Allyl chloride	ug/L	0.50U	1.0	10/18/10 12:18	
Benzene	ug/L	0.50U	1.0	10/18/10 12:18	
Bromochloromethane	ug/L	0.50U	1.0	10/18/10 12:18	
Bromodichloromethane	ug/L	0.27U	0.60	10/18/10 12:18	
Bromoform	ug/L	0.50U	1.0	10/18/10 12:18	
Bromomethane	ug/L	0.50U	1.0	10/18/10 12:18	
Carbon disulfide	ug/L	0.50U	1.0	10/18/10 12:18	
Carbon tetrachloride	ug/L	0.50U	1.0	10/18/10 12:18	
Chlorobenzene	ug/L	0.50U	1.0	10/18/10 12:18	
Chloroethane	ug/L	0.50U	1.0	10/18/10 12:18	
Chloroform	ug/L	0.50U	1.0	10/18/10 12:18	
Chloromethane	ug/L	0.62U	1.0	10/18/10 12:18	
Chloroprene	ug/L	0.50U	1.0	10/18/10 12:18	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/18/10 12:18	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/18/10 12:18	
Dibromochloromethane	ug/L	0.26U	0.50	10/18/10 12:18	
Dibromomethane	ug/L	0.50U	1.0	10/18/10 12:18	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/18/10 12:18	
Ethyl methacrylate	ug/L	0.50U	1.0	10/18/10 12:18	
Ethylbenzene	ug/L	0.50U	1.0	10/18/10 12:18	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/18/10 12:18	
Iodomethane	ug/L	3.1	1.0	10/18/10 12:18	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/18/10 12:18	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 308 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 130162

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Methacrylonitrile	ug/L	5.0U	10.0	10/18/10 12:18	
Methyl methacrylate	ug/L	5.0U	10.0	10/18/10 12:18	
Methylene Chloride	ug/L	2.5U	5.0	10/18/10 12:18	
Propionitrile	ug/L	5.0U	10.0	10/18/10 12:18	
Styrene	ug/L	0.50U	1.0	10/18/10 12:18	
Tetrachloroethene	ug/L	0.50U	1.0	10/18/10 12:18	
Toluene	ug/L	0.50U	1.0	10/18/10 12:18	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/18/10 12:18	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/18/10 12:18	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/18/10 12:18	
Trichloroethene	ug/L	0.50U	1.0	10/18/10 12:18	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/18/10 12:18	
Vinyl acetate	ug/L	1.0U	2.0	10/18/10 12:18	
Vinyl chloride	ug/L	0.50U	1.0	10/18/10 12:18	
Xylene (Total)	ug/L	0.50U	1.0	10/18/10 12:18	
1,2-Dichloroethane-d4 (S)	%	107	86-125	10/18/10 12:18	
4-Bromofluorobenzene (S)	%	98	70-114	10/18/10 12:18	
Dibromofluoromethane (S)	%	105	88-117	10/18/10 12:18	
Toluene-d8 (S)	%	103	87-113	10/18/10 12:18	

LABORATORY CONTROL SAMPLE: 130163

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.1	101	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	21.1	106	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	18.2	91	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	18.7	94	84.1-122.6	
1,1-Dichloroethane	ug/L	20	20.2	101	66.4-138.6	
1,1-Dichloroethene	ug/L	20	19.5	98	79.3-127.5	
1,1-Dichloropropene	ug/L	20	20.0	100	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	18.1	91	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	19.2	96	79.1-134.1	
1,2-Dichloroethane	ug/L	20	20.0	100	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.9	104	82.2-129.1	
1,3-Dichloropropane	ug/L	20	19.4	97	88.1-118.2	
2,2-Dichloropropane	ug/L	20	21.5	108	44-181.7	
2-Butanone (MEK)	ug/L	20	17.7	88	53.8-156.3	
2-Hexanone	ug/L	20	19.2	96	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	18.5	92	71.8-134.4	
Acetone	ug/L	20	20.3	101	47.2-184.1	
Acetonitrile	ug/L	200	193	97	65.2-133.1	
Acrolein	ug/L	200	185	92	41.8-131.7	
Acrylonitrile	ug/L	200	193	97	57.8-125.9	
Allyl chloride	ug/L	20	20.2	101	23.6-190.7	
Benzene	ug/L	20	21.0	105	77.3-132.8	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130163

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Bromochloromethane	ug/L	20	19.8	99	87.4-122.8	
Bromodichloromethane	ug/L	20	20.4	102	77.2-121.1	
Bromoform	ug/L	20	18.0	90	65.9-133.5	
Bromomethane	ug/L	20	20.4	102	48.2-223.9	
Carbon disulfide	ug/L	20	18.5	92	20.3-195.4	
Carbon tetrachloride	ug/L	20	21.0	105	69-155.5	
Chlorobenzene	ug/L	20	19.9	100	76.9-123.9	
Chloroethane	ug/L	20	19.0	95	46.7-157.8	
Chloroform	ug/L	20	20.0	100	69.7-132	
Chloromethane	ug/L	20	18.4	92	54.4-153.8	
Chloroprene	ug/L	20	20.0	100	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	20.5	103	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	20.8	104	73-121.6	
Dibromochloromethane	ug/L	20	20.2	101	65.4-126.2	
Dibromomethane	ug/L	20	19.3	97	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	20.7	104	63.1-143.7	
Ethyl methacrylate	ug/L	20	18.9	94	34.3-179.4	
Ethylbenzene	ug/L	20	20.1	100	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	18.8	94	74.4-153.6	
Iodomethane	ug/L	20	16.7	84	1-243.3	
Isobutyl Alcohol	ug/L	400	382	96	62.9-136.1	
Methacrylonitrile	ug/L	200	202	101	77.3-132.6	
Methyl methacrylate	ug/L	20	18.7	94	37.4-178.3	
Methylene Chloride	ug/L	20	18.4	92	65.7-137.3	
Propionitrile	ug/L	200	201	100	71-130.3	
Styrene	ug/L	20	20.6	103	76.5-118.5	
Tetrachloroethene	ug/L	20	19.8	99	71-134	
Toluene	ug/L	20	20.2	101	75-129	
trans-1,2-Dichloroethene	ug/L	20	20.1	101	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	20.0	100	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	18.9	95	36.1-177.4	
Trichloroethene	ug/L	20	20.1	101	81.1-122.4	
Trichlorofluoromethane	ug/L	20	21.3	106	75.4-124.6	
Vinyl acetate	ug/L	20	18.8	94	72.2-139	
Vinyl chloride	ug/L	20	21.3	106	70.2-136.9	
Xylene (Total)	ug/L	60	61.4	102	82.3-126	
1,2-Dichloroethane-d4 (S)	%			97	86-125	
4-Bromofluorobenzene (S)	%			106	70-114	
Dibromofluoromethane (S)	%			102	88-117	
Toluene-d8 (S)	%			103	87-113	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130888

130889

Parameter	Units	3520507007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,1,1,2-Tetrachloroethane	ug/L	0.500U	20	20	18.8	18.8	94	94	70-130	.09	40	
1,1,1-Trichloroethane	ug/L	0.500U	20	20	19.8	20.0	99	100	70-130	.8	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 310 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130888 130889											
Parameter	Units	3520507007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
1,1,2,2-Tetrachloroethane	ug/L	0.180U	20	20	17.2	17.2	86	86	70-130	.5	40
1,1,2-Trichloroethane	ug/L	0.500U	20	20	18.0	17.8	90	89	70-130	1	40
1,1-Dichloroethane	ug/L	0.500U	20	20	19.5	19.9	98	99	70-130	2	40
1,1-Dichloroethene	ug/L	0.500U	20	20	11.9	12.3	60	62	70-130	3	40 J(M1)
1,1-Dichloropropene	ug/L	0.500U	20	20	20.1	20.7	101	103	70-130	3	40
1,2,3-Trichloropropane	ug/L	0.360U	20	20	17.6	17.5	88	88	70-130	.5	40
1,2,4-Trichlorobenzene	ug/L	0.500U	20	20	18.6	18.7	93	94	70-130	.9	40
1,2-Dichloroethane	ug/L	0.500U	20	20	15.1	15.4	76	77	70-130	2	40
1,2-Dichloropropane	ug/L	0.500U	20	20	19.7	20.3	98	102	70-130	3	40
1,3-Dichloropropane	ug/L	0.500U	20	20	17.8	17.8	89	89	70-130	.08	40
2,2-Dichloropropane	ug/L	0.500U	20	20	19.0	19.0	95	95	70-130	.1	40
2-Butanone (MEK)	ug/L	5.00U	20	20	13.5	13.3	68	67	70-130	1	40 J(M1)
2-Hexanone	ug/L	5.00U	20	20	13.4	13.2	67	66	70-130	1	40 J(M1)
4-Methyl-2-pentanone (MIBK)	ug/L	5.00U	20	20	15.5	15.0	78	75	70-130	3	40
Acetone	ug/L	5.00U	20	20	7.2	7.9	25	29	70-130		40 J(M1)
Acetonitrile	ug/L	5.00U	200	200	159	157	79	78	70-130	1	40
Acrolein	ug/L	10.0U	200	200	89.3	87.2	45	44	70-130	2	40 J(M1)
Acrylonitrile	ug/L	5.00U	200	200	165	159	82	80	70-130	3	40
Allyl chloride	ug/L	0.500U	20	20	17.8	18.3	89	92	70-130	3	40
Benzene	ug/L	0.500U	20	20	20.6	20.9	103	105	70-130	2	40
Bromochloromethane	ug/L	0.500U	20	20	20.4	20.9	102	104	70-130	2	40
Bromodichloromethane	ug/L	0.270U	20	20	19.4	20.0	97	100	70-130	3	40
Bromoform	ug/L	0.500U	20	20	17.6	17.7	88	88	70-130	.4	40
Bromomethane	ug/L	0.500U	20	20	11.9	11.7	60	59	70-130	2	40 J(M1)
Carbon disulfide	ug/L	0.500U	20	20	19.2	19.4	94	95	70-130	.9	40
Carbon tetrachloride	ug/L	0.500U	20	20	20.6	20.8	103	104	70-130	1	40
Chlorobenzene	ug/L	0.500U	20	20	19.2	19.4	96	97	70-130	.8	40
Chloroethane	ug/L	0.500U	20	20	10.9	10.4	55	52	70-130	5	40 J(M1)
Chloroform	ug/L	1.34	20	20	18.6	19.2	87	89	70-130	3	40
Chloromethane	ug/L	0.620U	20	20	13.7	13.2	68	66	70-130	4	40 J(M1)
Chloroprene	ug/L	0.500U	20	20	17.7	18.0	89	90	70-130	2	40
cis-1,2-Dichloroethene	ug/L	0.500U	20	20	18.6	19.0	93	95	70-130	2	40
cis-1,3-Dichloropropene	ug/L	0.250U	20	20	20.5	20.9	102	104	70-130	2	40
Dibromochloromethane	ug/L	0.260U	20	20	19.9	19.6	100	98	70-130	1	40
Dibromomethane	ug/L	0.500U	20	20	19.7	19.9	98	100	70-130	1	40
Dichlorodifluoromethane	ug/L	0.500U	20	20	20.3	19.5	101	98	70-130	4	40
Ethyl methacrylate	ug/L	0.500U	20	20	15.5	14.7	77	73	70-130	5	40
Ethylbenzene	ug/L	0.500U	20	20	19.2	19.3	96	97	70-130	.8	40
Hexachloro-1,3-butadiene	ug/L	0.500U	20	20	18.5	18.8	92	94	70-130	2	40
Iodomethane	ug/L	0.500U	20	20	13.7	16.5	69	83	70-130	18	40 J(M1)
Isobutyl Alcohol	ug/L	10.0U	400	400	320	321	80	80	70-130	.4	40
Methacrylonitrile	ug/L	5.00U	200	200	144	146	72	73	70-130	1	40
Methyl methacrylate	ug/L	5.00U	20	20	19.2	18.7	96	94	70-130	2	40
Methylene Chloride	ug/L	2.50U	20	20	14.3	14.6	71	73	70-130	2	40
Propionitrile	ug/L	5.00U	200	200	168	170	84	85	70-130	1	40
Styrene	ug/L	0.500U	20	20	19.1	19.2	96	96	70-130	.5	40
Tetrachloroethene	ug/L	0.500U	20	20	18.7	18.9	94	95	70-130	1	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 311 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130888 130889											
Parameter	Units	3520507007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Toluene	ug/L	0.646 I	20	20	20.9	20.2	102	98	70-130	4	40
trans-1,2-Dichloroethene	ug/L	0.500U	20	20	19.3	19.4	96	97	70-130	.7	40
trans-1,3-Dichloropropene	ug/L	0.250U	20	20	18.9	18.6	94	93	70-130	1	40
trans-1,4-Dichloro-2-butene	ug/L	5.00U	20	20	16.5	15.1	83	76	70-130	9	40
Trichloroethene	ug/L	0.500U	20	20	21.4	21.4	107	107	70-130	.01	40
Trichlorofluoromethane	ug/L	0.500U	20	20	15.0	15.2	75	76	70-130	1	40
Vinyl acetate	ug/L	1.00U	20	20	14.4	14.2	72	71	70-130	1	40
Vinyl chloride	ug/L	0.500U	20	20	12.3	11.9	62	60	70-130	3	40 J(M1)
Xylene (Total)	ug/L	0.500U	60	60	56.8	56.8	95	95	70-130	.09	40
1,2-Dichloroethane-d4 (S)	%						74	74	86-125		J(S0)
4-Bromofluorobenzene (S)	%						101	100	70-114		
Dibromofluoromethane (S)	%						97	97	88-117		
Toluene-d8 (S)	%						103	103	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2258

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325009, 3519325010

METHOD BLANK: 130790

Matrix: Water

Associated Lab Samples: 3519325009, 3519325010

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/19/10 16:25	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/19/10 16:25	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/19/10 16:25	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/19/10 16:25	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/19/10 16:25	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/19/10 16:25	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/19/10 16:25	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/19/10 16:25	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/19/10 16:25	
2-Hexanone	ug/L	5.0U	10.0	10/19/10 16:25	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/19/10 16:25	
Acetone	ug/L	5.0U	10.0	10/19/10 16:25	
Acetonitrile	ug/L	5.0U	10.0	10/19/10 16:25	
Acrolein	ug/L	10.0U	20.0	10/19/10 16:25	
Acrylonitrile	ug/L	5.0U	10.0	10/19/10 16:25	
Allyl chloride	ug/L	0.50U	1.0	10/19/10 16:25	
Benzene	ug/L	0.50U	1.0	10/19/10 16:25	
Bromochloromethane	ug/L	0.50U	1.0	10/19/10 16:25	
Bromodichloromethane	ug/L	0.27U	0.60	10/19/10 16:25	
Bromoform	ug/L	0.50U	1.0	10/19/10 16:25	
Bromomethane	ug/L	0.50U	1.0	10/19/10 16:25	
Carbon disulfide	ug/L	0.50U	1.0	10/19/10 16:25	
Carbon tetrachloride	ug/L	0.50U	1.0	10/19/10 16:25	
Chlorobenzene	ug/L	0.50U	1.0	10/19/10 16:25	
Chloroethane	ug/L	0.50U	1.0	10/19/10 16:25	
Chloroform	ug/L	0.50U	1.0	10/19/10 16:25	
Chloromethane	ug/L	0.62U	1.0	10/19/10 16:25	
Chloroprene	ug/L	0.50U	1.0	10/19/10 16:25	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/19/10 16:25	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/19/10 16:25	
Dibromochloromethane	ug/L	0.26U	0.50	10/19/10 16:25	
Dibromomethane	ug/L	0.50U	1.0	10/19/10 16:25	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/19/10 16:25	
Ethyl methacrylate	ug/L	0.50U	1.0	10/19/10 16:25	
Ethylbenzene	ug/L	0.50U	1.0	10/19/10 16:25	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/19/10 16:25	
Iodomethane	ug/L	0.50U	1.0	10/19/10 16:25	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/19/10 16:25	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 313 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 130790

Matrix: Water

Associated Lab Samples: 3519325009, 3519325010

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Methacrylonitrile	ug/L	5.0U	10.0	10/19/10 16:25	
Methyl methacrylate	ug/L	5.0U	10.0	10/19/10 16:25	
Methylene Chloride	ug/L	2.5U	5.0	10/19/10 16:25	
Propionitrile	ug/L	5.0U	10.0	10/19/10 16:25	
Styrene	ug/L	0.50U	1.0	10/19/10 16:25	
Tetrachloroethene	ug/L	0.50U	1.0	10/19/10 16:25	
Toluene	ug/L	0.50U	1.0	10/19/10 16:25	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/19/10 16:25	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/19/10 16:25	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/19/10 16:25	
Trichloroethene	ug/L	0.50U	1.0	10/19/10 16:25	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/19/10 16:25	
Vinyl acetate	ug/L	1.0U	2.0	10/19/10 16:25	
Vinyl chloride	ug/L	0.50U	1.0	10/19/10 16:25	
Xylene (Total)	ug/L	0.50U	1.0	10/19/10 16:25	
1,2-Dichloroethane-d4 (S)	%	106	86-125	10/19/10 16:25	
4-Bromofluorobenzene (S)	%	96	70-114	10/19/10 16:25	
Dibromofluoromethane (S)	%	105	88-117	10/19/10 16:25	
Toluene-d8 (S)	%	104	87-113	10/19/10 16:25	

LABORATORY CONTROL SAMPLE: 130791

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.4	102	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	21.5	108	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	18.2	91	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	18.9	95	84.1-122.6	
1,1-Dichloroethane	ug/L	20	21.3	106	66.4-138.6	
1,1-Dichloroethene	ug/L	20	21.7	108	79.3-127.5	
1,1-Dichloropropene	ug/L	20	21.2	106	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	17.9	90	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	18.5	93	79.1-134.1	
1,2-Dichloroethane	ug/L	20	21.1	106	85.9-121.9	
1,2-Dichloropropane	ug/L	20	21.3	107	82.2-129.1	
1,3-Dichloropropane	ug/L	20	19.6	98	88.1-118.2	
2,2-Dichloropropane	ug/L	20	21.3	107	44-181.7	
2-Butanone (MEK)	ug/L	20	19.3	97	53.8-156.3	
2-Hexanone	ug/L	20	19.2	96	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	19.4	97	71.8-134.4	
Acetone	ug/L	20	20.2	101	47.2-184.1	
Acetonitrile	ug/L	200	217	108	65.2-133.1	
Acrolein	ug/L	200	206	103	41.8-131.7	
Acrylonitrile	ug/L	200	213	107	57.8-125.9	
Allyl chloride	ug/L	20	20.7	103	23.6-190.7	
Benzene	ug/L	20	21.8	109	77.3-132.8	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 130791

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Bromochloromethane	ug/L	20	21.4	107	87.4-122.8	
Bromodichloromethane	ug/L	20	21.5	107	77.2-121.1	
Bromoform	ug/L	20	18.2	91	65.9-133.5	
Bromomethane	ug/L	20	22.9	115	48.2-223.9	
Carbon disulfide	ug/L	20	20.7	104	20.3-195.4	
Carbon tetrachloride	ug/L	20	21.4	107	69-155.5	
Chlorobenzene	ug/L	20	20.2	101	76.9-123.9	
Chloroethane	ug/L	20	21.1	105	46.7-157.8	
Chloroform	ug/L	20	21.0	105	69.7-132	
Chloromethane	ug/L	20	20.3	102	54.4-153.8	
Chloroprene	ug/L	20	21.2	106	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	21.6	108	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	21.2	106	73-121.6	
Dibromochloromethane	ug/L	20	19.8	99	65.4-126.2	
Dibromomethane	ug/L	20	20.4	102	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	20.7	103	63.1-143.7	
Ethyl methacrylate	ug/L	20	20.2	101	34.3-179.4	
Ethylbenzene	ug/L	20	20.4	102	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	18.7	94	74.4-153.6	
Iodomethane	ug/L	20	17.1	85	1-243.3	
Isobutyl Alcohol	ug/L	400	438	110	62.9-136.1	
Methacrylonitrile	ug/L	200	220	110	77.3-132.6	
Methyl methacrylate	ug/L	20	19.4	97	37.4-178.3	
Methylene Chloride	ug/L	20	20.6	103	65.7-137.3	
Propionitrile	ug/L	200	211	106	71-130.3	
Styrene	ug/L	20	21.4	107	76.5-118.5	
Tetrachloroethene	ug/L	20	20.8	104	71-134	
Toluene	ug/L	20	20.1	100	75-129	
trans-1,2-Dichloroethene	ug/L	20	21.3	106	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	19.9	100	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	19.0	95	36.1-177.4	
Trichloroethene	ug/L	20	20.7	104	81.1-122.4	
Trichlorofluoromethane	ug/L	20	22.4	112	75.4-124.6	
Vinyl acetate	ug/L	20	20.1	101	72.2-139	
Vinyl chloride	ug/L	20	22.7	113	70.2-136.9	
Xylene (Total)	ug/L	60	63.0	105	82.3-126	
1,2-Dichloroethane-d4 (S)	%			101	86-125	
4-Bromofluorobenzene (S)	%			109	70-114	
Dibromofluoromethane (S)	%			104	88-117	
Toluene-d8 (S)	%			105	87-113	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130834

130835

Parameter	Units	3519452046 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	18.5	18.6	92	93	70-130	.6	40	
1,1,1-Trichloroethane	ug/L	0.50U	20	20	19.4	19.6	97	98	70-130	1	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 315 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130834 130835											
Parameter	Units	3519452046		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						RPD
				Conc.	Conc.	Result	Result	% Rec	% Rec	Limits	RPD
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	20	16.5	16.0	82	80	70-130	3 40
1,1,2-Trichloroethane	ug/L	0.50U	20	20	20	17.2	17.7	86	89	70-130	3 40
1,1-Dichloroethane	ug/L	0.50U	20	20	20	19.2	19.7	96	99	70-130	2 40
1,1-Dichloroethene	ug/L	0.50U	20	20	20	12.2	12.3	61	61	70-130	.9 40 J(M1)
1,1-Dichloropropene	ug/L	0.50U	20	20	20	20.6	20.8	103	104	70-130	1 40
1,2,3-Trichloropropane	ug/L	0.36U	20	20	20	17.1	17.6	86	88	70-130	3 40
1,2,4-Trichlorobenzene	ug/L	0.50U	20	20	20	14.8	17.2	74	86	70-130	15 40
1,2-Dichloroethane	ug/L	0.50U	20	20	20	14.8	15.2	74	76	70-130	3 40
1,2-Dichloropropane	ug/L	0.50U	20	20	20	19.7	19.6	99	98	70-130	.5 40
1,3-Dichloropropane	ug/L	0.50U	20	20	20	17.4	17.5	87	87	70-130	.3 40
2,2-Dichloropropane	ug/L	0.50U	20	20	20	19.3	19.2	96	96	70-130	.4 40
2-Butanone (MEK)	ug/L	5.0U	20	20	20	13.2	13.1	66	66	70-130	.4 40 J(M1)
2-Hexanone	ug/L	5.0U	20	20	20	12.8	13.3	64	67	70-130	4 40 J(M1)
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	20	14.7	15.1	73	76	70-130	3 40
Acetone	ug/L	5.0U	20	20	20	7.8 I	7.9 I	39	40	70-130	40 J(M1)
Acetonitrile	ug/L	5.0U	200	200	200	148	154	74	77	70-130	4 40
Acrolein	ug/L	10.0U	200	200	200	79.1	84.4	40	42	70-130	7 40 J(M1)
Acrylonitrile	ug/L	5.0U	200	200	200	153	159	77	79	70-130	3 40
Allyl chloride	ug/L	0.50U	20	20	20	17.6	18.4	88	92	70-130	4 40
Benzene	ug/L	0.50U	20	20	20	20.4	20.4	102	102	70-130	.1 40
Bromochloromethane	ug/L	0.50U	20	20	20	20.2	20.4	101	102	70-130	.7 40
Bromodichloromethane	ug/L	0.27U	20	20	20	19.2	19.4	96	97	70-130	1 40
Bromoform	ug/L	0.50U	20	20	20	17.3	17.4	86	87	70-130	.6 40
Bromomethane	ug/L	0.50U	20	20	20	10.8	11.8	54	59	70-130	9 40 J(M1)
Carbon disulfide	ug/L	0.58 I	20	20	20	19.3	20.1	93	97	70-130	4 40
Carbon tetrachloride	ug/L	0.50U	20	20	20	20.2	20.7	101	104	70-130	3 40
Chlorobenzene	ug/L	0.50U	20	20	20	19.2	19.4	96	97	70-130	1 40
Chloroethane	ug/L	0.50U	20	20	20	10.3	10.6	52	53	70-130	3 40 J(M1)
Chloroform	ug/L	0.50U	20	20	20	17.6	17.3	88	86	70-130	2 40
Chloromethane	ug/L	0.62U	20	20	20	12.8	13.9	64	69	70-130	8 40 J(M1)
Chloroprene	ug/L	0.50U	20	20	20	17.3	17.9	86	90	70-130	4 40
cis-1,2-Dichloroethene	ug/L	0.50U	20	20	20	19.0	18.9	95	95	70-130	.3 40
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	20	20.3	19.9	101	100	70-130	2 40
Dibromochloromethane	ug/L	0.26U	20	20	20	19.8	19.7	99	99	70-130	.5 40
Dibromomethane	ug/L	0.50U	20	20	20	18.7	19.2	94	96	70-130	3 40
Dichlorodifluoromethane	ug/L	0.50U	20	20	20	20.3	20.8	101	104	70-130	3 40
Ethyl methacrylate	ug/L	0.50U	20	20	20	15.3	15.2	76	76	70-130	.7 40
Ethylbenzene	ug/L	0.50U	20	20	20	19.6	19.3	98	97	70-130	2 40
Hexachloro-1,3-butadiene	ug/L	0.50U	20	20	20	16.6	17.8	83	89	70-130	7 40
Iodomethane	ug/L	0.50U	20	20	20	11.7	16.6	58	83	70-130	35 40 J(M1)
Isobutyl Alcohol	ug/L	10.0U	400	400	400	273	300	68	75	70-130	9 40 J(M1)
Methacrylonitrile	ug/L	5.0U	200	200	200	139	143	70	72	70-130	3 40
Methyl methacrylate	ug/L	5.0U	20	20	20	17.7	19.0	89	95	70-130	7 40
Methylene Chloride	ug/L	2.5U	20	20	20	14.4	14.5	72	72	70-130	.6 40
Propionitrile	ug/L	5.0U	200	200	200	159	166	79	83	70-130	5 40
Styrene	ug/L	0.50U	20	20	20	19.1	19.2	96	96	70-130	.5 40
Tetrachloroethene	ug/L	0.50U	20	20	20	19.0	19.2	95	96	70-130	1 40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 316 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130834 130835											
Parameter	Units	3519452046 Result	MS		MSD		MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Spike Conc.	MS Result	MSD Result				RPD	RPD
Toluene	ug/L	0.50U	20	20	19.9	20.1	99	101	70-130	1	40
trans-1,2-Dichloroethene	ug/L	0.50U	20	20	19.4	19.3	97	97	70-130	.5	40
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	18.1	18.0	90	90	70-130	.6	40
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	16.1	18.1	80	90	70-130	12	40
Trichloroethene	ug/L	0.50U	20	20	20.9	21.3	105	107	70-130	2	40
Trichlorofluoromethane	ug/L	0.50U	20	20	15.0	15.5	75	77	70-130	3	40
Vinyl acetate	ug/L	1.0U	20	20	13.7	13.9	68	70	70-130	2	40 J(M1)
Vinyl chloride	ug/L	0.50U	20	20	11.5	12.4	58	62	70-130	7	40 J(M1)
Xylene (Total)	ug/L	0.50U	60	60	60.4	57.7	101	96	70-130	5	40
1,2-Dichloroethane-d4 (S)	%						75	75	86-125		J(S0)
4-Bromofluorobenzene (S)	%						101	102	70-114		
Dibromofluoromethane (S)	%						96	97	88-117		
Toluene-d8 (S)	%						102	102	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2283

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325008, 3519325011, 3519325012, 3519325016, 3519325018, 3519325019, 3519325020

METHOD BLANK: 134513

Matrix: Water

Associated Lab Samples: 3519325008, 3519325011, 3519325012, 3519325016, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/27/10 20:30	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:30	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/27/10 20:30	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/27/10 20:30	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:30	
1,2-Dichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:30	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:30	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:30	
1,4-Dichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:30	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:30	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/27/10 20:30	
2-Hexanone	ug/L	5.0U	10.0	10/27/10 20:30	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/27/10 20:30	
Acetone	ug/L	5.0U	10.0	10/27/10 20:30	
Acetonitrile	ug/L	5.0U	10.0	10/27/10 20:30	
Acrolein	ug/L	10.0U	20.0	10/27/10 20:30	
Acrylonitrile	ug/L	5.0U	10.0	10/27/10 20:30	
Allyl chloride	ug/L	0.50U	1.0	10/27/10 20:30	
Benzene	ug/L	0.50U	1.0	10/27/10 20:30	
Bromochloromethane	ug/L	0.50U	1.0	10/27/10 20:30	
Bromodichloromethane	ug/L	0.27U	0.60	10/27/10 20:30	
Bromoform	ug/L	0.50U	1.0	10/27/10 20:30	
Bromomethane	ug/L	0.50U	1.0	10/27/10 20:30	
Carbon disulfide	ug/L	0.50U	1.0	10/27/10 20:30	
Carbon tetrachloride	ug/L	0.50U	1.0	10/27/10 20:30	
Chlorobenzene	ug/L	0.50U	1.0	10/27/10 20:30	
Chloroethane	ug/L	0.50U	1.0	10/27/10 20:30	
Chloroform	ug/L	0.50U	1.0	10/27/10 20:30	
Chloromethane	ug/L	0.62U	1.0	10/27/10 20:30	
Chloroprene	ug/L	0.50U	1.0	10/27/10 20:30	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:30	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/27/10 20:30	
Dibromochloromethane	ug/L	0.26U	0.50	10/27/10 20:30	
Dibromomethane	ug/L	0.50U	1.0	10/27/10 20:30	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/27/10 20:30	
Ethyl methacrylate	ug/L	0.50U	1.0	10/27/10 20:30	
Ethylbenzene	ug/L	0.50U	1.0	10/27/10 20:30	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/27/10 20:30	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 318 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 134513

Matrix: Water

Associated Lab Samples: 3519325008, 3519325011, 3519325012, 3519325016, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Iodomethane	ug/L	0.50U	1.0	10/27/10 20:30	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/27/10 20:30	
Methacrylonitrile	ug/L	5.0U	10.0	10/27/10 20:30	
Methyl methacrylate	ug/L	5.0U	10.0	10/27/10 20:30	
Methylene Chloride	ug/L	2.5U	5.0	10/27/10 20:30	
Propionitrile	ug/L	5.0U	10.0	10/27/10 20:30	
Styrene	ug/L	0.50U	1.0	10/27/10 20:30	
Tetrachloroethene	ug/L	0.50U	1.0	10/27/10 20:30	
Toluene	ug/L	0.50U	1.0	10/27/10 20:30	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:30	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/27/10 20:30	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/27/10 20:30	
Trichloroethene	ug/L	0.50U	1.0	10/27/10 20:30	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/27/10 20:30	
Vinyl acetate	ug/L	1.0U	2.0	10/27/10 20:30	
Vinyl chloride	ug/L	0.50U	1.0	10/27/10 20:30	
Xylene (Total)	ug/L	0.50U	1.0	10/27/10 20:30	
1,2-Dichloroethane-d4 (S)	%	114	86-125	10/27/10 20:30	
4-Bromofluorobenzene (S)	%	100	70-114	10/27/10 20:30	
Dibromofluoromethane (S)	%	102	88-117	10/27/10 20:30	
Toluene-d8 (S)	%	101	87-113	10/27/10 20:30	

LABORATORY CONTROL SAMPLE: 134514

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.3	101	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	19.5	98	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	20.6	103	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	19.4	97	84.1-122.6	
1,1-Dichloroethane	ug/L	20	19.2	96	66.4-138.6	
1,1-Dichloroethene	ug/L	20	18.8	94	79.3-127.5	
1,1-Dichloropropene	ug/L	20	21.3	106	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	18.3	91	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	20.8	104	79.1-134.1	
1,2-Dichlorobenzene	ug/L	20	20.0	100	91.7-127	
1,2-Dichloroethane	ug/L	20	19.9	100	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.5	102	82.2-129.1	
1,3-Dichloropropane	ug/L	20	20.0	100	88.1-118.2	
1,4-Dichlorobenzene	ug/L	20	19.5	97	91.9-121.7	
2,2-Dichloropropane	ug/L	20	21.8	109	44-181.7	
2-Butanone (MEK)	ug/L	20	24.5	123	53.8-156.3	
2-Hexanone	ug/L	20	21.2	106	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	19.0	95	71.8-134.4	
Acetone	ug/L	20	26.9	134	47.2-184.1	
Acetonitrile	ug/L	200	210	105	65.2-133.1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 319 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 134514

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Acrolein	ug/L	200	222	111	41.8-131.7	
Acrylonitrile	ug/L	200	203	102	57.8-125.9	
Allyl chloride	ug/L	20	20.9	105	23.6-190.7	
Benzene	ug/L	20	20.4	102	77.3-132.8	
Bromochloromethane	ug/L	20	20.0	100	87.4-122.8	
Bromodichloromethane	ug/L	20	21.8	109	77.2-121.1	
Bromoform	ug/L	20	17.1	85	65.9-133.5	
Bromomethane	ug/L	20	19.9	99	48.2-223.9	
Carbon disulfide	ug/L	20	23.8	119	20.3-195.4	
Carbon tetrachloride	ug/L	20	19.0	95	69-155.5	
Chlorobenzene	ug/L	20	19.6	98	76.9-123.9	
Chloroethane	ug/L	20	20.6	103	46.7-157.8	
Chloroform	ug/L	20	19.3	96	69.7-132	
Chloromethane	ug/L	20	21.8	109	54.4-153.8	
Chloroprene	ug/L	20	22.2	111	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	19.7	99	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	22.2	111	73-121.6	
Dibromochloromethane	ug/L	20	19.7	99	65.4-126.2	
Dibromomethane	ug/L	20	19.8	99	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	16.7	83	63.1-143.7	
Ethyl methacrylate	ug/L	20	19.3	96	34.3-179.4	
Ethylbenzene	ug/L	20	20.2	101	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	20.5	103	74.4-153.6	
Iodomethane	ug/L	20	25.6	128	1-243.3	
Isobutyl Alcohol	ug/L	400	450	112	62.9-136.1	
Methacrylonitrile	ug/L	200	205	103	77.3-132.6	
Methyl methacrylate	ug/L	20	21.9	109	37.4-178.3	
Methylene Chloride	ug/L	20	18.6	93	65.7-137.3	
Propionitrile	ug/L	200	206	103	71-130.3	
Styrene	ug/L	20	20.3	102	76.5-118.5	
Tetrachloroethene	ug/L	20	17.0	85	71-134	
Toluene	ug/L	20	20.2	101	75-129	
trans-1,2-Dichloroethene	ug/L	20	19.7	99	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	19.6	98	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	18.6	93	36.1-177.4	
Trichloroethene	ug/L	20	19.5	98	81.1-122.4	
Trichlorofluoromethane	ug/L	20	20.3	101	75.4-124.6	
Vinyl acetate	ug/L	20	22.9	114	72.2-139	
Vinyl chloride	ug/L	20	20.6	103	70.2-136.9	
Xylene (Total)	ug/L	60	59.1	98	82.3-126	
1,2-Dichloroethane-d4 (S)	%			98	86-125	
4-Bromofluorobenzene (S)	%			99	70-114	
Dibromofluoromethane (S)	%			100	88-117	
Toluene-d8 (S)	%			101	87-113	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134515 134516											
Parameter	Units	3520506003 Result	MS		MSD		MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Conc.	Spike Conc.	Result				RPD	RPD
1,1,1,2-Tetrachloroethane	ug/L	0.500U	20	20	20	20.9	105	107	70-130	2	40
1,1,1-Trichloroethane	ug/L	0.500U	20	20	20	19.6	98	110	70-130	11	40
1,1,2,2-Tetrachloroethane	ug/L	0.180U	20	20	20	19.0	95	97	70-130	2	40
1,1,2-Trichloroethane	ug/L	0.500U	20	20	20	19.2	96	97	70-130	1	40
1,1-Dichloroethane	ug/L	0.500U	20	20	20	19.8	99	104	70-130	5	40
1,1-Dichloroethene	ug/L	0.500U	20	20	20	20.1	101	101	70-130	.6	40
1,1-Dichloropropene	ug/L	0.500U	20	20	20	22.4	112	118	70-130	5	40
1,2,3-Trichloropropane	ug/L	0.360U	20	20	20	16.7	84	87	70-130	4	40
1,2,4-Trichlorobenzene	ug/L	0.500U	20	20	20	17.8	89	99	70-130	11	40
1,2-Dichlorobenzene	ug/L	0.500U	20	20	20	19.3	96	102	70-130	6	40
1,2-Dichloroethane	ug/L	0.500U	20	20	20	18.9	95	102	70-130	8	40
1,2-Dichloropropane	ug/L	0.500U	20	20	20	19.4	97	107	70-130	10	40
1,3-Dichloropropane	ug/L	0.500U	20	20	20	18.9	95	98	70-130	4	40
1,4-Dichlorobenzene	ug/L	0.500U	20	20	20	19.3	96	102	70-130	5	40
2,2-Dichloropropane	ug/L	0.500U	20	20	20	19.6	98	102	70-130	3	40
2-Butanone (MEK)	ug/L	5.00U	20	20	20	14.6	73	65	70-130	12	40 J(M1)
2-Hexanone	ug/L	5.00U	20	20	20	15.7	78	82	70-130	5	40
4-Methyl-2-pentanone (MIBK)	ug/L	5.00U	20	20	20	17.7	89	85	70-130	4	40
Acetone	ug/L	5.00U	20	20	20	12.7	57	61	70-130	6	40 J(M1)
Acetonitrile	ug/L	5.00U	200	200	200	183	92	96	70-130	4	40
Acrolein	ug/L	10.0U	200	200	200	185	93	103	70-130	11	40
Acrylonitrile	ug/L	5.00U	200	200	200	186	93	98	70-130	5	40
Allyl chloride	ug/L	0.500U	20	20	20	22.3	112	111	70-130	.8	40
Benzene	ug/L	0.500U	20	20	20	20.7	103	110	70-130	6	40
Bromochloromethane	ug/L	0.500U	20	20	20	18.2	91	98	70-130	8	40
Bromodichloromethane	ug/L	0.270U	20	20	20	20.7	103	114	70-130	10	40
Bromoform	ug/L	0.500U	20	20	20	15.6	78	81	70-130	4	40
Bromomethane	ug/L	0.500U	20	20	20	21.7	108	114	70-130	5	40
Carbon disulfide	ug/L	0.500U	20	20	20	24.9	122	132	70-130	8	40 J(M1)
Carbon tetrachloride	ug/L	0.500U	20	20	20	20.3	102	109	70-130	7	40
Chlorobenzene	ug/L	0.500U	20	20	20	19.9	99	104	70-130	4	40
Chloroethane	ug/L	0.500U	20	20	20	21.5	107	106	70-130	.8	40
Chloroform	ug/L	0.500U	20	20	20	20.1	101	112	70-130	11	40
Chloromethane	ug/L	0.620U	20	20	20	22.6	110	118	70-130	7	40
Chloroprene	ug/L	0.500U	20	20	20	23.4	117	115	70-130	1	40
cis-1,2-Dichloroethene	ug/L	0.500U	20	20	20	20.0	100	100	70-130	.03	40
cis-1,3-Dichloropropene	ug/L	0.250U	20	20	20	20.1	101	110	70-130	9	40
Dibromochloromethane	ug/L	0.260U	20	20	20	17.9	89	91	70-130	2	40
Dibromomethane	ug/L	0.500U	20	20	20	18.0	90	100	70-130	11	40
Dichlorodifluoromethane	ug/L	0.500U	20	20	20	22.8	114	122	70-130	7	40
Ethyl methacrylate	ug/L	0.500U	20	20	20	18.1	90	87	70-130	3	40
Ethylbenzene	ug/L	0.500U	20	20	20	20.5	102	106	70-130	4	40
Hexachloro-1,3-butadiene	ug/L	0.500U	20	20	20	18.2	91	104	70-130	14	40
Iodomethane	ug/L	0.500U	20	20	20	28.6	143	135	70-130	6	40 J(M1)
Isobutyl Alcohol	ug/L	10.0U	400	400	294	335	73	84	70-130	13	40
Methacrylonitrile	ug/L	5.00U	200	200	192	198	96	99	70-130	3	40
Methyl methacrylate	ug/L	5.00U	20	20	18.6	19.5	93	98	70-130	5	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 321 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134515 134516											
Parameter	Units	3520506003 Result	MS	MSD	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Spike Conc.						RPD	RPD
Methylene Chloride	ug/L	2.50U	20	20	18.5	19.9	91	98	70-130	7	40
Propionitrile	ug/L	5.00U	200	200	193	193	97	96	70-130	.3	40
Styrene	ug/L	0.500U	20	20	19.9	20.6	99	103	70-130	3	40
Tetrachloroethene	ug/L	0.500U	20	20	16.1	16.2	80	81	70-130	1	40
Toluene	ug/L	1.74	20	20	23.7	25.1	110	117	70-130	6	40
trans-1,2-Dichloroethene	ug/L	0.500U	20	20	20.6	21.7	103	109	70-130	5	40
trans-1,3-Dichloropropene	ug/L	0.250U	20	20	17.6	18.9	88	95	70-130	7	40
trans-1,4-Dichloro-2-butene	ug/L	5.00U	20	20	16.5	16.3	82	81	70-130	1	40
Trichloroethene	ug/L	0.500U	20	20	20.0	21.0	100	105	70-130	5	40
Trichlorofluoromethane	ug/L	0.500U	20	20	21.6	25.1	108	126	70-130	15	40
Vinyl acetate	ug/L	1.00U	20	20	15.2	15.8	76	79	70-130	4	40
Vinyl chloride	ug/L	0.500U	20	20	20.6	22.0	103	110	70-130	7	40
Xylene (Total)	ug/L	0.500U	60	60	60.0	62.3	100	104	70-130	4	40
1,2-Dichloroethane-d4 (S)	%						96	97	86-125		
4-Bromofluorobenzene (S)	%						99	96	70-114		
Dibromofluoromethane (S)	%						97	102	88-117		
Toluene-d8 (S)	%						103	104	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2284

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325021, 3519325022

METHOD BLANK: 134517

Matrix: Water

Associated Lab Samples: 3519325021, 3519325022

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/27/10 20:45	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:45	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/27/10 20:45	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/27/10 20:45	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:45	
1,2-Dichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:45	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:45	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:45	
1,4-Dichlorobenzene	ug/L	0.50U	1.0	10/27/10 20:45	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/27/10 20:45	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/27/10 20:45	
2-Hexanone	ug/L	5.0U	10.0	10/27/10 20:45	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/27/10 20:45	
Acetone	ug/L	5.0U	10.0	10/27/10 20:45	
Acetonitrile	ug/L	5.0U	10.0	10/27/10 20:45	
Acrolein	ug/L	10.0U	20.0	10/27/10 20:45	
Acrylonitrile	ug/L	5.0U	10.0	10/27/10 20:45	
Allyl chloride	ug/L	0.50U	1.0	10/27/10 20:45	
Benzene	ug/L	0.50U	1.0	10/27/10 20:45	
Bromochloromethane	ug/L	0.50U	1.0	10/27/10 20:45	
Bromodichloromethane	ug/L	0.27U	0.60	10/27/10 20:45	
Bromoform	ug/L	0.50U	1.0	10/27/10 20:45	
Bromomethane	ug/L	0.50U	1.0	10/27/10 20:45	
Carbon disulfide	ug/L	0.50U	1.0	10/27/10 20:45	
Carbon tetrachloride	ug/L	0.50U	1.0	10/27/10 20:45	
Chlorobenzene	ug/L	0.50U	1.0	10/27/10 20:45	
Chloroethane	ug/L	0.50U	1.0	10/27/10 20:45	
Chloroform	ug/L	0.50U	1.0	10/27/10 20:45	
Chloromethane	ug/L	0.62U	1.0	10/27/10 20:45	
Chloroprene	ug/L	0.50U	1.0	10/27/10 20:45	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:45	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/27/10 20:45	
Dibromochloromethane	ug/L	0.26U	0.50	10/27/10 20:45	
Dibromomethane	ug/L	0.50U	1.0	10/27/10 20:45	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/27/10 20:45	
Ethyl methacrylate	ug/L	0.50U	1.0	10/27/10 20:45	
Ethylbenzene	ug/L	0.50U	1.0	10/27/10 20:45	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/27/10 20:45	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 323 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 134517

Matrix: Water

Associated Lab Samples: 3519325021, 3519325022

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Iodomethane	ug/L	0.50U	1.0	10/27/10 20:45	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/27/10 20:45	
Methacrylonitrile	ug/L	5.0U	10.0	10/27/10 20:45	
Methyl methacrylate	ug/L	5.0U	10.0	10/27/10 20:45	
Methylene Chloride	ug/L	2.5U	5.0	10/27/10 20:45	
Propionitrile	ug/L	5.0U	10.0	10/27/10 20:45	
Styrene	ug/L	0.50U	1.0	10/27/10 20:45	
Tetrachloroethene	ug/L	0.50U	1.0	10/27/10 20:45	
Toluene	ug/L	0.50U	1.0	10/27/10 20:45	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/27/10 20:45	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/27/10 20:45	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/27/10 20:45	
Trichloroethene	ug/L	0.50U	1.0	10/27/10 20:45	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/27/10 20:45	
Vinyl acetate	ug/L	1.0U	2.0	10/27/10 20:45	
Vinyl chloride	ug/L	0.50U	1.0	10/27/10 20:45	
Xylene (Total)	ug/L	0.50U	1.0	10/27/10 20:45	
1,2-Dichloroethane-d4 (S)	%	102	86-125	10/27/10 20:45	
4-Bromofluorobenzene (S)	%	97	70-114	10/27/10 20:45	
Dibromofluoromethane (S)	%	100	88-117	10/27/10 20:45	
Toluene-d8 (S)	%	100	87-113	10/27/10 20:45	

LABORATORY CONTROL SAMPLE: 134518

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.2	101	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	19.5	98	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	21.2	106	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	19.2	96	84.1-122.6	
1,1-Dichloroethane	ug/L	20	18.8	94	66.4-138.6	
1,1-Dichloroethene	ug/L	20	18.2	91	79.3-127.5	
1,1-Dichloropropene	ug/L	20	20.8	104	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	20.2	101	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	20.4	102	79.1-134.1	
1,2-Dichlorobenzene	ug/L	20	19.9	99	91.7-127	
1,2-Dichloroethane	ug/L	20	19.4	97	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.0	100	82.2-129.1	
1,3-Dichloropropane	ug/L	20	19.8	99	88.1-118.2	
1,4-Dichlorobenzene	ug/L	20	20.3	102	91.9-121.7	
2,2-Dichloropropane	ug/L	20	20.5	102	44-181.7	
2-Butanone (MEK)	ug/L	20	22.5	113	53.8-156.3	
2-Hexanone	ug/L	20	21.9	110	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	19.0	95	71.8-134.4	
Acetone	ug/L	20	24.7	124	47.2-184.1	
Acetonitrile	ug/L	200	204	102	65.2-133.1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 324 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 134518

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Acrolein	ug/L	200	221	110	41.8-131.7	
Acrylonitrile	ug/L	200	198	99	57.8-125.9	
Allyl chloride	ug/L	20	21.3	107	23.6-190.7	
Benzene	ug/L	20	20.0	100	77.3-132.8	
Bromochloromethane	ug/L	20	19.0	95	87.4-122.8	
Bromodichloromethane	ug/L	20	20.3	101	77.2-121.1	
Bromoform	ug/L	20	18.6	93	65.9-133.5	
Bromomethane	ug/L	20	33.8	169	48.2-223.9	6p
Carbon disulfide	ug/L	20	23.1	116	20.3-195.4	
Carbon tetrachloride	ug/L	20	19.0	95	69-155.5	
Chlorobenzene	ug/L	20	19.2	96	76.9-123.9	
Chloroethane	ug/L	20	18.6	93	46.7-157.8	
Chloroform	ug/L	20	18.6	93	69.7-132	
Chloromethane	ug/L	20	20.3	102	54.4-153.8	
Chloroprene	ug/L	20	21.4	107	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	19.1	96	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	20.6	103	73-121.6	
Dibromochloromethane	ug/L	20	19.9	100	65.4-126.2	
Dibromomethane	ug/L	20	19.3	96	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	16.7	84	63.1-143.7	
Ethyl methacrylate	ug/L	20	19.5	97	34.3-179.4	
Ethylbenzene	ug/L	20	20.0	100	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	20.4	102	74.4-153.6	
Iodomethane	ug/L	20	14.6	73	1-243.3	
Isobutyl Alcohol	ug/L	400	406	102	62.9-136.1	
Methacrylonitrile	ug/L	200	204	102	77.3-132.6	
Methyl methacrylate	ug/L	20	19.3	96	37.4-178.3	
Methylene Chloride	ug/L	20	18.0	90	65.7-137.3	
Propionitrile	ug/L	200	201	101	71-130.3	
Styrene	ug/L	20	19.7	99	76.5-118.5	
Tetrachloroethene	ug/L	20	17.9	89	71-134	
Toluene	ug/L	20	19.8	99	75-129	
trans-1,2-Dichloroethene	ug/L	20	19.4	97	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	20.6	103	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	19.1	96	36.1-177.4	
Trichloroethene	ug/L	20	19.5	97	81.1-122.4	
Trichlorofluoromethane	ug/L	20	19.5	97	75.4-124.6	
Vinyl acetate	ug/L	20	20.8	104	72.2-139	
Vinyl chloride	ug/L	20	19.9	99	70.2-136.9	
Xylene (Total)	ug/L	60	59.8	100	82.3-126	
1,2-Dichloroethane-d4 (S)	%			100	86-125	
4-Bromofluorobenzene (S)	%			98	70-114	
Dibromofluoromethane (S)	%			100	88-117	
Toluene-d8 (S)	%			99	87-113	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134519 134520											
Parameter	Units	3520846001		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						RPD
				Conc.	Conc.	Result	Result	% Rec	% Rec	Limits	RPD
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	20	18.9	19.2	95	96	70-130	2 40
1,1,1-Trichloroethane	ug/L	0.50U	20	20	20	20.2	20.7	101	104	70-130	3 40
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	20	19.9	20.2	99	101	70-130	2 40
1,1,2-Trichloroethane	ug/L	0.50U	20	20	20	17.9	18.4	89	92	70-130	3 40
1,1-Dichloroethane	ug/L	0.50U	20	20	20	19.1	19.7	96	98	70-130	3 40
1,1-Dichloroethene	ug/L	0.50U	20	20	20	20.3	20.7	102	104	70-130	2 40
1,1-Dichloropropene	ug/L	0.50U	20	20	20	22.1	23.0	110	115	70-130	4 40
1,2,3-Trichloropropane	ug/L	0.36U	20	20	20	17.5	18.3	87	92	70-130	5 40
1,2,4-Trichlorobenzene	ug/L	0.50U	20	20	20	15.9	19.4	79	96	70-130	20 40
1,2-Dichlorobenzene	ug/L	0.50U	20	20	20	18.5	19.4	92	97	70-130	5 40
1,2-Dichloroethane	ug/L	0.50U	20	20	20	19.4	19.5	97	98	70-130	.5 40
1,2-Dichloropropane	ug/L	0.50U	20	20	20	20.2	20.3	101	101	70-130	.4 40
1,3-Dichloropropane	ug/L	0.50U	20	20	20	18.5	18.9	93	94	70-130	2 40
1,4-Dichlorobenzene	ug/L	0.50U	20	20	20	19.4	20.2	97	101	70-130	4 40
2,2-Dichloropropane	ug/L	0.50U	20	20	20	18.9	18.8	94	94	70-130	.4 40
2-Butanone (MEK)	ug/L	5.0U	20	20	20	16.2	16.3	81	81	70-130	.4 40
2-Hexanone	ug/L	5.0U	20	20	20	16.2	16.2	81	81	70-130	.1 40
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	20	16.4	16.6	82	83	70-130	1 40
Acetone	ug/L	5.0U	20	20	20	21.1	20.7	96	94	70-130	2 40
Acetonitrile	ug/L	5.0U	200	200	200	188	194	94	97	70-130	3 40
Acrolein	ug/L	10.0U	200	200	200	197	195	99	98	70-130	.8 40
Acrylonitrile	ug/L	5.0U	200	200	200	181	180	91	90	70-130	.8 40
Allyl chloride	ug/L	0.50U	20	20	20	22.0	22.0	110	110	70-130	.2 40
Benzene	ug/L	0.50U	20	20	20	20.6	21.2	102	105	70-130	3 40
Bromochloromethane	ug/L	0.50U	20	20	20	18.8	18.9	94	95	70-130	.7 40
Bromodichloromethane	ug/L	0.27U	20	20	20	19.5	19.7	98	99	70-130	.9 40
Bromoform	ug/L	0.50U	20	20	20	15.6	16.3	78	81	70-130	4 40
Bromomethane	ug/L	0.50U	20	20	20	22.8	27.1	114	136	70-130	17 40 J(M1)
Carbon disulfide	ug/L	0.50U	20	20	20	26.0	26.1	130	131	70-130	.6 40 J(M1)
Carbon tetrachloride	ug/L	0.50U	20	20	20	19.8	20.5	99	103	70-130	4 40
Chlorobenzene	ug/L	0.50U	20	20	20	18.8	19.1	94	96	70-130	2 40
Chloroethane	ug/L	0.50U	20	20	20	20.1	22.4	100	112	70-130	11 40
Chloroform	ug/L	0.50U	20	20	20	18.7	19.1	93	94	70-130	2 40
Chloromethane	ug/L	0.62U	20	20	20	23.8	25.8	119	129	70-130	8 40
Chloroprene	ug/L	0.50U	20	20	20	22.1	22.0	110	110	70-130	.4 40
cis-1,2-Dichloroethene	ug/L	0.50U	20	20	20	19.5	20.1	98	101	70-130	3 40
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	20	19.6	19.7	98	99	70-130	.7 40
Dibromochloromethane	ug/L	0.26U	20	20	20	17.7	18.4	89	92	70-130	4 40
Dibromomethane	ug/L	0.50U	20	20	20	18.8	18.9	94	94	70-130	.2 40
Dichlorodifluoromethane	ug/L	0.50U	20	20	20	24.5	25.7	123	129	70-130	5 40
Ethyl methacrylate	ug/L	0.50U	20	20	20	17.7	17.7	89	89	70-130	.002 40
Ethylbenzene	ug/L	0.50U	20	20	20	19.9	20.4	99	102	70-130	2 40
Hexachloro-1,3-butadiene	ug/L	0.50U	20	20	20	17.8	19.8	88	98	70-130	11 40
Iodomethane	ug/L	0.50U	20	20	20	19.1	30.1	95	150	70-130	45 40 J(D6), J(M1)
Isobutyl Alcohol	ug/L	10.0U	400	400	400	309	325	77	81	70-130	5 40
Methacrylonitrile	ug/L	5.0U	200	200	200	189	189	95	95	70-130	.002 40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 326 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134519 134520											
Parameter	Units	3520846001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Methyl methacrylate	ug/L	5.0U	20	20	18.2	17.9	91	89	70-130	2	40
Methylene Chloride	ug/L	2.5U	20	20	17.9	18.3	88	90	70-130	2	40
Propionitrile	ug/L	5.0U	200	200	184	185	92	93	70-130	.7	40
Styrene	ug/L	0.50U	20	20	19.3	20.0	96	100	70-130	3	40
Tetrachloroethene	ug/L	0.50U	20	20	15.8	16.2	79	81	70-130	3	40
Toluene	ug/L	1.1	20	20	20.3	20.8	96	99	70-130	2	40
trans-1,2-Dichloroethene	ug/L	0.50U	20	20	20.4	20.7	102	103	70-130	1	40
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	18.5	19.1	92	95	70-130	3	40
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	15.6	16.1	78	81	70-130	3	40
Trichloroethene	ug/L	0.50U	20	20	19.9	20.3	100	102	70-130	2	40
Trichlorofluoromethane	ug/L	0.50U	20	20	23.4	24.1	117	120	70-130	3	40
Vinyl acetate	ug/L	1.0U	20	20	16.4	15.8	82	79	70-130	4	40
Vinyl chloride	ug/L	0.50U	20	20	22.0	22.7	110	114	70-130	3	40
Xylene (Total)	ug/L	0.50U	60	60	58.3	60.5	97	101	70-130	4	40
1,2-Dichloroethane-d4 (S)	%						100	100	86-125		
4-Bromofluorobenzene (S)	%						95	96	70-114		
Dibromofluoromethane (S)	%						99	100	88-117		
Toluene-d8 (S)	%						100	100	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2287

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325013, 3519325014, 3519325015

METHOD BLANK: 134719

Matrix: Water

Associated Lab Samples: 3519325013, 3519325014, 3519325015

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/28/10 11:55	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/28/10 11:55	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/28/10 11:55	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/28/10 11:55	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/28/10 11:55	
1,2-Dichlorobenzene	ug/L	0.50U	1.0	10/28/10 11:55	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/28/10 11:55	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/28/10 11:55	
1,4-Dichlorobenzene	ug/L	0.50U	1.0	10/28/10 11:55	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/28/10 11:55	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/28/10 11:55	
2-Hexanone	ug/L	5.0U	10.0	10/28/10 11:55	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/28/10 11:55	
Acetone	ug/L	5.0U	10.0	10/28/10 11:55	
Acetonitrile	ug/L	5.0U	10.0	10/28/10 11:55	
Acrolein	ug/L	10.0U	20.0	10/28/10 11:55	
Acrylonitrile	ug/L	5.0U	10.0	10/28/10 11:55	
Allyl chloride	ug/L	0.50U	1.0	10/28/10 11:55	
Benzene	ug/L	0.50U	1.0	10/28/10 11:55	
Bromochloromethane	ug/L	0.50U	1.0	10/28/10 11:55	
Bromodichloromethane	ug/L	0.27U	0.60	10/28/10 11:55	
Bromoform	ug/L	0.50U	1.0	10/28/10 11:55	
Bromomethane	ug/L	0.50U	1.0	10/28/10 11:55	
Carbon disulfide	ug/L	0.50U	1.0	10/28/10 11:55	
Carbon tetrachloride	ug/L	0.50U	1.0	10/28/10 11:55	
Chlorobenzene	ug/L	0.50U	1.0	10/28/10 11:55	
Chloroethane	ug/L	0.50U	1.0	10/28/10 11:55	
Chloroform	ug/L	0.50U	1.0	10/28/10 11:55	
Chloromethane	ug/L	0.62U	1.0	10/28/10 11:55	
Chloroprene	ug/L	0.50U	1.0	10/28/10 11:55	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/28/10 11:55	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/28/10 11:55	
Dibromochloromethane	ug/L	0.26U	0.50	10/28/10 11:55	
Dibromomethane	ug/L	0.50U	1.0	10/28/10 11:55	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/28/10 11:55	
Ethyl methacrylate	ug/L	0.50U	1.0	10/28/10 11:55	
Ethylbenzene	ug/L	0.50U	1.0	10/28/10 11:55	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/28/10 11:55	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 328 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 134719

Matrix: Water

Associated Lab Samples: 3519325013, 3519325014, 3519325015

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Iodomethane	ug/L	0.50U	1.0	10/28/10 11:55	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/28/10 11:55	
Methacrylonitrile	ug/L	5.0U	10.0	10/28/10 11:55	
Methyl methacrylate	ug/L	5.0U	10.0	10/28/10 11:55	
Methylene Chloride	ug/L	2.5U	5.0	10/28/10 11:55	
Propionitrile	ug/L	5.0U	10.0	10/28/10 11:55	
Styrene	ug/L	0.50U	1.0	10/28/10 11:55	
Tetrachloroethene	ug/L	0.50U	1.0	10/28/10 11:55	
Toluene	ug/L	0.50U	1.0	10/28/10 11:55	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/28/10 11:55	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/28/10 11:55	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/28/10 11:55	
Trichloroethene	ug/L	0.50U	1.0	10/28/10 11:55	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/28/10 11:55	
Vinyl acetate	ug/L	1.0U	2.0	10/28/10 11:55	
Vinyl chloride	ug/L	0.50U	1.0	10/28/10 11:55	
Xylene (Total)	ug/L	0.50U	1.0	10/28/10 11:55	
1,2-Dichloroethane-d4 (S)	%	110	86-125	10/28/10 11:55	
4-Bromofluorobenzene (S)	%	96	70-114	10/28/10 11:55	
Dibromofluoromethane (S)	%	104	88-117	10/28/10 11:55	
Toluene-d8 (S)	%	99	87-113	10/28/10 11:55	

LABORATORY CONTROL SAMPLE: 134720

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	21.1	105	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	19.5	98	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	20.1	100	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	20.7	104	84.1-122.6	
1,1-Dichloroethane	ug/L	20	20.5	102	66.4-138.6	
1,1-Dichloroethene	ug/L	20	21.1	106	79.3-127.5	
1,1-Dichloropropene	ug/L	20	20.9	105	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	19.0	95	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	21.8	109	79.1-134.1	
1,2-Dichlorobenzene	ug/L	20	20.7	104	91.7-127	
1,2-Dichloroethane	ug/L	20	20.0	100	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.7	104	82.2-129.1	
1,3-Dichloropropane	ug/L	20	20.5	103	88.1-118.2	
1,4-Dichlorobenzene	ug/L	20	20.3	102	91.9-121.7	
2,2-Dichloropropane	ug/L	20	23.6	118	44-181.7	
2-Butanone (MEK)	ug/L	20	19.4	97	53.8-156.3	
2-Hexanone	ug/L	20	19.0	95	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	20.7	103	71.8-134.4	
Acetone	ug/L	20	19.7	98	47.2-184.1	
Acetonitrile	ug/L	200	209	104	65.2-133.1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 329 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 134720

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Acrolein	ug/L	200	226	113	41.8-131.7	
Acrylonitrile	ug/L	200	210	105	57.8-125.9	
Allyl chloride	ug/L	20	22.9	115	23.6-190.7	
Benzene	ug/L	20	20.8	104	77.3-132.8	
Bromochloromethane	ug/L	20	19.9	100	87.4-122.8	
Bromodichloromethane	ug/L	20	21.5	108	77.2-121.1	
Bromoform	ug/L	20	17.4	87	65.9-133.5	
Bromomethane	ug/L	20	22.1	110	48.2-223.9	
Carbon disulfide	ug/L	20	24.1	120	20.3-195.4	
Carbon tetrachloride	ug/L	20	20.2	101	69-155.5	
Chlorobenzene	ug/L	20	20.7	104	76.9-123.9	
Chloroethane	ug/L	20	21.5	107	46.7-157.8	
Chloroform	ug/L	20	21.3	106	69.7-132	
Chloromethane	ug/L	20	21.3	106	54.4-153.8	
Chloroprene	ug/L	20	22.2	111	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	21.0	105	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	22.3	112	73-121.6	
Dibromochloromethane	ug/L	20	18.9	94	65.4-126.2	
Dibromomethane	ug/L	20	20.5	103	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	19.1	95	63.1-143.7	
Ethyl methacrylate	ug/L	20	20.7	104	34.3-179.4	
Ethylbenzene	ug/L	20	21.0	105	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	21.8	109	74.4-153.6	
Iodomethane	ug/L	20	25.6	128	1-243.3	
Isobutyl Alcohol	ug/L	400	406	101	62.9-136.1	
Methacrylonitrile	ug/L	200	205	103	77.3-132.6	
Methyl methacrylate	ug/L	20	20.8	104	37.4-178.3	
Methylene Chloride	ug/L	20	20.5	103	65.7-137.3	
Propionitrile	ug/L	200	210	105	71-130.3	
Styrene	ug/L	20	21.9	109	76.5-118.5	
Tetrachloroethene	ug/L	20	17.8	89	71-134	
Toluene	ug/L	20	20.9	105	75-129	
trans-1,2-Dichloroethene	ug/L	20	21.3	107	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	20.7	103	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	20.4	102	36.1-177.4	
Trichloroethene	ug/L	20	20.3	102	81.1-122.4	
Trichlorofluoromethane	ug/L	20	20.0	100	75.4-124.6	
Vinyl acetate	ug/L	20	22.3	112	72.2-139	
Vinyl chloride	ug/L	20	20.4	102	70.2-136.9	
Xylene (Total)	ug/L	60	62.2	104	82.3-126	
1,2-Dichloroethane-d4 (S)	%			95	86-125	
4-Bromofluorobenzene (S)	%			101	70-114	
Dibromofluoromethane (S)	%			99	88-117	
Toluene-d8 (S)	%			102	87-113	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134721 134722											
Parameter	Units	3520572001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	19.1	19.5	95	98	70-130	2	40
1,1,1-Trichloroethane	ug/L	0.50U	20	20	20.4	20.1	102	101	70-130	1	40
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	18.8	19.0	94	95	70-130	1	40
1,1,2-Trichloroethane	ug/L	0.50U	20	20	18.7	19.3	93	96	70-130	3	40
1,1-Dichloroethane	ug/L	0.50U	20	20	20.8	20.8	104	104	70-130	.3	40
1,1-Dichloroethene	ug/L	0.50U	20	20	22.9	23.4	115	117	70-130	2	40
1,1-Dichloropropene	ug/L	0.50U	20	20	21.5	20.9	107	104	70-130	3	40
1,2,3-Trichloropropane	ug/L	0.36U	20	20	17.1	18.3	86	91	70-130	6	40
1,2,4-Trichlorobenzene	ug/L	0.50U	20	20	15.6	19.4	78	97	70-130	22	40
1,2-Dichlorobenzene	ug/L	0.50U	20	20	19.0	19.7	95	98	70-130	4	40
1,2-Dichloroethane	ug/L	0.50U	20	20	19.5	19.8	98	99	70-130	1	40
1,2-Dichloropropane	ug/L	0.50U	20	20	20.7	20.5	104	103	70-130	.9	40
1,3-Dichloropropane	ug/L	0.50U	20	20	19.2	19.4	96	97	70-130	.8	40
1,4-Dichlorobenzene	ug/L	0.50U	20	20	19.8	20.1	99	100	70-130	2	40
2,2-Dichloropropane	ug/L	0.50U	20	20	14.6	14.4	73	72	70-130	2	40
2-Butanone (MEK)	ug/L	5.0U	20	20	15.5	16.2	77	81	70-130	4	40
2-Hexanone	ug/L	5.0U	20	20	16.2	16.4	81	82	70-130	1	40
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	17.9	18.4	90	92	70-130	2	40
Acetone	ug/L	5.0U	20	20	17.2	16.7	76	73	70-130	3	40
Acetonitrile	ug/L	5.0U	200	200	195	194	97	97	70-130	.4	40
Acrolein	ug/L	10.0U	200	200	181	186	90	93	70-130	3	40
Acrylonitrile	ug/L	5.0U	200	200	197	196	98	98	70-130	.5	40
Allyl chloride	ug/L	0.50U	20	20	21.9	21.4	110	107	70-130	2	40
Benzene	ug/L	0.50U	20	20	21.2	21.0	106	105	70-130	1	40
Bromochloromethane	ug/L	0.50U	20	20	19.8	20.0	99	100	70-130	.8	40
Bromodichloromethane	ug/L	0.27U	20	20	19.3	19.1	97	96	70-130	1	40
Bromoform	ug/L	0.50U	20	20	15.9	16.3	80	82	70-130	2	40
Bromomethane	ug/L	0.50U	20	20	20.2	21.6	101	108	70-130	6	40
Carbon disulfide	ug/L	0.50U	20	20	26.3	26.2	131	130	70-130	.5	40 J(M1)
Carbon tetrachloride	ug/L	0.50U	20	20	20.4	20.3	102	102	70-130	.1	40
Chlorobenzene	ug/L	0.50U	20	20	20.0	19.9	100	100	70-130	.6	40
Chloroethane	ug/L	0.50U	20	20	22.5	23.6	112	118	70-130	5	40
Chloroform	ug/L	0.50U	20	20	19.3	19.3	96	96	70-130	.2	40
Chloromethane	ug/L	0.62U	20	20	26.3	26.3	131	131	70-130	.2	40 J(M1)
Chloroprene	ug/L	0.50U	20	20	22.2	22.1	111	110	70-130	.6	40
cis-1,2-Dichloroethene	ug/L	0.50U	20	20	20.6	20.6	103	103	70-130	.06	40
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	18.8	18.6	94	93	70-130	1	40
Dibromochloromethane	ug/L	0.26U	20	20	17.4	18.0	87	90	70-130	3	40
Dibromomethane	ug/L	0.50U	20	20	19.8	19.4	99	97	70-130	2	40
Dichlorodifluoromethane	ug/L	0.50U	20	20	25.8	25.8	129	129	70-130	.2	40
Ethyl methacrylate	ug/L	0.50U	20	20	19.4	19.5	97	98	70-130	.5	40
Ethylbenzene	ug/L	0.50U	20	20	20.2	20.3	101	101	70-130	.3	40
Hexachloro-1,3-butadiene	ug/L	0.50U	20	20	16.8	18.9	84	94	70-130	11	40
Iodomethane	ug/L	0.50U	20	20	16.5	19.7	82	98	70-130	18	40
Isobutyl Alcohol	ug/L	10.0U	400	400	319	359	80	90	70-130	12	40
Methacrylonitrile	ug/L	5.0U	200	200	204	202	102	101	70-130	.9	40
Methyl methacrylate	ug/L	5.0U	20	20	18.4	18.3	92	91	70-130	.9	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 331 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134721 134722											
Parameter	Units	3520572001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Methylene Chloride	ug/L	2.5U	20	20	20.5	20.5	102	102	70-130	.2	40
Propionitrile	ug/L	5.0U	200	200	195	192	97	96	70-130	1	40
Styrene	ug/L	0.50U	20	20	19.6	19.9	98	99	70-130	1	40
Tetrachloroethene	ug/L	0.50U	20	20	16.0	16.3	80	81	70-130	2	40
Toluene	ug/L	0.70 I	20	20	21.0	21.3	102	103	70-130	1	40
trans-1,2-Dichloroethene	ug/L	0.50U	20	20	21.7	21.6	109	108	70-130	.7	40
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	17.5	17.7	88	89	70-130	1	40
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	15.7	16.1	79	80	70-130	2	40
Trichloroethene	ug/L	0.50U	20	20	20.5	20.4	102	102	70-130	.4	40
Trichlorofluoromethane	ug/L	0.50U	20	20	25.6	25.2	128	126	70-130	1	40
Vinyl acetate	ug/L	1.0U	20	20	12.9	12.7	65	64	70-130	2	40 J(M1)
Vinyl chloride	ug/L	0.50U	20	20	25.8	25.5	129	128	70-130	.9	40
Xylene (Total)	ug/L	0.50U	60	60	60.1	60.6	100	101	70-130	.9	40
1,2-Dichloroethane-d4 (S)	%						98	101	86-125		
4-Bromofluorobenzene (S)	%						99	98	70-114		
Dibromofluoromethane (S)	%						100	99	88-117		
Toluene-d8 (S)	%						100	101	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2291

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029, 3519325030, 3519325031

METHOD BLANK: 135012

Matrix: Water

Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029, 3519325030, 3519325031

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/29/10 01:54	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/29/10 01:54	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/29/10 01:54	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/29/10 01:54	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/29/10 01:54	
1,2-Dichlorobenzene	ug/L	0.50U	1.0	10/29/10 01:54	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/29/10 01:54	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/29/10 01:54	
1,4-Dichlorobenzene	ug/L	0.50U	1.0	10/29/10 01:54	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/29/10 01:54	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/29/10 01:54	
2-Hexanone	ug/L	5.0U	10.0	10/29/10 01:54	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/29/10 01:54	
Acetone	ug/L	5.0U	10.0	10/29/10 01:54	
Acetonitrile	ug/L	5.0U	10.0	10/29/10 01:54	
Acrolein	ug/L	10.0U	20.0	10/29/10 01:54	
Acrylonitrile	ug/L	5.0U	10.0	10/29/10 01:54	
Allyl chloride	ug/L	0.50U	1.0	10/29/10 01:54	
Benzene	ug/L	0.50U	1.0	10/29/10 01:54	
Bromochloromethane	ug/L	0.50U	1.0	10/29/10 01:54	
Bromodichloromethane	ug/L	0.27U	0.60	10/29/10 01:54	
Bromoform	ug/L	0.50U	1.0	10/29/10 01:54	
Bromomethane	ug/L	0.50U	1.0	10/29/10 01:54	
Carbon disulfide	ug/L	0.50U	1.0	10/29/10 01:54	
Carbon tetrachloride	ug/L	0.50U	1.0	10/29/10 01:54	
Chlorobenzene	ug/L	0.50U	1.0	10/29/10 01:54	
Chloroethane	ug/L	0.50U	1.0	10/29/10 01:54	
Chloroform	ug/L	0.50U	1.0	10/29/10 01:54	
Chloromethane	ug/L	0.62U	1.0	10/29/10 01:54	
Chloroprene	ug/L	0.50U	1.0	10/29/10 01:54	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/29/10 01:54	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/29/10 01:54	
Dibromochloromethane	ug/L	0.26U	0.50	10/29/10 01:54	
Dibromomethane	ug/L	0.50U	1.0	10/29/10 01:54	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/29/10 01:54	
Ethyl methacrylate	ug/L	0.50U	1.0	10/29/10 01:54	
Ethylbenzene	ug/L	0.50U	1.0	10/29/10 01:54	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/29/10 01:54	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 333 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 135012

Matrix: Water

Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029, 3519325030, 3519325031

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Iodomethane	ug/L	3.2	1.0	10/29/10 01:54	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/29/10 01:54	
Methacrylonitrile	ug/L	5.0U	10.0	10/29/10 01:54	
Methyl methacrylate	ug/L	5.0U	10.0	10/29/10 01:54	
Methylene Chloride	ug/L	2.5U	5.0	10/29/10 01:54	
Propionitrile	ug/L	5.0U	10.0	10/29/10 01:54	
Styrene	ug/L	0.50U	1.0	10/29/10 01:54	
Tetrachloroethene	ug/L	0.50U	1.0	10/29/10 01:54	
Toluene	ug/L	0.50U	1.0	10/29/10 01:54	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/29/10 01:54	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/29/10 01:54	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/29/10 01:54	
Trichloroethene	ug/L	0.50U	1.0	10/29/10 01:54	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/29/10 01:54	
Vinyl acetate	ug/L	1.0U	2.0	10/29/10 01:54	
Vinyl chloride	ug/L	0.50U	1.0	10/29/10 01:54	
Xylene (Total)	ug/L	0.50U	1.0	10/29/10 01:54	
1,2-Dichloroethane-d4 (S)	%	102	86-125	10/29/10 01:54	
4-Bromofluorobenzene (S)	%	99	70-114	10/29/10 01:54	
Dibromofluoromethane (S)	%	102	88-117	10/29/10 01:54	
Toluene-d8 (S)	%	101	87-113	10/29/10 01:54	

LABORATORY CONTROL SAMPLE: 135013

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.4	102	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	20.1	101	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	20.8	104	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	20.5	102	84.1-122.6	
1,1-Dichloroethane	ug/L	20	20.6	103	66.4-138.6	
1,1-Dichloroethene	ug/L	20	20.4	102	79.3-127.5	
1,1-Dichloropropene	ug/L	20	20.5	103	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	20.6	103	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	21.9	109	79.1-134.1	
1,2-Dichlorobenzene	ug/L	20	20.8	104	91.7-127	
1,2-Dichloroethane	ug/L	20	19.8	99	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.6	103	82.2-129.1	
1,3-Dichloropropane	ug/L	20	20.9	105	88.1-118.2	
1,4-Dichlorobenzene	ug/L	20	21.0	105	91.9-121.7	
2,2-Dichloropropane	ug/L	20	21.2	106	44-181.7	
2-Butanone (MEK)	ug/L	20	20.4	102	53.8-156.3	
2-Hexanone	ug/L	20	19.0	95	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	19.5	97	71.8-134.4	
Acetone	ug/L	20	19.1	96	47.2-184.1	
Acetonitrile	ug/L	200	203	102	65.2-133.1	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 334 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 135013

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Acrolein	ug/L	200	219	109	41.8-131.7	
Acrylonitrile	ug/L	200	204	102	57.8-125.9	
Allyl chloride	ug/L	20	21.4	107	23.6-190.7	
Benzene	ug/L	20	20.6	103	77.3-132.8	
Bromochloromethane	ug/L	20	20.2	101	87.4-122.8	
Bromodichloromethane	ug/L	20	19.9	100	77.2-121.1	
Bromoform	ug/L	20	18.7	93	65.9-133.5	
Bromomethane	ug/L	20	25.1	125	48.2-223.9	
Carbon disulfide	ug/L	20	24.3	122	20.3-195.4	
Carbon tetrachloride	ug/L	20	19.9	100	69-155.5	
Chlorobenzene	ug/L	20	20.9	104	76.9-123.9	
Chloroethane	ug/L	20	18.9	95	46.7-157.8	
Chloroform	ug/L	20	19.5	98	69.7-132	
Chloromethane	ug/L	20	22.2	111	54.4-153.8	
Chloroprene	ug/L	20	20.9	104	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	20.7	104	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	20.5	103	73-121.6	
Dibromochloromethane	ug/L	20	19.8	99	65.4-126.2	
Dibromomethane	ug/L	20	20.3	101	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	19.8	99	63.1-143.7	
Ethyl methacrylate	ug/L	20	20.7	104	34.3-179.4	
Ethylbenzene	ug/L	20	21.1	105	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	20.8	104	74.4-153.6	
Iodomethane	ug/L	20	19.0	95	1-243.3	
Isobutyl Alcohol	ug/L	400	359	90	62.9-136.1	
Methacrylonitrile	ug/L	200	208	104	77.3-132.6	
Methyl methacrylate	ug/L	20	20.2	101	37.4-178.3	
Methylene Chloride	ug/L	20	20.6	103	65.7-137.3	
Propionitrile	ug/L	200	199	100	71-130.3	
Styrene	ug/L	20	21.2	106	76.5-118.5	
Tetrachloroethene	ug/L	20	17.8	89	71-134	
Toluene	ug/L	20	20.9	105	75-129	
trans-1,2-Dichloroethene	ug/L	20	20.9	105	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	20.6	103	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	20.1	101	36.1-177.4	
Trichloroethene	ug/L	20	20.4	102	81.1-122.4	
Trichlorofluoromethane	ug/L	20	20.3	102	75.4-124.6	
Vinyl acetate	ug/L	20	21.0	105	72.2-139	
Vinyl chloride	ug/L	20	21.1	106	70.2-136.9	
Xylene (Total)	ug/L	60	62.8	105	82.3-126	
1,2-Dichloroethane-d4 (S)	%			99	86-125	
4-Bromofluorobenzene (S)	%			100	70-114	
Dibromofluoromethane (S)	%			98	88-117	
Toluene-d8 (S)	%			100	87-113	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135014 135015											
Parameter	Units	3519452053 Result	MS		MSD		MS % Rec	MSD % Rec	% Rec Limits	Max	
			Spike Conc.	Conc.	Spike Conc.	Result				RPD	RPD
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	20	19.4	97	97	70-130	.5	40
1,1,1-Trichloroethane	ug/L	0.50U	20	20	20	19.8	99	99	70-130	.4	40
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	20	18.6	93	94	70-130	.9	40
1,1,2-Trichloroethane	ug/L	0.50U	20	20	20	19.1	95	93	70-130	3	40
1,1-Dichloroethane	ug/L	0.50U	20	20	20	20.4	102	104	70-130	1	40
1,1-Dichloroethene	ug/L	0.50U	20	20	20	22.9	114	115	70-130	.2	40
1,1-Dichloropropene	ug/L	0.50U	20	20	20	20.9	104	108	70-130	3	40
1,2,3-Trichloropropane	ug/L	0.36U	20	20	20	18.5	93	89	70-130	4	40
1,2,4-Trichlorobenzene	ug/L	0.50U	20	20	20	19.9	100	98	70-130	2	40
1,2-Dichlorobenzene	ug/L	0.50U	20	20	20	19.5	97	98	70-130	.3	40
1,2-Dichloroethane	ug/L	0.50U	20	20	20	18.7	94	96	70-130	3	40
1,2-Dichloropropane	ug/L	0.50U	20	20	20	20.3	102	104	70-130	2	40
1,3-Dichloropropane	ug/L	0.50U	20	20	20	19.0	95	94	70-130	.7	40
1,4-Dichlorobenzene	ug/L	0.50U	20	20	20	20.0	100	100	70-130	.5	40
2,2-Dichloropropane	ug/L	0.50U	20	20	20	18.3	92	92	70-130	.4	40
2-Butanone (MEK)	ug/L	5.0U	20	20	20	16.4	82	83	70-130	1	40
2-Hexanone	ug/L	5.0U	20	20	20	16.9	85	81	70-130	4	40
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	20	18.8	94	89	70-130	5	40
Acetone	ug/L	5.0U	20	20	20	16.4	72	65	70-130	9	40 J(M1)
Acetonitrile	ug/L	5.0U	200	200	200	196	97	97	70-130	.2	40
Acrolein	ug/L	10.0U	200	200	200	210	105	106	70-130	.5	40
Acrylonitrile	ug/L	5.0U	200	200	200	184	92	96	70-130	4	40
Allyl chloride	ug/L	0.50U	20	20	20	22.9	115	113	70-130	2	40
Benzene	ug/L	0.50U	20	20	20	20.9	104	105	70-130	.7	40
Bromochloromethane	ug/L	0.50U	20	20	20	19.7	99	100	70-130	2	40
Bromodichloromethane	ug/L	0.27U	20	20	20	18.8	94	97	70-130	3	40
Bromoform	ug/L	0.50U	20	20	20	16.3	81	82	70-130	.3	40
Bromomethane	ug/L	0.50U	20	20	20	29.4	147	165	70-130	11	40 J(M1)
Carbon disulfide	ug/L	0.50U	20	20	20	27.6	138	136	70-130	1	40 J(M1)
Carbon tetrachloride	ug/L	0.50U	20	20	20	20.0	100	103	70-130	3	40
Chlorobenzene	ug/L	0.50U	20	20	20	20.1	101	100	70-130	.3	40
Chloroethane	ug/L	0.50U	20	20	20	22.7	113	112	70-130	1	40
Chloroform	ug/L	0.50U	20	20	20	18.8	94	96	70-130	2	40
Chloromethane	ug/L	0.62U	20	20	20	25.0	125	122	70-130	3	40
Chloroprene	ug/L	0.50U	20	20	20	23.4	117	114	70-130	3	40
cis-1,2-Dichloroethene	ug/L	0.50U	20	20	20	21.0	105	104	70-130	.7	40
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	20	18.8	94	95	70-130	1	40
Dibromochloromethane	ug/L	0.26U	20	20	20	17.9	90	89	70-130	.3	40
Dibromomethane	ug/L	0.50U	20	20	20	19.1	95	94	70-130	2	40
Dichlorodifluoromethane	ug/L	0.50U	20	20	20	24.4	122	129	70-130	6	40
Ethyl methacrylate	ug/L	0.50U	20	20	20	20.2	101	97	70-130	4	40
Ethylbenzene	ug/L	0.50U	20	20	20	20.5	103	102	70-130	.6	40
Hexachloro-1,3-butadiene	ug/L	0.50U	20	20	20	18.9	94	93	70-130	1	40
Iodomethane	ug/L	0.50U	20	20	20	21.2	106	109	70-130	3	40
Isobutyl Alcohol	ug/L	10.0U	400	400	400	359	90	87	70-130	4	40
Methacrylonitrile	ug/L	5.0U	200	200	200	206	103	100	70-130	3	40
Methyl methacrylate	ug/L	5.0U	20	20	20	18.6	93	91	70-130	2	40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 336 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Peace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135014 135015											
Parameter	Units	3519452053 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Methylene Chloride	ug/L	2.5U	20	20	20.1	20.5	99	101	70-130	2	40
Propionitrile	ug/L	5.0U	200	200	198	185	99	93	70-130	7	40
Styrene	ug/L	0.50U	20	20	20.1	20.2	101	101	70-130	.4	40
Tetrachloroethene	ug/L	0.50U	20	20	16.4	16.5	82	82	70-130	.3	40
Toluene	ug/L	0.50U	20	20	20.6	20.5	103	102	70-130	.3	40
trans-1,2-Dichloroethene	ug/L	0.50U	20	20	21.1	21.2	106	106	70-130	.5	40
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	18.2	18.6	91	93	70-130	2	40
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	18.1	17.1	91	86	70-130	6	40
Trichloroethene	ug/L	0.50U	20	20	20.4	20.2	102	101	70-130	1	40
Trichlorofluoromethane	ug/L	0.50U	20	20	23.9	24.6	119	123	70-130	3	40
Vinyl acetate	ug/L	1.0U	20	20	18.6	18.3	93	91	70-130	2	40
Vinyl chloride	ug/L	0.50U	20	20	24.1	24.9	121	125	70-130	3	40
Xylene (Total)	ug/L	0.50U	60	60	61.1	60.5	102	101	70-130	1	40
1,2-Dichloroethane-d4 (S)	%						99	101	86-125		
4-Bromofluorobenzene (S)	%						100	100	70-114		
Dibromofluoromethane (S)	%						98	100	88-117		
Toluene-d8 (S)	%						99	101	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: MSV/2294

Analysis Method: EPA 8260

QC Batch Method: EPA 8260

Analysis Description: 8260 MSV

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325038

METHOD BLANK: 135449

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325038

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
1,1,1-Trichloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
1,1,2,2-Tetrachloroethane	ug/L	0.18U	0.50	10/29/10 14:54	
1,1,2-Trichloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
1,1-Dichloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
1,1-Dichloroethene	ug/L	0.50U	1.0	10/29/10 14:54	
1,1-Dichloropropene	ug/L	0.50U	1.0	10/29/10 14:54	
1,2,3-Trichloropropane	ug/L	0.36U	0.50	10/29/10 14:54	
1,2,4-Trichlorobenzene	ug/L	0.50U	1.0	10/29/10 14:54	
1,2-Dichloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
1,2-Dichloropropane	ug/L	0.50U	1.0	10/29/10 14:54	
1,3-Dichloropropane	ug/L	0.50U	1.0	10/29/10 14:54	
2,2-Dichloropropane	ug/L	0.50U	1.0	10/29/10 14:54	
2-Butanone (MEK)	ug/L	5.0U	10.0	10/29/10 14:54	
2-Hexanone	ug/L	5.0U	10.0	10/29/10 14:54	
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	10.0	10/29/10 14:54	
Acetone	ug/L	5.0U	10.0	10/29/10 14:54	
Acetonitrile	ug/L	5.0U	10.0	10/29/10 14:54	
Acrolein	ug/L	10.0U	20.0	10/29/10 14:54	
Acrylonitrile	ug/L	5.0U	10.0	10/29/10 14:54	
Allyl chloride	ug/L	0.50U	1.0	10/29/10 14:54	
Benzene	ug/L	0.50U	1.0	10/29/10 14:54	
Bromochloromethane	ug/L	0.50U	1.0	10/29/10 14:54	
Bromodichloromethane	ug/L	0.27U	0.60	10/29/10 14:54	
Bromoform	ug/L	0.50U	1.0	10/29/10 14:54	
Bromomethane	ug/L	0.50U	1.0	10/29/10 14:54	
Carbon disulfide	ug/L	0.50U	1.0	10/29/10 14:54	
Carbon tetrachloride	ug/L	0.50U	1.0	10/29/10 14:54	
Chlorobenzene	ug/L	0.50U	1.0	10/29/10 14:54	
Chloroethane	ug/L	0.50U	1.0	10/29/10 14:54	
Chloroform	ug/L	0.50U	1.0	10/29/10 14:54	
Chloromethane	ug/L	0.62U	1.0	10/29/10 14:54	
Chloroprene	ug/L	0.50U	1.0	10/29/10 14:54	
cis-1,2-Dichloroethene	ug/L	0.50U	1.0	10/29/10 14:54	
cis-1,3-Dichloropropene	ug/L	0.25U	0.50	10/29/10 14:54	
Dibromochloromethane	ug/L	0.26U	0.50	10/29/10 14:54	
Dibromomethane	ug/L	0.50U	1.0	10/29/10 14:54	
Dichlorodifluoromethane	ug/L	0.50U	1.0	10/29/10 14:54	
Ethyl methacrylate	ug/L	0.50U	1.0	10/29/10 14:54	
Ethylbenzene	ug/L	0.50U	1.0	10/29/10 14:54	
Hexachloro-1,3-butadiene	ug/L	0.50U	1.0	10/29/10 14:54	
Iodomethane	ug/L	3.2	1.0	10/29/10 14:54	
Isobutyl Alcohol	ug/L	10.0U	20.0	10/29/10 14:54	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 338 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

METHOD BLANK: 135449

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325038

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Methacrylonitrile	ug/L	5.0U	10.0	10/29/10 14:54	
Methyl methacrylate	ug/L	5.0U	10.0	10/29/10 14:54	
Methylene Chloride	ug/L	2.5U	5.0	10/29/10 14:54	
Propionitrile	ug/L	5.0U	10.0	10/29/10 14:54	
Styrene	ug/L	0.50U	1.0	10/29/10 14:54	
Tetrachloroethene	ug/L	0.50U	1.0	10/29/10 14:54	
Toluene	ug/L	0.50U	1.0	10/29/10 14:54	
trans-1,2-Dichloroethene	ug/L	0.50U	1.0	10/29/10 14:54	
trans-1,3-Dichloropropene	ug/L	0.25U	0.50	10/29/10 14:54	
trans-1,4-Dichloro-2-butene	ug/L	5.0U	10.0	10/29/10 14:54	
Trichloroethene	ug/L	0.50U	1.0	10/29/10 14:54	
Trichlorofluoromethane	ug/L	0.50U	1.0	10/29/10 14:54	
Vinyl acetate	ug/L	1.0U	2.0	10/29/10 14:54	
Vinyl chloride	ug/L	0.50U	1.0	10/29/10 14:54	
Xylene (Total)	ug/L	0.50U	1.0	10/29/10 14:54	
1,2-Dichloroethane-d4 (S)	%	104	86-125	10/29/10 14:54	
4-Bromofluorobenzene (S)	%	98	70-114	10/29/10 14:54	
Dibromofluoromethane (S)	%	99	88-117	10/29/10 14:54	
Toluene-d8 (S)	%	100	87-113	10/29/10 14:54	

LABORATORY CONTROL SAMPLE: 135450

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,1,1,2-Tetrachloroethane	ug/L	20	20.6	103	76.8-126.8	
1,1,1-Trichloroethane	ug/L	20	19.4	97	81.9-126.8	
1,1,2,2-Tetrachloroethane	ug/L	20	20.6	103	70.5-131.7	
1,1,2-Trichloroethane	ug/L	20	20.5	103	84.1-122.6	
1,1-Dichloroethane	ug/L	20	20.3	101	66.4-138.6	
1,1-Dichloroethene	ug/L	20	21.6	108	79.3-127.5	
1,1-Dichloropropene	ug/L	20	20.6	103	70.4-138.4	
1,2,3-Trichloropropane	ug/L	20	19.8	99	58.2-134.6	
1,2,4-Trichlorobenzene	ug/L	20	21.1	105	79.1-134.1	
1,2-Dichloroethane	ug/L	20	20.0	100	85.9-121.9	
1,2-Dichloropropane	ug/L	20	20.8	104	82.2-129.1	
1,3-Dichloropropane	ug/L	20	20.6	103	88.1-118.2	
2,2-Dichloropropane	ug/L	20	19.6	98	44-181.7	
2-Butanone (MEK)	ug/L	20	20.8	104	53.8-156.3	
2-Hexanone	ug/L	20	20.0	100	57.5-155.8	
4-Methyl-2-pentanone (MIBK)	ug/L	20	20.6	103	71.8-134.4	
Acetone	ug/L	20	22.2	111	47.2-184.1	
Acetonitrile	ug/L	200	206	103	65.2-133.1	
Acrolein	ug/L	200	228	114	41.8-131.7	
Acrylonitrile	ug/L	200	211	106	57.8-125.9	
Allyl chloride	ug/L	20	21.6	108	23.6-190.7	
Benzene	ug/L	20	20.7	104	77.3-132.8	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

LABORATORY CONTROL SAMPLE: 135450

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Bromochloromethane	ug/L	20	20.0	100	87.4-122.8	
Bromodichloromethane	ug/L	20	19.5	97	77.2-121.1	
Bromoform	ug/L	20	18.2	91	65.9-133.5	
Bromomethane	ug/L	20	27.6	138	48.2-223.9	
Carbon disulfide	ug/L	20	25.1	125	20.3-195.4	
Carbon tetrachloride	ug/L	20	19.5	98	69-155.5	
Chlorobenzene	ug/L	20	20.7	104	76.9-123.9	
Chloroethane	ug/L	20	20.4	102	46.7-157.8	
Chloroform	ug/L	20	19.2	96	69.7-132	
Chloromethane	ug/L	20	22.1	111	54.4-153.8	
Chloroprene	ug/L	20	21.2	106	31-185.4	
cis-1,2-Dichloroethene	ug/L	20	20.8	104	84-127.9	
cis-1,3-Dichloropropene	ug/L	20	20.6	103	73-121.6	
Dibromochloromethane	ug/L	20	18.8	94	65.4-126.2	
Dibromomethane	ug/L	20	19.9	100	85.3-121.7	
Dichlorodifluoromethane	ug/L	20	18.7	93	63.1-143.7	
Ethyl methacrylate	ug/L	20	22.1	110	34.3-179.4	
Ethylbenzene	ug/L	20	20.8	104	66.4-134.4	
Hexachloro-1,3-butadiene	ug/L	20	19.6	98	74.4-153.6	
Iodomethane	ug/L	20	15.8	79	1-243.3	
Isobutyl Alcohol	ug/L	400	422	105	62.9-136.1	
Methacrylonitrile	ug/L	200	222	111	77.3-132.6	
Methyl methacrylate	ug/L	20	20.6	103	37.4-178.3	
Methylene Chloride	ug/L	20	20.6	103	65.7-137.3	
Propionitrile	ug/L	200	212	106	71-130.3	
Styrene	ug/L	20	21.1	106	76.5-118.5	
Tetrachloroethene	ug/L	20	17.8	89	71-134	
Toluene	ug/L	20	20.9	105	75-129	
trans-1,2-Dichloroethene	ug/L	20	21.0	105	83.3-126.3	
trans-1,3-Dichloropropene	ug/L	20	20.2	101	67.6-130	
trans-1,4-Dichloro-2-butene	ug/L	20	19.6	98	36.1-177.4	
Trichloroethene	ug/L	20	19.9	99	81.1-122.4	
Trichlorofluoromethane	ug/L	20	20.7	103	75.4-124.6	
Vinyl acetate	ug/L	20	22.0	110	72.2-139	
Vinyl chloride	ug/L	20	22.8	114	70.2-136.9	
Xylene (Total)	ug/L	60	62.3	104	82.3-126	
1,2-Dichloroethane-d4 (S)	%			99	86-125	
4-Bromofluorobenzene (S)	%			101	70-114	
Dibromofluoromethane (S)	%			99	88-117	
Toluene-d8 (S)	%			100	87-113	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135451

135452

Parameter	Units	3520667001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
1,1,1,2-Tetrachloroethane	ug/L	0.50U	20	20	18.4	19.7	92	98	70-130	7	40	
1,1,1-Trichloroethane	ug/L	0.50U	20	20	19.7	21.0	98	105	70-130	6	40	

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 340 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135451 135452											
Parameter	Units	3520667001		MS	MSD	MS	MSD	MS	MSD	% Rec	Max
		Result	Conc.	Spike	Spike						RPD
				Conc.	Conc.	Result	Result	% Rec	% Rec	Limits	RPD
1,1,2,2-Tetrachloroethane	ug/L	0.18U	20	20	20	18.1	19.0	90	95	70-130	5 40
1,1,2-Trichloroethane	ug/L	0.50U	20	20	20	18.4	19.2	92	96	70-130	4 40
1,1-Dichloroethane	ug/L	0.50U	20	20	20	20.8	21.6	104	108	70-130	4 40
1,1-Dichloroethene	ug/L	0.50U	20	20	20	25.4	27.2	127	136	70-130	7 40 J(M1)
1,1-Dichloropropene	ug/L	0.50U	20	20	20	21.0	22.4	105	112	70-130	6 40
1,2,3-Trichloropropane	ug/L	0.36U	20	20	20	17.8	18.5	89	92	70-130	4 40
1,2,4-Trichlorobenzene	ug/L	0.50U	20	20	20	16.0	19.8	80	99	70-130	22 40
1,2-Dichloroethane	ug/L	0.50U	20	20	20	19.8	20.6	99	103	70-130	4 40
1,2-Dichloropropane	ug/L	0.50U	20	20	20	20.5	21.9	102	109	70-130	7 40
1,3-Dichloropropane	ug/L	0.50U	20	20	20	18.8	19.7	94	98	70-130	4 40
2,2-Dichloropropane	ug/L	0.50U	20	20	20	19.4	20.2	97	101	70-130	4 40
2-Butanone (MEK)	ug/L	5.0U	20	20	20	17.0	16.8	85	84	70-130	.9 40
2-Hexanone	ug/L	5.0U	20	20	20	16.4	17.6	82	88	70-130	7 40
4-Methyl-2-pentanone (MIBK)	ug/L	5.0U	20	20	20	18.7	19.5	94	97	70-130	4 40
Acetone	ug/L	5.0U	20	20	20	17.7	17.6	78	77	70-130	1 40
Acetonitrile	ug/L	5.0U	200	200	200	192	197	96	98	70-130	3 40
Acrolein	ug/L	10.0U	200	200	200	225	239	113	119	70-130	6 40
Acrylonitrile	ug/L	5.0U	200	200	200	195	204	98	102	70-130	4 40
Allyl chloride	ug/L	0.50U	20	20	20	21.7	23.0	109	115	70-130	6 40
Benzene	ug/L	0.50U	20	20	20	20.8	21.6	104	108	70-130	4 40
Bromochloromethane	ug/L	0.50U	20	20	20	18.8	19.9	94	99	70-130	6 40
Bromodichloromethane	ug/L	0.27U	20	20	20	18.6	19.4	93	97	70-130	4 40
Bromoform	ug/L	0.50U	20	20	20	14.5	16.3	72	82	70-130	12 40
Bromomethane	ug/L	0.50U	20	20	20	26.5	30.4	132	152	70-130	14 40 J(M1)
Carbon disulfide	ug/L	0.50U	20	20	20	24.7	25.9	123	129	70-130	5 40
Carbon tetrachloride	ug/L	0.50U	20	20	20	19.8	21.4	99	107	70-130	8 40
Chlorobenzene	ug/L	0.50U	20	20	20	19.1	20.2	96	101	70-130	6 40
Chloroethane	ug/L	0.50U	20	20	20	26.7	26.2	134	131	70-130	2 40 J(M1)
Chloroform	ug/L	0.50U	20	20	20	19.3	20.1	96	100	70-130	4 40
Chloromethane	ug/L	0.62U	20	20	20	27.6	28.5	138	143	70-130	3 40 J(M1)
Chloroprene	ug/L	0.50U	20	20	20	22.9	23.5	114	118	70-130	3 40
cis-1,2-Dichloroethene	ug/L	0.50U	20	20	20	21.3	22.2	106	111	70-130	4 40
cis-1,3-Dichloropropene	ug/L	0.25U	20	20	20	18.6	19.9	93	100	70-130	7 40
Dibromochloromethane	ug/L	0.26U	20	20	20	16.9	17.8	85	89	70-130	5 40
Dibromomethane	ug/L	0.50U	20	20	20	18.3	19.3	92	97	70-130	5 40
Dichlorodifluoromethane	ug/L	0.50U	20	20	20	24.8	25.7	124	129	70-130	4 40
Ethyl methacrylate	ug/L	0.50U	20	20	20	20.5	20.9	102	104	70-130	2 40
Ethylbenzene	ug/L	0.50U	20	20	20	19.4	21.1	97	106	70-130	8 40
Hexachloro-1,3-butadiene	ug/L	0.50U	20	20	20	16.3	19.2	81	96	70-130	16 40
Iodomethane	ug/L	0.50U	20	20	20	11.7	15.3	58	76	70-130	26 40 J(M1)
Isobutyl Alcohol	ug/L	10.0U	400	400	400	316	364	79	91	70-130	14 40
Methacrylonitrile	ug/L	5.0U	200	200	200	214	219	107	110	70-130	2 40
Methyl methacrylate	ug/L	5.0U	20	20	20	17.7	18.6	89	93	70-130	5 40
Methylene Chloride	ug/L	2.5U	20	20	20	20.7	21.8	102	108	70-130	5 40
Propionitrile	ug/L	5.0U	200	200	200	189	196	95	98	70-130	4 40
Styrene	ug/L	0.50U	20	20	20	19.3	20.5	96	102	70-130	6 40
Tetrachloroethene	ug/L	0.50U	20	20	20	15.4	16.5	77	83	70-130	7 40

Date: 01/05/2011 04:18 PM

REPORT OF LABORATORY ANALYSIS

Page 341 of 385

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, Inc..



QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135451 135452											
Parameter	Units	3520667001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
Toluene	ug/L	0.89 I	20	20	21.2	21.3	101	102	70-130	.8	40
trans-1,2-Dichloroethene	ug/L	0.50U	20	20	21.1	22.5	106	112	70-130	6	40
trans-1,3-Dichloropropene	ug/L	0.25U	20	20	18.1	19.3	91	97	70-130	7	40
trans-1,4-Dichloro-2-butene	ug/L	5.0U	20	20	17.1	17.2	85	86	70-130	.5	40
Trichloroethene	ug/L	0.50U	20	20	19.3	21.0	97	105	70-130	8	40
Trichlorofluoromethane	ug/L	0.50U	20	20	25.8	25.5	129	128	70-130	1	40
Vinyl acetate	ug/L	1.0U	20	20	18.4	18.8	92	94	70-130	2	40
Vinyl chloride	ug/L	0.50U	20	20	28.3	28.5	141	142	70-130	.7	40 J(M1)
Xylene (Total)	ug/L	0.50U	60	60	58.1	63.1	97	105	70-130	8	40
1,2-Dichloroethane-d4 (S)	%						103	104	86-125		
4-Bromofluorobenzene (S)	%						97	99	70-114		
Dibromofluoromethane (S)	%						98	100	88-117		
Toluene-d8 (S)	%						100	101	87-113		

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5629

Analysis Method: SM 2320B

QC Batch Method: SM 2320B

Analysis Description: 2320B Alkalinity

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 125357

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0	10/07/10 14:12	
Alkalinity, Total as CaCO ₃	mg/L	5.0U	5.0	10/07/10 14:12	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	5.0U	5.0	10/07/10 14:12	

LABORATORY CONTROL SAMPLE: 125358

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	250	241	96	90-110	

MATRIX SPIKE SAMPLE: 125360

Parameter	Units	3519243004 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	1270	250	1480	83	90-110	J(M1)

MATRIX SPIKE SAMPLE: 125362

Parameter	Units	3519451001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	16.8	250	266	100	90-110	

SAMPLE DUPLICATE: 125359

Parameter	Units	3519243004 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	993	956	4	20	
Alkalinity, Total as CaCO ₃	mg/L	1270	1230	3	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	278	275	1	20	

SAMPLE DUPLICATE: 125361

Parameter	Units	3519451001 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	ND	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	16.8	17.1	2	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	16.8	17.1	2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5776 Analysis Method: SM 2320B
QC Batch Method: SM 2320B Analysis Description: 2320B Alkalinity
Associated Lab Samples: 3519325015, 3519325016, 3519325018, 3519325019, 3519325020

METHOD BLANK: 130980 Matrix: Water
Associated Lab Samples: 3519325015, 3519325016, 3519325018, 3519325019, 3519325020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0	10/20/10 10:08	
Alkalinity, Total as CaCO ₃	mg/L	5.0U	5.0	10/20/10 10:08	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	5.0U	5.0	10/20/10 10:08	

LABORATORY CONTROL SAMPLE: 130981

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	250	247	99	90-110	

MATRIX SPIKE SAMPLE: 130983

Parameter	Units	3519325015 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	785	250	980	78	90-110	J(M1)

MATRIX SPIKE SAMPLE: 130985

Parameter	Units	3520168011 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	5.0U	250	252	100	90-110	

SAMPLE DUPLICATE: 130982

Parameter	Units	3519325015 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	785	771	2	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	785	771	2	20	

SAMPLE DUPLICATE: 130984

Parameter	Units	3520168011 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	5.0U	5.0U		20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5804

Analysis Method: SM 2320B

QC Batch Method: SM 2320B

Analysis Description: 2320B Alkalinity

Associated Lab Samples: 3519325025, 3519325029

METHOD BLANK: 132175

Matrix: Water

Associated Lab Samples: 3519325025, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0	10/22/10 08:28	
Alkalinity, Total as CaCO ₃	mg/L	5.0U	5.0	10/22/10 08:28	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	5.0U	5.0	10/22/10 08:28	

LABORATORY CONTROL SAMPLE: 132176

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	250	241	96	90-110	

MATRIX SPIKE SAMPLE: 132178

Parameter	Units	3519325025 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	710	250	950	96	90-110	

MATRIX SPIKE SAMPLE: 132180

Parameter	Units	3520549005 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	150	250	373	89	90-110	J(M1)

SAMPLE DUPLICATE: 132177

Parameter	Units	3519325025 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	710	721	2	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	710	721	2	20	

SAMPLE DUPLICATE: 132179

Parameter	Units	3520549005 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	150	153	2	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	150	153	2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5978

Analysis Method: SM 2320B

QC Batch Method: SM 2320B

Analysis Description: 2320B Alkalinity

Associated Lab Samples: 3519325039

METHOD BLANK: 137472

Matrix: Water

Associated Lab Samples: 3519325039

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0	11/04/10 09:59	
Alkalinity, Total as CaCO ₃	mg/L	5.0U	5.0	11/04/10 09:59	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	5.0U	5.0	11/04/10 09:59	

LABORATORY CONTROL SAMPLE: 137473

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	250	244	98	90-110	

MATRIX SPIKE SAMPLE: 137475

Parameter	Units	3521348005 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	136	250	377	96	90-110	

MATRIX SPIKE SAMPLE: 137477

Parameter	Units	3521433007 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Alkalinity, Total as CaCO ₃	mg/L	183	250	279	38	90-110	J(M1)

SAMPLE DUPLICATE: 137474

Parameter	Units	3521348005 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	136	136	.1	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	136	136	.1	20	

SAMPLE DUPLICATE: 137476

Parameter	Units	3521433007 Result	Dup Result	RPD	Max RPD	Qualifiers
Alkalinity, Carbonate (CaCO ₃)	mg/L	5.0U	5.0U		20	
Alkalinity, Total as CaCO ₃	mg/L	183	184	.6	20	
Alkalinity,Bicarbonate (CaCO ₃)	mg/L	183	184	.6	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5597

Analysis Method: SM 2540C

QC Batch Method: SM 2540C

Analysis Description: 2540C Total Dissolved Solids

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 124317

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Dissolved Solids	mg/L	5.0U	5.0	10/06/10 08:31	

LABORATORY CONTROL SAMPLE: 124318

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Dissolved Solids	mg/L	300	296	99	90-110	

SAMPLE DUPLICATE: 124319

Parameter	Units	3518753050 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	305	306	.3	20	

SAMPLE DUPLICATE: 124320

Parameter	Units	3519210001 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	155	157	1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5756 Analysis Method: SM 2540C
QC Batch Method: SM 2540C Analysis Description: 2540C Total Dissolved Solids
Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 130389 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Dissolved Solids	mg/L	5.0U	5.0	10/21/10 04:00	

LABORATORY CONTROL SAMPLE: 130390

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Dissolved Solids	mg/L	300	308	103	90-110	

SAMPLE DUPLICATE: 130391

Parameter	Units	3520108018 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	235	236	.4	20	

SAMPLE DUPLICATE: 135232

Parameter	Units	3520168020 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	31.0	39.0	23	20	J(D6)

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WET/5773	Analysis Method:	SM 2540C
QC Batch Method:	SM 2540C	Analysis Description:	2540C Total Dissolved Solids
Associated Lab Samples:	3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325017, 3519325018, 3519325019, 3519325020		

METHOD BLANK:	130937	Matrix:	Water
Associated Lab Samples:	3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325017, 3519325018, 3519325019, 3519325020		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Dissolved Solids	mg/L	5.0U	5.0	10/21/10 16:00	

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Dissolved Solids	mg/L	300	292	97	90-110	

Parameter	Units	3520441006 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	225	224	.4	20	

Parameter	Units	3519325012 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	2810	2940	4	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WET/5821	Analysis Method:	SM 2540C
QC Batch Method:	SM 2540C	Analysis Description:	2540C Total Dissolved Solids
Associated Lab Samples:	3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029		

METHOD BLANK:	132584	Matrix:	Water
Associated Lab Samples:	3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Dissolved Solids	mg/L	5.0U	5.0	10/23/10 10:20	

LABORATORY CONTROL SAMPLE: 132585

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Dissolved Solids	mg/L	300	303	101	90-110	

SAMPLE DUPLICATE: 132586

Parameter	Units	3519325023 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	770	772	.3	20	

SAMPLE DUPLICATE: 132587

Parameter	Units	3520569020 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	197	192	3	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WET/5959	Analysis Method:	SM 2540C
QC Batch Method:	SM 2540C	Analysis Description:	2540C Total Dissolved Solids
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

METHOD BLANK:	136761	Matrix:	Water
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Dissolved Solids	mg/L	5.0U	5.0	11/03/10 11:56	

LABORATORY CONTROL SAMPLE: 136762

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Dissolved Solids	mg/L	300	296	99	90-110	

SAMPLE DUPLICATE: 136763

Parameter	Units	3518484035 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	806	800	.7	20	

SAMPLE DUPLICATE: 136764

Parameter	Units	3520442016 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Dissolved Solids	mg/L	446	438	2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5579

Analysis Method: SM 2540D

QC Batch Method: SM 2540D

Analysis Description: 2540D Total Suspended Solids

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123482

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Suspended Solids	mg/L	5.0U	5.0	10/04/10 14:08	

LABORATORY CONTROL SAMPLE: 123483

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Suspended Solids	mg/L	80	84.0	105	90-110	

SAMPLE DUPLICATE: 123484

Parameter	Units	3519260001 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Suspended Solids	mg/L	5.0U	5.0U		20	

SAMPLE DUPLICATE: 123485

Parameter	Units	3519244001 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Suspended Solids	mg/L	5.0U	5.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WET/5755	Analysis Method:	SM 4500-S2E
QC Batch Method:	SM 4500-S2E	Analysis Description:	4500S2E Sulfide, Iodometric
Associated Lab Samples:	3519325006, 3519325007, 3519325008, 3519325009		

METHOD BLANK: 130376 Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfide	mg/L	1.0U	1.0	10/19/10 08:30	

LABORATORY CONTROL SAMPLE: 130377

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	6	5.2	87	80-120	

MATRIX SPIKE SAMPLE: 130379

Parameter	Units	3520338001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	1.0U	6	5.4	91	80-120	

SAMPLE DUPLICATE: 130378

Parameter	Units	3520338001 Result	Dup Result	RPD	Max RPD	Qualifiers
Sulfide	mg/L	1.0U	1.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5777 Analysis Method: SM 4500-S2E
QC Batch Method: SM 4500-S2E Analysis Description: 4500S2E Sulfide, Iodometric
Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

METHOD BLANK: 130986 Matrix: Water

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfide	mg/L	1.0U	1.0	10/20/10 10:00	

LABORATORY CONTROL SAMPLE: 130987

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	6	5.2	86	80-120	

MATRIX SPIKE SAMPLE: 131233

Parameter	Units	3520577002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	1.0U	6	5.4	85	80-120	

SAMPLE DUPLICATE: 131232

Parameter	Units	3520577002 Result	Dup Result	RPD	Max RPD	Qualifiers
Sulfide	mg/L	1.0U	1.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5852 Analysis Method: SM 4500-S2E
QC Batch Method: SM 4500-S2E Analysis Description: 4500S2E Sulfide, Iodometric
Associated Lab Samples: 3519325026, 3519325027, 3519325028

METHOD BLANK: 133278 Matrix: Water

Associated Lab Samples: 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfide	mg/L	1.0U	1.0	10/25/10 13:30	

LABORATORY CONTROL SAMPLE: 133279

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	6	5.1	86	80-120	

MATRIX SPIKE SAMPLE: 133281

Parameter	Units	3520683002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	1.0U	6	5.4	88	80-120	

SAMPLE DUPLICATE: 133280

Parameter	Units	3520683002 Result	Dup Result	RPD	Max RPD	Qualifiers
Sulfide	mg/L	1.0U	1.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5938

Analysis Method: SM 4500-S2E

QC Batch Method: SM 4500-S2E

Analysis Description: 4500S2E Sulfide, Iodometric

Associated Lab Samples: 3519325032, 3519325033

METHOD BLANK: 136050

Matrix: Water

Associated Lab Samples: 3519325032, 3519325033

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfide	mg/L	1.0U	1.0	11/01/10 11:50	

LABORATORY CONTROL SAMPLE: 136051

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	6	6.0	99	80-120	

MATRIX SPIKE SAMPLE: 136053

Parameter	Units	3520930001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	2.1	6	8.2	102	80-120	

SAMPLE DUPLICATE: 136052

Parameter	Units	3520930001 Result	Dup Result	RPD	Max RPD	Qualifiers
Sulfide	mg/L	2.1	2.2	7	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5951 Analysis Method: SM 4500-S2E
QC Batch Method: SM 4500-S2E Analysis Description: 4500S2E Sulfide, Iodometric
Associated Lab Samples: 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 136395 Matrix: Water

Associated Lab Samples: 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfide	mg/L	1.0U	1.0	11/02/10 09:00	

LABORATORY CONTROL SAMPLE: 136396

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	6	5.8	97	80-120	

MATRIX SPIKE SAMPLE: 136398

Parameter	Units	3521133001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Sulfide	mg/L	1.0U	6	6.0	94	80-120	

SAMPLE DUPLICATE: 136397

Parameter	Units	3521133001 Result	Dup Result	RPD	Max RPD	Qualifiers
Sulfide	mg/L	1.0U	1.0U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WET/5590

Analysis Method: SM10200

QC Batch Method: SM10200

Analysis Description: Chlorophyll & Pheophytin

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123887

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chlorophyll a	ug/L	1.0U	1.0	10/15/10 10:51	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6755

Analysis Method: EPA 300.0

QC Batch Method: EPA 300.0

Analysis Description: 300.0 IC Anions

Associated Lab Samples: 3519325001

METHOD BLANK: 129801

Matrix: Water

Associated Lab Samples: 3519325001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Sulfate	mg/L	2.5U	5.0	10/16/10 22:07	

LABORATORY CONTROL SAMPLE: 129802

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Sulfate	mg/L	50	49.6	99	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 129803

129804

Parameter	Units	3518484018 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Sulfate	mg/L	34.7	50	50	89.1	89.0	109	109	90-110	.01	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 129805

129806

Parameter	Units	3518484030 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Sulfate	mg/L	2.5U	50	50	48.3	48.2	96	96	90-110	.1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6773 Analysis Method: EPA 300.0
QC Batch Method: EPA 300.0 Analysis Description: 300.0 IC Anions
Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 130098 Matrix: Water
Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chloride	mg/L	2.5U	5.0	10/18/10 15:22	

LABORATORY CONTROL SAMPLE: 130099

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chloride	mg/L	50	45.4	91	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130102 130103

Parameter	Units	3519799033 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	35.9	50	50	89.1	89.2	106	107	90-110	.08	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6964

Analysis Method: EPA 300.0

QC Batch Method: EPA 300.0

Analysis Description: 300.0 IC Anions

Associated Lab Samples: 3519325002

METHOD BLANK: 134364

Matrix: Water

Associated Lab Samples: 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chloride	mg/L	2.5U	5.0	10/27/10 16:38	
Sulfate	mg/L	2.5U	5.0	10/27/10 16:38	

LABORATORY CONTROL SAMPLE: 134365

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chloride	mg/L	50	47.4	95	90-110	
Sulfate	mg/L	50	47.5	95	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134366

134367

Parameter	Units	3519325002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	26.2	50	50	78.0	78.0	104	104	90-110	.03	20	
Sulfate	mg/L	34.3	50	50	86.2	86.3	104	104	90-110	.07	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134368

134369

Parameter	Units	3520344002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	95.7	50	50	151	151	110	110	90-110	.08	20	
Sulfate	mg/L	6.4	50	50	54.2	54.3	96	96	90-110	.2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6986 Analysis Method: EPA 300.0
QC Batch Method: EPA 300.0 Analysis Description: 300.0 IC Anions
Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

METHOD BLANK: 134635 Matrix: Water
Associated Lab Samples: 3519325025, 3519325026, 3519325027, 3519325028, 3519325029

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chloride	mg/L	2.5U	5.0	10/28/10 12:46	
Sulfate	mg/L	2.5U	5.0	10/28/10 12:46	

LABORATORY CONTROL SAMPLE: 134636

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chloride	mg/L	50	46.6	93	90-110	
Sulfate	mg/L	50	45.7	91	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134637 134638

Parameter	Units	3519325025 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	16.4	100	100	113	113	97	96	90-110	.1	20	
Sulfate	mg/L	5.0U	100	100	90.7	90.7	90	90	90-110	.04	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 134639 134640

Parameter	Units	3520634003 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	132	50	50	185	185	106	106	90-110	.2	20	
Sulfate	mg/L	78.7	50	50	135	135	113	113	90-110	.2	20 J(M1)	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/7028	Analysis Method:	EPA 300.0
QC Batch Method:	EPA 300.0	Analysis Description:	300.0 IC Anions
Associated Lab Samples:	3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020		

METHOD BLANK:	135222	Matrix:	Water
Associated Lab Samples:	3519325011, 3519325012, 3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019, 3519325020		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chloride	mg/L	2.5U	5.0	10/29/10 13:21	
Sulfate	mg/L	2.5U	5.0	10/29/10 13:21	

LABORATORY CONTROL SAMPLE: 135223						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chloride	mg/L	50	46.7	93	90-110	
Sulfate	mg/L	50	46.1	92	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135224 135225												
Parameter	Units	3519325011	MS	MSD	MS	MSD	MS	MSD	% Rec	Max	Qual	
		Result	Spike Conc.	Spike Conc.								Result
Chloride	mg/L	158	250	250	414	415	102	103	90-110	.4	20	
Sulfate	mg/L	96.7	250	250	346	342	100	98	90-110	1	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 135246 135247												
Parameter	Units	3520870001	MS	MSD	MS	MSD	MS	MSD	% Rec	Max		
		Result	Spike	Spike								
Chloride	mg/L	62.5U	1250	1250	1220	1220	95	95	90-110	.03	20	
Sulfate	mg/L	62.5U	1250	1250	1230	1230	94	94	90-110	.1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/7064	Analysis Method:	EPA 300.0
QC Batch Method:	EPA 300.0	Analysis Description:	300.0 IC Anions
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

METHOD BLANK:	136160	Matrix:	Water
Associated Lab Samples:	3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037, 3519325039		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chloride	mg/L	2.5U	5.0	11/01/10 22:53	
Sulfate	mg/L	2.5U	5.0	11/01/10 22:53	

LABORATORY CONTROL SAMPLE: 136161

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chloride	mg/L	50	47.2	94	90-110	
Sulfate	mg/L	50	46.5	93	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136162 136163

Parameter	Units	3518484035 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	112	250	250	351	351	96	96	90-110	.1	20	
Sulfate	mg/L		250	250	286	285	92	91	90-110	.4	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136164 136165

Parameter	Units	3520945001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Chloride	mg/L	284	250	250	546	538	105	102	90-110	1	20	
Sulfate	mg/L	131	250	250	384	375	101	98	90-110	3	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6784

Analysis Method: EPA 335.4

QC Batch Method: EPA 335.4

Analysis Description: 335.4 Cyanide, Total

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

METHOD BLANK: 130416

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Cyanide	mg/L	0.0050U	0.010	10/19/10 16:10	

LABORATORY CONTROL SAMPLE: 130417

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Cyanide	mg/L	.05	0.047	93	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130418

130419

Parameter	Units	3520486001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	0.0050 U	.05	.05	0.047	0.048	93	94	90-110	.8	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 130420

130421

Parameter	Units	9279829001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	ND	.05	.05	0.047	0.048	92	95	90-110	3	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6870

Analysis Method: EPA 335.4

QC Batch Method: EPA 335.4

Analysis Description: 335.4 Cyanide, Total

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014, 3519325026, 3519325027, 3519325028

METHOD BLANK: 132339

Matrix: Water

Associated Lab Samples: 3519325011, 3519325012, 3519325013, 3519325014, 3519325026, 3519325027, 3519325028

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Cyanide	mg/L	0.0050U	0.010	10/22/10 15:54	

LABORATORY CONTROL SAMPLE: 132340

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Cyanide	mg/L	.05	0.049	97	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132341

132342

Parameter	Units	3519325011 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	0.0050 U	.05	.05	0.051	0.051	100	100	90-110	.6	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 132343

132344

Parameter	Units	3519452052 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	0.0050 U	.05	.05	0.051	0.050	100	99	90-110	1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7056 Analysis Method: EPA 335.4
QC Batch Method: EPA 335.4 Analysis Description: 335.4 Cyanide, Total
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

METHOD BLANK: 136069 Matrix: Water
Associated Lab Samples: 3519325032, 3519325033, 3519325034, 3519325035, 3519325036, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Cyanide	mg/L	0.0050U	0.010	11/08/10 14:41	

LABORATORY CONTROL SAMPLE: 136070

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Cyanide	mg/L	.05	0.048	97	90-110	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136071 136072

Parameter	Units	3521176002 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	0.0050 U	.05	.05	0.051	0.050	93	91	90-110	2	20	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 136073 136074

Parameter	Units	3519325032 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Cyanide	mg/L	0.025U	.05	.05	0.033 I	0.029 I	64	56	90-110		20	4p, J(M1)

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6598

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325001

METHOD BLANK: 125089

Matrix: Water

Associated Lab Samples: 3519325001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	10/07/10 14:28	

LABORATORY CONTROL SAMPLE: 125090

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	0.93	93	90-110	

MATRIX SPIKE SAMPLE: 125092

Parameter	Units	3518938021 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	1	1.0	100	90-110	

SAMPLE DUPLICATE: 125091

Parameter	Units	3518938021 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.020U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6599

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325002

METHOD BLANK: 125095

Matrix: Water

Associated Lab Samples: 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	10/07/10 12:47	

LABORATORY CONTROL SAMPLE: 125096

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	0.99	99	90-110	

MATRIX SPIKE SAMPLE: 125098

Parameter	Units	3519325002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.099	1	1.1	96	90-110	

SAMPLE DUPLICATE: 125097

Parameter	Units	3519325002 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.099	0.090	9	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6876

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325017

METHOD BLANK: 132456

Matrix: Water

Associated Lab Samples: 3519325006, 3519325007, 3519325008, 3519325009, 3519325011, 3519325012, 3519325017

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	10/22/10 13:03	

LABORATORY CONTROL SAMPLE: 132457

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.1	107	90-110	

MATRIX SPIKE SAMPLE: 132459

Parameter	Units	3520638009 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	1	0.95	95	90-110	

SAMPLE DUPLICATE: 132458

Parameter	Units	3520638009 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.020U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/6894	Analysis Method:	EPA 350.1
QC Batch Method:	EPA 350.1	Analysis Description:	350.1 Ammonia
Associated Lab Samples:	3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019		

METHOD BLANK:	133072	Matrix:	Water
Associated Lab Samples:	3519325013, 3519325014, 3519325015, 3519325016, 3519325018, 3519325019		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	10/25/10 08:09	

LABORATORY CONTROL SAMPLE: 133073

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.1	105	90-110	

MATRIX SPIKE SAMPLE: 133075

Parameter	Units	3520572020 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.99	1	2.0	97	90-110	

SAMPLE DUPLICATE: 133074

Parameter	Units	3520572020 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.99	0.98	.2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/6895	Analysis Method:	EPA 350.1
QC Batch Method:	EPA 350.1	Analysis Description:	350.1 Ammonia
Associated Lab Samples:	3519325020, 3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029		

METHOD BLANK:	133080	Matrix:	Water
Associated Lab Samples:	3519325020, 3519325023, 3519325024, 3519325025, 3519325026, 3519325027, 3519325028, 3519325029		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	10/25/10 08:50	

LABORATORY CONTROL SAMPLE: 133081

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.1	105	90-110	

MATRIX SPIKE SAMPLE: 133083

Parameter	Units	3519325020 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	25.5	5	28.9	68	90-110	J(M1)

SAMPLE DUPLICATE: 133082

Parameter	Units	3519325020 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	25.5	25.4	.2	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7046

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325032, 3519325037

METHOD BLANK: 135972

Matrix: Water

Associated Lab Samples: 3519325032, 3519325037

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	11/01/10 09:40	

LABORATORY CONTROL SAMPLE: 135973

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.1	105	90-110	

MATRIX SPIKE SAMPLE: 135975

Parameter	Units	3519325032 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	599	100	703	104	90-110	

SAMPLE DUPLICATE: 135974

Parameter	Units	3519325032 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	599	600	.1	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7283

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325033

METHOD BLANK: 141012

Matrix: Water

Associated Lab Samples: 3519325033

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	11/12/10 13:49	

LABORATORY CONTROL SAMPLE: 141013

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	0.98	98	90-110	

MATRIX SPIKE SAMPLE: 141015

Parameter	Units	3521739057 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.26	1	1.2	95	90-110	

SAMPLE DUPLICATE: 141014

Parameter	Units	3521739057 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.26	0.27	4	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7284

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325034, 3519325035

METHOD BLANK: 141022

Matrix: Water

Associated Lab Samples: 3519325034, 3519325035

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	11/12/10 14:33	

LABORATORY CONTROL SAMPLE: 141023

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	0.99	99	90-110	

MATRIX SPIKE SAMPLE: 141025

Parameter	Units	3521645001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	1	0.97	97	90-110	

SAMPLE DUPLICATE: 141024

Parameter	Units	3521645001 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.020U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7300

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325036

METHOD BLANK: 141533

Matrix: Water

Associated Lab Samples: 3519325036

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	11/15/10 08:08	

LABORATORY CONTROL SAMPLE: 141534

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.0	102	90-110	

MATRIX SPIKE SAMPLE: 141536

Parameter	Units	3521739062 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	0.076	1	0.83	76	90-110	J(M1)

SAMPLE DUPLICATE: 141535

Parameter	Units	3521739062 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	0.076	0.083	10	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/7301

Analysis Method: EPA 350.1

QC Batch Method: EPA 350.1

Analysis Description: 350.1 Ammonia

Associated Lab Samples: 3519325039

METHOD BLANK: 141537

Matrix: Water

Associated Lab Samples: 3519325039

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	0.020U	0.050	11/15/10 12:47	

LABORATORY CONTROL SAMPLE: 141538

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.0	101	90-110	

MATRIX SPIKE SAMPLE: 141540

Parameter	Units	3519325039 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1010	100	1090	76	90-110	M6

SAMPLE DUPLICATE: 141539

Parameter	Units	3519325039 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	1010	1010	.06	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/6549	Analysis Method:	EPA 351.2
QC Batch Method:	EPA 351.2	Analysis Description:	351.2 TKN
Associated Lab Samples: 3519325001, 3519325002			

METHOD BLANK: 123836 Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Kjeldahl, Total	mg/L	0.25U	0.50	10/06/10 12:15	

LABORATORY CONTROL SAMPLE: 123837

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Kjeldahl, Total	mg/L	20	20.6	103	90-110	

MATRIX SPIKE SAMPLE: 123839

Parameter	Units	3519320021 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Kjeldahl, Total	mg/L	0.67	20	20.3	98	90-110	

SAMPLE DUPLICATE: 123838

Parameter	Units	3519320021 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Kjeldahl, Total	mg/L	0.67	0.67	.4	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6543

Analysis Method: EPA 353.2

QC Batch Method: EPA 353.2

Analysis Description: 353.2 Nitrate + Nitrite, preserved

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123760

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, NO2 plus NO3	mg/L	0.025U	0.050	10/05/10 11:14	

LABORATORY CONTROL SAMPLE: 123761

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, NO2 plus NO3	mg/L	2	1.9	97	90-110	

MATRIX SPIKE SAMPLE: 123763

Parameter	Units	3519174001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, NO2 plus NO3	mg/L	1.2	2	3.1	93	80-120	

SAMPLE DUPLICATE: 123762

Parameter	Units	3519174001 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, NO2 plus NO3	mg/L	1.2	1.2	.6	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6550

Analysis Method: EPA 365.4

QC Batch Method: EPA 365.4

Analysis Description: 365.4 Phosphorus

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 123841

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Phosphorus, Total (as P)	mg/L	0.050U	0.10	10/06/10 12:50	

LABORATORY CONTROL SAMPLE: 123842

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Phosphorus, Total (as P)	mg/L	4	4.0	100	90-110	

MATRIX SPIKE SAMPLE: 123844

Parameter	Units	3519320021 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Phosphorus, Total (as P)	mg/L	0.24	4	4.2	98	80-120	

SAMPLE DUPLICATE: 123843

Parameter	Units	3519320021 Result	Dup Result	RPD	Max RPD	Qualifiers
Phosphorus, Total (as P)	mg/L	0.24	0.23	.7	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6583

Analysis Method: EPA 410.4

QC Batch Method: EPA 410.4

Analysis Description: 410.4 COD

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 124848

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chemical Oxygen Demand	mg/L	12.5U	25.0	10/06/10 17:55	

LABORATORY CONTROL SAMPLE: 124849

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chemical Oxygen Demand	mg/L	500	499	100	90-110	

MATRIX SPIKE SAMPLE: 124851

Parameter	Units	3518484011 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Chemical Oxygen Demand	mg/L	38.0	500	538	100	90-110	

SAMPLE DUPLICATE: 124850

Parameter	Units	3518484011 Result	Dup Result	RPD	Max RPD	Qualifiers
Chemical Oxygen Demand	mg/L	38.0	44.5	16	20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch:	WETA/7078	Analysis Method:	EPA 410.4
QC Batch Method:	EPA 410.4	Analysis Description:	410.4 COD
Associated Lab Samples:	3519325039		

METHOD BLANK: 136526 Matrix: Water

Associated Lab Samples: 3519325039

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Chemical Oxygen Demand	mg/L	12.5U	25.0	11/02/10 14:24	

LABORATORY CONTROL SAMPLE: 136527

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Chemical Oxygen Demand	mg/L	500	507	101	90-110	

MATRIX SPIKE SAMPLE: 136529

Parameter	Units	3521205001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Chemical Oxygen Demand	mg/L	12.5U	500	485	96	90-110	

SAMPLE DUPLICATE: 136528

Parameter	Units	3521205001 Result	Dup Result	RPD	Max RPD	Qualifiers
Chemical Oxygen Demand	mg/L	12.5U	12.5U		20	

QUALITY CONTROL DATA

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

QC Batch: WETA/6568

Analysis Method: SM 5310B

QC Batch Method: SM 5310B

Analysis Description: 5310B TOC

Associated Lab Samples: 3519325001, 3519325002

METHOD BLANK: 124340

Matrix: Water

Associated Lab Samples: 3519325001, 3519325002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Total Organic Carbon	mg/L	0.50U	1.0	10/06/10 09:30	

LABORATORY CONTROL SAMPLE: 124341

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Total Organic Carbon	mg/L	20	21.2	106	90-110	

MATRIX SPIKE SAMPLE: 124343

Parameter	Units	3518484011 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Total Organic Carbon	mg/L	11.6	20	34.1	113	80-120	

SAMPLE DUPLICATE: 124342

Parameter	Units	3518484011 Result	Dup Result	RPD	Max RPD	Qualifiers
Total Organic Carbon	mg/L	11.6	11.4	2	20	

QUALIFIERS

Project: Sarasota Central Landfill Comp
Pace Project No.: 3519325

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to changes in sample preparation, dilution of the sample aliquot, or moisture content.

ND - Not Detected at or above adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

S - Surrogate

1,2-Diphenylhydrazine (8270 listed analyte) decomposes to Azobenzene.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is NELAP accredited. Contact your Pace PM for the current list of accredited analytes.

LABORATORIES

PASI-G Pace Analytical Services - Green Bay
PASI-O Pace Analytical Services - Ormond Beach

ANALYTE QUALIFIERS

I	The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
1p	A heavy emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.
2p	An emulsion was generated during the extraction process employed in the preparation of the sample for this analysis.
3p	Reported result is estimated due to significant matrix interference to the related internal standard.
4p	Sample required a dilution due to matrix interference, which resulted in elevated reporting limits for the target compound(s).
5p	Sample was received with headspace.
6p	The continuing calibration for this compound is outside (HIGH) of method control limits. The result is estimated.
7p	The continuing calibration for this compound is outside of method control limits for this compound (high/low bias). However an acceptable Reporting Limit standard was analyzed at the end of the sequence demonstrating appropriate instrument sensitivity.
8p	The internal standard response associated with this result exceeds the lower control limit. However, the data is accepted based on surrogate compound recovery meeting control limits.
9p	The internal standard response associated with this result exceeds the upper control limit. However, the data is accepted based on surrogate compound recovery meeting control limits.
CH	The continuing calibration for this compound is outside of Pace Analytical acceptance limits. The results may be biased high.
D3	Sample was diluted due to the presence of high levels of non-target analytes or other matrix interference.
D4	Sample was diluted due to the presence of high levels of target analytes.
ES	The reported result is estimated because one or more of the constituent results are qualified as such.
J(D6)	Estimated Value. The relative percent difference (RPD) between the sample and sample duplicate exceeded laboratory control limits.
J(F3)	Estimated Value. The recovery of the second source standard used to verify the initial calibration curve for this analyte is outside the laboratory's control limit. The result is estimated.

QUALIFIERS

Project: Sarasota Central Landfill Comp

Pace Project No.: 3519325

ANALYTE QUALIFIERS

J(M0)	Estimated Value. Matrix spike recovery was outside laboratory control limits.
J(M1)	Estimated Value. Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.
J(P6)	Estimated Value. Matrix spike recovery was outside laboratory control limits due to a parent sample concentration notably higher than the spike level.
J(S0)	Estimated Value. Surrogate recovery outside laboratory control limits.
J(S2)	Estimated Value. Surrogate recovery outside laboratory control limits due to matrix interferences (confirmed by similar results from sample re-analysis).
J(S5)	Estimated Value. Surrogate recovery outside control limits due to matrix interferences (not confirmed by re-analysis).
J(SS)	Estimated Value. This analyte did not meet the secondary source verification criteria for the initial calibration. The reported result should be considered an estimated value.
L3	Analyte recovery in the laboratory control sample (LCS) exceeded QC limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.
M6	Matrix spike and Matrix spike duplicate recovery not evaluated against control limits due to sample dilution.
Q	Sample held beyond the accepted holding time.
S3	Surrogate recovery exceeded laboratory control limits. Analyte presence below reporting limits in associated samples. Results unaffected by high bias.
Z3	Methylene chloride is a common laboratory contaminant. Results for this analyte should be considered estimated unless the amount found in the sample is 3 to 5 times higher than that found in the method blank.
p2	Post-analysis pH measurement indicates pH > 2.