

FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798


Invoice

Description	Amount	Units	Extension
BTEXM8260	60.67	2	121.34
EPA6020 - Cadmium	17.69	1	17.69
EPA6020 - Chromium	17.69	1	17.69
EPA6020A - Lead	17.69	1	17.69
EPA6020- Arsenic	17.69	1	17.69
EPA8270-SIM-PAH	128.89	1	128.89
TOTAL			320.99

This invoice is due upon receipt.

VISA, MasterCard, and American Express cards will be accepted.

There will be a 1.5% service charge per month on all unpaid balances.



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Report Summary

Date Received: Apr 27, 2011

FCL Project Manager: June S. Flowers


Laboratory #	Sample Description	Analysis	Chemist	Location	Sample Matrix
148798GW1	MW-10R	EPA8260	CLS	Main Lab	Ground Water
		EPA8270	CLS	Main Lab	
		XBNE	CDG	Main Lab	
148798SO1	IDW	EPA6020	EVB	Main Lab	Soil
		EPA8260	CLS	Main Lab	

Certificate of Results

Sample integrity was certified prior to analysis. Test results meet all requirements of the NELAC Standards except as noted in the Quality Control Report. Uncertainties for these data are available on request. This report may not be reproduced in part; results relate only to items tested.



Jefferson S. Flowers, Ph.D.
President/Technical Director



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)


Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Analysis Report

Lab #: 148798GW1 Sampled: 04/26/11 08:20 AM Desc: MW-10R

Parameter	Result	Units	DF	MDL	PQL	QC Batch	Method	CAS #	Analyzed
Benzene	9.66	ug/L	1.00	0.500	1.00	10173329	EPA8260	71-43-2	04/29/11
Ethylbenzene	34.3	ug/L	1.00	0.500	1.00	10173329	EPA8260	100-41-4	04/29/11
Methyl-tert-butylether	1.00 U	ug/L	1.00	1.00	2.00	10173329	EPA8260	1634-04-4	04/29/11
Toluene	0.500 U	ug/L	1.00	0.500	1.00	10173329	EPA8260	108-88-3	04/29/11
Xylenes	2.81	ug/L	1.00	1.00	2.00	10173329	EPA8260	1330-20-7	04/29/11
Surr:1,2-Dichloroethane-d4 (77-129%)	115.67%		1.00	0.0100	0.0100	10173329	EPA8260		04/29/11
Surr:Bromofluorobenzene (82-122%)	104.43%		1.00	1.00	1.00	10173329	EPA8260	460-00-4	04/29/11
Surr:Toluene-d8 (88-111%)	88.93%		1.00	0.0100	0.0100	10173329	EPA8260		04/29/11
Base_Neutral_Extraction	1000	mL	1.00			10173376	XBNE		04/28/11
1-methyl-Naphthalene	60.7	ug/L	1.00	0.200	0.400	10173416	EPA8270	090-12-0	05/03/11
2-methyl-Naphthalene	40.3	ug/L	1.00	0.200	0.400	10173416	EPA8270	91-57-6	05/03/11
Acenaphthene	0.520	ug/L	1.00	0.200	0.400	10173416	EPA8270	83-32-9	05/03/11
Acenaphthylene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	208-96-8	05/03/11
Anthracene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	120-12-7	05/03/11
Benzo(a)anthracene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	56-55-3	05/03/11
Benzo(a)pyrene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	50-32-8	05/03/11
Benzo(b)fluoranthene	0.100 U	ug/L	1.00	0.100	0.200	10173416	EPA8270	205-99-2	05/03/11
Benzo(g,h,i)perylene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	191-24-2	05/03/11
Benzo(k)fluoranthene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	207-08-9	05/03/11
Chrysene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	218-01-9	05/03/11
Dibenz(a,h)anthracene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	53-70-3	05/03/11
Fluoranthene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	206-44-0	05/03/11
Fluorene	0.280 I	ug/L	1.00	0.200	0.400	10173416	EPA8270	86-73-7	05/03/11
Indeno(1,2,3-cd)pyrene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	193-39-5	05/03/11
Naphthalene	239	ug/L	1.00	0.200	0.400	10173416	EPA8270	91-20-3	05/03/11
Phenanthrene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	85-01-8	05/03/11
Pyrene	0.200 U	ug/L	1.00	0.200	0.400	10173416	EPA8270	129-00-0	05/03/11



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256


PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Lab #: 148798GW1 **Sampled:** 04/26/11 08:20 AM **Desc:** MW-10R

Parameter	Result	Units	DF	MDL	PQL	QC Batch	Method	CAS #	Analyzed
Surr:2-Fluorobiphenyl (18-110%)	80.38%		1.00	0.0100	0.0200	10173416	EPA8270	321-60-8	05/03/11
Surr:Nitrobenzene-d5 (54-138%)	83.26%		1.00	0.0100	0.400	10173416	EPA8270		05/03/11
Surr:Terphenyl-d14 (37-143%)	66.86%		1.00	0.0100	0.0200	10173416	EPA8270		05/03/11

Lab #: 148798SO1 **Sampled:** 04/26/11 12:00 PM **Desc:** IDW

Parameter	Result	Units	DF	MDL	PQL	QC Batch	Method	CAS #	Analyzed
Arsenic	1.53	mg/Kg	1.00	0.100	0.200	10173304	EPA6020	7440-38-2	04/28/11
Cadmium	0.100 U	mg/Kg	1.00	0.100	0.200	10173304	EPA6020	7440-43-9	04/28/11
Chromium	4.55	mg/Kg	1.00	0.100	0.200	10173304	EPA6020	7440-47-3	04/28/11
Lead	1.18	mg/Kg	1.00	0.100	0.200	10173304	EPA6020	7439-92-1	04/28/11
Benzene	3.00 U	ug/Kg	1.00	3.00	12.0	10173572	EPA8260	71-43-2	05/04/11
Ethylbenzene	3.00 U	ug/Kg	1.00	3.00	12.0	10173572	EPA8260	100-41-4	05/04/11
Methyl-tert-butylether	3.00 U	ug/Kg	1.00	3.00	12.0	10173572	EPA8260	1634-04-4	05/04/11
Toluene	3.00 U	ug/Kg	1.00	3.00	6.00	10173572	EPA8260	108-88-3	05/04/11
Xylenes	5.00 U	ug/Kg	1.00	5.00	20.0	10173572	EPA8260	1330-20-7	05/04/11
Surr:1,2-Dichloroethane-d4 (80-124%)	93.07%		1.00	0.0100	0.0200	10173572	EPA8260		05/04/11
Surr:Bromofluorobenzene (77-128%)	103.33%		1.00	0.0100	0.0200	10173572	EPA8260	460-00-4	05/04/11
Surr:Toluene-d8 (89-109%)	96.83%		1.00	0.0100	0.0200	10173572	EPA8260		05/04/11



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Report

Quality Control Batch: 10173304

Analyst: EVB

Blank	Result	Units
Arsenic	0.100U	mg/Kg
Cadmium	0.100U	mg/Kg
Chromium	0.100U	mg/Kg
Lead	0.100U	mg/Kg

Laboratory Control Sample

	Result	Units	Spike	%REC	%REC Lim
Arsenic	0.0915	mg/Kg	0.100	91.46	77.37-119.08
Cadmium	0.0956	mg/Kg	0.100	95.55	76.23-122.28
Chromium	0.0965	mg/Kg	0.100	96.54	79.67-125.91
Lead	0.106	mg/Kg	0.100	106.00	81.86-122.61

Matrix Spike

	Result	Units	Spike	%REC	%REC Lim	Sample
Arsenic	0.187	mg/Kg	0.150	123.65	46.94-143.33	0.00178
Cadmium	0.167	mg/Kg	0.150	111.41	56.41-141.39	0.100U
Chromium	0.155	mg/Kg	0.150	98.28	57.26-148.76	0.00720
Lead	0.160	mg/Kg	0.150	102.32	50.31-149.56	0.00687


Matrix Spike Duplicate

	Result	Units	Spike	%REC	%REC Lim	Sample	RPD	RPD Lim
Arsenic	0.178	mg/Kg	0.150	117.37	46.94-143.33	0.00178	5.17	25.63
Cadmium	0.155	mg/Kg	0.150	103.11	56.41-141.39	0.100U	7.74	23.21
Chromium	0.148	mg/Kg	0.150	93.63	57.26-148.76	0.00720	4.62	28.86
Lead	0.159	mg/Kg	0.150	101.50	50.31-149.56	0.00687	0.77	27.40

Quality Control Batch: 10173329

Analyst: CLS

Blank	Result	Units
Benzene	0.500U	ug/L
Ethylbenzene	0.500U	ug/L



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173329**Analyst: CLS**

Blank	Result	Units
Methyl-tert-butylether	1.00U	ug/L
Toluene	0.500U	ug/L
Xylenes	1.00U	ug/L
Surr:Bromofluorobenzene	29.0	ug/L
Surr:Toluene-d8	29.1	ug/L
Surr:1,2-Dichloroethane-d4	28.1	ug/L

Laboratory Control Sample


	Result	Units	Spike	%REC	%REC Lim
Benzene	10.4	ug/L	10.0	104.10	69.86-134.08
Ethylbenzene	10.1	ug/L	10.0	101.10	70.56-124.95
Methyl-tert-butylether	10.7	ug/L	10.0	107.10	58.57-146.52
Toluene	10.5	ug/L	10.0	105.20	75.34-126.10
Xylenes	32.0	ug/L	30.0	106.63	70.92-123.99
Surr:Bromofluorobenzene	29.0	ug/L	30.0	96.77	82.78-122.32
Surr:Toluene-d8	30.0	ug/L	30.0	100.03	88.96-111.11
Surr:1,2-Dichloroethane-d4	28.7	ug/L	30.0	95.50	77.68-129.99

Matrix Spike

	Result	Units	Spike	%REC	%REC Lim	Sample
Benzene	20.0	ug/L	20.0	100.10	74.48-134.24	0.500U
Ethylbenzene	306	ug/L	20.0	16.45	73.91-128.64	303
Methyl-tert-butylether	19.1	ug/L	20.0	95.40	66.43-141.40	1.00U
Toluene	159	ug/L	20.0	93.70	77.67-128.73	141
Xylenes	981	ug/L	60.0	33.57	74.03-127.32	960
Surr:Bromofluorobenzene	29.5	ug/L	30.0	98.27	53.82-151.16	
Surr:Toluene-d8	28.7	ug/L	30.0	95.77	88.05-109.81	
Surr:1,2-Dichloroethane-d4	30.5	ug/L	30.0	101.53	75.07-126.19	

Matrix Spike Duplicate

	Result	Units	Spike	%REC	%REC Lim	Sample	RPD	RPD Lim
Benzene	24.3	ug/L	20.0	121.35	74.48-134.24	0.500U	19.19	18.98



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256


PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173329**Analyst: CLS****Matrix Spike Duplicate**

	Result	Units	Spike	%REC	%REC Lim	Sample	RPD	RPD Lim
Ethylbenzene	296	ug/L	20.0	-34.35	73.91-128.64	303	3.38	17.78
Methyl-tert-butylether	24.9	ug/L	20.0	124.60	66.43-141.40	1.00U	26.55	23.14
Toluene	155	ug/L	20.0	73.15	77.67-128.73	141	2.61	18.11
Xylenes	946	ug/L	60.0	-23.42	74.03-127.32	960	3.55	18.64
Surr:Bromofluorobenzene	29.2	ug/L	30.0	97.43	53.82-151.16		0.85	13.51
Surr:Toluene-d8	28.7	ug/L	30.0	95.80	88.05-109.81		0.03	7.44
Surr:1,2-Dichloroethane-d4	29.3	ug/L	30.0	97.50	75.07-126.19		4.05	8.75

Quality Control Batch: 10173416**Analyst: CLS****Blank**

	Result	Units
Acenaphthene	0.200U	ug/L
Acenaphthylene	0.200U	ug/L
Anthracene	0.200U	ug/L
Benzo(a)anthracene	0.200U	ug/L
Benzo(a)pyrene	0.200U	ug/L
Benzo(b)fluoranthene	0.100U	ug/L
Benzo(g,h,i)perylene	0.200U	ug/L
Benzo(k)fluoranthene	0.200U	ug/L
Chrysene	0.200U	ug/L
Dibenz(a,h)anthracene	0.200U	ug/L
Fluoranthene	0.200U	ug/L
Fluorene	0.200U	ug/L
Indeno(1,2,3-cd)pyrene	0.200U	ug/L
Naphthalene	0.200U	ug/L
Phenanthrene	0.200U	ug/L
Pyrene	0.200U	ug/L
1-methyl-Naphthalene	0.200U	ug/L
2-methyl-Naphthalene	0.200U	ug/L



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173416

Analyst: CLS

Blank


Blank	Result	Units
Surr:Nitrobenzene-d5	52.0	ug/L
Surr:Terphenyl-d14	48.3	ug/L
Surr:2-Fluorobiphenyl	44.6	ug/L

Laboratory Control Sample

Laboratory Control Sample	Result	Units	Spike	%REC	%REC Lim
Acenaphthene	43.0	ug/L	50.0	85.92	19.43-110.82
Acenaphthylene	36.0	ug/L	50.0	72.02	70.00-130.00
Anthracene	39.1	ug/L	50.0	78.18	35.89-115.69
Benzo(a)anthracene	46.5	ug/L	50.0	93.00	41.93-129.99
Benzo(a)pyrene	39.6	ug/L	50.0	79.24	39.40-119.98
Benzo(b)fluoranthene	39.8	ug/L	50.0	79.64	31.36-135.33
Benzo(g,h,i)perylene	47.0	ug/L	50.0	93.90	46.39-130.38
Benzo(k)fluoranthene	41.0	ug/L	50.0	82.06	33.95-145.70
Chrysene	42.0	ug/L	50.0	84.04	46.72-133.45
Dibenz(a,h)anthracene	46.2	ug/L	50.0	92.40	37.64-133.88
Fluoranthene	52.0	ug/L	50.0	104.02	24.34-134.54
Fluorene	46.1	ug/L	50.0	92.26	19.06-111.20
Indeno(1,2,3-cd)pyrene	44.8	ug/L	50.0	89.50	37.85-131.90
Naphthalene	38.0	ug/L	50.0	75.92	22.76-100.00
Phenanthrene	47.6	ug/L	50.0	95.28	33.33-113.56
Pyrene	44.3	ug/L	50.0	88.68	42.73-127.70
1-methyl-Naphthalene	38.7	ug/L	50.0	77.36	27.35-113.22
2-methyl-Naphthalene	35.6	ug/L	50.0	71.18	19.08-101.21
Surr:Nitrobenzene-d5	49.0	ug/L	50.0	98.06	54.20-138.20
Surr:Terphenyl-d14	45.9	ug/L	50.0	91.86	37.97-143.50
Surr:2-Fluorobiphenyl	45.7	ug/L	50.0	91.36	18.75-110.25

Matrix Spike

Matrix Spike	Result	Units	Spike	%REC	%REC Lim	Sample
Acenaphthene	36.7	ug/L	50.0	72.86	24.01-102.00	0.260



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256


PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173416

Analyst: CLS

Matrix Spike	Result	Units	Spike	%REC	%REC Lim	Sample
Acenaphthylene	33.8	ug/L	50.0	67.50	70.00-130.00	0.200U
Anthracene	37.3	ug/L	50.0	74.64	36.36-119.37	0.200U
Benzo(a)anthracene	43.4	ug/L	50.0	86.86	22.93-131.37	0.200U
Benzo(a)pyrene	35.7	ug/L	50.0	71.44	28.55-113.30	0.200U
Benzo(b)fluoranthene	33.9	ug/L	50.0	67.80	33.90-132.20	0.100U
Benzo(g,h,i)perylene	43.5	ug/L	50.0	86.96	29.74-126.60	0.200U
Benzo(k)fluoranthene	42.6	ug/L	50.0	85.28	1.00-172.28	0.200U
Chrysene	43.3	ug/L	50.0	86.52	21.81-143.20	0.200U
Dibenz(a,h)anthracene	42.4	ug/L	50.0	84.88	14.09-128.46	0.200U
Fluoranthene	48.0	ug/L	50.0	96.06	20.92-125.32	0.200U
Fluorene	42.7	ug/L	50.0	85.34	15.17-114.79	0.200U
Indeno(1,2,3-cd)pyrene	41.0	ug/L	50.0	82.00	13.90-145.82	0.200U
Naphthalene	146	ug/L	50.0	54.16	9.17-110.12	119
Phenanthrene	48.4	ug/L	50.0	96.72	19.47-133.87	0.200U
Pyrene	45.3	ug/L	50.0	90.64	21.53-149.12	0.200U
1-methyl-Naphthalene	57.0	ug/L	50.0	53.34	23.63-87.85	30.4
2-methyl-Naphthalene	49.2	ug/L	50.0	58.16	23.63-87.85	20.1
Surr:Nitrobenzene-d5	46.3	ug/L	50.0	92.60	48.30-142.10	
Surr:Terphenyl-d14	42.5	ug/L	50.0	84.92	16.08-149.12	
Surr:2-Fluorobiphenyl	42.1	ug/L	50.0	84.10	16.08-95.65	

Matrix Spike Duplicate	Result	Units	Spike	%REC	%REC Lim	Sample	RPD	RPD Lim
Acenaphthene	36.6	ug/L	50.0	72.72	24.01-102.00	0.260	0.19	20.00
Acenaphthylene	33.3	ug/L	50.0	66.58	70.00-130.00	0.200U	1.37	20.00
Anthracene	40.2	ug/L	50.0	80.32	36.36-119.37	0.200U	7.33	20.00
Benzo(a)anthracene	45.8	ug/L	50.0	91.58	22.93-131.37	0.200U	5.29	20.00
Benzo(a)pyrene	36.4	ug/L	50.0	72.72	28.55-113.30	0.200U	1.78	20.00
Benzo(b)fluoranthene	35.9	ug/L	50.0	71.70	33.90-132.20	0.100U	5.59	20.00
Benzo(g,h,i)perylene	42.6	ug/L	50.0	85.20	29.74-126.60	0.200U	2.04	20.00



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598


E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173416		Analyst: CLS						
Matrix Spike Duplicate	Result	Units	Spike	%REC	%REC Lim	Sample	RPD	RPD Lim
Benzo(k)fluoranthene	43.2	ug/L	50.0	86.36	1.00-172.28	0.200U	1.26	20.00
Chrysene	42.4	ug/L	50.0	84.84	21.81-143.20	0.200U	1.96	20.00
Dibenz(a,h)anthracene	42.2	ug/L	50.0	84.44	14.09-128.46	0.200U	0.52	20.00
Fluoranthene	47.0	ug/L	50.0	93.98	20.92-125.32	0.200U	2.19	20.00
Fluorene	41.0	ug/L	50.0	82.06	15.17-114.79	0.200U	3.92	20.00
Indeno(1,2,3-cd)pyrene	40.5	ug/L	50.0	80.98	13.90-145.82	0.200U	1.25	20.00
Naphthalene	149	ug/L	50.0	58.74	9.17-110.12	119	1.55	20.00
Phenanthrene	48.7	ug/L	50.0	97.30	19.47-133.87	0.200U	0.60	20.00
Pyrene	49.1	ug/L	50.0	98.24	21.53-149.12	0.200U	8.05	20.00
1-methyl-Naphthalene	59.2	ug/L	50.0	57.62	23.63-87.85	30.4	3.68	20.00
2-methyl-Naphthalene	49.5	ug/L	50.0	58.70	23.63-87.85	20.1	0.55	20.00
Surr:Nitrobenzene-d5	46.7	ug/L	50.0	93.30	48.30-142.10		0.75	20.00
Surr:Terphenyl-d14	46.2	ug/L	50.0	92.36	16.08-149.12		8.39	20.00
Surr:2-Fluorobiphenyl	40.1	ug/L	50.0	80.14	16.08-95.65		4.82	20.00

Quality Control Batch: 10173572		Analyst: CLS						
Blank	Result	Units						
Benzene	3.00U	ug/Kg						
Ethylbenzene	3.00U	ug/Kg						
Methyl-tert-butylether	3.00U	ug/Kg						
Toluene	3.00U	ug/Kg						
Xylenes	5.00U	ug/Kg						
Surr:Bromofluorobenzene	28.3	ug/Kg						
Surr:Toluene-d8	31.0	ug/Kg						
Surr:1,2-Dichloroethane-d4	30.2	ug/Kg						
Laboratory Control Sample	Result	Units	Spike	%REC	%REC Lim			
Benzene	19.6	ug/Kg	20.0	98.05	69.94-134.60			
Ethylbenzene	20.1	ug/Kg	20.0	100.50	72.02-125.51			



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256


PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Quality Control Batch: 10173572

Analyst: CLS

Laboratory Control Sample

	Result	Units	Spike	%REC	%REC Lim
Methyl-tert-butylether	18.7	ug/Kg	20.0	93.40	49.79-151.70
Toluene	19.8	ug/Kg	20.0	98.85	73.77-128.25
Xylenes	60.6	ug/Kg	60.0	101.02	71.19-127.02
Surr:Bromofluorobenzene	30.2	ug/Kg	30.0	100.60	77.55-128.54
Surr:Toluene-d8	29.4	ug/Kg	30.0	98.00	89.44-109.59
Surr:1,2-Dichloroethane-d4	30.9	ug/Kg	30.0	103.13	80.22-124.14



FLOWERS CHEMICAL LABORATORIES INC.

P.O. Box 150597, Altamonte Springs FL 32715-0597
571 NW Mercantile Pl. Suite 111, Port St. Lucie FL 34986
P.O. Box 1200, Madison FL 32341
3980 Overseas Hwy Suite 103, Marathon FL 33050

Phone: 407-339-5984 / Fax: 407-260-6110
Phone: 772-343-8006 / Fax: 772-343-8089
Phone: 850-973-6878 / Fax: 850-973-6878
Phone: 305-743-8598 / Fax: 305-743-8598

E83018 (Main Lab)
E86562 (South Lab)
E82405 (North Lab)
E35834 (Keys Lab)

Golder-Jacksonville
9428 Baymeadows Rd, Suite 400
Jacksonville, FL 32256

PO #: 053-3906-B Revised 5/20/11
Client Project #: Majik Market-Kirk Blevins
Date Sampled: Apr 26, 2011
May 5, 2011; Invoice: 148798

Narrative Report

Sample Handling

Sample handling and holding time criteria were met for all samples. Samples collected by submitter. No unusual events occurred during analysis. Results are reported on a wet weight basis for aqueous matrices and on a dry weight basis for sludge and soil matrices unless otherwise noted. This report was revised on 5/20/11 to report the metals soil results by dry weight. Percent moisture = 12.18.

Quality Control

Enclosed analyses met method or FCL criteria, unless otherwise denoted on the sample results. Applied data qualifiers are defined below.

Additional Comments

The matrix spike recovery outliers for EPA8260 and EPA8270 only affect the sample spiked. The spiked samples were not taken from this sample delivery group. There was not sufficient sample vials for soil BTEXM MS/MSD. The LCS validates the QC batch.

Attachments

Chain of Custody

Qualifier	Meaning
U	Compound was analyzed for but not detected.
J	One or more QC samples associated with this data value exceeded QC limits.
J1	Surrogate recovery limits have been exceeded.
J2	No known quality control criteria exist for the component.
J3	Reported value failed to meet established quality control criteria for either precision or accuracy.
J4	Sample matrix interfered with the ability to make an accurate determination on the spiked sample.
Q	Sample held beyond the accepted holding time.
L	Off-scale high; reported concentration exceeds the highest standard.
V	Analyte was detected in both the sample and the associated method blank.
ZTNTC	Too numerous to count. Numeric value represents filtration volume.
A	Absent
P	Present
T	Value reported is less than the statistical method detection limit. Reported for informational purposes only.
M	Value reported is greater than the statistical method detection limit, but less than the reported MDL.
G	The greatest of the dilutions performed did not yield sufficient oxygen depletion for valid data.
S	The least of the dilutions performed did not yield sufficient oxygen residual for valid data.
O	Result is greater than (over) the specified value.
I	Reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
B	Results based upon colony plate count outside ideal range.
Y	The laboratory analysis was from an improperly preserved sample. The data may not be accurate.