

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

--- Project Information ---

12/06/95

Ms. Rose Stibila

Broward Testing Laboratory, Inc.

1034 N.E. 44th Court

Ft. Lauderdale, FL 33334

Page 1

Project Name: PASCO4

Class III Tank-1 Project #1826

rcd 4/22/96
w/ Permit
info

--- Sample Information ---

Station ID: Tank I

Lab ID: AA81497

Collector: NOT LISTED

Date Sampled: 10/30/95

Time Sampled: 11:45

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
622-8260 App. IX VOA Water ug/L	ug/L	SW8260		TITLE	11/02/95	GH
Acetone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Acetonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrolein	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrylonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Allyl chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Benzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromodichloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromoform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromomethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
2-Butanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Carbon disulfide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Carbon tetrachloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chlorobenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloroethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroprene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dibromo-3-chloropropane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,2-Dibromoethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromomethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,4-Dichloro-2-butene	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Dichlorodifluoromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,1-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromochloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethylbenzene	ug/L	SW8260	5.00	9.89	11/02/95	GH
Ethyl methacrylate	ug/L	SW8260	5.00	Not Det	11/02/95	GH
2-Hexanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 2

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Isobutyl alcohol	ug/L	SW8260	50.0	Not Det	11/02/95	GH
Methacrylonitrile	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methylene chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl iodide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl methacrylate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
4-Methyl-2-pentanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Pentachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Propionitrile	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Styrene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Tetrachloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Toluene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichlorofluoromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2,3-Trichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Vinyl acetate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Vinyl chloride	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Xylenes (total)	ug/L	SW8260	5.00	5.17	11/02/95	GH
sur-1,2-Dichloroethane-d4 %R 76-114	ug/L	SW8260	0	101	11/02/95	GH
sur-Toluene-d8 %R 88-110	ug/L	SW8260	0	98	11/02/95	GH
sur-Bromofluorobenzene 86-115	ug/L	SW8260	0	101	11/02/95	GH
624-300.0 Anions Water mg/L	mg/L	EPA300.0		_TITLE_	11/01/95	CBC
Fluoride	mg/L	EPA300.0	0.200	1.44	11/01/95	CBC
610-3010 ICP Digest Water		SW3010		Done	11/04/95	JDE
610-6010 App.IX Met. Water ug/L	ug/L	SW6010		_TITLE_	11/06/95	CLH
Antimony	ug/L	SW6010	25.0	Not Det	11/06/95	CLH
Arsenic	ug/L	SW6010	25.0	125	11/06/95	CLH
Barium	ug/L	SW6010	20.0	98.0	11/06/95	CLH
Beryllium	ug/L	SW6010	3.00	Not Det	11/06/95	CLH
Cadmium	ug/L	SW6010	5.00	Not Det	11/06/95	CLH
Chromium	ug/L	SW6010	10.0	43.0	11/06/95	CLH

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 3

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Cobalt	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
Copper	ug/L	SW6010	5.00	41.0	11/06/95	CLH
Lead	ug/L	SW6010	20.0	Not Det	11/06/95	CLH
Nickel	ug/L	SW6010	20.0	Not Det	11/06/95	CLH
Selenium	ug/L	SW6010	100	Not Det	11/06/95	CLH
Silver	ug/L	SW6010	15.0	Not Det	11/06/95	CLH
Thallium	ug/L	SW6010	50.0	Not Det	11/06/95	CLH
Vanadium	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
Zinc	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
610-7470 Mercury Water ug/L	ug/L	SW7470	0.200	Not Det	11/02/95	CDT
623-3520 Cont. Liq/Liq Ext.		SW3520		Done	11/01/95	LR
622-8270 App.IX S-VOA Water ug/L	ug/L	SW8270		<u>TITLE</u>	11/21/95	KRG
Acenaphthene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Acenaphthylene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Acetophenone	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Acetylaminofluorene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4-Aminobiphenyl	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
Aniline	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Anthracene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Aramite	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benz(a)anthracene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benzo(b)fluoranthene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benzo(k)fluoranthene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benzo(ghi)perylene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benzo(a)pyrene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Benzyl alcohol	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
bis(2-Chloroethoxy)methane	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
bis(2-Chloroethyl)ether	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
bis(2-Chloroisopropyl)ether	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
bis(2-Ethylhexyl)phthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4-Bromophenyl phenyl ether	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Butyl benzyl phthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
p-Chloroaniline	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
Chlorobenzilate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG

LAW ENVIROMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Page 4

Ms. Rose Stibila

Project Name: PASCO4

Broward Testing Laboratory, Inc.

Class III Tank-1 Project #1826

1034 N.E. 44th Court

Ft. Lauderdale, FL 33334

--- Sample Information ---

Date Sampled: 10/30/95

Station ID: Tank I

Time Sampled: 11:45

Lab ID: AA81497

Log In Date: 11/01/95

Collector: NOT LISTED

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
4-Chloro-3-methylphenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Chloronaphthalene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Chlorophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4-Chlorophenyl phenyl ether	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Chrysene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Diallate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Dibenz(a,h)anthracene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Dibenzofuran	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Di-n-butyl phthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,2-Dichlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,3-Dichlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,4-Dichlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
3,3'-Dichlorobenzidine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2,4-Dichlorophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2,6-Dichlorophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Diethyl phthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,4-Dioxane	ug/L	SW8270	100	Not Det	11/21/95	KRG
p-(Dimethylamino)azobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
7,12-Dimethylbenz[a]anthracene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
3,3'-Dimethylbenzidine	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
alpha,alpha-Dimethylphenethylamine	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
2,4-Dimethylphenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Dimethyl phthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,3-Dinitrobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4,6-Dinitro-2-methylphenol	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
2,4-Dinitrophenol	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
2,4-Dinitrotoluene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2,6-Dinitrotoluene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Di-n-octylphthalate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Diphenylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Ethyl methanesulfonate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Fluoranthene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Fluorene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Hexachlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG

LAW ENVIRO MENTAL NATIONAL LABC TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 5

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Hexachlorobutadiene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Hexachlorocyclopentadiene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Hexachloroethane	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Hexachlorophene	ug/L	SW8270	100	Not Det	11/21/95	KRG
Hexachloropropene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Indeno(1,2,3-cd)pyrene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Isodrin	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Isophorone	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Isosafrole (total)	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Kepone	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
Methapyrilene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
3-Methylcholanthrene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Methyl methanesulfonate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Methylnaphthalene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Methylphenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
3-Methylphenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4-Methylphenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Naphthalene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,4-Naphthoquinone	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1-Naphthylamine	ug/L	SW8270	100	Not Det	11/21/95	KRG
2-Naphthylamine	ug/L	SW8270	100	Not Det	11/21/95	KRG
2-Nitroaniline	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
3-Nitroaniline	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
4-Nitroaniline	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
Nitrobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2-Nitrophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
4-Nitrophenol	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
4-Nitroquinoline 1-oxide	ug/L	SW8270	50.0	Not Det	11/21/95	KRG
n-Nitrosodi-n-butylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosodiethylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosodimethylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosodiphenylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosodi-n-propylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosomethylethylamine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 6

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
n-Nitrosomorpholine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosopiperidine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
n-Nitrosopyrrolidine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
5-Nitro-o-toluidine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pentachlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pentachloronitrobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pentachlorophenol	ug/L	SW8270	30.0	Not Det	11/21/95	KRG
Phenacetin	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Phenanthrene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Phenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
p-Phenylenediamine	ug/L	SW8270	100	Not Det	11/21/95	KRG
2-Picoline	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pronamide	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pyrene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
Pyridine	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
Safrole	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,2,4,5-Tetrachlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2,3,4,6-Tetrachlorophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
o-Toluidine	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,2,4-Trichlorobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
2,4,5-Trichlorophenol	ug/L	SW8270	20.0	Not Det	11/21/95	KRG
2,4,6-Trichlorophenol	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
0,0,0-Triethyl phosphorothioate	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
1,3,5-Trinitrobenzene	ug/L	SW8270	10.0	Not Det	11/21/95	KRG
sur-2-Fluorophenol %R 21-100	ug/L	SW8270	0	71	11/21/95	KRG
sur-Phenol-d6 %R 10-94	ug/L	SW8270	0	76	11/21/95	KRG
sur-Nitrobenzene-d5 %R 35-114	ug/L	SW8270	0	85	11/21/95	KRG
sur-2-Fluorobiphenyl %R 43-116	ug/L	SW8270	0	74	11/21/95	KRG
sur-2,4,6-Tribromophenol %R 10-123	ug/L	SW8270	0	91	11/21/95	KRG
sur-Terphenyl-d14 %R 38-141	ug/L	SW8270	0	126	11/21/95	KRG
623-3520 Cont. Liq/Liq Ext. Pest		SW3520		Done	11/01/95	SC
621-8080 App. IX P/P Water ug/L	ug/L	SW8080		TITLE	12/03/95	AMP
Aldrin	ug/L	SW8080	0.420	Not Det	12/03/95	AMP
alpha-BHC	ug/L	SW8080	0.315	Not Det	12/03/95	AMP

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 7

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
beta-BHC	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
delta-BHC	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
gamma-BHC	ug/L	SW8080	0.420	Not Det	12/03/95	AMP
alpha-Chlordane	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
gamma-Chlordane	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
4,4'-DDD	ug/L	SW8080	1.05	Not Det	12/03/95	AMP
4,4'-DDE	ug/L	SW8080	0.420	Not Det	12/03/95	AMP
4,4'-DDT	ug/L	SW8080	1.05	Not Det	12/03/95	AMP
Dieldrin	ug/L	SW8080	0.210	Not Det	12/03/95	AMP
Endosulfan I	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
Endosulfan II	ug/L	SW8080	0.420	Not Det	12/03/95	AMP
Endosulfan sulfate	ug/L	SW8080	1.05	Not Det	12/03/95	AMP
Endrin	ug/L	SW8080	0.630	Not Det	12/03/95	AMP
Endrin aldehyde	ug/L	SW8080	1.05	Not Det	12/03/95	AMP
Heptachlor	ug/L	SW8080	0.315	Not Det	12/03/95	AMP
Heptachlor epoxide	ug/L	SW8080	0.525	Not Det	12/03/95	AMP
Methoxychlor	ug/L	SW8080	5.25	Not Det	12/03/95	AMP
Toxaphene	ug/L	SW8080	21.0	Not Det	12/03/95	AMP
Aroclor-1016	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
Aroclor-1221	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
Aroclor-1232	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
Aroclor-1242	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
Aroclor-1248	ug/L	SW8080	10.5	Not Det	12/03/95	AMP

LAW ENVIRO. NENTAL NATIONAL LABC. TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 8

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Class III Tank-1 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Tank I

Date Sampled: 10/30/95

Lab ID: AA81497

Time Sampled: 11:45

Collector: NOT LISTED

Log In Date: 11/01/95

Log In Time: 08:45


--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Aroclor-1254	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
Aroclor-1260	ug/L	SW8080	10.5	Not Det	12/03/95	AMP
sur-Dibutylchloredate %R 33-186	ug/L	SW8080	0	58	12/03/95	AMP
623-8150 Herbicides Ext W		SW8150		Done	11/03/95	DT
621-8150 App.IX Herb. Water ug/L	ug/L	SW8150		_TITLE_	11/29/95	GET
2,4-D	ug/L	SW8150	8.40	Not Det	11/29/95	GET
Dinoseb	ug/L	SW8150	4.20	Not Det	11/29/95	GET
2,4,5-T	ug/L	SW8150	0.840	Not Det	11/29/95	GET
2,4,5-TP (Silvex)	ug/L	SW8150	0.840	Not Det	11/29/95	GET
sur-DCAA %R 28-124	ug/L	SW8150	0	37	11/29/95	GET

Remarks:

Signed: _____



James M.G. Fucci
Laboratory Manager

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Page 1

Ms. Rose Stibila

Project Name: PASCO4

Broward Testing Laboratory, Inc.

Client #32660 Project #1826

1034 N.E. 44th Court

Ft. Lauderdale, FL 33334

--- Sample Information ---

Date Sampled:

Station ID: Trip Blank 07551

Time Sampled:

Lab ID: AA81498

Log In Date: 11/01/95

Collector:

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
622-8260 App.IX VOA Water ug/L	ug/L	SW8260		_TITLE_	11/02/95	GH
Acetone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Acetonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrolein	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrylonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Allyl chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Benzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromodichloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromoform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromomethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
2-Butanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Carbon disulfide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Carbon tetrachloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chlorobenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloroethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroprene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dibromo-3-chloropropane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,2-Dibromoethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromomethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,4-Dichloro-2-butene	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Dichlorodifluoromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,1-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromochloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethylbenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethyl methacrylate	ug/L	SW8260	5.00	Not Det	11/02/95	GH
2-Hexanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH

LAW ENVIRO NMENTAL NATIONAL LABC TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 2

Broward Testing Laboratory, Inc.

Project Name: PASCO4

1034 N.E. 44th Court

Client #32660 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Trip Blank 07551

Date Sampled:

Lab ID: AA81498

Time Sampled:

Collector:

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Isobutyl alcohol	ug/L	SW8260	50.0	Not Det	11/02/95	GH
Methacrylonitrile	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methylene chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl iodide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl methacrylate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
4-Methyl-2-pentanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Pentachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Propionitrile	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Styrene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Tetrachloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Toluene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichlorofluoromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2,3-Trichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Vinyl acetate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Vinyl chloride	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Xylenes (total)	ug/L	SW8260	5.00	Not Det	11/02/95	GH
sur-1,2-Dichloroethane-d4 %R 76-114	ug/L	SW8260	0	107	11/02/95	GH
sur-Toluene-d8 %R 88-110	ug/L	SW8260	0	97	11/02/95	GH

LAW ENVIRO. IMENTAL NATIONAL LABO .TORIES
TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Broward Testing Laboratory, Inc.

1034 N.E. 44th Court

Ft. Lauderdale, FL 33334

Page 3

Project Name: PASCO4

Client #32660 Project #1826

--- Sample Information ---

Station ID: Trip Blank 07551

Lab ID: AA81498

Collector:

Date Sampled:

Time Sampled:

Log In Date: 11/01/95

Log In Time: 08:45

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
sur-Bromofluorobenzene 86-115	ug/L	SW8260	0	99	11/02/95	GH

Remarks:

Signed: _____

James M.G. Tucci

Laboratory Manager



Lancaster Laboratories
Where quality is a science.

Page: 1 of 3

LLI Sample No. WW 2406712
Collected: 10/24/95 at 11:15 by CC

Submitted: 10/27/95 Reported: 11/ 8/95
Discard: 11/16/95

AA81171. Grab Water Sample

Class II Tank 2

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

P.O. CLASS II - TANK 2
Rel.

CAT NO.	ANALYSIS NAME	AS RECEIVED		
		RESULTS	LIMIT OF QUANTITATION	UNITS
1320	Appendix IX Organophosphates			See Page 2
9001	Dimethoate	< 1.	1.	ug/l
9002	Thionazin	< 0.1	0.1	ug/l

1 COPY TO Law Engineering & Env Svcs Inc ATTN: Ms. Kelli Silvia

Questions? Contact your Client Services Representative
F. Bradley Ayars at (717) 656-2300
05:26:33 D 0001 1 489822
055 0.00 00024000 ASR000

Respectfully Submitted
Jenifer E. Hess, B.S.
Group Leader Pesticides/PCBs



Lancaster Laboratories
2425 New Holland Pike
PO. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations.



#22



Lancaster Laboratories
Where quality is a science.

Page: 2 of 3

LLI Sample No. WW 2406712

Collected: 10/24/95 at 11:15 by CC

Submitted: 10/27/95 Reported: 11/ 8/95

Discard: 11/16/95

AA81171 Grab Water Sample

Class II Tank 2

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

P.O. CLASS II - TANK 2
Rel.

CAT NO.	ANALYSIS NAME	AS RECEIVED		
		RESULTS	LIMIT OF QUANTITATION	UNITS

Appendix IX Organophosphates

1317	Disulfoton	< 0.3	0.3	ug/l
3027	Methyl Parathion	< 0.2	0.2	ug/l
3028	Ethyl Parathion	< 0.2	0.2	ug/l
1318	Famphur	< 0.5	0.5	ug/l
1319	Phorate	< 0.3	0.3	ug/l

Questions? Contact your Client Services Representative
F. Bradley Ayars at (717) 656-2300

Respectfully Submitted
Jenifer E. Hess, B.S.
Group Leader Pesticides/PCBs



Lancaster Laboratories
2425 New Holland Pike
PO. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations



#221

**Lancaster Laboratories***Where quality is a science.*

LABORATORY CHRONICLE

Page: 3 of 3

LLI Sample No. WW 2406712

Collected: 10/24/95 at 11:15 by CC

Submitted: 10/27/95

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

AAB1171 Grab Water Sample

Class II Tank 2
SDG#:

CAT NO	ANALYSIS NAME	METHOD	TRIAL	DATE OF ANALYSIS	ANALYST
0817	Water Sample Pest. Extraction	SW-846 3510B	1	10/30/95	DTQ
1320	Appendix IX Organophosphates	SW-846 8141A	1	11/04/95	NS
9001	Dimethoate	SW-846 8141A		11/04/95	NS
9002	Thionazin	SW-846 8141A		11/04/95	NS



Lancaster Laboratories
2425 New Holland Pike
PO. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations



LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 1

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
622-8260 App. IX VOA Water ug/L	ug/L	SW8260		<u> TITLE </u>	11/02/95	GH
Acetone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Acetonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrolein	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrylonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Allyl chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Benzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromodichloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromoform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromomethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
2-Butanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Carbon disulfide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Carbon tetrachloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chlorobenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloroethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroprene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dibromo-3-chloropropane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,2-Dibromoethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromomethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,4-Dichloro-2-butene	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Dichlorodifluoromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,1-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromochloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethylbenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethyl methacrylate	ug/L	SW8260	5.00	Not Det	11/02/95	GH
2-Hexanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH

rc'd 1/22/96
w/ Permit
info

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 2

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Isobutyl alcohol	ug/L	SW8260	50.0	Not Det	11/02/95	GH
Methacrylonitrile	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methylene chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl iodide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl methacrylate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
4-Methyl-2-pentanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Pentachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Propionitrile	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Styrene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Tetrachloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Toluene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichlorofluoromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2,3-Trichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Vinyl acetate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Vinyl chloride	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Xylenes (total)	ug/L	SW8260	5.00	Not Det	11/02/95	GH
sur-1,2-Dichloroethane-d4 %R 76-114	ug/L	SW8260	0	106	11/02/95	GH
sur-Toluene-d8 %R 88-110	ug/L	SW8260	0	99	11/02/95	GH
sur-Bromofluorobenzene 86-115	ug/L	SW8260	0	100	11/02/95	GH
624-300.0 Anions Water mg/L	mg/L	EPA300.0		_TITLE_	11/01/95	CBC
Fluoride	mg/L	EPA300.0	0.200	Not Det	11/01/95	CBC
610-3010 ICP Digest Water		SW3010		Done	11/04/95	JDE
610-6010 App.IX Met. Water ug/L	ug/L	SW6010		_TITLE_	11/06/95	CLH
Antimony	ug/L	SW6010	25.0	Not Det	11/06/95	CLH
Arsenic	ug/L	SW6010	25.0	Not Det	11/06/95	CLH
Barium	ug/L	SW6010	20.0	Not Det	11/06/95	CLH
Beryllium	ug/L	SW6010	3.00	Not Det	11/06/95	CLH
Cadmium	ug/L	SW6010	5.00	Not Det	11/06/95	CLH
Chromium	ug/L	SW6010	10.0	Not Det	11/06/95	CLH

LAW ENVIRO. JENTIAL NATIONAL LABO TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 3

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Cobalt	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
Copper	ug/L	SW6010	5.00	11.0	11/06/95	CLH
Lead	ug/L	SW6010	20.0	Not Det	11/06/95	CLH
Nickel	ug/L	SW6010	20.0	Not Det	11/06/95	CLH
Selenium	ug/L	SW6010	100	Not Det	11/06/95	CLH
Silver	ug/L	SW6010	15.0	Not Det	11/06/95	CLH
Thallium	ug/L	SW6010	50.0	Not Det	11/06/95	CLH
Vanadium	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
Zinc	ug/L	SW6010	10.0	Not Det	11/06/95	CLH
610-7470 Mercury Water ug/L	ug/L	SW7470	0.200	Not Det	11/01/95	CDT
623-3520 Cont. Liq/Liq Ext.		SW3520		Done	10/28/95	BT
622-8270 App.IX S-VOA Water ug/L	ug/L	SW8270		_TITLE_	11/10/95	DH
Acenaphthene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Acenaphthylene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Acetophenone	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Acetylaminofluorene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4-Aminobiphenyl	ug/L	SW8270	20.6	Not Det	11/10/95	DH
Aniline	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Anthracene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Aramite	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benz(a)anthracene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benzo(b)fluoranthene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benzo(k)fluoranthene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benzo(ghi)perylene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benzo(a)pyrene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Benzyl alcohol	ug/L	SW8270	20.6	Not Det	11/10/95	DH
bis(2-Chloroethoxy)methane	ug/L	SW8270	10.3	Not Det	11/10/95	DH
bis(2-Chloroethyl)ether	ug/L	SW8270	10.3	Not Det	11/10/95	DH
bis(2-Chloroisopropyl)ether	ug/L	SW8270	10.3	Not Det	11/10/95	DH
bis(2-Ethylhexyl)phthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4-Bromophenyl phenyl ether	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Butyl benzyl phthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
p-Chloroaniline	ug/L	SW8270	20.6	Not Det	11/10/95	DH
Chlorobenzilate	ug/L	SW8270	10.3	Not Det	11/10/95	DH

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 4

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
4-Chloro-3-methylphenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Chloronaphthalene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Chlorophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4-Chlorophenyl phenyl ether	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Chrysene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Diallate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Dibenz(a,h)anthracene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Dibenzofuran	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Di-n-butyl phthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,2-Dichlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,3-Dichlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,4-Dichlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
3,3'-Dichlorobenzidine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2,4-Dichlorophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2,6-Dichlorophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Diethyl phthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,4-Dioxane	ug/L	SW8270	10.3	Not Det	11/10/95	DH
p-(Dimethylamino)azobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
7,12-Dimethylbenz[a]anthracene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
3,3'-Dimethylbenzidine	ug/L	SW8270	51.6	Not Det	11/10/95	DH
alpha,alpha-Dimethylphenethylamine	ug/L	SW8270	51.6	Not Det	11/10/95	DH
2,4-Dimethylphenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Dimethyl phthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,3-Dinitrobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4,6-Dinitro-2-methylphenol	ug/L	SW8270	51.6	Not Det	11/10/95	DH
2,4-Dinitrophenol	ug/L	SW8270	51.6	Not Det	11/10/95	DH
2,4-Dinitrotoluene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2,6-Dinitrotoluene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Di-n-octylphthalate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Diphenylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Ethyl methanesulfonate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Fluoranthene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Fluorene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Hexachlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH

LAW ENVIROI ENTAL NATIONAL LABO TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 5

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Hexachlorobutadiene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Hexachlorocyclopentadiene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Hexachloroethane	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Hexachlorophene	ug/L	SW8270	103	Not Det	11/10/95	DH
Hexachloropropene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Indeno(1,2,3-cd)pyrene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Isodrin	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Isophorone	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Isosafrole (total)	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Kepone	ug/L	SW8270	20.6	Not Det	11/10/95	DH
Methapyrilene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
3-Methylcholanthrene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Methyl methanesulfonate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Methylnaphthalene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Methylphenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
3-Methylphenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4-Methylphenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Naphthalene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,4-Naphthoquinone	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1-Naphthylamine	ug/L	SW8270	103	Not Det	11/10/95	DH
2-Naphthylamine	ug/L	SW8270	103	Not Det	11/10/95	DH
2-Nitroaniline	ug/L	SW8270	51.6	Not Det	11/10/95	DH
3-Nitroaniline	ug/L	SW8270	51.6	Not Det	11/10/95	DH
4-Nitroaniline	ug/L	SW8270	51.6	Not Det	11/10/95	DH
Nitrobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2-Nitrophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
4-Nitrophenol	ug/L	SW8270	51.6	Not Det	11/10/95	DH
4-Nitroquinoline 1-oxide	ug/L	SW8270	51.6	Not Det	11/10/95	DH
n-Nitrosodi-n-butylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosodiethylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosodimethylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosodiphenylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosodi-n-propylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosomethylethylamine	ug/L	SW8270	10.3	Not Det	11/10/95	DH

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 6

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
n-Nitrosomorpholine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosopiperidine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
n-Nitrosopyrrolidine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
5-Nitro-o-toluidine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Pentachlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Pentachloronitrobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Pentachlorophenol	ug/L	SW8270	30.9	Not Det	11/10/95	DH
Phenacetin	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Phenanthrene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Phenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
p-Phenylenediamine	ug/L	SW8270	103	Not Det	11/10/95	DH
2-Picoline	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Polychlorinated dibenzo-p-dioxins (PCDDs)	ug/L	SW8270	1.03	Not Det	11/10/95	DH
Polychlorinated dibenzofurans (PCDFs)	ug/L	SW8270	1.03	Not Det	11/10/95	DH
Pronamide	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Pyrene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
Pyridine	ug/L	SW8270	20.6	Not Det	11/10/95	DH
Safrole	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,2,4,5-Tetrachlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2,3,7,8-Tetrachlorodibenzo-p-dioxin	ug/L	SW8270	1.03	Not Det	11/10/95	DH
2,3,4,6-Tetrachlorophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
o-Toluidine	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,2,4-Trichlorobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
2,4,5-Trichlorophenol	ug/L	SW8270	20.6	Not Det	11/10/95	DH
2,4,6-Trichlorophenol	ug/L	SW8270	10.3	Not Det	11/10/95	DH
0,0,0-Triethyl phosphorothioate	ug/L	SW8270	10.3	Not Det	11/10/95	DH
1,3,5-Trinitrobenzene	ug/L	SW8270	10.3	Not Det	11/10/95	DH
sur-2-Fluorophenol %R 21-100	ug/L	SW8270	0	52	11/10/95	DH
sur-Phenol-d6 %R 10-94	ug/L	SW8270	0	55	11/10/95	DH
sur-Nitrobenzene-d5 %R 35-114	ug/L	SW8270	0	71	11/10/95	DH
sur-2-Fluorobiphenyl %R 43-116	ug/L	SW8270	0	71	11/10/95	DH
sur-2,4,6-Tribromophenol %R 10-123	ug/L	SW8270	0	80	11/10/95	DH
sur-Terphenyl-d14 %R 38-141	ug/L	SW8270	0	97	11/10/95	DH
623-3520 Cont. Liq/Liq Ext. Pest		SW3520		Done	10/28/95	BT

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 7

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
621-8080 App.IX P/P Water ug/L	ug/L	SW8080		<u>TITLE</u>	11/15/95	AMP
Aldrin	ug/L	SW8080	0.0440	Not Det	11/15/95	AMP
alpha-BHC	ug/L	SW8080	0.0330	Not Det	11/15/95	AMP
beta-BHC	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
delta-BHC	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
gamma-BHC	ug/L	SW8080	0.0440	Not Det	11/15/95	AMP
alpha-Chlordane	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
gamma-Chlordane	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
4,4'-DDD	ug/L	SW8080	0.110	Not Det	11/15/95	AMP
4,4'-DDE	ug/L	SW8080	0.0440	Not Det	11/15/95	AMP
4,4'-DDT	ug/L	SW8080	0.110	Not Det	11/15/95	AMP
Dieldrin	ug/L	SW8080	0.0220	Not Det	11/15/95	AMP
Endosulfan I	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
Endosulfan II	ug/L	SW8080	0.0440	Not Det	11/15/95	AMP
Endosulfan sulfate	ug/L	SW8080	0.110	Not Det	11/15/95	AMP
Endrin	ug/L	SW8080	0.0660	Not Det	11/15/95	AMP
Endrin aldehyde	ug/L	SW8080	0.110	Not Det	11/15/95	AMP
Heptachlor	ug/L	SW8080	0.0330	Not Det	11/15/95	AMP
Heptachlor epoxide	ug/L	SW8080	0.0550	Not Det	11/15/95	AMP
Methoxychlor	ug/L	SW8080	0.550	Not Det	11/15/95	AMP
Toxaphene	ug/L	SW8080	2.20	Not Det	11/15/95	AMP
Aroclor-1016	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
Aroclor-1221	ug/L	SW8080	1.10	Not Det	11/15/95	AMP

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 8

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Class III Tank 2 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Class III Tank 2 Leachate

Date Sampled: 10/24/95

Lab ID: AA81171

Time Sampled: 11:15

Collector: C.C.

Log In Date: 10/27/95

Log In Time: 14:33


--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Aroclor-1232	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
Aroclor-1242	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
Aroclor-1248	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
Aroclor-1254	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
Aroclor-1260	ug/L	SW8080	1.10	Not Det	11/15/95	AMP
sur-Dibutylchloroendate %R 33-186	ug/L	SW8080	0	59	11/15/95	AMP
623-8150 Herbicides Ext W		SW8150		Done	10/31/95	DT
621-8150 App.IX Herb. Water ug/L	ug/L	SW8150		_TITLE_	12/04/95	SW
2,4-D	ug/L	SW8150	2.00	Not Det	12/04/95	SW
Dinoseb	ug/L	SW8150	1.00	Not Det	12/04/95	SW
2,4,5-T	ug/L	SW8150	0.200	Not Det	12/04/95	SW
2,4,5-TP (Silvex)	ug/L	SW8150	0.200	Not Det	12/04/95	SW
sur-DCAA %R 28-124	ug/L	SW8150	0	46	12/04/95	SW

Remarks:

Signed:


James M.G. Tucci
Laboratory Manager

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 1

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Client #32660 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Date Sampled:

Station ID: Trip Blank

Time Sampled:

Lab ID: AA81172

Log In Date: 10/27/95

Collector:

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
622-8260 App.IX VOA Water ug/L	ug/L	SW8260		_TITLE_	11/02/95	GH
Acetone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Acetonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrolein	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Acrylonitrile	ug/L	SW8260	20.0	Not Det	11/02/95	GH
Allyl chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Benzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromodichloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromoform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Bromomethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
2-Butanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Carbon disulfide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Carbon tetrachloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chlorobenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloroethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroform	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Chloromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Chloroprene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dibromo-3-chloropropane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,2-Dibromoethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromomethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,4-Dichloro-2-butene	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Dichlorodifluoromethane	ug/L	SW8260	10.0	Not Det	11/02/95	GH
1,1-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,2-Dichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2-Dichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
cis-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
trans-1,3-Dichloropropene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Dibromochloromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethylbenzene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Ethyl methacrylate	ug/L	SW8260	5.00	Not Det	11/02/95	GH
2-Hexanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH

LAW ENVIRONMENTAL NATIONAL LABC TORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 2

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Client #32660 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Trip Blank

Date Sampled:

Lab ID: AA81172

Time Sampled:

Collector:

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
Isobutyl alcohol	ug/L	SW8260	50.0	Not Det	11/02/95	GH
Methacrylonitrile	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methylene chloride	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl iodide	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Methyl methacrylate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
4-Methyl-2-pentanone	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Pentachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Propionitrile	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Styrene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2,2-Tetrachloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Tetrachloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Toluene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,1-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,1,2-Trichloroethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichloroethene	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Trichlorofluoromethane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
1,2,3-Trichloropropane	ug/L	SW8260	5.00	Not Det	11/02/95	GH
Vinyl acetate	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Vinyl chloride	ug/L	SW8260	10.0	Not Det	11/02/95	GH
Xylenes (total)	ug/L	SW8260	5.00	Not Det	11/02/95	GH
sur-1,2-Dichloroethane-d4 %R 76-114	ug/L	SW8260	0	104	11/02/95	GH
sur-Toluene-d8 %R 88-110	ug/L	SW8260	0	98	11/02/95	GH

LAW ENVIRONMENTAL NATIONAL LABORATORIES

TEST DATA REPORT

12/06/95

--- Project Information ---

Ms. Rose Stibila

Page 3

Broward Testing Laboratory, Inc.

Project Name: PASCO3

1034 N.E. 44th Court

Client #32660 Project #1826

Ft. Lauderdale, FL 33334

--- Sample Information ---

Station ID: Trip Blank

Date Sampled:

Lab ID: AA81172

Time Sampled:

Collector:

Log In Date: 10/27/95

Log In Time: 14:33

--- Test Information ---

Analysis

Parameter	Units	Method	Det Lim	Result	Date	Tech
sur-Bromofluorobenzene 86-115	ug/L	SW8260	0	96	11/02/95	GH

Remarks:

Signed: _____

James M.G. Tucci

Laboratory Manager



Lancaster Laboratories
Where quality is a science.

Page: 1 of 3

LLI Sample No. WW 2409116
Collected: 10/30/95 at 11:45

Submitted: 11/ 2/95 Reported: 11/13/95
Discard: 11/21/95

AA81497 Grab Water Sample
Class III Landfill Tank -1
Pasco County Utilities

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

P.O. 1826
Rel.

CAT NO.	ANALYSIS NAME	AS RECEIVED		
		RESULTS	LIMIT OF QUANTITATION	UNITS
1320	Appendix IX Organophosphates			See Page 2

1 COPY TO Law Engineering & Env Svcs Inc ATTN: Ms. Kelli Silvia

Questions? Contact your Client Services Representative
F. Bradley Ayars at (717) 656-2300
08:23:39 D 0001 1 490462
332 0.00 00020000 ASR000

Respectfully Submitted
Jenifer E. Hess, B.S.
Group Leader Pesticides/PCBs



Lancaster Laboratories
2425 New Holland Pike
PO. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations.



#221
9/13/



Lancaster Laboratories
Where quality is a science.

Page: 2 of 3

LLI Sample No. WW 2409116
Collected: 10/30/95 at 11:45

Submitted: 11/ 2/95 Reported: 11/13/95
Discard: 11/21/95

AA81497 Grab Water Sample
Class III Landfill Tank -1
Pasco County Utilities

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

P.O. 1826
Rel.

		AS RECEIVED		
CAT NO.	ANALYSIS NAME	RESULTS	LIMIT OF QUANTITATION	UNITS
Appendix IX Organophosphates				
1317	Disulfoton	< 0.3	0.3	ug/l
3027	Methyl Parathion	< 0.2	0.2	ug/l
3028	Ethyl Parathion	< 0.2	0.2	ug/l
1318	Famphur	< 0.5	0.5	ug/l
1319	Phorate	< 0.3	0.3	ug/l

Questions? Contact your Client Services Representative
F. Bradley Ayars at (717) 656-2300

Respectfully Submitted
Jenifer E. Hess, B.S.
Group Leader Pesticides/PCBs



Lancaster Laboratories
2425 New Holland Pike
P.O. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations.



#22
9-11



Lancaster Laboratories
Where quality is a science.

LABORATORY CHRONICLE

Page: 3 of 3

LLI Sample No. WW 2409116
Collected: 10/30/95 at 11:45

Submitted: 11/02/95

AA81497 Grab Water Sample
Class III Landfill Tank -1
Pasco County Utilities
SDG#:

Account No: 07888
Law Engineering & Env Svcs Inc
National Laboratories Division
3355 MacLemore Drive
Pensacola FL 32514

CAT NO	ANALYSIS NAME	METHOD	TRIAL	DATE OF ANALYSIS	ANALYST
0817	Water Sample Pest. Extraction	SW-846 3510B	1	11/06/95	JSF
1320	Appendix IX Organophosphates	SW-846 8141A	1	11/09/95	DDS



Lancaster Laboratories
2425 New Holland Pike
P.O. Box 12425
Lancaster, PA 17605-2425
717-656-2300 FAX 717-656-2681

See reverse side for explanation of symbols and abbreviations



#22
9/10

3 - 7470

**WATER MERCURY LAB CONTROL SAMPLE
HG**

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Data File: HGWL81171

Level: (low/med) LOW

COMPOUND	SPIKE ADDED (mg/L)	LCS CONCENTRATION (mg/L)	% REC.	QC LIMITS # REC.
Mercury	0.00100	0.000990	99	80 - 120

hgwl8117

3 - 7470

MERCURY
WATER MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY
HG

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Sample No.: Class III Tank 2/AA81171

Level: (low/med) LOW

Class III Tank 2MS/ AA81171 COMPOUND	SPIKE ADDED (mg/L)	SAMPLE CONCENTRATION (mg/L)	MS CONCENTRATION (mg/L)	MS % REC. #	QC LIMITS REC.
Mercury	0.00100	0.00	0.00088	88	75 - 125

Class III Tank 2MSD/ AA81171 COMPOUND	SPIKE ADDED (mg/L)	MSD CONCENTRATION (mg/L)	MSD % REC.	% # RPD	% # RPD	QC LIMITS REC.
Mercury	0.00100	0.00091	91	3	20	75 - 125

Column to be used to flag recovery and RPD values with an asterisk.

* Values outside of QC limits.

RPD: 0 out of 1 outside limits.

Spike Recovery: 0 out of 2 outside limits.

hgwm8117

3 - 300

WATER FLUORIDE BY IC LAB CONTROL SAMPLE
Fluoride

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Data File: FLWL81171

Level: (low/med) LOW

COMPOUND	SPIKE ADDED (mg/L)	LCS CONCENTRATION (mg/L)	% REC.	QC LIMITS # REC.
Fluoride	1.0	0.99	99	76 - 116

flw8117

3 - SW8270

WATER SEMI-VOLATILE LAB CONTROL SAMPLE

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Data File: SWL81171

Level: (low/med) LOW

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	LCS % REC. #	QC LIMITS REC.
Acenaphthene	50.00	39.6	79	47 - 136
Acenaphthylene	50.00	41.3	83	39 - 135
Anthracene	50.00	42.8	86	43 - 133
Benzo(a)anthracene	50.00	37.5	75	48 - 139
Benzo(b)fluoranthene	50.00	35.7	71	24 - 159
Benzo(k)fluoranthene	50.00	35.7	71	11 - 162
Benzo(g,h,i)perylene	50.00	37.2	74	2 - 158
Benzo(a)pyrene	50.00	36.5	73	17 - 156
Benzoic acid	50.00	1.3	3	D - 250
Benzyl alcohol	50.00	8.6	17	D - 229
bis(2-Chloroethoxy)methane	50.00	38.8	78	33 - 154
bis(2-Chloroethyl)ether	50.00	33.2	66	12 - 140
bis(2-Chloroisopropyl)ether	50.00	72.4	145	36 - 166
bis(2-Ethylhexyl)phthalate	50.00	35.8	72	29 - 158
4-Bromophenyl phenyl ether	50.00	47.4	95	53 - 127
Butyl benzyl phthalate	50.00	40.4	81	36 - 152
4-Chloroaniline	50.00	37.3	75	D - 189
4-Chloro-3-methylphenol	50.00	35.8	72	44 - 136
2-Chloronaphthalene	50.00	41.7	83	60 - 118
2-Chlorophenol	50.00	37.6	75	28 - 134
4-Chlorophenyl phenyl ether	50.00	42.1	84	49 - 134
Chrysene	50.00	41.9	84	52 - 138
Dibenz(a,h)anthracene	50.00	37.9	76	D - 194

WATER SEMI-VOLATILE LAB CONTROL SAMPLE

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Data File: SWL81171

Level: (low/med) LOW

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	LCS % REC. #	QC LIMITS REC.
Dibenzofuran	50.00	41.1	82	27 - 127
Di-n-butylphthalate	50.00	37.9	76	37 - 118
1,2-Dichlorobenzene	50.00	36.7	73	32 - 124
1,3-Dichlorobenzene	50.00	33.0	66	10 - 138
1,4-Dichlorobenzene	50.00	35.2	70	20 - 124
3,3'-Dichlorobenzidine	50.00	49.4	99	D - 262
2,4-Dichlorophenol	50.00	29.6	59	39 - 135
Diethylphthalate	50.00	41.8	84	41 - 114
2,4-Dimethylphenol	50.00	2.4	5 *	32 - 113
Dimethylphthalate	50.00	40.8	82	48 - 112
4,6-Dinitro-2-methylphenol	50.00	17.0	34	D - 153
2,4-Dinitrophenol	50.00	0.6	1	D - 147
2,4-Dinitrotoluene	50.00	40.0	80	44 - 135
2,6-Dinitrotoluene	50.00	40.5	81	50 - 135
Di-n-octylphthalate	50.00	42.6	85	4 - 146
Fluoranthene	50.00	35.2	70	49 - 137
Fluorene	50.00	40.4	81	64 - 121
Hexachlorobenzene	50.00	49.0	98	34 - 152
Hexachlorobutadiene	50.00	35.5	71	33 - 116
Hexachloroethane	50.00	31.3	63	40 - 113
Indeno(1,2,3-cd)pyrene	50.00	38.3	77	D - 171

WATER SEMI-VOLATILE LAB CONTROL SAMPLE

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Data File: SWL81171

Level: (low/med) LOW

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	LCS % REC. #	QC LIMITS REC.
Isophorone	50.00	37.1	74	36 - 132
2-Methylnaphthalene	50.00	23.0	46	D - 163
2-Methylphenol	50.00	27.2	54	D - 169
4-Methylphenol	50.00	27.2	54	24 - 142
Naphthalene	50.00	37.8	76	42 - 128
2-Nitroaniline	100.00	37.1	37	D - 194
3-Nitroaniline	100.00	37.1	37	D - 194
4-Nitroaniline	50.00	39.3	79	D - 250
Nitrobenzene	50.00	38.0	76	35 - 140
2-Nitrophenol	50.00	35.3	71	31 - 139
4-Nitrophenol	50.00	4.8	10	D - 132
N-Nitrosodiphenylamine	50.00	57.4	115	D - 245
N-Nitroso-di-n-propylamine	50.00	34.6	69	30 - 144
Pentachlorophenol	50.00	2.1	4 *	14 - 176
Phenanthrene	50.00	41.9	84	60 - 120
Phenol	50.00	33.0	66	19 - 112
Pyrene	50.00	54.1	108	52 - 115
1,2,4-Trichlorobenzene	50.00	35.4	71	44 - 124
2,4,5-Trichlorophenol	50.00	37.8	76	38 - 142
2,4,6-Trichlorophenol	50.00	34.1	68	42 - 144

Column to be used to flag recovery with an asterisk

* Values outside of QC limits

File: slw51111

3 - 8270
SEMIVOLATILE
WATER MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Law Environmental, Inc.

Contract: 1826

Lab Code: LENL-P

Sample No.: Class III Tank 2/AA81171

Level: (low/med) LOW

Class III Tank 2MS/AA81171	SPIKE	SAMPLE	MS	MS	QC	
COMPOUND	ADDED	CONCENTRATION	CONCENTRATION	%	LIMITS	
	ug/L	ug/L	ug/L	REC.	#	REC.
Acenaphthene	111	0.00	70.3	63	47 -	136
Benzo(a)anthracene	111	0.00	74.7	67	48 -	139
Benzo(a)pyrene	111	0.00	66.2	60	17 -	156
bis(2-Chloroethoxy)methane	111	0.00	65.0	59	33 -	154
Butyl benzyl phthalate	111	0.00	75.1	68	36 -	152
4-Chloro-3-methylphenol	167	0.00	104.0	62	44 -	136
2-Chlorophenol	167	0.00	102.0	61	28 -	134
1,4-Dichlorobenzene	111	0.00	57.9	52	20 -	124
2,4-Dinitrotoluene	111	0.00	69.1	62	44 -	135
Hexachlorobenzene	111	0.00	86.1	77	34 -	152
Naphthalene	111	0.00	65.20	59	42 -	128
4-Nitrophenol	167	0.00	16.1	10	D -	132
N-Nitroso-di-n-propylamine	111	0.00	66.6	60	30 -	144
Pentachlorophenol	167	0.00	87.9	53	14 -	176
Phenol	167	0.00	78.9	47	19 -	112
Pyrene	111	0.00	97.3	88	52 -	115
1,2,4-Trichlorobenzene	111	0.00	58.9	53	44 -	124

Class III Tank 2MSD/AA81171	SPIKE	MSD	MSD	%	RPD	QC	
COMPOUND	ADDED	CONCENTRATION	%	REC.	RPD	LIMITS	
	ug/L	ug/L	REC.		#	REC.	
Acenaphthene	110	68.9	63	1	40	47 -	136
Benzo(a)anthracene	110	76.6	70	4	40	48 -	139
Benzo(a)pyrene	110	67.1	61	2	40	17 -	156
bis(2-Chloroethoxy)methane	110	65.3	59	2	40	33 -	154
Butyl benzyl phthalate	110	75.4	69	1	40	36 -	152
4-Chloro-3-methylphenol	165	107.0	65	4	40	44 -	136
2-Chlorophenol	165	104.0	63	3	40	28 -	134
1,4-Dichlorobenzene	110	61.7	56	7	40	20 -	124
2,4-Dinitrotoluene	110	70.1	64	3	40	44 -	135
Hexachlorobenzene	110	88.7	81	4	40	34 -	152
Naphthalene	110	64.2	58	0	40	42 -	128
4-Nitrophenol	165	29.5	18	60	* 40	D -	132
N-Nitroso-di-n-propylamine	110	66.7	61	1	40	30 -	144
Pentachlorophenol	165	109.0	66	23	40	14 -	176
Phenol	165	86.4	52	10	40	19 -	112
Pyrene	110	98.8	90	3	40	52 -	115
1,2,4-Trichlorobenzene	110	60.4	55	4	40	44 -	124

Column to be used to flag recovery and RPD values with an asterisk.

* Values outside of QC limits.

RPD: 1 out of 17 outside limits.

Spike Recovery: 0 out of 34 outside limits.

smw81171

WATER PESTICIDES/PCB LAB CONTROL SPIKE

Column I.D.: DB608

SF: PPWL7433

Time Analyzed: 00:31-01:24

SDG:

Date Analyzed: 11-15-95

Run Log #: 7433

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	% REC.#	QC LIMITS REC.	PQL
Aldrin	0.200	0.186	93	42 - 122	0.040
alpha-BHC	0.200	0.206	103	49 - 134	0.030
beta-BHC	0.200	0.223	112	25 - 147	0.040
delta-BHC	0.200	0.162	81	19 - 139	0.040
gamma-BHC (Lindane)	0.200	0.193	97	55 - 127	0.040
alpha-Chlordane	0.200	0.207	104	38 - 161	0.020
gamma-Chlordane	0.200	0.203	102	42 - 155	0.020
4,4'-DDD	0.400	0.394	99	36 - 141	0.100
4,4'-DDE	0.400	0.396	99	40 - 145	0.040
4,4'-DDT	0.400	0.367	92	59 - 148	0.100
Dieldrin	0.400	0.386	97	40 - 146	0.050
Endosulfan I	0.200	0.147	74	45 - 149	0.040
Endosulfan II	0.400	0.332	83	24 - 159	0.100
Endosulfan sulfate	0.400	0.385	96	36 - 144	0.100
Endrin	0.400	0.442	111	42 - 147	0.060
Endrin aldehyde	0.400	0.386	97	39 - 159	0.100
Heptachlor	0.200	0.183	92	43 - 111	0.030
Heptachlor epoxide	0.200	0.196	98	42 - 142	0.050
Methoxychlor	2.00	2.07	104	62 - 161	0.400
PCB-1016	10.00	NS	0	50 - 114	1.00
PCB-1260	10.00	NS	0	8 - 127	1.00
=====					
DBC	1.00	0.890	89	33 - 168	
TCMX	0.200	0.178	89	24 - 151	

NS - Not spiked.

PESTICIDES/PCB
WATER MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY
Law Environmental, Inc.

Column I.D.: DB - 608

SF: PPMW7433

SDG:

Data File MS.: CNOVY48

Data File MSD.: CNOVY49

MS Sample No.: AA81171

MSD Sample No.: AA81171

MS Pest Date Analyzed: 11/15/95

MSD Pest Date Analyzed: 11/15/95

MS Pest Time Analyzed: 03:10

MSD Pest Time Analyzed: 04:03

MS PCB Date Analyzed:

MSD PCB Date Analyzed:

MS PCB Time Analyzed:

MSD PCB Time Analyzed:

MS Pest C.F.: 2.27

MSD Pest C.F.: 2.27

MS PCB C.F.:

MSD PCB C.F.:

COMPOUND	SPIKE ADDED (ug/L)	SAMPLE CONCENTRATION (ug/L)	MS CONCENTRATION (ug/L)	MS % REC.#	QC LIMITS REC.
gamma-BHC (Lindane)	0.909	0.000	0.730	80	55 -127
Heptachlor	0.909	0.000	0.799	88	43 -111
Aldrin	0.909	0.000	0.718	79	42 -122
Dieldrin	2.27	0.000	1.930	85	40 -146
Endrin	2.27	0.000	1.630	72	42 -147
DDT	2.27	0.000	1.950	86	59 -148
AR1016					50 -114
AR1260					8 -127

DBC	2.27		1.560	69	33 -186
TCMX	0.455		0.373	82	24 -151

COMPOUND	SPIKE ADDED (ug/L)	MSD CONCENTRATION (ug/L)	MSD % REC.	% # RPD #	RPD	REC
gamma-BHC (Lindane)	0.909	0.717	79	2	40	55 -127
Heptachlor	0.909	0.792	87	1	40	43 -111
Aldrin	0.909	0.702	77	2	40	42 -122
Dieldrin	2.27	1.900	84	2	40	40 -146
Endrin	2.27	1.700	75	4	40	42 -147
DDT	2.27	1.940	85	1	40	59 -148
AR1016					40	50 -114
AR1260					40	8 -127

DBC	2.27	1.560	69			33 -186
TCMX	0.455	0.392	86			24 -151

Column to be used to flag recovery and RPD values with an asterisk.
* Values outside of QC limits.

RPD: 0 out of 6 outside limits.
Spike Recovery: 0 out of 12 outside limits.

Comments:

Herbicide-Water LAB-CONTROL-SPIKE

Lab Name: Law Environmental, Inc.

Data File: CDECW12

Date Analyzed: 12/04/95

Run Log no: 7453

Time Analyzed: 19:27

Column ID: DB-608

Compound	Spike Added (ug/L)	LCS Concentration (ug/L)	Spike Recov %	QC Recov Limits	PQL (ug/L)
Dalapon	25.00	7.37	29	D - 112	7.00
DCAA (surrogate) >>>>>>>>	10.00 surr	4.79	48	28 - 124	<> (surr)
MCPP	100.00	76.70	77	14 - 138	200.00
Dicamba	1.00	0.48	48	35 - 136	0.20
MCPA	100.00	73.20	73	12 - 146	200.00
Dichloroprop	10.00	4.44	44	35 - 131	2.00
2,4 D	10.00	4.47	45	4 - 143	2.00
2,4,5 TP (Silvex)	1.00	0.49	49	25 - 122	0.20
2,4,5 T	1.00	0.47	47	24 - 144	0.20
Dinoseb	5.00	0.45	9	D - 143	1.00
2,4 DB	10.00	4.56	46	9 - 122	3.00

=====

Herbicide-Water
MATRIX SPIKE/MATRIX SPIKE DUPLICATE Recovery
Law Environmental, Inc.

Column ID: DB-608

MS
Data File: CDECW13
Sample No: AA81171
Date Analyzed: 12/04/95
Time Analyzed: 20:03
Calib Factor: 2.0

Run Log No: 7453

MSD
Data File: CDEW146
Sample No: AA81171
Date Analyzed: 12/04/95
Time Analyzed: 20:39
Calib Factor: 2.0

Compound	Spike Added (ug/L)	Sample Conc (ug/L)	MS Conc (ug/L)	Spike Recov %	QC Recov Limits
2,4 D	10.00	0.0	8.99	90	4 -143
2,4,5 T	1.00	0.00	0.86	86	24 -144
2,4,5 TP (SILVEX)	1.00	0.00	0.83	83	25 -122
DCAA	10.00		9.12	91	28 -124

Compound	Spike Added (ug/L)	MSD Conc (ug/L)	Spike Recov %	RPD %	QC Recov Limits
2,4 D	10.00	8.68	87	4	4 -143
2,4,5 T	1.00	0.81	81	6	24 -144
2,4,5 TP (SILVEX)	1.00	0.80	80	4	25 -122
DCAA	10.00	8.85	89	3	28 -124

Column to be used to flag recovery and RPD values with an asterisk.
* Values outside of QC limits.

RPD: 0 out of 3 outside limits.
Spike Recovery: 0 out of 6 outside limits.

Comments: