684,"0001",,"RelativePercentDifference",,"The ""RelativePercentDifference"" field is missing information in this record.","a,a-Dimethylphenethylamine does not have a RPD because both LCS and LCSD were 0% recovery.",,,1
687,"0001",,"RelativePercentDifference",,"The ""RelativePercentDifference"" field is missing information in this record.","a,a-Dimethylphenethylamine does not have a RPD because both LCS and LCSD were 0% recovery.",,,1
1037,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1036,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1031,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1032,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1027,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1028,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1038,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1030,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1033,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1035,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1039,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1029,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1026,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1040,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.08 MG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1034,"0007",,"ReportingLimit",,"The Nitrite (N) Reporting Limit value of 0.1 MG/L

reported in this record for Method EPA 300.1 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2053,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2045,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2047,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2052,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2048,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2050,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2049,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2043,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2054,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2051,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2044,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2046,"0007",,"ReportingLimit",,"The Magnesium Reporting Limit value of 0.1 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.08 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2096,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2098,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2097,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

2102,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2099,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2095,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2094,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2101,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2092,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2093,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2091,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2100,"0007",,"ReportingLimit",,"The Potassium Reporting Limit value of 0.5 MG/L
reported in this record for Method EPA 6010 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 MG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
138,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
139,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
141,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
137,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
140,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
142,"0007",,"ReportingLimit",,"The 4-Bromophenyl phenyl ether Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
207,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.1 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
208,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
204,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
206,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
203,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
205,"0007",,"ReportingLimit",,"The 4-Methylphenol Reporting Limit value of 11.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
221,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
222,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
217,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
220,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
218,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
219,"0007",,"ReportingLimit",,"The 1,4-Dichlorobenzene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
239,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
246,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
238,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
249,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
240,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
237,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
247,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
248,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
245,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
250,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
243,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
242,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
244,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
241,"0007",,"ReportingLimit",,"The 1,4-Phenylenediamine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
278,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
274,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
277,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
253,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
268,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
259,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
272,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L

reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
254,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
275,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
264,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
276,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
266,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
280,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
252,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
251,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
270,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
258,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
279,"0007",,"ReportingLimit",,"The 1,2-Dibromoethane Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.1 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
320,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
319,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
317,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
334,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

323,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
318,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
330,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
325,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
331,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
329,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
332,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
333,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
326,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
327,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
322,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
321,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
324,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
328,"0007",,"ReportingLimit",,"The 1,2-Dichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
451,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
444,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
449,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
446,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
453,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
440,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
439,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
441,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
452,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
443,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
447,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
450,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
442,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
454,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
456,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
445,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
448,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
455,"0007",,"ReportingLimit",,"The Chlorobenzene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
470,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
458,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
466,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
463,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
467,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
461,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
465,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
462,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
460,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
468,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
459,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
457,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
469,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
464,"0007",,"ReportingLimit",,"The Phenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
478,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
483,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 22.2 UG/L

reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
476,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
473,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
472,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
480,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 22.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
471,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
481,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 23.2 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
477,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
482,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 22.2 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
484,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 22.2 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
474,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
479,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 22.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
475,"0007",,"ReportingLimit",,"The 2-Picoline Reporting Limit value of 20 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
539,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value
of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
542,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
544,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1

541,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
540,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
543,"0007",,"ReportingLimit",,"The Bis(2-chloroethoxy)methane Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
570,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
568,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
569,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
571,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
572,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
567,"0007",,"ReportingLimit",,"The Di-n-Octyl phthalate Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
583,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
582,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
586,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
584,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
585,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
581,"0007",,"ReportingLimit",,"The Hexachlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
593,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
592,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
594,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
595,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
588,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
589,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
597,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
590,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
596,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
591,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
587,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
598,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
599,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
600,"0007",,"ReportingLimit",,"The 3,3'-Dimethylbenzidine Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
613,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
609,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
610,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
614,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
611,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
612,"0007",,"ReportingLimit",,"The Anthracene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
652,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
649,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
650,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
654,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
651,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
653,"0007",,"ReportingLimit",,"The 1,2,4-Trichlorobenzene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
669,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
672,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
676,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
673,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
679,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
677,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
678,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.5

UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
681,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
671,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
682,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
674,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
670,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
675,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
680,"0007",,"ReportingLimit",,"The 2,4-Dinitrotoluene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
683,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
684,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
691,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 18 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
695,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 17.8 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
687,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
693,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 18.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
692,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 18 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
686,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

694,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 17.8 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
690,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
688,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
689,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
685,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 16 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
696,"0007",,"ReportingLimit",,"The a,a-Dimethylphenethylamine Reporting Limit value
of 17.8 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
700,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
713,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
708,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
701,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
709,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
714,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
705,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
704,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
706,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
697,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
712,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
711,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
699,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
698,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
702,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
703,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
707,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
710,"0007",,"ReportingLimit",,"The 1,4-Dioxane Reporting Limit value of 40 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
789,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
790,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
795,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
788,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
794,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
797,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
796,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
798,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
784,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
787,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
786,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
792,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
783,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
793,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
799,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
800,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
791,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
785,"0007",,"ReportingLimit",,"The Tetrachloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
809,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.5 UG/L reported
in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
810,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.5 UG/L reported
in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
812,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.4 UG/L reported
in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
813,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.4 UG/L reported
in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
811,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.6 UG/L reported
in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
814,"0007",,"ReportingLimit",,"The Pyrene Reporting Limit value of 4.4 UG/L reported

Page 17

in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
884,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
883,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
887,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
886,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
885,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
888,"0007",,"ReportingLimit",,"The 1-Naphthylamine Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
909,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
908,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
914,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
911,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
912,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
905,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
903,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
907,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
915,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

906,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
916,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
913,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
910,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
904,"0007",,"ReportingLimit",,"The Kepone Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 6 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
981,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
978,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
991,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
988,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
990,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
994,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
977,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
993,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
984,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
987,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
985,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
983,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
982,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
986,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
980,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
992,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
979,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
989,"0007",,"ReportingLimit",,"The trans-1,2-Dichloroethene Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1081,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1085,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1084,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1082,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1083,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1086,"0007",,"ReportingLimit",,"The Benzo(g,h,i)perylene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1101,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1099,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1104,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1108,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1102,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1112,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1106,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1105,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1111,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1109,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1100,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1107,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1110,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1103,"0007",,"ReportingLimit",,"The Indeno(1,2,3-cd)pyrene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1118,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1126,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1113,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1121,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1125,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of

4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1123,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1116,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1124,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1119,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1122,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1115,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1114,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1117,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1120,"0007",,"ReportingLimit",,"The Benzo(b)fluoranthene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1135,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1136,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1138,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1140,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1137,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1139,"0007",,"ReportingLimit",,"The Fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

1150,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1152,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1149,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1151,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1153,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1154,"0007",,"ReportingLimit",,"The Benzo(k)fluoranthene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1164,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1168,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1163,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1165,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1166,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1167,"0007",,"ReportingLimit",,"The Acenaphthylene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1180,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1178,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1182,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1177,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
1181,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1179,"0007",,"ReportingLimit",,"The Chrysene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1218,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1222,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1219,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1220,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1217,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1221,"0007",,"ReportingLimit",,"The Thionazin Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1234,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1231,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1224,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1227,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1233,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1230,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1232,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1228,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1226,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1229,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1223,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1225,"0007",,"ReportingLimit",,"The Methyl parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1239,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1244,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1236,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1246,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1245,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1237,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1238,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1240,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1235,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1242,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1243,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1241,"0007",,"ReportingLimit",,"The Phorate Reporting Limit value of 5.7 UG/L

reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1253,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.7 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1254,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.4 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1250,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1258,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.7 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1255,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.6 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1257,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.6 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1251,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1249,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1256,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5.4 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1248,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1247,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1252,"0007",,"ReportingLimit",,"The Disulfoton Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1269,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1267,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1266,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.057 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

1270,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1268,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1265,"0007",,"ReportingLimit",,"The Aldrin Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1289,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1292,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1294,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1290,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.057 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1293,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1291,"0007",,"ReportingLimit",,"The b-BHC Reporting Limit value of 0.056 UG/L
reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.05 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1327,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1329,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1324,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1320,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1328,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1323,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1322,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit
value of 5 UG/L reported in this record for Method EPA 8141 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L

(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1326,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1325,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1330,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1319,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1321,"0007",,"ReportingLimit",,"The Tetraethyl dithiopyrophosphate Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1370,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1367,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1366,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1368,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1365,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1369,"0007",,"ReportingLimit",,"The Benzo(a)pyrene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1380,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1377,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1374,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1384,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1378,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1383,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1371,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1375,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1382,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1381,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.6
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1376,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1379,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1372,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1373,"0007",,"ReportingLimit",,"The Chlorobenzilate Reporting Limit value of 4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1413,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1416,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1418,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1421,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1419,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1415,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1412,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value

of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1422,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1411,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1423,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1420,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1424,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
1414,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1417,"0007",,"ReportingLimit",,"The 4,6-Dinitro-2-methylphenol Reporting Limit value
of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1447,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1440,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1444,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1437,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1450,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1446,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1439,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1449,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

1448,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1445,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1438,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1443,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1442,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1441,"0007",,"ReportingLimit",,"The Dibenzo(a,h)anthracene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1478,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1477,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1474,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1476,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1473,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1475,"0007",,"ReportingLimit",,"The 1,3-Dichlorobenzene Reporting Limit value of 4.6
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1501,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1498,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1503,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1494,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
1493,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1510,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1505,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1508,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1497,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1509,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1495,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1506,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1500,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1499,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1502,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1507,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1504,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1496,"0007",,"ReportingLimit",,"The Carbon tetrachloride Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1521,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.6 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1516,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1517,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1518,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1520,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.4 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1513,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1515,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1519,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.6 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1512,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1522,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5.7 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1511,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1514,"0007",,"ReportingLimit",,"The Parathion Reporting Limit value of 5 UG/L reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1542,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1550,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1547,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1546,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1545,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1549,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.4

UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1541,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1543,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1544,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1548,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1539,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1540,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1537,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1538,"0007",,"ReportingLimit",,"The Benzo(a)anthracene Reporting Limit value of 4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1552,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1561,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1553,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1563,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1551,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1554,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1560,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

1559,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1555,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1562,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1564,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1556,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1557,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1558,"0007",,"ReportingLimit",,"The 4-Nitroquinoline-1-oxide Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1588,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.57 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1592,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.56 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1585,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.5 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1587,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.56 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1586,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.5 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1591,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.56 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1589,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.56 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1590,"0007",,"ReportingLimit",,"The Chlordane Reporting Limit value of 0.56 UG/L reported in this record for Method EPA 8081 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.2 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1631,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected

for dilution, if applicable).","Run with in house RLs.",,,1
1629,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value
of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
1632,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
1627,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value
of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
1630,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value
of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
1628,"0007",,"ReportingLimit",,"The 2,3,4,6-Tetrachlorophenol Reporting Limit value
of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected
for dilution, if applicable).","Run with in house RLs.",,,1
1661,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1659,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1662,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1660,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1663,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1664,"0007",,"ReportingLimit",,"The 4-Chloro-3-methylphenol Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1697,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1702,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.4 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1696,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1703,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.6 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1695,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1701,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.6 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1698,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1699,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.7 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1700,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.4 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1694,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1704,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5.7 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1693,"0007",,"ReportingLimit",,"The Dimethoate Reporting Limit value of 5 UG/L
reported in this record for Method EPA 8141 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1833,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1834,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1839,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1836,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1840,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1835,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1844,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1838,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1845,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of

22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1843,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1841,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1846,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1842,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1837,"0007",,"ReportingLimit",,"The Methyl methanesulfonate Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1871,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1876,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1877,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1882,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1869,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1872,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1868,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1874,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1866,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
1879,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

1865,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1870,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1875,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1873,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1878,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1880,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1867,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1881,"0007",,"ReportingLimit",,"The Chloroform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1929,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1930,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1937,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1933,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1936,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1935,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1941,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1931,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
1944,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1943,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1932,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1934,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1939,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1940,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1946,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1945,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1938,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
1942,"0007",,"ReportingLimit",,"The 1,1,1-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2461,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2454,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2450,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2449,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2457,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2456,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds

the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2448,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2446,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2455,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2462,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2453,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2445,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2447,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2459,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2460,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2451,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2452,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2458,"0007",,"ReportingLimit",,"The Bromoform Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2479,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2476,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2467,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2466,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1

UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2474,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2473,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2468,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2465,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2471,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2478,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2480,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2472,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2464,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2470,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2469,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2475,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2477,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2463,"0007",,"ReportingLimit",,"The Bromodichloromethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2494,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

2483,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2482,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2481,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2492,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2489,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2496,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2484,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2488,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2485,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2495,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2493,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2498,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2486,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2491,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2490,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2497,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
2487,"0007",,"ReportingLimit",,"The 1,1-Dichloroethane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2511,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2514,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2500,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2508,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2499,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2503,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2510,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2502,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2513,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2516,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2515,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2505,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2509,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2501,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2621,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2644,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2637,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2639,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2636,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2646,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2643,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2652,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2642,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2649,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2638,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2648,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2635,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2645,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2651,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2647,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2641,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1

UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2650,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2640,"0007",,"ReportingLimit",,"The 1,2-Dichloropropane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2679,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2675,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2676,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2678,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2684,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2673,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2682,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2686,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2687,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2677,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2683,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2671,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2674,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

2688,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2685,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2680,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2681,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2672,"0007",,"ReportingLimit",,"The 1,1,2-Trichloroethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2704,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2706,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2696,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2689,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2690,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2705,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2695,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2701,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2691,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2702,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2700,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
2699,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2697,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2703,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2693,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2698,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2694,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2692,"0007",,"ReportingLimit",,"The Trichloroethene Reporting Limit value of 1 UG/L
reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2751,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2746,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2745,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2738,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2740,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2741,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2743,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2736,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2747,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2752,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2748,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2750,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2742,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2735,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2739,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2749,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2737,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2744,"0007",,"ReportingLimit",,"The Methyl methacrylate Reporting Limit value of 5
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 2 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2755,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2762,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2758,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2763,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2753,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2760,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2765,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of

4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2754,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2761,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2757,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2756,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2764,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2759,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2766,"0007",,"ReportingLimit",,"The Pentachloronitrobenzene Reporting Limit value of
4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2776,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2778,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2775,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2780,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2777,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.6 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2779,"0007",,"ReportingLimit",,"The Acenaphthene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2794,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2790,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

2791,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.6
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2793,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2789,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2792,"0007",,"ReportingLimit",,"The Diethyl phthalate Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2819,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.6 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2817,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2820,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2821,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2818,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2822,"0007",,"ReportingLimit",,"The Phenanthrene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2862,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2860,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2864,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2863,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2859,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2861,"0007",,"ReportingLimit",,"The Fluorene Reporting Limit value of 4.6 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
2884,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2883,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2879,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2887,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2886,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2881,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2892,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2880,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2885,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2890,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2889,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.6
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2882,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2888,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2891,"0007",,"ReportingLimit",,"The Hexachlorobutadiene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 3 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2898,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2896,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2895,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2903,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2901,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2902,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2900,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2897,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2905,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2904,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2899,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2906,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2893,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2894,"0007",,"ReportingLimit",,"The Pentachlorophenol Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 12 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2944,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2947,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2943,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2945,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.6 UG/L

reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2946,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2948,"0007",,"ReportingLimit",,"The 2-Nitrophenol Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2988,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2987,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2983,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2985,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2986,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2984,"0007",,"ReportingLimit",,"The 2-Methylnaphthalene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3000,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3002,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3001,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.4 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2999,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2998,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2997,"0007",,"ReportingLimit",,"The 2-Chloronaphthalene Reporting Limit value of 4.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3098,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 57 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1

3090,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3096,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 55 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3088,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3093,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 56 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3087,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3094,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 56 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3095,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 57 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3097,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 57 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3092,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3089,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3091,"0007",,"ReportingLimit",,"The MCPP Reporting Limit value of 50 UG/L reported in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 7 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3133,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.6 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3131,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3135,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3136,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.1 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3132,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
3134,"0007",,"ReportingLimit",,"The Safrole Reporting Limit value of 11.1 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3146,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 55 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3145,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 57 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3147,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 57 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3139,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3141,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3140,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3138,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3137,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3142,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 50 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3143,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 56 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3148,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 57 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3144,"0007",,"ReportingLimit",,"The MCPA Reporting Limit value of 56 UG/L reported
in this record for Method EPA 8151 in Matrix AQUEOUS-Groundwater exceeds the project
library reporting limit of 3.5 UG/L (corrected for dilution, if applicable).","Run
with in house RLs.",,,1
3199,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3198,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2512,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2507,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2506,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2504,"0007",,"ReportingLimit",,"The 1,1-Dichloroethene Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2531,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2524,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2525,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2517,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2518,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2521,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2532,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2529,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2527,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2523,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2522,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2526,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of 1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
2520,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of

1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2534,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2530,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2528,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2533,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2519,"0007",,"ReportingLimit",,"The Trichlorofluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.8 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2539,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2551,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2537,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2547,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2542,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2540,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2549,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2544,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2552,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2541,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

2546,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2550,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2535,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2545,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2543,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2538,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2536,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2548,"0007",,"ReportingLimit",,"The Dichlorodifluoromethane Reporting Limit value of
1 UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 0.5 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2612,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2615,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2614,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2613,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.6 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2616,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2611,"0007",,"ReportingLimit",,"The Isophorone Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2619,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2624,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
2617,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2626,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2627,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2632,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2622,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2618,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2631,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2633,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2629,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2628,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2620,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2630,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2625,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2634,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
2623,"0007",,"ReportingLimit",,"The Isobutyl alcohol Reporting Limit value of 80
UG/L reported in this record for Method EPA 8260 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 50 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where

ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID

equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where

ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
3195,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater

exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3200,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.4
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3196,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3197,"0007",,"ReportingLimit",,"The 1,2-Dichlorobenzene Reporting Limit value of 4.6
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3227,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3225,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.6 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3228,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3224,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3226,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.4 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3223,"0007",,"ReportingLimit",,"The 2-Chlorophenol Reporting Limit value of 4.5 UG/L
reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds
the project library reporting limit of 4 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3270,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3279,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3283,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3275,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3281,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3278,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3273,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit

value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3269,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3286,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3271,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3274,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3272,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3276,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3277,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3284,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3280,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3282,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3285,"0007",,"ReportingLimit",,"The 1,2-Dibromo-3-chloropropane Reporting Limit
value of 2 UG/L reported in this record for Method EPA 8260 in Matrix
AQUEOUS-Groundwater exceeds the project library reporting limit of 0.1 UG/L
(corrected for dilution, if applicable).","Run with in house RLs.",,,1
3367,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3375,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of
23.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3376,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of
22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3371,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of
20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1

3368,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3373,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3366,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3365,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3378,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3370,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3372,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3377,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3369,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3374,"0007",,"ReportingLimit",,"The 1,3,5-Trinitrobenzene Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3393,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3400,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3399,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3405,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 22.2 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3398,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if applicable).","Run with in house RLs.",,,1
3402,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 22.5 UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if

applicable).","Run with in house RLs.",,,1
3397,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3403,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 23.2
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3396,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3404,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 22.2
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3406,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 22.2
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3395,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3394,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 20
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
3401,"0007",,"ReportingLimit",,"The 1,3-Dinitrobenzene Reporting Limit value of 22.5
UG/L reported in this record for Method EPA 8270 in Matrix AQUEOUS-Groundwater
exceeds the project library reporting limit of 10 UG/L (corrected for dilution, if
applicable).","Run with in house RLs.",,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID

'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where

ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where

ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where

ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a record where
ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA
8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine
must be reported for this method and libary MatrixID equivalent according to the
project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where

ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID

'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where

ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method

and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where

ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a record where
ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA
8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine
must be reported for this method and libary MatrixID equivalent according to the
project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where

ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID

equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID

'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID 'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where

ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where

ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where
ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a record where ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID 'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID 'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID 'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID 'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID 'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for

this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID

'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where

ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID 'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID 'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID 'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and

libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a record where

ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA
8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine
must be reported for this method and libary MatrixID equivalent according to the
project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID

'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where

ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and

libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Heptachlor epoxide analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Heptachlor epoxide must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Endosulfan sulfate analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Endosulfan sulfate must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where ReportableResult = YES for 3-Methylphenol analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 3-Methylphenol must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where

ReportableResult = YES for PCB-1260 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1260 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1254 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1254 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1221 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1221 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1232 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1232 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Diphenylamine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Diphenylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2-Diphenylhydrazine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Diphenylhydrazine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1248 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1248 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1016 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1016 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Demeton-S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for m,p-Cresols analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. m,p-Cresols must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Carbofuran analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbofuran must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 2,3,7,8-TCDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 2,3,7,8-TCDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for EPN analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Tetrachlorvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Tetrachlorvinphos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Thionazin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Thionazin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Phorate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phorate must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Demeton-O analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Disulfoton analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Disulfoton must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Aldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Aldrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for a-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. a-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for b-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. b-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for d-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. d-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Endosulfan II analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan II must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Tetraethyl dithiopyrophosphate analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Tetraethyl dithiopyrophosphate must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Bis(2-chloroisopropyl)ether analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. Bis(2-chloroisopropyl)ether must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 4,4'-DDT analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDT must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Famphur analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Famphur must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for PCB-1242 analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. PCB-1242 must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Endrin ketone analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin ketone must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Chlordane analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for g-BHC analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. g-BHC must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dimethoate analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dimethoate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dieldrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dieldrin must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8270 in MatrixID

'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Carbaryl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbaryl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Benzoic acid analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzoic acid must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Hexachlorophene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorophene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Endrin analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Methoxychlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Methoxychlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where

ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a record where
ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA
8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine
must be reported for this method and libary MatrixID equivalent according to the
project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Lithium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Lithium must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Molybdenum analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Molybdenum must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Strontium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Strontium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Titanium analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Titanium must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Boron analyzed by EPA 6010 in MatrixID
'AQUEOUS-Groundwater'. Boron must be reported for this method and libary MatrixID

equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8011 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2,3-Trichloropropane analyzed by EPA 8011 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dicofol analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Dicofol must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Hexachlorobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlordane- Technical analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlordane- Technical must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Kepone analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Kepone must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Alachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Alachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Propachlor analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Propachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Diallate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Diallate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Mirex analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Mirex must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Methyl parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Methyl parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Isodrin analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Isodrin must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlorobenzilate analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Chlorobenzilate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for a-Chlordane analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. a-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for g-Chlordane analyzed by EPA 8081 in MatrixID

'AQUEOUS-Groundwater'. g-Chlordane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Parathion analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Parathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Hexachlorocyclopentadiene analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. Hexachlorocyclopentadiene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Pentachloronitrobenzene analyzed by EPA 8081 in MatrixID
'AQUEOUS-Groundwater'. Pentachloronitrobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2-Dibromo-3-Chloropropane analyzed by EPA 8081 in
MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-Chloropropane must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Tetraethyl pyrophosphate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tetraethyl pyrophosphate must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Fensulfothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fensulfothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Malathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Malathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Fenitrothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenitrothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Simazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Simazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Terbufos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Terbufos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Phosphamidon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosphamidon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Ethoprop analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethoprop must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dicrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dicrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Merphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Merphos must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Atrazine analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Atrazine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where

ReportableResult = YES for EPN analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. EPN must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Leptophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Leptophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Azinphos ethyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos ethyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorpyrifos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Ronnel analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ronnel must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Naled analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Naled must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Aspon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Aspon must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Trichloronate analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichloronate must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Diazinon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Diazinon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Tokuthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Tokuthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Bolstar analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Bolstar must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Chlorfenvinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Chlorfenvinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Trichlorofon analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Trichlorofon must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Fenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Methyl chlorpyrifos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Methyl chlorpyrifos must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Ethion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Ethion must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Coumaphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Coumaphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dichlorvos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlorvos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Monocrotophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Monocrotophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Fonophos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Fonophos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Stirofos analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Stirofos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dichlofenthion analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. Dichlofenthion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 4-Nitrophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. 4-Nitrophenol must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Picloram analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Picloram must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Pentachlorophenol analyzed by EPA 8151 in MatrixID
'AQUEOUS-Groundwater'. Pentachlorophenol must be reported for this method and

libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where

ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 4,4'-DDD analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDD must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 4,4'-DDE analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 4,4'-DDE must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Phosmet analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Phosmet must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Endrin aldehyde analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Endrin aldehyde must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Heptachlor analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Heptachlor must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Crotoxyphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Crotoxyphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Mevinphos analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Mevinphos must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dioxathion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dioxathion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Carbophenothion analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbophenothion must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Toxaphene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Toxaphene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Demeton-O+S analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Demeton-O+S must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Azinphos methyl analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Azinphos methyl must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Carbazole analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Carbazole must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for Dinoseb analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. Dinoseb must be reported for this method and libary MatrixID
equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where
ReportableResult = YES for 1-Methylnaphthalene analyzed by EPA 8270 in MatrixID
'AQUEOUS-Groundwater'. 1-Methylnaphthalene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for 1-Chloronaphthalene analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1-Chloronaphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Benzidine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Benzidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for Endosulfan I analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Endosulfan I must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for 1,2-Dibromo-3-chloropropane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dibromo-3-chloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a record where ReportableResult = YES for N-Nitrosodiphenylamine plus Diphenylamine analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodiphenylamine plus Diphenylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where

ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where

ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and

libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where
ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a record where ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where

ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for
this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1

,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary

MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for n-Propylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for n-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 4-Chlorotoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Chlorotoluene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 1,4-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for n-Propylamine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. n-Propylamine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 1,3,5-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3,5-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for Bromobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 2-Chloroethylvinyl ether analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 2-Chloroethylvinyl ether must be reported for this method
and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for Pyridine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Pyridine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 1,2,4-Trichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for sec-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. sec-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 1,3-Dichloropropane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,3-Dichloropropane must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for cis-1,4-Dichloro-2-butene analyzed by EPA 8260 in
MatrixID 'AQUEOUS-Groundwater'. cis-1,4-Dichloro-2-butene must be reported for this
method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for cis-1,2-Dichloroethene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. cis-1,2-Dichloroethene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for Methyl-t-butyl ether analyzed by EPA 8260 in MatrixID

'AQUEOUS-Groundwater'. Methyl-t-butyl ether must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Xylene, m,p- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, m,p- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 1,3-Dichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 1,3-Dichloropropene- Total analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,3-Dichloropropene- Total must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 1,1-Dichloropropene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,1-Dichloropropene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 2,2-Dichloropropane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2,2-Dichloropropane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Hexachloroethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Hexachloroethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Bromochloromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Bromochloromethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Pentachloroethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 1,4-Dichloro-2-butene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,4-Dichloro-2-butene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 1,2,3-Trichlorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 1,2,3-Trichlorobenzene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Hexachlorobutadiene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Hexachlorobutadiene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Naphthalene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Naphthalene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for N-Nitrosodibutylamine analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. N-Nitrosodibutylamine must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for Xylene, o- analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. Xylene, o- must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where ReportableResult = YES for 2-Chlorotoluene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'. 2-Chlorotoluene must be reported for this method and libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where

ReportableResult = YES for 1,2-Dichlorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2-Dichlorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for o-Toluidine analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. o-Toluidine must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 1,2,4-Trimethylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,2,4-Trimethylbenzene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for t-Butylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. t-Butylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for Isopropylbenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Isopropylbenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for Nitrobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Nitrobenzene must be reported for this method and libary
MatrixID equivalent according to the project library.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a record where
ReportableResult = YES for 4-Isopropyltoluene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 4-Isopropyltoluene must be reported for this method and
libary MatrixID equivalent according to the project library.",,,,1
,"0012",,"Missing record",,"n-Propylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Butylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Chlorotoluene is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylamine is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3,5-Trimethylbenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bromobenzene is not reported as a spiked compound in the
LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chloroethylvinyl ether is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Pyridine is not reported as a spiked compound in the LCS
identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trichlorobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"sec-Butylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropane is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,4-Dichloro-2-butene is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,2-Dichloroethene is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Methyl-t-butyl ether is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, m,p- is not reported as a spiked compound in the
LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropene- Total is not reported as a spiked
compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"1,1-Dichloropropene is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,2-Dichloropropane is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachloroethane is not reported as a spiked compound in
the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bromochloromethane is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloroethane is not reported as a spiked compound
in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"1,4-Dichloro-2-butene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichlorobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobutadiene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodibutylamine is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, o- is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chlorotoluene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dichlorobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"o-Toluidine is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trimethylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"t-Butylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isopropylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nitrobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Isopropyltoluene is not reported as a spiked compound in the LCS identified as LabSampleID 070809LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and

matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Butylbenzene is not reported as a spiked compound in
the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Chlorotoluene is not reported as a spiked compound in
the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichlorobenzene is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"n-Propylamine is not reported as a spiked compound in
the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3,5-Trimethylbenzene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bromobenzene is not reported as a spiked compound in the
LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chloroethylvinyl ether is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Pyridine is not reported as a spiked compound in the
LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trichlorobenzene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"sec-Butylbenzene is not reported as a spiked compound in
the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropane is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"cis-1,4-Dichloro-2-butene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,2-Dichloroethene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Methyl-t-butyl ether is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this

method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, m,p- is not reported as a spiked compound in the
LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichlorobenzene is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropene- Total is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"1,1-Dichloropropene is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"2,2-Dichloropropane is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Hexachloroethane is not reported as a spiked compound in
the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bromochloromethane is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Pentachloroethane is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,4-Dichloro-2-butene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichlorobenzene is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobutadiene is not reported as a spiked compound
in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Naphthalene is not reported as a spiked compound in the
LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodibutylamine is not reported as a spiked
compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =

SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, o- is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chlorotoluene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"o-Toluidine is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trimethylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"t-Butylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isopropylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nitrobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Isopropyltoluene is not reported as a spiked compound in the LCSD identified as LabSampleID 070809LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Butylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Chlorotoluene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichlorobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylamine is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3,5-Trimethylbenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this

method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bromobenzene is not reported as a spiked compound in the
LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chloroethylvinyl ether is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Pyridine is not reported as a spiked compound in the LCS
identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trichlorobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"sec-Butylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropane is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,4-Dichloro-2-butene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,2-Dichloroethene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Methyl-t-butyl ether is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, m,p- is not reported as a spiked compound in the
LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropene- Total is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"1,1-Dichloropropene is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,2-Dichloropropane is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and

matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachloroethane is not reported as a spiked compound in
the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bromochloromethane is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloroethane is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichloro-2-butene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichlorobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobutadiene is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naphthalene is not reported as a spiked compound in the
LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodibutylamine is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, o- is not reported as a spiked compound in the
LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chlorotoluene is not reported as a spiked compound in
the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dichlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"o-Toluidine is not reported as a spiked compound in the
LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trimethylbenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"t-Butylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isopropylbenzene is not reported as a spiked compound in
the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same

matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nitrobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Isopropyltoluene is not reported as a spiked compound in the LCS identified as LabSampleID 070909LCSA32 for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Butylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Chlorotoluene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"n-Propylamine is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3,5-Trimethylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bromobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chloroethylvinyl ether is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pyridine is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"sec-Butylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"cis-1,4-Dichloro-2-butene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"cis-1,2-Dichloroethene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl-t-butyl ether is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, m,p- is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,3-Dichloropropene- Total is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,1-Dichloropropene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,2-Dichloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachloroethane is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bromochloromethane is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloroethane is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,4-Dichloro-2-butene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Hexachlorobutadiene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naphthalene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodibutylamine is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Xylene, o- is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2-Chlorotoluene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dichlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"o-Toluidine is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,4-Trimethylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"t-Butylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isopropylbenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nitrobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Isopropyltoluene is not reported as a spiked compound in the LCSD identified as LabSampleID 070909LCSA32D for Method EPA 8260 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 287909 for Method EPA 8011 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 287909 for Method EPA 8011 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 287910 for Method EPA 8011 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2,3-Trichloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 287910 for Method EPA 8011 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCS identified as LabSampleID 287985 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the LCS identified as LabSampleID 287985 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the LCS identified as LabSampleID 287985 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the LCS identified as LabSampleID 287985 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCS identified as LabSampleID 287985 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCSD identified as LabSampleID 287986 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the LCSD identified as LabSampleID 287986 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the LCSD identified as LabSampleID 287986 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the LCSD identified as LabSampleID 287986 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCSD identified as LabSampleID 287986 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Fenitrothion is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Simazine is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethoprop is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Merphos is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Atrazine is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos ethyl is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorpyrifos is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ronnel is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aspon is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichloronate is not reported as a spiked compound in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diazinon is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tokuthion is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bolstar is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in
the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichlorofon is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl chlorpyrifos is not reported as a spiked compound
in the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in
the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the
LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in
the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fonophos is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Stirofos is not reported as a spiked compound in the LCS
identified as LabSampleID 288035 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlofenthion is not reported as a spiked compound in
the LCS identified as LabSampleID 288035 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked
compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenitrothion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Simazine is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethoprop is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Merphos is not reported as a spiked compound in the LCSD
identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Atrazine is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos ethyl is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorpyrifos is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ronnel is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aspon is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichloronate is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diazinon is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tokuthion is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bolstar is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichlorofon is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl chlorpyrifos is not reported as a spiked compound in the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and

matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCSD
identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in
the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCSD
identified as LabSampleID 288036 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in
the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fonophos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Stirofos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlofenthion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288036 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked
compound in the LCS identified as LabSampleID 288038 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288038 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in
the LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked
compound in the LCS identified as LabSampleID 288038 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.

",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the
LCS identified as LabSampleID 288038 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 288038 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked
compound in the LCS identified as LabSampleID 288038 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCSD
identified as LabSampleID 288039 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound
in the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked
compound in the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCSD
identified as LabSampleID 288039 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCSD
identified as LabSampleID 288039 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCSD
identified as LabSampleID 288039 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in
the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked
compound in the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the
LCSD identified as LabSampleID 288039 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked
compound in the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked
compound in the LCSD identified as LabSampleID 288039 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound
in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound
in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound
in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked
compound in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent

as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in
the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in
the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in

the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCS
identified as LabSampleID 288040 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in
the LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked
compound in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the
LCS identified as LabSampleID 288040 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked
compound in the LCS identified as LabSampleID 288040 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"4-Nitrophenol is not reported as a spiked compound in

the LCS identified as LabSampleID 288046 for Method EPA 8151 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Picloram is not reported as a spiked compound in the LCS
identified as LabSampleID 288046 for Method EPA 8151 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachlorophenol is not reported as a spiked compound
in the LCS identified as LabSampleID 288046 for Method EPA 8151 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Nitrophenol is not reported as a spiked compound in
the LCSD identified as LabSampleID 288047 for Method EPA 8151 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Picloram is not reported as a spiked compound in the
LCSD identified as LabSampleID 288047 for Method EPA 8151 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachlorophenol is not reported as a spiked compound
in the LCSD identified as LabSampleID 288047 for Method EPA 8151 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound
in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound
in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked
compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"3-Methylphenol is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1260 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1254 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1221 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1232 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diphenylamine is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Diphenylhydrazine is not reported as a spiked
compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"PCB-1248 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1016 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-S is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"m,p-Cresols is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbofuran is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,3,7,8-TCDD is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetrachlorvinphos is not reported as a spiked compound
in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thionazin is not reported as a spiked compound in the

LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phorate is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Disulfoton is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl dithiopyrophosphate is not reported as a
spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bis(2-chloroisopropyl)ether is not reported as a spiked
compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Famphur is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent

as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1242 is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dimethoate is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbaryl is not reported as a spiked compound in the LCS
identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzoic acid is not reported as a spiked compound in the
LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorophene is not reported as a spiked compound in
the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCS

identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbazole is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dinoseb is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"1-Methylnaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Chloronaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzidine is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodiphenylamine plus Diphenylamine is not reported as a spiked compound in the LCS identified as LabSampleID 288083 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"3-Methylphenol is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1260 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1254 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1221 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1232 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diphenylamine is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Diphenylhydrazine is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1248 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1016 is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-S is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"m,p-Cresols is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbofuran is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,3,7,8-TCDD is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetrachlorvinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent

as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thionazin is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phorate is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Disulfoton is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl dithiopyrophosphate is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bis(2-chloroisopropyl)ether is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Famphur is not reported as a spiked compound in the LCSD
identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1242 is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCSD
identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCSD
identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dimethoate is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbaryl is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzoic acid is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorophene is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCSD
identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCSD
identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in
the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbazole is not reported as a spiked compound in the
LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dinoseb is not reported as a spiked compound in the LCSD

identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Methylnaphthalene is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Chloronaphthalene is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzidine is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodiphenylamine plus Diphenylamine is not reported as a spiked compound in the LCSD identified as LabSampleID 288084 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"3-Methylphenol is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1260 is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1254 is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1221 is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1232 is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in

the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diphenylamine is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Diphenylhydrazine is not reported as a spiked
compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"PCB-1248 is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1016 is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-S is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"m,p-Cresols is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbofuran is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,3,7,8-TCDD is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetrachlorvinphos is not reported as a spiked compound
in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thionazin is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phorate is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Disulfoton is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl dithiopyrophosphate is not reported as a
spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bis(2-chloroisopropyl)ether is not reported as a spiked
compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Famphur is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1242 is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dimethoate is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbaryl is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzoic acid is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorophene is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCS
identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in
the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbazole is not reported as a spiked compound in the
LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dinoseb is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Methylnaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Chloronaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzidine is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodiphenylamine plus Diphenylamine is not reported as a spiked compound in the LCS identified as LabSampleID 288122 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"3-Methylphenol is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1260 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1254 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1221 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1232 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diphenylamine is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Diphenylhydrazine is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1248 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1016 is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-S is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"m,p-Cresols is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbofuran is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,3,7,8-TCDD is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Tetrachlorvinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thionazin is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phorate is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Disulfoton is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl dithiopyrophosphate is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bis(2-chloroisopropyl)ether is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in

the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Famphur is not reported as a spiked compound in the LCSD
identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1242 is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in
the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCSD
identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCSD
identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dimethoate is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbaryl is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzoic acid is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorophene is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbazole is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dinoseb is not reported as a spiked compound in the LCSD
identified as LabSampleID 288123 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Methylnaphthalene is not reported as a spiked compound
in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Chloronaphthalene is not reported as a spiked compound
in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzidine is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the
LCSD identified as LabSampleID 288123 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked
compound in the LCSD identified as LabSampleID 288123 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"N-Nitrosodiphenylamine plus Diphenylamine is not
reported as a spiked compound in the LCSD identified as LabSampleID 288123 for
Method EPA 8270 with the same matrix equivalent as samples. This analyte must be
reported for this method and matrix according to the project library. Make sure
that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound
in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound
in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked
compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"3-Methylphenol is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1260 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1254 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1221 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to

the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1232 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diphenylamine is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Diphenylhydrazine is not reported as a spiked
compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"PCB-1248 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1016 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-S is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"m,p-Cresols is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbofuran is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"2,3,7,8-TCDD is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the

LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetrachlorvinphos is not reported as a spiked compound
in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thionazin is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phorate is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Disulfoton is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl dithiopyrophosphate is not reported as a
spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Bis(2-chloroisopropyl)ether is not reported as a spiked
compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the
same matrix equivalent as samples. This analyte must be reported for this method

and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Famphur is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"PCB-1242 is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in
the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dimethoate is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbaryl is not reported as a spiked compound in the LCS
identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzoic acid is not reported as a spiked compound in the

LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorophene is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbazole is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dinoseb is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Methylnaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1-Chloronaphthalene is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Benzidine is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"N-Nitrosodiphenylamine plus Diphenylamine is not reported as a spiked compound in the LCS identified as LabSampleID 288183 for Method EPA 8270 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Zinc is not reported as a spiked compound in the LCS identified as LabSampleID 288245 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aluminum is not reported as a spiked compound in the LCS

identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Iron is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lead is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Magnesium is not reported as a spiked compound in the
LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Manganese is not reported as a spiked compound in the
LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the
LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nickel is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Potassium is not reported as a spiked compound in the
LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Silver is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Sodium is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the
LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Thallium is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tin is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Antimony is not reported as a spiked compound in the LCS
identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Arsenic is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Barium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Beryllium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cadmium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chromium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cobalt is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Copper is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Vanadium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Calcium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Selenium is not reported as a spiked compound in the LCS identified as LabSampleID 288245R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the LCSD identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCSD identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to

the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Zinc is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246 for Method EPA 6010 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aluminum is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Iron is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lead is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Lithium is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Magnesium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Manganese is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Molybdenum is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Nickel is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Potassium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Silver is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Sodium is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Strontium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the
LCS identified as LabSampleID 9070485-BS2 for Method SM18 4500-NH3 C (Ammonia (N))
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the
LCS identified as LabSampleID 9070485-SRM1 for Method SM18 4500-NH3 C (Ammonia (N))
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1

,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in the LCS identified as LabSampleID 9070778-BS1 for Method EPA 9012 (CN - Total) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in the LCS identified as LabSampleID 9070778-BS2 for Method EPA 9012 (CN - Total) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in the LCS identified as LabSampleID 9070778-SRM1 for Method EPA 9012 (CN - Total) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the LCS identified as LabSampleID 9070899-BS1 for Method SM18 4500-NH3 C (Ammonia (N)) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the LCS identified as LabSampleID 9070899-BS2 for Method SM18 4500-NH3 C (Ammonia (N)) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the LCS identified as LabSampleID 9070899-SRM1 for Method SM18 4500-NH3 C (Ammonia (N)) with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0018",,"Extra record(s)",,"LCS and/or LCSD reported for Method SM18 2320 B and Matrix AQUEOUS-Groundwater but the project library does not require a LCS for this method and matrix equivalent.",,,,1
,"0018",,"Extra record(s)",,"LCS and/or LCSD reported for Method SM18 2540 C and Matrix AQUEOUS-Groundwater but the project library does not require a LCS for this method and matrix equivalent.",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070209 for method EPA 300.0 (Chloride). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070609 for method EPA 300.0 (Chloride). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070709 for method EPA 300.0 (Chloride). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070909 for method EPA 300.0 (Chloride). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070209 for method EPA 300.0 (Sulfate). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070609 for method EPA 300.0 (Sulfate). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070809 for method EPA 300.0 (Sulfate). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070209 for method EPA 300.1. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch IC2070809 for method EPA 300.1. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch 9070474 for method EPA 335.4. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch 070809A for method EPA 6010. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch 071409A for method EPA 6010. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to Water are reported in Method Batch 428371 for method EPA 7470. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to

Water are reported in Method Batch 428409 for method EPA 7470. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E4070609 for method EPA 8011. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E4071509 for method EPA 8011. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E2070709SW for method EPA 8081. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E2071509 for method EPA 8081. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch N1070709 for method EPA 8141. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch N1071409 for method EPA 8141. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E2070809 for method EPA 8151. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch E2071409 for method EPA 8151. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch M3070809 for method EPA 8260. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch M3070909 for method EPA 8260. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch S3070809 for method EPA 8270. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch S3070909 for method EPA 8270. ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch 9070474 for method EPA 9012 (CN - Total). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch 9070778 for method EPA 9012 (CN - Total). ",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch 9070485 for method SM18 4500-NH3 C (Ammonia (N)).
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike results with a matrix equivalent to
Water are reported in Method Batch 9070899 for method SM18 4500-NH3 C (Ammonia (N)).
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E4070609 for method EPA 8011.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E4071509 for method EPA 8011.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E2070709SW for method EPA 8081.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E2071509 for method EPA 8081.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch N1070709 for method EPA 8141.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch N1071409 for method EPA 8141.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E2070809 for method EPA 8151.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch E2071409 for method EPA 8151.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch M3070809 for method EPA 8260.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix

equivalent to Water are reported in Method Batch M3070909 for method EPA 8260.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch S3070809 for method EPA 8270.
",,,,1
,"0013",,"Missing record(s)",,"No Matrix Spike Duplicate results with a matrix
equivalent to Water are reported in Method Batch S3070909 for method EPA 8270.
",,,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Chloride) and matrix equivalent to Water in
MethodBatchID IC2070209. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Chloride) and matrix equivalent to Water in
MethodBatchID IC2070609. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Chloride) and matrix equivalent to Water in
MethodBatchID IC2070709. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Chloride) and matrix equivalent to Water in
MethodBatchID IC2070909. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Sulfate) and matrix equivalent to Water in
MethodBatchID IC2070209. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Sulfate) and matrix equivalent to Water in
MethodBatchID IC2070609. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.0 (Sulfate) and matrix equivalent to Water in
MethodBatchID IC2070809. ","PEL SOP does not include the analysis of a lab sample
duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.1 and matrix equivalent to Water in MethodBatchID
IC2070209. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 300.1 and matrix equivalent to Water in MethodBatchID
IC2070809. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 335.4 and matrix equivalent to Water in MethodBatchID
9070474. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 6010 and matrix equivalent to Water in MethodBatchID
070809A. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 6010 and matrix equivalent to Water in MethodBatchID
071409A. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 7470 and matrix equivalent to Water in MethodBatchID 428371.
","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 7470 and matrix equivalent to Water in MethodBatchID 428409.
","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 8011 and matrix equivalent to Water in MethodBatchID
E4070609. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are
reported for method EPA 8011 and matrix equivalent to Water in MethodBatchID
E4071509. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1

,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8081 and matrix equivalent to Water in MethodBatchID E2070709SW. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8081 and matrix equivalent to Water in MethodBatchID E2071509. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8141 and matrix equivalent to Water in MethodBatchID N1070709. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8141 and matrix equivalent to Water in MethodBatchID N1071409. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8151 and matrix equivalent to Water in MethodBatchID E2070809. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8151 and matrix equivalent to Water in MethodBatchID E2071409. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8260 and matrix equivalent to Water in MethodBatchID M3070809. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8260 and matrix equivalent to Water in MethodBatchID M3070909. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8270 and matrix equivalent to Water in MethodBatchID S3070809. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 8270 and matrix equivalent to Water in MethodBatchID S3070909. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 9012 (CN - Total) and matrix equivalent to Water in MethodBatchID 9070474. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method EPA 9012 (CN - Total) and matrix equivalent to Water in MethodBatchID 9070778. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method SM18 4500-NH3 C (Ammonia (N)) and matrix equivalent to Water in MethodBatchID 9070485. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0014",,"Missing record(s)",,"No Lab Duplicate results (where QCType = DUP) are reported for method SM18 4500-NH3 C (Ammonia (N)) and matrix equivalent to Water in MethodBatchID 9070899. ","PEL SOP does not include the analysis of a lab sample duplicate.",,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for 1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. 1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. Pentachloroethane must be reported for this method and libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for 2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. 2-Bromo-1-chloropropane must be reported for this method and libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for 4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'. 4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for

Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project
library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-15 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
Pentachloroethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
2-Bromo-1-chloropropane must be reported for this method and libary MatrixID

equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project
library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-16 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for

Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
Pentachloroethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
2-Bromo-1-chloropropane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project
library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-17 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the

project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
Pentachloroethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
2-Bromo-1-chloropropane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project
library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for

Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-18 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
Pentachloroethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
2-Bromo-1-chloropropane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project

library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-19 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
1,1,1,2-Tetrachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
1,1,1,2-Tetrachloroethane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Pentachloroethane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
Pentachloroethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
2-Bromo-1-chloropropane analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
2-Bromo-1-chloropropane must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
4-Bromofluorobenzene analyzed by EPA 8011 in MatrixID 'AQUEOUS-Groundwater'.
4-Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Decachlorobiphenyl analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
Decachlorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
2,4,5,6-Tetrachloro-m-xylene analyzed by EPA 8081 in MatrixID 'AQUEOUS-Groundwater'.
 2,4,5,6-Tetrachloro-m-xylene must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
4-Chloro-3-nitrobenzo-trifluoride analyzed by EPA 8141 in MatrixID
'AQUEOUS-Groundwater'. 4-Chloro-3-nitrobenzo-trifluoride must be reported for this
method and libary MatrixID equivalent according to the project library surrogate
table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Tributyl phosphate analyzed by EPA 8141 in MatrixID 'AQUEOUS-Groundwater'. Tributyl
phosphate must be reported for this method and libary MatrixID equivalent according
to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for

Page 208

1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Phenol-d6 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Phenol-d6 must be
reported for this method and libary MatrixID equivalent according to the project
library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Terphenyl-d14 analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. Terphenyl-d14
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
Dibromofluoromethane analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID MW-20 is missing a surrogate record for
p-Terphenyl analyzed by EPA 8270 in MatrixID 'AQUEOUS-Groundwater'. p-Terphenyl
must be reported for this method and libary MatrixID equivalent according to the
project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a surrogate record
for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a surrogate record
for Bromofluorobenzene analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Bromofluorobenzene must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0012",,"Missing record",,"Thallium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tin is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Titanium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Antimony is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Arsenic is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Barium is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Beryllium is not reported as a spiked compound in the
LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Boron is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cadmium is not reported as a spiked compound in the LCSD
identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Chromium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cobalt is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Copper is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Vanadium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Calcium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Selenium is not reported as a spiked compound in the LCSD identified as LabSampleID 288246R1 for Method EPA 6010 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Nitrophenol is not reported as a spiked compound in the LCS identified as LabSampleID 288266 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Picloram is not reported as a spiked compound in the LCS identified as LabSampleID 288266 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachlorophenol is not reported as a spiked compound in the LCS identified as LabSampleID 288266 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4-Nitrophenol is not reported as a spiked compound in the LCSD identified as LabSampleID 288267 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Picloram is not reported as a spiked compound in the LCSD identified as LabSampleID 288267 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachlorophenol is not reported as a spiked compound in the LCSD identified as LabSampleID 288267 for Method EPA 8151 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method

and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCS
identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the LCS
identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the LCS
identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in
the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked
compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the
LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.

",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 288398 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Toxaphene is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked compound in the LCSD identified as LabSampleID 288399 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor epoxide is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan sulfate is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicofol is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorobenzene is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane- Technical is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Kepone is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Alachlor is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Propachlor is not reported as a spiked compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diallate is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mirex is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl parathion is not reported as a spiked compound in
the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"b-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"d-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endosulfan II is not reported as a spiked compound in
the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Isodrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDT is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorobenzilate is not reported as a spiked compound in
the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"a-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin ketone is not reported as a spiked compound in
the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Parathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlordane is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix

equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"g-BHC is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dieldrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methoxychlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDD is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"4,4'-DDE is not reported as a spiked compound in the LCS
identified as LabSampleID 288400 for Method EPA 8081 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Endrin aldehyde is not reported as a spiked compound in
the LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Heptachlor is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Hexachlorocyclopentadiene is not reported as a spiked
compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Pentachloronitrobenzene is not reported as a spiked
compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Endosulfan I is not reported as a spiked compound in the
LCS identified as LabSampleID 288400 for Method EPA 8081 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-Chloropropane is not reported as a spiked
compound in the LCS identified as LabSampleID 288400 for Method EPA 8081 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked
compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the

LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenitrothion is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Simazine is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethoprop is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Merphos is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Atrazine is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos ethyl is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorpyrifos is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ronnel is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aspon is not reported as a spiked compound in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Trichloronate is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diazinon is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tokuthion is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bolstar is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichlorofon is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Methyl chlorpyrifos is not reported as a spiked compound
in the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix

according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the
LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fonophos is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Stirofos is not reported as a spiked compound in the LCS
identified as LabSampleID 288405 for Method EPA 8141 with the same matrix equivalent
as samples. This analyte must be reported for this method and matrix according to
the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlofenthion is not reported as a spiked compound in
the LCS identified as LabSampleID 288405 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tetraethyl pyrophosphate is not reported as a spiked
compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Fensulfothion is not reported as a spiked compound in
the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Malathion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenitrothion is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Simazine is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Terbufos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosphamidon is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethoprop is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dicrotophos is not reported as a spiked compound in the
LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix
equivalent as samples. This analyte must be reported for this method and matrix
according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Merphos is not reported as a spiked compound in the LCSD

identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Atrazine is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"EPN is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Leptophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos ethyl is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorpyrifos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ronnel is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Naled is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Aspon is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichloronate is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Diazinon is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Tokuthion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Bolstar is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Chlorfenvinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Trichlorofon is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fenthion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1

,"0012",,"Missing record",,"Methyl chlorpyrifos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Ethion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Coumaphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlorvos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Monocrotophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Phosmet is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Crotoxyphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Mevinphos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dioxathion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Carbophenothion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Demeton-O+S is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Azinphos methyl is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Fonophos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Stirofos is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Dichlofenthion is not reported as a spiked compound in the LCSD identified as LabSampleID 288406 for Method EPA 8141 with the same matrix equivalent as samples. This analyte must be reported for this method and matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked compound in the LCS identified as LabSampleID 288588 for Method EPA 8011 with the same matrix equivalent as samples. This analyte must be reported for this method

and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,2,3-Trichloropropane is not reported as a spiked
compound in the LCS identified as LabSampleID 288588 for Method EPA 8011 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,2-Dibromo-3-chloropropane is not reported as a spiked
compound in the LCSD identified as LabSampleID 288589 for Method EPA 8011 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"1,2,3-Trichloropropane is not reported as a spiked
compound in the LCSD identified as LabSampleID 288589 for Method EPA 8011 with the
same matrix equivalent as samples. This analyte must be reported for this method
and matrix according to the project library. Make sure that AnalyteType = SPK.
",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-BS1 for Method EPA 335.4 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-BS1 for Method EPA 9012 (CN - Total) with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-BS2 for Method EPA 335.4 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-BS2 for Method EPA 9012 (CN - Total) with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-SRM1 for Method EPA 335.4 with the same
matrix equivalent as samples. This analyte must be reported for this method and
matrix according to the project library. Make sure that AnalyteType = SPK. ",,,,1
,"0012",,"Missing record",,"Cyanide- Total is not reported as a spiked compound in
the LCS identified as LabSampleID 9070474-SRM1 for Method EPA 9012 (CN - Total) with
the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0012",,"Missing record",,"Ammonia (N) is not reported as a spiked compound in the
LCS identified as LabSampleID 9070485-BS1 for Method SM18 4500-NH3 C (Ammonia (N))
with the same matrix equivalent as samples. This analyte must be reported for this
method and matrix according to the project library. Make sure that AnalyteType =
SPK. ",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a surrogate record
for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a surrogate record
for Dibromofluoromethane analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Dibromofluoromethane must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID TRIP BLANK is missing a surrogate record
for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
1,4-Dichlorobenzene-d4 must be reported for this method and libary MatrixID
equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a surrogate
record for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.

2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a surrogate
record for Bromofluorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromofluorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a surrogate
record for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a surrogate
record for Dibromofluoromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Dibromofluoromethane must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1 is missing a surrogate
record for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene-d4 must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a surrogate
record for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a surrogate
record for Bromofluorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromofluorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a surrogate
record for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a surrogate
record for Dibromofluoromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Dibromofluoromethane must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-1A is missing a surrogate
record for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene-d4 must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a surrogate
record for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a surrogate
record for Bromofluorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromofluorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a surrogate
record for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a surrogate
record for Dibromofluoromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Dibromofluoromethane must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2 is missing a surrogate
record for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene-d4 must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a surrogate
record for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a surrogate

record for Bromofluorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromofluorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a surrogate
record for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a surrogate
record for Dibromofluoromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Dibromofluoromethane must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-2A is missing a surrogate
record for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene-d4 must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a surrogate
record for 2-Fluorobiphenyl analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
2-Fluorobiphenyl must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a surrogate
record for Bromofluorobenzene analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Bromofluorobenzene must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a surrogate
record for Nitrobenzene-d5 analyzed by EPA 8260 in MatrixID 'AQUEOUS-Groundwater'.
Nitrobenzene-d5 must be reported for this method and libary MatrixID equivalent
according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a surrogate
record for Dibromofluoromethane analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. Dibromofluoromethane must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1
,"0010",,"Missing record",,"ClientSampleID Trip Blank-3 is missing a surrogate
record for 1,4-Dichlorobenzene-d4 analyzed by EPA 8260 in MatrixID
'AQUEOUS-Groundwater'. 1,4-Dichlorobenzene-d4 must be reported for this method and
libary MatrixID equivalent according to the project library surrogate table.",,,,1