

## **Knauss, Elizabeth**

---

**From:** BOSEK, JOHN E <bosek.john@cleanharbors.com>  
**Sent:** Thursday, March 23, 2017 12:28 PM  
**To:** garcia.javier@epa.gov; Knauss, Elizabeth  
**Cc:** Desha, David A.  
**Subject:** Calibration logs and picture of reactivities room  
**Attachments:** 0707\_001.pdf; 20170323\_083433.jpg

Good afternoon Javier –

Once again it was nice meeting you yesterday and I passed along your regards to Jim Childress. Here are copies of the 2016 and 2017 MSA 4 gas meter calibration logs and a picture of the corrected reactive room issue as requested. I'll have the insurance and closure updates to you in just a little bit as well.

### ***Safety Starts with Me: Live It 3-6-5***

---

#### **John Bosek**

Facility General Manager  
Clean Harbors Florida LLC

7001 Kilo Avenue

Bartow, FL 33830

(o) 863.519.6331

(c) 863.559.1610

(f) 863.519.6361

[bosek.john@cleanharbors.com](mailto:bosek.john@cleanharbors.com)

[www.cleanharbors.com](http://www.cleanharbors.com)



**Clean Harbors Florida LLC**  
**MSA Multigas Detector**  
**2016 Calibration Log**

DATE	BATTERY CHECK	Calibrated Using Cal Gas	ZEROED	INITIAL BACKGROUND READINGS					Operator Name
				LEL	VOC	O2	H2S	CO	
01/21/16	/	/	/	∅	∅	20.8	∅	∅	MB
02/23/16	/	/	/	∅	∅	20.8	∅	∅	MB
3/22/16	/	/	/	∅	∅	20.8	∅	∅	MB
4/26/16	/	/	/	∅	∅	20.8	∅	∅	MB
5/27/16	/	/	/	∅	∅	20.8	∅	∅	MB
6/24/16	/	/	/	∅	∅	20.8	∅	∅	MB
7/25/16	/	/	/	∅	∅	20.8	∅	∅	MB
8/22/16	/	/	/	∅	∅	20.8	∅	∅	MB
9/26/16	/	/	/	∅	∅	20.8	∅	∅	MB
10/24/16	/	/	/	∅	∅	20.8	∅	∅	MB
11/28/16	/	/	/	∅	∅	20.8	∅	∅	MB
12/28/16	/	/	/	∅	∅	20.8	∅	∅	MB

**CALIBRATION GAS**

- 1. ISOBUTYLENE - LOT # 100978 EXP: N/A
- 2. H2S - LOT# 968933 EXP: N/A

MODEL MSA SISRIUS - SERIAL NUMBER A3-2713

**Clean Harbors Florida LLC**

**MSA Multigas Detector**

**2017 Calibration Log**

DATE	BATTERY CHECK	Calibrated Using Cal Gas	ZEROED	INITIAL BACKGROUND READINGS					Operator Name
				LEL	VOC	O2	H2S	CO	
1/31/17	✓	✓	✓	∅	∅	20.8	∅	∅	Mike B.
2/23/17	✓	✓	✓	∅	∅	20.8	∅	∅	Mike B.
3/20/17	✓	✓	✓	∅	∅	20.8	∅	∅	Mike B.

CALIBRATION GAS

- 1. ISOBUTYLENE – LOT # 100973      EXP: N/A
- 2. H2S – LOT# 968938                      EXP: N/A

MODEL MSA SISRIUS – SERIAL NUMBER A3-2713

