



FDEP Automated Data Evaluation

Southeast Landfill

Human Review Statement:

This is an automated data review produced by the FDEP Validator software. This report does not supercede FDEP rules or standards and may not contain all violations of FDEP rules and standards. This report should be reviewed by a qualified staff member.

Sample ID: **660-25638-1**

Matrix: W

Well Type: OT - Other

Monitoring Site: SLTRF POND

Sample Date: October 8, 2008

Evaluation Type: **QA/QC qualifier has been applied**

Review qualifier meaning and assess impact on qualified data

Parameter Name	Rslt	Qual	Units	QA/QC Qualifier	Qual meaning:
Di-n-butyl phthalate	1.5	JU	ug/L	J	Estimated value; value not accurate.
Comment: Accuracy limits exceeded or Precision limits exceeded					
Mercury	0.072	JU	ug/L	J	Estimated value; value not accurate.
Comment:					
Tetrachloroethylene	0.5	JU	ug/L	J	Estimated value; value not accurate.
Comment:					

Versions

Report:	3.0.0
Database:	3.7.51 6/24/2005 12:00

Wednesday, February 04, 2009

Page 1 of 5

FDEP Automated Data Evaluation

Southeast Landfill

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **660-25638-2**

Matrix: W

Well Type: OT - Other

Monitoring Site: SELF TANK TRUCK

Sample Date: October 8, 2008

Evaluation Type: **QA/QC qualifier has been applied**

Review qualifier meaning and assess impact on qualified data

Parameter Name	Rslt	Qual	Units
----------------	------	------	-------

Di-n-butyl phthalate	1.5	JU	ug/L
----------------------	-----	----	------

QA/QC Qualifier J

Comment: Accuracy limits exceeded or Precision limits exceeded

Qual meaning: Estimated value; value not accurate.

Tetrachloroethylene	0.5	JU	ug/L
---------------------	-----	----	------

QA/QC Qualifier J

Comment:

Qual meaning: Estimated value; value not accurate.

Versions	
Report:	3.0.0
Database:	3.7.51 6/24/2005 12:00

Wednesday, February 04, 2009

Page 2 of 5

FDEP Automated Data Evaluation

Southeast Landfill

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **660-25641-1**

Matrix: W

Well Type: OT - Other

Monitoring Site: Not Blank-DUP

Sample Date: October 8, 2008

Evaluation Type: **QA/QC qualifier has been applied**

Review qualifier meaning and assess impact on qualified data

Parameter Name	Rslt	Qual	Units
----------------	------	------	-------

Di-n-butyl phthalate	1.5	JU	ug/L
----------------------	-----	----	------

QA/QC Qualifier J

Comment: Accuracy limits exceeded or Precision limits exceeded

Qual meaning: Estimated value; value not accurate.

Tetrachloroethylene	0.5	JU	ug/L
---------------------	-----	----	------

QA/QC Qualifier J

Comment:

Qual meaning: Estimated value; value not accurate.

Versions	
Report:	3.0.0
Database:	3.7.51 6/24/2005 12:00

Wednesday, February 04, 2009

Page 3 of 5

FDEP Automated Data Evaluation

Southeast Landfill

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **660-25641-2**

Matrix: W

Well Type:

Monitoring Site: Equipment Blank

Sample Date: October 8, 2008

Evaluation Type: **QA/QC qualifier has been applied**

Review qualifier meaning and assess impact on qualified data

Parameter Name	Rslt	Qual	Units
----------------	------	------	-------

Di-n-butyl phthalate	1.5	JU	ug/L
----------------------	-----	----	------

QA/QC Qualifier J

Comment: Accuracy limits exceeded or Precision limits exceeded

Qual meaning: Estimated value; value not accurate.

Tetrachloroethylene	0.5	JU	ug/L
---------------------	-----	----	------

QA/QC Qualifier J

Comment:

Qual meaning: Estimated value; value not accurate.

Versions	
Report:	3.0.0
Database:	3.7.51 6/24/2005 12:00

Wednesday, February 04, 2009

Page 4 of 5

FDEP Automated Data Evaluation

Southeast Landfill

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **660-25641-3**

Matrix: W

Well Type:

Monitoring Site: Trip Blank

Sample Date: October 8, 2008

Evaluation Type: **QA/QC qualifier has been applied**

Review qualifier meaning and assess impact on qualified data

Parameter Name	Rslt	Qual	Units
Tetrachloroethylene	0.5	JU	ug/L

QA/QC Qualifier J

Comment: Accuracy limits exceeded or Precision limits exceeded

Qual meaning: Estimated value; value not accurate.

Versions
Report: 3.0.0
Database: 3.7.51 6/24/2005 12:00

Wednesday, February 04, 2009

Page 5 of 5