



# FDEP Automated Data Evaluation

WEST PASCO CLASS III LF

## Human Review Statement:

This is an automated data review produced by the FDEP Validator software. This report does not supercede FDEP rules or standards and may not contain all violations of FDEP rules and standards. This report should be reviewed by a qualified staff member.

Sample ID: **19670 08/28/2008**

Matrix: W

Well Type: BG - Background

Monitoring Site: 4MW-3A

Sample Date: August 28, 2008

---

Evaluation Type: **The sample was analyzed after the holding time had expired.**

Measured concentrations may not be representative of conditions at sample collection.

Parameter Name	Rslt	Qual	Units	
CHLORIDE	18.6		MG/L	Analysis performed outside 4.9 days. holding time by:

Comment: MANUALLY ENTERED

Versions
Report: 3.0.0
Database: 3.7.51 6/24/2005 12:00

Wednesday, March 04, 2009

Page 1 of 3

# FDEP Automated Data Evaluation

WEST PASCO CLASS III LF

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **2340 08/28/2008**

Matrix: W

Well Type: CO - Compliance

Monitoring Site: 4MW-7

Sample Date: August 28, 2008

---

Evaluation Type: **The sample was analyzed after the holding time had expired.**

Measured concentrations may not be representative of conditions at sample collection.

Parameter Name	Rslt	Qual	Units	
CHLORIDE	21.7		MG/L	Analysis performed outside holding time by: 5.0 days.
Comment: MANUALLY ENTERED				
NITRATE (AS N)	0.38		MG/l	Analysis performed outside holding time by: 7.8 days.
Comment:				

Versions
Report: 3.0.0
Database: 3.7.51 6/24/2005 12:00

Wednesday, March 04, 2009

Page 2 of 3

# FDEP Automated Data Evaluation

WEST PASCO CLASS III LF

This is an automated data review. Please see the Human Review Statement on page 1 of this report.

Sample ID: **2342 08/27/2008**

Matrix: W

Well Type: CO - Compliance

Monitoring Site: 4MW-9

Sample Date: August 27, 2008

---

Evaluation Type: **The sample was analyzed after the holding time had expired.**

Measured concentrations may not be representative of conditions at sample collection.

Parameter Name	Rslt	Qual	Units	
BROMOFORM	0.1	U	UG/L	Analysis performed outside holding time by: 1 days.
Comment: MANUALLY ENTERED				
CHLORIDE	32.3		MG/L	Analysis performed outside holding time by: 5.8 days.
Comment:				

Versions
Report: 3.0.0
Database: 3.7.51 6/24/2005 12:00

Wednesday, March 04, 2009

Page 3 of 3